

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Direct acting indicating analogue electrical measuring instruments and their accessories –

Part 9: Recommended test methods

Appareils de mesure électriques indicateurs analogiques à action directe et leurs accessoires –

Partie 9: Méthodes d'essai recommandées

IECNORM.COM : Click to view the full PDF of IEC 60051-9:2019



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2019 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Recherche de publications IEC -

webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

Electropedia - www.electropedia.org

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 000 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Glossaire IEC - std.iec.ch/glossary

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Direct acting indicating analogue electrical measuring instruments and their accessories –

Part 9: Recommended test methods

Appareils de mesure électriques indicateurs analogiques à action directe et leurs accessoires –

Partie 9: Méthodes d'essai recommandées

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 17.220.20

ISBN 978-2-8322-6495-9

Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.

CONTENTS

FOREWORD	3
INTRODUCTION	5
1 Scope	6
2 Normative references	6
3 Terms and definitions	6
4 Requirements for test equipment	7
4.1 General	7
4.2 Test equipment for measuring the intrinsic uncertainty of an instrument and the variations	7
4.3 Test equipment for overload tests, self-heating, damping, overshoot and response time	7
4.4 Test equipment for phase shift	7
4.5 Test equipment for temperatures	7
4.6 Test equipment for humidity	7
5 Test methods for determination of intrinsic uncertainty	7
6 Test methods for damping	8
7 Test methods for humidity	8
8 Tests with distorted AC quantities	8
9 Test methods for shock and vibration	8
10 Test method for external magnetic field	8
11 Test method for external electric field	8
12 Test methods for electrical safety	8
13 Test methods for electromagnetic compatibility (EMC)	9
14 Test methods for recurrent tests	9
15 Test methods for nonconformity	9
Bibliography	10

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**DIRECT ACTING INDICATING ANALOGUE ELECTRICAL
MEASURING INSTRUMENTS AND THEIR ACCESSORIES –****Part 9: Recommended test methods**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60051-9 has been prepared by IEC technical committee 85: Measuring equipment for electrical and electromagnetic quantities.

This fifth edition cancels and replaces the fourth edition published in 1988, Amendment 1:1994 and Amendment 2:1995. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) adding performance requirements for test equipment;
- b) updating the references to the applicable standards for test methods.

The text of this International Standard is based on the following documents:

FDIS	Report on voting
85/663/FDIS	85/672/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This International Standard is to be used in conjunction with IEC 60051-1:2016.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts in the IEC 60051 series, published under the general title *Direct acting indicating analogue electrical measuring instruments and their accessories*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IECNORM.COM : Click to view the full PDF of IEC 60051-9:2019

INTRODUCTION

IEC 60051 is published in separate parts according to the following structure and under the general title *Direct acting indicating analogue electrical measuring instruments and their accessories*.

Part 1: Definitions and general requirements common to all parts

Part 2: Special requirements for ammeters and voltmeters

Part 3: Special requirements for wattmeters and varmeters

Part 4: Special requirements for frequency meters

Part 5: Special requirements for phase meters, power factor meters and synchroscopes

Part 6: Special requirements for ohmmeters (impedance meters) and conductance meters

Part 7: Special requirements for multi-function instruments

Part 8: Special requirements for accessories

Part 9: Recommended test methods

IEC 60051-9 is not complete in itself and is read in conjunction with IEC 60051-1.

IECNORM.COM : Click to view the full PDF of IEC 60051-9:2019

DIRECT ACTING INDICATING ANALOGUE ELECTRICAL MEASURING INSTRUMENTS AND THEIR ACCESSORIES –

Part 9: Recommended test methods

1 Scope

This part of IEC 60051 applies to direct acting indicating analogue electrical measuring instruments and their accessories and gives guidance for applicable test methods and for the performance of test equipment.

This document does not apply to:

- special purpose instruments that are covered by their own IEC International Standards;
- special purpose devices that are covered by their own IEC International Standards when they are used as accessories.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60051-1:2016, *Direct acting indicating analogue electrical measuring instruments and their accessories – Part 1: Definitions and general requirements common to all parts*

IEC 60068-2-6, *Environmental testing – Part 2-6: Tests – Test Fc: Vibration (sinusoidal)*

IEC 61000-4-8, *Electromagnetic compatibility (EMC) – Part 4-8: Testing and measurement techniques – Power frequency magnetic field immunity test*

IEC 61010-1:2010, *Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 1: General requirements*

IEC 61010-2-030, *Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 2-030: Particular requirements for equipment having testing or measuring circuits*

IEC 61326-1:2012, *Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 1: General requirements*

IEC 61326-2-1, *Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 2-1: Particular requirements – Test configurations, operational conditions and performance criteria for sensitive test and measurement equipment for EMC unprotected applications*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60051-1:2016 apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

4 Requirements for test equipment

4.1 General

Test equipment shall have sufficient accuracy and stability for the determination of the intrinsic uncertainty of an instrument and the variations.

NOTE Calibration protocols can be used.

4.2 Test equipment for measuring the intrinsic uncertainty of an instrument and the variations

The maximum instrumental uncertainty of reference instruments used to determine the intrinsic uncertainty of an instrument and the variations shall be 0,2 times the class index of the instrument under test. In the event that a calibration protocol is used, the maximum calibration uncertainty shall be 0,2 times the class index of the equipment under test (EUT).

4.3 Test equipment for overload tests, self-heating, damping, overshoot and response time

The maximum uncertainty of the settings of the electrical quantities that will be used to test overload, self-heating, damping, overshoot and response time shall be $\pm 1\%$ of those settings.

4.4 Test equipment for phase shift

Test equipment for settings of the phase shift of power meters with class indexes of 0,5 or higher shall have an uncertainty of $\pm 1^\circ$ or better.

Test equipment for settings of the phase shift of power meters with class indexes below 0,5 shall have a maximum uncertainty of $\pm 0,1^\circ$ or better. The instrumental uncertainty of test equipment for phase meters and power factor meters shall comply with 4.2.

4.5 Test equipment for temperatures

Equipment for the setting of temperatures shall have a maximum uncertainty in the temperature setting of $\pm 1^\circ\text{C}$ in the case of EUT with a class index equal to or below 0,3. In the case of EUT with a class index above 0,3, the maximum uncertainty of the setting shall be $\pm 2^\circ\text{C}$.

4.6 Test equipment for humidity

Equipment for settings of humidity shall have a maximum uncertainty in the humidity settings of $\pm 3\%$

5 Test methods for determination of intrinsic uncertainty

The intrinsic uncertainty at defined test points shall be checked with slowly increasing and decreasing measuring values without any overshoot and without tapping. The test points for the type test can be defined by the manufacturer. During the routine tests, tests should be carried out at full scale at all ranges and the characteristics should be tested in one range at a minimum.

Tapping is only allowed for the determination of scale marks and for the determination of variations.

6 Test methods for damping

Damping shall be tested at 66 % of the scale length.

7 Test methods for humidity

The settings of the relative humidity in relation to temperatures shall be in accordance with the relevant requirements of IEC 60051-1. The minimum test duration for the determination of the variation shall be at least 48 h.

8 Tests with distorted AC quantities

Peak factors deviating from 1,414 1 should be simulated by a combination of the 3rd, 5th and 7th harmonic. Distortions should be simulated with the 7th harmonic.

9 Test methods for shock and vibration

Tests for mechanical shock shall be in accordance with the requirements for mechanical strength of IEC 61010-1.

The test method for vibration shall be in accordance with the requirements of IEC 60068-2-6.

Sweep frequency range: 10 Hz to 55 Hz to 10 Hz

Amplitude: 0,15 mm

Number of sweep cycles: 5

10 Test method for external magnetic field

For tests in accordance with IEC 60051-1:2016, 5.3 and Table 3, the test configuration shall comply with IEC 61000-4-8 but with a test level of 400 A/m.

11 Test method for external electric field

Electric fields should be simulated by two metallic plates placed in parallel at a distance of 1 m. The EUT shall be placed in the centre, between the plates and shall be turned around to find the worst-case position. Test points may be specified by the manufacturer. The cubature of this configuration should be at a minimum ten times higher than the cubature of the EUT to keep the field homogeneous.

12 Test methods for electrical safety

The test methods shall comply with IEC 61010-1 and IEC 61010-2-030.

Drop tests in accordance with IEC 61010-1:2010, 8.3, shall not be applied to products intended to be permanently installed in panels.

13 Test methods for electromagnetic compatibility (EMC)

The test methods shall comply with IEC 61326-1. The emission limits shall be in accordance with IEC 61326-1:2012, Table 2, and with IEC 61326-2-1. In the case of emission tests, the EUT shall comply with Class A in IEC 61326-1.

14 Test methods for recurrent tests

Time intervals and procedures of recurrent tests are at the discretion of the manufacturer or based on agreements between manufacturer and customer.

15 Test methods for nonconformity

Classifications according to IEC 60051-1:2016, 8.5 and 8.6, are procedures provided as guidance to establish the compliance of instruments and are based on a common agreement between manufacturer and customer.

IECNORM.COM : Click to view the full PDF of IEC 60051-9:2019

Bibliography

IEC 60068-2-27, *Environmental testing – Part 2-27: Tests – Test Ea and guidance: Shock*

IECNORM.COM : Click to view the full PDF of IEC 60051-9:2019

IECNORM.COM : Click to view the full PDF of IEC 60051-9:2019

SOMMAIRE

AVANT-PROPOS	13
INTRODUCTION.....	15
1 Domaine d'application	16
2 Références normatives	16
3 Termes et définitions	16
4 Exigences pour l'équipement d'essai	17
4.1 Généralités	17
4.2 Equipement d'essai pour la mesure de l'incertitude intrinsèque d'un instrument et des variations	17
4.3 Equipement d'essai pour essais de surcharge, autoéchauffement, amortissement, dépassement et temps de réponse	17
4.4 Equipement d'essai pour la mesure du déphasage	17
4.5 Equipement d'essai pour la mesure des températures.....	17
4.6 Equipement d'essai pour la mesure de l'humidité	17
5 Méthodes d'essai pour la détermination de l'incertitude intrinsèque	17
6 Méthodes d'essai pour la mesure de l'amortissement	18
7 Méthodes d'essai pour la mesure de l'humidité	18
8 Essais avec des grandeurs de courant alternatif distordues	18
9 Méthodes d'essai pour la mesure des chocs et des vibrations	18
10 Méthode d'essai pour la mesure du champ magnétique externe	18
11 Méthode d'essai pour la mesure du champ électrique externe	18
12 Méthodes d'essai pour la mesure de la sécurité électrique	18
13 Méthodes d'essai pour la mesure de la compatibilité électromagnétique (CEM)	19
14 Méthodes d'essai pour les essais récurrents.....	19
15 Méthodes d'essai pour la mesure des non-conformités	19
Bibliographie.....	20

COMMISSION ÉLECTROTECHNIQUE INTERNATIONALE

**APPAREILS DE MESURE ÉLECTRIQUES INDICATEURS
ANALOGIQUES À ACTION DIRECTE ET LEURS ACCESSOIRES –****Partie 9: Méthodes d'essai recommandées****AVANT-PROPOS**

- 1) La Commission Électrotechnique Internationale (IEC) est une organisation mondiale de normalisation composée de l'ensemble des comités électrotechniques nationaux (Comités nationaux de l'IEC). L'IEC a pour objet de favoriser la coopération internationale pour toutes les questions de normalisation dans les domaines de l'électricité et de l'électronique. À cet effet, l'IEC – entre autres activités – publie des Normes internationales, des Spécifications techniques, des Rapports techniques, des Spécifications accessibles au public (PAS) et des Guides (ci-après dénommés "Publication(s) de l'IEC"). Leur élaboration est confiée à des comités d'études, aux travaux desquels tout Comité national intéressé par le sujet traité peut participer. Les organisations internationales, gouvernementales et non gouvernementales, en liaison avec l'IEC, participent également aux travaux. L'IEC collabore étroitement avec l'Organisation Internationale de Normalisation (ISO), selon des conditions fixées par accord entre les deux organisations.
- 2) Les décisions ou accords officiels de l'IEC concernant les questions techniques représentent, dans la mesure du possible, un accord international sur les sujets étudiés, étant donné que les Comités nationaux intéressés sont représentés dans chaque comité d'études.
- 3) Les Publications de l'IEC se présentent sous la forme de recommandations internationales et sont agréées comme telles par les Comités nationaux de l'IEC. Tous les efforts raisonnables sont entrepris afin que l'IEC s'assure de l'exactitude du contenu technique de ses publications; l'IEC ne peut pas être tenue responsable de l'éventuelle mauvaise utilisation ou interprétation qui en est faite par un quelconque utilisateur final.
- 4) Dans le but d'encourager l'unification internationale, les Comités nationaux de l'IEC s'engagent à appliquer de façon transparente, dans toute la mesure possible, les publications de l'IEC dans leurs publications nationales et régionales. Toutes divergences entre toutes Publications de l'IEC et toutes publications nationales ou régionales correspondantes doivent être indiquées en termes clairs dans ces dernières.
- 5) L'IEC elle-même ne fournit aucune attestation de conformité. Des organismes de certification indépendants fournissent des services d'évaluation de conformité et, dans certains secteurs, accèdent aux marques de conformité de l'IEC. L'IEC n'est responsable d'aucun des services effectués par les organismes de certification indépendants.
- 6) Tous les utilisateurs doivent s'assurer qu'ils sont en possession de la dernière édition de cette publication.
- 7) Aucune responsabilité ne doit être imputée à l'IEC, à ses administrateurs, employés, auxiliaires ou mandataires, y compris ses experts particuliers et les membres de ses comités d'études et des Comités nationaux de l'IEC, pour tout préjudice causé en cas de dommages corporels et matériels, ou de tout autre dommage de quelque nature que ce soit, directe ou indirecte, ou pour supporter les coûts (y compris les frais de justice) et les dépenses découlant de la publication ou de l'utilisation de cette Publication de l'IEC ou de toute autre Publication de l'IEC, ou au crédit qui lui est accordé.
- 8) L'attention est attirée sur les références normatives citées dans cette publication. L'utilisation de publications référencées est obligatoire pour une application correcte de la présente publication.
- 9) L'attention est attirée sur le fait que certains des éléments de la présente Publication de l'IEC peuvent faire l'objet de droits de brevet. L'IEC ne saurait être tenue pour responsable de ne pas avoir identifié de tels droits de brevets et de ne pas avoir signalé leur existence.

La Norme internationale IEC 60051-9 a été établie par le comité d'études 85 de l'IEC: Équipement de mesure des grandeurs électriques et électromagnétiques.

Cette cinquième édition annule et remplace la quatrième édition parue en 1988, son Amendement 1:1994 et son Amendement 2:1995. Cette édition constitue une révision technique.

La présente édition inclut les modifications techniques majeures suivantes par rapport à l'édition précédente:

- a) ajout d'exigences de performance pour l'équipement d'essai;
- b) mise à jour des références aux normes de méthodes d'essai applicables.

Le texte de cette Norme internationale est issu des documents suivants:

FDIS	Rapport de vote
85/663/FDIS	85/672/RVD

Le rapport de vote indiqué dans le tableau ci-dessus donne toute information sur le vote ayant abouti à l'approbation de cette Norme internationale.

Cette Norme internationale doit être utilisée conjointement avec l'IEC 60051-1:2016.

Ce document a été rédigé selon les Directives ISO/IEC, Partie 2.

Une liste de toutes les parties de la série IEC 60051, publiées sous le titre général *Appareils de mesure électriques indicateurs analogiques à action directe et leurs accessoires*, peut être consultée sur le site web de l'IEC.

Le comité a décidé que le contenu de ce document ne sera pas modifié avant la date de stabilité indiquée sur le site web de l'IEC sous "<http://webstore.iec.ch>" dans les données relatives au document recherché. À cette date, le document sera:

- reconduit,
- supprimé,
- remplacé par une édition révisée, ou
- amendé.

IECNORM.COM : Click to view the full PDF of IEC 60051-9:2019

INTRODUCTION

La série IEC 60051 est publiée sous forme de plusieurs parties conformément à la structure suivante, sous le titre général *Appareils de mesure électriques indicateurs analogiques à action directe et leurs accessoires*.

Partie 1: Définitions et exigences générales communes à toutes les parties

Partie 2: Exigences particulières pour les ampèremètres et les voltmètres

Partie 3: Exigences particulières pour les wattmètres et les varmètres

Partie 4: Exigences particulières pour les fréquencemètres

Partie 5: Exigences particulières pour les phasemètres, les appareils de mesure de facteur de puissance et les synchronoscopes

Partie 6: Exigences particulières pour les ohmmètres (les impédancemètres) et les conductancemètres

Partie 7: Exigences particulières pour les appareils à fonctions multiples

Partie 8: Exigences particulières pour les accessoires

Partie 9: Méthodes d'essai recommandées

L'IEC 60051-9 ne se suffit pas à elle-même et est utilisée conjointement avec l'IEC 60051-1.

IECNORM.COM : Click to view the full PDF of IEC 60051-9:2019

APPAREILS DE MESURE ÉLECTRIQUES INDICATEURS ANALOGIQUES À ACTION DIRECTE ET LEURS ACCESSOIRES –

Partie 9: Méthodes d'essai recommandées

1 Domaine d'application

La présente partie de l'IEC 60051 s'applique aux appareils de mesure électriques indicateurs analogiques à action directe et à leurs accessoires, et fournit des recommandations pour les méthodes d'essai applicables ainsi que pour les performances de l'équipement d'essai.

Ce document ne s'applique pas aux:

- appareils spéciaux couverts par leurs propres Normes internationales IEC;
- dispositifs spéciaux couverts par leurs propres Normes internationales IEC, lorsqu'ils sont utilisés en tant qu'accessoires.

2 Références normatives

Les documents suivants cités dans le texte constituent pour tout ou partie de leur contenu, des exigences du présent document. Pour les références datées, seule l'édition citée s'applique. Pour les références non datées, la dernière édition du document de référence s'applique (y compris les éventuels amendements).

IEC 60051-1:2016, *Appareils de mesure électriques indicateurs analogiques à action directe et leurs accessoires – Partie 1: Définitions et exigences générales communes à toutes les parties*

IEC 60068-2-6, *Essais d'environnement – Partie 2-6: Essais – Essai Fc: Vibrations (sinusoïdales)*

IEC 61000-4-8, *Compatibilité électromagnétique (CEM) – Partie 4-8: Techniques d'essai et de mesure – Essai d'immunité au champ magnétique à la fréquence du réseau*

IEC 61010-1:2010, *Règles de sécurité pour appareils électriques de mesurage, de régulation et de laboratoire – Partie 1: Exigences générales*

IEC 61010-2-030, *Règles de sécurité pour appareils électriques de mesurage, de régulation et de laboratoire – Partie 2-030: Exigences particulières pour les appareils équipés de circuits d'essai ou de mesure*

IEC 61326-1:2012, *Matériel électrique de mesure, de commande et de laboratoire – Exigences relatives à la CEM – Partie 1: Exigences générales*

IEC 61326-2-1, *Matériel électrique de mesure, de commande et de laboratoire – Exigences relatives à la CEM – Partie 2-1: Exigences particulières – Configurations d'essai, conditions fonctionnelles et critères de performance pour essai de sensibilité et équipement de mesure pour les applications non protégées de la CEM*

3 Termes et définitions

Pour les besoins du présent document, les termes et les définitions de l'IEC 60051-1:2016 s'appliquent.

L'ISO et l'IEC tiennent à jour des bases de données terminologiques destinées à être utilisées en normalisation, consultables aux adresses suivantes:

- IEC Electropedia: disponible à l'adresse <http://www.electropedia.org/>
- ISO Online browsing platform: disponible à l'adresse <http://www.iso.org/obp>

4 Exigences pour l'équipement d'essai

4.1 Généralités

L'équipement d'essai doit présenter une exactitude et une stabilité suffisantes pour permettre la détermination de l'incertitude intrinsèque d'un instrument et des variations.

NOTE Des protocoles d'étalonnage peuvent être utilisés.

4.2 Equipement d'essai pour la mesure de l'incertitude intrinsèque d'un instrument et des variations

L'incertitude instrumentale maximale des instruments de référence utilisés pour déterminer l'incertitude intrinsèque d'un instrument et des variations doit être égale à 0,2 fois l'indice de classe de l'instrument soumis à essai. Lorsqu'un protocole d'étalonnage est utilisé, l'incertitude d'étalonnage maximale doit être égale à 0,2 fois l'indice de classe de l'équipement sous test (EST).

4.3 Equipement d'essai pour essais de surcharge, autoéchauffement, amortissement, dépassement et temps de réponse

L'incertitude maximale des paramétrages des grandeurs électriques qui sera utilisée pour les essais de surcharge, d'autoéchauffement, d'amortissement, de dépassement et de temps de réponse doit être égale à ± 1 % de ces paramétrages.

4.4 Equipement d'essai pour la mesure du déphasage

L'équipement d'essai destiné aux mesures du déphasage pour les mesureurs de puissance indices de classe de 0,5 ou plus doit présenter une incertitude maximale de $\pm 1^\circ$.

L'équipement d'essai destiné aux mesures du déphasage pour les mesureurs de puissance avec indices de classe inférieurs à 0,5 doit présenter une incertitude maximale de $\pm 1^\circ$. L'incertitude instrumentale de l'équipement d'essai pour les mesureurs de puissance et les appareils de mesure de facteur de puissance doit satisfaire aux exigences mentionnées en 4.2.

4.5 Equipement d'essai pour la mesure des températures

L'équipement destiné au paramétrage des températures doit présenter une incertitude maximale de température de $\pm 1^\circ\text{C}$ dans le cas d'un EST ayant un indice de classe inférieur ou égal à 0,3. Dans le cas d'un EST ayant un indice de classe supérieur à 0,3 l'incertitude maximale du paramétrage doit être de $\pm 2^\circ\text{C}$.

4.6 Equipement d'essai pour la mesure de l'humidité

L'équipement destiné aux paramétrages de l'humidité doit présenter une incertitude maximale des paramétrages d'humidité de ± 3 %.

5 Méthodes d'essai pour la détermination de l'incertitude intrinsèque

L'incertitude intrinsèque en des points d'essai définis doit être vérifiée en augmentant et en diminuant lentement les valeurs de mesure, sans dépasser les limites et sans discontinuité. Les points d'essai pour l'essai de type peuvent être définis par le fabricant. Au cours des