

IEC 60601-2-20
(Second edition – 2009)

Medical electrical equipment –

**Part 2-20: Particular requirements for the basic safety
and essential performance of infant transport incubators**

CORRIGENDUM 1

201.4.3.101 ESSENTIAL PERFORMANCE OF INFANT TRANSPORT INCUBATORS

Replace the existing Table 201.101 by the following:

Table 201.101 – Additional ESSENTIAL PERFORMANCE requirements

Requirement	Subclause
ESSENTIAL PERFORMANCE requirement 1	201.12.1.104 or generation of a visual and audible alarm in compliance with 201.9.6.2.1.102
ESSENTIAL PERFORMANCE requirement 2	201.12.1.106 or generation of a visual and audible alarm in compliance with 201.9.6.2.1.102

201.9.6.2.1.101 Sound level within the COMPARTMENT

In the first paragraph, replace the reference to "201.12.3.107" by "201.9.6.2.1.103".