

INTERNATIONAL STANDARD

IEC
60601-2-24

First edition
1998-02

Medical electrical equipment –

Part 2-24:

Particular requirements for the safety of infusion pumps and controllers

Appareils électromédicaux –

Partie 2-24:

Règles particulières de sécurité des pompes et régulateurs de perfusion



Reference number
IEC 60601-2-24:1998(E)

Numbering

As from 1 January 1997 all IEC publications are issued with a designation in the 60000 series.

Consolidated publications

Consolidated versions of some IEC publications including amendments are available. For example, edition numbers 1.0, 1.1 and 1.2 refer, respectively, to the base publication, the base publication incorporating amendment 1 and the base publication incorporating amendments 1 and 2.

Validity of this publication

The technical content of IEC publications is kept under constant review by the IEC, thus ensuring that the content reflects current technology.

Information relating to the date of the reconfirmation of the publication is available in the IEC catalogue.

Information on the revision work, the issue of revised editions and amendments may be obtained from IEC National Committees and from the following IEC sources:

- **IEC Bulletin**
- **IEC Yearbook**
On-line access*
- **Catalogue of IEC publications**
Published yearly with regular updates
(On-line access)*

Terminology, graphical and letter symbols

For general terminology, readers are referred to IEC 60050: *International Electrotechnical Vocabulary* (IEV).

For graphical symbols, and letter symbols and signs approved by the IEC for general use, readers are referred to publications IEC 60027: *Letter symbols to be used in electrical technology*, IEC 60417: *Graphical symbols for use on equipment. Index, survey and compilation of the single sheets* and IEC 60617: *Graphical symbols for diagrams*.

IEC publications prepared by the same technical committee

The attention of readers is drawn to the end pages of this publication which list the IEC publications issued by the technical committee which has prepared the present publication.

* See web site address on title page.

INTERNATIONAL STANDARD

IEC 60601-2-24

First edition
1998-02

Medical electrical equipment –

Part 2-24:

Particular requirements for the safety of infusion pumps and controllers

Appareils électromédicaux –

Partie 2-24:

Règles particulières de sécurité des pompes et régulateurs de perfusion

© IEC 1998 Copyright - all rights reserved — Droits de reproduction réservés

Aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'éditeur.

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission
Telefax: +41 22 919 0300

3, rue de Varembé Geneva, Switzerland
e-mail: inmail@iec.ch IEC web site <http://www.iec.ch>



Commission Electrotechnique Internationale
International Electrotechnical Commission
Международная Электротехническая Комиссия

PRICE CODE **XB**

For price, see current catalogue

CONTENTS

	Page
FOREWORD	4
INTRODUCTION	5
Clause	
SECTION ONE – GENERAL	
1 Scope and object	6
2 Terminology and definitions	7
3 General requirements	10
5 Classification	11
6 Identification, marking and documents	11
SECTION TWO – ENVIRONMENTAL CONDITIONS	
10 Environmental conditions	13
SECTION THREE – PROTECTION AGAINST ELECTRIC SHOCK HAZARDS	
14 Requirements related to classification	14
17 Separation	14
19 Continuous LEAKAGE CURRENTS and PATIENT AUXILIARY CURRENTS.....	14
SECTION FOUR – PROTECTION AGAINST MECHANICAL HAZARDS	
21 Mechanical strength	16
SECTION FIVE – PROTECTION AGAINST HAZARDS FROM UNWANTED OR EXCESSIVE RADIATION	
36 Electromagnetic compatibility	17
SECTION SIX – PROTECTION AGAINST HAZARDS OF IGNITION OF FLAMMABLE ANAESTHETIC MIXTURES	
SECTION SEVEN – PROTECTION AGAINST EXCESSIVE TEMPERATURES AND OTHER SAFETY HAZARDS	
44 Overflow, spillage, leakage, humidity, ingress of liquids, cleaning, sterilization, disinfection and compatibility	19
47 Electrostatic charges	20
49 Interruption of the power supply	20

Clause	Page
SECTION EIGHT – ACCURACY OF OPERATING DATA AND PROTECTION AGAINST HAZARDOUS OUTPUT	
50 Accuracy of operating data.....	21
51 Protection against hazardous output	38
SECTION NINE – ABNORMAL OPERATION AND FAULT CONDITIONS: ENVIRONMENTAL TESTS	
SECTION TEN – CONSTRUCTIONAL REQUIREMENTS	
54 General	42
56 Components and general assembly.....	44
Annexes	
L References – Publications mentioned in this standard.....	45
AA General guidance and rationale.....	47

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

**Part 2-24: Particular requirements for the safety
of infusion pumps and controllers**

FOREWORD

- 1) The IEC (International Electrotechnical Commission) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of the IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, the IEC publishes International Standards. Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. The IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of the IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested National Committees.
- 3) The documents produced have the form of recommendations for international use and are published in the form of standards, technical reports or guides and they are accepted by the National Committees in that sense.
- 4) In order to promote international unification, IEC National Committees undertake to apply IEC International Standards transparently to the maximum extent possible in their national and regional standards. Any divergence between the IEC Standard and the corresponding national or regional standard shall be clearly indicated in the latter.
- 5) The IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with one of its Standards.
- 6) Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. The IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-2-24 has been prepared by subcommittee 62D: Electro-medical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

The text of this Particular Standard is based on the following documents:

FDIS	Report on voting
62D/250/FDIS	62D/268/RVD

Full information on the voting for the approval of this Particular Standard can be found in the report on voting indicated in the above table.

Annex L is an integral part of this standard.

Annex AA is for information only.

In this Particular Standard the following print types are used:

- requirements, compliance with which can be tested, and definitions: in roman type;
- notes, explanations, advice, introductions, general statements, exceptions and references: in smaller type;
- *test specifications: in italic type;*
- TERMS DEFINED IN CLAUSE 2 OF THE GENERAL STANDARD IEC 60601-1 OR THIS PARTICULAR STANDARD: SMALL CAPITALS.

A bilingual version of this standard may be issued at a later date.

INTRODUCTION

This Particular Standard deals with the safety of INFUSION PUMPS and CONTROLLERS. The relationship between this Particular Standard, IEC 60601-1 (including amendments 1 and 2), and the Collateral Standards is explained in 1.3.

The safe use of infusion pumps and controllers is primarily the responsibility of the OPERATOR. It is also recognized that OPERATORS should be trained in the operation of MEDICAL ELECTRICAL EQUIPMENT and that safe use of the EQUIPMENT can only be achieved if it is operated in accordance with the manufacturer's instructions for use. The minimum specified safety requirements are considered to provide a practical degree of safety in operation. It is the responsibility of the manufacturer to ensure that the requirements of this Particular Standard are reliably implemented. This Particular Standard has been developed in accordance with these principles.

Safe use can be ensured only if the associated disposable parts, especially lines and syringes are consistent with the system. ISO 7886-2:1996, *Sterile hypodermic syringes for single use – Part 2: Syringes for use with power-driven syringe pumps* should be taken into account.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-24: Particular requirements for the safety of infusion pumps and controllers

SECTION ONE – GENERAL

The clauses and subclauses of this section of the General Standard and of this section of the Collateral Standard IEC 60601-1-2 apply, except as follows:

1 Scope and object

This clause of the General Standard and this clause of the Collateral Standard IEC 60601-1-2 apply, except as follows:

1.1* Scope

Addition:

This Particular Standard specifies the requirement for INFUSION PUMPS, INFUSION CONTROLLERS, SYRINGE PUMPS and PUMPS FOR AMBULATORY USE, as defined in 2.101 to 2.110. These devices are intended for use by medical staff and home PATIENTS as prescribed and medically indicated. These particular requirements do not apply to devices:

- 1) specifically intended for diagnostic or similar use (e.g. angiography or other pumps permanently controlled or supervised by the OPERATOR),
- 2) enteral infusion,
- 3) extracorporeal circulation of blood,
- 4) implantable or disposable devices,
- 5) EQUIPMENT specifically intended for diagnostic use within urodynamics (measurement of pressure-volume relationship of the urinary bladder when filled through a catheter with water);
- 6) EQUIPMENT specifically intended for diagnostic use within male impotence testing (measurement of amount of liquid infused, necessary to maintain a preset pressure level for maintaining penile erection: cavernosometry, cavernosography).

1.3 Particular standards

Addition:

This Particular Standard refers to IEC 60601-1:1988, *Medical electrical equipment – Part 1: General requirements for safety* as amended by its amendment 1 (1991) and amendment 2 (1995) and to the Collateral Standard IEC 60601-1-2:1993, *Medical electrical equipment – Part 1: General requirements for safety – 2. Collateral Standard: Electromagnetic compatibility – Requirements and tests*.

For brevity, Part 1 is referred to in this Particular Standard either as the General Standard or as the General Requirement(s) and IEC 60601-1-2 as the Collateral Standard.

The numbering of sections, clauses and subclauses of this Particular Standard corresponds to that of the General Standard. The changes to the text of the General Standard are specified by the use of the following words:

“Replacement” means that the clause or subclause of the General Standard is replaced completely by the text of this Particular Standard.

“Addition” means that the text of this Particular Standard is additional to the requirements of the General Standard.

“Amendment” means that the clause or subclause of the General Standard is amended as indicated by the text of this Particular Standard.

Subclauses or figures which are additional to those of the General Standard are numbered starting from 101, additional annexes are lettered AA, BB, etc., and additional items *aa*), *bb*), etc.

The term “this Standard” is used to make reference to the General Standard, the Collateral Standard and this Particular Standard taken together.

Where there is no corresponding section, clause or subclause in this Particular Standard, the section, clause or subclause of the General Standard, although possibly not relevant, applies without modification; where it is intended that any part of the General Standard, although possibly relevant, is not to be applied, a statement to that effect is given in this Particular Standard.

The requirements of this Particular Standard take priority over those of the General Standard.

The requirements are followed by specifications for the relevant tests.

Following the decision taken by subcommittee 62D at the meeting in Washington in 1979, a “General guidance and rationale” section giving some explanatory notes, where appropriate, about the more important requirements is included in annex AA.

Clauses or subclauses for which there are explanatory notes in annex AA are marked with an asterisk (*).

It is considered that a knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, this annex does not form part of the requirements of this Standard.

1.5 Collateral Standards

Addition:

This Particular Standard also refers to IEC 60601-1-2, which is applicable unless otherwise stated in a particular clause or subclause.

2 Terminology and definitions

This clause of the General Standard and of the Collateral Standard IEC 60601-1-2 apply, except as follows:

2.1.3

ACCESSORY

Addition:

Separate programmers are regarded as accessories and therefore a component part of the EQUIPMENT

2.1.5

APPLIED PART

Replacement:

entirety of all parts of the EQUIPMENT including the infusion liquid pathway that is intentionally in contact with the PATIENT being treated in NORMAL USE

2.2.18

PORTABLE EQUIPMENT

Replacement:

TRANSPORTABLE EQUIPMENT intended to be moved from one location to another while in use or between periods of use, by one or more persons or by other means

Additional definitions:

2.101

INFUSION PUMP

EQUIPMENT intended to regulate the flow of liquids into the PATIENT under positive pressure generated by the pump

The INFUSION PUMP may be of:

- type 1: continuous infusion flow only,
- type 2: non-continuous flow only,
- type 3: discrete delivery of a BOLUS,
- type 4: type 1 combined with type 3 and/or type 2 in the same EQUIPMENT,
- type 5: PROFILE PUMP.

2.102

VOLUMETRIC INFUSION PUMP

INFUSION PUMP in which the delivery rate is set by the OPERATOR and indicated by the EQUIPMENT in volume per unit of time, but excluding SYRINGE PUMPS

2.103

DRIP-RATE INFUSION PUMP

INFUSION PUMP in which the delivery rate is set by the OPERATOR and indicated by the EQUIPMENT as a number of drops per unit of time

2.104

INFUSION CONTROLLER

EQUIPMENT intended to regulate the flow of liquid into the PATIENT under positive pressure generated by gravitational force

2.105

VOLUMETRIC INFUSION CONTROLLER

INFUSION CONTROLLER in which the delivery rate is set by the OPERATOR and indicated by the EQUIPMENT in volume per unit of time

2.106

DRIP-RATE INFUSION CONTROLLER

INFUSION CONTROLLER in which the delivery rate is set by the OPERATOR and indicated by the EQUIPMENT as a number of drops per unit of time

2.107**SPECIAL USE EQUIPMENT**

EQUIPMENT in which the delivery rate is set by the OPERATOR and indicated by the EQUIPMENT in units other than those defined in 2.101 to 2.106

2.108**SYRINGE PUMP**

EQUIPMENT intended for controlled infusion of liquids into the PATIENT by means of one or more single action syringe(s) or similar container(s) (e.g. where the cartridge is emptied by pushing on its plunger)) and in which the delivery rate is set by the OPERATOR and indicated by the EQUIPMENT in volume per unit of time

2.109**INFUSION PUMP FOR AMBULATORY USE**

EQUIPMENT intended for the controlled infusion of liquids into the PATIENT and intended to be carried continuously by the PATIENT

2.110**PROFILE PUMP**

EQUIPMENT intended for controlled infusion of liquids into the PATIENT by means of a programmed sequence of delivery rates

2.111**REGION OF CONTROL**

that part of the EQUIPMENT within which flow regulation, flow shut-off or air detection occurs, within the body of the EQUIPMENT or remotely

2.112**ADMINISTRATION SET**

device(s) that convey(s) liquid from the supply via the EQUIPMENT to the PATIENT

2.113**PATIENT LINE**

that part of the ADMINISTRATION SET between the EQUIPMENT and the PATIENT

2.114**SUPPLY LINE**

that part of the ADMINISTRATION SET between the liquid supply and the EQUIPMENT

2.115**OCCCLUSION ALARM THRESHOLD (PRESSURE)**

value of the physical quantity at which the occlusion alarm is activated

2.116**KEEP OPEN RATE (KOR)**

low predetermined rate(s) to which the EQUIPMENT reverts under specified conditions with the object of keeping the PATIENT LINE open

NOTE – The abbreviation KVO (Keep-Vein-Open Rate) is commonly used as a synonym of KOR.

2.117**FREE FLOW**

flow in an ADMINISTRATION SET which is not controlled by the EQUIPMENT, for example, due to the unintended effects of gravity by the removal of the ADMINISTRATION SET from the EQUIPMENT

2.118**ADMINISTRATION SET CHANGE INTERVAL**

time recommended by the manufacturer of the EQUIPMENT for using the ADMINISTRATION SET

2.119

BOLUS

discrete quantity of liquid which is delivered in a short time

2.120

INTERMEDIATE RATE defined as follows:

- for volumetric infusion pumps and volumetric infusion controllers, set the rate to 25 ml/h;
- for drip-rate infusion pumps and drip-rate infusion controllers, set the rate to 20 drops/minute;
- for syringe pumps, set the rate to 5 ml/h;
- for special use equipment and infusion pumps for ambulatory use, set the rate specified by the manufacturer as typical for the equipment.

2.121

MINIMUM RATE

lowest rate selectable by the OPERATOR, but not less than 1 ml/h

NOTE – For INFUSION PUMPS FOR AMBULATORY USE it is the lowest selectable rate.

2.122

MAXIMUM INFUSION PRESSURE

maximum pressure which can be generated by the EQUIPMENT under conditions of total occlusion at the end of the PATIENT LINE

2.123

PATIENT END

that end of the PATIENT LINE where connection to the PATIENT takes place

3 General requirements

This clause of the General Standard applies, except as follows:

3.6* Addition

SINGLE FAULT CONDITIONS occurring in those protective systems specified in 51.5 and 51.102 shall become obvious to the OPERATOR within the ADMINISTRATION SET CHANGE INTERVAL. SINGLE FAULT CONDITIONS occurring in the protective system specified in clause 51.103 shall cause the cessation of delivery and the generation of an alarm within a time interval less than the volume of the ADMINISTRATION SET between the air detector and the venous cannula connected to it divided by the maximum flow rate of the pump.

NOTE – Acceptable methods of complying with this requirement are, for example:

- 1) a safety system check initiated and controlled by the EQUIPMENT, first at the beginning of the ADMINISTRATION SET CHANGE INTERVAL, and then repeated continuously as warranted;
- 2) one or more protective systems checks initiated by the OPERATOR and controlled by the EQUIPMENT within the ADMINISTRATION SET CHANGE INTERVAL, with the OPERATOR initiating checks before or during the infusion;
- 3) a safety system check carried out by the OPERATOR at least once within the ADMINISTRATION SET CHANGE INTERVAL (see 6.8.2 a) 24)).

The following are not regarded as SINGLE FAULT CONDITIONS, but are regarded as NORMAL USE CONDITIONS:

- leakage from the ADMINISTRATION SET and/or the liquid supply;
- depletion of the INTERNAL ELECTRICAL POWER SOURCE;
- mispositioning and/or incorrect filling of a drip chamber;

- air in the SUPPLY LINE or the REGION OF CONTROL;
- pulling on the PATIENT LINE (see ISO 8536-4).

5 Classification

This clause of the General Standard applies, except as follows:

5.2 Amendment:

Delete TYPE B APPLIED PART;

5.6 Amendment:

Delete all except for CONTINUOUS OPERATION.

6 Identification, marking and documents

This clause of the General Standard and this clause of the Collateral Standard IEC 60601-1-2 apply, except as follows:

6.1 Marking on the outside of EQUIPMENT or EQUIPMENT parts

Addition:

- aa) If detachable liquid reservoirs or PATIENT LINE(S) of specific sizes or brands, or containing specific concentrations of drugs need to be used to maintain safe NORMAL USE of the EQUIPMENT then relevant markings shall be fixed or indicated in a prominent place on the EQUIPMENT which either identify those conditions or provide location of such information.

Compliance is checked by inspection.

6.1 q) Physiological effects

Replacement:

The body of the EQUIPMENT shall be marked with the following:

- 1) symbol No. 14 of appendix D of the General Standard or a statement to refer the OPERATOR to the ACCOMPANYING DOCUMENTS;
- 2) an arrow or other appropriate symbol indicating the correct direction of flow if the ADMINISTRATION SET can be incorrectly loaded;
- 3) EQUIPMENT as defined in 2.103 and 2.106 shall additionally be marked as follows:
"Caution: this equipment controls the drip rate not the volume delivered."

Additional items:

6.1.201 of the Collateral Standard, IEC 60601-1-2

Addition:

Compliance is checked by inspection.

6.8 ACCOMPANYING DOCUMENTS

6.8.2 Instructions for use

a) Addition:

The instructions for use shall also include the following:

- 1) a list of the recommended ADMINISTRATION SET(S) to be used;
- 2) a warning of the consequences of the use of unsuitable ADMINISTRATION SET(S);
- 3) a list of particular ACCESSORIES recommended by the manufacturer for use with the EQUIPMENT;
- 4) permitted EQUIPMENT orientation and methods and precautions concerning its mounting, for example, stability on a pole;
- 5) instructions regarding loading, priming, changing and reloading the ADMINISTRATION SET(S), and the ADMINISTRATION SET CHANGE INTERVAL to maintain the specified performance;
- 6) instructions regarding the use of clamps on an ADMINISTRATION SET, the avoidance of FREE FLOW conditions and the procedure to be followed when changing liquid containers;
- 7) where gravity is relevant to performance, the acceptable height range of the liquid container above the PATIENT's heart;
- 8) the means provided to protect the PATIENT from air infusion;
- 9) a statement of the MAXIMUM INFUSION PRESSURE generated and the OCCLUSION ALARM THRESHOLD (PRESSURE)(S) of the EQUIPMENT;
- 10) a statement of the maximum time for activation of the occlusion alarm when the EQUIPMENT is operating at the MINIMUM RATE and the INTERMEDIATE RATE and at the minimum and maximum selectable OCCLUSION ALARM THRESHOLD (PRESSURE)(S);
- 11) a statement of the BOLUS volume generated as a result of the EQUIPMENT operating at the INTERMEDIATE RATE and reaching the minimum and maximum OCCLUSION ALARM THRESHOLD (PRESSURE) (see also 51.5 b));
- 12) a statement of the means provided (if any) to manage the BOLUS before occlusion release;
- 13) a statement to indicate to the OPERATOR if the EQUIPMENT cannot be used as PORTABLE EQUIPMENT;
- 14) precautions required with drop detectors, for example with respect to placement, cleanliness, liquid level, ambient light;
- 15) recommendations on any specific method of cleaning and maintaining the EQUIPMENT;
- 16) the typical operating time when the EQUIPMENT is operating from the INTERNAL ELECTRICAL POWER SOURCE at the INTERMEDIATE RATE;
- 17) a statement of KEEP OPEN RATE(S), and when initiated;
- 18) a list of alarms and their operating conditions;
- 19)* a warning that under certain circumstances the specified accuracy may not be maintained.
NOTE – The manufacturer must specify the parameters in which the device cannot maintain the specified accuracy; e.g. minimum/maximum viscosity of liquids, reaction time of the safety system, scope of the risk analysis, etc.
- 20)* reference to a guide on the SAFETY HAZARDS associated with the interconnection of other infusion systems or ACCESSORIES to the PATIENT LINE;
- 21) the rate obtained when the prime/purge or BOLUS control is operated, and a statement of any alarm disabled;
- 22) a warning statement on the possible SAFETY HAZARDS associated with external radio-frequency interference (RFI) or electromagnetic radiation which may affect the safe operation of the EQUIPMENT. This statement should include examples of typical EQUIPMENT which may generate such radiation;
- 23) the selectable rate range and the increments of selection;
- 24) guidance on tests to permit the OPERATOR to check the correct functioning of alarm(s) and the operational safety of the EQUIPMENT;

- 25) data as evaluated by the test methods of 50.101 to 50.108 at the rates indicated in table 102, including an explanation for the OPERATOR of the data presentation;
- 26) the time for which the electronic memory is retained following switch-off;
- 27) for SPECIAL USE EQUIPMENT, the conversion factor(s) for volume divided by unit of time;
- 28)* the maximum volume that may be infused under SINGLE FAULT CONDITIONS;
- 29) guidance on the safe operation of the EQUIPMENT if it is connected operationally to a remote control device;
- 30) information concerning type(s) of battery to be used and where available;
- 31) a statement of the meaning of claimed IP-classification.

6.8.201 of IEC 60601-1-2

Addition:

6.8.3 Technical description

Addition:

The technical description shall also include the following:

- aa) the sensitivity of the air detector, if included to comply with 51.9, over the specified range of rates for a single bubble;
- bb) the units of measurement used for calibration of the EQUIPMENT;
- cc) a description of any battery charging system;
- dd) a functional description of the means provided to protect the PATIENT from EQUIPMENT error resulting in overinfusion and, where applicable, in underinfusion;
- ee) the manufacturer shall disclose the ADMINISTRATION SET(s) used for all the tests in this standard.

Compliance is checked by inspection of the ACCOMPANYING DOCUMENTS.

SECTION TWO – ENVIRONMENTAL CONDITIONS

The clauses and subclauses of this section of the General Standard apply, except as follows:

10 Environmental conditions

This clause of the General Standard applies, except as follows:

Replacement:

- 10.2.1 a) An ambient temperature between +5 °C and +40 °C.
- 10.2.1 b) A relative humidity between 20 % and 90 %.

SECTION THREE – PROTECTION AGAINST ELECTRIC SHOCK HAZARDS

The clauses and subclauses of the General Standard apply, except as follows:

14 Requirements related to classification

This clause of the General Standard applies, except as follows:

Replacement:

14.6 b) EQUIPMENT shall be of Type BF or CF.

14.6 d) EQUIPMENT intended for DIRECT CARDIAC APPLICATION having one or more APPLIED PARTS of TYPE CF may have one or more additional APPLIED PARTS of TYPE BF which may be applied simultaneously if the requirements of 6.1 l) and 19.3 for such EQUIPMENT have been met.

17 Separation

This clause of the General Standard applies, except as follows:

Item c) is not applicable.

19 Continuous LEAKAGE CURRENTS and PATIENT AUXILIARY CURRENTS

This clause of the General Standard applies, except as follows:

19.4 Tests:

d) Measuring arrangement

Addition:

- 3) Measurement of the PATIENT LEAKAGE CURRENT shall be made from the APPLIED PART with the PATIENT LINE filled with saline solution (0,9 % NaCl), and with the PATIENT connection immersed in a container of saline solution (0,9 % NaCl) as indicated in figures 101 and 102.




Figure 101 – PATIENT LEAKAGE CURRENT external power supply

(MD = Measuring device, EUT = EQUIPMENT under test)




Figure 102 – PATIENT LEAKAGE CURRENT – INTERNAL ELECTRICAL POWER SOURCE

h) Measure of PATIENT LEAKAGE CURRENT

Addition:

- h) Measurement of PATIENT LEAKAGE CURRENT in a SINGLE FAULT CONDITION shall be performed utilizing the method described in 19.4 d) 3) of this Particular Standard.

SECTION FOUR – PROTECTION AGAINST MECHANICAL HAZARDS

The clauses and subclauses of this section of the General Standard apply, except as follows:

21 Mechanical strength

This clause of the General Standard applies, except as follows:

21.1 Replacement:

EQUIPMENT shall not present a safety hazard to the PATIENT as a result of external vibration. This requirement applies only to PORTABLE EQUIPMENT.

Compliance is checked by inspection and the following test:

Fit the EQUIPMENT with the manufacturer's recommended ADMINISTRATION SET and ACCESSORIES. Apply vibrations in a vertical direction and consecutively in two other directions perpendicular to each other in a horizontal plane and in accordance with the values given in table 101.

Table 101 – Vibration value

Frequency range Hz	Displacement or acceleration (peak value)	Number of sweep cycles in each direction
3 to 8	7,5 mm	4
8 to 300	2 g	4
Applied with a sweep rate of 1 octave/min.		

21.4 Replacement:

Remote parts including MAINS OPERATED adapters and parts not specified in 21.5 shall not present a safety hazard as a result of a free fall from a height of 1 m onto a hard surface. Subsequent to the fall of the remote part, when the EQUIPMENT is turned on for use, it shall either:

- function normally, or
- cease delivery and activate an alarm.

Compliance is checked by the following test:

The sample to be tested is allowed to fall freely once from each of three different starting attitudes from a height of 1 m onto a 50 mm thick hardwood board (e.g. hardwood with a density greater than 700 kg/m³) which lies flat on a rigid base (concrete block). After this test, no LIVE parts shall become accessible. Cracks not visible to the naked eye and surface cracks in fibre reinforced mouldings and the like shall be ignored. If the EQUIPMENT is operational after the test a dielectric strength test and LEAKAGE CURRENT tests according to clauses 19 and 20 and FUNCTIONAL TESTS at the INTERMEDIATE RATE shall be carried out.

21.6 Addition:

INFUSION PUMPS FOR AMBULATORY USE shall not present a SAFETY HAZARD as a result of a free fall from a height of 1 m onto a hard surface.

Compliance is checked by the test of 21.4.

**SECTION FIVE – PROTECTION AGAINST HAZARDS FROM UNWANTED
OR EXCESSIVE RADIATION**

The clauses and subclauses of this section of the General Standard and of this section of the Collateral Standard IEC 60601-1-2 apply, except as follows:

36* Electromagnetic compatibility

This clause of the Collateral Standard IEC 60601-1-2 applies, except as follows.

36.201* EMISSIONS**36.201.1 Radio frequency (RF) EMISSIONS**

36.201.1.3* This subclause of the Collateral Standard IEC 60601-1-2 does *not* apply.

36.201.1.4* This subclause of the Collateral Standard IEC 60601-1-2 applies, except as follows:

Amendment:

Only paragraph 2 of the Collateral Standard IEC 60601-1-2 applies.

36.201.1.5* This subclause of the Collateral Standard IEC 60601-1-2 does *not* apply.

36.201.1.6 High-frequency surgical equipment

This subclause of the Collateral Standard IEC 60601-1-2 does *not* apply.

36.201.1.7***36.201.2.1* VOLTAGE FLUCTUATIONS and harmonic distortion****36.201.2.2* Magnetic field EMISSIONS****36.202* IMMUNITY**

Addition:

The safe functioning of the EQUIPMENT as specified by the manufacturer shall not be impaired by one or more of the immunity tests, or the EQUIPMENT shall fail without creating a SAFETY HAZARD by these tests. In the latter case, the (non-hazardous) failure mode and the failure level to worst case shall be specified by the manufacturer.

Compliance is checked by the following test:

Set up the EQUIPMENT in NORMAL USE according to the manufacturer's instructions for use. Switch on the EQUIPMENT and select the INTERMEDIATE RATE. Carry out the test as described in this Particular Standard according to the test conditions described in this Particular Standard. By inspection and functional tests determine compliance with the additional requirement formulated in the previous paragraph. (In case of doubt and if the EQUIPMENT still continues to infuse liquid, carry out a functional test without changing any of the previously selected parameters, for a period of 1 h). Switch the EQUIPMENT off and then on again. Select the INTERMEDIATE RATE and carry out another functional test for a period of 1 h.

36.202.1* ELECTROSTATIC DISCHARGE

This subclause of the Collateral Standard IEC 60601-1-2 applies, except as follows (see annex AA also):

Amendment:

A level of 8 kV shall apply for contact discharge and a level of 15 kV shall apply for air discharge.

36.202.2 Radiated radiofrequency electromagnetic field

This subclause of the Collateral Standard IEC 60601-1-2 applies, except as follows:

Amendment:

36.202.2.1 Requirements

a)* This item applies except as follows:

The applicable level is not 3 V/m but 10 V/m.

b)* This item does *not* apply

c)* This item does *not* apply

d)*

36.202.2.2* Test conditions

c) This item does *not* apply

e)* This item does *not* apply

36.202.4* VOLTAGE DIPS, short interruptions and voltage variations on power supply input lines

36.202.5* Conducted disturbances, induced by radio-frequency fields above 9 kHz

36.202.6* Magnetic fields

This subclause of the Collateral Standard IEC 60601-1-2 applies, except as follows:

Amendment:

Level: 400 A/m

SECTION SIX – PROTECTION AGAINST HAZARDS OF IGNITION OF FLAMMABLE ANAESTHETIC MIXTURES

The clauses and subclauses of this section of the General Standard apply.

SECTION SEVEN – PROTECTION AGAINST EXCESSIVE TEMPERATURES AND OTHER SAFETY HAZARDS

The clauses and subclauses of this section of the General Standard apply, except as follows:

44 Overflow, spillage, leakage, humidity, ingress of liquids, cleaning, sterilization, disinfection and compatibility

This clause of the General Standard applies, except as follows:

44.3 Spillage

Replacement:

If an IPX1-classification or better is not claimed:

Replacement:

The EQUIPMENT shall be so designed that, taking into consideration ageing and rough handling of the EQUIPMENT, in the event of spillage (accidental wetting) no liquid is retained within the EQUIPMENT ENCLOSURE and the EQUIPMENT shall either continue to function normally or cease delivery and activate an alarm.

Compliance is checked by the following test:

Use the test in accordance with IEC 60529 with a test apparatus for DRIP-PROOF EQUIPMENT.

Place the EQUIPMENT in the position of NORMAL USE. Subject the EQUIPMENT to an artificial rainfall of 3 mm/min for 30 s, falling vertically from a height of 0,5 m above the top of the EQUIPMENT. Carry out the test using tap water. Covers and other parts, for example battery compartment covers, which can be removed without the aid of a TOOL are left in position during the test. Where carrying pouches are specified by the manufacturer as forming part of the protection against spillage then the test is carried out with the EQUIPMENT in the carrying pouch. Where no such specification exists then the carrying pouch is removed prior to the start of the test. Immediately after the 30 s exposure, remove visible moisture from the body of the EQUIPMENT. Immediately after the above test, determine by inspection that the water has not entered the EQUIPMENT. If water has entered the EQUIPMENT, repeat the test using saline solution (0,9 % NaCl). Carry out a functional test at the INTERMEDIATE RATE for a period of 1 h. Carry out the dielectric strength tests specified in 20.4.

44.4* Leakage

Replacement:

EQUIPMENT shall be so constructed that liquid which might leak from containers, tubing, couplings and the like does not impair the safe functioning of the EQUIPMENT nor wet uninsulated LIVE parts or electrical insulation which is liable to be adversely affected by such a liquid.

Compliance is checked by the following test:

Set up the EQUIPMENT in NORMAL USE and according to the manufacturer's instructions for use. By means of a pipette apply drops of the test solution to couplings, tubing connectors, seals and to parts of the ADMINISTRATION SET which might rupture. Moving parts are in operation or at rest whichever is the most unfavourable.

Immediately after application of the test solution, carry out the test(s) from 50.102 to 50.108 according to the classification of the EQUIPMENT, at the INTERMEDIATE RATE only. If the EQUIPMENT does not fall into one of the defined categories then use the appropriate test from 50.102 to 50.108. Carry out the tests of 51.103 and 51.104. Switch off the EQUIPMENT and allow it to stand for a minimum of 12 h under normal conditions ($20\text{ }^{\circ}\text{C} \pm 2\text{ }^{\circ}\text{C}$, $65\text{ \%} \pm 5\text{ \% RH}$). By means of functional tests determine that FREE FLOW does not occur. By inspection, check the function of controls and other parts which may have been adversely affected by the test solution.

Carry out the test with a test solution consisting of a 50 % dextrose solution.

44.6 Ingress of liquids

Addition:

If an IPX1-classification is claimed:

Covers and other parts, for example, battery compartment covers, which can be removed without the aid of a TOOL are left in position during the test. Where carrying pouches are specified by the manufacturer as forming part of the protection against ingress of liquids, then the test is carried out with the EQUIPMENT in the carrying pouch. Where no such specification exists then the carrying pouch is removed prior to the test.

47 Electrostatic charges

Not used. Transferred to clause 36.

49 Interruption of the power supply

This clause of the General Standard applies, except as follows:

49.2

Additions:

EQUIPMENT powered from the SUPPLY MAINS only shall give an audible alarm in the event of an accidental disconnection or a SUPPLY MAINS failure. Under such conditions, the audible alarm shall be maintained for at least 3 min or until power is restored, whichever is the less.

Compliance is checked by inspection and functional tests.

EQUIPMENT which utilizes an INTERNAL ELECTRICAL POWER SOURCE either as a primary or standby supply shall give an audible and visible warning 30 min before delivery ceases due to battery exhaustion. During this period, the EQUIPMENT shall give a continuous visible and an intermittent audible warning.

At least 3 min before the end of the battery life the EQUIPMENT shall give an audible and visible alarm and cease delivery. The alarm shall be maintained for the duration of the remaining battery lifetime.

Compliance is checked by inspection and functional tests when the EQUIPMENT is operated at the INTERMEDIATE RATE and with a fully charged battery.

SECTION EIGHT – ACCURACY OF OPERATING DATA AND PROTECTION AGAINST HAZARDOUS OUTPUT

The clauses and subclauses of this section of the General Standard apply, except as follows:

50 Accuracy of operating data

This clause of the General Standard applies except as follows:

Additions:

50.101* The EQUIPMENT shall maintain the manufacturer's stated accuracy or better over the recommended ADMINISTRATION SET CHANGE INTERVAL.

Compliance is checked, using the tests prescribed in 50.102 to 50.108, to verify the accuracy of the EQUIPMENT according to its defined type and the manufacturer's disclosure of accuracy. If the EQUIPMENT does not fall into one of the defined categories use the appropriate test from 50.102 to 50.108.

Definition of terms given in 50.102 to 50.108

rate r	the delivery rate selected by the operator
flow	the measured output in volume per unit of time
bolus	a discrete quantity of liquid which is delivered in a short time as an infusion but not part of a priming routine
sample interval S	the time between successive mass readings or drop counts
test period T	the total duration of the test from start to finish
analysis period T_0	designated as the first 2 h of the test period
analysis period T_1	designated as the second hour of test period
analysis period T_2	designated as the last hour of the test period
analysis period T_X	the analysis period specified as T_0 , T_1 or T_2
W	the total mass
W_i	the i^{th} mass sample over a specified analysis period
W_j	mass sample at the end of a specified analysis period or test period
W_k	mass sample at the start of a specified analysis period
A	overall mean percentage flow error measured over the analysis period T_1
B	overall mean percentage flow error measured over the analysis period T_2
P	observation window duration
$E_p(\text{max.})$	maximum measured error in observation window of specified duration

$E_p(\text{min.})$	minimum measured error in observation window of specified duration
shot pattern	a sequence of bolus deliveries which may occur at regular or irregular intervals
shot cycle I	the minimum time between successive repetitions of the shot or the shot pattern (from the start of the first shot pattern to the start of the second shot pattern)
density d	density of water (0,998 g/ml at 20 °C)

50.102* Accuracy tests for VOLUMETRIC INFUSION CONTROLLERS, VOLUMETRIC INFUSION PUMPS and SYRINGE PUMPS

The test apparatus shown in figures 104a and 104b is used. Carry out the tests using a test solution of ISO Class III water for medical use and installing an unused ADMINISTRATION SET. Set up the EQUIPMENT with the test solution in accordance with the manufacturer's instructions for use.

Ensure that EQUIPMENT which has a non-delivery segment within its operating cycle has this segment included in the test.

Set the required rate according to table 102. Set the sample interval S to 0,5 min. Begin the test period simultaneously with starting the EQUIPMENT.

Determine the test period T . This test period shall equal the recommended ADMINISTRATION SET CHANGE INTERVAL if there is sufficient fluid in the container. If not, calculate the duration of the test period by dividing the total fluid volume by the rate. Allow the EQUIPMENT to run for the test period T .

For VOLUMETRIC INFUSION PUMPS and SYRINGE PUMPS repeat the test at the INTERMEDIATE RATE for a period of 120 min at back pressures of $\pm 13,33$ kPa (± 100 mm Hg).

For VOLUMETRIC INFUSION CONTROLLERS repeat the test at the INTERMEDIATE RATE for a period of 120 min at a back pressure of $-13,33$ kPa (-100 mm Hg).

The manufacturer shall disclose the maximum deviation between the results under normal conditions and under back pressure conditions, if applicable.

For VOLUMETRIC INFUSION PUMPS repeat the test at the INTERMEDIATE RATE for a period of 120 min with the supply container below the pump mechanism at a distance of 0,5 m with the same ADMINISTRATION SET.

The manufacturer shall disclose the maximum deviation between the results under normal condition and under condition of negative head-height, if applicable.

If the EQUIPMENT incorporates a BOLUS facility carry out the tests specified in 50.106.

If the test of 50.102 cannot be applied because of design features within the EQUIPMENT, apply the most appropriate test from 50.103 to 50.108.

Calculate the actual flow Q_i for each sample interval for the analysis period $T_0(\text{min})$ from equation (1) (see figure 103).

Calculate $E_p(\text{max.})$ and $E_p(\text{min.})$ for the 2, 5, 11, 19 and 31 min observation windows from equations (2) and (3) over the analysis period $T_1(\text{min})$ of the second hour of the test period.

Except for SYRINGE PUMPS calculate $E_p(\text{max.})$ and $E_p(\text{min.})$ for the 2, 5, 11, 19 and 31 min observation windows from equations (2) and (3) over the analysis period $T_2(\text{min})$ of the last hour of the test period.

Plot the following graphs using a linear scale with scale ratios as follows (see Rationale), where r is the set rate (see figures AA.3.1 and AA.3.2):

For start-up graph, flow axis is:

maximum = $2 r$

minimum = $-0,2 r$

scale increment = $0,2 r$

time = 0 min to 120 min (10 min intervals)

For trumpet graph, flow axis is:

maximum = 15 %

minimum = -15 %

scale increment = 5 %

time = 0 min to 31 min (1 min intervals)

Plot flow Q_i (ml/h) against time T_0 (min) for the first 2 h of the test period, see example in figure 105. Indicate the rate by means of a broken line. Indicate flow Q_i by means of a solid line.

Plot percentage variation $E_p(\text{max.})$ and $E_p(\text{min.})$ against observation window duration P (min) and the overall mean percentage error A (derived from equation (4)) measured over the analysis period T_1 (min) of the second hour of the test period. See example in figure 106.

Indicate $E_p(\text{max.})$ and $E_p(\text{min.})$ and the overall mean percentage error A by means of a solid line. Indicate the zero error by means of a dotted line.

Plot percentage variation $E_p(\text{max.})$ and $E_p(\text{min.})$ against observation window duration P (min) and the overall mean percentage error B (derived from equation (5)) measured over the analysis period T_2 (min) of the last hour of the test period.

See example in figure 107.

Indicate $E_p(\text{max.})$ and $E_p(\text{min.})$ and the overall mean percentage error B by means of a solid line. Indicate the zero error by means of a dotted line. This graph is not applicable to SYRINGE PUMPS.

FORMULAE

Calculate flow using the expression:

$$Q_i = \frac{60 (W_i - W_{i-1})}{Sd} \quad (\text{ml/h}) \quad (1)$$

$i = 1, 2 \dots T_0/S$

where

W_i is the i^{th} mass sample from the analysis period T_0 (g) (corrected for evaporative loss);

T_0 is the analysis period (min);

S is the sample interval (min);

d is the density of water (0,998 g/ml at 20 °C).

Calculate $E_p(\text{max.})$ and $E_p(\text{min.})$ using the trumpet algorithm as follows:

For observation windows of duration $P = 2, 5, 11, 19$ and 31 min, within the analysis period T_x , there are a maximum of m observation windows, such that:

$$m = \frac{(t_x - P)}{S} + 1$$

where

m is the maximum number of observation windows;

P is the observation window duration;

S is the sample interval (min);

T_x is the analysis period (min).

The maximum $E_p(\text{max.})$ and minimum $E_p(\text{min.})$ percentage variations within an observation window of duration period P min are given by:

$$E_p(\text{max.}) = \text{MAX}_{j=1}^m \left[\frac{S}{P} \times \sum_{i=j}^{j+\frac{P}{S}-1} 100 \times \left(\frac{Q_i - r}{r} \right) \right] (\%) \quad (2)$$

$$E_p(\text{min.}) = \text{MIN}_{j=1}^m \left[\frac{S}{P} \times \sum_{i=j}^{j+\frac{P}{S}-1} 100 \times \left(\frac{Q_i - r}{r} \right) \right] (\%) \quad (3)$$

where

$$Q_i = \frac{60 (W_i - W_{i-1})}{Sd} \quad (\text{ml/h})$$

W_i is the i^{th} mass sample from the analysis period T_x (g) (corrected for evaporative loss);

r is the rate (ml/h);

S is the sample interval (min);

P is the observation window duration (min);

d is the density of water (0,998 g/ml at 20 °C).

Calculate the overall mean percentage flow error A using the following expression where A is measured over the analysis period T_1 (the second hour of the test period):

$$A = \frac{100 (Q - r)}{r} (\%) \quad (4)$$

where

$$Q = \frac{60 (W_j - W_k)}{T_1 d} \quad (\text{ml/h})$$

r is the rate (ml/h);

W_j is the mass sample at the end of the analysis period T_1 (g) ($j = 240$);

W_k is the mass sample at the start of the analysis period T_1 (g) ($k = 120$);

T_1 is the analysis period (min);

d is the density of water (0,998 g/ml at 20 °C).

Calculate the overall mean percentage flow error B using the following expression where B is measured over the analysis period T_2 (the last hour of the test period):

$$B = \frac{100 (Q - r)}{r} (\%) \quad (5)$$

where

$$Q = \frac{60 (W_j - W_k)}{T_2 d} \quad (\text{ml/h})$$


r is the rate (ml/h);

W_j is the mass sample at the end of the test period T_2 (g) (corrected for evaporative loss);

W_k is the mass sample at the start of the analysis period T_2 (g) (corrected for evaporative loss);

T_2 is the analysis period (min);

d is the density of water (0,998 g/ml at 20 °C).



IEC 255/98

Figure 103 – Analysis periods




Figure 104a – Test apparatus for VOLUMETRIC INFUSION PUMPS and VOLUMETRIC INFUSION CONTROLLERS




Figure 104 b – Test apparatus for SYRINGE PUMPS

NOTE – A balance accurate to five decimal places is required for PUMPS with low MINIMUM RATES.

Set height h_1 (collapsible bag, vented container) in accordance with the manufacturer's instructions for use. The needle (18G, 1,2 mm, ISO 7864) shall be positioned below the liquid surface.

The mean centre line of the pumping chamber to be at the same height as the tip of the needle (18G, 1,2 mm, ISO 7864).




Figure 105 – Start-up graph plotted from data gathered during the first 2 h of the test period




Figure 106 – Trumpet curve plotted from data gathered during the second hour of the test period




Figure 107 – Trumpet curve plotted from data gathered during the last hour of the ADMINISTRATION SET CHANGE INTERVAL

50.103* Accuracy tests for DRIP-RATE INFUSION CONTROLLERS and DRIP-RATE INFUSION PUMPS

The test apparatus shown in figure 108 is used. Carry out the tests using a test solution of ISO class III water for medical use and installing an unused ADMINISTRATION SET. Set up the EQUIPMENT with the test solution in accordance with the manufacturer's instructions for use. Set the required drip rate according to table 102. Set the sample interval to 1 min.

Begin the test period simultaneously with starting the EQUIPMENT.

Determine the test period T . If there is sufficient fluid in the container, this test period shall equal the recommended ADMINISTRATION SET CHANGE INTERVAL. If there is insufficient fluid, the duration of the test period shall be calculated by dividing the total fluid volume by the rate. Allow the EQUIPMENT to run for the test period T .

For DRIP-RATE INFUSION CONTROLLERS repeat the tests at the INTERMEDIATE RATE for a period of 120 min against a back pressure of $-13,33$ kPa (-100 mm Hg).

Compare the results obtained under back pressure conditions with those obtained previously. If the results show a significant deviation outside the tolerance in the ACCOMPANYING DOCUMENTS, then check that a warning statement is included in the ACCOMPANYING DOCUMENTS.

Calculate the actual drip rate Q_i at each sample interval for the analysis period T_0 from equation (1) (see figure 103).


Calculate $E_p(\text{max.})$ and $E_p(\text{min.})$ for the 1, 2, 5, 11, 19 and 31 min observation windows from equations (2) and (3) over the analysis period T_1 (min) of the second hour of the test period.

Calculate $E_p(\text{max.})$ and $E_p(\text{min.})$ for the 1, 2, 5, 11, 19 and 31 min observation windows from equations (2) and (3) over the analysis period T_2 (min) of the last hour of the test period.

For DRIP-RATE INFUSION PUMPS only repeat the tests at the INTERMEDIATE RATE for a period of 120 min against back pressures of $\pm 13,33$ kPa (± 100 mm Hg).

Plot the following graphs:


- Drip rate Q_i (drops/min) against time t (min) for the first 2 h of the test period. See example in figure 109. Indicate the set rate by means of a broken line. Indicate the drip rate Q_i by means of a solid line.
- Percentage variation $E_p(\text{max.})$ and $E_p(\text{min.})$ against observation window duration P (min) and the overall mean percentage error A (derived from equation (4)) measured over the analysis period T_1 (min) of the second hour of the test period. See example in figure 106. Indicate $E_p(\text{max.})$ and $E_p(\text{min.})$ and the overall mean percentage error A by means of a solid line. Indicate the zero error by means of a dotted line.
- Percentage variation $E_p(\text{max.})$ and $E_p(\text{min.})$ against observation window duration P (min) and the overall mean percentage error B (derived from equation (5)) measured over the analysis period T_2 of the last hour of the test period. See example in figure 107. Indicate $E_p(\text{max.})$ and $E_p(\text{min.})$ and the overall mean percentage error B by means of a solid line. Indicate the zero error by means of a dotted line.



IEC 261/98

Figure 108 – Test apparatus for DRIP-RATE INFUSION PUMPS and DRIP-RATE INFUSION CONTROLLERS

Set height h_1 (collapsible bag or vented container) in accordance with the manufacturer's instructions for use. Mean height of pumping chamber to be at the same height as tip of needle. Use configuration (a) if it is possible to place an independent drop detector on the drop chamber. Use configuration (b) (drop signal extracted from EUT) in other circumstances.



IEC 262/98

Figure 109 – Start-up graph plotted from data gathered during the first 2 h of the test period

FORMULAE

Calculate the drip rate using the expression:

$$Q_i = \frac{(N_i - N_{i-1})}{S} \text{ (drops/min)} \quad (6)$$

where

N_i is the i^{th} total drop count sample from the test period;

S is the sample interval (min).

Calculate $E_p(\text{max.})$ and $E_p(\text{min.})$ using the trumpet algorithm as follows:

For observation windows of duration $P = 1, 2, 5, 11, 19$ and 31 min, within the analysis period T_x , there are a maximum of m observation windows, such that:

$$m = \frac{(T_x - P)}{S} + 1$$

where

m is the maximum number of observation windows;

P is the observation window duration (min);

S is the sample interval (min);

T_x is the analysis period (min).

The maximum $E_p(\text{max.})$ and minimum $E_p(\text{min.})$ percentage variations within an observation window of duration period P (min) are given by:

$$E_p(\text{max.}) = \max_{j=1}^m \left[\frac{S}{P} \times \sum_{i=j}^{j+\frac{P}{S}-1} 100 \times \left(\frac{Q_i - r}{r} \right) \right] (\%) \quad (7)$$

$$E_p(\text{min.}) = \min_{j=1}^m \left[\frac{S}{P} \times \sum_{i=j}^{j+\frac{P}{S}-1} 100 \times \left(\frac{Q_i - r}{r} \right) \right] (\%) \quad (8)$$

where

$$Q_i = \frac{(N_i - N_{i-1})}{S} \text{ (drops/min)}$$

N_i is the i^{th} total drop count sample from the analysis period T_x ;

r is the drip rate (drops/min);

S is the sample interval (min);

P is the observation window duration (min).

Calculate the overall mean percentage drip rate error A using the following expression where A is measured over the analysis period T_1 (the second hour of the test period):

$$A = \frac{100 (Q - r)}{r} \quad (\%) \quad (9)$$

where

$$Q = \frac{(N_j - N_k)}{T_1} \quad (\text{drops/min})$$

r is the drip rate (drops/min);

N_j is the total drop count at the end of the analysis period T_1 ($j = 120$);

N_k is the total drop count at the start of the analysis period T_1 ($k = 60$);

T_1 is the analysis period (min).

Calculate the overall mean percentage drip rate error B using the following expression where B is measured over the analysis period T_2 (the last hour of the test period):

$$B = \frac{100 (Q - r)}{r} \quad (\%) \quad (10)$$

where

$$Q = \frac{(N_j - N_k)}{T_2} \quad (\text{drops/min})$$

r is the drip rate (drops/min);

N_j is the total drop count at the end of the test period T ;

N_k is the total drop count at the start of the analysis period T_2 ;

T_2 is the analysis period (min).

50.104* Accuracy tests for INFUSION PUMPS FOR AMBULATORY USE type 1

The test apparatus shown in figure 104b is used. Carry out the tests using a test solution of ISO class III water for medical use or a liquid which can be expected to give similar test results and installing an unused ADMINISTRATION SET. Set up the EQUIPMENT in accordance with the manufacturer's instructions for use. Prime the ADMINISTRATION SET and set the EQUIPMENT for the INTERMEDIATE RATE. Start the EQUIPMENT. Set the sample interval S to 15 min. Allow the EQUIPMENT to run for a time equivalent to half the container volume, or 24 h, whichever is the shorter as a stabilization period T_1 (min). Continue the test without stopping the EQUIPMENT for a further 25 h or until the liquid container is depleted. Measure the mass of infusate W_i delivered at each sample interval. Repeat the test at the MINIMUM RATE.

Calculate the mean flow from equation (6) for every two successive samples over the stabilization period T_1 .

Calculate $E_p(\text{max.})$ and $E_p(\text{min.})$ for the 15, 60, 150, 330, 570 and 930 min observation windows from equations (7) and (8) over the analysis period T_2 (min) starting from the end of the stabilization period to the end of the test.

Plot the following graphs:

- Flow Q_i ($\mu\text{l/h}$) against time (min) over the stabilization period T_1 at 30 min increments. Indicate the rate r ($\mu\text{l/h}$) by means of a broken line. Indicate flow Q_i by means of a solid line. See figure 110 as an example.
- Percentage variation $E_p(\text{max.})$ and $E_p(\text{min.})$ against observation window duration over the analysis period T_2 and the overall mean percentage error A (derived from equation (9)). Indicate the zero error by means of a broken line. Indicate $E_p(\text{max.})$ and $E_p(\text{min.})$ and the overall mean percentage error A by means of solid lines. See figure 111 as an example.




Figure 110 – Start-up graph over the stabilization period




Figure 111 – Trumpet curve plotted from data at the end of the stabilization period




Figure 112 – Start-up curve over the stabilization period for quasi-continuous output pumps




Figure 113 – Trumpet curve plotted from data at the end of the stabilization period for quasi-continuous pumps

FORMULAE

Calculate flow using the expression:

$$Q_i = \frac{60 (W_{2i} - W_{2(i-1)})}{2dS} \quad (\mu\text{l/h}) \quad (11)$$

where

$i = 1, 2, \dots, T_1/2S$;

W_i is the i^{th} mass sample from the stabilization period T_1 (mg) (corrected for evaporative loss);

T_1 is the stabilization period (min) (≈ 24 h);

S is the sample interval (min) (15 min);

d is the density of test liquid at 20 °C (g/ml).

Calculate $E_p(\text{max.})$ and $E_p(\text{min.})$ using the trumpet algorithm as follows:

For observation windows of duration $P = 15, 60, 150, 330, 570$ and 930 min, within the analysis period T_2 , there are a maximum of m observation windows, such that:

$$m = \frac{(T_2 - P)}{S} + 1$$

where

m is the maximum number of observation windows;

P is the observation window duration (min);

T_2 is the analysis period (min);

S is the sample interval (min) (15 min).

The maximum $E_p(\text{max.})$ and minimum $E_p(\text{min.})$ percentage variations, within an observation window of duration period P (min), are given by;

$$E_p(\text{max.}) = \text{MAX}_{j=1}^m \left[\frac{S}{P} \times \sum_{i=j}^{j+\frac{P}{S}-1} 100 \times \left(\frac{Q_i - r}{r} \right) \right] (\%) \quad (12)$$

$$E_p(\text{min.}) = \text{MIN}_{j=1}^m \left[\frac{S}{P} \times \sum_{i=j}^{j+\frac{P}{S}-1} 100 \times \left(\frac{Q_i - r}{r} \right) \right] (\%) \quad (13)$$

where

$$Q_i = \frac{60 (W_i - W_{i-1})}{Sd} \text{ (ml/h)}$$

W_i is the i^{th} mass sample from the analysis period T_2 (mg) (corrected for evaporative loss);

r is the set rate ($\mu\text{l/h}$);

S is the sample interval (min);

P is the observation window duration (min);

d is the density of test liquid at the test temperature (g/ml).

Calculate the overall percentage flow error A using the following expression, where A is measured over the analysis period T_2 :

$$A = \frac{100 (Q - r)}{r} (\%) \quad (14)$$

where

$$Q = \frac{60 (W_j - W_k)}{T_2 d} \text{ (}\mu\text{l/h)}$$

r is the set rate ($\mu\text{l/h}$);

W_j is the mass sample at the end of the analysis period T_2 (mg);

W_k is the mass sample at the start of the analysis period T_2 (mg);

T_2 is the analysis period (min);

d is the density of test liquid at the test temperature (g/ml).

50.105* Accuracy tests for INFUSION PUMPS FOR AMBULATORY USE type 2

The test apparatus shown in figure 104b is used. Carry out the tests using a test solution of ISO class III water for medical use or a liquid which can be expected to give similar test results and installing an unused ADMINISTRATION SET. Set up the EQUIPMENT in accordance with the manufacturer's instructions for use. Prime the ADMINISTRATION SET.

Determine the shot pattern of the pump output. Derive the shot cycle. Measure the time taken (in minutes) for 20 successive shot cycles at the INTERMEDIATE RATE (and ensure that there is sufficient liquid in the container for the subsequent 100 shots after the stabilization period).

Calculate the mean duration of the shot cycle I (min).

Derive sample interval S corresponding to the INTERMEDIATE RATE shot cycle I .

If the shot cycle I is greater than 0,5 min, then:

$$S = kI$$

where

S is the sample interval;

I is the shot cycle;

k is the integer constant = 1.

If the shot cycle I is less than 0,5 min, then

$$S = kI$$

where

S is the sample interval;

I is the shot cycle;

k is the minimum integer constant giving kI approximately equal to 0,5 min.

Synchronize the measuring equipment to measure the mass of infusate delivered in successive sequences of k shot cycles.

Set the EQUIPMENT for the INTERMEDIATE RATE.

Start the EQUIPMENT. Allow the EQUIPMENT to run for a time equivalent to half the container volume or 24 h, whichever is the shorter, as a stabilization period T_1 (min). Continue the test without stopping the EQUIPMENT for a further 100 sample intervals.

Measure the mass of infusate W_i delivered at each sample interval.

Choose any integer n so that:

$$nS \approx 30 \text{ (min)}$$

where

S is the sample interval (kI) (min);

n is the integer constant.

Calculate the mean flow from equation (15) for every successive nS samples, over the stabilization period T_1 .

Calculate $E_p(\text{max.})$ and $E_p(\text{min.})$ for $P = S, 2S, 5S, 11S, 19S$ and $31S$ min observation windows from equations (16) and (17) over the analysis period T_2 starting from the end of the stabilization period to the end of the test.

Plot flow as a function of elapsed time over the stabilization period T_1 defined above. Indicate the rate on the graph by means of a broken line. See figure 112 as an example.

Plot percentage variation $E_p(\text{max.})$ and $E_p(\text{min.})$ against observation window duration, over the analysis period T_2 and the overall mean percentage error A (derived from equation (18)).

Indicate the zero error by means of a broken line. Indicate $E_p(\text{max.})$ and $E_p(\text{min.})$ and the overall mean percentage error A by means of solid lines. See figure 113 as an example.

FORMULAE

Calculate flow using the expression:

$$Q_i = \frac{60 (W_{ni} - W_{n(i-1)})}{ndS} \quad (\mu\text{l/h}) \quad (15)$$

where

$i = 1, 2, \dots, T_1/nS$;

W_i is the i^{th} mass sample from the stabilization period T_1 (mg) (corrected for evaporative loss);

T_1 is the stabilization period (min) (≈ 24 h);

S is the sample interval (min) = (k/min);

n is the integer constant ($nS \approx 30$ min);

d is the density of test liquid at the test temperature (g/ml).

Calculate $E_p(\text{max.})$ and $E_p(\text{min.})$ using the trumpet algorithm as follows:

For consecutive observation windows $P = S, 2S, 5S, 11S, 19S$ and $31S$ min, within the analysis period T_2 , there are a maximum of m successive samples such that:

$$m = \frac{(T_2 - P)}{S} + 1$$

where

m is the maximum number of observation windows;

P is the observation window duration (min);

T_2 is the analysis period (min);

S is the sample interval (min).

The maximum $E_p(\text{max.})$ and minimum $E_p(\text{min.})$ percentage variations, within an observation window of duration period P (min), are given by:

$$E_p(\text{max.}) = \text{MAX}_{j=1}^m \left[\frac{S}{P} \times \sum_{i=j}^{j+\frac{P}{S}-1} 100 \times \left(\frac{Q_i - r}{r} \right) \right] (\%) \quad (16)$$

$$E_p(\text{min.}) = \text{MIN}_{j=1}^m \left[\frac{S}{P} \times \sum_{i=j}^{j+\frac{P}{S}-1} 100 \times \left(\frac{Q_i - r}{r} \right) \right] (\%) \quad (17)$$

where

$$Q_i = \frac{60 (W_i - W_{i-1})}{Sd} \quad (\mu\text{l/h})$$

W_i is the i^{th} mass sample from the analysis period T_2 (mg) (corrected for evaporative loss);

r is the set rate ($\mu\text{l/h}$);

S is the sample interval (min);

P is the observation window duration (min);

d is the density of test liquid at the test temperature (g/ml).

Calculate the overall percentage flow error A using the following expression, where A is measured over the analysis period T_2 :

$$A = \frac{100 (Q - r)}{r} (\%) \quad (18)$$

where

$$Q = \frac{60 (W_j - W_k)}{T_2 d} \quad (\mu\text{l/h})$$

r is the set rate ($\mu\text{l/h}$);

W is the total mass (mg) (corrected for evaporative loss);

W_j is the mass sample at the end of the analysis period T_2 (mg);

W_k is the mass sample at the start of the analysis period T_2 (mg);

T_2 is the analysis period (min);

d is the density of test liquid at the test temperature (g/ml).

50.106* Accuracy tests for pumps type 3

The test apparatus shown in figures 104a or 104b is used (as appropriate) using a test solution of ISO class III water for medical use or a liquid which can be expected to give similar test results and installing an unused ADMINISTRATION SET. Set up the EQUIPMENT with the recommended ADMINISTRATION SET in accordance with the manufacturer's instructions for use. Set the EQUIPMENT to supply a BOLUS at the minimum setting. Start the EQUIPMENT and weigh 25 successive BOLUS deliveries either demanded manually or by programme.

Calculate the mean and the percentage deviation from the set value. Select the deliveries with the maximum positive and maximum negative deviations from the set value. Express these as percentage deviations from the set value. Repeat the test with the EQUIPMENT at the maximum BOLUS setting.

50.107* Accuracy tests for pumps type 4

Pumps type 4 shall be tested according to 50.104, 50.105 and 50.106 as appropriate.

NOTE – Correction factors may be applicable to PUMPS FOR AMBULATORY USE type 4 where a continuous or quasi-continuous flow is maintained throughout the BOLUS delivery. These factors are disclosed in the ACCOMPANYING DOCUMENTS.

50.108* Accuracy tests for pumps type 5

Pumps type 5 shall be tested according to 50.102 to 50.106, as appropriate.

Table 102 – Set rates, BOLUS volumes and test apparatus for the accuracy tests of 50.102 to 50.108

EQUIPMENT	Set rates		Bolus		Test	
	Minimum	Inter-mediate	Minimum	Maximum	Apparatus (figure)	Sub-clause
DRIP-RATE INFUSION CONTROLLER	*	*			108	50.103
DRIP-RATE INFUSION PUMP	*	*			108	50.103
VOLUMETRIC INFUSION CONTROLLER	*	*			104a), 104b)	50.102
VOLUMETRIC INFUSION PUMP	*	*	*	*	104a), 104b)	50.102, (50.106)
SYRINGE PUMP	*	*	*	*	104b)	50.102, (50.106)
INFUSION PUMP FOR AMBULATORY USE						
Type 1	*	*			104b)	50.104
Type 2		*			104b)	50.105
DRIP-RATE, VOLUMETRIC, INFUSION PUMP, OF SYRINGE PUMP OR INFUSION PUMP FOR AMBULATORY USE						
Type 3			*	*	104a), 104b)	50.106
Type 4	*	*	*	*	104a), 104b)	50.104 and 50.106
Type 5	*	*	*	*	104a), 104b), 108	50.104 and 50.106

51 Protection against hazardous output

This clause of the General Standard applies except as follows:

51.1 Intentional exceeding of safety limits

Addition:

An example would be the priming/purge control of the EQUIPMENT.

Compliance is checked by inspection.

51.5 Incorrect output

Replacement:

a) Protection against overinfusion

Means shall be provided to prevent overinfusion under SINGLE FAULT CONDITIONS. An audible alarm shall be initiated in the event of overinfusion and the EQUIPMENT shall either cease delivery of infusion liquid or reduce the delivery rate to the KEEP OPEN RATE or less.

Compliance is checked by inspection and functional tests.

b)* Protection against overinfusion FREE FLOW conditions

Means shall be provided to protect the PATIENT from overinfusion as a result of FREE FLOW conditions. This requirement applies as soon as the ADMINISTRATION SET is installed in the EQUIPMENT in accordance with the manufacturer's instructions for use.

Remark: Refer also to 54.102 and 54.103.

Compliance is checked by inspection and functional tests, including, but not limited to, allowing the flow to stabilize the quick lowering of the collecting vessel by 50 cm and checking for evidence of FREE FLOW.

Additions:

51.101**a) maximum infusion pressure**

The EQUIPMENT shall not produce a MAXIMUM INFUSION PRESSURE capable of causing a rupture or a leak in the ADMINISTRATION SET.

Compliance is checked by inspection and functional tests.

b)* Protection against BOLUS volumes and occlusion

Means shall be provided to protect the PATIENT from BOLUS and underinfusion resulting from occlusion following activation of the occlusion alarm.

NOTE – An acceptable method of complying with this requirement is to activate an audible alarm and terminate the infusion liquid flow at the OCCLUSION ALARM THRESHOLD (PRESSURE).

Compliance is checked by the following test:

This test applies only to INFUSION PUMPS, VOLUMETRIC INFUSION PUMPS, DRIP-RATE INFUSION PUMPS, SPECIAL USE EQUIPMENT and SYRINGE PUMPS.

The test apparatus shown in figure 114 is used. Carry out the tests using a test solution of ISO class III water for medical use. Perform the test under normal conditions ($20\text{ }^{\circ}\text{C} \pm 2\text{ }^{\circ}\text{C}$, $65\% \pm 5\% \text{ RH}$). Operate the EQUIPMENT in NORMAL USE according to the manufacturer's instructions for use. Prime the ADMINISTRATION SET and the tubing connected to the pressure transducer.

Select the INTERMEDIATE RATE and the OCCLUSION ALARM THRESHOLD (PRESSURE) specified by the manufacturer, if the OCCLUSION ALARM THRESHOLD (PRESSURE) can be selected, set it to minimum. Connect the PATIENT END of the PATIENT LINE to the stopcock. Open the stopcock to the collecting vessel. Start the EQUIPMENT and allow the flow to become constant. Switch the stopcock and detect the OCCLUSION ALARM THRESHOLD (PRESSURE). Measure the time taken from switching the stopcock to activation of the occlusion alarm.

Inspect the ADMINISTRATION SET for ruptures or leaks. Empty the collecting vessel. Switch the stopcock and collect the BOLUS volume generated as a result of the occlusion until the pressure is reduced to atmospheric.

If the OCCLUSION ALARM THRESHOLD (PRESSURE) can be selected, repeat the test with it set to maximum.

If any OPERATOR action is given for 6.8.2 a) 12, a test shall be conducted of the means provided by the EQUIPMENT to release the BOLUS. This consists of performing the release as described before measuring the amount of the BOLUS remaining.

Verify by volume or mass that the result of the test is in accordance with the requirements of 51.5 a) and 51.5 b) and the disclosure statement in the ACCOMPANYING DOCUMENTS required by 6.8.2 a) 9) to 6.8.2 a) 12).

For INFUSION PUMPS FOR AMBULATORY USE carry out the following test:

The test apparatus shown in figure 114 is used. Carry out the tests using a test solution of ISO class III water for medical use. Perform the test under normal condition ($20\text{ }^{\circ}\text{C} \pm 2\text{ }^{\circ}\text{C}$, $65\% \pm 5\% \text{ RH}$). Operate the EQUIPMENT in NORMAL USE according to the manufacturer's instructions for use. Prime the ADMINISTRATION SET and the tubing connected to the pressure transducer.

Select the INTERMEDIATE RATE. Connect the PATIENT END of the PATIENT LINE to the pressure measuring system. Start the EQUIPMENT and detect the OCCLUSION ALARM THRESHOLD (PRESSURE).

Inspect the ADMINISTRATION SET for ruptures or leaks.

Stop the EQUIPMENT and disconnect the pressure measuring system. Vent the ADMINISTRATION SET to atmosphere. Close the PATIENT END of the PATIENT LINE. If the OCCLUSION ALARM THRESHOLD (PRESSURE) can be selected, set it to minimum. Restart the EQUIPMENT and wait until an occlusion alarm occurs or the EQUIPMENT stops. Collect the BOLUS volume generated as a result of the occlusion. If the OCCLUSION ALARM THRESHOLD (PRESSURE) can be selected, then repeat the test with it set to maximum. If any OPERATOR action is given for 6.8.2 a) 12), a test shall be conducted of the means provided by the EQUIPMENT to release the BOLUS. This consists of performing the release as described before measuring the amount of the BOLUS remaining. Verify by volume or mass that the result of the test is in accordance with the requirements of 51.101 a) and 51.101 b) and the disclosure statement in the ACCOMPANYING DOCUMENTS required by 6.8.2 a) 9) to 6.8.2 a) 12).




Figure 114 – Test apparatus to determine the OCCLUSION ALARM THRESHOLD (PRESSURE) and BOLUS volumes

51.102 Reverse delivery

During NORMAL USE and/or SINGLE FAULT CONDITION of the EQUIPMENT, continuous reverse delivery, which may cause a SAFETY HAZARD, shall not be possible.

Compliance is checked by inspection.

51.103 EQUIPMENT and drop sensor orientation

Safe operation of the EQUIPMENT shall not be affected by:

- the mispositioning or removal of a drop sensor, and
- operating the EQUIPMENT with a tilted or incorrectly filled drip chamber.

Under these conditions the EQUIPMENT shall either:

- maintain the accuracy of delivery, or
- stop the flow and generate an audible alarm.

Compliance is checked by the following functional test:

Operate the EQUIPMENT in NORMAL USE according to the manufacturer's instructions for use. Select any rate. Tilt the drip chamber from the vertical to a maximum of 20° in two orthogonal planes. By inspection determine the result of the test. By inspection determine the effects of mispositioning, removal or overfilling of a drip chamber.

51.104* Protection against air infusion

This requirement does not apply to SYRINGE PUMPS.

The EQUIPMENT shall protect the PATIENT from air infusion which may cause a SAFETY HAZARD due to air embolism.

Compliance is checked by inspection and functional tests in accordance with the manufacturer's specification (see 6.8.3 e)).

After the initiation of an air detection alarm it shall not be possible to recommence liquid delivery by a single action.

Compliance is checked by inspection and functional test.

51.105 ADMINISTRATION SETS – Operational characteristics

Should the manufacturer allow the use of a range of ADMINISTRATION SETS with different operational characteristics, then automatic means shall be provided or manual action(s) shall be necessary to prevent incorrect output.

Compliance is checked by inspection and functional test.

51.106 Audible and visual alarms

Unless specified elsewhere, the alarms required by this Particular Standard shall be so arranged that an audible alarm shall occur in all alarm situations.

This requirement does not apply to INFUSION PUMPS FOR AMBULATORY USE.

Compliance is checked by inspection and functional test.

51.107 Alarms required by clause 51 of this Particular Standard shall comply with the following. This requirement does not apply to INFUSION PUMPS FOR AMBULATORY USE:

- a) the audible alarm shall be able to produce a sound-pressure level (or, if adjustable a maximum level) of at least 65 dB(A) at 1 m, and shall not be by the OPERATOR externally adjustable below 45 dB(A) at 1 m;

- b) the audible alarm silence period of the EQUIPMENT in stand-alone operation shall not exceed 2 min;
- c) the visual alarm shall continue to operate during the audible alarm silence period;
- d) means shall be provided to enable the OPERATOR to check the operation of audible and visual alarms.

Compliance is checked by measuring the A-weighted sound pressure level, with an instrument complying with the requirements for a type 1 instrument laid down in IEC 60651 or IEC 60804, as follows.

The pump and the microphone are placed in free-field conditions (according to ISO 3744), at a height of 1,5 m from the reflecting plane. The distance between the pump and the microphone shall be 1 m. The background noise level shall be at least 10 dB(A) below the sound pressure level to be measured. During the test, microphone orientation should be toward, but in the lowest horizontal sound power direction from, the pump.

51.108 INFUSION PUMPS FOR AMBULATORY USE shall additionally include an alarm, if the EQUIPMENT is switched to a standby mode of operation for more than 1 h.

Compliance is checked by inspection and functional test.

51.109 Alarms required by 51.108, 51.110 and 49.2 shall comply with the following:

- a) the audible alarm shall produce a sound pressure level of at least 50 dB(A) at 1 m;
- b) the audible alarm output shall not be adjustable without either the use of a TOOL or by special means (e.g. pressing a sequence of switches);
- c) means shall be provided to enable the OPERATOR to check the operation of the alarms.

Compliance is checked by measuring the A-weighted sound power level, with an instrument complying with the requirements for a type 1 instrument laid down in IEC 60651 or IEC 60804, as follows.

The pump and the microphone are placed in free-field conditions (according to ISO 3744), at a height of 1,5 m from the reflecting plane. The distance between the pump and the microphone shall be 1 m. The background noise level shall be at least 10 dB(A) below the sound pressure level to be measured. During the test, microphone orientation should be toward, but in the lowest horizontal sound power direction from, the pump.

51.110 Audible means shall be provided to indicate to the OPERATOR the end of infusion.

This requirement does not apply to INFUSION PUMPS FOR AMBULATORY USE.

Compliance is checked by inspection and by functional test.

51.111 An audible warning shall be provided prior to the end of the infusion alarm.

This requirement applies only to SYRINGE PUMPS.

SECTION TEN – CONSTRUCTIONAL REQUIREMENTS

The clauses and subclauses of this section of the General Standard apply except as follows:

54 General

This clause of the General Standard applies, except as follows:

54.3 Inadvertent changing of settings

Replacement:

Means shall be provided to prevent accidental or unintended changes in rate settings.

Compliance is checked by inspection.

If manual means for priming/purging are provided, no single action by the OPERATOR shall initiate priming/purging to comply with the requirement of 51.1.

Compliance is checked by inspection and functional test.

Additions:

54.101 Fitting of the syringe

Means shall be provided to ensure correct clamping and location of a syringe barrel and plunger in the SYRINGE PUMP.

In the event of incorrect location of the plunger the SYRINGE PUMP shall not start.

Means shall be provided to prevent syphoning under SINGLE FAULT CONDITIONS.

An alarm shall be activated, if an attempt is made to remove the syringe while the SYRINGE PUMP is running.

EQUIPMENT shall be so designed that no SAFETY HAZARD to the PATIENT can occur due to pulling force on the PATIENT LINE.

Compliance is checked by inspection.

54.102 Fitting of the ADMINISTRATION SET

Where applicable, means shall be provided to ensure correct fitting of the ADMINISTRATION SET into the EQUIPMENT.

An alarm shall be activated, if an attempt is made to remove the ADMINISTRATION SET while the pump is running.

EQUIPMENT shall be so designed that no SAFETY HAZARD to the PATIENT can occur due to pulling force on the PATIENT LINE.

Compliance is checked by inspection.

54.103* Human errors

At least two distinctive and separate actions shall be required before FREE FLOW can occur in NORMAL USE. The first action shall stop the flow and initiate an audible alarm. This requirement does not apply to SYRINGE PUMPS and INFUSION PUMPS FOR AMBULATORY USE which use syringes.

Remark: Refer also to 51.5 b).

Compliance is checked by inspection and functional test.

54.104 EQUIPMENT shall be so designed that, if it is accidentally switched off and then switched on again by means of a functional control, the safety of the PATIENT shall be maintained.

Compliance is checked by inspection and functional test.

56 Components and general assembly

This clause of the General Standard applies, except as follows:

56.8 Indicators

Addition:

An indicator lamp (or means other than marking) shall be provided to indicate that the SUPPLY MAINS IS ON.

Compliance is checked by inspection.

IECNORM.COM : Click to view the full PDF of IEC 60601-2-24:1998

The appendices of the General Standard apply, except as follows:

Appendix L

References – Publications mentioned in this standard

Appendix L of the General Standard applies, except as follows:

Addition:

IEC 60521:1988, *Class 0.5, 1 and 2 alternating-current watthour meters*

IEC 60601-1:1988, *Medical electrical equipment – Part 1: General requirements for safety*
Amendment 1 (1991)
Amendment 2 (1995)

IEC 60601-1-2:1993, *Medical electrical equipment – Part 1: General requirements for safety – 2. Collateral Standard: Electromagnetic compatibility – Requirements and tests*

IEC 60651:1979, *Sound level meters*
Amendment 1 (1993)

IEC 60801-1:1984, *Electromagnetic compatibility for industrial-process measurement and control equipment – Part 1: General introduction*

IEC 60801-2:1991, *Electromagnetic compatibility for industrial-process measurement and control equipment – Part 2: Electrostatic discharge requirements*

IEC 60804:1985, *Integrating-averaging sound level meters*
Amendment 1 (1989)
Amendment 2 (1993)

IEC 61000-4-3:1995, *Electromagnetic compatibility (EMC) – Part 4: Testing and measurement techniques – Section 3: Radiated radio-frequency, electromagnetic field immunity test*

IEC 61000-4-4:1995, *Electromagnetic compatibility (EMC) – Part 4: Testing and measurement techniques – Section 4: Electrical fast transient/burst immunity test*

ISO 3696:1987, *Water for analytical laboratory use – Specification and test methods*

ISO 3744:1994, *Acoustics – Determination of sound power levels of noise sources using sound pressure – Engineering method in an essentially free field over a reflecting plane*

ISO 7864:1993, *Sterile hypodermic needles for single use*

ISO 7886-2:1996, *Sterile hypodermic syringes for single use – Part 2: Syringes for use with power-driven syringe pumps*

ISO 8536-4:1987, *Infusion equipment for medical use – Part 4: Infusion sets for single use*

The following Collateral Standards quoted in Amendment 2 of IEC 60601-1 do not apply:

IEC 60601-1-1:1992, *Medical electrical equipment – Part 1: General requirements for safety – 1. Collateral Standard: Safety requirements for medical electrical systems*
Amendment 1 (1995)

IEC 60601-1-3:1994, *Medical electrical equipment – Part 1: General requirements for safety – 3. Collateral Standard: General requirements for radiation protection in diagnostic X-ray equipment*

IEC 60601-1-4:1996, *Medical electrical equipment – Part 1: General requirements for safety – 4. Collateral Standard: Programmable electrical medical systems*

Withdrawing
IECNORM.COM : Click to view the full PDF of IEC 60601-2-24:1998

Annex AA (informative)

General guidance and rationale

AA.1 Rationale for the requirements of this Particular Standard

- 1.1 ADMINISTRATION SETS are not fully tested by this Particular Standard because they are outside its scope, but it is recognized that INFUSION PUMPS and CONTROLLERS can comply with this Particular Standard only if they are used together with compatible ADMINISTRATION SETS such as those recommended by the manufacturer. It is the responsibility of the OPERATOR to use such ADMINISTRATION SETS in order to avoid a SAFETY HAZARD resulting from the use of unsuitable ADMINISTRATION SETS. It is the responsibility of the manufacturer to recommend ADMINISTRATION SETS which comply with functional safety aspects.
- 3.6 In order to protect the PATIENT from a SAFETY HAZARD due to failure of the protective systems specified in clause 51, subclause 3.6 of this standard requires that SINGLE FAULT CONDITIONS occurring in these protective systems become immediately obvious to the OPERATOR while the EQUIPMENT is operational.
- One method of implementing this would be for the EQUIPMENT to continuously carry out self-check routines and alarm and stop infusing if a SINGLE FAULT CONDITION occurs (see example 3 of 3.6). However, it is recognized that this method might require expensive technology. Two other methods are, therefore, allowed. Example 2 allows the OPERATOR to initiate an automatic self-check procedure at any time before, during or after the infusion. Example 1 allows the OPERATOR to participate in an interactive procedure by following a safety check list described in the ACCOMPANYING DOCUMENTS.
- It is intended that, whichever method is employed, all primary sensors in the protective system should be included so that a true functional check is carried out.
- 6.8.2 a) 19) Examples of conditions under which the EQUIPMENT may fail to maintain the specified accuracy include short time periods, unusual infusion liquid characteristics, the use of excessively fine bore needles, inadequate protection against the extremes of environmental conditions, occlusion of the ADMINISTRATION SET upstream of the EQUIPMENT.
- 6.8.2 a) 20) Examples of a SAFETY HAZARD associated with interconnection of the infusion system or ACCESSORIES to the PATIENT LINE include the possible change in infusion rate due to such interconnections and the increased possibility of air infusion to the PATIENT, especially with gravity feed systems.
- 6.8.2 a) 28) The maximum infusion that may occur under SINGLE FAULT CONDITIONS may be declared as a percentage of the set rate or the BOLUS volume delivered before the EQUIPMENT stops.
- 44.4 Attention is drawn to the fact that leakage may occur from liquid reservoirs, ADMINISTRATION SETS and connectors above and in the EQUIPMENT, and that the liquid may be a viscous 50 % dextrose solution. Impairment of safety features due to leakage of this liquid may only occur after a period of time as the solution dries.

50.101 The ability of the EQUIPMENT to maintain the manufacturer's stated accuracy is the essential safety component of this requirement. This requirement for the EQUIPMENT does not take into account clinical criteria of the PATIENT, for example, age, weight, drugs used, etc.

Accuracy of these devices may be affected by extremes of back pressure.

50.102 to 50.108 Accuracy tests for INFUSION PUMPS and INFUSION CONTROLLERS

Data on performance following the start of infusion is important and must be shown by an unambiguous method so that the OPERATOR can select the appropriate EQUIPMENT to suit the clinical application. Graphs of the type shown in figures 105, 109 and 110 should be included in the INSTRUCTIONS OF USE. These graphs also give a good indication of the nature of the short-term flow fluctuations and are considered self-explanatory when studied with 50.102 to 50.105, as appropriate.

The type of presentation adopted enables OPERATORS to determine the start-up performance of the pump and the nature of its output, be it continuous, discontinuous, cyclical or otherwise. It is a matter of safety whether delivery starts in a reasonable time. OPERATORS will wish to be aware of likely delays in start-up and whether there are long periods of zero flow (or even reverse flow) during the pumping cycle.

Delays following start-up will vary greatly with:

- a) correct priming;
- b) backlash in the mechanism;
- c) the point at which a leadscrew is engaged (for SYRINGE PUMPS);
- d) set delivery rate;
- e) compliances within the syringe.

Following the attainment of the normal set delivery rate, it is important for OPERATORS to be aware of the short-term fluctuations in flow which may be expected from EQUIPMENT. Tests for this are conducted as described in 50.102 to 50.105 and example graphs are shown in figures 106, 107 and 111.

If these tests were carried out before delivery had stabilized, the results would normally be completely dependent on the first few minutes after start-up, and would give no useful information on expected performance at other times.

In establishing the accuracy of various pumps, the flow over a given period of time is measured. Parameters have been set to provide a safe standard to which the EQUIPMENT should comply. However, when the time interval over which the accuracy is measured is shortened, all pumps show considerable variations of flow pattern, for instance, on a minute-to-minute basis. This applies to all currently available EQUIPMENT: rotary and linear peristaltic, diaphragm and piston types, and even SYRINGE PUMPS. With certain EQUIPMENT it is possible to show errors of flow of $\pm 75\%$ over a 1 min cycle, and errors of $\pm 30\%$ over a 5 min cycle are not uncommon.

At the present time, certain drugs infused by such EQUIPMENT have a pharmacological and biological half-life of less than 5 min. For example, one of the agents commonly used to support the cardiac output in a severely ill PATIENT has a half-life of approximately 2,5 min. It is obvious that the use of such agents in concentrations which require low rates and where such demonstrated fluctuations occur, may lead to alarming and potentially dangerous responses by the PATIENT. It is therefore of vital importance that the OPERATOR is made aware that such fluctuations can occur so that he can make the necessary adjustments in both concentration and set delivery rate.

DRIP-RATE INFUSION CONTROLLERS are used only for intravenous infusions. They operate because the pressure created by the height of the liquid level in the container above the infusion site (usually about 90 cm H₂O = 8,83 kPa) is greater than the maximum venous pressure likely to be encountered in clinical practice (approximately 2,67 kPa (20 mm Hg)).

The maximum drip rate available with these devices is usually 100 drops per minute which, when using a 20 drops/ml set, is equivalent to a set rate of 300 ml/h. With an 18G, 1,2 mm needle 40 mm long, the pressure drop across the needle at 300 ml/h using water is approximately 3,33 kPa (2,5 mm Hg). With higher viscosity liquids, such as dextrose (50 %), these figures increase to 0,43 kPa (3,2 mm Hg) (with an 18G, 1,2 mm needle 40 mm long) and 2,86 kPa (21,4 mm Hg) (with a 21G, 0,8 mm needle 40 mm long) respectively.

In clinical practice, it would be inadvisable to attempt to use higher viscosity liquids or smaller gauge needles. Thus, the tests specified will allow realistic testing of the performance of the EQUIPMENT.

With DRIP-RATE INFUSION PUMPS it is necessary to investigate the effects of liquid viscosities on drip-rate accuracy as DRIP-RATE INFUSION PUMPS will be expected to pump a range of fluids of differing viscosities. The most viscous fluid, dextrose 50 %, will be pumped at drip rates not greater than an equivalent flow of 150 ml/h. If a 20 drops per ml ADMINISTRATION SET is used, this is a drip-rate of 50 drops per min. In order to simplify the test procedure and obviate the use of different infusion liquids, it has been decided to reproduce the additional back pressure that high viscosity liquids (such as dextrose 50 %) would produce by the use of fine gauge needles. Testing is carried out at the two extremes of back pressures, -13,33 kPa (-100 mm Hg) and +13,33 kPa (+100 mm Hg). An 18G, 1,2 mm needle 40 mm long (which produces no significant additional back pressure at any flow likely to be encountered) is used at -13,33 kPa (-100 mm Hg) and a 21G, 0,8 mm needle 40 mm long, at +13,33 kPa (+100 mm Hg). The use of a 21G, 0,8 mm needle 40 mm long produces additional back pressure that can be calculated by the Hagen-Poiseuille formula.

VOLUMETRIC INFUSION CONTROLLERS are similar to DRIP-RATE INFUSION CONTROLLERS in that they use gravity to supply the required infusion pressure. However, these devices are calibrated in volumetric units, for example, millilitres per hour (ml/h) and, although they count drops, they attempt to convert drops to volumes. This may be accomplished by the use of a special drop-forming orifice in the drip chamber and/or the use of liquid codes (programmed by the OPERATOR) to take account of the different characteristics of various solutions used in intravenous therapy. The volume of a drop is dependent on a number of factors which include drip rate, temperature, pressure, the material and condition of the drop-forming orifice, viscosity and surface tension of the liquid used. However, as the purpose of the test is to ensure that the infusion output is consistent with the selected value, tests carried out using ISO class III water for medical use and at the extremes of back pressure (negative back pressure only) are satisfactory.

VOLUMETRIC INFUSION PUMPS are designed to deliver precise volumes of liquids at medium and high set rates and shall be capable of pumping intravenously and using various sizes of needle and all types of liquids.

Variants of these pumps to cater for paediatric applications are designed to deliver precise volumes at low set rates (between 1 ml/h and 10 ml/h) and are calibrated in 0,1 ml/h increments. It is not considered necessary to test these pumps for accuracy of delivery below 1 ml/h because clinical applications would call for the use of a SYRINGE PUMP in these circumstances.

Because high-viscosity liquids may be used, these pumps are tested over the full range of set rates using water at a test back pressure of +13,33 kPa (+100 mm Hg) with a 21G, 0,8 mm needle 40 mm long to simulate the additional back pressure caused by the use of a high viscosity liquid such as dextrose 50 % (see Hagen-Poiseuille formula and calculations). Testing at –13,33 kPa (–100 mm Hg) is to simulate the negative back pressures that are sometimes encountered in clinical usage.

51.5 b) PATIENT movement has been known to cause FREE FLOW. During testing this may be investigated by allowing the flow to stabilize, then quickly lowering the collecting vessel 50 cm and checking for evidence of FREE FLOW. The above simulates PATIENT movement.

51.101 b) IEC 60601-1 allows testing to be carried out at temperatures between +10 °C and +40 °C. Test houses should be aware of the effects of the extremes of temperature on BOLUS volumes generated as a result of this test.

51.104 Infusion of 1 ml of air within 15 min is not considered to be a SAFETY HAZARD. Bubbles of less than 50 µl of air each are omitted in summing up the 1 ml.

54.103 Acceptable methods of maintaining PATIENT safety are either to maintain the previously selected mode of operation and set rate, or to cease delivery and initiate an audible alarm.

A functional control is one which is designed to either start or stop infusion, and may be separate from or combined with the mains switch.

The quality of insulation is investigated in clauses 17, 19, 20 and subclause 57.10.

AA.2 Rationale for the requirements of IEC 60601-1-2

36 It is well known that strong electromagnetic fields can interfere with electronic EQUIPMENT. INFUSION PUMPS and CONTROLLERS are known to be affected in the same way. In particular, there have been reports of interference from radio transmitters in ambulances and from electromagnetic fields generated by diathermy EQUIPMENT. The increased use of mobile telephones, particularly those operating at frequencies of 450 MHz and 900 MHz, has likewise been shown to cause problems with the operation of INFUSION PUMPS and CONTROLLERS.

As the use of automatic INFUSION PUMPS escalates in new environments which are often alien to the sensitive electronic circuitry employed, it is important for the OPERATOR to be aware of the possible hazards which may arise. Examples of such SAFETY HAZARDS include unpredictable cessation of infusion and reversion to a purge mode of operation.

It is also important that the manufacturer is aware that external interference may also change or destroy internal or external feedback loops that regulate various physical variables within the EQUIPMENT. A known example of this is an oscillating action of the infusion mechanism due to such external interference. Generally, the EQUIPMENT reacts in an unpredictable manner.

36.201 It is important to realise that for emissions not only CISPR 11 is required by the Collateral Standard IEC 60601-1-2, but also CISPR 14. CISPR 14 is mandatory because it is mentioned in the list with normative references in annex BBB of the Collateral Standard.