

Standard means for the reporting of the acoustic output
of medical diagnostic ultrasonic equipment

CORRIGENDUM 1

Page 6

3 Terms, definitions and symbols

3.7

beam axis

In the bracketed note on the origin of the definition, instead of:

[IEV 62127-1, definition 3.8 modified]

read:

[IEC 62127-1, definition 3.8 modified]

3.9

beamwidth midpoint

Delete the right bracket at the end of the note.

3.12

external transducer aperture

Replace “NOTE 1” with “NOTE”.

3.43

ultrasonic transducer element group dimensions

Delete “NOTE 2” and replace “NOTE 1” with “NOTE”.

Page 15

Table 1 – List of symbols

Add the following symbol, term and reference:

f_{awf}

zero-crossing acoustic-working frequency

IEC 62127-1

Page 16

4 Requirements

4.1 General

After the 2nd bullet point, replace “NOTE” with “NOTE 1”.

After the 5th bullet point, replace “NOTE” with “NOTE 2”.