



IEC 62366-1

Edition 1.0 2015-02

# INTERNATIONAL STANDARD

## NORME INTERNATIONALE

**Medical devices –  
Part 1: Application of usability engineering to medical devices**

**Dispositifs médicaux –  
Partie 1: Application de l'ingénierie de l'aptitude à l'utilisation aux dispositifs médicaux**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

ICS 11.040

ISBN 978-2-8322-2281-2

**Warning! Make sure that you obtained this publication from an authorized distributor.  
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

## CONTENTS

|                                                                                                   |    |
|---------------------------------------------------------------------------------------------------|----|
| FOREWORD .....                                                                                    | 4  |
| INTRODUCTION .....                                                                                | 6  |
| 1 * Scope .....                                                                                   | 7  |
| 2 Normative references .....                                                                      | 7  |
| 3 Terms and definitions .....                                                                     | 7  |
| 4 Principles .....                                                                                | 12 |
| 4.1 General requirements .....                                                                    | 12 |
| 4.1.1 * USABILITY ENGINEERING PROCESS .....                                                       | 12 |
| 4.1.2 * RISK CONTROL as it relates to USER INTERFACE design .....                                 | 13 |
| 4.1.3 Information for SAFETY as it relates to USABILITY .....                                     | 13 |
| 4.2 * USABILITY ENGINEERING FILE .....                                                            | 14 |
| 4.3 Tailoring of the USABILITY ENGINEERING effort .....                                           | 14 |
| 5 * USABILITY ENGINEERING PROCESS .....                                                           | 14 |
| 5.1 * Prepare USE SPECIFICATION .....                                                             | 14 |
| 5.2 * Identify USER INTERFACE characteristics related to SAFETY and potential USE<br>ERRORS ..... | 15 |
| 5.3 * Identify known or foreseeable HAZARDS and HAZARDOUS SITUATIONS .....                        | 15 |
| 5.4 * Identify and describe HAZARD-RELATED USE SCENARIOS .....                                    | 15 |
| 5.5 * Select the HAZARD-RELATED USE SCENARIOS for SUMMATIVE EVALUATION .....                      | 16 |
| 5.6 * Establish USER INTERFACE SPECIFICATION .....                                                | 16 |
| 5.7 * Establish USER INTERFACE EVALUATION plan .....                                              | 16 |
| 5.7.1 General .....                                                                               | 16 |
| 5.7.2 * FORMATIVE EVALUATION planning .....                                                       | 17 |
| 5.7.3 * SUMMATIVE EVALUATION planning .....                                                       | 17 |
| 5.8 * Perform USER INTERFACE design, implementation and FORMATIVE<br>EVALUATION .....             | 18 |
| 5.9 * Perform SUMMATIVE EVALUATION of the USABILITY of the USER INTERFACE .....                   | 19 |
| 5.10 USER INTERFACE OF UNKNOWN PROVENANCE .....                                                   | 19 |
| Annex A (informative) General guidance and rationale .....                                        | 20 |
| A.1 General guidance .....                                                                        | 20 |
| A.2 Rationale for requirements in particular clauses and subclauses .....                         | 20 |
| ANNEX B (informative) Examples of possible HAZARDOUS SITUATIONS related to<br>USABILITY .....     | 38 |
| Annex C (normative) Evaluation of a USER INTERFACE OF UNKNOWN PROVENANCE (UOUP) .....             | 41 |
| C.1 General .....                                                                                 | 41 |
| C.2 USABILITY ENGINEERING PROCESS for USER INTERFACE OF UNKNOWN<br>PROVENANCE .....               | 41 |
| C.2.1 * USE SPECIFICATION .....                                                                   | 41 |
| C.2.2 * Review of POST-PRODUCTION information .....                                               | 42 |
| C.2.3 HAZARDS and HAZARDOUS SITUATIONS related to USABILITY .....                                 | 42 |
| C.2.4 RISK CONTROL .....                                                                          | 42 |
| C.2.5 RESIDUAL RISK evaluation .....                                                              | 42 |
| Annex D (informative) Types of MEDICAL DEVICE use, with examples .....                            | 43 |
| Annex E (informative) Reference to the essential principles .....                                 | 45 |
| Bibliography .....                                                                                | 46 |
| Index of defined terms .....                                                                      | 49 |

|                                                                                                                                              |    |
|----------------------------------------------------------------------------------------------------------------------------------------------|----|
| Figure 1 – Relationship of the types of use .....                                                                                            | 8  |
| Figure A.1 – Model of USER-MEDICAL DEVICE interaction .....                                                                                  | 24 |
| Figure A.2 – Relationship of TASKS and functions within a USE SCENARIO .....                                                                 | 25 |
| Figure A.3 – Relationship of TASKS and functions and USE ERROR within a HAZARD-RELATED USE SCENARIO .....                                    | 26 |
| Figure A.4 – The relationship between the RISK MANAGEMENT PROCESS (ISO 14971:2007) and the USABILITY ENGINEERING PROCESS (IEC 62366-1) ..... | 32 |
| Figure D.1 – Interrelationships between the different types of MEDICAL DEVICE use, with examples .....                                       | 44 |
| Table B.1 – Glossary of relevant RISK MANAGEMENT terms .....                                                                                 | 38 |
| Table B.2 – Examples of HARM due to RISK caused by USE ERROR(S) or poor USABILITY (1 of 3) .....                                             | 38 |
| Table E.1 – Correspondence between this document and the essential principles .....                                                          | 45 |

# INTERNATIONAL ELECTROTECHNICAL COMMISSION

## MEDICAL DEVICES –

### Part 1: Application of usability engineering to medical devices

#### FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 62366-1 has been prepared by a joint working group of subcommittee 62A: Common aspects of electrical medical equipment used in medical practice, of IEC technical committee 62: Electrical medical equipment in medical practice, and ISO technical committee 210: Quality management and corresponding general aspects for MEDICAL DEVICES.

It is published as double logo standard.

This first edition of IEC 62366-1, together with the first edition of IEC 62366-2, cancels and replaces the first edition of IEC 62366 published in 2007 and its Amendment 1 (2014).

Part 1 has been updated to include contemporary concepts of USABILITY ENGINEERING, while also streamlining the process. It strengthens links to ISO 14971:2007 and the related methods of RISK MANAGEMENT as applied to SAFETY related aspects of medical device user interfaces. Part 2 contains tutorial information to assist manufactures in complying with Part 1, as well as offering more detailed descriptions of USABILITY ENGINEERING methods that can be applied

more generally to MEDICAL DEVICES that go beyond safety-related aspects of MEDICAL DEVICE USER INTERFACES.

The text of this standard is based on the following documents:

| FDIS         | Report on voting |
|--------------|------------------|
| 62A/977/FDIS | 62A/988/RVD      |

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table. In ISO, the standard has been approved by 26 P-members out of 26 having cast a vote.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this International Standard, the following print types are used:

- Requirements and definitions: roman type.
- *Means to assess compliance: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type
- TERMS DEFINED IN CLAUSE 3 OR AS NOTED: SMALL CAPITALS.

The requirements are followed by means to assess compliance.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

Clauses and subclauses for which a rationale is provided in informative Annex A are marked with an asterisk (\*).

A list of all parts of the IEC 62366 series, published under the general title *Medical devices*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under “http://webstore.iec.ch” in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of National Committees and Member Bodies is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC or ISO publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for mandatory implementation nationally not earlier than 3 years from the date of publication.

## INTRODUCTION

Medical practice is increasingly using MEDICAL DEVICES for observation and treatment of PATIENTS. USE ERRORS caused by inadequate MEDICAL DEVICE USABILITY have become an increasing cause for concern. Many of the MEDICAL DEVICES developed without applying a USABILITY ENGINEERING (HUMAN FACTORS ENGINEERING) PROCESS are non-intuitive, difficult to learn and difficult to use. As healthcare evolves, less skilled USERS including PATIENTS themselves are now using MEDICAL DEVICES and MEDICAL DEVICES are becoming more complicated. The design of the USER INTERFACE to achieve adequate USABILITY requires a different PROCESS and skill set than that of the technical implementation of the USER INTERFACE.

The USABILITY ENGINEERING PROCESS is intended to identify and minimise USE ERRORS and thereby reduce use-associated RISKS. Some, but not all, forms of incorrect use are suited to control by the MANUFACTURER. The USABILITY ENGINEERING PROCESS is related to the RISK MANAGEMENT PROCESS as indicated in Figure A.4.

This International Standard describes a USABILITY ENGINEERING PROCESS to provide acceptable RISK related to USABILITY of a MEDICAL DEVICE. It is intended to be useful not only for MANUFACTURERS of MEDICAL DEVICES, but also for technical committees responsible for the preparation of particular MEDICAL DEVICE standards.

This International Standard strictly focuses on applying the USABILITY ENGINEERING PROCESS to optimize MEDICAL DEVICE USABILITY as it relates to SAFETY. The companion technical report (IEC 62366-2<sup>1</sup>) is comprehensive and has a broader focus. It focuses not only on USABILITY as it relates to SAFETY, but also on how USABILITY relates to attributes such as TASK accuracy, completeness and EFFICIENCY, and USER satisfaction.

NOTE SAFETY is freedom from unacceptable RISK. Unacceptable RISK can arise from USE ERROR, which can lead to exposure to direct physical HAZARDS or loss or degradation of clinical functionality.

MANUFACTURERS can choose to implement a USABILITY ENGINEERING program focused narrowly on SAFETY or more broadly on SAFETY and other attributes, such as those cited above. A broader focus might also be useful to address specific USABILITY ENGINEERING expectations, such as the need to confirm that USERS can successfully perform non-SAFETY-related TASKS. A MANUFACTURER might also implement a broader program to realize the commercial benefits of a MEDICAL DEVICE that not only is safe to use but also offers superior USABILITY.

<sup>1</sup> IEC 62366-2, *Medical devices – Part 2: Guidance on the application of usability engineering to medical devices* (in preparation).

## MEDICAL DEVICES –

### Part 1: Application of usability engineering to medical devices

#### 1 \* Scope

This part of IEC 62366 specifies a PROCESS for a MANUFACTURER to analyse, specify, develop and evaluate the USABILITY of a MEDICAL DEVICE as it relates to SAFETY. This USABILITY ENGINEERING (HUMAN FACTORS ENGINEERING) PROCESS permits the MANUFACTURER to assess and mitigate RISKS associated with CORRECT USE and USE ERRORS, i.e., NORMAL USE. It can be used to identify but does not assess or mitigate RISKS associated with ABNORMAL USE.

NOTE 1 SAFETY is freedom from unacceptable RISK. Unacceptable RISK can arise from USE ERROR, which can lead to exposure to direct physical HAZARDS or loss or degradation of clinical functionality.

NOTE 2 Guidance on the application of USABILITY ENGINEERING to MEDICAL DEVICES is available in IEC 62366-2<sup>2</sup>, which addresses not only SAFETY but also aspects of USABILITY not related to SAFETY.

If the USABILITY ENGINEERING PROCESS detailed in this International Standard has been complied with, then the USABILITY of a MEDICAL DEVICE as it relates to SAFETY is presumed to be acceptable, unless there is OBJECTIVE EVIDENCE to the contrary.

NOTE 3 Such OBJECTIVE EVIDENCE can subsequently originate from POST-PRODUCTION surveillance.

#### 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

NOTE 2 Informative references are listed in the bibliography beginning on page 46.

ISO 14971:2007, *Medical devices – Application of risk management to medical devices*

<sup>2</sup> IEC 62366-2, *Medical devices – Part 2: Guidance on the application of usability engineering to medical devices* (in preparation).

## SOMMAIRE

|                                                                                                                                             |           |
|---------------------------------------------------------------------------------------------------------------------------------------------|-----------|
| AVANT-PROPOS .....                                                                                                                          | 54        |
| INTRODUCTION .....                                                                                                                          | 57        |
| 1 * Domaine d'application .....                                                                                                             | 58        |
| 2 Références normatives .....                                                                                                               | 58        |
| 3 Termes et définitions .....                                                                                                               | 58        |
| 4 Principes .....                                                                                                                           | 64        |
| 4.1 Exigences générales .....                                                                                                               | 64        |
| 4.1.1 * PROCESSUS D'INGENIERIE DE L'APTITUDE A L'UTILISATION .....                                                                          | 64        |
| 4.1.2 * MAITRISE DU RISQUE concernant la conception de l'INTERFACE<br>UTILISATEUR .....                                                     | 65        |
| 4.1.3 Informations relatives à la SECURITE concernant l'APTITUDE A<br>L'UTILISATION .....                                                   | 65        |
| 4.2 * DOSSIER D'INGENIERIE DE L'APTITUDE A L'UTILISATION .....                                                                              | 65        |
| 4.3 Adaptation de l'effort d'INGENIERIE DE L'APTITUDE A L'UTILISATION .....                                                                 | 66        |
| 5 * PROCESSUS D'INGENIERIE DE L'APTITUDE A L'UTILISATION .....                                                                              | 66        |
| 5.1 * Préparer la SPECIFICATION D'UTILISATION .....                                                                                         | 66        |
| 5.2 * Identifier les caractéristiques de l'INTERFACE UTILISATEUR relatives à la<br>SECURITE et les ERREURS D'UTILISATION potentielles ..... | 66        |
| 5.3 * Identifier les PHENOMENES DANGEREUX et SITUATIONS DANGEREUSES connus<br>ou prévisibles .....                                          | 67        |
| 5.4 * Identifier et décrire les SCENARIOS D'UTILISATION RELATIFS AU PHENOMENE<br>DANGEREUX .....                                            | 67        |
| 5.5 * Sélection du SCENARIO D'UTILISATION RELATIF AU PHENOMENE DANGEREUX<br>pour l'EVALUATION SOMMATIVE .....                               | 67        |
| 5.6 * Établir la SPECIFICATION DE L'INTERFACE UTILISATEUR .....                                                                             | 68        |
| 5.7 * Établir le plan d'EVALUATION DE L'INTERFACE UTILISATEUR .....                                                                         | 68        |
| 5.7.1 Généralités .....                                                                                                                     | 68        |
| 5.7.2 * PLANIFICATION DE L'EVALUATION FORMATIVE .....                                                                                       | 69        |
| 5.7.3 * Planification de l'EVALUATION SOMMATIVE .....                                                                                       | 70        |
| 5.8 * Effectuer la conception de l'INTERFACE UTILISATEUR, sa mise en œuvre et<br>son EVALUATION FORMATIVE .....                             | 70        |
| 5.9 * Effectuer l'EVALUATION SOMMATIVE de l'APTITUDE A L'UTILISATION de<br>l'INTERFACE UTILISATEUR .....                                    | 71        |
| <b>5.10 INTERFACE UTILISATEUR D'ORIGINE INCONNUE .....</b>                                                                                  | <b>72</b> |
| Annexe A (informative) Lignes directrices générales et justifications .....                                                                 | 73        |
| A.1 Lignes directrices générales .....                                                                                                      | 73        |
| A.2 Justifications pour les exigences dans des articles et paragraphes<br>particuliers .....                                                | 73        |
| ANNEXE B (informative) Exemples de SITUATIONS DANGEREUSES possibles relatives à<br>l'APTITUDE A L'UTILISATION .....                         | 97        |
| Annexe C (normative) Évaluation d'une INTERFACE UTILISATEUR D'ORIGINE INCONNUE<br>(UOUP) .....                                              | 100       |
| C.1 Généralités .....                                                                                                                       | 100       |
| C.2 PROCESSUS D'INGENIERIE DE L'APTITUDE A L'UTILISATION pour l'INTERFACE<br>UTILISATEUR D'ORIGINE INCONNUE .....                           | 101       |
| C.2.1 * SPECIFICATION D'UTILISATION .....                                                                                                   | 101       |
| C.2.2 * Examen des informations relatives à la POSTPRODUCTION .....                                                                         | 101       |

|                                                                                                                                                                |                                                                                        |     |
|----------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------|-----|
| C.2.3                                                                                                                                                          | PHENOMENES DANGEREUX et SITUATIONS DANGEREUSES liés à l'APTITUDE A L'UTILISATION ..... | 101 |
| C.2.4                                                                                                                                                          | MAITRISE DU RISQUE .....                                                               | 101 |
| C.2.5                                                                                                                                                          | Évaluation du RISQUE RESIDUEL .....                                                    | 101 |
| Annexe D (informative) Types d'utilisation de DISPOSITIF MEDICAL avec exemples .....                                                                           |                                                                                        | 102 |
| Annexe E (informative) Référence aux principes essentiels .....                                                                                                |                                                                                        | 105 |
| Bibliographie.....                                                                                                                                             |                                                                                        | 106 |
| Index des termes définis .....                                                                                                                                 |                                                                                        | 109 |
| Figure 1 – Relation entre les types d'utilisation .....                                                                                                        |                                                                                        | 59  |
| Figure A.1 – Modèle d'interaction UTILISATEUR-DISPOSITIF MEDICAL.....                                                                                          |                                                                                        | 78  |
| Figure A.2 – Relations des TACHES et des fonctions dans un SCENARIO D'UTILISATION.....                                                                         |                                                                                        | 80  |
| Figure A.3 – Relations des TACHES, des fonctions et des ERREURS D'UTILISATION dans un SCENARIO D'UTILISATION RELATIF AU PHENOMENE DANGEREUX .....              |                                                                                        | 81  |
| Figure A.4 – Relation entre le PROCESSUS DE GESTION DES RISQUES (ISO 14971:2007) et le PROCESSUS D'INGENIERIE DE L'APTITUDE A L'UTILISATION (IEC 62366-1)..... |                                                                                        | 89  |
| Figure D.1 – Interrelations entre les différents types d'utilisation de DISPOSITIF MEDICAL avec exemples.....                                                  |                                                                                        | 104 |
| Tableau B.1 – Glossaire des termes relatifs à la GESTION DES RISQUES .....                                                                                     |                                                                                        | 97  |
| Tableau B.2 – Exemples de DOMMAGES dus à des RISQUES causés par une ou plusieurs ERREURS D'UTILISATION ou une faible APTITUDE A L'UTILISATION (1 de 3).....    |                                                                                        | 97  |
| Tableau E.1 – Correspondance entre le présent document et les principes essentiels .....                                                                       |                                                                                        | 105 |

## COMMISSION ÉLECTROTECHNIQUE INTERNATIONALE

### DISPOSITIFS MÉDICAUX –

#### Partie 1: Application de l'ingénierie de l'aptitude à l'utilisation aux dispositifs médicaux

##### AVANT-PROPOS

- 1) La Commission Electrotechnique Internationale (IEC) est une organisation mondiale de normalisation composée de l'ensemble des comités électrotechniques nationaux (Comités nationaux de l'IEC). L'IEC a pour objet de favoriser la coopération internationale pour toutes les questions de normalisation dans les domaines de l'électricité et de l'électronique. À cet effet, l'IEC – entre autres activités – publie des Normes internationales, des Spécifications techniques, des Rapports techniques, des Spécifications accessibles au public (PAS) et des Guides (ci-après dénommés "Publication(s) de l'IEC"). Leur élaboration est confiée à des comités d'études, aux travaux desquels tout Comité national intéressé par le sujet traité peut participer. Les organisations internationales, gouvernementales et non gouvernementales, en liaison avec l'IEC, participent également aux travaux. L'IEC collabore étroitement avec l'Organisation Internationale de Normalisation (ISO), selon des conditions fixées par accord entre les deux organisations.
- 2) Les décisions ou accords officiels de l'IEC concernant les questions techniques représentent, dans la mesure du possible, un accord international sur les sujets étudiés, étant donné que les Comités nationaux de l'IEC intéressés sont représentés dans chaque comité d'études.
- 3) Les Publications de l'IEC se présentent sous la forme de recommandations internationales et sont agréées comme telles par les Comités nationaux de l'IEC. Tous les efforts raisonnables sont entrepris afin que l'IEC s'assure de l'exactitude du contenu technique de ses publications; l'IEC ne peut pas être tenue responsable de l'éventuelle mauvaise utilisation ou interprétation qui en est faite par un quelconque utilisateur final.
- 4) Dans le but d'encourager l'uniformité internationale, les Comités nationaux de l'IEC s'engagent, dans toute la mesure possible, à appliquer de façon transparente les Publications de l'IEC dans leurs publications nationales et régionales. Toutes divergences entre toutes Publications de l'IEC et toutes publications nationales ou régionales correspondantes doivent être indiquées en termes clairs dans ces dernières.
- 5) L'IEC elle-même ne fournit aucune attestation de conformité. Des organismes de certification indépendants fournissent des services d'évaluation de conformité et, dans certains secteurs, accèdent aux marques de conformité de l'IEC. L'IEC n'est responsable d'aucun des services effectués par les organismes de certification indépendants.
- 6) Tous les utilisateurs doivent s'assurer qu'ils sont en possession de la dernière édition de cette publication.
- 7) Aucune responsabilité ne doit être imputée à l'IEC, à ses administrateurs, employés, auxiliaires ou mandataires, y compris ses experts particuliers et les membres de ses comités d'études et des Comités nationaux de l'IEC, pour tout préjudice causé en cas de dommages corporels et matériels, ou de tout autre dommage de quelque nature que ce soit, directe ou indirecte, ou pour supporter les coûts (y compris les frais de justice) et les dépenses découlant de la publication ou de l'utilisation de cette Publication de l'IEC ou de toute autre Publication de l'IEC, ou au crédit qui lui est accordé.
- 8) L'attention est attirée sur les références normatives citées dans cette publication. L'utilisation de publications référencées est obligatoire pour une application correcte de la présente publication.
- 9) L'attention est attirée sur le fait que certains des éléments de la présente Publication de l'IEC peuvent faire l'objet de droits de brevet. L'IEC ne saurait être tenue pour responsable de ne pas avoir identifié de tels droits de brevets et de ne pas avoir signalé leur existence.

La Norme internationale IEC 62366-1 a été établie par un groupe de travail mixte du sous-comité 62A: Aspects généraux des équipements électriques utilisés en pratique médicale, du comité d'études 62 de l'IEC: Équipements électriques dans la pratique médicale, et du comité technique 210 de l'ISO: Management de la qualité et aspects généraux correspondants des DISPOSITIFS MEDICAUX.

Elle est publiée sous double logo.

La première édition de l'IEC 62366-1, ainsi que la première édition de l'IEC 62366-2, annule et remplace la première édition de l'IEC 62366 parue en 2007 et son Amendement 1 (2014).

La Partie 1 a été mise à jour afin d'inclure des concepts contemporains d'INGENIERIE DE L'APTITUDE A L'UTILISATION, tout en rationalisant le processus. Elle renforce aussi les liens avec l'ISO 14971:2007 et les méthodes connexes de GESTION DES RISQUES appliquées aux aspects

relatifs à la SECURITE des INTERFACES UTILISATEUR des DISPOSITIFS MEDICAUX. La Partie 2 contient des informations didactiques pour aider les fabricants à se conformer à la Partie 1 et fournit des descriptions plus détaillées des méthodes d'INGENIERIE DE L'APTITUDE A L'UTILISATION qui peuvent être appliquées d'une façon plus générale aux DISPOSITIFS MEDICAUX et qui vont au-delà des aspects relatifs à la sécurité des INTERFACES UTILISATEUR des DISPOSITIFS MEDICAUX.

Le texte de cette norme est issu des documents suivants:

| FDIS         | Rapport de vote |
|--------------|-----------------|
| 62A/977/FDIS | 62A/988/RVD     |

Le rapport de vote indiqué dans le tableau ci-dessus donne toute information sur le vote ayant abouti à l'approbation de cette norme. A l'ISO, la norme a été approuvée par 26 membres P sur 26 ayant voté.

Cette publication a été rédigée selon les Directives ISO/IEC, Partie 2.

Dans la présente Norme internationale, les caractères d'imprimerie suivants sont utilisés:

- Exigences et définitions: caractères romains.
- *Moyens d'évaluation de la conformité: caractères italiques.*
- Indications de nature informative apparaissant hors des tableaux, comme les notes, les exemples et les références: petits caractères romains. Le texte normatif à l'intérieur des tableaux est également en petits caractères
- TERMES DEFINIS A L'ARTICLE 3 OU COMME NOTES: PETITES MAJUSCULES.

Les exigences sont suivies par des moyens d'évaluation de la conformité.

Dans la présente norme, la conjonction "ou" est utilisée avec la valeur d'un "ou inclusif"; ainsi un énoncé est vrai si une combinaison des conditions quelle qu'elle soit est vraie.

Les formes verbales utilisées dans la présente norme sont conformes à l'usage donné à l'Annexe H des Directives ISO/IEC, Partie 2. Pour les besoins de la présente norme, l'auxiliaire:

- "devoir" mis au présent de l'indicatif signifie que la satisfaction à une exigence ou à un essai est obligatoire pour la conformité à la présente norme;
- "il convient" signifie que la satisfaction à une exigence ou à un essai est recommandée mais n'est pas obligatoire pour la conformité à la présente norme;
- "pouvoir" mis au présent de l'indicatif est utilisé pour décrire un moyen admissible pour satisfaire à une exigence ou à un essai.

Les articles et paragraphes pour lesquels une justification est donnée dans l'Annexe A informative sont repérés par un astérisque (\*).

Une liste de toutes les parties de la série CEI 62366, publiée sous le titre général *Dispositifs médicaux*, figure sur le site web de l'IEC.

Le comité a décidé que le contenu de cette publication ne sera pas modifié avant la date de stabilité indiquée sur le site web de l'IEC sous "<http://webstore.iec.ch>" dans les données relatives à la publication recherchée. À cette date, la publication sera

- reconduite,
- supprimée,
- remplacée par une édition révisée, ou
- amendée.