

**IEC 80601-2-35**  
(First edition – 2009)

**Medical electrical equipment –**

**Part 2-35: Particular requirements for the basic safety  
and essential performance of heating devices using blankets,  
pads or mattresses and intended for heating in medical use**

**CORRIGENDUM 1**

**201.4.3.101 Additional requirements for ESSENTIAL PERFORMANCE**

*In Table 201.101, replace the reference to "201.12.3.103" by "201.12.3.102" as follows.*

**Table 201.101 – \*Additional ESSENTIAL PERFORMANCE requirements**

| Requirement                         | Subclause                                                                                 |
|-------------------------------------|-------------------------------------------------------------------------------------------|
| ESSENTIAL PERFORMANCE requirement 1 | 201.12.4.104 or generation of a TECHNICAL ALARM CONDITION in compliance with 201.12.3.102 |

**201.15.4.2.1.101 \*Temperature and overload control devices – Additional requirements  
for application**

*In the fifth paragraph, replace the reference to "42 °C" by "41 °C".*

**201.15.4.101.1 Securing the fixation of heating elements and internal wiring**

*In the fifth paragraph, replace the reference to "201.15.4.1.102.2" by "201.15.4.101.2".*