

IEC 80601-2-35
(Second edition – 2009)

Medical electrical equipment –

**Part 2-35: Particular requirements for the basic safety
and essential performance of heating devices using blankets,
pads or mattresses and intended for heating in medical use**

CORRIGENDUM 1

201.4.3.101 Additional requirements for ESSENTIAL PERFORMANCE

In Table 201.101, replace the reference to "201.12.3.103" by "201.12.3.102" as follows:

Table 201.101 – *Additional ESSENTIAL PERFORMANCE requirements

Requirement	Subclause
ESSENTIAL PERFORMANCE requirement 1	201.12.4.104 or generation of a TECHNICAL ALARM CONDITION in compliance with 201.12.3.102

201.15.4.2.1.101 *Temperature and overload control devices – Additional requirements for application

In the fifth paragraph, replace the reference to "42 °C" by "41 °C".

201.15.4.101.1 Securing the fixation of heating elements and internal wiring

In the fifth paragraph, replace the reference to "201.15.4.1.102.2" by "201.15.4.101.2".