

TECHNICAL REPORT



Summary of requirements and tests to products in the scope of IEC 60601-2-66

IECNORM.COM : Click to view the full PDF of IEC TR 62809:2019 RLV



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2019 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

IECNORM.COM : Click to view the full text of IEC 60929:2019 PLV

TECHNICAL REPORT



Summary of requirements and tests to products in the scope of IEC 60601-2-66

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 11.180.15; 17.140.50

ISBN 978-2-8322-7556-6

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD	3
INTRODUCTION.....	2
1 Scope	6
2 Normative references	6
3 Terms and definitions	6
4 Summary of requirements and tests.....	7
Bibliography.....	19
Table 1 – Summary of requirements and tests	8

IECNORM.COM : Click to view the full PDF of IEC TR 62809:2019 RLV

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**SUMMARY OF REQUIREMENTS AND TESTS FOR
PRODUCTS IN THE SCOPE OF IEC 60601-2-66****FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This redline version of the official IEC Standard allows the user to identify the changes made to the previous edition. A vertical bar appears in the margin wherever a change has been made. Additions are in green text, deletions are in strikethrough red text.

The main task of IEC technical committees is to prepare International Standards. However, a technical committee may propose the publication of a Technical Report when it has collected data of a different kind from that which is normally published as an International Standard, for example "state of the art".

IEC 62809, which is a Technical Report, has been prepared by IEC technical committee 29: Electroacoustics.

This second edition cancels and replaces the first edition published in 2013. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) introduction of the term FITTED OSPL90 (FITTED MAXIMUM OUTPUT SOUND PRESSURE LEVEL) (201.3.206 of IEC 60601-2-66:2019);
- b) the allowable maximum output sound pressure level is now based on FITTED MAXIMUM OUTPUT SOUND PRESSURE LEVEL (201.9.6 of IEC 60601-2-66:2019);
- c) ESSENTIAL PERFORMANCE is based on risk analysis (201.4.3).

The text of this Technical Report is based on the following documents:

Draft TR	Report on voting
29/1015/DTR	29/1019/RVDTR

Full information on the voting for the approval of this Technical Report can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

During the preparation of IEC 60601-2-66, members of the involved technical committee and working group voiced concerns about the complexity of the document and its structure as part of the IEC 60601 series. Members felt distracted from the technical content by this complexity during reviews of the document stages. There was also concern that groups in the hearing ~~instrument~~ aid community would have problems to understand and apply the standard and that this could be an issue with its acceptability.

In order to have a broad consensus for the new standard, it was agreed that the standard should be supported by this Technical Report, which should enable members of the community and the industry to have a basic understanding of the requirements of the standard, without the need to study the complete standard document and the documents that are referenced in it.

IEC 60601-2-66 was published to address the specific requirements for safety of hearing ~~instruments~~ aids, and it is entitled “Particular requirements for the basic safety and essential performance of hearing ~~instruments~~ aids and hearing ~~instruments~~ aid systems”. It was published because IEC 60601-1 is a general standard intended to address a wide range of medical electrical equipment – including large scale facilities such as MRI machines, for example – and thus has large sections that are not relevant to low-voltage, low power, subminiature hearing ~~instruments~~ aids.

If IEC 60601-2-66 was not published, test and regulatory organizations would probably have difficulty applying IEC 60601-1, because it does not contain specific guidance for hearing ~~instruments~~ aids. This Technical Report contains all the requirements from IEC 60601-2-66 which relate to hearing ~~instruments~~ aids and reduces discussion with those that do not relate to hearing ~~instruments~~ aids.

It includes specific references to the applicable requirements within IEC 60601-1, and it is suggested that hearing ~~instrument~~ aid designers and manufacturers along with test and regulatory organizations read this Technical Report as an overview of IEC 60601-2-66.

SUMMARY OF REQUIREMENTS AND TESTS FOR PRODUCTS IN THE SCOPE OF IEC 60601-2-66

1 Scope

This document, which is a Technical Report, provides an overview of the requirements and tests of IEC 60601-2-66 in combination with the applicable sections of IEC 60601-1, and the collateral standards of the IEC 60601 series.

NOTE The IEC 60601 series consists of three levels of standards: IEC 60601-1, known as the general standard, several IEC 60601-1-X documents, known as the collateral standards, and a series of particular standards covering requirements for specific types of equipment (IEC 60601-2-X).

It is intended to assist various groups involved in the product lifecycles process – like designers and suppliers – to get an overview of the basic requirements without studying all involved standard documents in detail. The table includes not all but just the more common requirements and tests.

It is crucial to understand that the summary in this document cannot serve as an input for a product requirement specification or as a test plan without consulting ~~the standard document~~ IEC 60601-2-66 itself. This document alone cannot be used to establish or assess compliance to ~~the standard~~ IEC 60601-2-66.

The summary in Table 1 below does not preclude the user from reading the referenced standards in their entirety for a thorough knowledge of the basic safety of hearing ~~instruments~~ aids and hearing ~~instrument~~ aid systems.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*
IEC 60601-1:2005/AMD1:2012

IEC 60601-2-66:2012, *Medical electrical equipment – Part 2-66: Particular requirements for the basic safety and essential performance of hearing ~~instruments~~ aids and hearing ~~instrument~~ aid systems*

3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

4 Summary of requirements and tests

The reference column in Table 1 shows the clause number of IEC 60601-2-66:20122019 and, if applicable, the reference to IEC 60601-1:2005, or other documents. References to the particular standard IEC 60601-2-66:20122019 start with the number 201, while references to the general standard, IEC 60601-1:2005, start directly with the clause or subclause number.

Other documents will be referred to explicitly. Some detailed references, for example describing tools, are placed in the text instead of in the reference column.

IECNORM.COM : Click to view the full PDF of IEC TR 62809:2019 RLV

Table 1 – Summary of requirements and tests

Reference	Subject	Requirements and tests
201.1.1 201.8	ACCESSORY	Requirements: ACCESSORIES TO HEARING INSTRUMENTS AIDS (e.g. remote-control units, audio streamers, battery chargers, power supplies) need to have documented proof of compliance to IEC 60065, IEC 60950-1, IEC 60601-1 IEC 62368-1 or other applicable IEC safety standards. They form a HEARING INSTRUMENT AID SYSTEM when connected to the HEARING INSTRUMENT AID. Wherever this connection has an influence on the compliance to the requirements of IEC 60601-2-66, the HEARING INSTRUMENT AID has to pass the requirements while being connected to the ACCESSORY. If this connection results in additional requirements to the ACCESSORY, these requirements have to shall be fulfilled beyond the applicable IEC standards. Programming interfaces or ACCESSORIES in a clinical application are covered by IEC 60601 (all parts). For HEARING INSTRUMENTS AIDS that are supplied by an external power source: If a particular separate power supply is specified, then the relevant tests are performed with the HEARING INSTRUMENTS AIDS connected to it. If a generic separate power supply is specified, then the specification in the ACCOMPANYING DOCUMENTS is inspected. Tests: Inspection of ACCESSORY documentation, test configuration
201.4.1 201.4 Clause 4 Clause 5	Type tests	TYPE TESTS are performed on a representative sample of the item being tested. If multiple products are under consideration, which have a similar mechanical and electrical architecture, then an engineering analysis by the MANUFACTURER may justify a single representative sample for a family of products. Testing conditions shall consider not just NORMAL USE but also reasonably foreseeable misuse. Misuse and faults shall be the subject and results of the RISK ANALYSIS. The instructions for use have to shall be considered in testing conditions. The equipment is tested under the least favorable working conditions.
4.2 4.5 ISO 14971	Risk management	Requirements: A RISK MANAGEMENT complying with ISO 14971 shall be performed. Where IEC 60601-2-66 specifies requirements addressing particular RISKS, alternative means of addressing these RISKS are acceptable provided that the MANUFACTURER can justify that the RESIDUAL RISKS are the same or lower. Tests: Inspection of the RISK MANAGEMENT FILE.
201.4.3 4.3	Essential performance	Requirements; After a careful consideration of the clauses within this document, it was decided that they all deal with BASIC SAFETY as defined in the general standard. Manufacturers have the ability to identify functions of HEARING AIDS which are considered ESSENTIAL PERFORMANCE in accordance with their RISK MANAGEMENT PROCESS.
4.4	Expected service life	Requirements: The MANUFACTURER shall state the EXPECTED SERVICE LIFE of the HEARING INSTRUMENT AID in the RISK MANAGEMENT FILE. Tests: Inspection of the RISK MANAGEMENT FILE.

Reference	Subject	Requirements and tests
4.8 4.9	Components	Requirements: Components the failure of which could result in a HAZARDOUS SITUATION shall be used in accordance with their specified ratings. The reliability of components that are used as MEANS OF PROTECTION shall be assessed. They shall comply with the applicable safety requirements of a relevant IEC or ISO standard (options see 4.8). Tests: Inspection and, where necessary, test.
7.1.3	Durability of markings	Requirements: Markings shall be removable only with a TOOL or by appreciable force and shall be sufficiently durable to remain CLEARLY LEGIBLE during EXPECTED SERVICE LIFE. Tests: Markings are rubbed by hand, without undue pressure, first for 15 s with a cloth rag soaked with distilled water, then for 15 s with a cloth rag soaked with methylated spirit and then for 15 s with a cloth rag soaked with isopropyl alcohol.
201.7.1.2 201.7.2.1	Legibility of markings	Requirements: Markings shall be CLEARLY LEGIBLE when it is placed in the hand. The serial number and other markings shall be legible utilizing an optical aid if necessary. If the size of the HEARING-INSTRUMENT AID does not allow affixation of all required markings, the markings shall be recorded in full in the ACCOMPANYING DOCUMENTS or instructions for use (IFU). Tests: Inspection and verification of the applicable requirements.
201.7.2.2 7.3.3 7.4.2 7.8.2	Markings	Requirements: HEARING-INSTRUMENTS AIDS shall be marked with: – the name or trademark of the MANUFACTURER; – a MODEL OR TYPE REFERENCE. HEARING-INSTRUMENTS AIDS shall be marked visibly on the outside or inside when the battery drawer is open at a user accessible location (e.g. battery drawer), with: – if needed: identification of right and left HEARING-INSTRUMENT AID. Right = red. Left = blue; – serial number. HEARING-INSTRUMENTS AIDS worn in the ear: The marking on the instrument HEARING AID may be reduced to the serial number and the identification of right and left. The type of battery and the mode of insertion shall be marked on HEARING-INSTRUMENTS AIDS unless the design of the battery compartment prevents incorrect replacement of a battery Different positions of control devices and switches shall be indicated by figures, letters or other visual means. If the change of setting of a control could result in a RISK, such controls shall be provided with an indicating device or an indication of the direction in which the magnitude of the function changes. The color red shall be used only for emergency controls. Tests: Inspection and verification of the applicable requirements.

Reference	Subject	Requirements and tests
201.7.2.17	Protective packaging	Requirements: If special handling measures have to be taken during transport or storage, the packaging shall be marked accordingly. Tests: Inspection and verification of the applicable requirements.
201.7.8.1 201.7.9 7.6.1	ACCOMPANYING DOCUMENTS	Requirements: HEARING INSTRUMENTS AIDS shall be accompanied by instructions containing at least: – the purpose and INTENDED USE of the HEARING INSTRUMENT AID; – instructions for use, operating functions and a technical description; – easily understood diagrams, illustrations, or photographs of the fully assembled and ready-to-operate HEARING INSTRUMENT AID including all controls, visual information signals, and indicators, proper connection of the PATIENT to the HEARING INSTRUMENT AID, ACCESSORIES and other equipment; – identification of the HEARING INSTRUMENT AID: • name or trade name of the MANUFACTURER and an address to which the PATIENT can refer; • MODEL OR TYPE REFERENCE; – a list of detachable and replaceable parts as well as ACCESSORIES; – any restrictions on locations or environments in which the HEARING INSTRUMENT AID can be used; – identification of any known side effects associated with the use of HEARING INSTRUMENT AID that may warrant consultation with a physician e.g. accumulation of cerumen; – advice to the PATIENT to contact the MANUFACTURER or the MANUFACTURER's representative: • for assistance, if needed, in setting up, using or maintaining the HEARING INSTRUMENT AID or HEARING INSTRUMENTS AID SYSTEM; • to report unexpected operation or events; – a description and illustration on how to replace and/or recharge batteries; – colors of indicator lights and their meanings; – the meanings of the symbols used for marking (symbols see 7.6.2). – how to dispose of batteries of the HEARING INSTRUMENTS AIDS and of any part that may provide a RISK associated with the disposal; – information about cleaning and maintenance, where applicable: • the procedure to follow for washing the ear mould; • replacing tubing, filters and other replaceable parts; • storing the HEARING INSTRUMENT AID; • special adequate maintenance for rechargeable batteries; • information on how and where to obtain repair service; – if a HEARING INSTRUMENTS AIDS is difficult to retrieve from the ear canal a method to detect its location and to retrieve it shall be provided; – warning and safety notices in a specifically identified section of the instructions for use; if a warning or safety notice applies only to a specific instruction or action it should precede the instruction to which it applies; • for HEARING INSTRUMENTS AIDS in pediatric applications: Warning to keep small parts (HEARING INSTRUMENTS AIDS, batteries and detachable parts) that can be swallowed out of children's reach; • for HEARING INSTRUMENTS able to provide more than 132 dB SPL: warning to the professional OPERATOR fitting the HEARING INSTRUMENT that there may be a RISK of impairing the remaining hearing of the PATIENT; • for HEARING INSTRUMENTS AIDS that do not comply with requirements for explosive or oxygen-enriched atmospheres: warning not to use the HEARING INSTRUMENTS AIDS in such areas where there is danger of explosion;

Reference	Subject	Requirements and tests
		• warning that the specific HEARING-INSTRUMENT-MUST AID shall only be used by the intended person and not by others; • for HEARING-INSTRUMENTS AID with wireless transmission: warning to check first before using the HEARING AID SYSTEM in areas where electronics or wireless devices are restricted; • statement required about the special needs of particular PATIENT groups e.g. small children or mentally disabled persons; • warning about common conditions that could damage the HEARING-INSTRUMENT AID such as dropping, immersing in liquid, strong electromagnetic fields or excessive heat; • other warnings that may result from the risk assessment, e.g. a warning if parts could remain in the ear and what to do; • the permissible environmental conditions of transport and storage of a HEARING AID after it has been removed from its protective packaging and subsequently between uses; • for each warning and safety sign, the nature of the HAZARD, likely consequences that could occur if the advice is not followed, and the precautions for reducing the RISK; • if leakage from a battery would result in a RISK, a warning to remove the battery; • if the HEARING-INSTRUMENT AID can be externally connected, a warning only to connect to equipment that conforms to relevant international safety standards; – permissible environmental conditions of transport and storage of a HEARING-INSTRUMENT AID after it has been removed from its protective packaging and subsequently between uses; – the ACCOMPANYING DOCUMENTS shall be written at a level consistent with the education, training and any special needs of the person(s) for whom they are intended; – the instructions for use shall be in a language that is acceptable to the intended PATIENT; – Warning to the HEARING AID PROFESSIONAL; • for HEARING AID able to provide more than 132 dB SPL: warning to the professional OPERATOR fitting the HEARING AID that there may be a RISK of impairing the remaining hearing of the PATIENT; • the developed SPL in the ears of children can be substantially higher than in average adults. RECD measured to correct target of fitted OSPL90 is recommended; Parts of the ACCOMPANYING DOCUMENTS may be provided electronically. If so, the RISK MANAGEMENT PROCESS shall include consideration of which information also needs to be provided as hard copy. Tests: Inspection and verification of the applicable requirements.
201.8 201.8.7 201.5.7	Protection against electrical HAZARDS	Requirements: HEARING-INSTRUMENTS AID that are normally used in a HOME HEALTHCARE ENVIRONMENT with connections to electrical equipment in compliance with the relevant standard IEC 60065, IEC 60950-1, IEC 62368-1 or other applicable IEC safety standards shall pass the PATIENT LEAKAGE CURRENT requirements of 100 μA RMS at 275 V AC, as described in 201.8.7. Tests: PATIENT LEAKAGE CURRENT of 201.8.7, measurement after the humidity treatment of 201.5.7 and the drop test of 201.15.3.4.

Reference	Subject	Requirements and tests
201.16 201.5.9.2.1	ACCESSIBLE PARTS	Requirements: The voltage to earth or to ACCESSIBLE PARTS other than HEARING INSTRUMENTS AIDS shall not exceed 42,4 V peak AC or 60 V DC in NORMAL CONDITION or in SINGLE FAULT CONDITION. The DC limit of 60 V applies to DC with not more than 10 % peak-to-peak ripple. If the ripple exceeds that amount, the 42,4 V peak limit applies. The energy power shall not exceed 240 VA for longer than 60 s or the stored energy available shall not exceed 20 J at a potential up of 2 V or more. Tests: Inspection of product documentation in case this inspection is not conclusive: measurement. If internal parts exceed these limits, access is tested by test finger (201.5.9.2.1); small finger probe (Figure 1 of IEC 60601-1-11:2010), test pin (Figure 8 of IEC 60601-1:2005) and/or metal test rod (8.4.2).
201.8.4.2 201.8.7.4.8	Accessible contacts	Requirements: If an internally supplied HEARING INSTRUMENTS AIDS rated at 1,6 4,5 V DC or less has ACCESSIBLE PARTS, the DC current flowing in a realistic worst-case configuration between those contacts shall not exceed 10 µA and the RISK ASSESSMENT shall cover the particular design and application. Tests: Measurement and inspection of the RISK ASSESSMENT
201.9	Mechanical hazards	Requirements: Rough surfaces, sharp corners and edges of HEARING INSTRUMENT AID and HEARING INSTRUMENT AID SYSTEMS (e.g. moulded edges, battery doors and connector flanges) that could cause injury or damage shall be avoided or covered. Test: inspection of the HEARING INSTRUMENT AID, HEARING INSTRUMENT AID SYSTEMS and the RISK MANAGEMENT FILE.
201.9.6 201.13 IEC 60118-7:2005 IEC 60118-0	Acoustic energy	Requirements: HEARING INSTRUMENTS AIDS shall be designed in a way that users cannot be unintentionally exposed to a sound pressure level above 132 dB in NORMAL and SINGLE FAULT CONDITION SPL above the fitted OSPL90 in NORMAL CONDITION. The acceptable level increase in SINGLE FAULT condition, which shall be evaluated in the RISK MANAGEMENT FILE. The fitted OSPL90 shall be indicated on either the HEARING AID or the fitting software or otherwise. The maximum deviation shall be evaluated in the RISK MANAGEMENT FILE. HEARING AIDS with a possible MAXIMUM OUTPUT SOUND PRESSURE LEVEL of 132 dB SPL require a special warning notice for the HEARING AID PROFESSIONAL (see 201.7). Tests: Inspection of circuitry, firmware and software as applicable. Test in normal use and under SINGLE FAULT CONDITION. Measurements of sound pressure level are described in IEC 60118-7:2005 Inspection of the RISK MANAGEMENT FILE
201.9.101	Entanglement	Requirements: Cables and lanyards of HEARING INSTRUMENTS AIDS or ACCESSORIES worn by the PATIENT around the neck shall not pose a RISK of injury or strangulation. The disconnection force shall be no greater than 30 N. Tests: Application of the pull force.

Reference	Subject	Requirements and tests
201.9.102	Parts remaining in the ear canal	Requirements: A HEARING-INSTRUMENT AID that can be worn in the ear canal shall be safely retrievable by the PATIENT. HEARING-INSTRUMENTS AIDS shall be designed in a way that parts do not come loose during use, insertion and retrieval from the ear canal. Any part which is exposed to a pull force during the removal of a HEARING-INSTRUMENT AID from the ear canal shall resist a force of at least 3 N without coming loose from the instrument HEARING AID. Tests: Application of the pull force test, inspection of user instructions.
201.10 10.4	LED	Requirements: LEDs shall be class I according to IEC 60825-1. Tests: Inspection of the component specification and the circuitry.
201.11 201.13	Maximum temperature	Requirements: Maximum temperature of the HEARING-INSTRUMENTS AIDS shall not exceed 41 °C in normal use (in case of higher temperature, see 201.11.1.1) and 50 48 °C under fault conditions. If the hearing aid can operate in ambient temperatures above 43 °C, the maximum case temperature is permitted to be equal to the ambient temperature. Tests: Inspection of energy source and circuitry causing self-heating of the hearing aid. In case of doubt, a measurement shall be performed: the maximum temperature of the HEARING AID shall not exceed 43 °C, when tested in ambient temperature of 25 °C to 35 °C.
201.11.6.5	Ingress of water	Requirements: If the RISK ASSESSMENT requires protection against harmful ingress of water or particulate matter, the IP class of the HEARING-INSTRUMENT AID shall be not less than the level required for safe operation as detailed in IEC 60529. Tests: Apply IEC 60529.
201.11.6.6	Cleaning	Requirements: HEARING-INSTRUMENTS AIDS and HEARING-INSTRUMENT AID SYSTEMS and their parts and ACCESSORIES shall be capable of withstanding, without damage or deterioration, the cleaning or disinfection processes (such as cerumen removal), as specified in the instructions for use. Tests: Evaluate the effects of multiple cleanings during the EXPECTED SERVICE LIFE of the HEARING-INSTRUMENTS AIDS and HEARING-INSTRUMENTS AID SYSTEMS, their parts and ACCESSORIES and assure that no unacceptable RISK will occur. The results of the evaluation shall be documented in the RISK MANAGEMENT FILE.
11.6.8	Compatibility with substances	Requirements: When applicable, the MANUFACTURER shall address in the RISK MANAGEMENT PROCESS the RISKS associated with compatibility with substances used with the HEARING-INSTRUMENT AID. Tests: Inspection of the RISK MANAGEMENT FILE.
11.7	Biocompatibility	Requirements: HEARING-INSTRUMENTS AIDS shall be biocompatible. Tests: Assessment and documentation according to the guidance and principles given in ISO 10993-series (all parts).

Reference	Subject	Requirements and tests
201.7.1.1 201.12.2 12.2 IEC 60601-1-6	Usability	Requirements: USABILITY ENGINEERING PROCESS according to IEC 60601-1-6 and IEC 62366 shall be performed. The results have to be considered in the design of a product, the labelling and instructions. HEARING INSTRUMENTS AIDS should be designed to be simple to use and not require reference to complex instructions. Typical PRIMARY OPERATING FUNCTIONS are: – critical functions: • placing and removing the HEARING INSTRUMENT AID; • fitting a HEARING INSTRUMENT AID; • testing of essential physical HEARING INSTRUMENT AID parameters; – frequently used functions: • changing battery; • cleaning; • switching on /off; • adjust volume, program and other essential parameters. Tests: Inspection of the results of the USABILITY ENGINEERING PROCESS and the instructions; trial test the product if necessary.
201.12.4.4	Incorrect output	Requirements: When a control adjusts the intended maximum power output, output power shall not increase if the control is disconnected or defective. Software controlled maximum power settings shall not exceed the selected value as a result of corrupt data transfer between programmer and HEARING INSTRUMENT AID. Tests: Inspection of circuitry, firmware, software and the RISK MANAGEMENT FILE; trial test where necessary.
12.4.3	Incorrect output	Requirements: The MANUFACTURER shall address in the RISK MANAGEMENT PROCESS the RISKS associated with accidental selection of an excessive output value. Tests: Inspection of the RISK MANAGEMENT FILE.

Reference	Subject	Requirements and tests
201.13 13 11.3	HAZARDOUS SITUATIONS and fault conditions	Requirements: The following HAZARDOUS SITUATIONS shall not occur: – emission of flames, molten metal, poisonous or ignitable substance in hazardous quantities; – deformation of ENCLOSURES to such an extent that risks result; – temperatures of HEARING INSTRUMENTS AIDS that are likely to be touched, exceeding 50 °C when measured and adjusted as described in 201.11.1.3 of the general standard shall not exceed 43 °C. - exceeding the allowable values for “other components and materials” identified in Table 22 of the general standard times 1,5 minus 12,5 °C. Tests: Inspection of circuit, application of faults. Faults need not be applied if a) the construction or the supply circuit limits the power dissipation in SINGLE FAULT CONDITION to less than 15 W or the energy dissipation to less than 900 J, - or b) if parts are completely contained within a fire enclosure according to the general standard. See 201.13.2.1 for the description of SINGLE FAULT and NORMAL CONDITIONS. NOTE The requirements to HEARING INSTRUMENTS AIDS that are intended to be used in oxygen enriched or explosive atmospheres are not contained in this particular standard IEC 60601-2-66.

Reference	Subject	Requirements and tests
201.14 14	Software	Requirements: – Embedded and fitting software shall conform to IEC 62304. – The classification of software according to IEC 62304 shall be the result of the RISK ASSESSMENT. Aspects to be considered beyond the requirements of IEC 62304: 14.2 Documentation The required processes for RISK MANAGEMENT and software development have to shall be correctly implemented and maintained. Documentation that the PROCESS steps have been performed needs to be filed in the RISK MANAGEMENT FILE. A document control system needs to be in place. 14.3 RISK MANAGEMENT plan In addition to elements of the RISK MANAGEMENT plan required by ISO 14971, a PEMS VALIDATION plan is required. 14.6.1 Identification of known and foreseeable HAZARDS PEMS have extra initiating causes for HAZARDS. 14.11 PEMS VALIDATION PEMS VALIDATION is intended to assure that the right product is built. Unexpected interactions between functions might occur that can only be discovered by validation. It can include tests for a high volume of data, heavy loads or stresses, human factors, security, performance, configuration compatibility, fault testing, documentation and safety. Independence is needed to avoid conflicts of interest and because the assumptions of the designer should not influence or limit the extent of the PEMS VALIDATION. Examples of level of independence include: – separate person; – separate management; – separate organization. Tests: Application of the requirements in Clause 201.14.2 to 201.14.13 (inclusive), by inspection of the RISK MANAGEMENT FILE. Compliance with the IEC 62304 software design and the RISK MANAGEMENT PROCESS by inspection of external or internal audit reports or certificates.
201.15.2	Serviceability	Requirements: Parts of HEARING INSTRUMENTS AIDS subject to mechanical wear, electrical and environmental degradation or ageing that could result in an unacceptable RISK if allowed to continue unchecked for too long a period shall be accessible for inspection, replacement and maintenance. Tests: Inspection of such parts and their location.

Reference	Subject	Requirements and tests
201.15.3	Mechanical strength	Requirements: HEARING INSTRUMENTS AIDS or its parts shall have adequate mechanical strength and shall not result in an unacceptable RISK due to moulding stress or when subjected to mechanical stress caused by pushing, impact, dropping, and rough handling. Mechanical design requirements for instruments HEARING AIDS intended for use by infants under 36 months: a) battery doors shall be constructed to: • require a TOOL to remove the battery, or • require a force of at least 10 N in the least favorable direction to remove the battery, b) any detachable part of the HEARING INSTRUMENT AID (e.g. ear hook, tube, type plate, programming cover) shall not be removable: • without a TOOL, or • with a force lower than 10 N in the direction of least resistance. Tests: Inspection of the product and the application of the described forces.
201.15.3.4	Drop test	Requirement: HEARING INSTRUMENTS AIDS shall not result in an unacceptable RISK as a result of a free fall, such as increased LEAKAGE CURRENT or mechanical hazards. Test: The sample to be tested is allowed to fall freely once from each of six different starting orientations from 4,5 1,0 m height onto a hard wood surface.
201.15.3.7	Environmental influences	Requirements: The selection and treatment of materials used in the construction of HEARING INSTRUMENTS AIDS shall take account of the INTENDED USE, the EXPECTED SERVICE LIFE and the conditions for transport and storage. The HEARING INSTRUMENTS AIDS shall be so designed and constructed that during its EXPECTED SERVICE LIFE any corrosion, ageing, mechanical wear, or degradation of biological materials due to the influence of moisture, sweat, humidity, hair care products or toiletries shall not reduce its mechanical properties in a way that results in an unacceptable RISK. Tests: Inspection of HEARING INSTRUMENTS AIDS, the ACCOMPANYING DOCUMENTS and of the specifications of materials used, of the processing specifications for these materials and relevant tests or calculations.
201.15.4.3.1	Housing	Requirements: Battery compartments shall be designed to prevent accidental short circuiting of the battery where such short circuits could result in a HAZARDOUS SITUATION. If a HAZARDOUS SITUATION might develop by the incorrect connection or replacement of a battery, the equipment shall be fitted with a means of preventing incorrect polarity. Tests: Inspection of circuit and construction.
15.4.3.3	Overcharging	Requirements: Where overcharging of any battery could result in an unacceptable RISK, the design shall prevent overcharging. Tests: Inspection of the design documentation.

Reference	Subject	Requirements and tests
201.15.4.3.101	Batteries	Requirements: Batteries, used to supply HEARING INSTRUMENTS AIDS, shall comply with the relevant international standards. The design of the electronic circuit shall avoid overheating of a wrongly inserted battery above 50 °C. Tests: Inspection of circuit and construction, if necessary, testing. Inspection of the instructions for use.
201.17	Electromagnetic compatibility	Requirements: The MANUFACTURER shall address in the RISK MANAGEMENT PROCESS the RISKS associated with the introduction by the HEARING INSTRUMENT AID of electromagnetic phenomena into the environment that might degrade the performance of other devices, electrical equipment and systems. Tests: Electromagnetic compatibility is tested according to IEC 60118-13. If the HEARING INSTRUMENT has a wireless transmitter, emissions are furthermore tested according to relevant international radio standards. Electromagnetic compatibility shall be tested in accordance with IEC 60601-1-2. For HEARING AIDS, additional testing for immunity to digital wireless interference shall be tested according to IEC 60118-13. Pass/fail criteria shall be defined according to the generic standard IEC 60601-1-2. Inspection of the RISK MANAGEMENT FILE.

IECNORM.COM : Click to view the full PDF of IEC TR 62809:2019 RLV

Bibliography

IEC 60065:~~2001~~2014, *Audio, video and similar electronic apparatus – Safety requirements*

~~IEC 60118-7:2005, *Electroacoustics – Hearing aids – Part 7: Measurement of the performance characteristics of hearing aids for production, supply and delivery quality assurance purposes*~~

IEC 60118-0:2015, *Electroacoustics – Hearing aids – Part 0: Measurement of the performance characteristics of hearing aids*

IEC 60118-13:~~2011~~2016, *Electroacoustics – Hearing aids – Part 13: Electromagnetic compatibility (EMC)*

IEC 60529, *Degrees of protection provided by enclosures (IP Code)*

IEC 60601 (all parts), *Medical electrical equipment*

~~IEC 60601-1-1:2000, *Medical electrical equipment – Part 1-1: General requirements for safety – Collateral standard: Safety requirements for medical electrical systems*~~

IEC 60601-1-2, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests*

~~IEC 60529:1989, *Degrees of protection provided by enclosures (IP Code)*~~

~~IEC 60825-1:2007, *Safety of laser products – Part 1: Equipment classification and requirements*~~

~~IEC 60601-1-11:2010, *Medical electrical equipment – Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment*~~

IEC 60601-1-6:2010, *Medical electrical equipment – Part 1-6: General requirements for basic safety and essential performance – Collateral standard: Usability*

IEC 60825-1:2014, *Safety of laser products – Part 1: Equipment classification and requirements*

IEC 60950-1:~~2005~~, *Information technology equipment – Safety – Part 1: General requirements*

IEC 62304:2006, *Medical device software – Software life-cycle processes*

IEC 62366-1, *Medical devices – Part 1: Application of usability engineering to medical devices*

IEC 62368-1:2018, *Audio/video, information and communication technology equipment – Part 1: Safety requirements*

ISO 10993 (all parts), *Biological evaluation of medical devices*

ISO 14971:~~2000~~2007, *Medical devices – Application of risk management to medical devices*

IECNORM.COM : Click to view the full PDF of IEC TR 62809:2019 RLV

TECHNICAL REPORT

Summary of requirements and tests to products in the scope of IEC 60601-2-66

IECNORM.COM : Click to view the full PDF of IEC TR 62809:2019 RLV

CONTENTS

FOREWORD	3
INTRODUCTION.....	5
1 Scope	6
2 Normative references	6
3 Terms and definitions	6
4 Summary of requirements and tests.....	7
Bibliography.....	18
Table 1 – Summary of requirements and tests	8

IECNORM.COM : Click to view the full PDF of IEC TR 62809:2019 RLV

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**SUMMARY OF REQUIREMENTS AND TESTS FOR
PRODUCTS IN THE SCOPE OF IEC 60601-2-66****FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. However, a technical committee may propose the publication of a Technical Report when it has collected data of a different kind from that which is normally published as an International Standard, for example "state of the art".

IEC 62809, which is a Technical Report, has been prepared by IEC technical committee 29: Electroacoustics.

This second edition cancels and replaces the first edition published in 2013. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) introduction of the term FITTED OSPL90 (FITTED MAXIMUM OUTPUT SOUND PRESSURE LEVEL) (201.3.206 of IEC 60601-2-66:2019);
- b) the allowable maximum output sound pressure level is now based on FITTED MAXIMUM OUTPUT SOUND PRESSURE LEVEL (201.9.6 of IEC 60601-2-66:2019).
- c) ESSENTIAL PERFORMANCE is based on risk analysis (201.4.3).

The text of this Technical Report is based on the following documents:

Draft TR	Report on voting
29/1015/DTR	29/1019/RVDTR

Full information on the voting for the approval of this Technical Report can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

A bilingual version of this publication may be issued at a later date.

IECNORM.COM : Click to view the full PDF of IEC TR 62809:2019 RLV

INTRODUCTION

During the preparation of IEC 60601-2-66, members of the involved technical committee and working group voiced concerns about the complexity of the document and its structure as part of the IEC 60601 series. Members felt distracted from the technical content by this complexity during reviews of the document stages. There was also concern that groups in the hearing aid community would have problems to understand and apply the standard and that this could be an issue with its acceptability.

In order to have a broad consensus for the new standard, it was agreed that the standard should be supported by this Technical Report, which should enable members of the community and the industry to have a basic understanding of the requirements of the standard, without the need to study the complete standard document and the documents that are referenced in it.

IEC 60601-2-66 was published to address the specific requirements for safety of hearing aids, and it is entitled “Particular requirements for the basic safety and essential performance of hearing aids and hearing aid systems”. It was published because IEC 60601-1 is a general standard intended to address a wide range of medical electrical equipment – including large scale facilities such as MRI machines, for example – and thus has large sections that are not relevant to low-voltage, low power, subminiature hearing aids.

If IEC 60601-2-66 was not published, test and regulatory organizations would probably have difficulty applying IEC 60601-1, because it does not contain specific guidance for hearing aids. This Technical Report contains all the requirements from IEC 60601-2-66 which relate to hearing aids and reduces discussion with those that do not relate to hearing aids.

It includes specific references to the applicable requirements within IEC 60601-1, and it is suggested that hearing aid designers and manufacturers along with test and regulatory organizations read this Technical Report as an overview of IEC 60601-2-66.

SUMMARY OF REQUIREMENTS AND TESTS FOR PRODUCTS IN THE SCOPE OF IEC 60601-2-66

1 Scope

This document, which is a Technical Report, provides an overview of the requirements and tests of IEC 60601-2-66 in combination with the applicable sections of IEC 60601-1, and the collateral standards of the IEC 60601 series.

NOTE The IEC 60601 series consists of three levels of standards: IEC 60601-1, known as the general standard, several IEC 60601-1-X documents, known as the collateral standards, and a series of particular standards covering requirements for specific types of equipment (IEC 60601-2-X).

It is intended to assist various groups involved in the product lifecycles process – like designers and suppliers – to get an overview of the basic requirements without studying all involved standard documents in detail. The table includes not all but just the more common requirements and tests.

It is crucial to understand that the summary in this document cannot serve as an input for a product requirement specification or as a test plan without consulting IEC 60601-2-66 itself. This document alone cannot be used to establish or assess compliance to IEC 60601-2-66.

The summary in Table 1 below does not preclude the user from reading the referenced standards in their entirety for a thorough knowledge of the basic safety of hearing aids and hearing aid systems.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*
IEC 60601-1:2005/AMD1:2012

IEC 60601-2-66:2019, *Medical electrical equipment – Part 2-66: Particular requirements for the basic safety and essential performance of hearing aids and hearing aid systems*

3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

4 Summary of requirements and tests

The reference column in Table 1 shows the clause number of IEC 60601-2-66:2019 and, if applicable, the reference to IEC 60601-1:2005, or other documents. References to the particular standard IEC 60601-2-66:2019 start with the number 201, while references to the general standard, IEC 60601-1:2005, start directly with the clause or subclause number.

Other documents will be referred to explicitly. Some detailed references, for example describing tools, are placed in the text instead of in the reference column.

IECNORM.COM : Click to view the full PDF of IEC TR 62809:2019 RLV

Table 1 – Summary of requirements and tests

Reference	Subject	Requirements and tests
201.1.1 201.8	ACCESSORY	Requirements: ACCESSORIES TO HEARING AIDS (e.g. remote-control units, audio streamers, battery chargers, power supplies) need to have documented proof of compliance to IEC 60065, IEC 60950-1, IEC 62368-1 or other applicable IEC safety standards. They form a HEARING AID SYSTEM when connected to the HEARING AID. Wherever this connection has an influence on the compliance to the requirements of IEC 60601-2-66, the HEARING AID has to pass the requirements while being connected to the ACCESSORY. If this connection results in additional requirements to the ACCESSORY, these requirements shall be fulfilled beyond the applicable IEC standards. Programming interfaces or ACCESSORIES in a clinical application are covered by IEC 60601 (all parts). For HEARING AIDS that are supplied by an external power source: If a particular separate power supply is specified, then the relevant tests are performed with the HEARING AIDS connected to it. If a generic separate power supply is specified, then the specification in the ACCOMPANYING DOCUMENTS is inspected. Tests: Inspection of ACCESSORY documentation, test configuration
201.4.1 201.4 Clause 4 Clause 5	Type tests	TYPE TESTS are performed on a representative sample of the item being tested. If multiple products are under consideration, which have a similar mechanical and electrical architecture, then an engineering analysis by the MANUFACTURER may justify a single representative sample for a family of products. Testing conditions shall consider not just NORMAL USE but also reasonably foreseeable misuse. Misuse and faults shall be the subject and results of the RISK ANALYSIS. The instructions for use shall be considered in testing conditions. The equipment is tested under the least favorable working conditions.
4.2 4.5 ISO 14971	Risk management	Requirements: A RISK MANAGEMENT complying with ISO 14971 shall be performed. Where IEC 60601-2-66 specifies requirements addressing particular RISKS, alternative means of addressing these RISKS are acceptable provided that the MANUFACTURER can justify that the RESIDUAL RISKS are the same or lower. Tests: Inspection of the RISK MANAGEMENT FILE.
201.4.3 4.3	Essential performance	Requirements; After a careful consideration of the clauses within this document, it was decided that they all deal with BASIC SAFETY as defined in the general standard. Manufacturers have the ability to identify functions of HEARING AIDS which are considered ESSENTIAL PERFORMANCE in accordance with their RISK MANAGEMENT PROCESS.
4.4	Expected service life	Requirements: The MANUFACTURER shall state the EXPECTED SERVICE LIFE of the HEARING AID in the RISK MANAGEMENT FILE. Tests: Inspection of the RISK MANAGEMENT FILE.
4.8 4.9	Components	Requirements: Components the failure of which could result in a HAZARDOUS SITUATION shall be used in accordance with their specified ratings. The reliability of components that are used as MEANS OF PROTECTION shall be assessed. They shall comply with the applicable safety requirements of a relevant IEC or ISO standard (options see 4.8). Tests: Inspection and, where necessary, test.

Reference	Subject	Requirements and tests
7.1.3	Durability of markings	Requirements: Markings shall be removable only with a TOOL or by appreciable force and shall be sufficiently durable to remain CLEARLY LEGIBLE during EXPECTED SERVICE LIFE. Tests: Markings are rubbed by hand, without undue pressure, first for 15 s with a cloth rag soaked with distilled water, then for 15 s with a cloth rag soaked with methylated spirit and then for 15 s with a cloth rag soaked with isopropyl alcohol.
201.7.1.2 201.7.2.1	Legibility of markings	Requirements: Markings shall be CLEARLY LEGIBLE when it is placed in the hand. The serial number and other markings shall be legible utilizing an optical aid if necessary. If the size of the HEARING AID does not allow affixation of all required markings, the markings shall be recorded in full in the ACCOMPANYING DOCUMENTS or instructions for use (IFU). Tests: Inspection and verification of the applicable requirements.
201.7.2.2 7.3.3 7.4.2 7.8.2	Markings	Requirements: HEARING AIDS shall be marked with: – the name or trademark of the MANUFACTURER; – a MODEL OR TYPE REFERENCE. HEARING AIDS shall be marked visibly on the outside or at a user accessible location (e.g. battery drawer), with: – if needed: identification of right and left HEARING AID. Right = red Left = blue; – serial number. HEARING AIDS worn in the ear: The marking on the HEARING AID may be reduced to the serial number and the identification of right and left. The type of battery and the mode of insertion shall be marked on HEARING AIDS unless the design of the battery compartment prevents incorrect replacement of a battery Different positions of control devices and switches shall be indicated by figures, letters or other visual means. If the change of setting of a control could result in a RISK, such controls shall be provided with an indicating device or an indication of the direction in which the magnitude of the function changes. The color red shall be used only for emergency controls. Tests: Inspection and verification of the applicable requirements.
201.7.2.17	Protective packaging	Requirements: If special handling measures have to be taken during transport or storage, the packaging shall be marked accordingly. Tests: Inspection and verification of the applicable requirements.
201.7.8.1 201.7.9 7.6.1	ACCOMPANYING DOCUMENTS	Requirements: HEARING AIDS shall be accompanied by instructions containing at least: – the purpose and INTENDED USE of the HEARING AID; – instructions for use, operating functions and a technical description; – easily understood diagrams, illustrations, or photographs of the fully assembled and ready-to-operate HEARING AID including all controls, visual information signals, and indicators, proper

		connection of the PATIENT to the HEARING AID, ACCESSORIES and other equipment; – identification of the HEARING AID: • name or trade name of the MANUFACTURER and an address to which the PATIENT can refer; • MODEL OR TYPE REFERENCE; – a list of detachable and replaceable parts as well as ACCESSORIES; – any restrictions on locations or environments in which the HEARING AID can be used; – identification of any known side effects associated with the use of HEARING AID that may warrant consultation with a physician e.g. accumulation of cerumen; – advice to the PATIENT to contact the MANUFACTURER or the MANUFACTURER's representative: • for assistance, if needed, in setting up, using or maintaining the HEARING AID or HEARING AID SYSTEM; • to report unexpected operation or events; – a description and illustration on how to replace and/or recharge batteries; – colors of indicator lights and their meanings; – the meanings of the symbols used for marking (symbols see 7.6.2). – how to dispose of batteries of the HEARING AIDS and of any part that may provide a RISK associated with the disposal; – information about cleaning and maintenance, where applicable: • the procedure to follow for washing the ear mould; • replacing tubing, filters and other replaceable parts; • storing the HEARING AID; • special adequate maintenance for rechargeable batteries; • information on how and where to obtain repair service; – if a HEARING AIDS is difficult to retrieve from the ear canal a method to detect its location and to retrieve it shall be provided; – warning and safety notices in a specifically identified section of the instructions for use; if a warning or safety notice applies only to a specific instruction or action it should precede the instruction to which it applies; • for HEARING AIDS in pediatric applications: Warning to keep small parts (HEARING AIDS, batteries and detachable parts) that can be swallowed out of children's reach; • for HEARING AIDS that do not comply with requirements for explosive or oxygen-enriched atmospheres: warning not to use the HEARING AIDS in such areas where there is danger of explosion; • warning that the specific HEARING AID shall only be used by the intended person and not by others; • for HEARING AIDS with wireless transmission: warning to check first before using the HEARING AID SYSTEM in areas where electronics or wireless devices are restricted; • statement required about the special needs of particular PATIENT groups e.g. small children or mentally disabled persons; • warning about common conditions that could damage the HEARING AID such as dropping, immersing in liquid, strong electromagnetic fields or excessive heat; • other warnings that may result from the risk assessment, e.g. a warning if parts could remain in the ear and what to do; • the permissible environmental conditions of transport and storage of a HEARING AID after it has been removed from its protective packaging and subsequently between uses; • for each warning and safety sign, the nature of the HAZARD, likely consequences that could occur if the advice is not followed, and the precautions for reducing the RISK; • if leakage from a battery would result in a RISK, a warning to remove the battery; • if the HEARING AID can be externally connected, a warning only to connect to equipment that conforms to relevant international safety standards;
--	--	---

Reference	Subject	Requirements and tests
		- permissible environmental conditions of transport and storage of a HEARING AID after it has been removed from its protective packaging and subsequently between uses; - the ACCOMPANYING DOCUMENTS shall be written at a level consistent with the education, training and any special needs of the person(s) for whom they are intended; - the instructions for use shall be in a language that is acceptable to the intended PATIENT; - Warning to the HEARING AID PROFESSIONAL; - for HEARING AIDS able to provide more than 132 dB SPL: warning to the professional OPERATOR fitting the HEARING AID that there may be a RISK of impairing the remaining hearing of the PATIENT; - the developed SPL in the ears of children can be substantially higher than in average adults. RECD measured to correct target of fitted OSPL90 is recommended; Parts of the ACCOMPANYING DOCUMENTS may be provided electronically. If so, the RISK MANAGEMENT PROCESS shall include consideration of which information also needs to be provided as hard copy. Tests: Inspection and verification of the applicable requirements.
201.8 201.8.7 201.5.7	Protection against electrical HAZARDS	Requirements: HEARING AIDS that are normally used in a HOME HEALTHCARE ENVIRONMENT with connections to electrical equipment in compliance with the relevant standard IEC 60065, IEC 60950-1, IEC 62368-1 or other applicable IEC safety standards shall pass the PATIENT LEAKAGE CURRENT requirements of 100 μA RMS at 275 V AC, as described in 201.8.7. Tests: PATIENT LEAKAGE CURRENT of 201.8.7, measurement after the humidity treatment of 201.5.7 and the drop test of 201.15.3.4.
201.16 201.5.9.2.1	ACCESSIBLE PARTS	Requirements: The voltage to earth or to ACCESSIBLE PARTS other than HEARING AIDS shall not exceed 42,4 V peak AC or 60 V DC in NORMAL CONDITION or in SINGLE FAULT CONDITION. The DC limit of 60 V applies to DC with not more than 10 % peak-to-peak ripple. If the ripple exceeds that amount, the 42,4 V peak limit applies. The power shall not exceed 240 VA for longer than 60 s or the stored energy available shall not exceed 20 J at a potential of to 2 V or more. Tests: Inspection of product documentation in case this inspection is not conclusive: measurement. If internal parts exceed these limits, access is tested by test finger (201.5.9.2.1); small finger probe (test pin (Figure 8 of IEC 60601-1:2005) and/or metal test rod (8.4.2).
201.8.4.2 201.8.7.4.8	Accessible contacts	Requirements: If an internally supplied HEARING AIDS rated at 4,5 V DC or less has ACCESSIBLE PARTS, the DC current flowing in a realistic worst-case configuration between those contacts shall not exceed 10 μA and the RISK ASSESSMENT shall cover the particular design and application. Tests: Measurement and inspection of the RISK ASSESSMENT
201.9	Mechanical hazards	Requirements: Rough surfaces, sharp corners and edges of HEARING AID and HEARING AID SYSTEMS (e.g. moulded edges, battery doors and connector flanges) that could cause injury or damage shall be avoided or covered. Test: inspection of the HEARING AID, HEARING AID SYSTEMS and the RISK MANAGEMENT FILE.