



Medical suction equipment –

Part 1:

Electrically powered suction equipment – Safety requirements

TECHNICAL CORRIGENDUM 2

Matériel d'aspiration médical –

Partie 1: Matériel électrique d'aspiration – Prescriptions de sécurité

RECTIFICATIF TECHNIQUE 2

Technical corrigendum 2 to International Standard ISO 10079-1:1991 was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Sub-Committee SC 8, *Suction devices for hospital and emergency care use*.

Page 3, Clause 1.3.5

Add a Note as follows.

NOTE – A series of push buttons or a rotary switch or similar by which it is possible to set different vacuum levels constitute a vacuum regulator.

Page 15, Clause 9.52.6

Delete the second paragraph starting with "Compliance shall..." and add the following:

Except for equipment intended for use in transport or in the field, when the requirements of 52.8 shall apply, compliance shall be checked by the test given in 52.7.

Page 17, Clause 10.56.8 6)

Add a Note as follows.

NOTE – If equipment intended for field and transport use is fitted with a vacuum regulator with a variable control, a vacuum indicator is required in order to comply with 56.8 1).