



# INTERNATIONAL STANDARD ISO 10271:2001 TECHNICAL CORRIGENDUM 1

Published 2005-01-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

## Dental metallic materials — Corrosion test methods

### TECHNICAL CORRIGENDUM 1

*Produits dentaires métalliques — Méthodes pour les essais de corrosion*

*RECTIFICATIF TECHNIQUE 1*

Technical Corrigendum 1 to ISO 10271:2001 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 2, *Prosthetic materials*.

Page 4, subclause 4.1.6.4.1

In the third paragraph, replace the first three sentences:

“Remove at least 0,1 mm from each surface using standard metallurgical procedures unless samples are to be tested in the as-received condition. Use fresh abrasive paper for each alloy. Finish with ASTM 600 or FEPA 1200 wet silicon carbide paper.”

by

“Finish each surface as specified by the manufacturer or in a manner consistent with good dental technology.”

In the fourth paragraph, first sentence, replace

“water”

by

“distilled water”.