
**Health and safety in welding and allied
processes — Sampling of airborne
particles and gases in the operator's
breathing zone —**

**Part 2:
Sampling of gases**

*Hygiène et sécurité en soudage et techniques connexes —
Échantillonnage de particules en suspension et gaz dans la zone
respiratoire des opérateurs —*

Partie 2: Échantillonnage des gaz



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 10882-2:2000

© ISO 2000

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.ch
Web www.iso.ch

Printed in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this part of ISO 10882 may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

International Standard ISO 10882-2 was prepared by the European Committee for Standardization (CEN) in collaboration with ISO Technical Committee TC 44, *Welding and allied processes*, Subcommittee SC 9, *Health and safety*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

Throughout the text of this standard, read "...this European Standard..." to mean "...this International Standard...".

ISO 10882 consists of the following parts, under the general title *Health and safety in welding and allied processes* — *Sampling of airborne particles and gases in the operator's breathing zone*:

- Part 1: *Sampling of airborne particles*
- Part 2: *Sampling of gases*

Annexes A and B of this part of ISO 10882 are for information only.

Contents

Page

Foreword.....	v
Introduction	vi
1 Scope	1
2 Normative references	2
3 Terms and definitions.....	2
4 Description of measurement methods	6
4.1 General.....	6
4.2 Direct reading electrical apparatus	7
4.3 Detector tubes	7
4.4 Indirect methods involving laboratory analysis	8
5 Requirements	9
6 Assessment strategy	9
7 Measurement strategy	10
7.1 General.....	10
7.2 Personal exposure measurements	10
7.3 Fixed point measurements	10
7.4 Selection of measurement conditions and measurement pattern.....	10
8 Sampling	11
8.1 Sampling position.....	11
8.2 Sampling equipment.....	11
8.3 Sample filtration.....	12
8.4 Multiple sampling.....	12
8.5 Volume of sampling line.....	12
8.6 Flow rate	12
8.7 Handling of temperature, pressure and humidity data	12
9 Measurement of individual gases and vapours.....	13
9.1 General.....	13
9.2 Ozone (0,01 ppm to 3 ppm)	13
9.3 Carbon monoxide (3 ppm to 500 ppm)	13
9.4 Carbon dioxide (500 ppm to 10 %)	14
9.5 Nitric oxide (1 ppm to 100 ppm) and nitrogen dioxide (0,3 ppm to 250 ppm)	14
9.6 Vapours.....	15
10 Recording of test data and presentation of results	16
Annex A (Informative) Measurement of individual gases and vapours	17
Annex B (informative) An example of a test report	18
B.1 Basic data	18
B.2 Process data.....	19
B.3 Sampling data and test results.....	20
Bibliography	21

Foreword

The text of EN ISO 10882-2:2000 has been prepared by Technical Committee CEN/TC 121 "Welding", the secretariat of which is held by DS, in collaboration with Technical Committee ISO/TC 44 "Welding and allied processes".

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2001, and conflicting national standards shall be withdrawn at the latest by March 2001.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

STANDARDSISO.COM : Click to view the full PDF of ISO 10882-2:2000

Introduction

Gases encountered during welding and allied processes are so numerous that it would be impracticable to cover them all in this European Standard. Depending on the process, they can include:

- a) fuel gases which are used in gas welding and cutting which on combustion produce carbon dioxide and in some instances carbon monoxide;
- b) shielding gases such as argon, helium, carbon dioxide or mixtures of these gases, which can be toxic or asphyxiant;
- c) gases produced by the action of heat upon the welding flux or slag, e.g. carbon dioxide and carbon monoxide;
- d) gases produced by the action of heat or ultraviolet radiation upon the atmosphere surrounding the welding arc, e.g. nitric oxide, nitrogen dioxide and ozone; and
- e) vapours produced as a result of thermal degradation of surface coatings in the welding or cutting of metals treated with paint, primer, sealer or other substances. Vapours can also be produced as a result of degradation of solvent vapour from degreasing operations, but their measurement is not dealt with in this standard because good working practices will avoid their production.

The scope of this part of EN ISO 10882 has been limited to those gases which are produced by welding operations. In particular, fuel, oxidant and shielding gases used in welding and allied processes are not covered, since the hazards associated with their use (e.g. asphyxiation, explosion) are different from those arising from the gases dealt with in this guide.

This part of EN ISO 10882 gives a generalised description of measurement methods suitable for the assessment of personal exposure to gases produced by welding and allied processes; gives details of relevant European Standards which specify required characteristics, performance requirements and test methods; augments guidance provided in EN 689 on assessment strategy and measurement strategy; lists basic sampling requirements; and provides specific information about the availability of direct reading electrical apparatus, detector tubes and indirect methods involving laboratory analysis for individual gases.

It has been assumed in the drafting of this standard that the execution of its provisions, and the interpretation of the results obtained, is entrusted to appropriately qualified and experienced people.

1 Scope

This part of EN ISO 10882 provides guidance for the determination of personal exposure to gases and vapours in welding and allied processes. It applies to the following thermal processes used to join, cut, surface or remove metals:

- (111) Manual metal arc welding (metal arc welding with covered electrode); shielded metal arc welding /USA/
- (114) Self-shielded tubular-cored arc welding
- (131) Metal inert gas welding; MIG welding; gas metal arc welding /USA/
- (135) Metal active gas welding; MAG welding; gas metal arc welding /USA/
- (136) Tubular-cored metal arc welding with active gas shield; flux cored arc welding /USA/
- (137) Tubular-cored metal arc welding with inert gas shield; flux cored arc welding /USA/
- (141) Tungsten inert gas arc welding; TIG welding; gas tungsten arc welding /USA/
- (15) Plasma arc welding;
- (31) Oxy-fuel gas welding; oxy-fuel gas welding /USA/
- (52) Laser beam welding;
- (912) Flame brazing; torch brazing /USA/
- (97) Braze welding;
- arc and flame gouging;
- arc and laser cutting processes;
- flame, plasma and laser and plasma cutting processes;
- metal-spraying (see EN ISO 4063).

The following gases and vapours which can be produced or be present during welding and allied processes are covered:

- ozone (O_3);
- carbon monoxide (CO);
- carbon dioxide (CO_2);
- nitric oxide (NO) and nitrogen dioxide (NO_2);
- vapours produced in the welding or cutting of metals having paint or other surface coatings.

Fuel, oxidant and shielding gases used in welding and allied processes are not covered.

The general background level of gases and vapours in the workplace atmosphere influences personal exposure, and therefore the role of fixed point measurements is also considered.

2 Normative references

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed here after. For dated references, subsequent amendments to or revisions of any of these incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies.

EN 175, *Personal protection — Equipment for eye and face protection during welding and allied processes.*

EN 482, *Workplace atmospheres — General requirements for the performance of procedures for the measurement of chemical agents.*

EN 689:1995, *Workplace atmospheres — Guidance for the assessment of exposure to chemical agents for comparison with limit values and measurement strategy.*

EN 838, *Workplace atmospheres — Requirements and test methods for diffusive samplers for the determination of gases and vapours.*

EN 1076, *Workplace atmospheres — Pumped sorbent tubes for the determination of gases and vapours — Requirements and test methods.*

EN 1231, *Workplace atmospheres — Short term detector tube measurement systems — Requirements and test methods.*

EN 1232, *Workplace atmospheres — Pumps for personal sampling of chemical agents — Requirements and test methods.*

EN 1540, *Workplace atmospheres — Terminology.*

EN ISO 4063, *Welding and allied processes — Nomenclature of processes and reference numbers.*

EN ISO 10882-1, *Health and safety in welding and allied processes — Sampling of airborne particles and gases in the operator's breathing zone — Part 1: Sampling of airborne particles.*

prEN 45544-1, *Workplace atmospheres — Electrical apparatus for the direct detection and direct concentration measurement of toxic gases and vapours — Part 1: General requirements and test methods.*

ISO 3534-1, *Statistics — Vocabulary and symbols — Part 1: Probability and general statistical terms.*

ISO 6879, *Air quality — Performance characteristics and related concepts for air quality methods.*

ISO 8756, *Air quality — Handling of temperature, pressure and relative humidity data.*

3 Terms and definitions

For the purposes of this part of EN ISO 10882, the following terms and definitions apply:

3.1

air sampling

process consisting of the collection, withdrawal or isolation of a fractional part of a larger volume of air. It can include the simultaneous isolation of selected components. (EN 1540)

3.2

bias

consistent deviation of the measured value from the value of the air quality characteristic itself or the accepted reference value. (ISO 6879)

3.3**breathing zone**

the space around the worker's face from where he takes his breath. For technical purposes a more precise definition is as follows: hemisphere (generally accepted to be 0,3 m in radius) extending in front of the human face, centred on the mid point of a line joining the ears; the base of the hemisphere is a plane through this line, the top of the head and the larynx. The definition is not applicable when respiratory protective equipment is used. (EN 1540)

NOTE This definition is not strictly applicable when a welder's face shield is used. In such circumstances the breathing zone should be considered to extend only behind the welder's face shield.

3.4**chemical agent**

any chemical element or compound, on its own or admixed as it occurs in the natural state or as produced by any work activity, whether or not produced intentionally and whether or not placed on the market.

NOTE This definition is taken from the "Council Directive 98/24/EC of 7 April 1998 on the protection of the health and safety of workers from the risks related to chemical agents at work". (EN 1540)

3.5**detector tube, diffusive; indicator tube, diffusive; colorimetric tube, diffusive**

a diffusion tube, similar in construction to a pumped detector tube. The length of the stain produced provides a measure of the exposure dose of a specified chemical agent in air, stated in ppm hours.

3.6**detector tube, pumped; indicator tube, pumped; colorimetric tube, pumped**

a glass tube containing chemical reagents in which a colour change may be produced when a sample of the atmosphere is drawn through it. The length of the stain produced provides a measure of the concentration of a specified chemical agent in air.

3.7**diffusion tube; diffusive tube; tube type diffusive sampler**

a diffusive sampler with a cross-sectional area which is small in relation to the internal air gap, across which the gas or vapour passes by diffusion to the sorbent.

3.8**diffusive badge; badge type diffusive sampler; passive badge**

a diffusive sampler in which the gas or vapour passes to the sorbent by permeation through a thin solid membrane or diffusion across a porous membrane. The cross-sectional area is large in relation to the internal air gap.

3.9**diffusive sampler; passive sampler**

a device which is capable of taking samples of gases or vapours from the atmosphere at a rate controlled by a physical process such as gaseous diffusion through a static air layer or permeation through a membrane, but which does not involve the active movement of air through the sampler. (EN 838)

3.10**direct reading electrical apparatus; direct reading instrument**

apparatus in which the presence of a gas or vapour causes a change that is manifest as an automatically generated electrical signal. When applied to a calibrated indicating or recording meter, this gives a direct measure of the concentration of the relevant gas or vapour.

3.11**exposure (by inhalation)**

a situation in which a chemical or biological agent is present in air which is inhaled by a person. (EN 1540)

3.12**fixed apparatus**

an apparatus which is intended to have all parts permanently installed. (prEN 45544-1)

3.13

harness

An assembly that provides a means of maintaining a welder's face shield in position on the head. (EN 175)

3.14

interferent

any component of the air sample, excluding the constituent(s) to be measured, affecting the instrument reading. (EN 1540)

3.15

limit value

reference figure for the concentration of a chemical or biological agent in air (EN 1540)

3.16

long term detector tube

a detector tube that provides a means of obtaining a measurement of the time weighted average concentration of a specified chemical agent in air

3.17

operator

a person who performs welding and allied processes.

3.18

overall uncertainty (of a measuring procedure or of an instrument)

quantity used to characterize as a whole the uncertainty of the result given by an apparatus or a measuring procedure. It is expressed, as a percentage, by a combination of bias and precision, usually according to the formula :

$$\frac{|\bar{x} - x_{ref}| + 2s}{x_{ref}} \times 100$$

where

$\bar{x}$ is the mean value of results of a number (n) of repeated measurements;

x_{ref} is the true or accepted reference value of concentration;

s is the standard deviation of measurements. (EN 1540)

3.19

personal sampler

a device attached to a person that samples air in the breathing zone. (EN 689)

3.20

personal sampling

the process of sampling carried out using a personal sampler. (EN 1540)

3.21

portable apparatus

spot reading or continuously sensing apparatus that has been designed to be readily carried from place to place and to be used whilst being carried. Portable apparatus is battery powered. (prEN 45544-1)

3.22

precision

the closeness of agreement of results obtained by applying the method several times under prescribed conditions. (ISO 6879)

3.23**reference period**

the specified period of time stated for the limit value of a specific chemical agent. (EN 1540)

3.24**sampler**

a device for sampling.

3.25**screening measurements of time weighted average concentration**

measurements performed to obtain relatively crude information on the exposure level in order to decide whether an exposure problem exists at all and if so to appraise its possible seriousness. They can also be used to determine if the exposure is well below or well above the limit value.

3.26**screening measurements of variation of concentration in time/and or space**

measurements performed to provide information on the likely pattern of concentration of chemical agents. They can be used to identify locations and periods of elevated exposure and to set the duration and frequency of sampling for measurements for comparison with limit values. Emission sources can be located and the effectiveness of ventilation or other technical measures can be estimated.

3.27**selectivity**

degree of independence from interferences. (EN 482)

3.28**short term detector tube**

a detector tube that provides a means of obtaining a rapid measurement (typically up to 15 min) of the concentration of a specified chemical agent in air. (EN 1231)

3.29**sorbent tube, pumped**

a tube, usually made of metal or glass, containing an active sorbent or reagent-impregnated support, through which sampled atmosphere is passed at a rate controlled by an air sampling pump. (EN 1076)

3.30**time weighted average (TWA) concentration**

the concentration of a chemical agent in the atmosphere, averaged over the reference period.

3.31**transportable apparatus**

an apparatus not intended to be portable, but which can be readily moved from one place to another. (prEN 45544-1)

3.32**true value**

the value which characterizes a quantity perfectly defined in the conditions which exist when that quantity is considered. (ISO 3534-1)

NOTE

The true value of a quantity is a theoretical concept and, in general, cannot be known exactly.

3.33**welder's face shield**

a welder's shield worn on the head and in front of the face, usually secured in position by a harness, to give protection to the eyes and face when fitted with appropriate filter(s). (EN 175)

3.34**welder's hand shield**

a welder's shield held in the hand to give protection to the eyes and face when fitted with appropriate filter(s). (EN 175)

3.35

welding episode

a period during which the operator carries out welding and allied processes, including welding related operations except when these generate a significant quantity of airborne particles e.g. during lengthy periods of grinding.

3.36

welding protector

a device which provides protection to the wearer harmful optical radiation and other specific hazards generated by welding and allied processes.

NOTE It may be a welder's shield, welder's goggles or welder's spectacles.

3.37

welding related operations

operations other than welding and allied processes carried out by the operator.

3.38

work pattern

the sequence of activities carried out by the worker during the period under consideration. (EN 1540)

3.39

workplace

the defined area or areas in which the work activities are carried out. (EN 1540)

3.40

worst case measurements

screening measurements of time weighted average concentration made to identify work activity during which highest exposure occurs.

4 Description of measurement methods

4.1 General

Personal exposure to gases and vapours in welding and allied processes is generally determined using:

- direct reading electrical apparatus;
- detector tubes (short term or long term); or
- indirect methods involving laboratory analysis.

Direct reading electrical apparatus or detector tubes are generally most applicable for measurement of gases. Indirect methods, which involve laboratory analysis of samples collected using a suitable solid or liquid sorbent, are most applicable for the determination of vapours which can be produced in the welding or cutting of metals having paint or other coatings.

A complex mixture of particulates and gases is produced in welding and allied processes, and, whatever method of analysis is selected, it is necessary to confirm that techniques which might have been used successfully in other applications are suitable for the welding situation.

In selecting any of the methods described, due regard should be paid to the possibility of interference with the determinations of one gas or vapour by the presence of another, which could result in either enhancement or reduction of the result.

4.2 Direct reading electrical apparatus

4.2.1 Applicability

Direct reading electrical apparatus is more widely applicable than other means for the measurement of personal exposure to gases in welding and allied processes, as it can be accurately calibrated and gives instantaneous results. It can be used to obtain a continuous record of concentrations throughout the sample period, and this can be integrated to determine the time weighted average concentrations.

Direct reading electrical apparatus is therefore useful for making screening measurements of variation of concentration in time, screening measurements of time weighted average concentration, measurements for comparison with limit values and periodic measurements.

4.2.2 Operating principles

Direct reading electrical apparatus usually operates by aspiration of a sample of the atmosphere in the operator's breathing zone into the apparatus through a sampling line, e.g. by means of a hand operated or electric air sampling pump. Direct reading electrical apparatus is also available which involves transfer of the gas or vapour from the atmosphere to the sensor by diffusion. However, such apparatus is generally unsuitable for making measurements of personal exposure to gases in welding and allied processes because it cannot be used with a sampling line and it is usually too large to fix in position behind the welder's face shield.

Measurement is made directly, or after reaction with solids, liquids or gases, usually by spectrophotometry or using an electrochemical sensor. The electrical signal produced is applied to an indicating or recording meter, which is normally calibrated to give a direct measure of the concentration of the relevant gas or vapour.

4.2.3 Availability

Fixed, transportable and portable apparatus is available for all the gases and some of the vapours covered in this guide. However, such apparatus is generally too large to be worn and it is necessary to place it in a remote position and aspirate samples through an extended sampling line. Small self-contained personal monitors are available for some gases and vapours, but these are primarily intended for protection of the operator against acute hazards, some incorporating alarms that can be set to operate at predetermined level. Such apparatus is normally less suitable than other means for measuring time weighted average concentrations.

4.3 Detector tubes

4.3.1 Applicability

The relative overall uncertainty and selectivity exhibited by detector tubes varies for different gases and vapours. Furthermore, measurements made using detector tubes usually exhibit a greater relative overall uncertainty than those obtained using direct reading apparatus or indirect methods involving laboratory analysis. Detector tubes are therefore most useful for screening measurements of time weighted average concentration rather than measurements for comparison with limit values or periodic measurements.

4.3.2 Pumped detector tubes

Pumped detector tubes are supplied in a sealed condition, and it is necessary to break off both ends of the tubes immediately before use. Use of a pumped detector tube to measure personal exposure to a gas or vapour in welding and allied processes involves aspiration of a sample of the atmosphere in the operator's breathing zone through the tube using an appropriate air sampling pump. If sampling is to be carried out by directly positioning the detector tube in the breathing zone, it is necessary to protect the operator from the exposed broken end by the attachment of a 10 mm length of plastics tubing. The air sampling pump and the detector tube, which together constitute a functional pumped detector tube measurement system, are calibrated for use with each other. It is therefore necessary that they are supplied by the same manufacturer.

Short term detector tubes are available for all the gases and many of the vapours covered in this standard. A discontinuously operating, hand operated or battery powered air sampling pump is used with the tubes, for which sampling periods are typically up to a few minutes. Repeated measurements are therefore necessary in order to determine time weighted average concentrations.

Long term detector tubes are available for most of the gases and many of the vapours covered in this standard. A continuously operating, battery powered air sampling pump is used with the tubes, for which sampling periods may be up to 8 h. Long term detector tubes are therefore much better suited than short term detector tubes for making measurements of 8 h time weighted average concentrations.

4.3.3 Diffusive detector tubes

Diffusive detector tubes are supplied in a sealed condition, and it is necessary to open them at the sample intake end immediately before use by breaking off the end at the appropriate breaking point. Use of a diffusive detector tube to measure the exposure dose of a gas or vapour to which an operator is subjected simply involves mounting it in a tube holder and positioning it in the breathing zone for an appropriate sampling period, which may be up to 8 h.

Diffusive detector tubes are only available for a few of the gases and vapours covered in this standard.

4.4 Indirect methods involving laboratory analysis

4.4.1 Applicability

Indirect methods involving laboratory analysis have separate sampling and analysis stages. They are the most widely applicable for measurement of 8 h time weighted average concentrations of vapours, such as those produced in the welding or cutting of metals having paint or other coatings, because they exhibit good relative overall uncertainty and selectivity. However, they are seldom applicable for the determination of the gases covered in this guide, which are better determined using direct reading electrical apparatus.

4.4.2 Pumped sampler methods

4.4.2.1 Pumped sorbent tube methods

Sorbent tubes made of glass are supplied in a sealed condition, and it is necessary to break off both ends of tubes immediately before use. Sorbent tubes made of metal have removable end caps. Use of a sorbent tube to measure personal exposure to gases or vapours in welding and allied processes involves aspiration of a sample of the atmosphere in the operator's breathing zone through the tube using a continuously operating, battery powered air sampling pump. Sampling periods may be up to 8 h. If sampling is to be carried out by positioning a sorbent tube made of glass in the breathing zone, it is necessary to protect the operator from the exposed broken end by the attachment of a 10 mm length of plastics tubing. After sampling, both ends of the sorbent tube are closed with end caps.

Sorbent tube methods are available for some of the gases covered in this guide, and for most of the vapours that can be produced in significant amounts during welding or cutting of metals having paint or other coatings. The analytical technique used varies according to which gases and vapours are determined. However, the analyte is typically desorbed with solvent (solvent desorption) or by heat (thermal desorption), usually followed by gas or high performance liquid chromatography or occasionally spectrophotometry.

The sampling of reactive compounds (e.g. aldehydes) can also be made using reagent-impregnated filters or silica gel tubes. Such methods may be used for making exposure measurements in welding and allied processes.

4.4.2.2 Liquid sorbent methods

Some gases and vapours may be determined by laboratory analysis of liquid sorbent, contained in an absorption tube known as a bubbler, through which the test atmosphere is drawn at an appropriate rate using an air sampling pump. In some instances a reagent solution is used, with which the gas or vapour reacts to yield the species to be measured in the subsequent analysis. Alternatively, the gas or vapour of interest may simply be absorbed in the liquid through which the air is drawn.

However, various practical difficulties associated with the use of liquid sorbent methods make them generally unsuitable for use in the assessment of personal exposure to gases and vapours in welding and allied processes.

4.4.2.3 Gas sampling bag methods

Certain gases may be aspirated into specially designed gas sampling bags for subsequent laboratory analysis e.g. by gas chromatography. The manufacturer's instructions should be followed when using such methods.

4.4.3 Diffusive sampler methods

4.4.3.1 Applicability

Diffusive samplers are particularly convenient devices for personal sampling in welding and allied processes, since there is no requirement for an air sampling pump to be worn or a sampling line to be attached to the operator. Diffusive samplers may be used to measure the exposure dose of a gas or vapour to which an operator is subjected. This simply involves positioning the device in the breathing zone for an appropriate sampling period, which may be up to 8 h. The analytical technique used varies for different gases and vapours. Occasionally, the colour change produced in a diffusive badge may give an indication of the concentration of a gas or vapour, but more commonly analysis involves solvent or thermal desorption, followed by measurement using gas or high performance liquid chromatography.

4.4.3.2 Diffusive badge methods

The sampling characteristics of diffusive badges are such that a representative sample of the gas or vapour might not be collected if there is insufficient air movement across the face of the device. In general, air movement resulting from normal work activity is sufficient to enable diffusive badges to be used for personal sampling, but they might be unsuitable for sampling inside welder's face shields, where there can be limited air movement.

Diffusive badge methods are available for many of the vapours that can be produced during welding or cutting of metals having paint or other coatings. Diffusive badges are also available for the gases covered in this standard, but their use is not well documented.

4.4.3.3 Diffusion tube methods

The sampling characteristics of diffusion tubes are such that the amount of gas or vapour collected is not affected significantly by air movement across the face of the device. They should therefore be suitable for sampling gases and vapours in the operator's breathing zone.

Diffusion tube methods are available for many of the vapours that can be produced during welding or cutting of metals having paint or other coatings. However, devices are only available for a few of the gases covered in this standard.

5 Requirements

Procedures used for assessment of personal exposure to gases and vapours in welding and allied processes shall comply with the provisions of EN 482.

6 Assessment strategy

EN 689 gives guidance for the assessment of exposure by inhalation to chemical agents in workplace atmospheres. The generalised assessment strategy it describes should be observed.

This could involve (i) an initial appraisal of the likelihood of personal exposure to gases and vapours in welding and allied processes, e.g. by using known information about the process and workplace factors; (ii) a basic survey to provide quantitative information about likely exposure, e.g. by using existing exposure data from comparable processes; and (iii) a more detailed survey involving workplace measurements.

7 Measurement strategy

7.1 General

EN 689 should be referred to for generalised guidance on measurement strategy. It highlights the need to take an approach which makes the most efficient use of resources by making full use of screening measurements and worst case measurements when it is suspected that exposure levels are well below or above the limit values. The following guidance, specific for measurement of personal exposure to gases and vapours in welding and allied processes, is given to supplement that given in EN 689.

7.2 Personal exposure measurements

The highest concentrations of gases usually occur in the immediate vicinity of the operator and it is therefore essential that personal exposure measurements are performed in the operator's breathing zone. Welder's face shields can provide some degree of protection from exposure by physically deflecting the welding plume away from the breathing zone. Personal sampling should therefore take place behind such welding protectors, if used, rather than in the conventional position on the lapel. Measurements of very short duration, e.g. for screening purposes, can be made by directly positioning the sampling line or sampler in the breathing zone and maintaining it in position by hand. However, to obtain time weighted average concentrations of gases and vapours, e.g. for comparison with limit values, it will usually be necessary for the operator to wear special apparatus that enables the sampler to be maintained in position in the breathing zone throughout the sampling period, without impeding normal work activity.

Personal exposure to gases and vapours in welding and allied processes is very variable over a single work period. The pattern of exposure depends on the nature of the job and its location, the use of hygiene controls, the work technique of the operator and the work pattern. Many of these variables are subject to the control of the operator and sampling strategies designed to assess the exposure of one individual as representative of a group carrying out similar work are often not appropriate. Each operator should therefore be the subject of a separate assessment, which does not necessarily imply a separate measurement of exposure.

7.3 Fixed point measurements

Fixed point measurements are required to characterise the background level of gases and vapours in the workplace. They can be useful for assessment of the personal exposure to gases and vapours of workers in adjacent locations or in overhead cranes, and they can give an indication of the efficiency of ventilation.

7.4 Selection of measurement conditions and measurement pattern

7.4.1 General

It is necessary to devise the sampling procedure so as to cause the least possible interference with the operator and the normal performance of his job and to provide samples that are representative of normal working conditions and that are compatible with subsequent methods of analysis. The pattern of measurement should take into consideration practical issues, such as the frequency and duration of welding episodes, and the nature of the measurement task.

7.4.2 Screening measurements of time weighted average concentration and worst case measurements

Screening measurements of time weighted average concentration may be carried out in the initial stages of any survey to assess the effectiveness of control measures. Sampling shall be repeated during representative welding episodes to obtain clear information about the level and pattern of exposure. If results indicate that concentrations of gases or vapours are well below the limit value there is adequate control, and measurements of time weighted average concentrations for comparison with limit values are not necessary. If results indicate that the concentration of gas or vapour is above the limit value, control measures could be inadequate. In such instances, control measures in place should be reviewed and screening measurements repeated after improvements have been made.

Detector tubes and direct reading electrical apparatus are generally most applicable for making screening measurements of time weighted average concentration.

7.4.3 Measurements for comparison with limit values and periodic measurements

Direct reading electrical apparatus is generally most applicable for making measurements of time weighted average concentrations of gases in welding and allied processes. Indirect methods involving laboratory analysis are generally most applicable for making measurements of time weighted average concentrations of vapours.

When appropriate, measurements for comparison with short term limit values and associated periodic measurements should be made over a sampling period equal to the reference period, which is normally between 5 and 30 minutes.

The best estimates of time weighted average concentrations of chemical agents are typically obtained by making long term measurements in the breathing zone over the entire working period. However, personal exposure to gases and vapours in welding and allied processes is intermittent. Highest exposure occurs during welding episodes, and episodes in between give rise to minimal exposure. Time weighted average concentrations over an 8 h reference period should therefore be calculated from the results of a series of measurements made during a number of representative welding episodes, as described in Annex B of EN 689:1995.

8 Sampling

8.1 Sampling position

8.1.1 Personal sampling

Sampling should be performed in the breathing zone:

If a welder's face shield is used, the sampler should be positioned adjacent to the operator's nose and mouth, at mouth level, a maximum distance of 50 mm to the right or left of the mouth.

If an alternative welding protector is used, e.g. a welder's hand shield, the sampler should be positioned as close as possible to the operator's nose and mouth.

If the sampler cannot be positioned directly in the breathing zone, the sampled air should be drawn from the breathing zone through small bore tubing fastened to a convenient point on the operator's body, e.g. the waist or the small of the back. Connections can then be made to personal monitors or other devices attached to the operator's body, or, via a sampling line long enough not to impede the operator's movements, to remote direct reading electrical apparatus.

See 4.3.2 and 4.4.2.1 regarding protection of the broken tube end of glass tubes. See Annex B of EN ISO 10882-1 for examples of arrangements for mounting a filter holder behind a welder's face shield, which might be necessary if the sampled air is to be filtered to remove airborne particles before its introduction into the sampler or measurement system (see 8.3).

8.1.2 Fixed point sampling

If fixed point sampling to determine the general background level of gases and vapours in the workplace atmosphere is to be carried out, the sampling position selected should be sufficiently remote from welding and allied processes that results are not directly affected by emission sources. Ventilation, local circumstances and the information that is being sought should be taken into consideration.

8.2 Sampling equipment

8.2.1 Direct reading electrical apparatus

Direct reading electrical apparatus used for the assessment of personal exposure to gases and vapours in welding and allied processes should comply with the provisions of prEN 45544-1.

8.2.2 Detector tubes

Detector tubes used for the assessment of personal exposure to gases and vapours in welding and allied processes should comply with the provisions of EN 1231.

8.2.3 Pumped sorbent tubes

Sorbent tube methods for the assessment of personal exposure to gases and vapours in welding and allied processes should comply with the provisions of prEN 1076. Air sampling pumps used in such methods should comply with the provisions of EN 1232.

8.2.4 Diffusive samplers

Diffusive detector tubes used for the assessment of personal exposure to gases and vapours in welding and allied processes and diffusive samplers used in indirect methods should comply with the provisions of EN 838.

8.2.5 Construction materials

Filters, filter holders, sampling lines, connectors etc., with which the sampled air might come into contact before its introduction into the direct reading electrical apparatus, detector tube or sampler, should be made of material that is inert to the gases being sampled, to prevent adsorption or reaction. The sampling line should be resistant to kinking and, for personal sampling, it should also be flexible. Tubing manufactured from plasticized polyvinylchloride (PVC) is normally suitable, except for ozone (see 9.2.1).

8.3 Sample filtration

When using a measurement method that can be affected by particulate matter, the sampled air should be filtered to remove airborne particles before its introduction into the sampler or measurement system. This might not be necessary for direct reading electrical apparatus if it incorporates an internal filter.

Filtration should be carried at the point of sampling, if possible. This is particularly important for ozone measurement, as it is necessary to keep the sample line clean to prevent against low results. If filtration at the point of sampling is not possible, a filter should be inserted into the sampling line at a convenient point between the sampling position and the analytical equipment, but in this case there should be regular checking and cleaning of the sample lines.

8.4 Multiple sampling

When the concentration of more than one gas is to be measured, successive samples may be taken through a single sample line, or it might be possible to make successive measurements on a single sample.

Where simultaneous sampling is required, either multiple sample lines should be used or branch sampling lines should be taken from a single sampling line.

8.5 Volume of sampling line

The length and internal volume of the sampling line, and of filter holders if used, should be kept to a minimum and the residence time of the air in them should not exceed a few seconds. If a discrete volume of air is sampled, the volume of the sampling line plus filter holder, if used, should not exceed 5 % of the volume of the discrete sample.

8.6 Flow rate

Direct reading electrical apparatus with an integral air sampling pump should be used in accordance with the manufacturer's instructions.

For other methods, the flow rate at which the atmosphere is sampled should be compatible with the detector tube or sampler used. Refer to the manufacturer's instructions.

8.7 Handling of temperature, pressure and humidity data

ISO 8756 gives procedures for adjusting air quality measurements for changes in temperature, pressure and humidity during the sampling period and states the reference conditions of temperature, pressure and humidity to be used in reporting the results. The requirements of ISO 8756 should be taken into consideration when using the measurement methods referred to in this standard.

9 Measurement of individual gases and vapours

9.1 General

This clause gives specific guidance on measurement methods available for the gases and vapours covered in this part of EN ISO 10882. Values quoted in parentheses after each gas represent the concentration range of principal interest, and this has influenced the choice of methods referred to in the standard. The information given in 9.2 through 9.6 is summarised in Annex B.

9.2 Ozone (0,01 ppm to 3 ppm)

9.2.1 Special sampling requirements

For ozone sampling, it is essential that filters, filter holders, sampling line, connectors etc., with which the sampled air could come into contact before its introduction into the measurement system, are made of polytetrafluoroethylene (PTFE). PTFE filters should be changed frequently to prevent the build up of particulate deposits which could react with ozone and reduce the measured concentration.

To ensure that the sample collection system will not reduce ozone in the sampled air, it is necessary to pass air containing not less than 1 ppm ozone from an ozone generator through the collection system for at least 10 min before sampling commences and after each filter change.

9.2.2 Direct reading electrical apparatus

Direct reading electrical apparatus used for measurement of ozone most commonly works on one of the following principles:

- a) Photometry of the chemiluminescence produced by reaction of ozone with ethylene. The reaction is specific for ozone and is extremely sensitive;
- b) Absorption of ultraviolet radiation. The method is made specific for ozone by differential measurements on the air sample before and after removal of the ozone by a catalytic converter. Unlike other systems, it can be calibrated directly from knowledge of the absorption path length and the absorption coefficient for ozone.

The apparatus should be calibrated with air of known ozone concentration from an ozone generator calibrated against a primary standard with the complete sampling train in place.

9.2.3 Detector tubes

The use of short term detector tubes for ozone is limited, e.g. to screening measurements.

9.2.4 Indirect methods involving laboratory analysis

Indirect methods are not generally applicable to the measurement of ozone. However, a validated indirect method for ozone has been described [1] involving oxidation of nitrite on an impregnated glass fibre filter and subsequent determination of the resulting nitrate.

9.3 Carbon monoxide (3 ppm to 500 ppm)

9.3.1 Direct reading electrical apparatus

Direct reading electrical apparatus used most commonly for measurement of carbon monoxide works on one of the following principles:

- a) Dispersive infra-red absorption and non-dispersive infra-red absorption used with or without filters to reduce interference by carbon dioxide. Interfering compounds include acetylene, nitrous oxide, olefins, carbon dioxide and water vapour. However, carbon dioxide is unlikely to interfere with

dispersive methods or non-dispersive methods used with a suitable filter at concentrations of carbon dioxide less than 3 %.

- b) Diffusion of the carbon monoxide through a semi-permeable membrane, at a rate proportional to the concentration, followed by electrochemical oxidation of the gas at a potential-controlled electrode and measurement of the current produced.
- c) Gas chromatography.

The apparatus should be calibrated with certified gas mixtures of known carbon monoxide concentration. (Suitable concentrations of carbon monoxide in air and carbon monoxide in nitrogen are available).

9.3.2 9Detector tubes

Short term detector tubes, long term detector tubes, and diffusive detector tubes are available for carbon monoxide, and are useful for measurement of personal exposure in welding and allied processes.

9.3.3 Indirect methods involving laboratory analysis

Indirect measurement of carbon monoxide by gas chromatography has been described [1] after collection in a gas sampling bag.

9.4 Carbon dioxide (500 ppm to 10 %)

9.4.1 Origin

Carbon dioxide could be present as a combustion or decomposition product of the welding or cutting process, or because of its use as a shielding gas or from exhaled breath.

9.4.2 Direct reading electrical apparatus

Direct reading electrical apparatus used most commonly for measurement of carbon dioxide works on by non-dispersive infra-red absorption. It should be calibrated using standard gas mixtures of known carbon dioxide concentration.

9.4.3 Detector tubes

Short term detector tubes, long term detector tubes, and diffusive detector tubes are available for carbon dioxide, and are useful for measurement of personal exposure in welding and allied processes.

9.4.4 Indirect methods involving laboratory analysis

Indirect measurement of carbon dioxide by gas chromatography has been described [1] after collection in a gas sampling bag.

9.5 Nitric oxide (1 ppm to 100 ppm) and nitrogen dioxide (0,3 ppm to 250 ppm)

9.5.1 General

Nitric oxide and nitrogen dioxide are considered together since they generally occur together in welding and allied processes. They may be determined separately or together, as oxides of nitrogen, using related methods.

9.5.2 Direct reading electrical apparatus

Direct reading electrical apparatus used most commonly for measurement of nitric oxide and nitrogen dioxide works on one of the following principles:

- a) Measurement of chemiluminescence produced by the reaction between nitric oxide and ozone (which yields activated nitrogen dioxide). Nitrogen dioxide present in the sample may be determined by the difference between the nitric oxide concentration before and after conversion of nitrogen dioxide to nitric oxide;
- b) Measurement of the signal generated by electrochemical reaction of nitric oxide and nitrogen dioxide at catalytically-active, potential-controlled electrodes in aqueous sulphuric acid. By using two sensing electrodes at appropriate potentials, each gas may be determined in the presence of the other.

The apparatus should be calibrated using suitable certified gas mixtures, such as nitric oxide in nitrogen and nitrogen dioxide in nitrogen.

9.5.3 Detector tubes

Short term detector tubes and long term detector tubes are available for measurement of both nitrogen dioxide and oxides of nitrogen. Short term detector tubes are particularly useful for making screening measurements. By using tubes for both nitrogen dioxide and oxides of nitrogen, nitric oxide may be estimated by difference. Diffusive detector tubes are available for nitrogen dioxide.

9.5.4 Indirect methods involving laboratory analysis

Indirect measurement of nitrogen dioxide by ion chromatography has been described [1] after collection using a pumped sorbent tube containing triethanolamine impregnated molecular sieve (TEA-IMS), and a similar method has been described [2] with analysis by spectrophotometry. A diffusive badge method has also been described [3] for nitrogen dioxide with collection on triethanolamine coated screens and analysis by spectrophotometry.

Indirect measurement of nitric oxide by ion chromatography has been described [1] after collection using in a three stage pumped sorbent tube. TEA-IMS in the front section of the tube collects nitrogen dioxide, and TEA-IMS in the rear section of the tube collects nitrogen dioxide produced after reaction of nitric oxide with oxidizer positioned between the two TEA-IMS sections. A similar method has been described [2] with analysis by spectrophotometry.

9.6 Vapours

9.6.1 General

A complex mixture of vapours and particulates can be produced in the welding or cutting of metals having paint or other surface coatings, the composition of which will be dependent upon the nature of the surface coating. Substances which may be determined [4] include the following:

Acids	e.g. formic acid, acetic acid, aliphatic acids, benzoic acid, hydrochloric acid, phthalic anhydride
Phenols	e.g. phenol, alkyl phenols, bisphenol-A
Aldehydes	e.g. formaldehyde, acetaldehyde, acrolein, butyraldehyde, aliphatic aldehydes
Esters	e.g. methylmethacrylate, butylmethacrylate
Hydrocarbons	e.g. aliphatic and aromatic hydrocarbons, alkylbenzenes, polyaromatic hydrocarbons
Di-isocyanates	e.g. toluene di-isocyanate

These are too numerous for detailed consideration in this part of EN ISO 10882, but some general guidance on sampling and analysis is given in 9.6.2 through 9.6.4.

9.6.2 Direct reading electrical apparatus

Direct reading electrical apparatus is available for many of the volatile organics mentioned in 9.6.1. However, because of its poor specificity and the complex mixture of chemical agents involved, such apparatus is not generally useful for measurement of personal exposure to vapours in welding and allied processes.

9.6.3 Detector tubes

Detector tubes are available for many of the volatile organics mentioned in 9.6.1, but most exhibit relatively poor specificity. Detector tubes therefore have limited use in estimating personal exposure to vapours in welding and allied processes.

9.6.4 Indirect methods involving laboratory analysis

Indirect methods are most useful techniques for determining personal exposure to vapours in welding and allied processes.

Vapours are typically sampled using diffusive sorbent tubes, pumped sorbent tubes or impregnated filters. Many methods have been published [5, 6, 7], and most should be applicable to determination of personal exposure to vapours in welding and allied processes, although it is necessary to use a pre-filter with pumped sorbent tubes to prevent particulate matter clogging the tube.

Volatile organics adsorbed on airborne particles are collected on filters, as described in EN ISO 10882-1. Analysis is generally by solvent desorption followed by high performance liquid chromatography or gas chromatography.

10 Recording of test data and presentation of results

In order to obtain the full benefit of measurements of gases and vapours, it is necessary to record as much information as possible about the process itself, the materials used and the particular circumstances of each test. Such information is essential for the correct interpretation of results by occupational health/hygiene experts.

An example of the type of information that should be recorded in the test report is given in Annex C. Basic data related to the test are recorded in C.1, process specific data are recorded in C.2, and sampling data and test results are recorded in C.3, together with calculated 8-hour Time Weighted Average exposures.