
**Sterilization of health care products —
Chemical indicators —**

Part 4:

**Class 2 indicators as an alternative to the
Bowie and Dick-type test for detection of
steam penetration**

Stérilisation des produits de santé — Indicateurs chimiques —

*Partie 4: Indicateurs de Classe 2 comme alternative à l'essai de Bowie
et Dick pour la détection de la pénétration de la vapeur*



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 11140-4:2007

© ISO 2007

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 General requirements	3
5 Indicator system format	4
6 Performance requirements	4
7 Packaging and labelling	6
8 Quality assurance	7
Annex A (normative) Determination of indicator strength during and after steam sterilization	8
Annex B (normative) Standard test cycles	10
Annex C (normative) Estimation of visual difference between colour of the substrate and of the changed or unchanged indicator system by determination of relative reflectance density	15
Annex D (normative) Determination of uniform colour change on exposure to saturated steam	19
Annex E (normative) Determination of equivalence of the alternative indicator to the Bowie and Dick test	20
Annex F (normative) Determination of reproducibility of fail conditions created in a standard test pack by air injection, air leak and retained air systems	22
Annex G (normative) Determination of indicator colour change on exposure to dry heat	26
Annex H (normative) Determination of shelf life of product	27
Annex I (normative) Accelerated ageing of test samples	28
Annex J (normative) Steam exposure apparatus and steam for test purposes	29
Annex K (normative) Standard test pack	32
Annex L (normative) Air injection system	33
Bibliography	35

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 11140-4 was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*.

This second edition cancels and replaces the first edition (ISO 11140-4:2001) which has been technically revised.

ISO 11140 consists of the following parts, under the general title *Sterilization of health care products — Chemical indicators*:

- *Part 1: General requirements*
- *Part 3: Class 2 indicator systems for use in the Bowie and Dick-type steam penetration test*
- *Part 4: Class 2 indicators as an alternative to the Bowie and Dick-type test for detection of steam penetration*
- *Part 5: Class 2 indicators for Bowie and Dick-type air removal tests*

Introduction

The Bowie and Dick test is a performance test for steam sterilizers for wrapped goods and porous loads. As such it is performed during the demonstration of conformance of steam sterilizers to EN 285 and as a routine test of performance in ISO 17665-1. The test method is described in EN 285.

A failure of the Bowie and Dick test is symptomatic of a number of potential problems with the sterilizer that could compromise the uniform sterilization of a load to be processed. This failure is not conclusive proof that the fault in the sterilizer is due to air retention, air leakage or non-condensable gases, and it can be necessary to investigate other causes of failure.

The Bowie and Dick test was conceived as a test for successful air removal from high-vacuum porous-load sterilizers used in the sterilization of health care products ^[1]. A successful Bowie and Dick test indicates rapid and even penetration of steam into the test pack. The presence of air within the pack, due to an inefficient air removal stage, an air leak during this stage or non-condensable gases in the steam supply, are circumstances which can lead to failure of the test. The result of the test may also be affected by other factors which inhibit steam penetration. The test does not necessarily demonstrate either achievement of the required temperature or maintenance of that temperature for the required time to achieve sterilization.

A test pack for the Bowie and Dick test consists of two components:

- a) a small standardized test load;
- b) a chemical indicator system to detect the presence of steam.

The Bowie and Dick test as originally described ^[1] utilized huckaback towels as the material for the test load. The test as described in EN 285 uses cotton sheets for this purpose.

Indicators intended as an alternative to the Bowie and Dick test use different materials for the test load and employ indicator systems specifically formulated for use with the defined test load. Because a range of different tests in different countries have historically been termed the Bowie and Dick test, the term "Bowie and Dick-type test" is used in this part of ISO 11140.

This part of ISO 11140 specifies the performance of the indicator system in combination with the test load with which it is intended to be used. The test load may be presented with the indicator system already incorporated and intended for single use, or it may be intended for multiple use with a new indicator system to be inserted prior to each use.

The indicator for which the performance is specified in this part of ISO 11140 is intended to indicate when steam penetration has been inadequate. The performance of the indicator specified in this part of ISO 11140 should be equivalent, but not necessarily identical, to the performance obtained in the Bowie and Dick-type test as described in ISO 11140-3. Equivalence should be regarded as providing a similar response to steam penetration with any differences being predictable and such that the necessary level of assurance of satisfactory steam penetration is provided. An indicator meeting this specification is not intended to identify which of the potential causes of poor steam penetration was responsible for the failure indicated by the test.

STANDARDSISO.COM : Click to view the full PDF of ISO 11140-4:2007

Sterilization of health care products — Chemical indicators —

Part 4:

Class 2 indicators as an alternative to the Bowie and Dick-type test for detection of steam penetration

1 Scope

This part of ISO 11140 specifies the performance for a Class 2 indicator to be used as an alternative to the Bowie and Dick-type test for steam sterilizers for wrapped health care goods (instruments, etc., and porous loads).

NOTE The Bowie and Dick-type test is used for routine testing of steam sterilizers and validation of steam sterilization processes.

An indicator complying with this part of ISO 11140 incorporates a specified material which is used as a test load. This test load may, or may not, be re-usable. This part of ISO 11140 does not specify requirements for the test load, but specifies the performance of the indicator in combination with the test load with which it is intended to be used. The indicator specified in this part of ISO 11140 is intended to identify poor steam penetration but does not necessarily indicate the cause of this poor steam penetration.

This part of ISO 11140 does not include test methods to establish the suitability of these indicator systems for use in sterilizers in which the air removal stage does not include evacuation below atmospheric pressure.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5-1, *Photography — Density measurements — Part 1: Terms, symbols and notations*

ISO 5-3, *Photography — Density measurements — Part 3: Spectral conditions*

ISO 5-4:1995, *Photography — Density measurements — Part 4: Geometric conditions for reflection density*

ISO 187:1990, *Paper, board and pulps — Standard atmosphere for conditioning and testing and procedure for monitoring the atmosphere and conditioning of samples*

ISO 2248, *Packaging — Complete, filled transport packages — Vertical impact test by dropping*

ISO 10012-1, *Quality assurance requirements for measuring equipment — Part 1: Metrological confirmation system for measuring equipment*

ISO 11140-1:2005, *Sterilization of health care products — Chemical indicators — Part 1: General requirements*

ISO/CIE 10526:1999, *CIE standard illuminants for colorimetry*

IEC 60584-2:1982, *Thermocouples. Part 2: Tolerances*

IEC 60584-2/am1:1989, *Amendment 1 — Thermocouples. Part 2: Tolerances*

IEC 60751:1983, *Industrial platinum resistance thermometer sensors*

IEC 60751/am1:1986, *Amendment 1 — Industrial platinum resistance thermometer sensors*

EN 285:2006, *Sterilization — Steam sterilizers — Large sterilizers*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 11140-1 and the following apply.

3.1

air pocket

concentration of residual, induced or injected air or non-condensable gases in the standard test pack

3.2

chamber reference temperature

temperature measured at a defined reference point within the steam exposure apparatus

NOTE The defined reference point is usually located in the chamber drain or active chamber discharge.

3.3

exposure time

period for which the chamber reference temperature lies within the sterilization temperature band

3.4

pre-assembled pack

indicator in which the indicator system is incorporated into the test load during the manufacturing process and which is supplied ready for use

3.5

reference fault period

period of 30 s commencing when the chamber reference temperature attains the set operating temperature

3.6

sterilization temperature

minimum temperature of the sterilization temperature band

NOTE The use of the word “sterilization” within this and other definitions is not intended to imply that sterilizing conditions will take place under the test cycle conditions.

3.7

sterilization temperature band

range of temperatures from the sterilization temperature to the maximum allowable temperature which may prevail throughout the load during the holding time

NOTE These temperatures are usually stated in whole degrees centigrade.

3.8

temperature depression

thermodynamic temperature difference in kelvin given by (chamber reference temperature, in degrees centigrade) minus (temperature in the standard test pack, in degrees centigrade)

3.9**test equilibration time**

time elapsed after the chamber reference temperature attains the set operating temperature until the temperature within the standard test pack is the same as the chamber reference temperature, within the limits of accuracy of the temperature-measuring equipment

3.10**user-assembled pack**

indicator in which the user combines the indicator system with the test load prior to use

4 General requirements

4.1 The requirements of ISO 11140-1 apply.

4.2 Test samples shall be conditioned in accordance with ISO 187 prior to testing for performance.

4.3 Compliance with the requirements of this part of ISO 11140 shall be determined by establishing conformity with the performance requirements of Clause 6.

4.4 The indicator shall have sufficient strength to withstand steam sterilization and subsequent handling.

Compliance shall be tested in accordance with Annex A.

4.5 Test cycles for demonstrating compliance with the requirements of this part of ISO 11140 shall employ sub-atmospheric, trans-atmospheric and super-atmospheric air removal stages (see Table 1 and B.1, B.2 and B.3) except when the indicator, or indicator system, is intended solely for use with one type of air removal system, in which case only the specified air removal system needs to be used during compliance testing.

4.6 A thermometric recording instrument(s) shall be used in conjunction with temperature sensors to record the temperatures measured in the locations specified in the tests described in this part of ISO 11140. The temperature-measuring equipment used in all test methods for demonstrating compliance with this part of ISO 11140 shall meet the following requirements.

- a) Temperature sensors shall be either platinum resistance and comply with Class A of IEC 60751:1983 and IEC 60751 Amendment 1:1986 or a thermocouple and comply with one of the tables of tolerance class 1 of IEC 60584-2:1982 and IEC 60584-2 Amendment 1:1989.
- b) The performance characteristic of the temperature sensor shall not be affected by the environment in which it is used, e.g. pressure, steam or vacuum.
- c) The temperature sensors shall have a response time in water of $\tau_{90} \leq 0,5$ s.
- d) The temperature measured by all temperature sensors when immersed in a temperature source at a temperature known to within $\pm 0,1$ K, and within the sterilization temperature band, shall not differ by more than 0,5 K.
- e) The recording instrument shall record the temperature from a minimum of 6 sensors. The sampling interval shall not exceed 2,5 s. All data sampled shall be used for the interpretation of results.
- f) The scale range shall include 0 °C to 150 °C. For analogue instruments, the minor mark interval shall not exceed 1 K, the resolution shall be not less than 0,5 K and the chart speed shall be not less than 15 mm/min. Digital instruments shall register and record in increments of not more than 0,1 K.
- g) The limit of error of the recording instrument between 0 °C and 150 °C (excluding temperature sensors) shall not exceed 0,25 % when tested in an ambient temperature of (20 ± 3) °C. The additional error due to change in the environmental temperature shall not exceed 0,04 K/K.
- h) Calibration shall be carried out using a working or reference standard that is traceable to a national standard or a primary standard. The instrument shall have a valid test certificate.

5 Indicator system format

5.1 When the indicator system is one in which the indicator reagent is distributed on a substrate, it shall meet the following requirements.

- a) The indicator reagent shall be distributed to cover not less than 30 % of the surface area of the substrate. The distance between adjacent areas of indicator reagent shall not exceed 20 mm.

The pattern of indicator reagent distribution should permit clear interpretation of the colour change.

- b) The substrate shall have a colour which is uniform to visual observation.
- c) The indicator system shall have a difference in relative reflectance density of not less than 0,3 between the colour of the substrate and either the changed indicator or unchanged indicator as specified by the manufacturer.

Compliance shall be tested in accordance with Annex C.

5.2 When the indicator system depends on migration of the indicator reagent to demonstrate change, the pattern of indicator reagent distribution before and after use shall permit clear interpretation of the result.

5.3 When the indicator system is intended for use with a user-assembled pack, the indicator system shall permit writing in permanent ink to be made legibly on both processed and unprocessed materials. Those markings made before processing shall remain legible after processing.

5.4 When the indicator system is provided by the manufacturer already incorporated into the test load, the material of either the indicator or the indicator system, as appropriate, shall permit writing to be made after processing.

6 Performance requirements

6.1 The indicator, when tested in combination with the test load specified by the manufacturer, shall show a uniform colour change complying with 5.1 c) after exposure to saturated steam at 134 °C for 3,5 min, or at 121 °C for 15 min or at any other time/temperature combination specified by the manufacturer when the temperature tolerance shall be $\begin{pmatrix} +1,5 \\ 0 \end{pmatrix}$ °C and the time tolerance shall be ± 5 s.

Compliance shall be tested in accordance with Annex D using the steam exposure apparatus. The steam exposure apparatus shall be operated with the standard test cycles described in Annex B as shown in Table 1.

Indicators intended for use only with specific air removal cycles shall be tested with those specific cycles only (see 5.7 of ISO 11140-1:2005).

Indicators intended to be used over a wide range of sterilization temperatures, e.g. both for cycles operating at 121 °C and for those operating at 134 °C, may not give the same depth or intensity of colour change at both temperatures. This should be regarded as in compliance if:

- a) all other performance characteristics required by this part of ISO 11140 are met;
- b) the nature of the colour change is unambiguously defined in the instructions for use (see 5.8 of ISO 11140-1:2005).

6.2 The indicator shall show no colour change, incomplete colour change, or uneven colour change when exposed to a test cycle previously demonstrated to produce a reference fault condition (a fault response). Exposure to a reference fault condition shall produce a fault response regardless of the means of creating the reference fault condition, i.e. the system used to produce the fault may use air retention, air leak or air injection. The test cycles used to generate the reference fault conditions shall be as shown in Table 1. The

chamber reference temperatures and holding times shall include 134 °C for 3,5 min, or 121 °C for 15 min or another time/temperature combination specified by the manufacturer (see 6.1) when the temperature tolerance shall be $\begin{pmatrix} +1,5 \\ 0 \end{pmatrix}$ °C and the time tolerance shall be ± 5 s.

Compliance shall be tested in accordance with Annex E.

Compliance with the fault condition reproducibility shall be demonstrated in accordance with Annex F.

Table 1 — Schedule of test cycles to be used

Test condition	Standard test cycle of Annex B		
	B.1 Sub-atmospheric pulsing	B.2 Trans-atmospheric pulsing	B.3 Super-atmospheric pulsing
"Pass" cycle (see 6.1)	√	√	√
"Fail" cycle – modified air removal stage (see 6.2)	√	√	x
"Fail" cycle – induced leak (see 6.2)	√	x	x
"Fail" cycle – air injection (see 6.2)	√	x	√
√ = test required; x = test not required.			

6.3 The indicator system shall show no discernible colour change after exposure to dry heat at (140 ± 2) °C for not less than 30 min.

With some indicators, the indicator system can show a slight colour change after exposure to dry heat; this shall be acceptable if the change that occurs is slight or markedly different from that brought about by exposure to steam in accordance with 6.1 and within the limits specified by the manufacturer.

Compliance shall be tested in accordance with Annex G.

6.4 Indicators intended for use only with a sterilization temperature of 121 °C shall be tested by exposure to dry heat at (130 ± 2) °C for not less than 45 min if the indicator will not withstand heating to 140 °C.

Compliance shall be tested in accordance with Annex G.

6.5 Indicator systems intended for use with re-usable user-assembled packs shall not visibly transfer indicator reagent to the material of the test load during processing. Pre-assembled packs and indicator systems intended for use with single-use user-assembled packs shall not transfer indicator reagent to the material of the test load during processing to an extent which impairs the utility of the product.

Compliance shall be demonstrated by visual examination after testing in accordance with the requirements of 6.1 and Annex D.

6.6 The indicator shall comply with the requirements of this part of ISO 11140 for the duration of the shelf life specified by the manufacturer.

If any change in the indicator occurs during ageing, it shall be different from the change on exposure to saturated steam (as described in 6.1) and shall either inactivate the indicator system so that no further change can take place or not affect the performance of the indicator system with respect to the requirements of 6.1 and 6.2.

Compliance shall be tested in accordance with Annex H or by performance testing after accelerated ageing in accordance with Annex I.

7 Packaging and labelling

7.1 Each indicator, or indicator system, shall be marked with:

- a) the sterilization temperature(s) at which the product is designed to be used;
- b) a unique code from which the manufacturing history can be traced;
- c) the expiry date under the specified storage conditions;
- d) at least the information summarized in Figure 1.

Adjacent to each heading there shall be a clear space not less than 5 mm × 20 mm for the user to enter the required information at the time of use or, if the size of the indicator system does not permit this, each indicator or indicator system shall be supplied with a means of retaining the indicator or indicator system as a permanent record which shall be printed with the information given in Figure 1. The means of retention shall permit writing in permanent ink to be made in association with the indicator.

Cycle No. 	Site
Machine No. 	Department
Date 	Operator
Supervisor 	Result

NOTE This is an example of a suitable format. Other formats and/or text can be used.

Figure 1 — Provision for recording information to be provided on or with each indicator

7.2 When the indicator is supplied assembled, i.e. with the indicator system within the test load, the exterior of the test load shall be marked with the sterilization temperature(s) at which the product is suitable for use, the manufacturer's name, batch number and date of manufacture. In addition, either a means of uniquely identifying the individual indicator or an area on the outside of the test load on to which the operator can write the number of the machine tested and date shall be provided.

When a manufacturer provides similar products which are intended only for specific sterilization cycles, the product shall include identification sufficient to enable the user to determine, from the instructions for use, any restrictions on the use of the product. The identification shall be on the indicator or indicator system and, if not visible to the user before use, shall also be on the outside of the test load.

7.3 The transport package shall be such that the product can be removed easily. The package shall protect the product to the extent necessary to ensure that the indicator retains its performance throughout the stated shelf life when stored and transported in accordance with the manufacturer's instructions.

The manufacturer shall retain documentary evidence demonstrating compliance.

7.4 The outside of each package shall be marked with the sterilization temperature(s) at which the product is suitable for use.

7.5 The information supplied by the manufacturer (see 5.8 of ISO 11140-1:2005) shall include sufficient instructions on the use of the indicator to enable correct interpretation of the test results.

7.6 When requested by the purchaser, the manufacturer shall supply a certificate of conformity to the requirements of this part of ISO 11140 for each batch of product supplied.

8 Quality assurance

8.1 The quality system shall ensure that the performance requirements given in Clause 6 are maintained.

8.2 Suitable records shall be maintained to ensure that, if necessary, faulty batches can be recalled from use.

8.3 The manufacturing and distribution records shall be retained for a period of five years, or twice the declared shelf life of the product, whichever is greater. An example of the requirements for maintaining records is given in ISO 9001:2000 ^[6].

Annex A (normative)

Determination of indicator strength during and after steam sterilization

A.1 Apparatus

A.1.1 Steam exposure apparatus, as specified in Annex J.

A.2 Procedure

A.2.1 Expose the indicator to three successive test cycles at the stated sterilization temperature of the indicator or indicator system. The indicator shall be tested using both the standard test cycles specified in B.1 and B.2 (see 4.5) unless the indicator is intended for use with only one type of air removal stage, in which case the appropriate test cycle shall be used. The rate of pressure change of evacuation during the air removal pulse and during the drying stage shall be not less than 400 kPa min⁻¹.

The rate of pressure change shall be determined as follows (see Figure A.1):

$$p_3 = 0,125(7p_1 + p_2) \quad (\text{A.1})$$

$$p_4 = 0,5(p_1 + p_2) \quad (\text{A.2})$$

$$\frac{\Delta p}{\Delta t} = \frac{(p_3 - p_4)}{(t_4 - t_3)} \quad (\text{A.3})$$

where

p_1 is the maximum absolute pressure attained during the last air removal pulse and during the operating stage, in kilopascals;

p_2 is the minimum absolute pressure attained during the last air removal pulse (prior to the admission of steam to the operating pressure required for the chamber reference temperature to attain the sterilization temperature) and the drying stage, in kilopascals;

p_3 is the pressure calculated from Equation A.1, in kilopascals;

p_4 is the pressure calculated from Equation A.2, in kilopascals;

t_3 is the time at p_3 , in minutes;

t_4 is the time at p_4 , in minutes;

$\frac{\Delta p}{\Delta t}$ is the rate of pressure change, in kilopascals per minute.

A.2.2 Remove the pre-assembled or user-assembled indicator from the exposure apparatus and examine the indicator for visible damage, including, for example, opening or distortion of seals. Record the result.

A.2.3 If the indicator has remained intact, perform a drop test in accordance with ISO 2248 from a height of 1 m onto a firm horizontal surface. Record the result.

NOTE Concrete or terrazzo surfaces are suitable.




Figure A.1 — Determination of the rate of pressure change

A.2.4 Carry out this test with three samples for each of three separate production batches. All nine samples can be processed simultaneously.

A.2.5 Damage occurring during the drop test, which can be demonstrated as not impairing the interpretation of the indicator in normal use shall not constitute a failure. For re-usable test loads, any damage occurring during the drop test which can be demonstrated as not impairing the subsequent re-use of the test load shall not constitute a failure.

Annex B (normative)


Standard test cycles

B.1 Cycle 1: Air removal by sub-atmospheric pulsing

The standard test cycle for sub-atmospheric air removal shall consist of the following steps:

- evacuation of the chamber to 5,0 kPa;
- steam admission to 97,0 kPa;
- repetition of steps a) and b) a further three times;
- if air injection is being used, it shall take place and be completed during steam admission to the exposure time at a pressure of between 75 kPa and 105 kPa (indicated by an arrow on Figure B.1);
- steam admission to set operating pressure (see specific requirements for steam admission stage in B.4);
- exposure time;
- evacuation to 5,0 kPa;
- air admission.

The actual pressures achieved at the set points shall be determined by the tolerance permitted for the steam exposure apparatus (see Annex J).



Key

- set operating pressure (kPa)
- air injection


Figure B.1 — Standard test cycle — Sub-atmospheric air removal

B.2 Cycle 2: Air removal by trans-atmospheric pulsing

The standard test cycle for air removal by trans-atmospheric pulsing shall consist of the following steps:

- a) evacuation of the chamber to 5,0 kPa;
- b) steam admission to 150 kPa;
- c) evacuation of the chamber to 50 kPa;
- d) repetition of steps b) and c) a further three times;
- e) steam admission to set operating pressure minus 10,0 kPa;
- f) evacuation of chamber to 110 kPa to 120 kPa;
- g) repetition of steps e) and f) one further time;
- h) steam admission to set operating pressure (see specific requirements for steam admission stage in B.4);
- i) exposure time;
- j) evacuation to 5,0 kPa;
- k) air admission.

The actual pressures achieved at the set points shall be determined by the tolerance permitted for the steam exposure apparatus (see Annex J).



Key


- 1 set operating pressure
- 2 set operating pressure minus 10 kPa

Figure B.2 — Standard test cycle — Trans-atmospheric air removal

B.3 Cycle 3: Air removal by super-atmospheric pulsing

The standard test cycle for air removal by super-atmospheric pulsing shall consist of the following steps:

- a) evacuation of the chamber to 5,0 kPa;
- b) steam admission to 95 kPa;
- c) evacuation of the chamber to 5,0 kPa;
- d) repetition of steps b) and c);
- e) steam admission to set operating pressure minus 20,0 kPa;
- f) evacuation of the chamber to 105 kPa to 120 kPa;
- g) repetition of steps e) and f) a further two times;
- h) if air injection is being used, it shall take place during steam admission to the exposure time at a pressure of between 120 kPa and 130 kPa (indicated in Figure B.3);
- i) steam admission to set operating pressure (see specific requirements for steam admission stage in B.4);
- j) exposure time;
- k) evacuation to 5,0 kPa;
- l) air admission.



Key

- 1 set operating pressure (kPa)
- 2 set operating pressure minus 20 kPa
- 3 air injection

Figure B.3 — Standard test cycle — Super-atmospheric air removal

B.4 Acceptance limits during steam admission

B.4.1 The rate of pressure rise during steam admission, to set operating pressure over the range 100 kPa or lowest pressure at the bottom of the last super-atmospheric pulse to the set operating pressure of the exposure time, shall be between 100 kPa min⁻¹ and 250 kPa min⁻¹ as indicated in Figure B.4.

B.4.2 Select the operating temperature of the exposure time such that it corresponds to the temperature as stated for the indicator. Set the operating pressure such that it corresponds to a saturated steam temperature of (selected operating temperature, in degrees centigrade + 0,2 °C).

B.4.3 The integral [Integrated Come-up Exposure (ICE)] between the chamber reference temperature when the chamber reaches 100 kPa, or at the bottom of the last super-atmospheric pulse, whichever is the greater, and the set temperature during the steam admission period, bounded by the chamber reference temperature on the graph, shall not exceed $T_R \times (12T_R/6)$

where T_R is the set temperature in degrees centigrade minus 100 °C.

EXAMPLES

At a set temperature 134 °C the integral shall not exceed $34 \times \left(\frac{12 \times 34}{6} \right) = 2\,312 \text{ s}\cdot\text{K}$ or

at a set temperature 121 °C the integral shall not exceed $21 \times \left(\frac{12 \times 21}{6} \right) = 882 \text{ s}\cdot\text{K}$.

NOTE 1 These limits are intended to ensure that steam admission does not contribute to excessive exposure of the indicator to typical conditions. ICE shall be calculated using the equation:

$$\text{ICE} = \sum_{t_0}^{t_1} (T_1 - T_0) \cdot dt \quad (\text{B.1})$$

where

T_1 is the chamber reference temperature at time t , in degrees centigrade;


T_0 is 100 °C or the lowest temperature of the last positive pulse, in degrees centigrade;

dt is 1 s;

t_0 is the time at which chamber reference temperature attains T_0 , in seconds;

t_1 is the time after t_0 at which the chamber reference temperature, T_c , attains set operating temperature, T_R , e.g. 134 °C, in seconds.

NOTE 2 Each of the test cycles described within this Annex consists of three principal stages: air removal, exposure time (equivalent to the sterilization stage) and evacuation stage. The temperatures attained during the air removal stage and the duration of the air removal stage can have a significant effect on indicator performance. The cycles described are not intended to imitate any of the many cycles which are commercially available. They are intended to provide an overall range of effects within which most commercially available cycles will occur.



Key

- 1 maximum rate of pressure rise during steam admission $\frac{(p_R - p_0)}{t_{1min}}$
 - 2 rate of pressure rise
 - 3 minimum rate of pressure rise during steam admission $\frac{(p_R - p_0)}{t_{1max}}$
 - 4 Integrated Come-up Exposure: area bounded by T_0 and the curve traced by T_c over the time t_0 to t_1
- t_{1max} is the come-up time at the minimum allowed rate of pressure rise of 100 kPa min^{-1}
- t_{1min} is the come-up time at the maximum allowed rate of pressure rise of 250 kPa min^{-1}
- p_R is the pressure of saturated steam, corresponding to the set operating temperature, in kilopascals
- p_0 is the pressure of saturated steam, corresponding to the temperature T_0 , in kilopascals
- T_0 is $100 \text{ }^\circ\text{C}$ or the lowest temperature of the last positive pulse, in degrees centigrade
- T_c is the chamber reference temperature, in degrees centigrade

Figure B.4 — Steam admission

Annex C (normative)

Estimation of visual difference between colour of the substrate and of the changed or unchanged indicator system by determination of relative reflectance density

C.1 Principle

The relative reflectance density, as defined in ISO 5-1, of the changed indicator and the substrate shall be determined in accordance with the methods given below which are based on ISO 5-3 and ISO 5-4, to which reference shall also be made.

Relative reflectance density, D_{Rf} , is calculated as follows

$$D_{Rf} = -\log_{10} R_f \quad (C.1)$$

$$R_f = \Phi_c / \Phi_{ce} \quad (C.2)$$

where

Φ_c is the reflected flux from the indicator;

Φ_{ce} is the reflected flux from the substrate.

To completely define a type of density spectrally, it is necessary to specify the light source, optics and spectral response of the measuring system.

C.2 Apparatus

C.2.1 Steam exposure apparatus, as specified in Annex J.

C.2.2 Illuminant, with a relative spectral power distribution of the incident flux, as specified in CIE standard illuminant D₆₅ (see ISO/CIE 10526:1999).

NOTE This is regarded as equivalent to "Daylight – cloudy northern sky".

C.2.3 Photoelectric reflectance photometer, giving within 0,3 % an indicated reading proportional to the intensity of light reflected from the surface under test. The instrument has the following characteristics.

C.2.3.1 Optical geometry

The optical geometry shall conform to the requirements of ISO 5-4; this includes illumination of the specimen at angles between 40° and 50°, viewed along the normal (0°) with an angle of acceptance (observer angle) of 10°.

The dimensions of the measurement aperture of the instrument shall permit the measurement aperture to be entirely filled with substrate or indicator reagent.

To minimize measurement errors, the optical system should be equipped with a polarizing filter if the surface to be measured is highly reflecting, e.g. a plastic-coated surface.

C.2.3.2 Spectral response

For the visual reflectance density, the combined spectral sensitivity of the receiver and spectral characteristics of the components on the efflux section of the measuring instrument shall match the spectral luminance efficiency in photopic vision, designated $V_{(\lambda)}$. The product of $V_{(\lambda)}$ and the reflection densitometer illuminance E_A , wavelength by wavelength, defines the spectral products required of the measuring instrument in order to provide comparison of visual densities. The spectral product of the measuring instrument shall be within $\pm 20\%$ of the values given in Table C.1.

NOTE These conditions assume that there is no fluorescence in the optical elements of the instrument or the sample.

C.2.3.3 Calibration

Reflectance density is determined using a perfectly-reflecting and perfectly-diffusing material as a reference standard. Such a material does not exist, but the response that would theoretically be obtained from such a material can be compared with a suitable secondary reference standard, e.g. compressed barium sulfate or enamelled metal plaques which can then be used to calibrate the densitometer.

The measuring instrument shall be calibrated against reference samples previously calibrated by a National Reference Laboratory.

The instrument shall indicate values within $\pm 3\%$ of the calibrated values of the reference samples.

C.2.3.4 Background

While readings of the reflectance density of the substrate and the indicator are being made, the sample shall be in contact with a backing material which is spectrally non-selective and diffuse-reflecting and which has an ISO reflection density as defined in ISO 5-4:1995, Annex A, greater than 1,50.

C.3 Test method

C.3.1 Sample conditioning

Samples shall be conditioned to, and in equilibrium with, $(23 \pm 2)^\circ\text{C}$ and $(50 \pm 5)\%$ relative humidity when tested.

Standardized conditions are recommended because some materials change density with variations in temperature and relative humidity.

C.3.2 Procedure

Expose the indicator system within the indicator to a cycle of the steam exposure apparatus at the specified operating temperature for the indicator, to produce a uniform colour change in the indicator reagent in accordance with 6.1.

Determine the relative reflectance density of the indicator reagent on the substrate by using the substrate as the reference reflectance.

Carry out this measurement on three samples for each of three batches of the indicator system and on aged samples (see Annex H and Annex I) in accordance with 6.6.

C.4 Test report

The test report shall contain at least the following information:

- a) name and address of the indicator manufacturer;
- b) batch numbers of the individual batches of indicator tested;
- c) make, model and serial number of the test instrument;
- d) calibration details traceable to a national standards authority;
- e) temperature chart records of the steam exposure to which the indicators were exposed;
- f) mean and range of the relative reflectance density measurements;
- g) date of test;
- h) identification of the test operator.

STANDARDSISO.COM : Click to view the full PDF of ISO 11140-4:2007

Table C.1 — Values of spectral product required of the reflectance photometer
at the given values of wavelength and illuminance

Wavelength nm	Reflection densitometer illuminance E_A	Visual density spectral product Π_v
340	4	
350	5	
360	6	
370	8	
380	10	
390	12	
400	15	< 1 000
410	18	1 322
420	21	1 914
430	25	2 447
440	29	2 811
450	33	3 090
460	38	3 346
470	43	3 582
480	48	3 818
490	54	4 041
500	60	4 276
510	66	4 513
520	72	4 702
530	79	4 825
540	86	4 905
550	93	4 957
560	100	4 989
570	107	5 000
580	114	4 989
590	122	4 956
600	129	4 902
610	136	4 827
620	144	4 731
630	151	4 593
640	158	4 433
650	165	4 238
660	172	4 013
670	179	3 749
680	185	3 490
690	192	3 188
700	198	2 901
710	204	2 622
720	210	2 334
730	216	2 041
740	222	1 732
750	227	1 431
760	232	1 146
770	237	≤ 1 000

Annex D (normative)

Determination of uniform colour change on exposure to saturated steam

D.1 Apparatus

D.1.1 Steam exposure apparatus, as specified in Annex J and set to an operating cycle previously demonstrated to produce satisfactory air removal and rapid steam penetration in a standard test pack (see Annex K).

D.1.2 Standard test pack, as specified in Annex K.

D.1.3 Temperature sensors and thermometric recording instruments, as specified in the requirements for test instrumentation given in 4.6.

D.2 Procedure

D.2.1 Expose the indicator (or the indicator system combined with its test load for user-assembled packs) to a cycle of the steam exposure apparatus with the holding time and temperature preset to those specified in 6.1.

D.2.2 At the end of the operating cycle, remove the indicator from the steam exposure apparatus and examine for compliance with 6.1.

D.2.3 Carry out the test on three samples for each of three production batches using operating cycles with a sub-atmospheric air removal stage, and on further sets of samples with operating cycles employing a trans-atmospheric and a super-atmospheric air removal stage (see Table 1 and Annex B). The obtained samples can be used in the tests according to Annex C.

D.2.4 Before and after each series of three tests, run an operating cycle containing a standard test pack, monitored with temperature sensors, to verify the operating cycle is performing within the required limits.

D.2.5 Run the tests required on any one production batch on separate operating cycles. The steam exposure apparatus door cools down between consecutive tests, thus the highest reproducibility is obtained if the time period for the unloading-loading-start sequence is kept short and constant.

Annex E (normative)

Determination of equivalence of the alternative indicator to the Bowie and Dick test

E.1 Principle

The method provides three different systems for comparing the sensitivity of the alternative indicator to the Bowie and Dick test performed using a standard test pack (see Annex K) and thermometric measurement. The three systems produce similar but not identical results in a Bowie and Dick test pack and rely on the air pocket produced when:

- a) air is injected into the steam exposure apparatus;
- b) residual air is allowed to remain within the chamber of the steam exposure apparatus;
- c) air is allowed to leak into the chamber of the steam exposure apparatus (sub-atmospheric air removal stages only).

E.2 Apparatus

E.2.1 Steam exposure apparatus, as specified in Annex J, fitted with an air injection system as specified in Annex L, and fitted with a valved system, including a non-return valve, to provide a controlled leak of ambient air into the chamber while the latter is at pressures below ambient, and a flow meter to measure the rate of air ingress.

E.2.2 Standard test pack, as specified in Annex K.

E.2.3 Thermometric recording instrument, as specified in 4.6 to record the temperature from a minimum of six temperature sensors. The temperature sensors shall be introduced into the chamber through the temperature-sensor entry connection and fitting.

E.3 Procedure

E.3.1 Perform the tests on the standard test loads at the operating temperature(s) of the indicator.

E.3.2 Remove the wrapping from the standard test pack and place not less than five, and not more than ten, temperature sensors in the test pack, of which one shall be placed at the geometric centre of the test pack. The others shall be arranged in a pattern around the geometric centre of the test pack to detect a temperature depression occurring within a radius of 30 mm of the geometric centre. Place one temperature sensor at the defined reference point within the chamber to measure the chamber reference temperature. Reassemble the test pack as described in Annex K. As the coolest spot within the standard test pack will not be predictably at the exact geometric centre, the additional temperature sensors in the standard test pack are used to improve the reproducibility of the test results. In setting up the standard test pack, the use of a chemical indicator test sheet complying with ISO 11140-3 ^[3], placed within the pack, may be helpful in visualizing the position of the air pocket and determining the optimum position for the temperature sensors.

E.3.3 Provided at least two of the sensors in the standard test pack show a temperature depression, the lowest depression shall be used to calculate the reference fault condition.

E.3.4 Use the following conditions to produce a reference fault condition and when testing the indicator.

- Inject the volume of air into the chamber at the time specified in Annex B. The volume and the rate of injection should be determined by prior trial.
- Determine the air leak into the chamber by the method given in EN 285. The rate of air leak required should be determined by prior trial.
- Modify the air removal stage. The reduction in the pressure range and, if necessary, in the number of pulses in the air removal stage, should be determined by prior trial.

E.3.5 During the tests, vary the orientation of the product between replicate tests (within any limits given in the instructions for use of the product).

E.3.6 Carry out the test three times for each of three production batches, using operating cycles with a sub-atmospheric air removal stage, and on further sets of samples with operating cycles employing a trans-atmospheric air removal stage and a super-atmospheric air removal stage as specified in Table 1 and Annex B.

E.3.7 Before and after each series of three tests, run an operating cycle containing a standard test pack and monitor with thermocouples to verify that the operating cycle is performing within the required limits.

E.3.8 Run the three tests required on any one production batch on separate operating cycles to ensure that the results are consistent, within the variation that inevitably occurs between successive cycles on apparatus of this sort.

E.3.9 After each test, inspect the indicator system visually for compliance with 6.2.

E.3.10 The position of all temperature sensors in the test pack (see E.3.2) shall be described in the test report.

Annex F (normative)

Determination of reproducibility of fail conditions created in a standard test pack by air injection, air leak and retained air systems

F.1 General

It is not possible to determine simultaneously the effect in a standard test pack and in an alternative indicator, since the two systems are each intended to be used in a sterilizer chamber which contains no other load (i.e. is empty except for the chamber furniture).

It is therefore necessary to create a system in which a standard set of operating conditions can be shown reproducibly to yield the desired results with a standard test pack; the alternative indicator can then be tested under these conditions to determine whether the results obtained correspond with those which would have been expected from a standard test pack.

Although many of the variables of the operating cycle for the steam exposure apparatus can be controlled with great precision, the key variable, i.e. the disposition of air within the chamber and load, is the least amenable to control, the least stable, and cannot be independently measured simultaneously.

No difficulty should be experienced in producing operating cycles which show satisfactory steam penetration in a standard test pack. It is only when air is present that it becomes less predictable.

A statistical basis is required so that, with the minimum number of cycles and a suitable means to measure variability, appropriate acceptance criteria can be determined:

- for cycles intended to produce a reference fault condition;
- for cycles intended to produce satisfactory steam penetration.

F.2 Criteria for acceptability of test cycles

F.2.1 General

Use the conditions described in F.2.2 and F.2.3 as the limiting conditions within which the temperature trace from the standard test pack shall lie. For each cycle run, these criteria shall be met. In addition, by integrating the area between the chamber reference temperature and the test pack temperature for the hold period, it is possible to calculate the Reference Integrated Fault (RIF) which provides a simple means of comparing cycles [see F.2.2 j)].

F.2.2 Reference fault condition

For cycles to evaluate the ability of the indicator to detect inadequate steam penetration, the results of thermometric monitoring of the steam exposure apparatus and the standard test pack shall meet the following criteria as shown in Figure F.1.

- a) The operating cycle, including the steam admission stage, shall meet all the criteria given in Annex B.
- b) The elapsed time between the chamber attaining the set operating pressure and the chamber reference temperature attaining the set temperature shall not exceed 5 s. In a correctly functioning steam exposure apparatus, any difference should be due solely to the difference in response time of the pressure and temperature sensors.


- c) At the time the chamber reference temperature, T_c , attains the set temperature, T_r , the temperature measured in the standard test pack, T_p , shall show a temperature depression, $(T_c - T_p)$, of 2 K or greater.
- d) The temperature depression shall remain at 2 K or greater throughout the reference fault period.
- e) The test equilibration time shall be 90 s.
- f) The temperature depression, $(T_c - T_p)$, at the beginning of the reference fault period shall be no greater than 7 K.
- g) The temperature depression, $(T_c - T_p)$, at the end of the reference fault period shall be no greater than 4 K.
- h) The temperature depression, $(T_c - T_p)$, at the end of the minimum permitted equilibration time shall be no greater than 2 K.
- i) The temperature depression, $(T_c - T_p)$, at the end of the exposure time or 10 min, whichever is shorter, shall be no greater than 1 K.
- j) The Reference Integrated Fault (RIF), determined as the area bounded by the trace of the chamber reference temperature, T_c , and the temperature at the centre of the standard test pack, T_p , shall be between 120 s-K and 525 s-K for a cycle at 134 °C for 3,5 min, and 120 s-K and 1 080 s-K for a cycle at 121 °C for 15 min.
- k) The steam entry rate shall be 100 kPa/min to 250 kPa/min.

F.2.3 No fault condition

A “no fault” cycle is one in which all indicators would be expected to provide evidence of a satisfactory cycle.

For cycles to evaluate the ability of the indicator to detect adequate steam penetration, the results of thermometric monitoring of the steam exposure apparatus and the standard test pack shall meet the following criteria as shown in Figure F.1.

- a) The operating cycle, including the steam admission stage, shall meet all the criteria given in Annex B.
- b) The elapsed time between the chamber attaining the set operating pressure and the chamber reference temperature attaining the set temperature shall not exceed 5 s. In a correctly functioning steam exposure apparatus, any difference should be due solely to the difference in response time of the pressure and temperature sensors.
- c) At the time, t_1 , at which the chamber reference temperature, T_c , attains the set temperature, T_r , the temperature measured in the standard test pack, T_p , shall show a temperature depression, $(T_c - T_p)$, of 1 K or less.
- d) By the end of the reference fault period there shall be no detectable temperature difference between the centre of the test pack, T_p , and the chamber reference temperature, T_c , (within the limits of accuracy of the measuring equipment).



Key

1 reference fault period

T_r set temperature

T_s minimum value of T_p at the time t_1 ("no fault" condition) ($T_r - 1$)

T_f maximum value of T_p at the time t_1 and t_2 ("fault" condition) ($T_r - 2$)

T_m minimum value of T_p at the time t_1 ("fault" condition) ($T_r - 7$)

T_x minimum value of T_p at the time t_2 ("fault" condition) ($T_r - 4$)

T_p temperature at the centre of the standard test pack

T_c chamber reference temperature

t_1 time at which the chamber reference temperature attains set temperature, T_r

t_2 end of the reference fault period = ($t_1 + 30$) s

t_3 end of the minimum permitted test equilibration time = ($t_1 + 90$) s

t_4 end of the set exposure time

A_F area within which the plot of the test pack temperature, T_p , shall lie for a "fault" condition

A_G area within which the plot of the test pack temperature, T_p , shall lie for a "no fault" condition

A_H area within which the plot of the test pack temperature, T_p , would indicate failure to obtain a correctly defined cycle for either "fault" or "no fault" conditions

Figure F.1 — Integral of the fault condition

F.3 Statistical evaluation of reproducibility

F.3.1 Run ten consecutive cycles on each of two days.

F.3.2 From the cycle records confirm that each cycle meets the acceptance criteria and determine the value of the Reference Integrated Fault. [See F.2.2 j).]

F.3.3 Calculate the mean and estimate of variance of the Reference Integrated Fault for the two sets of runs. It is not necessary to assume, or demonstrate, that the data are normally distributed (Gaussian). Using the F statistic shows this is not critical since the test is robust.

F.3.4 Carry out the F test (variance ratio test):

$$F = \frac{\left(\text{estimate of variances}_{\text{day 1}} \right)^2}{\left(\text{estimate of variances}_{\text{day 2}} \right)^2} \quad (\text{F.1})$$

with $(n_1 - 1)$ and $(n_2 - 1)$ degrees of freedom.

Standard F tables are drawn up for 1-sided hypothesis, so for H_1 ($\text{variance}_{\text{day 1}} = \text{variance}_{\text{day 2}}$) and $\alpha = 0,05$, the upper limit is given by δ_1 / δ_2 and $(n_1 - 1)$ and $(n_2 - 1)$ degrees of freedom, and the lower limit is given by δ_2 / δ_1 and $(n_2 - 1)$ and $(n_1 - 1)$ degrees of freedom.

F.3.5 Compare the calculated value of F with the critical region obtained from the tabulated values of F . The same tables can also be used to calculate the confidence interval.

Annex G (normative)

Determination of indicator colour change on exposure to dry heat

G.1 Apparatus

G.1.1 Dry heat oven, capable of maintaining a steady temperature of $(140 \pm 2) ^\circ\text{C}$ and which maintains a relative humidity of less than 5 % throughout the test period.

G.2 Procedure

G.2.1 General

Carry out both tests three times for each of three separate production batches of the indicator system.

Several test samples may be exposed simultaneously.

G.2.2 Test 1

G.2.2.1 Pre-heat the oven to the operating temperature.

G.2.2.2 Place the indicator systems in the oven and subject them to dry heat as required in 6.3 and 6.4. Remove the samples and examine for colour change after the required exposure period.

G.2.3 Test 2

G.2.3.1 Fit the indicator system in combination with its specified test load (the indicator) with a temperature sensor to monitor the temperature of the indicator system and subject it to dry heat at $(140 \pm 2) ^\circ\text{C}$ to determine the time required for the indicator to reach $134 ^\circ\text{C}$. This time is the heat-up time.

Indicators intended for use only at a sterilization temperature of $121 ^\circ\text{C}$ may be subjected to $(130 \pm 1) ^\circ\text{C}$ and the time required to attain $121 ^\circ\text{C}$ shall be determined.

G.2.3.2 Dry new samples of the indicator system in combination with its specified test load (the indicator) to constant mass at a temperature greater than $100 ^\circ\text{C}$, or by equivalent means at a lower temperature using a suitable desiccant.

G.2.3.3 Transfer the dried indicators to the oven without undue delay and in a manner that will prevent rehydration of the indicator, and subject them to dry heat at $(140 \pm 2) ^\circ\text{C}$ for [(heat-up time) + 30] min.

Indicators intended for use only at a sterilization temperature of $121 ^\circ\text{C}$ may be subjected to dry heat at $(130 \pm 2) ^\circ\text{C}$.

Annex H

(normative)

Determination of shelf life of product

H.1 The testing of the product for determination of the shelf life shall be performed in accordance with a written protocol which shall be established before the commencement of the study.

H.2 The samples of the product shall be stored in their normal packaging at or above the maximum temperature and relative humidity recommended for storage. These conditions shall be controlled and monitored.

H.3 All performance requirements shall be met during and on completion of the storage period. This shall be verified by testing.

H.4 All results of the storage trial shall be retained for a period of at least five years from completion of the trial. After this period, a summary report should be retained for as long as the product is commercially available.