



INTERNATIONAL STANDARD ISO 11981:1999
TECHNICAL CORRIGENDUM 1

Published 2005-04-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Ophthalmic optics — Contact lenses and contact lens care products — Determination of physical compatibility of contact lens care products with contact lenses

TECHNICAL CORRIGENDUM 1

Optique ophtalmique — Lentilles de contact et produits d'entretien des lentilles de contact — Détermination de la compatibilité physique des produits d'entretien des lentilles de contact avec les lentilles de contact

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to ISO 11981:1999 was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*.

Pages 1 and 2, Clause 2

Replace the normative references such that they are undated and so that their most recent edition applies.

Delete footnotes 1) and 2).

Page 4, subclause 5.2

Add the following text after first paragraph.

“Taking into account recent reports on the incompatibility of certain silicones contained in hydrogels used systems for contact lenses. It is recommended that separate compatibility tests be performed on these types of materials with contact lens care systems.”