
Tissue paper and tissue products —
Part 9:
Determination of ball burst strength

Papier tissue et produits tissue —

Partie 9: Détermination de la résistance à l'éclatement, méthode à la balle



STANDARDSISO.COM : Click to view the full PDF of ISO 12625-9:2015



COPYRIGHT PROTECTED DOCUMENT

© ISO 2015

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

	Page
Foreword	iv
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	2
5 Apparatus	2
5.1 General	2
5.2 Clamping system	2
5.3 Penetration system	3
5.4 Force-measuring system	3
5.5 Drive mechanism	3
6 Sampling	4
7 Conditioning	4
8 Preparation of test pieces	4
8.1 General	4
8.2 Preparation of test pieces	4
9 Procedure	4
10 Calculation	5
11 Test report	5
Annex A (informative) Precision	6
Annex B (informative) Adaptor	9
Bibliography	10

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 6, *Paper, board and pulps*, Subcommittee SC 2, *Test methods and quality specifications for paper and board*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement) in collaboration with by the European Committee for Standardization (CEN) Technical Committee CEN/TC 172, *Pulp, paper and board*.

This second edition cancels and replaces the first edition, ISO 12625-9:2005, which has been technically revised.

The following changes have been made:

- [Clause 4](#) was reformulated;
- internal diameter of the two concentric rings was reduced to 50 mm;
- description of the procedure in [Clause 9](#) was simplified;
- precision data in [Annex A](#) was added;
- editorial updating.

ISO 12625 consists of the following parts, under the general title *Tissue paper and tissue product*:

- *Part 1: General guidance on terms;*
- *Part 3: Determination of thickness, bulking thickness apparent bulk density and bulk;*
- *Part 4: Determination of tensile strength, stretch at maximum force and tensile energy absorption;*
- *Part 5: Determination of wet tensile strength;*
- *Part 6: Determination of grammage;*
- *Part 7: Determination of optical properties — Measurement of brightness and colour with D65/10° (outdoor daylight);*

- *Part 8: Water-absorption time and water-absorption capacity, basket-immersion test method;*
- *Part 9: Determination of ball burst strength;*
- *Part 11: Determination of wet ball burst strength;*
- *Part 12: Determination of tensile strength of perforated lines — Calculation of perforation efficiency;*
- *Part 15: Determination of optical properties — Measurement of brightness and colour with C/2° (indoor daylight);*
- *Part 16: Determination of optical properties — Opacity (paper backing) — Diffuse reflectance method*

STANDARDSISO.COM : Click to view the full PDF of ISO 12625-9:2015

Introduction

This part of ISO 12625 is applicable to tissue papers and tissue products. In principle, application to other paper types is possible, but not covered by this part of ISO 12625.

It is expressly stated that the detection of impurities and contraries in tissue and tissue products be applied according to ISO 15755.

For the determination of moisture content in tissue paper and tissue products, ISO 287 and ISO 638 are applied.

STANDARDSISO.COM : Click to view the full PDF of ISO 12625-9:2015

Tissue paper and tissue products —

Part 9: Determination of ball burst strength

1 Scope

This part of ISO 12625 specifies a test method for the determination of the resistance to mechanical penetration (ball burst strength procedure) of tissue paper and tissue products.

Currently, two types of clamping devices are available on the market with two different diameters, one is with 50 mm and one is with 89 mm (see [Annex B](#)). This part of ISO 12625 applies for a 50 mm clamping device to be able to measure all sample sizes of tissue paper and tissue products and to be consistent with ISO 12625-11.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 186, *Paper and board — Sampling to determine average quality*

ISO 187, *Paper, board and pulps — Standard atmosphere for conditioning and testing and procedure for monitoring the atmosphere and conditioning of samples*

ISO 7500-1, *Metallic materials — Verification of static uniaxial testing machines — Part 1: Tension/compression testing machines — Verification and calibration of the force-measuring system*

ISO 12625-1, *Tissue paper and tissue products — Part 1: General guidance on terms*

ISO 12625-6, *Tissue paper and tissue products — Part 6: Determination of grammage*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 12625-1 and the following apply.

3.1 grammage *g*

mass of a unit area of tissue paper or tissue product as determined by the procedure specified in ISO 12625-6

Note 1 to entry: The grammage is expressed in gram per square metre (g/m^2).

3.2 bursting force *F*

maximum force that a test piece of tissue paper or tissue product can withstand under the test conditions, applied at right angle to its surface

Note 1 to entry: The bursting force is expressed in newton (N).

3.3 burst index

X

bursting force of tissue paper or tissue product divided by the grammage

Note 1 to entry: The burst index is expressed in newton square metre per gram ($\text{N}\cdot\text{m}^2/\text{g}$).

4 Principle

The test consists of measuring the resistance force of a highly polished stainless steel burst ball, moving at a constant speed, penetrating perpendicularly a tissue test piece that is clamped between the two concentric rings.


5 Apparatus

5.1 General

The apparatus shall be placed on a horizontal surface, free from externally induced vibrations.

5.2 Clamping system

The clamping system is designed to clamp the test piece firmly and uniformly between two concentric annular plane, parallel surfaces with same internal diameter. The clamping rings can be activated mechanically or pneumatically. The principle of the clamping system, using a pneumatic fixture is illustrated in [Figure 1](#).



Key

- | | | | |
|---|--|---|-------------------------------|
| 1 | load cell | 5 | movable ring |
| 2 | travel | 6 | pneumatic cylinder |
| 3 | stationary ring | 7 | probe |
| 4 | suitable band made of rubber (e.g. polychloroprene rubber) | a | polished stainless steel ball |

Figure 1 — Principle of the clamping system, pneumatic fixture


The clamping pressure shall be sufficient to prevent slippage during the test while avoiding damaging the test pieces (e.g. including a weakness point near or under the clamping ring that could influence the results).

The clamping surfaces of the clamping rings shall be covered with a band made of a suitable commercial grade rubber material (e. g. polychloropren rubber), typically 1,0 mm to 2,0 mm thick having an IRHD hardness (International Rubber Hardness Degree) of 70 IRHD to 85 IRHD.

The inner edges of the annular rubber ring shall be coincident with the inner diameters of the clamping rings and the band shall be at least 12,5 mm wide.

The internal diameter of the two concentric rings shall be $(50,0 \pm 0,2)$ mm (see [Annex B](#)).

For the dimensions of the clamping system, see [Figure 2](#).



Key

- 1 test area
- 2 clamp ring covered with suitable material (e.g. polychloroprene rubber)
- a suitable band made of rubber (e.g. polychloroprene rubber)

Figure 2 — Dimensions of the clamping system

5.3 Penetration system

The penetration (burst) system shall consist of a spherical ball permanently attached to the end of a hard solid rod designed to transmit force applied to the ball from the test piece. The penetration ball shall be made of highly polished stainless steel and shall have the following dimensions:

- diameter: $(16 \pm 0,2)$ mm;
- sphericity: better than $1 \mu\text{m}$.

The penetration ball shall be centred with respect to the annular clamps. The diameter of the rod shall be smaller than the diameter of the ball.

5.4 Force-measuring system

The force-measuring system shall measure loads with an accuracy of $\pm 1 \%$ of the reading or $\pm 0,1$ N, whichever is greater and shall be calibrated and verified to conform to the requirements according to ISO 7500-1. The load cell system shall have a measuring range of 0,1 N to 50 N. The readout system shall have a display that shows the maximum force at burst.

5.5 Drive mechanism

The drive mechanism shall consist of a static uniaxial testing machine with a suitable mechanism for moving either the clamping system or the penetration system so that they approach each other at a constant rate in a direction perpendicular to the clamping system plane. The penetration test speed shall be (125 ± 5) mm/min.

6 Sampling

If the tests are being made to evaluate a lot, the sample shall be selected in accordance with ISO 186.

If the tests are being made on another type of sample make sure that the specimens taken are representative of the sample. Each test piece shall be free from perforations and faults not normally inherent in the tissue.

7 Conditioning

Condition the samples according to ISO 187 and keep them in the standard atmosphere throughout the test.

8 Preparation of test pieces

8.1 General

Test pieces shall be free of areas containing creases, dirt, or visible damage.

8.2 Preparation of test pieces

Test pieces shall be larger than the clamping area of the testing machine.

Prepare at least 10 test pieces of the sample conditioned as described in [Clause 7](#).

The exact area of the portion of the specimen tested is defined by the inner diameter of the clamping system.

8.2.1 For converted tissue products, testing shall be done on the product as received, regardless of the number of plies which are supplied as a product unit. Generally, a single finished product sheet is suitable for use as a test piece.

8.2.2 Tissue that has not been converted into a finished product shall be tested as a single ply unless otherwise agreed between the parties concerned.

Toilet tissue is frequently produced in rolls of perforated sheets. In testing such toilet tissue it is convenient to remove a specimen consisting of three connected sheets. The outer two sheets should be used for moving the test piece into the clamping system, while the middle sheet is subjected to the test.

9 Procedure

9.1 Turn on the power of the testing machine at least 30 min prior to testing.

9.2 Clamp the conditioned test piece with its top side facing towards the burst ball without applying any tension to it. If the top side cannot be identified, ensure that the sample is clamped with the same side of all test pieces facing toward the burst ball.

9.3 Set the rate at which the penetration system and test piece approach each other to (125 ± 5) mm/min.

9.4 Activate the testing machine.

9.5 After rupture of the test piece by the penetration burst ball, record the maximum force as measured during the test to three significant figures. This value is the resistance to penetration (i.e. bursting force, F).

9.6 Perform the above sequence with the remaining pieces until 10 valid results are obtained.

10 Calculation

Determine the average value, $\bar{F}$, and the standard deviation of the resistance to penetration (bursting force) for the 10 test pieces tested.

Determine the burst index, X , expressed in newton square metres per gram, using Formula (1):

$$X = \frac{\bar{F}}{g} \quad (1)$$

where

$\bar{F}$ is the mean bursting force of the specimens, expressed in newton;

g is the grammage of the tissue, expressed in gram per square metre.

11 Test report

The test report shall include the following information:

- a) a reference to this part of ISO 12625;
- b) the date and place of testing;
- c) the conditioning atmosphere used;
- d) a complete identification of the sample tested and the sampling method used;
- e) the results, expressed as average bursting force result ($\bar{F}$), reported in newton rounded to three significant figures;
- f) the standard deviation of the 10 individual burst force results, reported to two significant figures, along with the maximum and minimum F values measured;
- g) if required, the burst index X to two significant figures, expressed in newton square metres per gram;
- h) the diameter of clamping ring;
- i) any departure from this part of ISO 12625 and other circumstances that might have affected the test results.

Annex A (informative)

Precision

A.1 General

In December 2009, an international interlaboratory test was performed on six converted tissue products by 10 different laboratories according to this part of ISO 12625 with a 50 mm clamp. In addition, an 89 mm clamping diameter was included in this interlaboratory test to compare the two settings.

The repeatability and reproducibility limits reported are estimates of the maximum difference which should be expected in 19 of 20 instances, when comparing two test results for material similar to those described under similar test conditions. These estimates might not be valid for different materials or different test conditions.

NOTE Repeatability and reproducibility limits are calculated by multiplying the repeatability and reproducibility standard deviations by 2,77, where $R = 1,96\sqrt{2} \cdot s$.

A.2 Ball burst 50 mm diameter clamp — Repeatability and reproducibility data

Table A.1 — Precision of repeatability data

Sample	Mean bursting force $\bar{F}$ N	Number of laboratories p	Repeatability standard deviation s_r N	Coefficient of variation $C_{V,r}$ %	Repeatability limit r N
household towel (type 1) 2 plies	9,36	10	1,02	10,92	2,83
bathroom tissue 2 plies	3,72	10	0,50	13,34	1,38
kitchen towel	5,29	10	0,69	13,12	1,92
bathroom tissue 4 plies	6,25	10	0,94	14,98	2,59
household towel (type 2) 2 plies	3,84	10	0,63	16,38	1,74
handkerchief 4 plies	6,44	10	0,77	11,99	2,14

Table A.2 — Precision of reproducibility data

Sample	Mean bursting force $\bar{F}$ N	Number of laboratories p	Reproducibility standard deviation s_R N	Coefficient of variation $C_{V,R}$ %	Reproducibility limit R N
household towel (type 1) 2 plies	9,36	10	1,28	13,67	3,55
bathroom tissue 2 plies	3,72	10	0,51	13,71	1,43
kitchen towel	5,29	10	0,68	12,86	1,88
bathroom tissue 4 plies	6,25	10	0,96	15,37	2,66
household towel (type 2) 2 plies	3,84	10	0,66	17,21	1,83
handkerchief 4 plies	6,44	10	0,87	13,50	2,40

A.3 Ball burst 89 mm diameter clamp – Repeatability and reproducibility data

50 mm and 89 mm diameter clamping settings have been compared during this interlaboratory test.

In this study, no significant differences between the two diameters in terms of average values as well as repeatability and reproducibility data were detected.

Table A.3 — Precision of repeatability data

Sample	Mean bursting force $\bar{F}$ N	Number of laboratories p	Repeatability standard deviation s_r N	Coefficient of variation $C_{V,r}$ %	Repeatability limit r N
household towel (type 1) 2 plies	9,14	10	1,06	11,61	2,94
bathroom tissue 2 plies	3,78	10	0,39	10,31	1,08
kitchen towel	5,77	10	0,58	9,97	1,59
bathroom tissue 4 plies	6,00	10	0,80	13,29	2,21
household towel (type 2) 2 plies	3,8	10	0,70	18,31	1,93
handkerchief 4 plies	5,99	10	0,66	11,09	1,84

Table A.4 — Precision of reproducibility data

Sample	Mean bursting force $\bar{F}$ N	Number of laboratories p	Reproducibility standard deviation s_R N	Coefficient of variation $C_{V,R}$ %	Reproducibility limit R N
house-hold towel (type 1) 2 plies	9,14	10	1,09	11,92	3,03
bathroom tissue 2 plies	3,78	10	1,19	11,38	1,19
kitchen towel	5,77	10	0,65	11,26	1,81
bathroom tissue 4 plies	6,00	10	0,83	13,83	2,29
household towel (type 2) 2 plies	3,81	10	0,72	18,89	1,99
handkerchief 4 plies	5,99	10	0,71	11,85	1,98