
**Medical devices — Symbols to be used with
medical device labels, labelling and
information to be supplied**

*Dispositifs médicaux — Symboles à utiliser avec les étiquettes, l'étiquetage
et les informations à fournir relatifs aux dispositifs médicaux*



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 15223:2000

© ISO 2000

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 734 10 79
E-mail copyright@iso.ch
Web www.iso.ch

Printed in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

International Standard ISO 15223 was prepared by Technical Committee ISO/TC 210, *Quality management and corresponding general aspects for medical devices*.

This first edition cancels and replaces ISO/TR 15223:1998.

Annex A of this International Standard is for information only.

Introduction

This International Standard considers certain items of information that may be considered by regulatory authorities to be essential for the safe and proper use of medical devices. As such, the items are required by laws and regulations of certain political jurisdictions to be presented with the device. This information may be required on the device itself, as part of the label of the device on its packaging, or provided with the device in an information document.

There is a considerable degree of international harmonization of the information to be provided. However, there is no harmonization with regard to the language to be used when presenting this information. This presents potential problems to manufacturers, users and regulatory authorities.

Device manufacturers, desiring to minimize the indirect costs not associated with healthcare purposes, seek to minimize costs of labelling by reducing or rationalizing labelling variants. In the European Union alone, there are thirteen languages that may be required. This presents a major problem of design and logistics. In addition, technical translation can present difficulties in transferring the precise meaning from one language to another.

Users may be presented with devices labelled in a number of different languages. This may cause confusion and delay in locating the appropriate language. It may also create confusion as to precise meanings for multilingual users.

Regulatory authorities may be presented with labelling not in their national language and have difficulty in ascertaining the safety and fitness for use of a device required in emergencies or other exceptional circumstances.

This International Standard proposes solutions to these problems through the use of internationally recognized symbols, with precisely defined meanings that transcend language.

Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied

1 Scope

This International Standard identifies symbols conventionally used to convey information essential for proper use to the user and others for safe and effective use of medical devices. This International Standard is primarily intended to be used by:

- manufacturers of medical devices who market their products in a number of countries having different language requirements for medical device labelling;
- users of medical devices who draw their supplies from a number of sources and may have varied language capabilities;
- those responsible for postmarket surveillance;
- health care regulatory authorities, testing organizations, certification bodies and other organizations responsible for implementing regulations affecting medical devices and having responsibility for postmarket surveillance.

This International Standard may also be of assistance to:

- manufacturers having to cope with space limitations on small labels;
- distributors of medical devices or other representatives of manufacturers;
- health care authorities responsible for training as well as those being trained.

NOTE This International Standard deals with a small number of symbols that may be used when appropriate on the device itself, its package or in the accompanying documentation. Many other standards, such as IEC 60601-1, specify additional symbols that are applicable to particular kinds or groups of devices, or to particular situations.

2 Terms and definitions

For the purposes of this International Standard, the following terms and definitions apply.

NOTE This International Standard does not introduce new concepts. The following terms and definitions are provided for guidance. In particular circumstances, the legal definitions expressed by relevant statutes should be applied.

2.1

information essential for proper use

information that is essential for the safe use of the device for the patient, user or others

NOTE This information could, for example, include the degree of microbial cleanliness, up to and including sterility, when this is necessary with regard to the essential purpose. It could also include information that would facilitate traceability in the interest of postmarket surveillance by manufacturers and postmarket vigilance by regulatory authorities. It may include storage and handling instructions.

2.2

transition period

period during which a symbol and its referent appear in association in order to familiarize distributors, users and others with the symbol

3 Symbols

When appropriate, certain information essential for proper use shall be indicated on the medical device, on its package, or in the accompanying documents by using the corresponding symbols given in Table 1.

Table 1 — Symbols to convey information essential for proper use



No.	Symbol	Referent
3.1		Biological risk
3.2		Do not re-use
3.3		Consult operating instructions ^a
3.4		Caution, consult accompanying documents ^{b, c}
3.5		Fragile, handle with care
3.6		Keep away from sunlight

Table 1 (continued)







No.	Symbol	Referent
3.7		Protect from heat and radioactive sources
3.8		Keep dry
3.9		Lower limit of temperature
3.10		Upper limit of temperature
3.11		Temperature limitation
3.12		Use by ^d
3.13		Date of manufacture ^e

Table 1 (continued)

No.	Symbol	Referent
3.14	LOT	Batch code
3.15	REF	Catalogue number
3.16	SN	Serial number
3.17	CONTROL	Control ^f
3.18	CONTROL -	Negative control ^g
3.19	CONTROL +	Positive control ^h
3.20	STERILE	Sterile

Table 1 (continued)

No.	Symbol	Referent
3.21		Sterilized using aseptic processing techniques
3.22		Sterilized using ethylene oxide
3.23		Sterilized using irradiation
3.24		Sterilized using steam or dry heat
^a This symbol advises the reader to consult the operating instructions for information needed for the proper use of the device. See also symbol 3.4. ^b This symbol advises the reader to consult the accompanying documents for important safety-related information such as warnings and precautions that cannot, for a variety of reasons, be presented on the device itself. See also symbol 3.3. ^c The referent given is compiled from all of the sources where this symbol appears in conjunction with medical devices. It is recommended that this referent be used during the transition period (see A.2). ^d The symbol is accompanied by a date to indicate that the device should not be used after the end of the year, month, or day shown. The date could be a year, year and month, or year, month and day, as appropriate. See ISO 8601 for date formats. ^e This symbol is accompanied by the date that the device was manufactured. The date could be a year, year and month, or year, month and day, as appropriate. See ISO 8601 for date formats. ^f This symbol would appear on the labelling of material that is used as part of the quality control procedure for an <i>in vitro</i> diagnostic device. ^g This is a variant of symbol 3.17 used to indicate a negative control. ^h This is a variant of symbol 3.17 used to indicate a positive control.		

Annex A (informative)

Guidance on the creation and use of symbols to convey information essential for proper use

A.1 Origins of symbols

Within the ISO framework of standards and technical reports, all symbols are to be standardized through Technical Committee ISO/TC 145 and included in ISO 7000. ISO/TC 145 works closely with IEC/TC 3, which has the responsibility for standardization of graphical symbols relating to electrotechnical standardization. Symbols standardized by IEC/TC 3 are published in IEC 60417-1. This mechanism allows for coordination across product and service sectors for a common set of standardized symbols.

New symbols may be proposed to ISO/TC 145 from any ISO technical committee and may subsequently be published in a standard by other technical committees.

Some of the symbols included in this International Standard originated in the medical device sector, while others were already in ISO 7000. ISO/TC 210 has proposed those symbols unique for medical devices and has chosen those from ISO 7000 that may be particularly useful to convey information essential for proper use of medical devices.

One of the origins of symbols for medical devices is the European Standard EN 980. Each of the symbols from EN 980 has been included in ISO 7000 and this International Standard. Other symbols have been proposed directly from the medical device sector through ISO/TC 210.

A.2 Transition period

It is recommended that the symbols proposed as acceptable for wider use in this International Standard should appear together with the relevant meaning in a language understandable to the end user. This recommendation may be relaxed for a given market area under the following conditions, whichever applies soonest. The manufacturer may demonstrate that the symbol and its meaning have appeared as recommended for a continuous period as required by the relevant regulatory authority in the market concerned, or may satisfactorily demonstrate that 75 % of typical end-users recognize the symbol and can give the meaning without prompting.

“Appear together” in the context of this transition period should be construed as meaning at the same time in association with the same device. This is to allow manufacturers to use symbols on small packaging, while including the wording on other information that is provided with the device.

An example of an appropriate transition period for a consumer market is provided in EN 71-6, which was developed by CEN/TC 52.

A.3 Proposals for additional symbols

Members of the medical device sector are encouraged to propose additional symbols that may have widespread utility in transcending language. Proposals should be submitted to ISO/TC 210 through the national member bodies of ISO.

In making a proposal, the sponsor should provide a formally designed symbol following the principals for the creation of graphical symbols contained in ISO 3461-1, as well as a documented definition. The presentation of the symbol should take into account the accepted style of approved and existing symbols. Where such symbols, in the opinion of ISO/TC 210, properly fit the scope of this International Standard, and can be seen as having widespread utility in transcending language, they will be proposed to ISO/TC 145 for inclusion in ISO 7000. Once accepted in this manner, they will be included in the subsequent revision of this International Standard.

STANDARDSISO.COM : Click to view the full PDF of ISO 15223:2000