
**Cardiovascular implants and artificial
organs — Requirements for single-use
tubing packs for cardiopulmonary bypass
and extracorporeal membrane
oxygenation (ECMO)**

*Implants cardiovasculaires et organes artificiels — Exigences pour les
paquets de tubes à usage unique pour pontage cardiopulmonaire et
oxygénation des membranes extracorporelles*



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 15676:2005

© ISO 2005

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 15676 was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, Subcommittee SC 2, *Cardiovascular implants and extracorporeal systems*.

Introduction

The intent of this International Standard is to ensure that medical grade tubing in single-use tubing packs for the transfer of blood and fluid during the period of cardiopulmonary bypass (CPB) and extracorporeal membrane oxygenation (ECMO) is adequately tested for both safety and function. The user commonly provides the specifications for the tubing pack. Furthermore, the purpose of this International Standard is to ensure that the tubing pack characteristics be appropriately disclosed in the labelling and manufacturing information package. Tubing performance characteristics are specifically addressed within the context of this International Standard as a component part of a single-use tubing pack.

This International Standard therefore contains recommended procedures to evaluate such medical grade tubing intended for use during CPB procedures and ECMO. Test procedures to determine the material characteristics, the useful life of the tubing when used in a roller pump, and cleanliness are described. The limits for these characteristics are not specified.

This International Standard also includes minimum reporting requirements. Ready identification of the performance characteristics should assist the user in the selection of such medical grade tubing for the procedure appropriate to the patient and procedure. This information may be useful in a clinic's quality control process that aims to improve the safety of CPB and ECMO procedures.

This International Standard makes reference to other International Standards, which references methods for the determination of characteristics common to medical devices.

Requirements for animal and clinical studies are not included in this International Standard. Such studies, however, may be part of a manufacturer's quality system.

This International Standard contains only those requirements that are specific to such medical grade tubing for use during CPB and ECMO. Non-specific requirements are covered by reference to other International Standards listed in the Normative References section.

Cardiovascular implants and artificial organs — Requirements for single-use tubing packs for cardiopulmonary bypass and extracorporeal membrane oxygenation (ECMO)

1 Scope

This International Standard specifies requirements for single-use tubing packs for cardiopulmonary bypass and extracorporeal membrane oxygenation (ECMO). This International Standard is applicable to all medical tubing intended for cardiopulmonary bypass (CPB) and/or extracorporeal membrane oxygenation (ECMO), but specific requirements and tests are included for tubing intended for use with peristaltic pumps during (short-term, i.e. < 6 h duration) CPB surgery, or (long-term, i.e. > 24 h) ECMO procedures. The sterility and non-pyrogenicity provisions of this International Standard are applicable to tubing packs labelled as “sterile”.

This International Standard is applicable only to the tubing aspects for multifunctional systems that may have integral components such as blood gas exchangers (oxygenators), reservoirs, blood filters, defoamers, blood pumps, etc.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 34-1, *Rubber, vulcanized or thermoplastic — Determination of tear strength — Part 1: Trouser, angle and crescent test pieces*

ISO 527-1, *Plastics — Determination of tensile properties — Part 1: General principles*

ISO 9352, *Plastics — Determination of resistance to wear by abrasive wheels*

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing*

ISO 10993-7, *Biological evaluation of medical devices — Part 7: Ethylene oxide sterilization residuals*

ISO 10993-11, *Biological evaluation of medical devices — Part 11: Tests for systemic toxicity*

ISO 11134, *Sterilization of health care products — Requirements for validation and routine control — Industrial moist heat sterilization*

ISO 11135, *Sterilization of health care products — Ethylene oxide — Requirements for development, validation and routine control of a sterilization process for medical devices*

ISO 11137-1, *Sterilization of health care products — Radiation — Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices*

ISO 11137-2, *Sterilization of health care products — Radiation — Part 2: Establishing the sterilization dose*

ISO 11607-1, *Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems*

ISO 11607-2, *Packaging for terminally sterilized medical devices — Part 2: Requirements for forming, sealing and assembly processes*

ISO 14937, *Sterilization of health care products — General requirements for characterization of a sterilizing agent and the development, validation and routine control of a sterilization process for medical devices*

ASTM D792-00, *Standard test methods for density and specific gravity (relatively density) of plastics by displacement*

ASTM D1044-99, *Standard test method for resistance of transparent plastics to surface abrasion*

ASTM D2240-04, *Standard test method for rubber property — Durometer hardness*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

durometer hardness

measure of hardness of elastic materials by Shore A range

3.2

elongation

increase in linear dimension

3.3

tensile strength

force per unit of original cross section on elongation to rupture

3.4

tear strength

measure of stress needed to continue rupturing a sheet of rubber or plastic, usually after an initial cut

3.5

tubing pack

consists of tubing sections joined by extracorporeal connectors and/or connected to extracorporeal devices intended for CPB or ECMO applications

3.6

specific gravity

ratio of the mass of a body to the mass of an equal volume of water at 4 °C

3.7

spallation

phenomenon whereby particles dislodge from a surface under cyclical stress

3.8

brittle point

temperature at which 50 % of test samples exhibit cracking or breakage after linear impact at a specified speed

4 Requirements

4.1 Biological characteristics

4.1.1 Sterility and non-pyrogenicity

The blood pathway shall be sterile and non-pyrogenic.

Compliance shall be verified in accordance with 5.2.1.

4.1.2 Biocompatibility

All parts of the tubing pack that may come in direct contact with the patient's blood pathway shall be biocompatible with respect to their intended use.

Compliance shall be verified in accordance with 5.2.2.

4.2 Physical characteristics

4.2.1 General

When tested in accordance with 5.3.1, the blood pathway shall not leak.

4.2.2 Dimensions

The dimensions of the tubing (e.g. inner diameter wall thickness, segment lengths) shall conform to the specifications of the user.

4.2.3 Material properties

The tubing shall be tested by the manufacturer or extruder to determine that the material properties listed in this subclause conform to the manufacturer's specifications as reported in 6.4 b). These tests shall be conducted using standard test methods as provided in 5.3.3. Upon request, the manufacturer should make them available in a technical data sheet. The material properties include:

- a) durometer hardness;
- b) ultimate elongation;
- c) tensile strength;
- d) brittle point;
- e) specific gravity;
- f) tear strength.

4.3 Performance characteristics

4.3.1 Priming volume

The priming volume shall be measured or calculated and reported in 6.2 e). Results shall indicate the priming volume over the entire range of tubing size provided by the manufacturer. Testing shall be performed according to the manufacturer's protocol.

Some of these tests may be combined and performed at the same time.

4.3.2 Life to failure testing

The labelled anticipated lifetime of the roller pump boot tubing should be a figure not exceeding the lifetime of tubing as determined using the test specified in 5.4.1. The tubing shall be tested under the operating variables specified by the manufacturer in 6.2 c) for each available size and wall thicknesses of tubing. The results of these tests shall be reported as mean and standard deviation in 6.3 d).

4.3.3 Spallation

When tubing intended for use in a peristaltic pump is tested in accordance with 5.4.2, the spalled particles shall not exceed the level specified by the manufacturer for an 8 h period in 6.4 c). A measurement shall be taken at 1 h, and again at 6 h, to model early and late onset of spallation.

5 Tests and measurements

5.1 General

5.1.1 Tests and measurements shall be performed with the device under test prepared according to the manufacturer's instructions for intended clinical use.

5.1.2 Operating variables shall be those specified by the manufacturer for intended clinical use unless otherwise specified.

5.1.3 According to the intended use of the tubing, the temperature of test liquids shall be 4 °C, 30 °C and 39 °C to reflect typical and extreme use conditions.

5.1.4 If the relationship between variables is non-linear, sufficient determinations shall be made to permit valid interpolation between data points.

5.1.5 The test or measurement procedures are to be regarded as reference procedures. Other procedures can be accepted provided that the alternative procedure has been shown to be of comparable precision and reproducibility.

5.2 Biological characteristics

5.2.1 Sterility and non-pyrogenicity

Sterility and non-pyrogenicity shall be determined in accordance with the requirements of ISO 11134, ISO 11135, ISO 11137-1, ISO 11137-2 and ISO 14937, as applicable.

5.2.2 Biocompatibility

Biocompatibility shall be determined in accordance with the requirements of ISO 10993-1 and ISO 10993-11. If the product is sterilized with ethylene oxide, biocompatibility shall also be tested in accordance with the requirements of ISO 10993-7.

5.3 Physical characteristics

5.3.1 Determination of blood pathway integrity

5.3.1.1 The test shall be performed with air or water at the appropriate pressures. The test shall be performed to assure freedom from leaking.

5.3.1.2 Subject the tubing to a positive pressure of 1,5 times the manufacturer's rated pressure or, if none is given, to a pressure of 60 kPa gauge and maintain this pressure for 6 h or for the intended time of use specified by the manufacturer. Using air pressure decay or visual inspection, check for leakage.

5.3.2 Connections

The connections shall withstand a pull force of 15 N for 15 s without separating. Testing shall be performed as specified in the manufacturer's protocol.

5.3.3 Tubing material property testing

Tubing material property testing shall be determined in accordance with the requirements of ISO 34-1, ISO 527-1, ISO 9352, ASTM D792-00, ASTM D1044-99, and ASTM D2240-04, as applicable or consistent with the requirements of the end user.

5.4 Performance characteristics

5.4.1 Determination of tubing life

5.4.1.1 The test liquid shall be glycerine-isotonic salt solution to simulate blood viscosity. The testing range for viscosities shall be $1,0 \times 10^{-3}$ Pa·s (1,0 cP), $2,0 \times 10^{-3}$ Pa·s (2,0 cP), and $3,0 \times 10^{-3}$ Pa·s (3,0 cP).

5.4.1.2 The manufacturer shall conduct the test with a conventional dual-roller pump, reservoir, tubing, measurement and control equipment specified by the manufacturer. Tubing of each internal diameter and wall thickness shall be tested. The operating variables of pump speed, back pressure, liquid temperature, and method of setting pump occlusion shall be described, monitored and kept constant over the course of the test.

5.4.1.3 A failure is a leak in the tubing wall greater than 1 ml/min of liquid.

5.4.2 Spallation in tubing used in roller pumps

5.4.2.1 In order to obtain consistency across various testing sites, the minimum volume of water in the circuit shall be provided and actual volume contained at test onset shall be reported. The flow rate(s) shall be reported, so that the volume/hour of fluid contacting the tubing wall can be estimated, as in accepted methods for quantifying wear debris generation.

5.4.2.2 The test liquid shall be glycerine-isotonic salt solution to simulate blood viscosity and the test shall be conducted at 20 °C for CPB and at 39 °C for ECMO pre-filtered through a 5 µm filter.

5.4.2.3 The manufacturer shall test tubing of each internal diameter and wall thickness with the test equipment described in 5.4.1.2.

5.4.2.4 The circuit shall be run for 1 h intervals with the longest test lasting 8 h.

5.4.2.4.1 For CPB, the circuit shall be run for 6 h. A measurement shall be taken at 1 h, 2 h, 4 h, and 6 h, to reflect spallation.

5.4.2.4.2 For ECMO, measurements should be taken every 24 h for the length of time specified by the manufacturer.

5.4.2.5 A litre of pump liquid shall be siphoned off after a 1 h interval through a 10 µm filter. The filter shall be dried and weighed.

5.4.2.6 The mass of recovered spall particles shall be reported in milligrams recovered per litre of starting liquid volume, for each time point.

6 Information supplied by the manufacturer

6.1 Information to be given on the tubing pack

6.1.1 Information to be given on the unit container

The following shall be given on the unit container:

- a) manufacturer's name and address;
- b) description of contents;
- c) model designation;
- d) statement on sterility and non-pyrogenicity;
- e) batch, lot or serial number designation;
- f) statement "read instructions before use";
- g) special handling or storage conditions;
- h) statement on single-use;
- i) expiry date.

6.1.2 Information to be given on the shipping container

The following shall be provided on the shipping container:

- a) the manufacturer's name and address;
- b) description of contents, including number of units;
- c) model designation;
- d) statement on sterility and non-pyrogenicity;
- e) special handling, storage or unpacking instructions;
- f) batch, lot or serial number.

6.2 Information to be given on the accompanying documents

Each shipping container shall contain an "Instructions for use" leaflet with the following information:

- a) manufacturer's address and telephone number (fax number and internet site address optional);
- b) model designation;
- c) operating instructions;
- d) placement, type and securing of tubing connections;
- e) volume per length for all sizes of tubing;
- f) pressure limitations for blood pathways;
- g) statement that the following are available upon request:
 - 1) sterilization method;
 - 2) list of the materials of the blood pathway.

6.3 Information to be given in the accompanying documents in a prominent form

The following information shall be provided in a prominent form in the accompanying documents:

- a) inner diameter and wall thickness;
- b) pressure limitations;
- c) flow rate limitations;
- d) expected life under manufacturer's operating conditions, reported as mean and standard deviation;
- e) other device limitations, e.g. material incompatibility with known volatile anaesthetic agents, solvents, or disinfectants.

6.4 Information to be provided by manufacturer upon request

The following information shall be provided by the manufacturer upon request:

- a) results of the testing conducted under 5.3.3;
- b) physical characteristics (cited in 4.2.3);
- c) spallation data.

7 Packaging

Packaging shall comply with the appropriate requirements of ISO 11607-1 and ISO 11607-2.