
**Plastics — Determination of average
molecular weight and molecular
weight distribution of polymers using
size-exclusion chromatography —**

**Part 1:
General principles**

*Plastiques — Détermination de la masse moléculaire moyenne
et de la distribution des masses moléculaires de polymères par
chromatographie d'exclusion stérique —*

Partie 1: Principes généraux



STANDARDSISO.COM : Click to view the full PDF of ISO 16014-1:2019



COPYRIGHT PROTECTED DOCUMENT

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	2
5 Reagents	2
5.1 Eluent	2
5.2 Reagent for column evaluation	2
5.3 Molecular weight standards	2
5.4 Reagent for flow rate marker (internal standard)	3
5.5 Additives	3
6 Apparatus	3
6.1 General	3
6.2 Eluent reservoir	3
6.3 Pumping system	3
6.4 Injector	4
6.5 Columns	5
6.5.1 General	5
6.5.2 Determination of theoretical plate number	5
6.5.3 Determination of resolution factor	6
6.5.4 Determination of asymmetry factor	7
6.6 Detector	7
6.7 Tubing	8
6.8 Temperature control	8
6.9 Recorder and plotter	8
6.10 Data-processing system	8
6.11 Other components	8
7 Procedure	8
8 Data acquisition and processing	9
8.1 Data acquisition	9
8.2 Evaluation of data and correction of chromatograms	9
8.3 Data processing	9
8.3.1 Baseline determination	9
8.3.2 Determination of calculation range	9
9 Expression of results	11
9.1 Calibration curve	11
9.2 Calculation of average molecular weight	12
9.3 Differential molecular weight distribution curve	13
9.4 Cumulative molecular weight distribution curve	13
10 Precision	14
11 Test report	14
11.1 General	14
11.2 Apparatus and measurement parameters	14
11.3 Calibration of the system	15
11.4 Calibration curve	15
11.5 Results	15
Annex A (informative) Supplementary information	16
Annex B (informative) Narrow molecular weight distribution standards	17

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 61, *Plastics*, Subcommittee SC 5, *Physical-chemical properties*.

This third edition cancels and replaces the second edition (ISO 16014-1:2012), which has been technically revised. The main changes compared to the previous edition are as follows:

- publication dates of references have been removed;
- the term “system peak” has been added to [Figure 4](#);
- molecular mass has been changed to molecular weight according to IUPAC rule.

A list of all parts in the ISO 16014 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Plastics — Determination of average molecular weight and molecular weight distribution of polymers using size-exclusion chromatography —

Part 1: General principles

1 Scope

This document specifies a general method for determining the average molecular weight and the molecular weight distribution of polymers using size-exclusion chromatography (SEC).

The average molecular weight and the molecular weight distribution are calculated from a calibration curve constructed using polymer standards if using one of the SEC techniques described in ISO 16014-2 to ISO 16014-4 or from a calibration curve constructed using absolute molecular weight data if using size-exclusion chromatography coupled with light-scattering detection (SEC-LS) as described in ISO 16014-5.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 472, *Plastics — Vocabulary*

ISO 16014-2, *Plastics — Determination of average molecular weight and molecular weight distribution of polymers using size-exclusion chromatography — Part 2: Universal calibration method*

ISO 16014-3, *Plastics — Determination of average molecular weight and molecular weight distribution of polymers using size-exclusion chromatography — Part 3: Low-temperature method*

ISO 16014-4, *Plastics — Determination of average molecular weight and molecular weight distribution of polymers using size-exclusion chromatography — Part 4: High-temperature method*

ISO 16014-5, *Plastics — Determination of average molecular weight and molecular weight distribution of polymers using size-exclusion chromatography — Part 5: Light-scattering method*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 472 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

light-scattering detection

LS detection

technique for determining the mass or size of polymer molecules dissolved in solution by measuring the light scattered by the polymer molecules

4 Principle

A polymer sample is dissolved in a suitable solvent to make a dilute solution. This solution is injected into the mobile phase and onto the SEC column, which is packed with non-adsorbing material made up of small particles having pores of similar or varying size. As the polymer sample passes through the column, the polymer molecules are separated from each other according to the difference in their molecular masses, or more precisely, the difference in their molecular sizes (i.e. their hydrodynamic volume). In SEC, the larger-size molecules cannot permeate into the pores, and thus elute faster, while smaller molecules can permeate into the pores and elute more slowly. The polymer concentration in the eluate is continuously monitored by a concentration-sensitive detector (coupled to a light-scattering detector if SEC-LS is being used) to give an SEC chromatogram.

In the SEC techniques described in ISO 16014-2 to ISO 16014-4, the molecular weight at any elution time on the SEC chromatogram is determined from a calibration curve which is constructed using reference polymer standards with a narrow molecular weight distribution. In SEC-LS, described in ISO 16014-5, a calibration curve constructed using absolute molecular weight data obtained from the SEC-LS chromatogram at any elution time is used. The average molecular weight and the molecular weight distribution of the unknown polymer are calculated by using the molecular weight and concentration data corresponding to each elution time.

5 Reagents

5.1 Eluent

The required purity of the eluent used for SEC varies with the application, but in general the solvent should be free of particulate matter and substances that react with the polymer or interfere with detection of the polymer. Additives such as antioxidants and salts can be used to prevent the degradation of the eluent, the aggregation of polymer molecules, the adsorption of the polymer on the packing material and for other purposes. A mixed eluent may also be used in the SEC techniques described in ISO 16014-2 to ISO 16014-4 to modify the solubility and the refractive index, or to reduce the cost of the mobile phase. A mixed eluent cannot be used for SEC-LS measurements, however, because the polymer selectively adsorbs the components of the mixed eluent, thus giving an erroneous result.

5.2 Reagent for column evaluation

A low molecular weight compound is used for the determination of the theoretical plate number, asymmetry factor and resolution factor of the column.

5.3 Molecular weight standards

The test methods described in ISO 16014-2 to ISO 16014-4 are not absolute methods but relative ones; they require a calibration curve for the calculation of the average molecular weight and the molecular weight distribution from the SEC chromatogram. This calibration curve is plotted using standards of known molecular weight and narrow molecular weight distribution. The value of M_w and/or M_n of the standard determined by an absolute method, such as light scattering, membrane osmometry, vapour pressure osmometry, ultracentrifugation or end-group analysis. The polydispersity M_w/M_n is calculated by dividing the absolute value of M_w by the absolute value of M_n . The polydispersity of the polymer standards shall lie within the following ranges:

$$M_p < 2 \cdot 10^3 \quad M_w/M_n < 1,20$$

$$2 \cdot 10^3 \leq M_p < 10^6 \quad M_w/M_n < 1,10$$

$$10^6 \leq M_p \quad M_w/M_n < 1,20$$

where

M_w is the mass-average molecular weight;

M_n is the number-average molecular weight;

M_p is the molecular weight at peak maximum.

M_p can be calculated from [Formula \(1\)](#):

$$M_p = (M_n \times M_w)^{1/2} \quad (1)$$

NOTE Some commercially available molecular weight standards specify the value of M_w and M_n but not M_p . In such cases, [Formula \(1\)](#) can be used to provide the value of M_p , provided the molecular weight distribution of the polymer sample is a logarithmic normal distribution.

Some examples of commercially available molecular weight standards are given in [Annex B](#).

In the case of SEC-LS (see ISO 16014-5), such molecular weight standards are unnecessary because SEC-LS is an absolute method.

5.4 Reagent for flow rate marker (internal standard)

A low molecular weight compound is used to monitor the accuracy of the elution time, i.e. to evaluate whether or not the data are within the specification.

5.5 Additives

Additives to the eluents may be used to improve SEC performance and prevent sample degradation and the like.

6 Apparatus

6.1 General

A schematic diagram of an SEC system is shown in [Figure 1](#). The essential components are an eluent reservoir, a pumping system, an injector, column(s), a detector, tubing, a recorder, a temperature-control system, and a data-processing system. For SEC-LS measurements, a light-scattering detector, i.e. a molecular-weight-sensitive detector, is coupled to the normal (concentration-sensitive) detector. Any light-scattering detector that meets the requirements specified for this method may be used.

6.2 Eluent reservoir


The eluent reservoir should preferably have sufficient capacity to hold the amount of eluent required for column calibration and successive measurements. Dissolved air in the eluent should preferably be removed before use by placing the solvent in a suitable container designed to reduce the pressure and placing this container in an ultrasonic bath, or by using a vacuum degasser between the reservoir and the pumping system. Particles in the eluent may be removed by membrane filtration. It is desirable in addition to bubble an inert gas through the eluent in the reservoir and blanket the surface of the eluent with the gas, and to shield the reservoir from light.

6.3 Pumping system

A constant, pulseless flow of eluent through the column is desirable. The flow rate should preferably be approximately 1 cm³/min for a column of approximately 8 mm inner diameter. The SEC system

shall have an overall flow-rate precision of within $\pm 0,3\%$. Lower flow rates are recommended for high molecular weight and/or shear-sensitive polymers and viscous eluents. To keep the flow rate constant, temperature control providing a stability of at least $\pm 1\text{ K}$ is required for the pumping system.

The flow rate shall be monitored by the use of an internal standard, or by a direct method such as volume or mass measurements (see A.2), and corrected in the event of significant deviations. In the methods described in ISO 16014-2 to ISO 16014-4, knowledge of the value of the flow rate is not required because the method is a relative one in which the result is calculated from a calibration curve constructed from measurements on molecular weight standards.



Key

- 1 eluent reservoir
- 2 pump
- 3 injector
- 4 columns
- 5 detector
- 6 computer
- 7 recorder
- 8 display
- 9 plotter
- 10 to waste

Figure 1 — Schematic diagram of SEC system

6.4 Injector

In addition to having an eluent bypass capability, the injector shall be able to hold the sample solution and inject the sample solution into the columns with minimum band broadening and minimum pressure change.

To maintain the required precise flow rate, temperature control equipment, or a precise air conditioner, is required for the injection system.

6.5 Columns

6.5.1 General

The function of the columns is to separate the sample molecules according to differences in their molecular size (mass). Columns usually consist of a stainless-steel tube with end fittings, filters and a porous packing material. There is no limitation on the column length or diameter or on the packing-material particle size.

6.5.2 Determination of theoretical plate number

Use a low molecular weight compound, such as ethylbenzene, to obtain a peak (see [Figure 2](#)) and calculate the theoretical plate number N of the set of columns from [Formulae \(2\)](#) or [\(3\)](#):

$$N = 5,54 \times (t_e / W_{1/2})^2 \quad (2)$$

$$N = 16 \times (t_e / W)^2 \quad (3)$$


where

t_e is the elution time to the peak maximum;

$W_{1/2}$ is the peak width at half height;

W is the distance between the points of intersection of the two tangents to the peak with the baseline.

Requirements for the theoretical plate number are specified in ISO 16014-3 and ISO 16014-4.



Key

1	injection
H	height of peak
$H_{1/10}$	one-tenth of peak height
$H_{1/2}$	half of peak height
W	distance between the points of intersection of the two tangents to the peak with the baseline
$W_{1/2}$	peak width at half height
a	width of the leading half of the peak at 10 % peak height
b	width of the trailing half of the peak at 10 % peak height
t_e	elution time to the peak maximum
X	elution time, min
Y	refractive index, mV

Figure 2 — SEC chromatogram of a low molecular weight compound

6.5.3 Determination of resolution factor

The resolution factor R of the set of columns can be calculated from [Formula \(4\)](#) by the use of the calibration curve (see [9.1](#) and [Figure 5](#)) and a molecular weight standard (see [5.3](#) and [Figure 3](#)) with a narrow molecular weight distribution that elutes at a point close to the apex of the sample peak:


$$R = -1 / (D \times W_{STD}) \quad (4)$$

where

D is the slope of the calibration curve at the point corresponding to the apex of the sample peak;

W_{STD} is the peak width at the baseline of the molecular weight standard.

Requirements for the resolution factor are specified in ISO 16014-3 and ISO 16014-4.

**Key**

1 injection

 W_{STD} peak width at the baseline of the molecular weight standard

X time

Y signal strength

Figure 3 — SEC chromatogram of a narrow molecular weight distribution standard**6.5.4 Determination of asymmetry factor**

The asymmetry factor, A_S , of the set of columns can be calculated from [Formula \(5\)](#), using data obtained from the peak produced by a low molecular weight compound such as ethylbenzene (see [Figure 2](#)):

$$A_S = (a + b) / (2 \times a) \quad (5)$$

where

A_S is the asymmetry factor;

a is the width of the leading half of the peak at 10 % peak height;

b is the width of the trailing half of the peak at 10 % peak height.

Requirements for the asymmetry factor are specified in ISO 16014-3 and ISO 16014-4.

6.6 Detector

The detector is used to continuously monitor the concentration of the polymer in the eluent coming off the columns. There are several types of commercially available concentration-sensitive detector, such as the refractive index detector, ultraviolet/visible detector, infrared detector and fluorescence detector. For SEC-LS measurements, a light-scattering detector connected directly to the SEC system is used.

The volume of the flow cell shall be sufficiently small so as to maintain the narrow molecular weight distribution of the molecules separated by the columns and to maintain the overall theoretical plate number and the resolution factor of the set of columns determined in [6.5.2](#) and [6.5.3](#).

The sensitivity of the concentration-sensitive detector used in ISO 16014-2 to ISO 16014-4 shall be such that it can detect a difference in refractive index of 10^{-8} or a difference in UV absorbance of 10^{-4} . The signal/noise ratio shall normally be greater than 200, although a lower ratio is admissible in the case of extremely broad molecular weight distributions or low-concentration measurements on extremely high molecular weight samples. In such cases, the signal/noise ratio shall, however, be greater than 20. Signal drift shall be less than 10 % of the peak height per hour, at the appropriate maximum sensitivity level. The sensitivity of the LS detector is not specified because it can vary significantly, depending on the type of detector and the experimental conditions, for instance.

NOTE Many copolymers can have different molecular compositions, and this can cause problems with detectors such as UV, RI and LS detectors. As the SEC methods described in ISO 16014-1 to ISO 16014-4 are relative methods, they are applicable to such copolymers. SEC-LS (see ISO 16014-5), on the other hand, is not applicable to copolymers whose molecular composition can vary (see [A.1](#)).

6.7 Tubing

The inner diameter and length (including swage length) of the tubing used to connect the sample injector to the first column, the columns to each other and the last column to the detector shall be as small and short as possible to prevent the separated fractions from remixing and to ensure that the performance requirements specified in [6.5.1](#) are met. The inner diameter of the tubing used from the injector to the detector shall be 0,05 cm or less. Care shall be taken, however, not to use tubing of too small an inner diameter so as to avoid rupture of the polymer chain and turbulence in the detector cell.

6.8 Temperature control

The temperature of the columns, pumping system, injection system and tubing shall be kept constant within a narrow range as described in the appropriate subclause for each component. In the case of the detector, the temperature shall be controlled to meet the performance requirements for SEC.

6.9 Recorder and plotter

The SEC curve shall be recorded or plotted clearly enough to assess whether parameters such as peak height, baseline level, signal drift and peak separation are suitable for data processing.

6.10 Data-processing system

A data-processing system capable of data acquisition, generation of calibration curves, calculation of the required molecular weights and molecular weight distributions, and presentation of appropriate data and/or graphics is required. This system shall be capable of collecting, analysing and reporting data in the manner specified in the relevant parts of the ISO 16014 series.

It is desirable that the SEC or SEC-LS chromatogram is generated in real time, but the data may also be stored for subsequent processing off line.

6.11 Other components

In addition to the components described above, a column guard filter, a pressure monitor, a pulse damper or other related components may be used, if necessary.

7 Procedure

The procedure includes setting up the SEC or SEC-LS apparatus and the data acquisition and processing system, preparing solutions of molecular weight standards, test solutions, and solutions for determining column performance, filtering the solutions and injecting them.

Details of the exact procedure to be used are given in the relevant part of the ISO 16014 series.

8 Data acquisition and processing

8.1 Data acquisition

Data acquisition shall be carried out from the onset of sample elution to the end of the sample peak or when the signal drops back to the baseline. The number of data points or readings shall be at least 50 per decade of molecular weight. Care shall be taken to use enough data points to provide accurate estimates of the peak area and the elution time at the peak apex, as well as an accurate molecular weight distribution curve and average molecular weight derived from the curve.

8.2 Evaluation of data and correction of chromatograms

SEC chromatograms shall be monitored to determine whether the error in the elution time of the internal standard or that in the volume or mass of eluent is less than or equal to $\pm 0,3\%$ of the calibration value for each run.

If the error is greater than $\pm 0,3\%$, the data shall be rejected and the measurement repeated. Peak-broadening corrections are not required.

NOTE The elution-time and elution-volume/mass monitoring methods are described in [A.2](#) of [Annex A](#).

8.3 Data processing

8.3.1 Baseline determination

If the detector signal drops to the baseline before the system peak as shown in [Figure 4 a\)](#), the baseline shall be assumed to be a straight line from t_a to t_b .

If the detector signal does not recover to the baseline before the system peak as shown in [Figure 4 b\)](#), the baseline shall be assumed to be a straight line connecting the point t_a just before sample elution and the point t_c just after the system peak.

The baseline of an SEC-LS chromatogram shall be assumed to be the same as that of a chromatogram obtained using one of the SEC methods described in ISO 16014-2 to ISO 16014-4, i.e. a straight line connecting the point just before sample elution with the point just after baseline recovery. In general, an SEC-LS chromatogram is modified by eliminating spike noise and smoothing the baseline to remove short-term noise before making a decision on the baseline.

8.3.2 Determination of calculation range


If the sample does not contain components of molecular weight $< 1\ 000$, as shown in [Figure 4 a\)](#), the range between points t_a and t_b on the baseline shall be used for the calculation.

If the sample contains components of molecular weight $< 1\ 000$, and these low molecular weight components make up $< 30\%$ of the total polymer peak area, as shown in [Figure 4 b\)](#), one of the following two procedures shall be used for the calculation.

- Calculate the area below the curve from point t_a to a point $t_{1\ 000}$ corresponding to a molecular weight of 1 000.
- Calculate the area from point t_a to t_d which covers the entire polymer, including oligomers and monomer but excluding additives. Point t_d is determined by producing a chromatogram of the eluent alone.

If the sample contains components of molecular weight <1 000, and these low molecular weight components make up >30 % of the total polymer peak area, as shown in [Figure 4 c\)](#), the method described in this document is not recommended.

NOTE The detector response can vary at low molecular weight, and the presence of a significant proportion of low molecular weight material makes any calculation unreliable.



Key

1	system peak	t_d	elution time of molecular weight d
t_a	elution time of molecular weight a	$t_{1\,000}$	elution time of molecular weight 1 000
t_b	elution time of molecular weight b	X	time
t_c	elution time of molecular weight c	Y	signal strength

NOTE 1 System peak is signal peculiar to the SEC using a refractive index (RI) detector.

NOTE 2 The system peak signal appears at the total penetration limit of the columns. These peaks are not part of the sample, but of the overall system.

Figure 4 — Typical SEC curves

9 Expression of results

9.1 Calibration curve

For the SEC methods described in ISO 16014-2 to ISO 16014-4, prepare the calibration curve by plotting elution times against $\lg M_p$ as shown in [Figure 5](#). The values of M_p are obtained using one of the following ways:

- from the data sheets for the standard materials;
- by calculation, using [Formula \(1\)](#), from the values of M_w and M_n given in the data sheets for the standard materials;
- by calculation, using [Formula \(1\)](#), from the values of M_w or M_n and M_w/M_n given in the data sheets for the standard materials.

Polynomials containing terms up to t^3 are widely used to describe calibration curves. The addition of higher powers might improve the fit of the curve to the data:

$$\lg M = A_0 + A_1 t \quad (6)$$

$$\lg M = A_0 + A_1 t + A_2 t^2 + A_3 t^3 \quad (7)$$

where


M is the molecular weight;

A_0, A_1, A_2, A_3 are the coefficients;

t is the elution time.

Other methods, or a combination of methods, may be used to improve the fit.

For the preparation of a universal calibration curve, see ISO 16014-2. For the preparation of a calibration curve for SEC-LS, see ISO 16014-5.

**Key**

t elution time (min)

$\lg M$ natural logarithm of the molecular weight

Figure 5 — Calibration curve

9.2 Calculation of average molecular weight

Calculate the molecular weight, M_i , and signal intensity, H_i , at each elution time using the calibration curve (see 9.1) and the SEC chromatogram of the polymer sample for which the baseline and the calculation range have been determined (see 8.3.1 and 8.3.2), as follows:

- calculate the molecular weight, M_i , at the i th elution time, t_i , by inserting t_i in [Formula \(6\)](#) or [Formula \(7\)](#);
- calculate the signal intensity, H_i , at the i th elution time, t_i , by subtracting the baseline signal intensity from the total detector signal intensity at the elution time, t_i .

The average molecular weight and the polydispersity can be calculated from the values of M_i and H_i using [Formulae \(8\)](#) to [\(12\)](#), where n denotes the n th set of data:

$$M_n = \frac{\sum_{i=1}^n H_i}{\sum_{i=1}^n (H_i / M_i)} \quad (8)$$

$$M_w = \frac{\sum_{i=1}^n (H_i \times M_i)}{\sum_{i=1}^n H_i} \quad (9)$$

$$M_z = \frac{\sum_{i=1}^n (H_i \times M_i^2)}{\sum_{i=1}^n (H_i \times M_i)} \quad (10)$$

$$M_v = \left[\frac{\sum_{i=1}^n H_i \times M_i^a}{\sum_{i=1}^n H_i} \right]^{1/a} \quad (11)$$

$$\text{Polydispersity} = M_w / M_n \quad (12)$$

9.3 Differential molecular weight distribution curve

The differential molecular weight distribution curve is prepared by plotting $dW_i/d(\lg M_i)$ against $\lg M_i$ as shown in Figure 6. W_i is calculated from Formulae (13) to (15):


$$\Delta W_i = \frac{H_i}{\sum_{i=1}^n H_i} \quad (13)$$

$$w_i = \Delta W_i \times \frac{1}{I} \quad (14)$$

$$\frac{dW_i}{d(\lg M_i)} = -w_i \times \frac{dt_i}{d(\lg M_i)} \quad (15)$$

where I is the data acquisition interval.

If the sample contains components of molecular weight $<1\ 000$, and these low molecular weight components make up $<30\ %$ of the sample, draw a vertical line at the point corresponding to $M_{1\ 000}$.



Key

Y $dW/d(\lg M)$

Figure 6 — Differential molecular weight distribution curve

9.4 Cumulative molecular weight distribution curve

The cumulative molecular weight distribution curve is prepared by plotting the mass fraction C_i versus $\lg M_i$ as shown in Figure 7, C_i being calculated from Formula (16):

$$C_i = \sum_{j=1}^i (\Delta W_j + \Delta W_j) / 2 \quad (16)$$