

---

---

**Wheat flour and durum wheat  
semolina — Determination of colour  
by diffuse reflectance colorimetry**

*Farine de blé tendre et semoule de blé dur — Détermination de la  
couleur par colorimétrie de réflectance diffuse*

STANDARDSISO.COM : Click to view the full PDF of ISO 16624:2020



STANDARDSISO.COM : Click to view the full PDF of ISO 16624:2020



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Fax: +41 22 749 09 47  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

|                                                                          | Page      |
|--------------------------------------------------------------------------|-----------|
| Foreword .....                                                           | iv        |
| <b>1 Scope .....</b>                                                     | <b>1</b>  |
| <b>2 Normative references .....</b>                                      | <b>1</b>  |
| <b>3 Terms and definitions .....</b>                                     | <b>1</b>  |
| <b>4 Principle .....</b>                                                 | <b>1</b>  |
| <b>5 Apparatus .....</b>                                                 | <b>2</b>  |
| <b>6 Sample preparation .....</b>                                        | <b>2</b>  |
| <b>7 Procedure .....</b>                                                 | <b>2</b>  |
| 7.1 General .....                                                        | 2         |
| 7.2 Setting of the colorimeter .....                                     | 3         |
| 7.3 Colorimetric determination .....                                     | 3         |
| <b>8 Expression of the results .....</b>                                 | <b>3</b>  |
| <b>9 Precision .....</b>                                                 | <b>3</b>  |
| 9.1 General .....                                                        | 3         |
| 9.2 Repeatability limit, $r$ .....                                       | 3         |
| 9.3 Reproducibility limit, $R$ .....                                     | 3         |
| 9.4 Critical difference, $d_C$ .....                                     | 4         |
| 9.4.1 General .....                                                      | 4         |
| 9.4.2 Comparison of two groups of measurements in one laboratory .....   | 4         |
| 9.4.3 Comparison of two groups of measurements in two laboratories ..... | 4         |
| 9.5 Uncertainty, $U$ .....                                               | 4         |
| <b>10 Test report .....</b>                                              | <b>5</b>  |
| <b>Annex A (informative) Results of the interlaboratory test .....</b>   | <b>6</b>  |
| <b>Bibliography .....</b>                                                | <b>12</b> |

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 4, *Cereals and pulses*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 338, *Cereal and cereal products*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

# Wheat flour and durum wheat semolina — Determination of colour by diffuse reflectance colorimetry

## 1 Scope

This document specifies a method for the determination of colour in durum wheat semolina and wheat flour by diffuse reflectance colorimetry.

It is applicable to industrial semolina and flour.

The method can be applicable to flour obtained from experimental mill.

## 2 Normative references

There are no normative references in this document.

## 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

### 3.1

#### **colour metric space**

expression of the colour of an object or of a light source by some parameters expressed by figures

### 3.2

#### **illuminant**

light source characterized by a spectral curve, where the energy relative distribution is defined in the field of wavelengths that are able to influence the object colour vision

## 4 Principle

The principle is based on the measurement of colour directly on semolina and flour by a reflectance colorimeter.

The colour of wheat milling product (semolina and flour) is due to the pigments naturally present in wheat grains. These pigments (xanthophyll's and carotenoids) are responsible for the colour visually perceived in milling products.

## 5 Apparatus

**5.1 Reflectance colorimeter**<sup>1)</sup> with head of measurement suitable for carrying out measurements of absolute chromaticity. Consists of a setting system and a cell samples-driver.

The colorimeter shall be characterized by the following technical characteristics:

- system of measurement with pulsed xenon lamp diffused to the light and receipt of the radiation reflected to 0° (geometry d/0°);
- circular surface of measurement;
- measurements of chromaticity expressed as  $L^*$ ,  $a^*$ ,  $b^*$  (see CIE 1976) with the use of the illuminant CIE D<sub>65</sub> (illuminant D<sub>65</sub>: representing one of the relative spectral distribution of the day-light energy that corresponds to a proximal colour temperature similar to 6 504 K);
- time of measurement 1 s;
- possibility of calibration with reference plate;
- repeatability within a DE \* 0,6 (30 measurements effected to an interval of 10 s on the reference plate).

**5.2 Accessory samples-driver and cell for the measurement of granular materials.**

The dimensions of the cell that defines the quantity of sample submitted to the test are: external diameter = 60 mm, diameter inside hole = 22 mm, thickness = 9 mm.

**5.3 Reference plate in porcelain for the initial setting of the colorimeter.**

## 6 Sample preparation

Before the analysis, the samples shall be carefully homogenized.

Sampling is not part of the method specified in this document. A recommended sampling method is given in ISO 24333.

## 7 Procedure

### 7.1 General

Before each series of measurements, the apparatus shall be calibrated.

The colorimeter calibration shall be made through opaque stable materials (as ceramics, glaze etc.) samples supplied by manufacturers. When the colorimeter (5.1) is used, a further calibration for better measurement accuracy can be performed using a reference material similar to the colour of the samples to be measured.

Before the calibration, verify the integrity of the reference plate (5.3) used as a reference (e.g. for the absence of linings or colour not homogeneous). Also, for the setting, verify that the coordinates are those reported on the reference standard.

1) Laboratories involved in the ring test nearly all used a colorimeter CR 400 or CR 410 Minolta. These are fit apparatus responding to the required technical characteristics. Minolta is a trade name and is an example of a suitable apparatus available commercially. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of these products.

## 7.2 Setting of the colorimeter

To set the colorimeter (5.1), put the head of measurement on the centre of the reference plate (5.3) and select the colour metric space  $y, x, z$  with illuminant  $D_{65}$  in the system CIE Lab.

At the end, return in the colour metric space  $L^*, a^*, b^*$ , and therefore prepare the colorimeter to the reading of the samples.

The reference plate shall carefully be cleaned after every setting, being careful to avoid any rubbing or lining. After the setting, the plate shall be stored in the dark.

## 7.3 Colorimetric determination

From a homogeneous sample of at least 100 g, withdraw a share, fill with care the cell for granular samples and measure. Every determination shall be performed twice.

The cell has to be emptied completely and cleared between two determinations. The sample shall be mixed each time.

## 8 Expression of the results

The colour metric space CIELAB (1976) is the most used for measuring the colour of objects.

$L^*$  indicates the lightness and extends from 0 (black) to 100 (white);  $a^*$  and  $b^*$  are chromaticity indexes, respectively;  $a^*$ : red/green ( $a > 0$  red;  $a < 0$  green);  $b^*$ : yellow/blue ( $a > 0$  yellow;  $a < 0$  blue).

The results are expressed as yellow index ( $b^*$ ). It is possible to have a datum of dark coloration through the measure “100- $L^*$ ” and of red coloration through the measure “ $a^*$ .”

The results are expressed with two decimal places.

## 9 Precision

### 9.1 General

The values derived for the interlaboratory tests, given in Annex A, may not be applicable to concentration ranges and matrices other than those given, i.e. for  $L^*$  between 83,37 and 92,47;  $a^*$ : -2,17 and -0,18;  $b^*$ : 8,56 and 27,54.

### 9.2 Repeatability limit, $r$

The absolute difference between two independent single test results, obtained using the same method on identical test material in the same laboratory by the same operator using the same equipment within a short interval of time, will in not more than 5 % of cases be greater than the repeatability limit  $r$  given in Table 1.

**Table 1 — Repeatability limit**

| Parameter                               | $L^*$                 | $a^*$                     | $b^*$                |
|-----------------------------------------|-----------------------|---------------------------|----------------------|
| Repeatability standard deviation, $s_r$ | $-0,0079L^* + 0,7688$ | 0,03                      | $0,0100b^* - 0,0186$ |
| Repeatability limit, $r$                | $-0,0219L^* + 2,1296$ | $2,77 \times 0,03 = 0,08$ | $0,0277b^* - 0,0515$ |

### 9.3 Reproducibility limit, $R$

The absolute difference between two single test results, obtained using the same method on identical test material in different laboratories with different operators using different equipment, will in not more than 5 % of cases be greater than the reproducibility limit  $R$  given in Table 2.

**Table 2 — Reproducibility limit**

| Parameter                                 | $L^*$                      | $a^*$                     | $b^*$                    |
|-------------------------------------------|----------------------------|---------------------------|--------------------------|
| Reproducibility standard deviation, $s_R$ | $-0,033\ 6L^* + 4,680\ 8$  | 0,63                      | $0,038\ 2b^* + 0,057\ 1$ |
| Reproducibility limit, $R$                | $-0,093\ 1L^* + 12,965\ 8$ | $2,77 \times 0,63 = 1,75$ | $0,105\ 8b^* + 0,158\ 2$ |

## 9.4 Critical difference, $d_C$

### 9.4.1 General

Critical difference is the difference between two averaged values obtained from two test results under repeatability conditions.

### 9.4.2 Comparison of two groups of measurements in one laboratory

The critical difference ( $d_{C,r}$ ) between two averaged values obtained in one laboratory from two test results under repeatability conditions is equal to [Formula \(1\)](#):

$$d_{C,r} = 2,77 s_r \sqrt{\frac{1}{2n_1} + \frac{1}{2n_2}} = 2,77 s_r \sqrt{\frac{1}{2}} = 1,98 s_r \quad (1)$$

where

$s_r$  is the standard deviation of repeatability;

$n_1$  and  $n_2$  are the number of test results corresponding to each of the averaged values ( $n_1 = n_2 = 2$ ).

### 9.4.3 Comparison of two groups of measurements in two laboratories

The critical difference ( $d_{C,R}$ ) between two averaged values obtained in two different laboratories from two test results under repeatability conditions is equal to [Formula \(2\)](#):

$$d_{C,R} = 2,77 \sqrt{s_R^2 - s_r^2 \left(1 - \frac{1}{2n_1} - \frac{1}{2n_2}\right)} = 2,77 \sqrt{s_R^2 - 0,5 s_r^2} \quad (2)$$

where

$s_r$  is the standard deviation of repeatability;

$s_R$  is the standard deviation of reproducibility;

$n_1$  and  $n_2$  are the number of test results corresponding to each of the averaged values ( $n_1 = n_2 = 2$ ).

## 9.5 Uncertainty, $U$

The measurement uncertainty,  $U$ , is a parameter representing the distribution of the values that may reasonably be attributed to the result.

It is possible to evaluate measurement uncertainties using data obtained from studies carried out in accordance with ISO 5725-2. The reproducibility standard deviation obtained during an interlaboratory test is a valid basis to evaluate measurement uncertainty because, by definition, uncertainty characterizes the dispersion of values that can be reasonably attributed to the parameter. The calculated expanded standard uncertainty shall be plus or minus twice the standard deviations of reproducibility given in this document.



## 10 Test report

The test report shall contain at least the following information:

- a) all the required information for the complete identification of the sample;
- b) the method used, with reference to this document, i.e. ISO 16624;
- c) the final result obtained;
- d) all operational details not specified in this document, or considered optional, as well as any incident that could have influenced the test(s) result(s);
- e) the data of the test.

STANDARDSISO.COM : Click to view the full PDF of ISO 16624:2020

## Annex A (informative)

### Results of the interlaboratory test

The values for repeatability limit and reproducibility limit have been derived from the results of interlaboratory test carried out in accordance with ISO 5725-1, ISO 5725-2 and ISO 5725-6.

An international interlaboratory test was organized by the CRA-Cereal Quality Research Unit (CRA-QCE), Italy.

This international interlaboratory test was carried out on four industrial samples of durum wheat semolina (FD) and four industrial samples of common wheat flour (FT).


Twenty-seven laboratories took part in this test and all of them sent their results. Specifically, the test was attended by 15 Italian laboratories, 5 French laboratories, 5 Spanish laboratories, 2 Portugal laboratories and 1 Philippine laboratory.

Each laboratory analysed each sample five times in duplicate. The processing of the data showed that the accuracy of the results is the same when comparing the average of the ten or the two first measurements. The statistic results presented were therefore determined from the analysis of each sample in duplicate.

The results are given in [Tables A.1, A.2 and A.3](#) and [Figures A.1, A.2 and A.3](#).

**Table A.1 — Statistical results of the interlaboratory test for  $L^*$**

| $L^*$                                                                                                        | FD4   | FD3   | FD2   | FT1   | FD1   | FT3   | FT4   | FT2   |
|--------------------------------------------------------------------------------------------------------------|-------|-------|-------|-------|-------|-------|-------|-------|
| Number of laboratories after outlier elimination                                                             | 20    | 25    | 24    | 26    | 26    | 25    | 22    | 23    |
| Global mean                                                                                                  | 83,37 | 86,01 | 87,55 | 88,45 | 89,63 | 91,44 | 92,05 | 92,47 |
| Repeatability standard deviation, $s_r$                                                                      | 0,14  | 0,06  | 0,05  | 0,08  | 0,07  | 0,05  | 0,04  | 0,05  |
| Repeatability coefficient of variation, $C_{V,r}$ , in %                                                     | 0,17  | 0,07  | 0,06  | 0,09  | 0,08  | 0,06  | 0,05  | 0,05  |
| Limit of repeatability, $r = (2,77 \times s_r)$                                                              | 0,39  | 0,16  | 0,14  | 0,21  | 0,19  | 0,15  | 0,12  | 0,13  |
| Reproducibility standard deviation, $s_R$                                                                    | 1,87  | 1,81  | 1,76  | 1,64  | 1,67  | 1,62  | 1,63  | 1,52  |
| Reproducibility coefficient of variation, $C_{V,R}$ , in %                                                   | 2,24  | 2,10  | 2,02  | 1,86  | 1,86  | 1,78  | 1,77  | 1,65  |
| Limit of reproducibility, $R = (2,77 \times s_R)$                                                            | 5,18  | 5,01  | 4,89  | 4,55  | 4,63  | 4,50  | 4,51  | 4,22  |
| <b>Key</b><br>FD: industrial samples of durum wheat semolina<br>FT: industrial samples of common wheat flour |       |       |       |       |       |       |       |       |

**Key**

- 1 reproducibility standard deviation,  $s_R = -0,033\ 6x + 4,680\ 8 R^2 = 0,8904$   
 2 repeatability standard deviation,  $s_r = -0,007\ 9x + 0,768\ 8 R^2 = 0,5972$

**Figure A.1 — Relationship between precision standard deviations and  $L^*$**

Figure A.1 shows that repeatability and reproducibility standard deviations are dependent of the average value  $L^*$ .

Reproducibility limit,  $R = 2,77 \times (-0,033\ 6L^* + 4,680\ 8) = -0,093\ 1L^* + 12,965\ 8$

Repeatability limit,  $r = 2,77 \times (-0,007\ 9L^* + 0,768\ 8) = -0,021\ 9L^* + 2,129\ 6$

**Table A.2 — Statistical results of the interlaboratory test for  $a^*$**

| $a^*$                                                                                                        | FD2   | FD3   | FD1   | FD4   | FT2   | FT4   | FT3   | FT1   |
|--------------------------------------------------------------------------------------------------------------|-------|-------|-------|-------|-------|-------|-------|-------|
| Number of laboratories after outlier elimination                                                             | 24    | 25    | 22    | 23    | 25    | 25    | 26    | 25    |
| Global mean                                                                                                  | -2,17 | -2,13 | -2,00 | -1,67 | -1,16 | -0,71 | -0,52 | -0,18 |
| Repeatability standard deviation, $s_r$                                                                      | 0,02  | 0,03  | 0,02  | 0,05  | 0,02  | 0,01  | 0,02  | 0,02  |
| Repeatability coefficient of variation, $C_{V,r}$ in %                                                       | -1,05 | -1,30 | -0,93 | -3,02 | -1,47 | -1,60 | -2,93 | -8,56 |
| Limit of repeatability, $r = (2,77 \times s_r)$                                                              | 0,06  | 0,08  | 0,05  | 0,14  | 0,05  | 0,03  | 0,04  | 0,04  |
| Reproducibility standard deviation, $s_R$                                                                    | 0,63  | 0,71  | 0,59  | 0,75  | 0,55  | 0,63  | 0,61  | 0,52  |
| <b>Key</b><br>FD: industrial samples of durum wheat semolina<br>FT: industrial samples of common wheat flour |       |       |       |       |       |       |       |       |

Table A.2 (continued)

| $a^*$                                                                                                        | FD2    | FD3    | FD1    | FD4    | FT2    | FT4    | FT3     | FT1     |
|--------------------------------------------------------------------------------------------------------------|--------|--------|--------|--------|--------|--------|---------|---------|
| Reproducibility coefficient of variation, $C_{VR}$ , in %                                                    | -29,24 | -33,22 | -29,54 | -44,75 | -47,17 | -88,76 | -118,70 | -290,44 |
| Limit of reproducibility, $R = (2,77 \times s_R)$                                                            | 1,76   | 1,96   | 1,64   | 2,07   | 1,52   | 1,75   | 1,69    | 1,44    |
| <b>Key</b><br>FD: industrial samples of durum wheat semolina<br>FT: industrial samples of common wheat flour |        |        |        |        |        |        |         |         |




Figure A.2 — Relationship between precision standard deviations and mean values and  $a^*$

Figure A.2 shows that repeatability and reproducibility standard deviations are independent of the average value  $a^*$ .

Repeatability standard deviation,  $s_r = 0,03$


Reproducibility standard deviation,  $s_R = 0,63$

Repeatability limit,  $r = 2,77 \times s_r = 0,08$

Reproducibility limit,  $R = 1,75$

Table A.3 — Statistical results of the interlaboratory for  $b^*$ 

| $b^*$                                                                                                        | FT4  | FT3  | FT1   | FT2   | FD1   | FD2   | FD3   | FD4   |
|--------------------------------------------------------------------------------------------------------------|------|------|-------|-------|-------|-------|-------|-------|
| Number of laboratories after outlier elimination                                                             | 25   | 25   | 25    | 26    | 26    | 26    | 23    | 20    |
| Global mean                                                                                                  | 8,56 | 9,25 | 10,43 | 10,59 | 18,17 | 22,98 | 26,33 | 27,54 |
| Repeatability standard deviation, $s_r$                                                                      | 0,07 | 0,07 | 0,09  | 0,12  | 0,11  | 0,23  | 0,20  | 0,32  |
| Repeatability coefficient of variation, $C_{V,r}$ in %                                                       | 0,76 | 0,77 | 0,82  | 1,17  | 0,60  | 0,98  | 0,75  | 1,15  |
| Limit of repeatability, $r = (2,77 \times s_r)$                                                              | 0,18 | 0,20 | 0,24  | 0,34  | 0,30  | 0,63  | 0,55  | 0,88  |
| Reproducibility standard deviation, $s_R$                                                                    | 0,39 | 0,41 | 0,40  | 0,53  | 0,65  | 1,07  | 0,95  | 1,15  |
| Reproducibility coefficient of variation, $C_{V,R}$ in %                                                     | 4,57 | 4,48 | 3,87  | 5,00  | 3,57  | 4,68  | 3,63  | 4,19  |
| Limit of reproducibility, $R = (2,77 \times s_R)$                                                            | 1,08 | 1,15 | 1,12  | 1,47  | 1,80  | 2,98  | 2,65  | 3,20  |
| <b>Key</b><br>FD: industrial samples of durum wheat semolina<br>FT: industrial samples of common wheat flour |      |      |       |       |       |       |       |       |

**Key**

- 1 reproducibility standard deviation,  $s_R = 0,038\ 2x + 0,057\ 1$   $R^2 = 0,927\ 9$   
 2 repeatability standard deviation,  $s_r = 0,010\ 0x - 0,018\ 6$   $R^2 = 0,815\ 3$

**Figure A.3 — Relationship between precision standard deviations and mean values and  $b^*$** 

Figure A.3 shows that repeatability and reproducibility standard deviations are dependent of the average value  $b^*$ .

Reproducibility limit,  $R = 2,77 \times (0,038\ 2b^* + 0,057\ 1) = 0,105\ 8b^* + 0,158\ 2$

Repeatability limit,  $r = 2,77 \times (0,010\ 0b^* - 0,018\ 6) = 0,027\ 7b^* - 0,051\ 5$

The conversion of precision data into formulae for the calculation of repeatability, reproducibility and critical difference are given in [Tables A.4](#) and [A.5](#).

**Table A.4 —  $L^*$**

| $L^*$ | $s_r$ | $r$  | $s_R$ | $R$  | $d_{C,r}$ | $d_{C,R}$ |
|-------|-------|------|-------|------|-----------|-----------|
| 83,37 | 0,11  | 0,31 | 1,88  | 5,21 | 0,22      | 5,19      |
| 84,00 | 0,11  | 0,29 | 1,86  | 5,15 | 0,21      | 5,13      |
| 84,50 | 0,10  | 0,28 | 1,84  | 5,10 | 0,20      | 5,09      |
| 85,00 | 0,10  | 0,27 | 1,82  | 5,05 | 0,19      | 5,04      |
| 85,50 | 0,09  | 0,26 | 1,81  | 5,01 | 0,18      | 5,00      |
| 86,00 | 0,09  | 0,25 | 1,79  | 4,96 | 0,18      | 4,95      |
| 86,50 | 0,09  | 0,24 | 1,77  | 4,92 | 0,17      | 4,90      |
| 87,00 | 0,08  | 0,23 | 1,76  | 4,87 | 0,16      | 4,86      |
| 87,50 | 0,08  | 0,21 | 1,74  | 4,82 | 0,15      | 4,81      |
| 88,00 | 0,07  | 0,20 | 1,72  | 4,78 | 0,15      | 4,77      |
| 88,50 | 0,07  | 0,19 | 1,71  | 4,73 | 0,14      | 4,72      |
| 89,00 | 0,07  | 0,18 | 1,69  | 4,68 | 0,13      | 4,68      |
| 89,50 | 0,06  | 0,17 | 1,67  | 4,64 | 0,12      | 4,63      |
| 90,00 | 0,06  | 0,16 | 1,66  | 4,59 | 0,11      | 4,58      |
| 90,50 | 0,05  | 0,15 | 1,64  | 4,54 | 0,11      | 4,54      |
| 91,00 | 0,05  | 0,14 | 1,62  | 4,50 | 0,10      | 4,49      |
| 91,50 | 0,05  | 0,13 | 1,61  | 4,45 | 0,09      | 4,45      |
| 92,00 | 0,04  | 0,12 | 1,59  | 4,40 | 0,08      | 4,40      |
| 92,47 | 0,04  | 0,11 | 1,57  | 4,36 | 0,08      | 4,36      |

**Table A.5 —  $b^*$**

| $b^*$ | $s_r$ | $r$  | $s_R$ | $R$  | $d_{C,r}$ | $d_{C,R}$ |
|-------|-------|------|-------|------|-----------|-----------|
| 8,56  | 0,07  | 0,19 | 0,38  | 1,06 | 0,13      | 1,06      |
| 9,00  | 0,07  | 0,20 | 0,40  | 1,11 | 0,14      | 1,10      |
| 10,00 | 0,08  | 0,23 | 0,44  | 1,22 | 0,16      | 1,21      |
| 11,00 | 0,09  | 0,25 | 0,48  | 1,32 | 0,18      | 1,31      |
| 12,00 | 0,10  | 0,28 | 0,52  | 1,43 | 0,20      | 1,41      |
| 13,00 | 0,11  | 0,31 | 0,55  | 1,53 | 0,22      | 1,52      |
| 14,00 | 0,12  | 0,34 | 0,59  | 1,64 | 0,24      | 1,62      |
| 15,00 | 0,13  | 0,36 | 0,63  | 1,75 | 0,26      | 1,72      |
| 16,00 | 0,14  | 0,39 | 0,67  | 1,85 | 0,28      | 1,82      |
| 17,00 | 0,15  | 0,42 | 0,71  | 1,96 | 0,30      | 1,93      |
| 18,00 | 0,16  | 0,45 | 0,74  | 2,06 | 0,32      | 2,03      |
| 19,00 | 0,17  | 0,47 | 0,78  | 2,17 | 0,34      | 2,13      |
| 20,00 | 0,18  | 0,50 | 0,82  | 2,27 | 0,36      | 2,23      |
| 21,00 | 0,19  | 0,53 | 0,86  | 2,38 | 0,38      | 2,33      |
| 22,00 | 0,20  | 0,56 | 0,90  | 2,49 | 0,40      | 2,43      |
| 23,00 | 0,21  | 0,59 | 0,94  | 2,59 | 0,42      | 2,53      |