
**Respiratory protective devices —
Methods of test and test equipment —
Part 2:
Determination of breathing resistance**

*Appareils de protection respiratoire — Méthodes d'essai et
équipement d'essai —*

Partie 2: Détermination de la résistance respiratoire



STANDARDSISO.COM : Click to view the full PDF of ISO 16900-2:2017



COPYRIGHT PROTECTED DOCUMENT

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Prerequisites	2
5 General test requirements	2
6 Method 1: Static breathing resistance for filters and respiratory interfaces (RI)	2
6.1 General	2
6.2 Equipment	2
6.3 Procedure	3
6.3.1 Procedure for filters for respiratory protective devices	3
6.3.2 Procedure for respiratory interfaces	4
7 Method 2: Dynamic breathing resistance	5
7.1 General	5
7.2 Equipment	5
7.3 Procedure for complete RPD	5
7.3.1 General	5
7.3.2 Breathing resistance measurement	5
8 Test report	5
Annex A (normative) Application of uncertainty of measurement	6
Bibliography	8

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 94, *Personal safety — Protective clothing and equipment*, Subcommittee SC 15, *Respiratory protective devices*.

This second edition cancels and replaces the first edition (ISO 16900-2:2009), which has been technically revised.

The main changes compared to the previous edition are as follows:

- relevant respiratory protective device (RPD) headform or torso have been used and previously specified headform and concentric breathing tube assembly have been deleted;
- Annex B has been deleted;
- air volume flow rate has been corrected to the standardized condition of 1 013 hPa and 20 °C.

A list of all parts in the ISO 16900 series can be found on the ISO website.

Introduction

This document is intended as a supplement to the relevant performance standards for respiratory protective devices (RPDs). Test methods are specified for complete devices or parts of devices. If deviations from the test method given in this document are necessary, these deviations will be specified in the relevant performance standard.

The following definitions apply in understanding how to implement an ISO International Standard and other normative ISO deliverables (TS, PAS, IWA):

- “shall” indicates a requirement;
- “should” indicates a recommendation;
- “may” is used to indicate that something is permitted;
- “can” is used to indicate that something is possible, for example, that an organization or individual is able to do something.

STANDARDSISO.COM : Click to view the full PDF of ISO 16900-2:2017

Respiratory protective devices — Methods of test and test equipment —

Part 2: Determination of breathing resistance

1 Scope

This document specifies the method(s) of test for breathing resistance for

- respiratory protective devices (RPDs),
- filters for RPDs, and
- respiratory interfaces (RI).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 16900-5, *Respiratory protective devices — Methods of test and test equipment — Part 5: Breathing machine, metabolic simulator, RPD headforms and torso, tools and verification tools*

ISO 16900-10, *Respiratory protective devices — Methods of test and test equipment — Part 10: Resistance to ignition, flame, radiant heat and heat*

ISO 16972, *Respiratory protective devices — Terms, definitions, graphical symbols and units of measurement*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 16972 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

3.1

static breathing resistance

differential pressure caused by an RPD when the breathing gas is passed through the device at a constant flow

3.2

dynamic breathing resistance

differential pressure caused by an RPD when the breathing gas is moved by a breathing machine, adjusted to a specified minute ventilation including breathing frequency, waveform and tidal volume

4 Prerequisites

In order to implement this document, at least the following parameters shall be specified in the performance standard:

- the number of specimens;
- the selection and type(s) of support such as RPD headform/RPD torso;
- the state in which the RI or RPD shall be tested;
- any prior conditioning or testing;
- flow rate (static test method);
- flow rate (minute ventilation), including breathing frequency, tidal volume and waveform (dynamic test method);
- any deviations from the method(s).

5 General test requirements

Unless otherwise specified, the values stated in this document are expressed as nominal values. Except for temperature limits, values which are not stated as maxima or minima shall be subject to a tolerance of $\pm 5\%$. Unless otherwise specified, the ambient conditions for testing shall be between $16\text{ }^{\circ}\text{C}$ and $32\text{ }^{\circ}\text{C}$ and $(50 \pm 30)\%$ RH. Any temperature limits specified shall be subject to an accuracy of $\pm 1\text{ }^{\circ}\text{C}$.

For each of the required measurements performed in accordance with this document, a corresponding estimate of the uncertainty of measurement should be evaluated. This estimate of uncertainty should be applied and stated when reporting test results, in order to enable the user of the test report to assess the reliability of the result (see [Annex A](#)).

All flow rate values shall be corrected to $1\,013\text{ hPa}$ and $20\text{ }^{\circ}\text{C}$.

NOTE Uncertainty of measurement can be calculated in accordance with JCGM 100[1].

6 Method 1: Static breathing resistance for filters and respiratory interfaces (RI)

6.1 General

The filter or RI is mounted on a support and air is drawn through the device at a constant flow rate.

NOTE 1 The convention of reporting breathing resistance is that if, during the inhalation resistance test, the pressure inside an RI or downstream of a filter relative to atmosphere is negative, the result is prefixed with a “-”. If the relative pressure inside an RI is positive, the result is prefixed with a “+”.

NOTE 2 The pressure drop across a filter is reported without any prefix.

6.2 Equipment

6.2.1 Pressure gauge, calibrated in the appropriate range relevant to the performance standard.

6.2.2 Flowmeter(s), calibrated in the appropriate volumetric flow rate range and corrected for the ambient temperature and ambient atmospheric pressure during use.

6.2.3 Ambient temperature and ambient atmospheric pressure measuring equipment.

6.2.4 Regulated blower/compressed air source and/or a variable suction device.

6.2.5 Support for the device (e.g. filter holder or relevant RPD headform).

NOTE 1 The filter holder can be provided by the RPD manufacturer.

NOTE 2 It is important that the support for the device does not reduce the effective working area of the filter or RPD.

6.3 Procedure

6.3.1 Procedure for filters for respiratory protective devices

Air is drawn through the filter and filter holder system (see [Figure 1](#)). The pressure drop between ambient and a pressure port fitted at a suitable point between the filter holder and the connection to the suction device is measured.

Ensure that the filter is in the state designated by the performance standard.

Mount the filter in a leak-tight manner as indicated in [Figure 1](#). Draw the appropriate airflow, as specified in the performance standard, through the filter holder system. Measure and record the pressure drop, Δp_S .

Remove the filter. Draw the same airflow through the filter holder system. Measure and record the pressure drop, Δp_H .

Calculate and report the breathing resistance of the filter, Δp_F , according to [Formula \(1\)](#) at the specified flow rate.

$$\Delta p_F = \Delta p_S - \Delta p_H \quad (1)$$




Figure 1 — Typical arrangement for measurement of breathing resistance of filters

6.3.2 Procedure for respiratory interfaces

6.3.2.1 General

Air is drawn through the RI and support system. The support system is an RPD headform and trachea tube assembly specified in ISO 16900-5. The pressure drop between ambient and the pressure port in the trachea tube assembly is measured.

Ensure that the RI is in the state designated by the performance standard.

6.3.2.2 Inhalation resistance

Connect a suitable suction device to the trachea tube assembly.

Mount the RI to the support system in a leak-tight manner as it would be worn by a wearer while standing upright. Draw the appropriate airflow, as specified in the performance standard, through the RI and support system. Measure and record the pressure drop, Δp_{S-in} .

Remove the RI. Draw the same airflow through the support system. Measure and record the pressure drop, Δp_{H-in} .

Calculate and report the inhalation resistance of the RI, Δp_{RI-in} , according to [Formula \(2\)](#) at the specified flow rate.

$$\Delta p_{RI-in} = \Delta p_{S-in} - \Delta p_{H-in} \quad (2)$$

If an inhalation valve is fitted, measure and record the inhalation resistance with the support, as described above, in each of the following four additional orientations:

- a) facing vertically, upwards;
- b) facing vertically, downwards;
- c) lying on the left side;
- d) lying on the right side.

6.3.2.3 Exhalation resistance

Connect a suitable blower to the trachea tube assembly.

Mount the RI to the support system in a leak-tight manner as it would be worn by a wearer while standing upright. Blow the appropriate airflow, as specified in the performance standard, through the RI and support system. Measure and record the pressure drop, Δp_{S-ex} .

Remove the RI. Blow the same airflow through the support system. Measure and record the pressure drop, Δp_{H-ex} .

Calculate and report the exhalation resistance of the RI, Δp_{RI-ex} , according to [Formula \(3\)](#) at the specified flow rate.

$$\Delta p_{RI-ex} = \Delta p_{S-ex} - \Delta p_{H-ex} \quad (3)$$

If an exhalation valve is fitted, measure and record the exhalation resistance with the support system, as described above, in each of the following four additional orientations:

- a) facing vertically, upwards;
- b) facing vertically, downwards;
- c) lying on the left side;

d) lying on the right side.

7 Method 2: Dynamic breathing resistance

7.1 General

The RPD is mounted on a support and connected to a breathing machine operating at a specified breathing frequency, waveform and tidal volume.

NOTE The convention of reporting breathing resistance is that if, during the inhalation resistance test, the pressure inside the RI relative to atmosphere is negative, the result is prefixed with a “-”. If the relative pressure inside the RI is positive, the result is prefixed with a “+”.

7.2 Equipment

7.2.1 Breathing machine, set to minute volume specified in the performance standard.

7.2.2 Support for the RPD, as specified in ISO 16900-5 or ISO 16900-10, as appropriate.

7.2.3 Pressure gauge and recording device, calibrated in the appropriate range to the performance standard. The response time of this pressure gauge, including the recording device, shall be less than 30 ms for a response of 10 % to 90 % of the full scale deflection of the range used.

7.3 Procedure for complete RPD

7.3.1 General

Connect the headform to a suitable breathing machine and connect the pressure gauge to the pressure sampling point.

7.3.2 Breathing resistance measurement

Measure the peak inhalation and exhalation pressure offsets of the support system, Δp_{H-in} and Δp_{H-ex} , at the relevant breathing minute ventilation.

Ensure that the RPD is in the state designated by the performance standard.

Fit the RPD on the appropriate headform and torso.

Measure the peak inhalation and exhalation pressure of the RPD and support system, Δp_{S-in} and Δp_{S-ex} , at the relevant breathing minute ventilation.

Calculate and report the peak inhalation or exhalation resistance of the RPD (Δp_{RPD-in} or Δp_{RPD-ex}) according to [Formula \(4\)](#) or [Formula \(5\)](#) at the specified flow rate.

$$\Delta p_{RPD-in} = \Delta p_{S-in} - \Delta p_{H-in} \quad (4)$$

$$\Delta p_{RPD-ex} = \Delta p_{S-ex} - \Delta p_{H-ex} \quad (5)$$

8 Test report

The test report shall include information regarding those parameters specified in [Clause 4](#), together with the measured breathing resistances and pressure drops. The report shall include a note stating all flow rates have been corrected to 1 013 hPa and 20 °C.


Annex A (normative)

Application of uncertainty of measurement

A.1 Determination of compliance

In order to determine compliance or otherwise of the measurement made in accordance with this document, the following protocol shall be applied.

If the test result plus or minus the uncertainty of measurement, U , falls completely inside or outside of the specification zone for the particular test given in this document, then the result shall be deemed to be a straightforward pass or fail (see [Figures A.1](#) and [A.2](#)).



Key

- 1 lower specification limit
- 2 specification zone
- 3 upper specification limit
- 4 uncertainty of measurement, U
- 5 measured value

Figure A.1 — Result pass