



Published 1983-08-01

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Enclosed-scale general purpose thermometers AMENDMENT 1

Amendment 1 to International Standard ISO 1771-1981 was developed by Technical Committee ISO/TC 48, *Laboratory glassware and related apparatus*. It was submitted directly to the ISO Council, in accordance with clause 6.11.2, part 1 of the Directives for the technical work of ISO.

Page 2

Clause 11 :

Add "11.1" as a sub-clause number at the beginning of the first sentence.

Delete existing sub-clause 11 f) and replace the text with :

"11.2 In addition to 11.1, if the purchaser requires the thermometer to bear an identifying number, the following inscription shall also be durably and legibly marked on the thermometer :

manufacturer's identification or serial number, the last two digits of which should indicate the year of manufacture."

UDC 536.513 : 542.2

Ref. No. : ISO 1771-1981/A1-1983 (E)

Descriptors : laboratory equipment, laboratory glassware, temperature measuring instruments, thermometers, graduations, accuracy, dimensions, specifications.

© International Organization for Standardization, 1983 •

Printed in Switzerland

Price group A