
**Dentistry — Analysis of fluoride
concentration in aqueous solutions by
use of fluoride ion-selective electrode**

*Médecine bucco-dentaire — Analyse de la concentration en fluorure
dans les solutions aqueuses en utilisant une électrode sélective des
ions fluorures*

STANDARDSISO.COM : Click to view the full PDF of ISO 19448:2018



STANDARDSISO.COM : Click to view the full PDF of ISO 19448:2018



COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	1
5 Reagents and materials	2
6 Apparatus	2
7 Preparation	3
7.1 TISAB solution	3
7.2 TISAB blank solution	3
7.3 Fluoride standard solutions	3
7.4 Calibration curve	4
8 Procedure	5
8.1 Direct analysis technique	5
8.2 Standard addition technique	6
8.3 Dilution analysis technique	7
8.4 Multiple sample addition technique	7
8.5 Internal standards technique	7
9 Technical notes about the analytical techniques	7
10 Calculations	8
10.1 Direct analysis technique	8
10.2 Standard additions technique	8
10.3 Dilution analysis technique	9
10.4 Multiple sample addition technique	9
10.5 Internal standards technique	9
11 Test report	9
12 Trouble shooting	9
Bibliography	11

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 7, *Oral care products*.

Introduction

Fluoride is used in dentistry primarily for caries prevention. Fluoride is the active ingredient for caries prevention in many dental products such as dentifrices, gels, oral rinses, and fluoride releasing varnishes. The fluoride concentration in these products is an integral part of other standards where the product is intended to be analysed for fluoride content. This document provides methods based on the use of fluoride ion-selective electrode technology for the analysis of the total fluoride content in aqueous samples. The methods describe uses for fluoride ion-selective electrode technology and anticipate that the sample-specific preparations such as digestion, distillation, etc. have been described in the standards specific to product types.

STANDARDSISO.COM : Click to view the full PDF of ISO 19448:2018

STANDARDSISO.COM : Click to view the full PDF of ISO 19448:2018

Dentistry — Analysis of fluoride concentration in aqueous solutions by use of fluoride ion-selective electrode

1 Scope

This document specifies test methods for the quantification of fluoride concentrations in dental products including dentifrices, gels, oral rinses, fluoride releasing varnishes, and other fluoride containing products. The methods are based on fluoride ion-selective electrode technology for the analysis of fluoride in aqueous samples derived from dental products.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 78-2, *Chemistry — Layouts for standards — Part 2: Methods of chemical analysis*

ISO 835, *Laboratory glassware — Graduated pipettes*

ISO 1942, *Dentistry — Vocabulary*

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 78-2 and ISO 1942 apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

4 Principle

This test method covers the determination of fluoride concentration in aqueous solutions derived from dental products. The following test methods are presupposed to use fluoride ion-selective electrodes (F-ISE).

The sensing element of the fluoride ion-selective electrode consists of a single crystal of europium (II)-doped lanthanum fluoride, LaF_3 . When the sensing element is in contact with a solution containing fluoride ions, an electrode potential develops across the sensing element. This potential, which depends on the activity of free fluoride ion in solution $\{\text{F}^-\}$, is measured against a constant reference electrode potential with a digital pH/mV meter or ISE (concentration) potentiometer. A potential (E) is developed which is proportional to the logarithm of the activity of the fluoride ion $\{\text{F}^-\}$ in solution following the Nernst Equation.

$$E = E_0 + S \cdot \log \{\text{F}^-\} \quad (1)$$

where

E is the measured electrode potential;

E_0 is the reference potential (a constant);

$\{F^-\}$ is the fluoride ion activity level in solution;

S is the electrode slope (58,2 mV per decade change of $\{F^-\}$ at 20 °C).

The level of fluoride ion, $\{F^-\}$, is the activity or “effective concentration” of free fluoride ion in solution. The fluoride ion activity is related to free fluoride ion concentration, $[F]$, by the activity coefficient, γ . The activity coefficient is related to the ionic strength of the solution. For these measurements, the ionic strength of the samples and standards are held constant by the use of total ionic strength adjusting buffer solution (TISAB), which also chelates fluoride-binding cations and establishes the solution pH at the optimal level for the fluoride ion-selective electrode sensitivity. When using TISAB to set the ionic strength of the solution, the electric potential is proportional to the concentration of the fluoride ion in solution. TISAB may come in several variations. Which one to use is specific to the sample preparation procedure. One has to be aware that other interferences may exist and it is up to the user and sample preparation to properly identify them.

A calibration curve is drawn, plotting measured potential versus the log of the fluoride ion concentration of known standard solutions. Then, the potential for the unknown sample is measured and the concentration of F^- in the sample can be calculated from the calibration curve.

The sample will have been pretreated as specified for the particular sample type, and the final composition is an aqueous sample of at least 1 ml volume. Samples that contain more than 0,01 mol/l F (190 mg/l F) shall be diluted to concentrations less than 0,01 mol/l F with deionized water.

5 Reagents and materials

5.1 Chemicals to prepare total ionic strength adjusting buffer (TISAB) solution.

5.1.1 Cyclohexylene dinitrilotetraacetic acid (CDTA) or 1,2-diaminocyclohexane-N,N,N',N'-tetraacetic acid.

5.1.2 Acetic acid (CH_3COOH) reagent grade.

5.1.3 Sodium chloride (NaCl) salt.

5.1.4 Sodium hydroxide (NaOH), at 5 mol/l.

5.1.5 Deionized water, in accordance with ISO 3696, grade 2.

5.2 Chemicals to prepare sodium fluoride standards.

5.2.1 Sodium fluoride (NaF), previously dried at least 4 h at 150°C.

5.2.2 Deionized water, in accordance with ISO 3696, grade 2.

6 Apparatus

6.1 Analytical balance, accurate to 0,000 1 g.

6.2 Magnetic stir bar.

6.3 Magnetic stir plate.

6.4 Pipette, of nominal capacity 5,0 ml to smallest scale division: 0,05 ml; 1,0 ml to 0,01 ml; and 0,10 ml to 0,01 ml, in accordance with ISO 835, Class A.

6.5 Fluoride ion-selective electrode (F-ISE) with separate reference electrode, or combination F-ISE/reference electrode pair.

6.6 Plastic vial (or beaker or small container), 20 or more ml capacity.

NOTE Fluoride does not react with plastic but does react with glass.

6.7 pH/mV electrometer (pH meter), with a sensitivity of $\pm 0,1$ mV.

7 Preparation

7.1 TISAB solution

7.1.1 In a 1 l beaker with about 500 ml deionized H₂O, dissolve 57 ml CH₃COOH, 58 g NaCl, and 4 g CDTA or 1,2-diaminocyclohexane-N,N,N',N'-tetraacetic acid.

7.1.2 Cool to room temperature and adjust pH to 5,0–5,5 with 5 mol/l NaOH.

7.1.3 Cool to room temperature, transfer to 1 l volumetric flask, and dilute to volume with deionized H₂O.

NOTE Preparation of TISAB solution is described in ASTM D1179–16 [\[1\]](#), 18.1.

TISAB is the buffering system that is used to prepare the sample for analysis via the fluoride ion-selective electrode (F-ISE). The F-ISE has several limitations that can limit its use. It is sensitive only to ionic fluoride in solution, and cannot respond to complexes of fluoride such as calcium-fluoride. Multivalent cations such as calcium, magnesium, iron, aluminium, etc., can form fluoride complexes in solution which reduces the ionic fluoride concentration. Additionally, the F-ISE is sensitive to variations in sample pH and to the ionic strength of the sample. For accurate analyses, the samples and the standards must have the same pH and ionic strength. The fluoride complexes must be disrupted such that all the fluoride in the sample and standard is in its ionic form. To address these limitations, a total ionic strength adjusting buffering system (TISAB) that adjusts the sample pH and ionic strength, and also contains chelators to bind the cations that can form fluoride complexes, is used.

TISAB II® (Orion Research) is available from Thermo Fisher Scientific, Beverly, MA 01915 USA¹⁾.

7.2 TISAB blank solution

Mix equal volumes of TISAB and deionized water for the blank solution.

7.3 Fluoride standard solutions

7.3.1 1×10^{-2} mol/l NaF standard solution: dissolve 209,9 mg accurately weighed NaF into 200 ml deionized H₂O, then quantitatively transfer to a 500 ml volumetric flask, and dilute to volume with deionized H₂O.

1) This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of the product named. Equivalent products may be used if they can be shown to lead to the same results.

7.3.2 1×10^{-3} mol/l NaF standard solution: pipette 25,0 ml of 10^{-2} mol/l NaF standard solution to 250 ml volumetric flask. Dilute to volume with deionized H₂O.

7.3.3 1×10^{-4} mol/l NaF standard solution: pipette 25,0 ml 10^{-3} mol/l solution to 250 ml volumetric flask. Dilute to volume with deionized H₂O.

7.3.4 1×10^{-5} mol/l NaF standard solution: pipette 25,0 ml 10^{-4} mol/l solution to 250 ml volumetric flask. Dilute to volume with deionized H₂O.

7.4 Calibration curve

7.4.1 Pipette 5,0 ml of each standard solution into a plastic vial (or beaker or small container).

7.4.2 Add 5,0 ml of TISAB buffer solution to each plastic vial.

7.4.3 Add magnetic stir bar and mix thoroughly.

7.4.4 Insert the F-ISE and reference electrode(s) (6.5) into the standard solution under moderate stirring.

7.4.5 Record the millivolt reading to 0,1 mV at the steady potential difference with the mV electrometer (6.7).

The F-ISE is considered stable when the drift in millivolts is less than $\pm 0,5$ mV/min. The time required to attain stable millivolt response is typically longer for more dilute samples. It is not unusual for a 1×10^{-5} mol/l solution to take 2 min to become stable; however, it should not take longer than 5 min to reach a stable response.

NOTE When measuring the millivolt response of the standard solutions, determine the response for each standard solution starting with the most dilute and proceed to the most concentrated. These steps prevent contamination of dilute standards from carry over from the concentrated standards, and shortens the response time required to reach stable values.


7.4.6 Remove the electrode(s) from the solution, rinse and blot dry and replace into the same solution.

7.4.7 Conduct at least two measurements for millivolt readings until the difference between the two readings is less than 0,2 mV.

7.4.8 Repeat 7.4.1 to 7.4.7 for each of the other standard solutions.

7.4.9 Construct a calibration curve of electrode potential (millivolts) measured versus the $\log_{10}$ of the fluoride ion concentration of the standard. See Figure 1 where the calibration curve is a straight line

with a coefficient of determination (r^2) of 0,999 90, with a slope of -58,2 mV/log [F] concentration and intercept of -569,1 mV.



Key

X log[F] mol/l
Y potential (mV)

Figure 1 — Fluoride ion-selective electrode calibration curve

The calibration curve shall be linear, with a coefficient of determination (r^2) greater than 0,95, and a slope between -56 mV and -60 mV per decade difference in fluoride concentration.

8 Procedure

8.1 Direct analysis technique

8.1.1 Dilute each sample solution with equal volume of TISAB (50:50) into a separate plastic vial (or beaker or small container) and mix thoroughly.

8.1.2 Insert the F-ISE and reference electrode(s) into the liquid in the plastic vial containing the sample solution while stirring.

8.1.3 Record the millivolt reading to 0,1 mV at the steady potential difference with the mV electrometer.

8.1.4 Remove the electrode(s) from the sample solution, rinse with deionized water and gently blot dry the electrode(s).

8.1.5 Conduct at least two measurements for millivolt readings until the difference between the two readings is less than 0,2 mV.

8.1.6 Repeat [8.1.1](#) to [8.1.5](#) for each of the other sample solutions.

8.1.7 If the apparent fluoride concentration is greater than 1×10^{-2} mol/l, dilute the sample with TISAB blank solution by a factor of 100 and reanalyse the sample.

8.2 Standard addition technique

This technique is suitable for samples that may have bound fluoride ion, or when the unknown sample is contaminated with other substances that alter the assay.

8.2.1 Prepare a standard addition series of four sample solutions of equal and known volume (4,0 ml) in separate plastic vials (or beakers or small containers).

8.2.2 To one sample, add 1,0 ml deionized water.

8.2.3 Add 1,0 ml of NaF standard solution of differing concentrations, such as 1×10^{-3} mol/l, 5×10^{-3} mol/l, or 1×10^{-2} mol/l, to each of the remaining sample solutions.

8.2.4 Dilute each with 5,0 ml of TISAB (50:50) and mix thoroughly.

8.2.5 Insert the F-ISE/reference electrode(s) into the liquid in the first plastic vial containing the sample solution while stirring.

8.2.6 Record the millivolt reading to 0,1 mV at the steady potential difference with the mV electrometer.

8.2.7 Remove the electrode(s) from the sample solution, rinse with deionized water and gently blot dry the electrode(s).

8.2.8 Conduct at least two measurements for millivolt readings until the difference between the two readings is less than 0,2 mV.

8.2.9 If the apparent fluoride concentration of the first sample in the standard addition series is greater than 1×10^{-2} mol/l, dilute the sample with TISAB blank solution by a factor of 100 and return to [8.2.1](#).

8.2.10 Remove the F-ISE/reference electrode(s) from the sample and rinse with deionized water and gently blot dry the electrode(s).

8.2.11 Insert the F-ISE/reference electrode(s) into the liquid in the second sample solution of the series.

8.2.12 Record the millivolt reading to 0,1 mV at the steady potential difference with the mV electrometer.

8.2.13 Remove the electrode(s) from the sample solution, rinse with deionized water and gently blot dry the electrode(s).

8.2.14 Conduct at least two measurements for millivolt readings until the difference between the two readings is less than 0,2 mV.

8.2.15 Repeat [8.2.11](#) to [8.2.14](#) for the third and fourth sample solutions in the series.

8.3 Dilution analysis technique

This technique is suitable for samples of small volume.

8.3.1 Place 4,950 ml TISAB blank solution in a plastic vial with stir bar.

8.3.2 Record the millivolt reading of the TISAB blank solution to within ± 1 mV at the steady potential difference with the mV electrometer.

8.3.3 Add 50 μ l sample to the 4,950 ml blank while stirring.

8.3.4 Determine and record a stable mV potential within $\pm 0,1$ mV.

8.3.5 Repeat [8.3.1](#) to [8.3.4](#) for each of the other sample solutions.

8.4 Multiple sample addition technique

This technique is suitable for samples of complex matrix, and only small amount of samples are available or the samples are of high concentration.

8.4.1 Place 4,950 ml TISAB blank solution in a plastic vial with stir bar.

8.4.2 Record the millivolt reading of the TISAB blank solution to within $\pm 0,1$ mV at the steady potential difference with the mV electrometer.

8.4.3 With the F-ISE/reference electrode(s) in the solution, add 50 μ l sample to the 4,950 ml blank while stirring.

8.4.4 Determine and record a stable mV potential within $\pm 0,1$ mV.

8.4.5 Add an additional 50 μ l sample to the solution while stirring.

8.4.6 Determine and record a stable mV potential within $\pm 0,1$ mV.

8.4.7 Repeat the procedures described in [8.4.5](#) and [8.4.6](#) twice more.

8.5 Internal standards technique

This technique is suitable for samples of complex matrix.

8.5.1 Prepare a set of internal standards where these standards receive the same steps for preparation as the samples. Follow the steps in [8.1](#) to determine the potential of the internal standards and samples.

9 Technical notes about the analytical techniques

9.1 Clean and carefully blot dry the electrode(s) between each measurement.

9.2 Store electrode(s) according to manufactures recommendations.

9.3 Measurements shall be made at a stable temperature.

9.4 Samples and standards shall be at the same temperature.

10 Calculations


10.1 Direct analysis technique

10.1.1 Use the calibration curve of standard solutions (7.4) to determine the apparent fluoride concentration ([F_{dil}]) in the sample solution in mol/l.

10.2 Standard additions technique

10.2.1 Use the calibration curve of standard solutions (7.4) to determine the fluoride ion concentration for the sample solution in the standard addition series in mol/l.

10.2.2 Construct a plot of added fluoride (mol/l) for each sample versus the measured fluoride ion concentration for those samples. The line shall be linear with a coefficient of determination (r²) greater than 0,95. The X-intercept of the linear regression of this plot is the fluoride ion concentration of the unknown sample in mol/l. Alternatively, the fluoride ion concentration of the unknown can be calculated from the linear regression by dividing the Y-intercept by the slope. See Figure 2. The absolute value of the x intercept is the fluoride concentration in the sample. In this example, the standard addition line is linear with the equation $y = 0,467\ 1\ x - 0,006\ 8$ and a coefficient of determination (r²) of 0,998 3. The fluoride concentration of the slurry = 0,014 5 mol/l (276,3 µg/l) and because the dilution factor was 4, to make the slurry, the fluoride concentration of the NaF dentifrice is 4 x the slurry concentration = 0,058 2 mol/l (1 105 µg/l).



NOTE Technique for a NaF dentifrice slurry, that was not hydrolyzed or centrifuged, of 1 part dentifrice and 3 parts deionized water. The fluoride concentration of the slurry = 0,014 5 mol/l (276,3 µg/l) and because the dilution factor was 4 to make the slurry the fluoride concentration of the NaF dentifrice is 4 × the slurry concentration = 0,058 2 mol/l (1 105 µg/l).

Figure 2 — Fluoride analysis via standard additions