
**Fine ceramics (advanced ceramics,
advanced technical ceramics) —
Thermophysical properties of ceramic
composites — Determination of
specific heat capacity**

*Céramiques techniques — Propriétés thermophysiques des composites
céramiques — Détermination de la capacité thermique spécifique*



STANDARDSISO.COM : Click to view the full PDF of ISO 19628:2017



COPYRIGHT PROTECTED DOCUMENT

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Method A – drop calorimetry	2
4.1 Principle	2
4.2 Apparatus	2
4.3 Standard reference materials	2
4.4 Test specimens	2
4.5 Calibration of calorimeter	3
4.5.1 General	3
4.5.2 Electrical calibration	3
4.5.3 Calibration using standard reference material	3
4.6 Test procedures	3
4.6.1 Test without a crucible	3
4.6.2 Test with a crucible	4
4.6.3 Description of test	4
4.7 Calculations	5
4.7.1 General	5
4.7.2 Determination of the calorimetric calibration factor	5
4.7.3 Determination of mean specific heat capacity $\bar{C}_p$	5
5 Method B – differential scanning calorimetry	6
5.1 Principle	6
5.1.1 General	6
5.1.2 Stepwise heating method	6
5.1.3 Continuous heating method	7
5.2 Apparatus	8
5.3 Standard reference materials, SRM	8
5.4 Test specimens	8
5.5 Temperature calibration	8
5.6 Test procedure for the determination of C_p	8
5.6.1 General	8
5.6.2 Method 1: Measurements requiring the knowledge of the K factor	9
5.6.3 Method 2: measurements requiring the use of a reference standard material (SRM)	11
5.7 Calculation of results	14
5.7.1 Method requiring the knowledge of the K factor	14
5.7.2 Method using an SRM	16
6 Test report	17
Annex A (normative) Drop calorimetry – determination of the calibration factor using standard reference material	18
Annex B (informative) Standard reference material	20
Annex C (informative) Materials for calorimeter calibrations	25
Bibliography	26

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html

This document was prepared by Technical Committee ISO/TC 206, *Fine ceramics*.

Fine ceramics (advanced ceramics, advanced technical ceramics) — Thermophysical properties of ceramic composites — Determination of specific heat capacity

1 Scope

This document describes two methods for the determination of the specific heat capacity of ceramic matrix composites with continuous reinforcements (1D, 2D, 3D).

Unidirectional (1D), bi-directional (2D) and tridirectional (XD, with $2 < x \leq 3$).

The two methods are:

- method A: drop calorimetry;
- method B: differential scanning calorimetry.

They are applicable from ambient temperature up to a maximum temperature, depending on the method: method A can be used up to 2 250 K, while method B is limited to 1 900 K.

NOTE Method A is limited to the determination of an average value of the specific heat capacity over a given temperature range and can give a larger spread of results.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 19634, *Fine ceramics (advanced ceramics, advanced technical ceramics) — Ceramic composites — Notations and symbols*

IEC 60584-1, *Thermocouples — Part 1: Reference tables*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 19634 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1

specific heat capacity

C_p

amount of heat required to raise the temperature of a mass unit of material by 1 K at constant temperature and pressure

$$C_p = \frac{1}{m} \frac{dQ}{dT}$$

where Q is the heat required for a test-piece of mass m .

3.2

mean specific heat capacity

$\overline{C_p}$

amount of heat required to raise the temperature of a mass unit of a material from temperature T_1 to temperature T_2 at a constant pressure, divided by the temperature increase ($T_2 - T_1$) expressed in K

3.3

representative volume element

RVE

minimum volume which is representative of the material considered

4 Method A – drop calorimetry

4.1 Principle

A test piece is dropped from a conditioning chamber at a constant temperature T_1 to another chamber at a constant temperature T_2 .

The mean specific heat capacity is determined from the measured amount of heat required to maintain the temperature constant in the second chamber. Transfer of the test piece shall be done under conditions as close as possible to adiabatic conditions.

4.2 Apparatus

4.2.1 Drop calorimeter, there are several types of drop calorimeters. They include one (or more) conditioning chambers and measuring chambers, which can be operated under controlled atmosphere and which are all equipped with a temperature control system that allows a temperature stability of less than 1 K.

The conditioning chamber shall have a homogeneous temperature zone size greater than the test specimen size. The measuring chamber shall have a homogeneous temperature zone of a sufficient length to accept several specimens and a sufficient thermal inertia to limit the temperature disturbance, due to the drop.

Heat transfer by radiation during the drop shall be avoided as far as possible.

4.2.2 Balance, with an accuracy of 0,1 mg for test pieces over 10 mg and an accuracy of 0,01 mg for test pieces below 10 mg.

4.2.3 Temperature detectors, thermocouples in accordance with IEC 60584-1 shall be used for the measurement of temperature up to 1 920 K.

For higher temperatures, infrared detectors or any other suitable device may be used.

4.2.4 Data acquisition system, the sampling period during the test shall be less than 0,5 s.

4.3 Standard reference materials

Standard reference materials which can be used for calibration purposes are listed in [Annex B](#).

4.4 Test specimens

The test specimens shall be representative of the material.

This criterion is generally met by test specimens containing the maximum number of representative volume elements compatible with the volume of the crucible. If this number is less than five, several solutions are possible:

- a) the test specimens should have an exact number of representative volume elements;
- b) the material should be cut into specimens; a number of similar test specimens should be tested and an average value determined.

4.5 Calibration of calorimeter

4.5.1 General

Calibration of calorimeters may be done according to two different methods. The first consists of dissipating a known amount of thermal power using a calibrated resistor introduced in the second chamber of the calorimeter. In the second method a reference specimen with known specific heat capacity is dropped according to the procedure described in 4.6.

4.5.2 Electrical calibration

The calibration factor is the ratio of a known amount of thermal power dissipated in the resistor to the steady-state calorimetric output signal, and is measured at temperature T_2 .

NOTE 1 The method using power dissipation in a resistor is limited to 1 350 K.

NOTE 2 This method can only be used if the sensitivity of the calorimeter is not affected by the filling of the measuring chamber.

4.5.3 Calibration using standard reference material

This calibration is called “drop calibration”. A specimen made of a standard reference material with a known specific heat capacity is dropped according to the test procedures described in 4.6. (See [Annex B](#) for standard reference material.) The calibration factor is determined according to [Annex A](#).

4.6 Test procedures

4.6.1 Test without a crucible

4.6.1.1 Test with drop calibration

The test without a crucible and with drop calibration is done in the following order:

R, T, R, T, R, T, R

where

R is the test of standard reference material;

T is the test of test specimen.

Carry out each test as described in 4.6.3.

4.6.1.2 Test with electrical calibration

The test without a crucible and with calibration using power dissipation in a resistor is done in the following order:

— calibration of calorimeter;

- test on three test specimens.

Carry out each test as described in [4.6.3](#).

NOTE The avoidance of interaction between the test specimen and the calorimetric conditioning and measuring chambers can require the use of a sealed crucible.

4.6.2 Test with a crucible

4.6.2.1 General

The mass of all empty crucibles used for the test shall not differ by more than 5 %.

4.6.2.2 Test with drop calibration

The test with a crucible and with drop calibration is carried out in the following order:

C, C + R, C + T, C, C + R, C + T, C, C + R, C + T, C

where

C is the test with the empty crucible;

C + R is the test of crucible plus standard reference material;

C + T is the test of crucible plus test specimen.

Carry out each test as described in [4.6.3](#).

4.6.2.3 Test with electrical calibration

The test with a crucible and with calibration using power dissipation in a resistor is done in the following order:

- calibration of calorimeter;
- carry out the following sequence:

C, C + T, C, C + T, C, C + T, C

where

C is the test with the empty crucible;

C + T is the test with crucible plus test specimen.

Carry out each test as described in [4.6.3](#).

4.6.3 Description of test

The test piece (test specimen, standard material or empty crucible) and reference material shall be dried at $(110 \pm 5) ^\circ\text{C}$ until the difference in weight of two successive weighings is lower than 0,2 mg:

- measure the mass when a crucible is not used with an accuracy of $\pm 0,1$ mg or $\pm 0,1$ %, whichever is the smaller;
- when a crucible is used, measure the mass of each assembly dropped (empty crucible, crucible and standard reference material, crucible and test specimen);

- place the test piece (test specimen, standard material or empty crucible) in the conditioning chamber at temperature T_1 and wait for a sufficient period (around 15 min) to reach thermal equilibrium of the test piece with its environment;
- measure T_1 and T_2 ;
- start recording the calorimetric signal before the test piece is dropped;
- drop the test piece;
- stop the recording when the steady-state output signal is reached.

4.7 Calculations

4.7.1 General

The change in heat Q corresponding to the drop of the test piece is related to the area A under the calorimetric output signal by the following equation.

$$Q = K \cdot A$$

where K is the calorimeter calibration factor.

4.7.2 Determination of the calorimetric calibration factor

4.7.2.1 Electrical calibration

See [Annex A](#).

$$K = \frac{\text{heat dissipated}}{\text{area under the calorimetric output signal}} = \frac{H}{A}$$

4.7.2.2 With standard reference material

See [Annex B](#).

4.7.3 Determination of mean specific heat capacity $\overline{C_p}$

The mean specific heat capacity is determined using the following formula:

$$\overline{C_p}(T_1, T_2) = \frac{1}{m_i} \frac{Q_i(T_1, T_2)}{(T_2 - T_1)}$$

where

T_1 is the initial temperature at which test pieces are conditioned;

T_2 is the calorimeter temperature;

$Q_i(T_1, T_2)$ is the heat variation between T_1 and T_2 ;

m_i is the mass of the test piece, determined by weighing;

$\overline{C_p}(T_1, T_2)$ is the mean specific heat capacity between T_1 and T_2 .

The subscript i has a different meaning depending on the type of test piece:

- i = c for an empty crucible;
- i = t for a test piece;
- i = t + c for a test piece and crucible.

without crucible

$$\overline{C}_{pt} = \frac{K \cdot A_t}{m_t (T_2 - T_1)}$$

with crucible

$$\overline{C}_{pt} = \frac{K(A_{c+t} - A_c)}{m_t (T_2 - T_1)}$$

where

- A_t is the value of integration of calorimetric output signal of test specimen;
- A_c is the value of integration of calorimetric output signal of crucible;
- A_{c+t} is the value of integration of calorimetric output signal of test specimen plus crucible.

5 Method B – differential scanning calorimetry

5.1 Principle

5.1.1 General

The method consists in measuring the difference in power needed to raise the temperature of the test specimen in its crucible and of an empty identical crucible using the same heating programme, which may be stepwise heating or continuous heating.

Stepwise heating allows only the determination of the mean specific heat capacity $\overline{C}_p(T_1, T_2)$ over a temperature range (T_1, T_2) , whereas continuous heating allows determination of the specific heat capacity C_p at a given temperature.

5.1.2 Stepwise heating method

The mean specific heat capacity $\overline{C}_p(T_1, T_2)$ is measured in a temperature interval defined by two isothermal levels, T_1 and T_2 . The heat, Q_E , which is necessary to change the temperature from T_1 to T_2 is determined by integrating the thermal power, P_E , with respect to time. The corresponding heat, Q_E , is:

$$Q_E = \int_0^t P_E dt = (m_t \overline{C_p}(T_1, T_2) + C_c + C_o)(T_2 - T_1)$$

where

m_t is the mass of the test specimen;

$\overline{C_p}(T_1, T_2)$ is the mean specific heat capacity of the test specimen;

C_o is the heat capacity of the calorimeter;

C_c is the heat capacity of the crucible.

Another experiment for the determination of the baseline is performed using an identical imposed heating sequence with the empty crucible. The corresponding heat, Q_B , is given by:

$$Q_B = \int_0^t P_B dt = [C_c + C_o](T_2 - T_1)$$

From the above equations, the mean specific heat capacity can be calculated as:

$$\overline{C_p}(T_1 - T_2) = \frac{Q_E - Q_B}{m_t(T_2 - T_1)}$$

5.1.3 Continuous heating method

Temperature is increased linearly versus time at a constant heating rate β . Using the same notation as in 5.1.2, the thermal power P_E supplied to the system at every moment is:

$$K \cdot S_{c+t} = (m_t C_p + C_c + C_o) \beta$$

Another experiment for the determination of the baseline is performed with the empty crucible. The corresponding thermal power is given by

$$K \cdot S_c = (C_c + C_o) \beta$$

The specific heat capacity can be calculated from:

$$C_p = \frac{K(S_{c+t} - S_c)}{m_t \beta}$$

where

K is the calibration factor;

S_c, S_{c+t} are the output signals;

$K \cdot S_c$ and $K \cdot S_{c+t}$ are the thermal powers supplied to the system.

5.2 Apparatus

5.2.1 Differential scanning calorimeter.

5.2.1.1 There are two types of differential scanning calorimeters operating on power compensation and heat flux principles, both designed to operate under adiabatic conditions.

Both comprise two measuring cells housed in a furnace which provides overall system heating. One cell contains the test specimen and its crucible, the other contains an empty crucible only.

5.2.1.2 Power compensation type: each cell has an additional heater to compensate for the temperature variations from the overall heating programme. The power which is supplied to either cell heater to maintain equal temperatures during heating is measured.

5.2.1.3 Heat flux type: power is exchanged between each cell and its respective surrounding during the heating programme. The difference in power exchange between the two cells is measured.

5.2.2 **Balance**, with an accuracy better than 0,1 mg.

5.2.3 **Temperature detectors**, thermocouples in accordance with IEC 60584-1 shall be used for the measurement of temperature.

5.2.4 **Data acquisition system**, the time duration between two successive measurements shall be less than 0,5 s.

5.3 Standard reference materials, SRM

Standard reference materials shall be used for calibration. An example is given in [Annex B](#).

5.4 Test specimens

The test specimens shall be representative of the material.

This criterion is generally met by test specimens containing the maximum number of representative volume elements compatible with the volume of the crucible. If this number is less than five, several solutions are possible:

- a) the test specimens should have an exact number of representative volume elements;
- b) the material should be cut into pieces, and a number of similar test pieces should be tested and an average value determined.

5.5 Temperature calibration

A temperature calibration curve for the furnace using the same heating rate as for the determination of the specific heat capacity is established by using the melting points of standard reference materials (see, for example, [Annex C](#)).

Thermocouples shall be calibrated in accordance with IEC 60584-1.

5.6 Test procedure for the determination of C_p

5.6.1 General

Depending on the necessity or not of using a calibration factor K for the calorimeter, two methods can be used:

Method 1: measurements requiring the knowledge of the K factor; in this case, care shall be taken in order to ensure that the calibration is valid for all the measurements to be taken.

NOTE Generally, this can be done by running a test using a test specimen with well-known properties.

Method 2: measurements requiring the use of a reference standard material during a series of tests.

5.6.2 Method 1: Measurements requiring the knowledge of the K factor

5.6.2.1 Determination of the K factor

The calibration factor K is obtained by electrical calibration. It is determined from the ratio of a known amount of power dissipated in a resistor to the steady-state calorimetric output signal.

5.6.2.2 Measurements with the specimen for the determination of the C_p

5.6.2.2.1 General

A series of measurements shall always be referenced to a baseline measurement performed under experimental conditions identical to the other measurements in the series. The type of crucible used depends on the type of test specimen and on the temperature range, and shall be the same for the series of measurements. The mass of all empty crucibles used in the series shall not differ by more than 5 %.

5.6.2.2.2 Test sequence for the stepwise heating method (see Figure 1)

5.6.2.2.2.1 Generation of the baseline

See Figure 1.

- a) weigh the two empty crucibles to the nearest 0,1 mg;
- b) place the two crucibles in the calorimeter;
- c) set the calorimeter heating rate, initial and final temperature, and cooling rate;


NOTE Generally, the heating rate is in the range 1 K/min to 20 K/min.

- d) heat to an initial temperature, and wait for the temperature to be stabilized at the initial temperature;
- e) heat at a constant rate to final temperature of the first step while recording the calorimeter output signal, until the final temperature is reached and stabilized in order to obtain a baseline;
- f) repeat c) to e) for the number of steps required;
- g) cool down to initial temperature;
- h) remove the crucibles from the measurement cell.

5.6.2.2.2.2 Measurements using a test specimen

See Figure 1.

Weigh the test specimen and place it in the crucible to be located in the measurement cell. Repeat operations b) to h) from [5.6.2.2.2.1](#) on generation of the baseline. Repeat this procedure for a minimum of three test specimens.



Key

- 1 output signal
- 2 temperature
- 3 time
- 4 time

Figure 1 — Stepwise heating method (with calibration factor)

5.6.2.2.3 Test sequence for the continuous heating method

5.6.2.2.3.1 Generation of the baseline


See Figure 2.

- a) weigh the two empty crucibles to the nearest 0,1 mg;
 - b) place the two crucibles in the calorimeter;
 - c) set the calorimeter heating rate, initial and final temperature, and cooling rate;
- NOTE Generally, the heating rate is in the range 1 K/min to 20 K/min.
- d) heat to an initial temperature, and wait for the temperature to be stabilized at the initial temperature;
 - e) heat at a constant rate to final temperature of the first step, while recording the calorimeter output signal, until the final temperature is reached and stabilized in order to obtain a baseline;
 - f) cool down to initial temperature;
 - g) remove the crucibles from the measurement cell.

5.6.2.2.3.2 Measurements using a test specimen

See Figure 2.

Weigh the test specimen and place it in the crucible to be located in the measurement cell. Repeat operations b to g of the above paragraph on generation of the baseline. Repeat this procedure for a minimum of three test specimens.



Key

1	output signal	4	time
2	temperature	5	test piece
3	time	6	baseline

Figure 2 — Continuous heating method (with calibration factor)

5.6.3 Method 2: measurements requiring the use of a reference standard material (SRM)

5.6.3.1 General

The two methods described in [5.6.3.2](#) and [5.6.3.3](#) each require:

- measurements with two empty crucibles for the generation of the baseline;
- measurements with one empty crucible and one crucible with the SRM;
- measurements with one empty crucible and one crucible with the test specimen.

5.6.3.2 Test sequence for the stepwise heating method

5.6.3.2.1 Generation of the baseline

See Figure 3.

- a) weigh the two empty crucibles to the nearest 0,1 mg;
- b) place the two crucibles in the calorimeter;
- c) set the calorimeter heating rate, initial and final temperature, and cooling rate;
NOTE Generally, the heating rate is in the range 1 K/min to 20 K/min.
- d) heat to an initial temperature, and wait for the temperature to be stabilized at the initial temperature;
- e) heat at a constant rate to final temperature of the first step, while recording the calorimeter output signal, until the final temperature is reached and stabilized in order to obtain a baseline;
- f) repeat c) to e) for the number of steps required;
- g) cool down to initial temperature;
- h) remove the crucibles from the measurement cell.

5.6.3.2.2 Measurement with a test specimen or with a SRM

See Figure 3.

Weigh the test specimen or the SRM and place it in the crucible to be located in the measurement cell. Repeat operations b) to h) from [5.6.3.2.1](#) on generation of the baseline. Repeat this procedure for a minimum of three test specimens.




Figure 3 — Stepwise heating method (with standard reference material)

5.6.3.3 Test sequence for continuous heating method

5.6.3.3.1 Generation of the baseline


See Figure 4.

- a) weigh the two empty crucibles to the nearest 0,1 mg;
 - b) place the two crucibles in the calorimeter;
 - c) set the calorimeter heating rate, initial and final temperature, and cooling rate;
- NOTE Generally, the heating rate is in the range 1 K/min to 20 K/min.
- d) heat to an initial temperature, and wait for the temperature to be stabilized at the initial temperature;
 - e) heat at a constant rate to final temperature of the first step, while recording the calorimeter output signal, until the final temperature is reached and stabilized in order to obtain a baseline;
 - f) cool down to initial temperature;
 - g) remove the crucibles from the measurement cell.

5.6.3.3.2 Measurements using a test specimen or an SRM

See Figure 4.

Weigh the test specimen or the SRM and place it in the crucible to be located in the measurement cell. Repeat operations b) to g) from 5.6.3.3.1 on generation of the baseline. Repeat this procedure for a minimum of three test specimens.



Key	
1	output signal
2	temperature
3	time
4	time
5	test piece
6	baseline
7	reference

Figure 4 — Continuous heating method (with standard reference material)

5.7 Calculation of results

5.7.1 Method requiring the knowledge of the *K* factor

5.7.1.1 Stepwise heating method

See Figure 1.

The use of a computer with adapted software greatly simplifies data acquisition and treatment, and is recommended.

For the considered temperature interval, the shaded areas, *A*, are the integrals of the output signal, *s*, with respect to time.

For the baseline run with an empty crucible, one obtains:

$$\bar{K} \cdot A_c = (C_c + C_o)(T_2 - T_1)$$

For crucible plus test specimen:

$$\bar{K} \cdot A_{c+t} = (m_t \bar{C}_{pt} + C_c + C_o)(T_2 - T_1)$$

From these two equations one obtains:

$$\bar{C}_{pt}(T_1, T_2) = \frac{\bar{K}(A_{c+t} - A_c)}{m_t(T_2 - T_1)}$$

where

$\bar{K}$ is the mean calorimeter calibration factor for the temperature range T_1 to T_2 , in watts per volt ($W \cdot V^{-1}$);

C_o is the heat capacity of the calorimeter, in joules per kelvin ($J \cdot kg^{-1} \cdot K^{-1}$);

C_c is the heat capacity of the crucible, in joules per kelvin ($J \cdot kg^{-1} \cdot K^{-1}$);

$\bar{C}_{pt}$ is the mean specific heat capacity of the test specimen, in joules per kilogram and per kelvin ($J \cdot kg^{-1} \cdot K^{-1}$);

m_t is the mass of the test specimen, in kilograms (kg).

5.7.1.2 Continuous heating method

See Figure 2.

Using the same notations as in 5.7.1.1 plus C_{pt} as the specific heat capacity of the test specimen, one obtains for each temperature:

— for the empty crucible:

$$KS_c = (C_c + C_o)\beta$$

— for crucible plus test specimen:

$$KS_{c+t} = (m_t C_{pt} + C_c + C_o)\beta$$

The specific heat capacity of the test specimen for each temperature, T , is then obtained from:

$$C_{pt} = \frac{K(S_{c+t} - S_c)}{m_t \cdot \beta}$$

where

β is the constant heating rate in ($K \cdot s^{-1}$);

K is the calibration factor at a temperature, T ;

S_c, S_{c+t} are the output signals of the calorimeter in volts (V).

5.7.2 Method using an SRM

5.7.2.1 Stepwise heating method

See Figure 3.

The use of a computer with adapted software greatly simplifies data acquisition and treatment, and is recommended.

For the considered temperature interval, the shaded areas, A , are the integrals of the output signal, s , with respect to time.

For the baseline run with an empty crucible, one obtains:

$$\overline{KA}_c = (C_c + C_o)(T_2 - T_1)$$

— for crucible plus test piece:

$$\overline{KA}_{c+t} = (m_t \overline{C}_{pt} + C_c + C_o)(T_2 - T_1)$$

— for crucible plus reference piece:

$$\overline{KA}_{c+r} = (m_r \overline{C}_{pr} + C_c + C_o)(T_2 - T_1)$$

where

$\overline{K}$ is the mean calorimeter calibration factor in watts per volt ($W \cdot V^{-1}$);

C_o is the heat capacity of the calorimeter in joules per kilogram and per kelvin ($J \cdot kg^{-1} \cdot K^{-1}$);

C_c is the heat capacity of the crucible in joules per kilogram and per kelvin ($J \cdot kg^{-1} \cdot K^{-1}$);

$\overline{C}_{pt}$ is the mean specific heat capacity of the test piece in joules per kilogram and per kelvin ($J \cdot kg^{-1} \cdot K^{-1}$);

$\overline{C}_{pr}$ is the mean specific heat capacity of the reference piece in joules per kilogram and per kelvin ($J \cdot kg^{-1} \cdot K^{-1}$);

m_t is the mass of the test specimen in kilograms (kg);

m_r is the mass of the standard reference material in kilograms (kg).

The mean specific heat capacity of the test piece is then obtained from:

$$\overline{C}_{pt}(T_1, T_2) = \overline{C}_{pr} \frac{m_r (A_{c+t} - A_c)}{m_t (A_{c+r} - A_c)}$$

5.7.2.2 Continuous heating method

See Figure 4.

Using the same notations as in 5.7.2.1 plus C_{pt} as the specific heat capacity of the test specimen and C_{pr} as the specific heat capacity of the standard reference material, one obtains for each temperature:

— for the empty crucible:

$$KS_c = (C_c + C_o) \beta$$

— for crucible plus test specimen:

$$KS_{c+t} = (m_t C_{pt} + C_c + C_o) \beta$$

— for crucible plus standard reference material:

$$KS_{c+r} = (m_r C_{pr} + C_c + C_o) \beta$$

The specific heat capacity of the test specimen for each temperature, T , is then obtained from:

$$\overline{C}_{pt} = \overline{C}_{pr} \frac{m_r (S_{c+t} - S_c)}{m_t (S_{c+t} - S_c)}$$

where

β is the constant heating rate in (K_s^{-1})

K is the calibration factor at a temperature, T

S, S_{c+t}, S_{c+r} are the output signals of the calorimeter in volts (V).

6 Test report


The test report shall contain at least the following information:

- a) name of the testing establishment;
- b) date of the test; a unique identification of the report and of each page, the customer's and signatory's names and addresses;
- c) reference to this document, i.e. "Determined in accordance with ISO 19628";
- d) type of method of measurement used;
- e) brief details of the equipment used, or any procedures adopted, including necessary deviations;
- f) calibration procedures if applicable;
- g) material type, manufacturing code, batch number;
- h) methods of production of test specimens from supplied material (if appropriate);
- i) relevant test parameters (as required for each test);
- j) results of individual tests on required number of test specimens;
- k) mean results and standard deviation (if required);
- l) comments about the test or the test results;
- m) value of specific heat capacity at each required temperature, or value of the mean specific heat capacity between two temperatures calculated on at least three test specimens.

Annex A (normative)

Drop calorimetry – determination of the calibration factor using standard reference material

A test specimen made from a standard reference material (mass m_r) is dropped from a temperature T_1 into the calorimeter at temperature T_2 . [Figure A.1](#) shows an example of a heat dissipation curve on a reference material.



Key

- 1 output signal
- 2 time
- 3 A_r in $\mu V \cdot s$

Figure A.1 — Heat dissipation curve as a function of time

The calibration factor K of the calorimeter at temperature T expressed in watt/volts is equal to:

$$K = \frac{H_r}{A_r}$$

where

A_r is the area under the peak in volt seconds;

H_r is the amount of heat required for the reference test specimen to heat from temperature T_1 to temperature T_2 .

The calibration coefficient K varies with temperature and shall be measured for each temperature.

STANDARDSISO.COM : Click to view the full PDF of ISO 19628:2017

Annex B (informative)

Standard reference material

The most convenient standard reference material, for this method the α -alumina or sapphire, certified by the National Institute of Science and Technology of USA (NIST) under reference SRM 720 up to 2 250 K.

This Standard Reference Material (SRM) is intended for use in calibrating or checking calorimeters used to measure either enthalpy or heat capacity within the range of 10 to 2 250 K.

The material furnished is synthetic sapphire cylinders, cut from centerless-ground rods grown by the Vernieul process and obtained from the Union Carbide Corporation.

The enthalpy and heat-capacity data have been derived from high-temperature enthalpy and low-temperature heat-capacity measurements. These data are presented in both tabular and equation format.

The enthalpy values are accurate to $\pm 0,1$ % from 70 to 1173 K and the heat-capacity values have an accuracy ranging from $\pm 0,1$ % at 70 K to $\pm 0,3$ % at 1 200 K. Below 70 K, the inaccuracy in heat-capacity and enthalpy values increase gradually to ± 10 % at 10 K, because, with decreasing temperature, the heat capacity of sapphire diminishes at a much faster rate than does that of the sample container (mainly copper). The precision of the heat-capacity measurement between 100 and 380 K is estimated to be $\pm 0,02$ %. The precision of the enthalpy measurement from 273,15 to 1 173 K is estimated to be $\pm 0,02$ %. For the temperature range 1 173 to 2 250 K, the precision of the enthalpy measurement is estimated to be $\pm 0,03$ %, and the accuracy of the measured enthalpy is estimated to be $\pm 0,2$ %, to a large extent reflecting the uncertainty in temperature measurements at these high temperatures. Above 1 700 K, a detectable weight loss was observed in an open container due to evaporation of material.

Relative enthalpy and heat capacity^b

Temp ^a K	$H_T - H_0$ J·mol ⁻¹	C_p J·mol ⁻¹ ·K ⁻¹
10	0,023	0,009 1
15	0,115	0,030 7
20	0,364	0,073 2
25	0,898	0,146
30	1,905	0,265
35	3,646	0,443
40	6,460	0,697
45	10,77	1,046
50	17,11	1,507
60	38,18	2,793
70	74,68	4,592
80	131,7	6,901
90	214,2	9,679
100	326,6	12,855
110	472,4	16,347
^a Temperatures expressed on IPTS-68 scale.		
^b Molecular weight = 101,961 3.		

Relative enthalpy (continued)

Temp ^a K	H _T -H ₀ K J·mol ⁻¹	C _p J·mol ⁻¹ ·K ⁻¹
120	654,3	20,07
130	874,3	23,95
140	1 133,7	27,93
150	1 433,1	31,95
160	1 772,7	35,95
170	2 152,0	39,90
180	2 570,3	43,75
190	3 026,7	47,50
200	3 519,9	51,12
210	4 048,7	54,61
220	4 611,6	57,95
230	5 207,1	61,14
240	5 833,9	64,18
250	6 490,3	67,08
260	7 175,0	69,82
270	7 886,3	72,12
273,15	8 115,6	73,21
280	8 622,8	74,87
290	9 383,2	77,20
298,15	10 020	79,01
300	10 166	79,41
310	10 971	81,51
320	11 796	83,49
330	12 641	85,37
340	13 503	87,16
350	14 383	88,84
360	15 280	90,45
370	16 192	91,97
380	17 119	93,41
390	18 060	94,78
400	19 014	96,08
410	19 982	97,32
420	20 961	98,50
430	21 951	99,62
440	22 953	100,69
450	23 965	101,71
460	24 987	162,68
470	26 018	103,60
480	27 059	104,48
490	28 108	105,33
500	29 165	106,13
^a Temperatures expressed on IPTS-68 scale.		
^b Molecular weight = 101,961 3.		