
**Molecular in vitro diagnostic
examinations — Specifications for
pre-examination processes for venous
whole blood —**

**Part 3:
Isolated circulating cell free DNA
from plasma**

*Analyses de diagnostic moléculaire in vitro — Spécifications relatives
aux processus préanalytiques pour le sang total veineux —*

Partie 3: ADN libre circulant extrait du plasma



STANDARDSISO.COM : Click to view the full PDF of ISO 20186-3:2019



COPYRIGHT PROTECTED DOCUMENT

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 General consideration	5
5 Outside the laboratory	5
5.1 Specimen collection	5
5.1.1 Information about the specimen donor/patient	5
5.1.2 Selection of the venous whole blood collection tube by the laboratory	6
5.1.3 Venous whole blood collection from the donor/patient and stabilization procedures	6
5.1.4 Information about the specimen and storage requirements at the blood collection facility	7
5.2 Transport requirements	7
6 Inside the laboratory	8
6.1 Specimen reception	8
6.2 Storage requirements for blood specimens	8
6.3 Plasma preparation	9
6.4 Storage requirements for plasma samples	9
6.5 Isolation of the ccfdDNA	10
6.5.1 General	10
6.5.2 Using blood collection tubes with stabilizers	10
6.5.3 Using blood collection tubes without stabilizers	11
6.6 Quantity and quality assessment of isolated ccfdDNA	11
6.7 Storage of isolated ccfdDNA	11
6.7.1 General	11
6.7.2 ccfdDNA isolated with commercially available kits	12
6.7.3 ccfdDNA isolated with the laboratory's own protocols	12
Annex A (informative) Impact of pre-examination process steps on circulating cell free DNA profiles in venous whole blood plasma	13
Bibliography	16

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 212, *Clinical laboratory testing and in vitro diagnostic test systems*.

A list of all parts in the ISO 20186 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Molecular in vitro diagnostics has enabled a significant progress in medicine. Further progress is expected by new technologies analysing profiles of nucleic acids, proteins, and metabolites in human tissues and body fluids. However, the profiles of these molecules can change drastically during the pre-examination process, including the specimen collection, transport, storage and processing. Consequently, this makes the outcome from diagnostics or research unreliable or even impossible because the subsequent examination might not determine the real situation in the patient, but an artificial profile generated during the pre-examination processes.

Circulating cell free DNA (ccfDNA) profiles can change significantly after blood collection (e.g. release of genomic DNA from cells in blood, ccfDNA degradation and fragmentation and ccfDNA quantity change). Therefore, special measures need to be taken to secure good quality specimens for ccfDNA examination. Studies have been undertaken to determine the important influencing factors^[23].

Standardization of the entire workflow from specimen collection to the ccfDNA examination is needed.

This document standardizes the steps of the pre-examination phase of circulating cell free DNA prepared from plasma of venous whole blood.

In this document, the following verbal forms are used:

- “shall” indicates a requirement;
- “should” indicates a recommendation;
- “may” indicates a permission;
- “can” indicates a possibility or a capability.

STANDARDSISO.COM : Click to view the full PDF of ISO 20186-3:2019

Molecular in vitro diagnostic examinations — Specifications for pre-examination processes for venous whole blood —

Part 3: Isolated circulating cell free DNA from plasma

1 Scope

This document provides recommendations and requirements on the handling, storage, processing and documentation of venous whole blood specimens intended for circulating cell free DNA (ccfDNA) examination during the pre-examination phase before an analytical test is performed. This document covers specimens collected in venous whole blood collection tubes.

This document is applicable to any molecular in vitro diagnostic examination performed by medical laboratories. It is also intended to be used by laboratory customers, in vitro diagnostics developers and manufacturers, biobanks, institutions and commercial organizations performing biomedical research, and regulatory authorities.

Different dedicated measures are taken for stabilizing blood genomic DNA, which are not described in this document. Blood genomic DNA is covered in ISO 20186-2.

Different dedicated measures are taken for preserving DNA in circulating exosomes, which are not described in this document.

NOTE ccfDNA obtained from blood by the procedures cited in this document can contain DNA originally present in exosomes^{[8][9]}.

DNA in pathogens present in blood is not covered by this document.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 15189:2012, *Medical laboratories — Requirements for quality and competence*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

analyte

component represented in the name of a measurable quantity

[SOURCE: ISO 17511:2003, 3.2, modified — The example has been deleted.]

3.2

backflow

flow of a liquid opposite to the usual or desired direction

3.3

blood collection set

intravenous device specialized for venipuncture consisting of a stainless steel beveled needle and tube (tubing) with attached plastic wings and fitting connector

Note 1 to entry: The connector attaches to an additional blood collection device, e.g. a blood collection tube.

3.4

blood collection tube

tube used for blood collection, usually with a vacuum which forces blood from the vein through the needle into the tube

3.5

ccfDNA

circulating cell free DNA

extracellular human DNA present in blood and plasma

Note 1 to entry: ccfDNA can include DNA present in vesicles such as exosomes^{[8][9]}

3.6

ccfDNA profile

circulating cell free DNA profile

amount of different ccfDNA molecules, present in blood and plasma that can be measured in the absence of any losses, inhibition and interference

3.7

closed system

non-modifiable system provided by the vendor including all necessary components for the pre-examination and/or examination (i.e. hardware software, procedures and reagents)

3.8

DNA

deoxyribonucleic acid

polymer of deoxyribonucleotides occurring in a double-stranded (dsDNA) or single-stranded (ssDNA) form

[SOURCE: ISO 22174:2005, 3.1.2]

3.9

DNase

deoxyribonuclease

enzyme that catalyzes the degradation of DNA into smaller components

3.10

examination

analytical test

set of operations having the object of determining the value or characteristics of a property

Note 1 to entry: Processes that start with the isolated analyte and include all kinds of parameter testing or chemical manipulation for quantitative or qualitative examination.

[SOURCE: ISO 15189:2012, 3.7, modified — "analytical test" has been added as additional preferred term; Notes to entry have been deleted; new Note 1 to entry has been added.]

3.11**examination performance**
analytical test performance
analytical performance

ability of an examination procedure to measure or detect a particular analyte

Note 1 to entry: Analytical performance is determined from analytical performance studies used to assess the ability of an in vitro diagnostic examination procedure to measure or detect a particular analyte.

Note 2 to entry: Analytical performance includes such characteristics as analytical sensitivity, detection limit, analytical specificity (interference and cross-reactivity), trueness, precision and linearity.

[SOURCE: ISO/TS 17822-1:2014, 3.2, modified — Two preferred terms have been added.]

3.12**examination provider**
analytical test provider

entity that provides the specific analytical test

3.13**needle holder**

barrel used in routine venipuncture procedures to hold the blood collection tube in place and to protect the phlebotomist from direct contact with blood

3.14**pre-examination processes**
preanalytical phase
preanalytical workflow

processes that start, in chronological order, from the clinician's request and include the examination request, preparation and identification of the patient, collection of the primary sample(s), transportation to and within the medical laboratory, isolation of analytes, and end when the analytical examination begins

Note 1 to entry: The pre-examination phase includes preparative processes, e.g. ccfDNA isolation procedures, which influence the outcome of the intended examination.

[SOURCE: ISO 15189:2012, 3.15, modified — An additional term has been added and more detail have been included in the definition; Note 1 to entry has been added.]

3.15**primary sample**
specimen

discrete portion of a body fluid, breath, hair or tissue taken for examination, study or analysis of one or more quantities or properties assumed to apply for the whole

[SOURCE: ISO 15189:2012, 3.16, modified — Notes to entry have been deleted.]

3.16**proficiency testing**

evaluation of participant performance against pre-established criteria by means of inter-laboratory comparisons

[SOURCE: ISO/IEC 17043:2010, 3.7, modified — Notes have been deleted.]

3.17**RNA****ribonucleic acid**

polymer of ribonucleotides occurring in a double-stranded or single-stranded form

[SOURCE: ISO 22174:2005, 3.1.3]

3.18

**RNase
ribonuclease**

enzyme that catalyzes the degradation of RNA into smaller components

3.19

room temperature

temperature in the range of 18 °C to 25 °C for the purposes of this document

Note 1 to entry: Local or national regulations can have different definitions.

3.20

sample

one or more parts taken from a primary sample

[SOURCE: ISO 15189:2012, 3.24, modified — The example has been deleted.]

3.21

stability

ability of a specimen or sample, when stored under specified conditions, to maintain a stated property value within specified limits for a specified period of time

[SOURCE: ISO Guide 30:2015, 2.1.15, modified — The words “reference material” have been replaced by “specimen or sample”.]

3.22

validation

confirmation, through the provision of objective evidence, that the requirements for a specific intended use or application have been fulfilled

Note 1 to entry: The term “validated” is used to designate the corresponding status.

[SOURCE: ISO 9000:2015, 3.8.13, modified — Notes 1 and 3 to entry have been deleted, Note 2 to entry has been renumbered as Note 1 to entry.]

3.23

venous whole blood

blood collected after directly puncturing a vein, usually with a needle and syringe, or other collection device

3.24

verification

confirmation, through the provision of objective evidence, that specified requirements have been fulfilled

Note 1 to entry: The term “verified” is used to designate the corresponding status.

Note 2 to entry: Confirmation can comprise activities such as

- performing alternative calculations;
- comparing a new design specification with a similar proven design specification;
- undertaking tests and demonstrations;
- reviewing documents prior to issue.

[SOURCE: ISO 9000:2015, 3.8.12, modified — Notes 1 and Note 2 to entry have been deleted; Note 3 to entry has been renumbered as Note 1 to entry; new Note 2 to entry has been added.]

3.25

workflow

series of activities necessary to complete a task

4 General consideration

For general statements on medical laboratory quality management systems and in particular on specimen collection, reception and handling (including avoidance of cross contaminations) see ISO 15189:2012, 4.2, 5.4.4, 5.4.6 or ISO/IEC 17020:2012, Clause 8 and 7.2. The requirements on laboratory equipment, reagents, and consumables according to ISO 15189:2012, 5.3 shall be followed; ISO 15189:2012, 5.5.1.2 and 5.5.1.3 and ISO/IEC 17020:2012, 6.2 can also apply.

All steps of a diagnostic workflow can influence the final examination result. Thus, the entire workflow, including specimen/sample storage and transport conditions, and its impact on the stability of biomolecules intended to be examined shall be verified and validated. Workflow steps which cannot always be controlled shall be documented and their impact on the examination performance shall be investigated and mitigation measures shall be established to allow the required examination performance. In these cases, risk assessment is recommended.

CcfDNA profiles can change significantly after blood collection. The post-collection release of genomic DNA from cells in blood can change the ccfDNA profile significantly (see A.1). Additional post-collection effects can also occur, e.g. ccfDNA fragmentation^{[10][11][12][13]}. All these post-collection changes can vary individually in specimens from different donors or patients, and they can also depend on pathophysiological conditions^{[10][14][15][16]}. This can impact the validity and reliability of the examination results (see A.2).

Before or during the design of an examination, it shall therefore be investigated and ensured that the ccfDNA profile(s) intended to be analysed is/are not compromised in a manner impacting the examination performance. This can be done, e.g. by applying the intended examination to specimens/samples which underwent time course studies reflecting the individual pre-examination process steps such as transport and storage and by implementing measures to prevent or reduce impacts by the identified pre-analytical variables, e.g. by using blood collection tubes with stabilizers.

Safety procedures for handling and transport shall be in place. Safety requirements on transport and handling shall be considered (see ISO 15189 and ISO 15190).

During the whole pre-examination process, precautions shall be taken to avoid cross contamination between different samples/specimens, e.g. by using single-use material whenever feasible or appropriate cleaning procedures between processing of different specimens/samples.

If a commercial product is not used in accordance with the manufacturer's instructions, responsibility for its validation, verification, use and performance lies with the user.

5 Outside the laboratory

5.1 Specimen collection

5.1.1 Information about the specimen donor/patient

The documentation shall include the ID of the specimen donor/patient, which can be in the form of a code.

The documentation should include, but is not limited to:

- a) the relevant health status of the specimen donor or patient [e.g. healthy, disease type, concomitant disease, demographics (e.g. age and gender)];

NOTE In particular, e.g. cancer, inflammation, diabetes, hepatic disease, coronary disease, respiratory syndrome, trauma, after exhaustive exercise^[10], in elderly patients suffering from acute or chronic disease, first trimester of pregnancy, placental disorders as pre-term labour, pre-eclampsia and malimplantation have been reported to affect both ccfDNA quantity and fragmentation^{[10][14][15][16]}.

- b) the information about medical treatment and special treatment prior to blood collection (e.g. anaesthetics, medications, fasting status);

- c) the type and purpose of the proposed examination requested;
- d) the appropriate consent from the specimen donor/patient.

See also ISO 15189:2012, 5.4.4.

5.1.2 Selection of the venous whole blood collection tube by the laboratory

The ccfDNA profile can be influenced by inadequate venous whole blood collection procedures and inappropriate storage/shipping conditions, plasma separation as well as by ccfDNA isolation procedures. Specifically, the post-collection release of genomic DNA from white blood cells can change the ccfDNA profile significantly (see [A.1](#)). This can impact the validity of the examination results (see [A.2](#)).

Venous whole blood should be collected in appropriate collection devices.

Blood should be collected in appropriate venous whole blood collection tubes containing ccfDNA profile stabilizers as post-collection release of genomic DNA from blood cells or other ccfDNA profile changes can cause impacts on the intended examination. The tubes' catalogue and lot number should be documented.

Blood collection tubes not containing any ccfDNA profile stabilizers should be used only if allowed by the examination provider's instructions. In these cases, EDTA blood collection tubes should be used in preference to other collection tubes^{[12][17]}. EDTA prevents clotting but not the release of DNA from blood cells^[17]. Consult the specifications by the examination provider for details. Where the ccfDNA examination provider requires usage of a dedicated blood collection tube, this shall be used.

Induced clotting process in serum tubes can lead to a leucocytes lysis causing the release of DNA, thus changing the native ccfDNA profile. Therefore the use of serum tubes should be avoided^[18].

5.1.3 Venous whole blood collection from the donor/patient and stabilization procedures

- a) The identity of the person collecting the specimen and the date and time of blood collection shall be documented.
- b) For the labelling (sample/specimen identification) of the blood collection tube a routine procedure [ISO 15189:2012, 5.4.4.3, e)] or a procedure with additional information (e.g. 2D-barcode) shall be used.
- c) Standard venipuncture technique can be used. Steps for preventing possible backflow into the donor's/patient's body can be required. The manufacturer's instructions for using the blood collection tubes shall be followed. A blood collection set and needle holder can be required when using ccfDNA profile stabilizer containing tubes. In this case, the instructions of the collection set and needle holder manufacturer shall be followed.

NOTE There is no known specific effect of venous whole blood draw procedure on the ccfDNA profile. Routine procedures can therefore be used.

- d) Blood collection tubes shall be filled in accordance with the manufacturer's instructions and attention should be drawn to the correct positioning of the collection tube during the blood draw as well as the required blood volume.

NOTE 1 Prolonged tourniquet during the venipuncture process or any hard-draw collection, such as retracting and/or pulling the syringe plunger too hard, can lead to increased hemolysis.

NOTE 2 Underfilled/overfilled blood collection tubes alter the dilution factor with the stabilizer and/or anticoagulant, which can lead to changed ccfDNA profiles.

- e) Blood collection tube manufacturer's instructions, for mixing or inverting the tube immediately after blood collection, shall be followed. Mixing or inverting the blood collection tube shall be done gently to avoid the destruction of cells in blood with subsequent release of DNA. If no information

about mixing or inverting is given by the manufacturer's instructions, each tube should be inverted 8 times to 10 times.

NOTE 3 Wrong and/or insufficient mixing can be one of the most important pre-examination variables. Unless additives in the blood collection tubes are homogeneously mixed with the specimen, the ccfDNA profile and the ccfDNA quality can be compromised, which can impact the validity and reliability of the examination results.

f) Any tampering with and/or additions to the specimen shall be documented.

Until clean plasma is obtained, special care need to be taken to avoid lysis of blood cells thus leading to the release of cellular DNA changing the native ccfDNA profile of the donor/patient. Therefore, the specimen shall not be frozen or shaken vigorously^[11].

5.1.4 Information about the specimen and storage requirements at the blood collection facility

5.1.4.1 General

The cellular DNA released by the ongoing blood cell lysis after blood collection contaminates the sample^[11] and can affect the validity and reliability of the examination result (see A.1 and A.2). The documentation regarding the specimen shall include the date and time of blood collection^{[11][18]}.

It shall be documented that the required storage conditions have been followed. The temporary storage duration in the blood collection facility contributes to the total duration for storage.

5.1.4.2 Using blood collection tubes with stabilizers

Blood tubes with ccfDNA profile stabilizers should be used. For storing the specimens collected in blood tubes with ccfDNA profile stabilizers, the dedicated blood collection tube manufacturer's instructions on storage conditions shall be followed (e.g. temperature and storage duration). Where the examination provider's instructions are more stringent, these shall be followed. The storage conditions (temperature and duration etc.) shall be documented.

5.1.4.3 Using blood collection tubes without stabilizers

5.1.4.3.1 Blood collection tubes without ccfDNA profile stabilizers should only be used, if the ordered examination specifications allow the usage of such tubes. In these cases, the examination provider's instructions on storage conditions shall be followed. This can require documentation of storage conditions (temperature and duration etc.).

5.1.4.3.2 When using blood collection tubes without ccfDNA profile stabilizers and no requirements on the storage conditions are available, the specimen should be transferred immediately to 2 °C to 8 °C in order to minimize the release of DNA from cells into the blood. The storage duration allowed at 2 °C to 8 °C shall be validated by analysing the potential impact on the examination. The storage conditions (temperature and duration etc.) shall be documented.

NOTE Some studies using blood collection tubes with EDTA have shown that storage at 2 °C to 8 °C for up to 6 h^[11] had no negative impact on results of applied examination(s)^{[19][22]}.

5.2 Transport requirements

It shall be required that the specified transport conditions be followed and documented. Any deviations therefrom shall be documented.

Temperature monitoring should be applied in a suitable manner.

NOTE 1 Temperature strips are available to detect any extreme temperature occurrence during transit.

When using blood collection tubes with ccfDNA profile stabilizers, the dedicated tube's manufacturer's instructions on transport conditions shall be followed (transport and temperature etc.). Where the examination provider's instructions are more stringent, these shall be followed. The transportation conditions (duration and temperature etc.) shall be documented.

When using blood collection tubes without ccfDNA profile stabilizers, the examination provider's instructions on transport conditions shall be followed. This can require the documentation of transport conditions (duration and temperature etc.).

When using blood collection tubes without ccfDNA profile stabilizers and no examination provider's instructions are available, the specimen should be transported at 2 °C to 8 °C in order to minimize the release of DNA from cells into the blood. The transport duration allowed at 2 °C to 8 °C shall be validated by analysing the potential impact on the examination. The transportation conditions (duration and temperature etc.) shall be documented.

NOTE 2 Some studies have shown that storage at 2 °C to 8 °C for up to 6 h^[11] had no negative impact on results of applied examination(s). See also [Table 1](#).

If the blood collection tube manufacturer or the examination provider requires a dedicated packaging, specimens shall be transported accordingly. If there are no such requirements, the specimens should be packed in tissues, airbags, paper or the like to protect it from shaking during transport, including accidental dropping of the package.

The specimen shall not be frozen or shaken strongly^[11].

See also ISO 15189:2012, 5.4.5.

The transport duration to the laboratory contributes to the total duration for storage.

6 Inside the laboratory

6.1 Specimen reception

The specimen reception date and time shall be documented as well as the name of the person receiving the specimen. Non-conformities of labelling, transport conditions and blood volume differences to specifications, leaking/broken tubes etc. shall be documented.

NOTE This includes for example a note, when specimens have been accidentally frozen or collected in another tube than the indicated one.

Where there are nonconformities in labelling, transport conditions, overall storage, and transport duration or blood volume that could affect the validity and reliability of the examination result, a new specimen should be obtained.

The correct identity of the specimen shall be checked. This should include the clinical information (see [5.1.1](#) and [5.1.3](#)) of the specimen, hospital admission number and/or donor/patient ID, name of the patient, date of birth of the patient.

6.2 Storage requirements for blood specimens

The storage temperature and time interval between specimen receipt and sample processing for obtaining plasma shall be documented.

Storage temperature and total storage duration shall not exceed specifications identified in [5.1.4](#), [5.2](#) and [Table 1](#).

The specimen total storage duration shall include the duration for storage at the blood collection facility ([5.1.4](#)), for transportation to the laboratory ([5.2](#)) and for further storage at the laboratory or other institutions. Any specified maximum storage duration given by the blood collection tube manufacturer

or the provider of the examination shall not be exceeded. If such specifications are not available, the maximum storage duration shall be validated and generally kept to a minimum.

6.3 Plasma preparation

When using blood collection tubes with ccfDNA profile stabilizers, the manufacturer's instructions to perform the plasma preparation shall be followed.

When using blood collection tubes without ccfDNA profile stabilizers, and if instructions for plasma preparation are available from the provider of the dedicated examination procedure, these shall be followed.

When using blood collection tubes without ccfDNA profile stabilizers, and no instructions are available from the provider of the dedicated examination procedure, unstabilized blood specimen should be centrifuged at 1 600*g* to 2 500*g* at 2 °C to 8 °C for 10 min. Plasma shall be carefully transferred into a new tube without disturbing the plasma cellular interface layer in order to avoid contamination with genomic DNA and cellular RNA derived from leucocytes. A second centrifugation should be performed on the supernatant of the first centrifugation step at 14 000*g* to 16 000*g* at 2 °C to 8 °C for 10 min. If the second centrifugation step is performed, the supernatant shall be carefully transferred into a new tube without disturbing the pellet^{[12][21]}. If high *g*-force centrifugation is not possible, e.g. due to lack of an appropriate centrifuge, the examination shall be verified when carrying out the second centrifugation at a lower *g*-force e.g. 3 000*g* to 5 000*g* for 20 min at 2 °C to 8 °C.

6.4 Storage requirements for plasma samples

The storage temperature and time interval between the plasma generation and the isolation of the ccfDNA shall be documented, including any deviations therefrom. Any specified maximum storage duration given by the blood collection tube manufacturer or the provider of the dedicated examination procedure shall not be exceeded. If such specification is not available, the maximum storage duration shall be validated and generally kept to a minimum.

The plasma samples should be processed for analytical down-stream tests immediately. Depending on the examination specifications, for short-term storage, plasma may be stored at 2 °C to 8 °C for a maximum of 24 h. For long-term storage, plasma should be stored frozen at ≤ -20 °C^{[11][21]}. The plasma storage conditions (i.e. duration and temperature) shall be documented.

The freezer temperature shall be continuously monitored and recorded by validated instruments, such as a circular temperature chart or an electronic thermometer.

Samples should not be stored in a “frost-free” freezer. The temperature is cycled several times a day causing nucleic acid target degradation.

Frozen plasma samples shall not be thawed more than once^{[11][21][22]}. Therefore, the plasma samples should be aliquoted into cryo-vials or other suitable vials if further testing is needed^[10]. See also [Table 1](#).

Where dedicated examination provider's instructions on storage of plasma are available, these shall be followed and documented (see [Table 1](#)).

Table 1 — Summary of storage requirements for venous whole blood collection tubes with or without ccfDNA profile stabilizers

Blood collection tube	Blood collection, transport and storage		Plasma storage	
	Duration	Temperature (°C)	Duration	Temperature (°C)
With ccfDNA profile stabilizers	According to blood collection tube manufacturer's or examination provider's instructions ^{a,b,c}		According to blood collection tube manufacturer's or examination provider's instructions ^a	
EDTA blood collection tubes without ccfDNA profile stabilizers	Examination provider's instructions ^{c,d}	Examination provider's instructions ^{c,d} 2 °C to 8 °C ^{c,e}	Examination provider's instructions ^f ≤ 24 h ^g	Examination provider's instructions ^f 2 °C to 8 °C ^g
			Long term storage	Examination provider's instructions ^f ≤ -20 °C ^g
^a If more stringent than blood tube manufacturer's instructions. ^b Requirement according to 5.1.4.2. ^c Requirement according to 5.2. ^d Requirement according to 5.1.4.3.1. ^e Requirement according to 5.1.4.3.2, if there are no examination provider's instructions. ^f Requirement according to 6.4. ^g Requirement according to 6.4, if there are no examination provider's instructions.				

6.5 Isolation of the ccfDNA

6.5.1 General

To avoid a cross contamination with amplified material, the isolation of ccfDNA should not be performed in the same area as the amplification and post-amplification steps of the examination, unless a closed system is used, which is designed to avoid cross-contamination.

CcfDNA is usually of shorter length than genomic DNA, and contains ccfDNA lengths between a few bp (basepairs) up to several kbp (kilobasepairs). Therefore, specific dedicated ccfDNA isolation procedures should be used.

Different ccfDNA isolation procedures can show significantly different ccfDNA yields^{[21][23][24]}, and size distribution patterns^{[21][23]} from the same sample. Therefore, this aspect should be specifically considered during the validation process.

6.5.2 Using blood collection tubes with stabilizers

When processing blood from tubes containing blood ccfDNA profile stabilizers, kits specified by the manufacturer of the blood collection tube should be used for the isolation of ccfDNA. The blood collection tube and kit manufacturer's instructions for isolating the ccfDNA shall be followed.

If the specifications of the examination provider require the use of a dedicated commercially available kit, then this shall be used instead in accordance with the instructions of the examination provider.

Alternative isolation procedures can be used, if no examination provider's instructions are available and if they are verified for the same requirements and validated for the same intended use. In this case, the instructions for the validated alternative for isolating the ccfDNA shall be followed.

NOTE 1 When using alternative isolation procedures, dedicated measures and technologies can be needed in order to avoid carrying over ccfDNA stabilization molecules to the final ccfDNA eluate. Stabilization molecules carry over can lead to an inhibition of the examination reaction.

NOTE 2 Dedicated procedures can be included in the ccfDNA isolation kit manufacturer's instructions for processing frozen plasma samples.

6.5.3 Using blood collection tubes without stabilizers

6.5.3.1 When using blood collection tubes not containing any ccfDNA profile stabilizer, the examination provider's instructions or validated alternatives for ccfDNA isolation shall be followed.

6.5.3.2 When using blood collection tubes not containing any ccfDNA profile stabilizers, and when there are no examination provider's instructions available, the laboratory shall validate the entire ccfDNA isolation process. Specific ccfDNA isolation kits or alternatives validated to the same requirements shall be used for reliably isolating the different fragment lengths of the ccfDNA profile. The kit manufacturer's instructions or the instructions for the validated alternative for isolating the ccfDNA shall be followed.

The reagents and consumables coming in contact with the ccfDNA shall be DNase-free.

6.6 Quantity and quality assessment of isolated ccfDNA

Where required, the ccfDNA quantity and quality should be checked according to the examination provider's instructions or according to validated procedures by generally accepted physical, chemical and biochemical procedures. These may include one or more of the following.

- a) Quantity: ccfDNA is usually at very low concentration, which makes the use of UV absorbance reading such as spectrophotometers unreliable and therefore should be avoided. Often, ccfDNA isolation procedures use carrier nucleic acids [e.g. carrier RNA of a neutral sequence such as Poly(A) or Poly(C)]; this carrier will additionally interfere with the UV absorbance reading. Appropriate methods for ccfDNA quantification are therefore required, such as qPCR (quantitative polymerase chain reaction), targeting a known sequence of a single copy gene.

Other methods, such as digital polymerase chain reaction (dPCR) and fluorometric or chip-based methods, are available as well.

- b) Quality: Due to the low concentration and the heterogeneity of ccfDNA profiles, there is no generic method for quality assessment. Depending on the examination requirements dedicated quality assessment test may therefore be required to be performed, e.g. percentage of fetal DNA within the total ccfDNA^[29].

The ccfDNA isolation performance should be tested in a ccfDNA proficiency testing program where available^[30].

6.7 Storage of isolated ccfDNA

6.7.1 General

For long-term storage, usually the isolated ccfDNA is frozen. However, for ccfDNA preservation other validated methods for archiving can also be used^[31].

For long-term storage, aliquots of the isolated ccfDNA should be generated to avoid repeated freezing and thawing^[11] or repeated recovery from other archiving systems.

For small ccfDNA amounts, storage vessels with reduced nucleic acid adsorption to the tube wall should be used.

Unintended freeze-drying of the isolated ccfDNA during long-term storage due to water evaporation should be avoided as the ccfDNA can degrade and the recovery from the tube can be difficult or even impossible. Therefore, appropriate storage vessels, such as screw-capped cryogenic vials, avoiding water evaporation during long-term storage, should be used and documented.

Traceability shall be ensured. For long-term storage, a validated system to organize and uniquely mark aliquots at the intended storage temperature making them easily retrievable and identifiable should be in place. Labels suitable for storage temperature with readable 1D- or 2D-barcodes or pre-printed tubes with unique codes provided by manufacturers are recommended to avoid loss or confusion of sample identity.

The freezer temperature shall be continuously monitored and recorded by validated instruments, such as a circular temperature chart or an electronic thermometer.

Samples should not be stored in a “frost-free” freezer. The temperature is cycled several times a day causing nucleic acid target degradation.

6.7.2 ccfDNA isolated with commercially available kits

For storing the isolated ccfDNA before the examination, the ccfDNA isolation kit provider's specific instructions should be followed. Where the examination provider's instructions are most stringent, these shall be followed.

6.7.3 ccfDNA isolated with the laboratory's own protocols

If the laboratory's own validated ccfDNA isolation procedures are used, the isolated ccfDNA should be assayed immediately. Where the isolated ccfDNA cannot be assayed immediately, the laboratory shall have verified procedures in place on how to store the isolated ccfDNA before the examination.

For long-term storage, isolated ccfDNA should be eluted in an appropriate buffer and stored at $\leq -20\text{ }^{\circ}\text{C}$ ^[11]. Other validated methods for archiving can also be used^[31].

Annex A (informative)

Impact of pre-examination process steps on circulating cell free DNA profiles in venous whole blood plasma

A.1 Post collection changes of blood ccfDNA profiles

Upon storage and transport of venous whole blood collected in blood collection tubes without ccfDNA profile stabilizer, blood cells undergo apoptosis or mechanical lysis and release fragmented genomic DNA which thus becomes cell-free. These post collection changes artificially modify the native ccfDNA profile as it was in the donor's body. This can lead to unreliable or wrong ccfDNA examination results. The artificial post collection release of ccfDNA can be measured by comparing ccfDNA quantities in plasma samples obtained immediately after blood collection with ccfDNA quantities in plasma samples obtained from stored blood.

[Figure A.1](#) shows the post collection change of 18S ccfDNA quantities in venous whole blood specimens at different storage durations, with or without ccfDNA profile stabilizers.

Venous whole blood specimens were collected from healthy donors ($n = 6$) in EDTA blood collection tubes without any ccfDNA profile stabilizer and either left untreated or were stabilized by a ccfDNA profile stabilizer reagent immediately after blood collection. Plasma was either obtained immediately after blood collection/stabilization (t_0) or after blood storage for 1, 3 or 6 days at room temperature. Plasma was obtained according to the examination provider's instructions from EDTA blood collection tubes. A first centrifugation was performed at 1 900g for 15 min at room temperature. Plasma was transferred into a new tube. A second centrifugation of this plasma was performed at 1 900g for 10 min at room temperature. From stabilized blood, plasma was obtained according to the tube manufacturer's instruction which were identical to those for EDTA blood. ccfDNA was isolated from the obtained plasma samples by using an automated ccfDNA extraction procedure. The 18S ribosomal DNA (rDNA) copy numbers in the isolated ccfDNA were determined by real-time PCR (66 bp/500 bp amplicon).

Blood collected and stored in EDTA blood collection tubes without any ccfDNA profile stabilizer showed a significant increase of 18S rDNA copies in the isolated ccfDNA in a range of 5-10 fold after 3 days and up to 20-150-fold after 6 days of storage at room temperature.

An increase was also observed after one day of storage. In contrast, blood preserved by a ccfDNA profile stabilizer did not show these artefactual post collection changes.

The release of cellular DNA can be minimized by using blood collection tubes containing a ccfDNA profile stabilizing reagent, which prevents or minimizes the increase of ccfDNA quantities over time during blood storage and transport.

Blood collection tubes with ccfDNA profile stabilizers can be an important contributor to avoid impacts on examinations by artefactual post collection changes in the native whole blood ccfDNA profile.