
**Determination of particle size
distribution — Single particle light
interaction methods —**

**Part 1:
Light scattering aerosol spectrometer**

*Détermination de la distribution granulométrique — Méthodes
d'interaction lumineuse de particules uniques —*

Partie 1: Spectromètre d'aérosol en lumière dispersée



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 21501-1:2009



COPYRIGHT PROTECTED DOCUMENT

© ISO 2009

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Terms and definitions	1
3 Requirements	3
3.1 Size range	3
3.2 Counting efficiency.....	3
3.2.1 General	3
3.2.2 Lower size limit	4
3.2.3 Upper size limit	4
3.3 Size resolution	4
3.4 Sizing accuracy	5
3.5 Sampling flow rate	5
3.6 Effective detection flow rate	5
3.7 Maximum particle number concentration	5
4 Test method	5
4.1 Size calibration.....	5
4.2 Effective detection flow rate	6
4.3 Maximum particle number concentration	7
4.4 Size resolution	8
4.5 Counting efficiency.....	9
Annex A (informative) Principle of the instruments	11
Annex B (informative) Particle size standards	18
Annex C (informative) Effects of the LSAS parameters on the particle size and particle number concentration determination	21
Annex D (informative) Representative sampling	22
Annex E (informative) Example of an LSAS calibration with DEMS-classified PSL particles	24
Bibliography	26

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 21501-1 was prepared by Technical Committee ISO/TC 24, *Particle characterization including sieving*, Subcommittee SC 4, *Particle characterization*.

ISO 21501 consists of the following parts, under the general title *Determination of particle size distribution — Single particle light interaction methods*:

- *Part 1: Light scattering aerosol spectrometer*
- *Part 2: Light scattering liquid-borne particle counter*
- *Part 3: Light extinction liquid-borne particle counter*
- *Part 4: Light scattering airborne particle counter for clean spaces*

Introduction

Monitoring particle size distributions and particle number concentrations is required in various fields, e.g. in filter manufacturing, in the electronic industry, in the pharmaceutical industry, in the chemical industry, in the manufacture of precision machines and in medical operations. The aerosol spectrometer is a useful instrument for the determination of the size distribution and number concentration of particles suspended in a gas. The purpose of this part of ISO 21501 is to provide the calibration procedure and the validation method for aerosol spectrometers, so as to improve the accuracy of the measurement result by aerosol spectrometers in general, and to minimize the difference in the results measured by different instruments.

STANDARDSISO.COM : Click to view the full PDF of ISO 21501-1:2009

STANDARDSISO.COM : Click to view the full PDF of ISO 21501-1:2009

Determination of particle size distribution — Single particle light interaction methods —

Part 1: Light scattering aerosol spectrometer

1 Scope

This part of ISO 21501 specifies characteristics of a light scattering aerosol spectrometer (LSAS) which is used for measuring the size, number concentration and number/size distribution of particles suspended in a gas. The light scattering technique described in this part of ISO 21501 is based upon single particle measurements. The size range of particles measured by this method is between approximately 0,06 µm to 45 µm in diameter.

Instruments that conform to this part of ISO 21501 are used for the determination of the particle size distribution and particle number concentration at relatively high concentrations of up to 10^{11} particles/m³.

Application fields include:

- characterization of metered dose inhalers (MDI), dry powder inhalers (DPI) and nebulizers in pharmacy;
- production control of active agents;
- cut-off determination: impactors, cyclones and impingers;
- atmospheric aerosols: bio-aerosols, stables/composting facilities, nebulized droplets, measurements in street tunnels;
- fractional separation efficiency determination of filters.

For the above-mentioned applications, aerosol spectrometers should determine the particle size distribution, particle number concentration, size resolution and sizing accuracy as accurately as possible. These aerosol spectrometers are not suitable for the classification of clean rooms.

2 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

2.1

particle

discrete element of the material regardless of size

[ISO 2395:1990]

2.2

aerosol

suspension in a gaseous medium of solid particles, liquid particles or solid and liquid particles having a negligible falling velocity

[ISO 4225:1994]

NOTE In general, one divides the atmospheric aerosol into three size categories: the superfine range $x < 0,1 \mu\text{m}$, the fine range $0,1 \mu\text{m} < x < 1 \mu\text{m}$ and the coarse range $x > 1 \mu\text{m}$, where x is the particle diameter.

2.3

particle size

size of a sphere having the same physical properties in the method of analysis as the particle being described

NOTE 1 See also equivalent particle diameter (2.4).

NOTE 2 There is no single definition of particle size. Different methods of analysis are based on the measurement of different physical properties. The physical property to which the equivalent diameter refers shall be indicated using a suitable subscript or reference to the documentary measurement standard according to which the particle size was measured.

In ISO 9276 the symbol x is used to denote the particle size or the diameter of a sphere. However, it is recognized there that the symbol d is also widely used to designate these values. Therefore the symbol x may be replaced by d where it appears.

2.4

equivalent particle diameter

diameter of the sphere with defined characteristics which behaves under defined conditions in exactly the same way as the particle being described

2.5

light scattering equivalent particle diameter

x_{sca}
equivalent diameter of a homogeneous sphere of a reference substance (e.g. latex) which scatters defined incident light with the same radiation efficiency into a defined solid angle element

2.6

number concentration density distribution

density (frequency) distribution of the particle number concentration represented as a function of the particle size

2.7

particle concentration

indication of, e.g. particle numbers, particle mass, particle surface related to the unit volume of the carrier gas

NOTE For the exact concentration indication, information on the gaseous condition (temperature and pressure) or the reference to a standard volume indication is necessary.

2.8

coincidence error

probability of the presence of more than one particle inside the sensing zone simultaneously

NOTE Coincidence error is related to particle number concentration and size of sensing zone.

2.9

counting efficiency

relation of the concentration determined from the counting rate of the measuring instrument and the real concentration of the aerosol at the inlet of the instrument

2.10**border zone error**

particle sizing error that occurs when particles pass through the optical border of the sensing zone

3 Requirements**3.1 Size range**

The measuring size range is defined by the lower and upper size limit of quantification.

3.2 Counting efficiency**3.2.1 General**


The counting efficiency is the relation of the particle number concentration $C_{N, \text{measured}}$ — determined from the counting rate of a device and corrected for possible coincidence errors — to the real particle number concentration $C_{N, \text{actual}}$ of the aerosol at the inlet of the device. The counting efficiency $[\eta(x)]$ is a function of the particle size and is expressed as the ratio:

$$\eta(x) = \frac{C_{N, \text{measured}}(x)}{C_{N, \text{actual}}(x)} \quad (1)$$

The counting efficiency is also a function of signal processing, the homogeneous illumination of the measuring volume and the extent to which the particles enter the measuring volume and flow rate.

Figure 1 shows a graphical representation of counting efficiency. In an ideal case, the counting efficiency in the middle of the measuring range — as represented here — has the value one. If an experimental examination results in a value deviating from one, then this is to be accounted for as a correction to the measurement result. Usually, one defines the lower and/or upper size limit of the measuring range for the particle size with the particle diameters for which the counting efficiency shows the value 0,5.

For a proper evaluation of a measuring instrument, it is useful to determine the complete counting efficiency curve, or to indicate the particle diameters corresponding to values of the counting efficiency (e.g. at 0,1 and 0,9) besides those corresponding to a counting efficiency value of 0,5. The counting efficiency can be determined according to 4.5.

**Key**

X particle size
Y counting efficiency $\eta(x)$

- 1 lower size limit
- 2 upper size limit
- 3 size range

Figure 1 — Graphical representation of the counting efficiency ^[5]

3.2.2 Lower size limit

The lower size limit for the particle size is defined by convention to be the smallest diameter with which the counting efficiency shall be $0,5 \pm 0,15$ (50 % $\pm$ 15 %; lower size limit of the measuring range).

3.2.3 Upper size limit

The upper size limit for the particle size is defined by convention to be the largest diameter with which the counting efficiency shall be $0,5 \pm 0,15$ (50 % $\pm$ 15 %; upper size limit of measuring range). This is of particular interest if the aerosol inlet is situated horizontally in the LSAS, as particle losses can occur in the LSAS through impact and sedimentation.

3.3 Size resolution

The size resolution indicates which neighbouring particle sizes a particle measuring instrument can still differentiate between and record separately. Aerosol spectrometers should determine the particle size distribution and the particle number concentration as accurately as possible with high size resolution and good sizing accuracy. The size resolution depends on particle size.

Almost all measuring instruments determine the particle number concentration in a limited number of size classes which are firmly specified by the instrument design (e.g. instrument geometry, evaluation electronics, evaluation software). In practical operation, the size resolution of an LSAS cannot be better than the width of its size classes.

The size resolution can be determined according to 4.4.

3.4 Sizing accuracy

The sizing accuracy depends on the particle size. Therefore, the sizing accuracy can be evaluated for any particle size as follows

$$\varepsilon(x) = \frac{x_s - x_r}{x_r} \times 100 \quad (2)$$

where

$\varepsilon(x)$ is the ratio of particle size difference, in %;

x_r is the particle size of the certified reference material, in μm ;

x_s is the particle size indicated by the LSAS, in μm .

The sizing accuracy of an LSAS describes the difference between the actual calibration particle size and the particle size indicated by the instrument. The correlation between the particle size and size class stated by the manufacturer (channel number) is normally based upon a calibration of the instrument with a known test aerosol [mostly polymer latex (PL) particles]. Refer to Annex E.

3.5 Sampling flow rate

The sampling flow rate is the volumetric flow rate through the LSAS. Any error in the volume flow will affect the reported particle number concentration with a proportional relationship. The error in the sampling flow rate shall be within 5 %. Typical sampling flow rate is about 0,5 l/min to 5 l/min.

3.6 Effective detection flow rate

The effective detection flow rate is the volumetric flow rate through the optically and/or aerodynamically limited sensing zone. For particle number concentration measurements, the effective detection flow rate shall be evaluated. If the geometrical dimensions of the measuring volume are exactly known, the effective detection flow rate can be determined by measuring the transit time of the particles through the measuring volume. Otherwise, the effective detection flow rate has to be determined by a calibration experiment according to 4.2.

3.7 Maximum particle number concentration

The LSAS must be able to measure in high particle number concentrations up to 10^{11} particles/ m^3 and the maximum particle number concentration shall be specified by the manufacturer. According to 4.3, the coincidence loss at the maximum particle number concentration of an LSAS shall be equal to or less than 10 %.

4 Test method

4.1 Size calibration

The light scattered by the individual particles is detected and transformed into a voltage pulse. These pulses are classified according to their height in a multi-channel analyser (MCA). The result is a count rate histogram. This histogram is transferred into a particle size distribution by applying the calibration curve. This calibration curve, which is instrument specific, shall be determined either experimentally or theoretically. Normally, aerosol spectrometers are calibrated with monodisperse test aerosols of known (traceable) size and refractive index.

For the production of monodisperse test aerosols, several generation principles can be used. The size distribution of test aerosols with small variances are obtained if aqueous suspensions of PL particles are nebulized, dried and drawn through the instrument. Dry powder monodisperse PL particles are also useful for calibration.




Figure 2 — Comparison of a theoretical calibration curve (line) and experimental data with monodisperse aerosol particles (vertical bars) ^[14]

Other techniques of providing narrow distributed aerosols are the vibrating-orifice generator or passing a polydisperse aerosol through a differential mobility analyser, which extracts particles within a narrow size range according to their electrical mobility.

4.2 Effective detection flow rate

The effective detection flow rate is obtained by relating the instrument particle count rate to a reference particle number concentration measurement using an instrument with a higher particle number concentration range than the instrument to be calibrated and that is traceable to internationally accepted standards.




Figure 3 — Example for effective detection flow rate and counting efficiency test set-up

The effective detection flow rate is as follows.

$$q_E = \frac{N}{C_{N, \text{ref}}} \quad (3)$$

where

q_E is the effective detection flow rate, in volume per time;

N is the measured particle count rate by LSAS, in number per time;

$C_{N, \text{ref}}$ is the measured particle number concentration by reference instruments, in number per volume.

Effective detection flow rate calibration is necessary for counting efficiency evaluation. If the effective flow rate is unknown, the effective detection flow rate calibration shall be performed with particle sizes much smaller than the dimensions of the optical sensing zone, because larger particles effectively increase the resulting measuring volume. However, the particle size must be such that both the LSAS under test and the reference device have a counting efficiency close to 100 %.

4.3 Maximum particle number concentration

The maximum particle number concentration of an LSAS is determined by the coincidence loss, which is normally limited to 10 %. The coincidence loss is determined by the sampling flow rate, the time required for particles to pass through the sensing zone and the electrical signal processing time. These values are determined by the design of the LSAS. Calculation of coincidence loss is as follows:

$$L = \left[1 - \exp(-q_E \times t \times C_{N, \text{max}}) \right] \times 100 \quad (4)$$

where

L is the coincidence loss, in %;

q_E is the effective detection flow rate, in m³/s;


t is the time of passing through the effective sensing region plus electrical processing time, in s;

$C_{N, \text{max}}$ is the maximum particle number concentration, in particles per m³.

There are several practical ways to ensure sufficiently low coincidence loss.

- A defined and reproducible concentration change of the same aerosol can be performed. If the number concentration measured by the LSAS shows an equivalent change according to the defined conditions, both concentrations are below the maximum particle number concentration.
- A reference instrument may be used to measure the reference concentration $C_{N, \text{ref}}$. In this case, the particle size must be chosen such that the counting efficiency of both the LSAS and the reference instrument is close to 100 %. The number concentration range of the reference instrument must be greater than the maximum particle number concentration of the LSAS. As long as $C_{N, \text{LSAS}}/C_{N, \text{ref}}$ is greater than 0,9, the number concentration is below the maximum particle number concentration of the LSAS.

In both cases, increasing the particle number concentration until a coincidence loss of 10 % is found will determine the maximum particle number concentration of the LSAS.



Key

X particle number concentration from defined concentration changes or $C_{N, \text{ref}}$

Y LSAS particle number concentration

Figure 4 — Example for a concentration comparison measurement to determine the coincidence loss and the maximum particle number concentration

4.4 Size resolution

The size resolution depends on the particle size. It can be determined by using monodisperse test aerosols with known geometric standard deviation (σ_{PSL}). Determine the median MCA channel (Ch2 in Figure 5). The lower channel (Ch1) and upper channel (Ch3) are selected so that the relative counts are $61 \% \pm 10 \%$ of the median channel. Using the calibration curve, determine the particle sizes corresponding to Ch1 and Ch3. Calculate the absolute value of the differences in particle sizes between the PL particle size and the particle sizes corresponding to Ch1 and Ch3. The larger difference in particle size is the observed standard deviation σ . Calculate the percentage of size resolution of the LSAS using Equation (5). (For further information, refer to Annex E.)

$$R = \frac{\sqrt{\sigma^2 - \sigma_{\text{P}}^2}}{x} \times 100 \quad (5)$$


where

R is the size resolution, in %;

σ is the observed standard deviation of the LSAS, in μm ;

σ_{P} is the standard deviation of the particle size of the calibration particles, as reported by the reference material producer, in μm ;

x is the particle size of the calibration particle, in μm .

**Key**

X MCA channel
Y relative count

- 1 count rate distribution
2 lower side resolution
3 upper side resolution

Figure 5 — Size resolution

4.5 Counting efficiency

Counting efficiency can be experimentally determined by using the measuring set-up shown in 4.2.

For the test of the counting efficiency according to 3.2, it is assumed that $C_{N, \text{actual}} = C_{N, \text{ref}}$.

It must be verified that


- the LSAS under test and the reference counting device are below the 10 % coincidence limit,
- the effective detection flow rate of the LSAS is correct,
- the LSAS under test and the reference counter are calibrated correctly concerning particle size, the optics is not polluted and the electronics is adjusted correctly.

If $Y > 1$ in Figure 1, then $C_{N, \text{ref}} \neq C_{N, \text{actual}}$ or a), b) and c) are not adhered to.

If $Y = 1$ in Figure 1, then the LSAS under test has 100 % counting efficiency.

If $Y < 1$ in Figure 1, then this is due to the sensitivity of the sensor (point 1 in Figure 1; see also point B in Figure 6) or to transport losses (point 2 of Figure 1).

If Y decreases with increasing $C_{N, \text{ref}}$, then this is due to coincidence losses (see point A in Figure 6).



Key

X reference particle number concentration $C_{N, \text{ref}}$ in particles/cm³

Y counting efficiency $C_{N, \text{LSAS}}/C_{N, \text{ref}}$

○ 0,156 µm PL particles

◆ 0,234 µm PL particles

△ 0,312 µm PL particles

Figure 6 — Counting efficiency is affected by coincidence A, incorrect flow rate (4.2) and decreasing sensitivity B of the instrument

Annex A (informative)

Principle of the instruments

A.1 Light scattering

A.1.1 General

The measuring method of an LSAS is based on the Lorentz-Mie theory. The particle characteristic, the diameter, is determined for individual particles which are assumed to be spherical, and the number of measured particles is registered at the same time.

If light with the wavelength λ meets a spherical particle with a diameter d and a refractive index n , then the light is scattered in different directions (Figure A.1). The scattering of light at the particle is caused by diffraction, refraction and reflection. The polarization plane of the incident light wave is also affected.

The intensity I of the light scattered from the single particles depends on the incident light intensity I_0 , the polarization angle Φ , the detection angle of the scattered light Θ , the refractive index n , the light wavelength λ and the particle diameter d .

$$I = I_0 \times f(\Phi, \Theta, n, \lambda, d) \quad (\text{A.1})$$

By means of the scattering parameter α , introduced by Mie


$$\alpha = \frac{\pi \times d}{\lambda} \quad (\text{A.2})$$

the relation between the sphere circumference $\pi \times d$ to the wavelength λ is used in Equation (A.1):

$$I = I_0 \times f(\Phi, \Theta, n, \alpha) \quad (\text{A.3})$$

With regard to the particle-size-depending scattering power, three ranges can be distinguished in terms of the scattering parameter α (see Figure A.2).

- a) **Rayleigh range:** $\alpha \ll 1$; here the scattering power rises with the sixth power of the particle diameter, see References [6] and [17]. The scattered light will be proportional to d^6/λ^4 . This means if in the Rayleigh range one wants to be able to measure a particle half as large as before (lower size limit), then doubling the supplied quantity of light is not enough. The required quantity of light must be possibly 64-times stronger than for a particle twice as large.
- b) **Mie range:** $0,1 \leq \alpha \leq 10$; here the relation between the scattered light intensity and the particle size is not monotonic for certain optical configurations (Figure A.2).
- c) **Fraunhofer range:** $\alpha \gg 1$; here a quadratic relation between the scattering power and particle diameter is valid.

**Key**


- 1 incident light
- 2 scattering area
- α scattering parameter
- λ wavelength of light
- θ scattering angle
- n refractive index
- r radius
- Φ polarization angle

Figure A.1 — Principle of the scattering incident light by particle

A.1.2 Theoretical response function

Although it is advisable to calibrate an optical instrument experimentally by means of test aerosols of known size and refractive index, theoretical response functions give a general indication of the characteristics of an optical system. By means of electromagnetic theory, the response function of optical systems, which describe the power of light scattered through the collecting aperture, can be calculated as a function of the diameter of a spherical particle. The parameter of these functions is the refractive index of the particle material. Commercial optical systems can be divided roughly into instruments using low-angle scattering, those collecting scattered light in the forward direction (diffraction lobe) and those employing right-angle scattering. With respect to the light source, one has to distinguish between polychromatic incandescent light and monochromatic laser light.

The influence of monochromatic and polychromatic light on the response function is demonstrated in Figure A.2, where the partial scattering cross-section is plotted against the particle diameter for monochromatic and white light and a mean scattering angle of $\theta = 45^\circ$. For monochromatic light, the curves show typical oscillations. These fluctuations can be smoothed by using either white light and a mean scattering angle between 45° and 90° , or monochromatic light and a wide angle collecting aperture. In the diameter range smaller than the wavelength, the response functions are monotonic even for monochromatic light.



a) Example using monochromatic illumination

b) Example using polychromatic illumination


Key

- X particle size, in μm
 Y relative scattered light intensity
 n refractive indexes of particles

Figure A.2 — Examples for theoretical response function of a 45 light scattering system using monochromatic and polychromatic illumination adapted from Reference [6]

A.2 Operating principle

The operational principle of an LSAS ^[5] is based on the particle passing individually through a light beam (Figure A.3) or through an intensively lighted measuring volume (Figure A.6) and on collecting the light scattered by the particles. At the same time, the intensity of the scattered light is interpreted as a measure of the particle size. From the number of the counted scattered light impulses, one can conclude with known sample volume flow and defined measuring period the particle number concentration. Figure A.3 shows the set-up in principle. Laser diodes, He-Ne lasers or intensive white sources of light (e.g. xenon high-pressure lamps) are used. The scattered light is collected by optics under a certain solid angle (receiver aperture) and focussed onto a detector (photo-multiplier or photodiode).



Key

- 1 aerosol supply
- 2 measuring chamber
- 3 light source
- 4 measuring volume
- 5 light trap
- 6 detector
- 7 amplifier
- 8 aerosol outlet

Figure A.3 — Example set-up of an LSAS [5]

Figure A.4 contains an example of an amplified detector output signal $U(t)$ as a function of time t . The scattered light pulses are classified according to their height and summed as counting events in appropriate classes.




Figure A.4 — Example of the output signal of an LSAS [5]

For the application of the measuring principle, it is presupposed that there is always only one particle in the measuring volume and that the signal processing for this particle is finished before the next scattered light impulse appears at the detector output.


The height of the scattered light impulses does not depend only on the particle size but also on the particle material and the particle shape, so that finally one determines an equivalent scattered light diameter [4] defined by the calibration with spherical latex particles.

A.3 Definition of the sensing zone

Figures A.5 and A.6 illustrate how the sensing zone can be defined to determine where the scattered light is collected by the photo-detector. Where the sensing zone is aerodynamically defined, the sample volume flow rate is equal to the effective detection flow rate, but where the sensing zone is optically defined, then the effective detection flow rate is smaller than the sample volume flow rate.

Optically defined measuring volumes can be made very small and therefore are useful for the measurement of particle size distributions at high particle number concentrations with as little coincidence error as possible. A particle number concentration of 10^{12} particles/m³ can be shown by a simple calculation to be equivalent to a single particle in the volume enclosed by a cube with an edge length of 100 µm. Thus, should a particle number concentration of 10^{11} particles/m³ need to be measured, the edge length of the measuring volume cannot be larger than 100 µm for measurements to be free from coincidence error. The indication of the measuring volume size is very helpful for the user in order to estimate the maximum particle number concentration that can be measured without coincidence errors.


In Figure A.6, the problem of the border zone error can be seen. A particle which is partially illuminated in the border zone scatters only a fraction of the light of a particle of the same size which is situated in the measuring volume centre. The border zone error becomes larger with increasing particle size.



Key

- 1 sample air
- 2 aerosol nozzle
- 3 clean air nozzle
- 4 aerosol flow
- 5 clean air curtain
- 6 light beam
- 7 measuring volume
- 8 suction nozzle

Figure A.5 — Example of an aerodynamic definition of the measuring volume [5]



Key

- 1 aerosol channel
- 2 light beam
- 3 measuring volume
- 4 detection
- 5 detail of border zone error

Figure A.6 — Optical definition of the sensing zone for an LSAS

Annex B (informative)

Particle size standards

B.1 Particle size reference materials

The unambiguous realization of particle size in terms of the SI-unit length is possible only with microscopic methods and for spherical particles. Nevertheless, laboratories using LSAS also have needs for metrological quality assurance, which can be met, for example, by reference materials. For the generation of reference aerosols for the test and calibration of measuring instruments, two different approaches can in principle be taken.

On the one hand, spherical, monodisperse reference particles are commercially available in defined sizes as bulk or as aqueous suspensions. The size of the particles can be examined, for example, by light or electron microscopically and is usually certified by the manufacturer. These particles can be regarded as certified reference materials.

Alternatively, monodisperse reference particles can be generated directly in the aerosol phase. These test aerosols are generated *in situ* for immediate use and the users must satisfy themselves as to the reliability of the methods used.

B.2 Certified reference materials for dispersion with aerosol generators


The currently available reference particles are very limited in terms of particle composition. There are monodisperse polymer latex particles which are available as suspensions or in bulk.

In order to be able to be used for calibration purposes, the particles must be aerosolized. The user must both achieve and demonstrate a complete and reproducible dispersion. Otherwise, incompletely dispersed particle agglomerates can result in erroneous calibration results.

In addition, aerosolization can introduce an electrostatic charge on to the particles. The influence of this charge on the calibration experiment must be determined. If necessary, the calibration aerosol must be electrostatically equilibrated using suitable methods.

Bulk particles can be dispersed using a solid aerosol generator such as a brush feeder [4], a small-scale Venturi feeder or a belt feeder [4], as relatively small quantities are required. These methods are limited to particle sizes $> 0,5 \mu\text{m}$, since for smaller particles incomplete deagglomeration is achieved.

Polymer latex suspensions usually contain about 10 % solids. After sufficient dilution, they are nebulized, e.g. with the help of a nozzle nebulizer [4]. The dilution factor must be suitably large to ensure that there is no more than one latex particle in each water droplet generated during nebulization. Consequently, this means that the vast majority of the droplets do not contain latex particles. Furthermore, after evaporation of the water, these droplets result in residual particles from the impurities in the water and from the stabilizer substances added to the latex suspension. The particle number concentration of test aerosols generated in such a way is therefore limited to a few hundred particles per cubic centimetre, while the particle number concentration of the smaller residual particles can amount to as many as $10^6/\text{cm}^3$. Such test aerosols are therefore not suitable for testing instruments which do not measure particle size, as such instruments will be unable to distinguish between the latex and residual particles.

**Key**

- X particle diameter, in μm
 Y particle number distribution

Figure B.1 — Number distribution $q_0(x)$ of a latex calibration aerosol [5]

In Figure B.1, the number distribution of a latex calibration aerosol measured with an LSAS is represented. One recognizes a narrow, almost monodisperse peak in the distribution slightly above $0,5 \mu\text{m}$. The larger particles existing in small numbers are agglomerates from droplets, which contained more than one latex particle. The distribution of the residual particles begins only below $0,2 \mu\text{m}$ and is shown by this measurement to be a comparatively small part of the distribution.

B.3 Generation of test aerosols from materials other than reference materials

The generation of monodisperse test aerosols with a vibrating orifice generator [4] belongs to the secondary particle size standards. At this instrument, a periodic instability is superimposed onto a liquid jet by a vibrating orifice, which under certain operating conditions leads to the break-up of the jet into equally sized large drops. The size of these drops is determined by the volume flow of the liquid through the orifice and the vibrating frequency of the orifice. By dissolving the aerosol substance in a volatile solvent, the diameter of the ultimately generated particles, after subsequent evaporation of the solvent from the generated droplets, can be adjusted within broad limits by varying the concentration of the desired material in the solvent. Usable aerosol substances are, for example, heavy-volatile hydrocarbons which can be dissolved in alcohols or salts in aqueous solution. Particles can be generated within a size range of approximately $1 \mu\text{m}$ to $50 \mu\text{m}$, with the particle number concentration of the order of $100 \text{ particles}/\text{cm}^3$. Since each vibration of the orifice generates a single droplet, the instrument could in theory be a source of a standard particle number concentration. However, undefined losses of the generated particles in the transport system make this difficult in practice.

Monodisperse aerosols can also be generated by classifying polydisperse aerosols. For fine and superfine particles, differential mobility analysers are used for this purpose. With these devices — within the particle size range of between approximately 5 nm and $1 \mu\text{m}$ — a fraction of particles having the same electrical mobility can be extracted from a polydisperse aerosol. If the number distribution of the polydisperse aerosol is not too broad and its median diameter is selected accordingly, all extracted particles carry only a single electrical charge and the monodisperse aerosol is also monodisperse. The particle number concentration of the monodisperse aerosol fraction depends on the particle size distribution of the polydisperse aerosol and the size selected for the classified particles. This particle number concentration range is typically in the range of 100 to $100\,000 \text{ particles}/\text{cm}^3$.

Differential mobility-analysers can also be used to separate latex test aerosols generated from suspensions from the residual particles. Since only some of the latex particles delivered to the mobility analyser carry a single electrical charge, the particle number concentration of the classified aerosol is significantly smaller than that of the aerosol before classification.

STANDARDSISO.COM : Click to view the full PDF of ISO 21501-1:2009

Annex C

(informative)

Effects of the LSAS parameters on the particle size and particle number concentration determination

C.1 Calibration curve

A monotonic, steadily increasing calibration curve is important for high size resolution and classification accuracy.

Due to the influence of optical properties of the particle material, particle shape (spherical or irregular) and the optical set-up (light source, detection angle), experimentally measured calibration curves may differ from theoretical predictions.

It is advisable to calibrate an optical instrument experimentally by means of test aerosols of known size and refractive index. The output of such a calibrated instrument will then be related to light scattering equivalent particle diameter.

A steadily increasing calibration curve can be achieved by using white light and a 90° scattered light detection, or by collecting the scattered light over a large angle range.

C.2 Border zone error

Border zone error results in undersized particles. Border zone error increases with increasing particle size.

C.3 Coincidence error

Coincidence error results in the underestimation of particle number concentration and the overestimation of particle diameter, as described in Reference [9].

C.4 Size resolution

Size resolution is a measure of the ability of an instrument to distinguish between particles of different sizes. With low size resolution, neither the mean nor the modal diameter of a particle distribution can be determined reliably.

C.5 Sizing accuracy

Poor sizing accuracy results in the mean and/or the modal diameter of a particle size distribution being determined incorrectly.

C.6 Counting efficiency

Counting efficiency directly affects the measurement of particle number concentration and consequently of the reported particle size distribution.

C.7 Effective detection flow rate

The tolerances and/or the variations of the effective detection flow rate directly affect the determination of particle number concentration.