
Dentistry — Periodontal probes —

Part 1:

General requirements

Médecine bucco-dentaire — Sondes parodontales —

Partie 1: Exigences générales



STANDARDSISO.COM : Click to view the full PDF of ISO 21672-1:2012



COPYRIGHT PROTECTED DOCUMENT

© ISO 2012

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 21672-1 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 4, *Dental instruments*.

ISO 21672 consists of the following parts, under the general title *Dentistry — Periodontal probes*:

- *Part 1: General requirements*
- *Part 2: Designation*

STANDARDSISO.COM : Click to view the full PDF of ISO 21672-1:2012

Dentistry — Periodontal probes —

Part 1: General requirements

1 Scope

This part of ISO 21672 specifies general requirements and test methods for periodontal probes.

It is applicable to periodontal probes made of austenitic and martensitic stainless steel.

It is not applicable to periodontal probes with working ends made completely of plastics, nor to HAUER probes and periodontal probes with a defined probing force.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, *Dentistry — Vocabulary*

ISO 6507-1, *Metallic materials — Vickers hardness test — Part 1: Test method*

ISO 6892-1, *Metallic materials — Tensile testing — Part 1: Method of test at room temperature*

ISO 7153-1, *Surgical instruments — Metallic materials — Part 1: Stainless steel*

ISO 13402, *Surgical and dental hand instruments — Determination of resistance against autoclaving, corrosion and thermal exposure*

3 Terms, definitions and symbols

3.1 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942 and the following apply.

3.1.1

periodontal probe

dental hand instrument designed to measure subgingival pocket depth, used in dentistry for diagnostic purposes and assessment of the condition of periodontal pockets

3.1.2

furcation probe

periodontal probe designed to probe the furcations of premolars and molars

3.2 Symbols

b_1 blade thickness

b_2 blade width

d_1 diameter of spherical or round end

d_2 diameter of shank

h_1 blade height

h_2 shank height

R_1 radius of shank

R_2 radius of working end


α blade angle

β offset angle

4 Classification


For the purposes of this part of ISO 21672, periodontal probes are classified according to their intended function and their shapes into the following five types:

- Type 1, in accordance with Figure 1;
- Type 2, in accordance with Figure 2;
- Type 3, in accordance with Figure 3;
- Type 4, in accordance with Figure 4;
- Type 5, furcation probe, in accordance with Figure 5.




a Enlarged view of the working end.

Figure 1 — Dimensions of Type 1 probes



- ^a Enlarged view of the two possible versions for the working end: spherical (d_1) or oval (b_1 , b_2).

Figure 2 — Dimensions of Type 2 probes



- ^a Enlarged view of the two possible versions for the working end: spherical (d_1) or oval (b_1 , b_2).

Figure 3 — Dimensions of Type 3 probes




Figure 4 — Dimensions of Type 4 probes




Figure 5 — Dimensions of Type 5 probes

5 Requirements

5.1 Design and dimensions

Periodontal probes shall have one of the designs shown in Figure 1 to Figure 5 and the dimensions given in Table 1.

Table 1 — Dimensions for periodontal probes

Dimensions in millimetres

Angles in degrees

Type of probe	Quantity and tolerance									
	b_1^a $\pm 0,1$	b_2^a $\pm 0,1$	d_1 $\pm 0,05$	d_2 $\pm 0,05$	h_1 ± 2	h_2 ± 2	R_1 ± 1	R_2 ± 1	α ± 5	β ± 5
1	—	—	0,6	0,4	6,6	7,5	2	—	60	—
2	0,5	1,5	0,5	—	12,0	—	2	—	60	—
3	0,5	1,5	0,5	—	8,4	4,5	2	—	60	—
4	—	—	0,5	—	9,5	7,0	2	—	60	—
5	—	—	0,4	—	6,2	6,3	4,7	6,5	35	20

^a The working end of Type 2 and Type 3 has a cross-section that is either
 — circular, or
 — rectangular and flat with rounded corners.

5.2 Working end

5.2.1 Graduated scale

The working end of the periodontal probe shall have a graduated scale, in order to allow the measurement of periodontal pockets. The tolerances of each graduation, measured from the distal end of the tip, shall be $\pm 0,2$ mm. The graduated scale may be colour-coded or non-colour-coded.

The periodontal probes specified in Table 2 shall have a graduated scale in accordance with Table 2. Graduated scales for other periodontal probes are at the discretion of the manufacturer.

Table 2 — Graduated scales for periodontal probes


Dimensions in millimetres

Type	Pattern name	Figure	Graduated scale
1	WHO, PCP 11,5	Figure 1	3,5 – 5,5 – 8,5 – 11,5
2	Merritt, PCP 10	Figure 2	1 – 2 – 3 – 5 – 7 – 8 – 9 – 10
3	Goldman-Fox	Figure 3	1 – 2 – 3 – 5 – 7 – 8 – 9 – 10
4	PCP 11 or PCP 12 or PCP N 22 or QOW	Figure 4	3 – 6 – 8 – 11 or 3 – 6 – 9 – 12 or 2 – 4 – 6 – 8 – 10 – 12
5	Nabers	Figure 5	3 – 6 – 9 – 12

5.2.2 Shape of the working end

The shape of the working end of periodontal probes shall have one of the following profiles (see Figure 6):


- a) spherical end (circular cross-section);
- b) domed end (rectangular cross-section);
- c) flat end with rounded corners (rectangular cross-section).



a) Spherical end



b) Domed end



c) Flat end with rounded corners

Figure 6 — Profiles of the working end

5.2.3 Material of the working end

The working ends of periodontal probes shall be made of stainless steel complying with ISO 7153-1. Other stainless steel materials may also be used provided the periodontal probes made of these materials meet the requirements of Clause 5.

5.2.4 Hardness and tensile strength of the working end

5.2.4.1 Martensitic stainless steel — Vickers hardness

The Vickers hardness of the working end of the periodontal probe made of martensitic stainless steel shall be within the following range: 500 HV 1 to 650 HV 1.

The Vickers hardness shall be measured in accordance with ISO 6507-1 or, alternatively, ISO 15087-1:1999, Annex B.

5.2.4.2 Austenitic stainless steel — Tensile strength

The tensile strength of the working end of the periodontal probe made of austenitic stainless steel shall be within the following range: 1 600 MPa to 1 800 MPa based on wire diameter.

The wire diameter is at the discretion of the manufacturer.

The tensile strength shall be measured in accordance with ISO 6892-1.

5.2.5 Rigidity

The working end of the periodontal probe shall not fracture when a rigidity test is performed in accordance with 6.4.

The tip of the working end shall not exhibit permanent deformation of more than 0,05 mm measured along the axis of the handle.

5.3 Handle

The material of the handle of the periodontal probe, selected at the discretion of the manufacturer, shall meet the requirements of 6.2.

5.4 Surfaces

5.4.1 All surfaces

All surfaces shall be visibly free from pores, crevices, grinding marks, residual scales, acid, grease and residual grinding and polishing materials.

Surfaces shall be tested in accordance with 6.1.

5.4.2 Satin finish

If a satin finish is used on any part of the instrument, it shall be both uniform and smooth and shall reduce glare.

Satin finishes shall be tested in accordance with 6.1.

5.5 Maximum overall length

The maximum overall length of the periodontal probe shall be less than 178 mm.

NOTE This size limitation is intended to make the instrument compliant with current designs of autoclave trays and cassettes.

5.6 Resistance to reprocessing procedures

When tested in accordance with 6.2.2, there shall be no visible signs of deterioration. When tested in accordance with 6.3, there shall be no deterioration in tensile load or torque and no change in hardness value or rigidity.

5.7 Resistance to corrosion and thermal exposure

When tested in accordance with 6.2.3, there shall be no visible signs of corrosion. When tested in accordance with 6.3, there shall be no deterioration in tensile load or torque and no change in hardness value or rigidity.

5.8 Union of working end and handle

The tested working end of the periodontal probe shall not separate from the handle or become loose under tensile load when tested in accordance with 6.3.1, and under torque when tested in accordance with 6.3.2.

6 Test methods

6.1 Visual inspection

Conduct visual inspection at normal acuity at no more than $\times 10$ magnification.

6.2 Resistance tests

6.2.1 Test sequence and cycles

Perform one of the two tests given in 6.2.2 and 6.2.3, in one continuous sequence of operation, for five cycles. After completing the test, rub the instrument vigorously with a cloth to remove blemishes.

6.2.2 Autoclave test

Carry out the autoclave test as specified in ISO 13402.

6.2.3 Boiling water test and thermal exposure test

Carry out the boiling water test and the thermal exposure test as specified in ISO 13402.

NOTE The combination of the boiling water test and thermal exposure test is used to evaluate the leakage resistance of hollow handles and the tolerance to dry-heat sterilization methods.

6.3 Test of union of working end and handle

6.3.1 Under tensile force

Subject the union between the working end and handle to a tensile force of 50 N, applied in the direction parallel to the centre line of the handle, for a duration of at least 5 s.

6.3.2 Under torque

Subject the union between the working end and the handle to a torque of 5 N-cm for a duration of at least 5 s.

6.4 Rigidity test

Subject the working end to a load of 50 g, applied in the direction parallel to the centre line of the handle, with the test apparatus illustrated in Figure 7 for a duration of 60 s.