
**Dentistry — Membrane materials for
guided tissue regeneration in oral and
maxillofacial surgery — Contents of a
technical file**

*Art dentaire — Membranes pour régénération de tissus en chirurgie
buccale et maxillo-faciale — Contenu du dossier technique*



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 22803:2004

© ISO 2004

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 22803 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 8, *Dental implants*.

Introduction

Different materials used for the preservation of masticatory function, such as dental restorative materials and dental implants are subject to standards and regulations, either in existence or in preparation, designed to evaluate the performance of these products.

Membrane materials for periodontal tissue reconstruction in oral and maxillofacial surgery are not covered by the procedures for evaluating and testing dental restorative materials and dental implants, thus it is necessary to develop a new International Standard for these materials.

The aim of this International Standard is to define the content of a technical file that demonstrates safety and effectiveness of membrane materials used in oral and maxillofacial surgery.

STANDARDSISO.COM : Click to view the full PDF of ISO 22803:2004

Dentistry — Membrane materials for guided tissue regeneration in oral and maxillofacial surgery — Contents of a technical file

1 Scope

This International Standard gives the requirements for a technical file on the evaluation of the chemical, physical, mechanical, biological and clinical aspects and behaviour of membrane materials, whether resorbable, partially resorbable or non-resorbable, which are used

- for guided tissue regeneration in oral and maxillofacial surgery to correct a morphological defect or abnormality,
- in contact with teeth and/or dental implants,
- for prevention of epithelial migration in periodontal surgery,
- for the augmentation of bone prior to the planned insertion of dental implants,
- and/or for augmentation of bone for stabilization of dental prostheses.

This International Standard is not applicable to materials whose primary intended use is to deliver a medicinal product, autografts and allografts, or materials intended to act through pharmacological, immunological or metabolic means.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, *Dentistry — Vocabulary*¹⁾

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing*

ISO 10993-7, *Biological evaluation of medical devices — Part 7: Ethylene oxide sterilization residuals*

ISO 11134, *Sterilization of health care products — Requirements for validation and routine control — Industrial moist heat sterilization*

ISO 11135, *Sterilization of health care products — Ethylene oxide — Requirements for development, validation and routine control of a sterilization process for medical devices*

ISO 11137, *Sterilization of health care products — Requirements for validation and routine control — Radiation sterilization*

ISO 11607, *Packaging for terminally sterilized medical devices*

ISO 14155-1, *Clinical investigation of medical devices for human subjects — Part 1: General requirements*

ISO 14937, *Sterilization of health care products — General requirements for characterization of a sterilizing agent and the development, validation and routine control of a sterilization process for medical devices*

1) Revision of ISO 1942-1:1989, ISO 1942-2:1989, ISO 1942-3:1989, ISO 1942-4:1989 and ISO 1942-5:1989.

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO 15223, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied*

EN 1041, *Information supplied by the manufacturer with medical devices*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942 and the following apply.

3.1
periodontal tissue
all tissues constituting the dental periodontium, i.e. alveolar bone, gingival tissue, periodontal ligament and cementum

3.2
biocompatibility
<material action> capacity of a material to fulfill its function with an appropriate response for a specific application in the recipient

3.3
biocompatibility
<material reaction> quality of being accepted in a specific living environment without adverse or unwanted side effects

[ISO 1942-1:1989/Amd.5:1993, definition 1.200]

3.4
biomaterial
<general purpose> material intended to interface with the biological system to evaluate, treat, augment or replace tissue, organ or function of the organism

3.5
biomaterial
<tailored preparation> material specially prepared and/or presented to exhibit bioacceptability, biocompatibility or positive biocompatibility

[ISO 1942-1:1989/Amd.5:1993, definition 1.204]

NOTE The implantable materials referred to in this International Standard are all biomaterials.

3.6
membrane material
medical device specifically prepared as a material which, when placed into tissue, carries out a barrier function

NOTE The sheet may be occlusive or selectively permeable to cells, macromolecules and/or fluid.

3.7
barrier
structure which, when placed into tissue, prevents the intermixing of the cell population on each side of the structure and/or prevents the prolapse of tissue

3.8
packing
surgical placement of a biomaterial to fill an intrabony cavity or defect

3.9**augmentation**

surgical placement of autogenous bone and/or of a biomaterial, resorbable or non-resorbable, to increase the volume of a bone or bridging of a defect

3.10**resorbable**

ability of a membrane material to undergo progressive elimination by cellular activity and/or dissolution in a biological environment

3.11**tissue regeneration**

reproduction or reconstruction of a lost or injured tissue by induction, conduction or healing process

3.12**guided tissue regeneration****GTR**

formation of tissue among which the orientation, the function, the volume and the place are pre-modelled by an exogenous mean

3.13**guided bone regeneration****GBR**

bone formation specifically obtained by GTR principles

4 General description

Membrane materials are widely used in periodontology and oral and maxillofacial surgery, and have characteristics which are unique to these applications, for example:

- a) placement in contact with teeth and their supporting tissues;
- b) prevention of oral mucosal epithelial migration into surgically treated defects;
- c) predictable loss of structural integrity and mechanical properties over time in oral sites into which dental implants are to be subsequently placed;
- d) known behaviour of the material should it become inadvertently exposed to the oral or paranasal cavities subsequent to placement;
- e) augmentation of the volume or dimension of bone for enhanced stabilization of dentures or for placement of dental implants.

The development of membrane materials shall be considered with regard to the properties required for the intended purpose, taking into account the effects of manufacture, handling, sterilization and storage. Possible reactions (intended or not) of membrane materials with human tissues and body fluids, other materials, other implants, substances, gases, radiation and electromagnetic fields shall be considered.

Membrane materials for periodontal tissue reconstruction in oral and maxillofacial surgery are used as either a barrier or a covering of packing materials.

During their use, these materials can induce an acute inflammatory reaction, which must be evaluated by specific tests.

5 Contents of a technical file

5.1 General

The contents of a technical file shall include the following information about the membrane material:

- a) details of its chemical composition;
- b) its intended performance;
- c) its preclinical and clinical evaluations;
- d) details of its manufacture, sterilization and packaging;
- e) certain other items of information necessary for the user.

5.2 Chemical composition

Refer to ISO 10993-13, ISO 10993-14, ISO 10993-15 and ISO 10993-18 for guidance. All polymer characterization shall be completed after recommended sterilization procedures have been applied.

5.3 Intended performance

The intended performance of a membrane material shall be described and documented by addressing the following:

- general description of the product;
- functional characteristics: resorbable, partially resorbable or non-resorbable;
- typical intended applications: simple barrier (in this case give physical and mechanical characteristics, including deformability);
- completely submersed or not during use;
- intended conditions of use;
- whether the device is intended to be used with fixation;
- included physical additives (e.g. titanium);
- included chemical additives (e.g. mineral salts);
- included material degradation when undergoing multiple cycles of reprocessing;
- reference to published standards to which the device conforms, with particular regard to safety.

Account should be taken of

- published standards,
- published clinical and scientific literature,
- validated test results.

5.4 Preclinical and clinical evaluation

5.4.1 General

Following an appropriate risk analysis as part of a risk management programme in accordance with ISO 14971, membrane materials shall be evaluated to demonstrate that their intended performance is achieved in GTR and/or GBR. The extent to which the intended performance has been achieved shall be determined and documented. Safety shall be demonstrated by preclinical and clinical evaluations and testing, as appropriate.

5.4.2 Preclinical evaluation

All details of the preclinical evaluation shall be provided. Preclinical evaluation shall include both the physical and biological properties of the membrane materials.

Among its physical properties, the physical strength of the membrane in an artificial medium over a period of time, including its tear resistance, shall be determined, using appropriate methods.

The preclinical biological evaluation shall be carried out in accordance with ISO 10993-1 by

- a compilation and critical analysis of the relevant scientific literature; and
- if necessary, analysis of data obtained from tests performed.

Laboratory and, if necessary, animal tests should be designed to establish that the materials meet the requirements for their use in periodontal, oral and maxillofacial surgery. Such tests should include *in vitro* and *in vivo* studies aimed at establishing the biocompatibility of the membranes with the periodontal tissues and bone, their ability to inhibit or guide the migration of oral epithelial cells, their biodegradability by the host tissues and their behaviour when exposed to the oral environment, including the ability of oral and/or nasal micro-organisms to colonise them. Evidence to show that the membrane is compatible with devices and substances that would be encountered during intended uses shall be included.

5.4.3 Clinical evaluation

Details of the clinical evaluation shall be provided.

Membrane materials shall undergo clinical evaluation by

- a) a compilation and critical analysis of the relevant literature covering the intended clinical use of the materials, and/or
- b) an analysis of data obtained from clinical investigations of the specific membrane material, especially with regard to evidence of new bone formation, bone resorption and/or periodontal tissue regeneration.

If a clinical investigation is carried out, it shall be managed in accordance with the requirements of ISO 14155-1.

5.5 Manufacture

Manufacturing processes shall be described and evidence of their validation shall be provided.

Membrane materials shall be manufactured in such a way that the intended performance is achieved.

5.6 Sterilization

5.6.1 Products supplied sterile

The sterility assurance level of terminally sterilized membrane materials shall be 10^{-6} or better. Sterilization processes shall be validated and routinely controlled.

If membrane materials are to be sterilized by ethylene oxide, ISO 11135 shall apply; by irradiation, ISO 11137; by steam, ISO 11134; by any other method, validation shall be carried out according to ISO 14937.

If the manufacturer states that resterilization is acceptable, the maximum permissible number of cycles shall be stated and at least one validated method shall be specified.

For membrane materials that are supplied non-sterile, the manufacturer shall specify at least one validated method of sterilization such that their functional safety is not adversely affected. If multiple sterilizations are not allowed, this shall be stated in the information provided by the manufacturer (see 5.8).