

---

---

## Dentistry — Dental amalgam

*Médecine bucco-dentaire — Amalgame dentaire*

STANDARDSISO.COM : Click to view the full PDF of ISO 24234:2021



STANDARDSISO.COM : Click to view the full PDF of ISO 24234:2021



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2021

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

|                                                                              |            |
|------------------------------------------------------------------------------|------------|
| <b>Foreword</b>                                                              | <b>v</b>   |
| <b>Introduction</b>                                                          | <b>vii</b> |
| <b>1 Scope</b>                                                               | <b>1</b>   |
| <b>2 Normative references</b>                                                | <b>1</b>   |
| <b>3 Terms and definitions</b>                                               | <b>2</b>   |
| <b>4 Requirements</b>                                                        | <b>3</b>   |
| 4.1 Chemical composition and purity of the dental amalgam alloy              | 3          |
| 4.2 Purity of the dental mercury                                             | 3          |
| 4.3 Foreign material and large particles in the dental amalgam alloy powder  | 3          |
| 4.4 Accuracy and variability of pre-proportioned masses                      | 4          |
| 4.4.1 For dental mercury sachets                                             | 4          |
| 4.4.2 For dental amalgam alloy tablets                                       | 4          |
| 4.5 Properties of the dental amalgam                                         | 4          |
| 4.5.1 General                                                                | 4          |
| 4.5.2 Creep                                                                  | 4          |
| 4.5.3 Dimensional changes during hardening                                   | 4          |
| 4.5.4 Compressive fracture stress at 2 h                                     | 4          |
| 4.5.5 Compressive fracture stress at 24 h                                    | 4          |
| 4.6 Appearance of the mixed dental amalgam before setting                    | 4          |
| 4.7 Corrosion resistance of the dental amalgam                               | 5          |
| <b>5 Sampling</b>                                                            | <b>5</b>   |
| <b>6 Test methods</b>                                                        | <b>5</b>   |
| 6.1 Chemical composition and purity of the dental amalgam alloy              | 5          |
| 6.1.1 Principle                                                              | 5          |
| 6.1.2 Test sample                                                            | 5          |
| 6.1.3 Apparatus                                                              | 5          |
| 6.1.4 Procedure                                                              | 5          |
| 6.1.5 Expression of results                                                  | 5          |
| 6.1.6 Report                                                                 | 6          |
| 6.2 Purity of the dental mercury                                             | 6          |
| 6.2.1 Principle                                                              | 6          |
| 6.2.2 Sample                                                                 | 6          |
| 6.2.3 Apparatus                                                              | 6          |
| 6.2.4 Procedure                                                              | 7          |
| 6.2.5 Expression of results                                                  | 7          |
| 6.2.6 Report                                                                 | 7          |
| 6.3 Foreign material and large particles in the dental amalgam alloy powder  | 7          |
| 6.3.1 Principle                                                              | 7          |
| 6.3.2 Test sample                                                            | 7          |
| 6.3.3 Apparatus                                                              | 8          |
| 6.3.4 Test procedure                                                         | 8          |
| 6.3.5 Expression of the results                                              | 8          |
| 6.3.6 Report                                                                 | 9          |
| 6.4 Determination of the accuracy and variability of pre-proportioned masses | 9          |
| 6.4.1 Principle                                                              | 9          |
| 6.4.2 Test sample                                                            | 9          |
| 6.4.3 Apparatus                                                              | 9          |
| 6.4.4 Test procedure                                                         | 9          |
| 6.4.5 Treatment of data                                                      | 10         |
| 6.4.6 Report                                                                 | 10         |
| 6.5 Properties of the dental amalgam                                         | 11         |
| 6.5.1 Principle                                                              | 11         |

|          |                                                                                                                                          |           |
|----------|------------------------------------------------------------------------------------------------------------------------------------------|-----------|
| 6.5.2    | Mould for the preparation of test-pieces for determining creep, dimensional change during hardening and compressive fracture stress..... | 11        |
| 6.5.3    | Sample.....                                                                                                                              | 14        |
| 6.5.4    | Test-piece production.....                                                                                                               | 14        |
| 6.5.5    | Procedure for the determination of creep.....                                                                                            | 16        |
| 6.5.6    | Procedure for the determination of dimensional change during hardening.....                                                              | 17        |
| 6.5.7    | Procedure for the determination of compressive fracture stress .....                                                                     | 19        |
| 6.6      | Appearance of the mixed dental amalgam before setting .....                                                                              | 20        |
| 6.6.1    | Principle .....                                                                                                                          | 20        |
| 6.6.2    | Apparatus.....                                                                                                                           | 20        |
| 6.6.3    | Test procedure .....                                                                                                                     | 21        |
| 6.6.4    | Expression of the results.....                                                                                                           | 21        |
| 6.6.5    | Report.....                                                                                                                              | 22        |
| 6.7      | Corrosion resistance of the dental amalgam.....                                                                                          | 22        |
| 6.7.1    | Principle .....                                                                                                                          | 22        |
| 6.7.2    | Sampling.....                                                                                                                            | 22        |
| 6.7.3    | Test procedure .....                                                                                                                     | 23        |
| 6.7.4    | Treatment of results.....                                                                                                                | 23        |
| 6.7.5    | Report.....                                                                                                                              | 23        |
| <b>7</b> | <b>Report .....</b>                                                                                                                      | <b>24</b> |
| <b>8</b> | <b>Marking and labelling .....</b>                                                                                                       | <b>24</b> |
| 8.1      | Information .....                                                                                                                        | 24        |
| 8.1.1    | General.....                                                                                                                             | 24        |
| 8.1.2    | Dental amalgam alloy products .....                                                                                                      | 24        |
| 8.1.3    | Dental mercury.....                                                                                                                      | 25        |
| 8.2      | Labelling for a package containing dental mercury .....                                                                                  | 25        |
| 8.3      | Labelling of the outer surface of package or container used for shipping dental mercury .....                                            | 26        |
| 8.4      | Manufacturer's instructions.....                                                                                                         | 26        |
| 8.5      | Precautionary notes .....                                                                                                                | 26        |
|          | <b>Bibliography .....</b>                                                                                                                | <b>27</b> |

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 1, *Filling and restorative materials*.

This third edition cancels and replaces the second edition (ISO 24234:2015), which has been technically revised.

The main changes compared to the previous edition are as follows.

- Pre-capsulated dental amalgam products have been removed from the scope of this document.
- A requirement for corrosion resistance has been added.
- In previous editions of this document, the presence of a limited number of foreign body particles in the dental amalgam alloy powder was permitted. Now, as a requirement, foreign body particles are not permitted to be present in the dental amalgam alloy powder.
- The roughness parameter used to specify the finish required on working surfaces of test-piece moulds has been changed from  $R_k$  to  $R_a$ .
- An instruction to lightly abrade the ends of the cylindrical test-pieces, if required for removing flash, has been deleted.
- The requirement for early compression strength has been altered. Measurement of the value is made at 2 h and not at 1 h.
- An additional four items of information have been added to each of the test reports.
- The edition number of the manufacturer's instructions and information, and the date of its introduction have been added as a requirement to the manufacturer's instructions.
- For each test method used to determine conformity to a requirement, a new subclause, "Principle", has been added in which a brief summary is present to explain the method adopted.

- For each test method used to determine conformity to a requirement, a new subclause, “Test report”, has been added.
- A new clause “7 Report” has been added which provides details of the evaluation that are to accompany a statement or claim of conformity to this document overall.

Any feedback or questions on this document should be directed to the user’s national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

STANDARDSISO.COM : Click to view the full PDF of ISO 24234:2021

## Introduction

Continuing concern about the use of dental mercury and a move in some countries to limit its use to pre-capsulated products led to the development of ISO 20749. The scope of ISO 20749 is restricted to pre-capsulated products alone. Consequently, it is appropriate to remove pre-capsulated dental amalgam products from the scope of this document.

Dental amalgam alloy supplied as a free-flowing powder and as tablets remain in use in some countries. For their use, dental mercury is required and the supply of dental mercury sachets (also referred to as pillows) continues to be consistent with the objective to restrict the supply of dental mercury only in sealed capsules containing a mass suitable for a single restoration. All such products are within the scope of this revision.

**NOTE** In some jurisdictions only pre-capsulated products are allowed to be used. ISO TC 106, *Dentistry*, must consider global use and not restrict the standards it produces to the position prevailing in individual states or regional blocks. For as long as product types within the scope of this document are in legal use in other nations, this standard will continue to be required.

Specific qualitative and quantitative test methods for demonstrating freedom from unacceptable biological hazard are not included in this document. It is recommended that, for the assessment of possible biological hazards, reference should be made to ISO 10993-1 and ISO 7405.

[STANDARDSISO.COM](https://standardsiso.com) : Click to view the full PDF of ISO 24234:2021

# Dentistry — Dental amalgam

## 1 Scope

This document specifies the requirements and test methods for dental amalgam alloy powder and dental mercury that are suitable for the preparation of dental amalgam together with the requirements and test methods for that dental amalgam and the requirements for packaging and marking.

NOTE Two of the requirements apply only to dental mercury (as supplied). All of the other requirements apply to the dental amalgam alloy (as supplied) and dental amalgam.

This document is not applicable to dental amalgam alloy powder and dental mercury supplied in a pre-capsulated form.

This document is not applicable to other metallic materials in which an alloy powder reacts with an alloy that is liquid at ambient temperature to produce a solid metallic material intended for dental restoration.

This document applies to products used to make dental amalgam restorations, supplied to the user in the following forms: dental amalgam alloy as a fine free flowing powder, or as a fine powder compacted into tablets and dental mercury in dental mercury sachets (sometimes referred to as dental mercury pillows). The mass of dental mercury in these sachets is limited to the amount required to make a small to medium-sized restoration in a single tooth.

This document is not applicable to dental mercury that is supplied in a primary container in an undivided mass that exceeds the amount suitable for a small to medium-sized restoration.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 286-2, *Geometrical product specifications (GPS) — ISO code system for tolerances on linear sizes — Part 2: Tables of standard tolerance classes and limit deviations for holes and shafts*

ISO 1942, *Dentistry — Vocabulary*

ISO 3310-1, *Test sieves — Technical requirements and testing — Part 1: Test sieves of metal wire cloth*

ISO 3864-2, *Graphical symbols — Safety colours and safety signs — Part 2: Design principles for product safety labels*

ISO 4287, *Geometrical Product Specifications (GPS) — Surface texture: Profile method — Terms, definitions and surface texture parameters*

ISO 7488, *Dentistry — Mixing machines for dental amalgam*

ISO 13897, *Dentistry — Dental amalgam reusable mixing-capsules*

ISO 15223-1:2016, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements*

ISO 23325, *Dentistry — Corrosion resistance of dental amalgam*

*UN Recommendations on the Transport of Dangerous Goods, Model Regulations*. United Nations, New York and Geneva, 21<sup>st</sup> Edition, 2019, eISBN 978-92-1-004112-6

*Globally Harmonized System of Classification and Labelling of Chemicals (GHS)*. United Nations, New York and Geneva, 8<sup>th</sup> Edition, 2019, eISBN 978-92-1-004083-9

### 3 Terms and definitions

For the purposes of this document, terms and definition given in ISO 1942 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

#### 3.1 dental amalgam alloy

alloy in fine particles, composed mainly of silver, tin and copper, which when mixed with *dental mercury* (3.2) produces a dental amalgam for dental restoration

[SOURCE: ISO 20749:2017, 3.1, modified — "for dental restoration" has been added at the end of the definition.]

#### 3.2 dental mercury

mercury supplied for use in the preparation of dental amalgam

[SOURCE: ISO 20749:2017, 3.2]

#### 3.3 pre-capsulated product

product supplied in a sealed capsule that contains measured amounts of *dental amalgam alloy* (3.1) powder and *dental mercury* (3.2) with masses that are appropriate for the production of a mass of dental amalgam that is considered to be suitable for a single small or medium size restoration in a single tooth

Note 1 to entry: The dental amalgam alloy powder and dental mercury are separated by a barrier that is broken immediately prior to mixing, allowing their contact. The capsule remains sealed until mixing has been completed.

[SOURCE: ISO 20749:2017, 3.3]

#### 3.4 dental amalgam alloy tablet

quantity of *dental amalgam alloy* (3.1) powder that has been compacted to form a single entity for the purpose of providing a pre-dosed quantity of the alloy that, when mixed with an appropriate mass of *dental mercury* (3.2), produces a mass of dental amalgam that is considered to be suitable for a single small or medium size restoration in a single tooth

Note 1 to entry: During mixing, the tablet is intended to break apart, forming a fine powder.

[SOURCE: ISO/TS 20746:2016, 3.4]

#### 3.5 dental mercury sachet

dental mercury pillow  
measured quantity of *dental mercury* (3.2) supplied in a sachet (for use in a reusable mixing capsule) in a mass that, when mixed with an appropriate mass of *dental amalgam alloy* (3.1) powder, produces a mass of dental amalgam that is considered to be suitable for a single small or medium size restoration in a single tooth

Note 1 to entry: The sachet is intended to rupture during mixing to allow the dental mercury to come into contact with the dental amalgam alloy powder.

[SOURCE: ISO/TS 20746:2016, 3.5, modified — "dental amalgam alloy" has been replaced with "dental amalgam alloy powder" in the definition and the term "dental mercury pillow" has been added.]

### 3.6

#### **mixing machine for dental amalgam**

DEPRECATED: amalgamator

electrically powered mixing machine that operates using an oscillating action for mixing *dental amalgam alloy* (3.1) and *dental mercury* (3.2) (in a capsule) to produce a dental amalgam

[SOURCE: ISO/TS 17988: 2020, 3.12]

## 4 Requirements

### 4.1 Chemical composition and purity of the dental amalgam alloy

The manufacturer shall declare every element that is present in a concentration greater than, or equal to a mass fraction of 0,1 %. All alloying elements present in concentrations greater than a mass fraction of 0,5 % shall be given by name with mass fraction values rounded to the nearest whole percentage point. Alloying elements that are present in concentrations between a mass fraction of 0,1 % and 0,5 % shall be named without a percentage value.

Test in accordance with 6.1.

The chemical composition shall comply with Table 1.

The total mass fraction for other elements present in concentrations greater than a mass fraction of 0,01 % but below a mass fraction of 0,1 % that are not declared as alloying elements shall not exceed a mass fraction of 0,1 %.

**Table 1 — Requirements for chemical composition of the dental amalgam alloy**

| Element   | Mass fraction<br>% |
|-----------|--------------------|
| Silver    | ≥40                |
| Tin       | ≤32                |
| Copper    | ≤30                |
| Indium    | ≤5                 |
| Palladium | ≤1                 |
| Platinum  | ≤1                 |
| Zinc      | ≤2                 |
| Mercury   | ≤3                 |

### 4.2 Purity of the dental mercury

Elements other than dental mercury shall not be present in a concentration greater than a mass fraction of 0,01 % in total. Test in accordance with 6.2.

### 4.3 Foreign material and large particles in the dental amalgam alloy powder

When conformity to this requirement is determined in accordance with 6.3, the proportion of the dental amalgam alloy powder that occurs as particles that have a size greater than 150 µm shall not exceed a mass fraction of 0,1 %.

When tested in accordance with 6.3, no particles of foreign matter shall be found on the sieve.

## 4.4 Accuracy and variability of pre-proportioned masses

### 4.4.1 For dental mercury sachets

The arithmetic mean of the mass of the dental mercury in the sachet shall be within  $\pm 2,0$  % of the manufacturer's stated mass, when tested in accordance with 6.4.

The coefficient of variation of the mass of the dental mercury in the sachets shall not exceed 1,5 %, when tested in accordance with 6.4.

### 4.4.2 For dental amalgam alloy tablets

The arithmetic mean of the mass of the dental amalgam alloy tablet shall be within  $\pm 2,0$  % of the manufacturer's stated mass, when tested in accordance with 6.4.

The coefficient of variation of the mass of the dental amalgam alloy tablets shall not exceed 1,5 %, when tested in accordance with 6.4.

## 4.5 Properties of the dental amalgam

### 4.5.1 General

**Table 2 — Properties of the dental amalgam**

| Maximum creep | Permitted dimensional change during hardening | Minimum compressive fracture stress at 2 h | Minimum compressive fracture stress at 24 h |
|---------------|-----------------------------------------------|--------------------------------------------|---------------------------------------------|
| %             | %                                             | MPa                                        | MPa                                         |
| 2,0           | -0,10 to +0,15                                | 100                                        | 350                                         |

### 4.5.2 Creep

When conformity to this requirement is determined in accordance with 6.5, the results for either three out of three, or four out of five test-pieces shall meet the requirement in Table 2.

### 4.5.3 Dimensional changes during hardening

When conformity to this requirement is determined in accordance with 6.5, the results for at least four out of five test-pieces shall meet the requirement in Table 2.

### 4.5.4 Compressive fracture stress at 2 h

When conformity to this requirement is determined in accordance with 6.5, the results for at least four out of five test-pieces or eight out of 10 test-pieces shall meet the requirement in Table 2.

### 4.5.5 Compressive fracture stress at 24 h

When conformity to this requirement is determined in accordance with 6.5, the results for at least four out of five test-pieces or eight out of 10 test-pieces shall meet the requirement in Table 2.

## 4.6 Appearance of the mixed dental amalgam before setting

When conformity to this requirement is determined in accordance with 6.6, the dental amalgam alloy and dental mercury being mixed according to the manufacturer's instructions, the dental amalgam shall form a coherent plastic mass with a shiny surface before packing and remain a coherent body after packing is completed.

#### 4.7 Corrosion resistance of the dental amalgam

When conformity to this requirement is determined in accordance with 6.7, the mean value (in newtons) of 10 valid results for corrosion test-pieces shall not be less than 80 % of the mean value (in newtons) of 10 valid results for control test-pieces.

### 5 Sampling

Procure material in packages that have been produced for retail and that are from a single lot.

To evaluate a dental amalgam alloy, procure a mass of dental amalgam alloy sufficient to conduct all the testing needed to evaluate the alloy itself and to make the required number of test-pieces, including the maximum number of test-pieces allowed to replace any that are rejected. To make the dental amalgam test-pieces, procure a sufficient number of dental mercury sachets from a single lot that has been produced for retail. These sachets shall comply with the requirements for dental mercury of this document.

A minimum of 200 g is advisable. There is waste amalgam with the production of each test-piece and an allowance for this is needed.

To evaluate a dental mercury sachet product for conformity to the requirements for dental mercury (see 4.2 and 4.4.1) 30 sachets are required.

### 6 Test methods

#### 6.1 Chemical composition and purity of the dental amalgam alloy

##### 6.1.1 Principle

Chemical analysis of the dental amalgam alloy using an instrumented technique for metallic materials.

##### 6.1.2 Test sample

10 g of dental amalgam alloy powder or dental amalgam alloy tablets, as appropriate.

##### 6.1.3 Apparatus

**Recognized, instrumented analytical instrument**, with sensitivity adequate to determine the composition of the dental amalgam alloy for each of the elements declared by the manufacturer in compliance with 4.1.

NOTE Inductively coupled plasma atomic emission spectroscopy (ICP-AES) is an example of a suitable analytical procedure.

##### 6.1.4 Procedure

Determine the composition of the dental amalgam alloy for the elements declared by the manufacturer in compliance with 4.1. Other elements may be detected during the analysis, being undeclared or impurities. Determine the concentration of each of these as a mass fraction percentage.

##### 6.1.5 Expression of results

Record all alloying elements detected in concentrations greater than a mass fraction of 0,01 % and their mass fraction percentages.

For other elements that are detected in concentrations greater than a mass fraction of 0,01 % and below a mass fraction of 0,1 % but are not alloying elements (declared as such by the manufacturer in

compliance with [4.1](#)), sum these values and record the sum as the mass fraction percentage of other elements. For an element, that is not a declared alloying element detected in a concentration greater than a mass fraction of 0,1 %, record this value and the name of the element.

### 6.1.6 Report

#### 6.1.6.1 Test report

A test report shall be prepared. At least the following information shall be included:

- a) the name of the dental amalgam alloy product and its lot number;
- b) the name and address of the manufacturer;
- c) the International Standard used (i.e. ISO 24234:2021);
- d) analytical method used;
- e) any irregularities in the test procedure;
- f) the mass fraction percentages for those elements that are alloying elements according to [Table 1](#) and declared as such by the manufacturer, as recorded in [6.1.5](#);
- g) if any other element is declared by the manufacturer as an alloying element, report this and its mass fraction percentage as recorded in [6.1.5](#);
- h) each undeclared element found in a concentration greater than a mass fraction of 0,1 % by name and the mass fraction percentage as recorded in [6.1.5](#);
- i) the sum of the mass fraction percentages of undeclared elements present in concentrations greater than a mass fraction of 0,01 % as recorded in [6.1.5](#);
- j) the name and address of the organization responsible for the testing (e.g. test house, university, department of manufacturer);
- k) the date of testing.

#### 6.1.6.2 Conformity

Report whether the product does or does not conform with the requirement for composition and purity of the dental amalgam alloy in accordance with [4.1](#)

## 6.2 Purity of the dental mercury

### 6.2.1 Principle

Chemical analysis of the dental mercury by using an instrumented technique for metallic materials.

### 6.2.2 Sample

One dental mercury sachet.

### 6.2.3 Apparatus

**6.2.3.1 Recognized, instrumented analytical instrument**, with sensitivity adequate to determine elements present as impurities in dental mercury, in compliance with [4.2](#).

NOTE ICP-AES is an example of a suitable analytical procedure.

**6.2.3.2 Surgical scalpel.****6.2.3.3 Watch glass.****6.2.4 Procedure**

Cut open the dental mercury sachet and empty its contents onto the watch glass.

Determine the purity of the dental mercury in compliance with [4.2](#) by using the analytical instrument to determine the concentration of any element (other than mercury) that is present in a concentration greater than a mass fraction of 0,000 5 %. Record these elements and their concentrations as mass fraction percentages.

**6.2.5 Expression of results**

Other than the value for dental mercury, sum the values recorded for elements that are present.

**6.2.6 Report****6.2.6.1 Test report**

A test report shall be prepared. At least the following information shall be included:

- a) the name of dental mercury product and its lot number;
- b) the name and address of the manufacturer;
- c) the International Standard used (i.e. ISO 24234:2021);
- d) the analytical method used;
- e) any irregularities in the test procedure;
- f) mass fraction percentages for those elements that are present in the dental mercury in concentrations greater than a mass fraction of 0,000 5 % (see [6.2.5](#));
- g) the sum of these concentrations as a mass fraction percentage;
- h) the name and address of the organization responsible for testing (e.g. test house, university, department of the manufacturer);
- i) the date of testing.

**6.2.6.2 Conformity**

Report whether the product does or does not conform to the requirement for purity of dental mercury in accordance with [4.2](#).

**6.3 Foreign material and large particles in the dental amalgam alloy powder****6.3.1 Principle**

Foreign particles, separated from the dental amalgam alloy powder by sieving, are identified by visual inspection. The large dental amalgam alloy particles (defined as >150 µm in size) separated from the sample (a known mass of dental amalgam alloy powder) are weighed.

**6.3.2 Test sample**

A (10,0 ± 0,1) g sample of dental amalgam alloy powder.

### 6.3.3 Apparatus

**6.3.3.1 Chemical balance**, having a resolution and accuracy to 1 mg.

**6.3.3.2 Sieve**, having a mesh size 150 µm that conforms to ISO 3310-1 with collection pan and cover.

**6.3.3.3 Tweezers**, with pointed ends.

**6.3.3.4 Weighing boat**, or similar.

**6.3.3.5 Stereomicroscope**, set at ×10 magnification.

**6.3.3.6 Mixing machine for dental amalgam**, complying with ISO 7448 (for use with dental amalgam alloy tablets).

**6.3.3.7 Reusable dental amalgam mixing capsule**, complying with ISO 13897 (for use with dental amalgam alloy tablets).

### 6.3.4 Test procedure

For free-flowing dental amalgam alloy powder, weigh out a  $(10 \pm 0,1)$  g sample to an accuracy of 1 mg and record this as  $m_p$ .

For dental amalgam alloy tablets, place a tablet in the reusable dental amalgam mixing capsule. Break the tablet in the capsule into the constituent powder particles by using the mixing machine at the setting and for half of the mixing time recommended (to produce an acceptable dental amalgam) by the manufacturer of the tablet. If the manufacturer's recommendations include any other action to break-up the tablet (e.g. use a pestle), incorporate this at the appropriate time. Repeat this using a sufficient number of tablets to obtain  $(10 \pm 0,1)$  g of powder. Weigh this sample to an accuracy of 1 mg and record as  $m_p$ .

Place the sample on the sieve. Hold the sieve assembly (consisting of collecting pan, sieve and cover) in one hand and tap it gently against the other hand at a rate of approximately twice a second for 120 s. Use the stereomicroscope to inspect the sieve for any foreign material. Record the number of foreign material particles.

Remove these foreign particles. Then, transfer the dental amalgam alloy particles remaining on the sieve to the balance. Weigh the dental amalgam alloy particles to an accuracy of 1 mg and record as  $m_r$ .

### 6.3.5 Expression of the results

Calculate  $w$ , the proportion of the dental amalgam alloy powder present as particles that have a size greater than 150 µm (expressed as a percentage of the mass of the sample), as follows:

$$w = \frac{m_r}{m_p} \times 100(\%)$$

where

$m_r$  is the mass of dental amalgam alloy particles remaining on the sieve;

$m_p$  is the mass of the powder sample.

### 6.3.6 Report

#### 6.3.6.1 Test report

A test report shall be prepared. At least the following information shall be included:

- a) the name of the dental amalgam alloy product and its batch number;
- b) the name and address of the manufacturer;
- c) the International Standard used (i.e. ISO 24234:2021);
- d) the test method used;
- e) any irregularities in the test procedure;
- f) whether foreign material was found on the sieve and the number of these particles;
- g) the proportion of the dental amalgam alloy that is present as particles greater than 150 µm in size, expressed as a percentage of the mass of the test sample (see 6.3.5);
- h) the name and address of the organization responsible for the testing (e.g. test house, university, department of manufacturer);
- i) the date of testing.

#### 6.3.6.2 Conformity

Report whether the product does or does not conform to the requirement for foreign matter and for large particles in accordance with 4.3.

### 6.4 Determination of the accuracy and variability of pre-proportioned masses

#### 6.4.1 Principle

The value of both parameters is obtained by weighing.

#### 6.4.2 Test sample

25 dental mercury sachets, or 25 dental amalgam alloy tablets, as is appropriate.

#### 6.4.3 Apparatus

##### 6.4.3.1 Surgical scalpel (for dental mercury sachets).

##### 6.4.3.2 Watch glass.

##### 6.4.3.3 Chemical balance, having a resolution and accuracy to 1 mg.

#### 6.4.4 Test procedure

##### 6.4.4.1 Dental mercury sachets

Select 25 dental mercury sachets at random.

Determine the mass of the dental mercury in each of these as follows.

Remove all dental mercury from one of these sachets. Cut it open using the scalpel and empty the dental mercury onto the watch glass. Weigh this dental mercury and the empty sachet separately, to the nearest 1 mg. Record these masses.

Weigh the other 24 sachets with the dental mercury *in situ* individually to the nearest 1 mg. Subtract the mass of the empty sachet (found previously) to obtain the mass of dental mercury in each of these. Record these masses.

#### 6.4.4.2 Dental amalgam alloy tablets

Select 25 dental amalgam alloy tablets at random. Weigh these individually to the nearest 1 mg. Record the masses.

#### 6.4.5 Treatment of data

Determine the arithmetic mean and standard deviation for the set of 25.

Calculate the coefficient of variation,  $C_v$ :

$$C_v = \frac{s}{\bar{x}} \times 100(\%)$$

where

$s$  is the standard deviation;

$\bar{x}$  is the arithmetic mean.

#### 6.4.6 Report

##### 6.4.6.1 Test report

A test report shall be prepared. At least the following information shall be included:

- the name of the dental mercury or dental amalgam alloy tablet product (as appropriate) and its lot number;
- the name and address of the manufacturer;
- the International Standard used (i.e. ISO 24234:2021);
- the test method used;
- any irregularities in the test procedure;
- all values for the masses of dental mercury in the sachets or dental amalgam alloy tablets, each to 1 mg (see [6.4.4.1](#) or [6.4.4.2](#), as appropriate);
- the arithmetic mean, the standard deviation and the coefficient of variation;
- any unusual features observed;
- the name and address of the organization responsible for the testing (e.g. test house, university, department of manufacturer);
- the date of testing.

##### 6.4.6.2 Conformity

Report whether the product does or does not conform to the requirement for the accuracy and consistency of the pre-proportioned mass, in accordance with [4.4](#).

## 6.5 Properties of the dental amalgam


### 6.5.1 Principle

All three properties for the requirement described in 4.5 are determined using cylindrical test-pieces. To produce consistency in packing throughout the body of the test-piece and consistency from one to the next, a standardized procedure is used to produce these test-pieces.

### 6.5.2 Mould for the preparation of test-pieces for determining creep, dimensional change during hardening and compressive fracture stress

#### 6.5.2.1 General

The mould and its component parts are shown in Figures 1 to 5.





#### Key

- 1 holder
- 2 spacer No. 1
- 3 spacer No. 2
- 4 plunger No. 2
- 5 test-piece
- 6 die
- 7 plunger No. 1

NOTE The dimensions for each of the components are given in the figures that follow.

**Figure 1 — Vertical section through the mould for making cylindrical dental amalgam test-pieces, showing the assembled mould with a test-piece in place**

Dimensions in millimetres



**Key**

1 slot

**Figure 2 — Holder**

Dimensions in millimetres







**Figure 3 — Spacer No. 1 (left) and spacer No. 2 (right)**

Dimensions in millimetres




Figure 4 — Plunger No. 2 (left), the die (centre) and plunger No. 1 (right)

Dimensions in millimetres




Figure 5 — Cap used for the production of test-pieces used for the measurement of dimensional change during hardening

To assist the operator in judging whether the correct quantity of dental amalgam has been inserted into the die, for the test-piece to be within the permitted range for length [i.e.  $(8 \pm 1)$  mm], circumferential datum lines may be scribed at 9 mm, 11 mm and 13 mm from one end of plunger No. 1. This end (from which the distances to the scribed lines are measured) shall be in contact with the dental amalgam. Though such datum lines are not mandatory, their use is recommended.

The diameters of the plungers shall be subject to a shaft (or in this case a plunger) clearance (with a tolerance) of h7 according to ISO 286-2. For a plunger that is nominally 4,000 mm in diameter, its diameter shall be between 0 µm and 18 µm less than 4,000 mm. Thus, the diameter of the plunger is to be between 3,982 mm and 4,000 mm.

The diameter of the hole in the die is subject to a clearance (with a tolerance) of F7 according to ISO 286-2. For a hole that is nominally 4,000 mm in diameter, its diameter shall be between 10 µm and 20 µm more than 4,000 mm. Thus, the diameter of the hole shall be between 4,010 mm and 4,020 mm.

#### 6.5.2.2 Materials and working surface finishing for construction of the apparatus

Make the holder, the spacers and the cap of cold-rolled or stainless steel. Make the die and the plungers of hardened tool steel or hardened stainless steel. Hone the working surfaces of the die and the plungers to an arithmetic mean roughness ( $R_a$ ) not greater than 6,3 µm when tested in accordance with ISO 4287.

#### 6.5.2.3 Assembly of the apparatus

For the production of creep and compressive strength test-pieces, assemble the holder, spacers No. 1 and No. 2, the die and plunger No. 2 as shown in [Figure 1](#). At this point in time, do not insert plunger No. 1.

NOTE Plunger No. 1 is inserted after the dental amalgam mix is placed in the die.

Particular measuring instruments used in the test for dimensional change during hardening (e.g. interferometers) may require an impression on the end surface of the test-piece. It is produced by the cap that is shown in [Figure 5](#). For the production of test-pieces for the measurement of dimensional change during hardening, include the cap in the assembly if this is appropriate for the measuring instrument that is to be used. In that case, position the cap on top of the plunger No. 2.

#### 6.5.3 Sample

Sufficient dental amalgam alloy powder or tablets to produce the required number of test-pieces for each property determination. An appropriate number of dental mercury sachets.

#### 6.5.4 Test-piece production

##### 6.5.4.1 Temperature

Prepare test-pieces at  $(23 \pm 2)^\circ\text{C}$ .

##### 6.5.4.2 Apparatus

**6.5.4.2.1 Dental amalgam mixing machine**, complying with ISO 7448 and recommended by the manufacturer of the dental amalgam alloy product.

**6.5.4.2.2 Reusable dental amalgam mixing-capsule**, complying with ISO 13897.

**6.5.4.2.3 Timer**, with an accuracy and resolution to 1 s.

**6.5.4.2.4 Tweezers**, with pointed ends.

**6.5.4.2.5 Hand-instrument for dental amalgam packing.**

**6.5.4.2.6 Light source**, with an illuminance  $\geq 1\,000$  lux.

**6.5.4.2.7 Mould**, shown in [Figures 1](#) to [5](#).

#### 6.5.4.2.8 Apparatus to apply $(176 \pm 13)$ N to the mould plunger No. 1.

6.5.4.2.9 Air oven, set at a temperature of  $(37 \pm 1)$  °C.

#### 6.5.4.3 Mixing

Use the ratio by mass of the dental amalgam alloy to the mass of dental mercury that is recommended by the manufacturer. Use a reusable mixing capsule (with a pestle, if needed) to contain the mix. Use any other mixing accessory that is required, as recommended by the manufacturer. If more than one mix is required to make the test-piece, produce these mixes simultaneously using equipment of the same type for each mix. However, if the last mix can be produced within the working time of the first mix, mixing these masses sequentially on a single piece of equipment is permitted.

Mix a mass of dental amalgam sufficient to make a 4 mm diameter cylindrical test-piece  $(8 \pm 1)$  mm in length.

NOTE The mass of a 4 mm diameter dental amalgam cylinder 8 mm in length is approximately 1,2 g.

Use the setting and the mixing time that are recommended by the manufacturer for the mass of dental amalgam alloy that is being mixed [see 8.4 b)].

#### 6.5.4.4 Packing

Using tweezers, place the coherent mass of mixed dental amalgam on top of the die cavity and insert immediately with several thrusts of a hand-instrument for dental amalgam packing. Do not express mercury during this insertion. Then insert plunger No. 2 and proceed, following the schedule in Table 3.

After ejection, the test-piece shall not be trimmed.

Inspect the surfaces of the test-piece for any defects. Use visual inspection without magnification. Carry out this inspection at an illuminance of at least 1 000 lux and at a distance not exceeding 250 mm. A person making the inspection shall have nominally normal visual acuity. Corrective (non-magnifying) non-tinted lenses may be worn. If the test-piece is defective, replace it.

Transfer the test-piece to air maintained at  $(37 \pm 1)$  °C.

**Table 3 — Schedule for the preparation of the cylindrical test-pieces**

| Procedure                                                                                                                                        | Time<br>s |
|--------------------------------------------------------------------------------------------------------------------------------------------------|-----------|
| End of mixing at                                                                                                                                 | 0         |
| Insert the mixed mass into the die cavity, then plunger No. 1 and apply a force of $(176 \pm 13)$ N to produce a pressure of $(14 \pm 1)$ MPa at | 30        |
| Release the force and remove spacer No. 2 at                                                                                                     | 45        |
| Reapply the force at                                                                                                                             | 50        |
| Re-release the force at                                                                                                                          | 90        |
| Carefully remove excess dental mercury and eject the test-piece at                                                                               | 120       |

NOTE 1 If the cap (see Figure 5) is not present in the assembled apparatus and plunger No. 1 has circumferential datum lines scribed on its cylindrical surface, the test-piece will have a length that is  $(8 \pm 1)$  mm if the 13 mm datum line alone can be seen.

NOTE 2 If the cap (see Figure 5) is present in the assembled apparatus and both 11 mm and 13 mm datum lines can be seen but the 9 mm line cannot, the test-piece will have a length that is  $(8 \pm 1)$  mm.

## 6.5.5 Procedure for the determination of creep

### 6.5.5.1 Apparatus

**6.5.5.1.1 Instrument for determining creep**, to apply and sustain a compressive stress of  $(36,0 \pm 0,2)$  MPa continuously for a period not less than 4 h. The instrument is to maintain the test-piece at a temperature of  $(37,0 \pm 0,5)$  °C during the test period. The accuracy of the creep measurement shall be 0,01 mm.

NOTE The application of  $(456,0 \pm 2,5)$  N force to a 4 mm diameter test-piece produces  $(36,0 \pm 0,2)$  MPa stress.

**6.5.5.1.2 Micrometer screw gauge**, or similar measuring instrument, with an accuracy of 0,01 mm;

**6.5.5.1.3 Air oven**, or incubator to maintain a temperature of  $(37 \pm 1)$  °C.

### 6.5.5.2 Test-pieces

Make five test-pieces. After ejection from the mould and inspection (see 6.5.4.4), immediately transfer the test-piece to air maintained at  $(37 \pm 1)$  °C. One hour after the test-piece has been removed from the mould take it from the oven or incubator and measure its length to determine whether this is acceptable  $(8 \pm 1)$  mm. If it is acceptable, return it to the oven or incubator. If it is not acceptable, reject and replace this test-piece.

Checking acceptability at this time is recommended to avoid a lengthy delay, should it be found to have an unacceptable length when this is measured at 7 d.

Store for  $(7,0 \pm 0,2)$  d.

A minimum of three test-pieces are tested and for each the creep stress is applied for 4 h. Because the tolerance in time for the application of the creep force is  $\pm 0,2$  d (i.e. 4,8 h) after 7 d of storage, it will be necessary to stagger the production of test-pieces if fewer than three sets of creep test apparatus are available.

### 6.5.5.3 Test procedure

At the end of the storage period, remove the test-piece from the oven or incubator and measure the length to the nearest 0,01 mm. Record this as the original length.

Directly after measuring the original length [i.e. at  $(7,0 \pm 0,2)$  d] apply a compressive force normally and uniformly over the cylinder ends (of the first test-piece) to produce a stress of  $(36,0 \pm 0,2)$  MPa. This stress is applied continuously for 4 h at a test temperature of  $(37,0 \pm 0,5)$  °C. Measure the change in length of the test-piece to an accuracy of 0,01 mm between  $(1,00 \pm 0,05)$  h and  $(4,0 \pm 0,1)$  h after the force is first applied. Record the new length of the test-piece.

Measure and test two more test-pieces.

If necessary, in accordance with 6.5.5.4, test both the remaining test-pieces.

### 6.5.5.4 Expression of the results

For each test-piece, calculate the creep strain,  $\varepsilon_c$ , as a percentage of the original length to the nearest 0,1 %, as follows:

$$\varepsilon_c = \frac{\Delta l}{l_0} \times 100(\%)$$

where

$\Delta l$  is the change in length between 1 h and 4 h, to an accuracy of 0,01 mm;

$l_0$  is the original length, to an accuracy of 0,01 mm.

If all three results conform to requirement in [Table 2](#), it is not necessary to test the other two test-pieces.

If two or all three test-pieces fail to conform to the requirement in [Table 2](#), the product fails to conform to the requirement for creep. As a consequence, it is not necessary to test the remaining two test-pieces.

If one of the three test-pieces fails to meet the requirement in [Table 2](#), test two more test-pieces.

Test no more than five test-pieces. Record all results.

### 6.5.5.5 Report

#### 6.5.5.5.1 Test report

A test report shall be prepared. At least the following information shall be included:

- a) the name of the dental amalgam alloy product and its lot number;
- b) the name and address of the manufacturer;
- c) the International Standard used (i.e. ISO 24234:2021);
- d) the test method used;
- e) any irregularity during test-piece production or testing;
- f) results for all test-pieces that were subjected to creep loading (see [6.5.5.4](#));
- g) any unusual features observed;
- h) the name and address of the organization responsible for the testing (e.g. test house, university, department of manufacturer);
- i) the date of testing.

#### 6.5.5.5.2 Conformity

Report whether the product does or does not conform to the requirement for creep, in accordance with [4.5.2](#).

### 6.5.6 Procedure for the determination of dimensional change during hardening

#### 6.5.6.1 Apparatus

**6.5.6.1.1 Instrument for measuring the dimensional change during hardening**, that does not subject the test-piece to a restraint greater than 20 mN during the test and with which the change in test-piece length can be measured to an accuracy of 0,5  $\mu$ m.

**6.5.6.1.2 Micrometer screw gauge**, or similar measuring instrument, with an accuracy of 0,01 mm.

**6.5.6.1.3 Air oven or incubator** to maintain a temperature of  $(37 \pm 1) ^\circ\text{C}$ . This is required only if the test-piece is removed from the apparatus after the initial measurement at 5 min and reinserted at 24 h to make the second measurement.

### 6.5.6.2 Test-pieces

Make five test-pieces.

### 6.5.6.3 Test procedure

Place the test-piece in the instrument immediately after its production. Measure the dimensional change that occurs between  $(5,0 \pm 0,1)$  min and  $(24,0 \pm 0,1)$  h from the end of mixing, to an accuracy of  $0,5 \mu\text{m}$ . Record this.

Maintain the test-piece at a temperature of  $(37 \pm 1) ^\circ\text{C}$  during the 24 h test period.

At  $(24,0 \pm 0,1)$  h, remove the test-piece from the instrument and measure the test-piece length to an accuracy of  $0,01 \text{ mm}$ . If the length of the test-piece is not within the specified length,  $(8 \pm 1) \text{ mm}$ , reject the result and replace the test-piece. Using this replacement test-piece, repeat the preceding test procedure in [6.5.6.3](#).

Rejection of a test-piece because its length is inadequate does not constitute a test failure and further replacements are permitted to obtain five test-pieces with acceptable lengths.

Test all five test-pieces.

During the 24 h test period, the test-piece may remain in the measuring instrument and the change in length monitored continuously, or it may be removed from the measuring instrument after the first measurement, held at  $37 ^\circ\text{C}$  without an applied force then returned to the measuring instrument for the second measurement.

If the test-piece is retained in the measuring instrument for the full test period, it is necessary to complete the test before making the next test-piece.

### 6.5.6.4 Expression of the results

Calculate the dimensional change during hardening,  $\varepsilon_d$ , as a percentage of the test-piece length to the nearest  $0,01 \%$ , as follows, and record these results:

$$\varepsilon_d = \frac{\Delta l_d}{l_d} \times 100(\%)$$

where

$\Delta l_d$  is the dimensional change between 5 min and 24 h;

$l_d$  is the length at 24 h.

### 6.5.6.5 Report

#### 6.5.6.5.1 Test report

A test report shall be prepared. At least the following information shall be included:

- the name of the dental amalgam alloy product and its lot number;
- the name and address of the manufacturer;
- the International Standard used (i.e. ISO 24234:2021);
- the test method used;
- results for all test-pieces (see [6.5.6.4](#));
- any irregularity during test-piece production or testing;

- g) any unusual features observed;
- h) the name and address of the organization responsible for the testing (e.g. test house, university, department of manufacturer);
- i) the date of testing.

#### 6.5.6.5.2 Conformity

Report whether the product does or does not conform to the requirement for dimensional change during hardening in accordance with [4.5.3](#).

### 6.5.7 Procedure for the determination of compressive fracture stress

#### 6.5.7.1 Apparatus

**6.5.7.1.1 Universal mechanical testing machine**, configured for compressive testing, with at least 10 kN frame and load cell capacity, and with a resolution and accuracy better than 10 N.

**6.5.7.1.2 Micrometer screw gauge**, or similar measuring instrument, with an accuracy of 0,01 mm.

**6.5.7.1.3 Air oven or incubator**, to maintain a temperature of  $(37 \pm 1) ^\circ\text{C}$ .

#### 6.5.7.2 Test-pieces

Make 20 test-pieces.

#### 6.5.7.3 Test procedure

##### 6.5.7.3.1 General

Immediately after ejection from the mould, transfer the test-piece to an air environment maintained at  $(37 \pm 1) ^\circ\text{C}$ . Store it in this environment until 30 min before it is to have force applied. At this time, remove the test-piece from the oven or incubator and place it on a clean surface in air at  $(23 \pm 2) ^\circ\text{C}$  to allow it to cool and equilibrate with the test temperature.

During this equilibration period, measure the diameter of the test-piece to an accuracy of 0,01 mm and record the value. Measure the length to determine whether the length of the test-piece is within the specified length of  $(8 \pm 1) \text{ mm}$ . If it is not, reject the test-piece and make a replacement.

Determine the compressive strength using the mechanical testing machine. During the test, maintain the test-piece at a temperature of  $(23 \pm 2) ^\circ\text{C}$ . Apply an increasing compressive force normally and uniformly over the circular ends of the test-piece at a constant crosshead speed of  $(0,5 \pm 0,1) \text{ mm/min}$ .

For each test-piece, record the force that produces failure and calculate the compressive fracture stress to the nearest 5 MPa.

##### 6.5.7.3.2 Compressive fracture stress at 2 h

Determine the compressive fracture stress of five test-pieces at  $(120 \pm 5) \text{ min}$  after mixing.

If only three test-pieces conform to the requirement in [Table 2](#) for compressive fracture stress at 2 h, determine the compressive fracture stress of five more test-pieces.

Test no more than 10 test-pieces at 2 h.

#### 6.5.7.3.3 Compressive fracture stress at 24 h

Determine the compressive fracture stress of five test-pieces at  $(24 \pm 1)$  h after mixing.

If only three test-pieces conform to the requirement in [Table 2](#) for the compressive fracture stress at 24 h, determine the compressive fracture stress of five more test-pieces.

Test no more than 10 test-pieces at 24 h.

#### 6.5.7.4 Expression of the results

Record the compressive fracture stress for all test-pieces that were loaded to failure.

#### 6.5.7.5 Report

##### 6.5.7.5.1 Test report

A test report shall be prepared. At least the following information shall be included:

- a) the name of the dental amalgam alloy product and its lot number;
- b) the name and address of the manufacturer;
- c) the International Standard used (i.e. ISO 24234:2021);
- d) the test method used;
- e) results for all test-pieces that were loaded to failure (see [6.5.7.4](#));
- f) any irregularity during test-piece production or testing;
- g) any unusual features observed;
- h) the name and address of the organization responsible for the testing (e.g. test house, university, department of manufacturer);
- i) the date of testing.

##### 6.5.7.5.2 Conformity

Report whether the product does or does not conform to the requirement for compressive fracture stress at 2 h, in accordance with [4.5.4](#).

Report whether the product does or does not conform to the requirement for compressive fracture stress at 24 h, in accordance with [4.5.5](#).

### 6.6 Appearance of the mixed dental amalgam before setting

#### 6.6.1 Principle

Visual inspection is used to determine the appearance of the surface of the mixed dental amalgam, whether a coherent mass exists initially and whether a coherent mass has been maintained during packing.

#### 6.6.2 Apparatus

**6.6.2.1 Reusable dental amalgam mixing capsule**, that complies with ISO 13897.