

---

---

**Geometrical product specification  
(GPS) — Surface texture: Areal —**

Part 606:

**Nominal characteristics of non-contact  
(focus variation) instruments**

*Spécification géométrique des produits (GPS) — État de surface:  
Surfacique —*

*Partie 606: Caractéristiques nominales des instruments sans contact  
(à variation de focale)*



STANDARDSISO.COM : Click to view the full PDF of ISO 25178-606:2015



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2015, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Ch. de Blandonnet 8 • CP 401  
CH-1214 Vernier, Geneva, Switzerland  
Tel. +41 22 749 01 11  
Fax +41 22 749 09 47  
copyright@iso.org  
www.iso.org

# Contents

|                                                                                         | Page      |
|-----------------------------------------------------------------------------------------|-----------|
| Foreword.....                                                                           | iv        |
| Introduction.....                                                                       | vi        |
| <b>1 Scope</b> .....                                                                    | <b>1</b>  |
| <b>2 Normative references</b> .....                                                     | <b>1</b>  |
| <b>3 Terms and definitions</b> .....                                                    | <b>1</b>  |
| 3.1 Terms and definitions related to all areal surface texture measurement methods..... | 2         |
| 3.2 Terms and definitions related to <i>x</i> - and <i>y</i> -scanning systems.....     | 9         |
| 3.3 Terms and definitions related to optical systems.....                               | 11        |
| 3.4 Terms and definitions related to optical properties of the workpiece.....           | 13        |
| 3.5 Terms and definitions specific to focus variation instruments.....                  | 13        |
| <b>4 Description of the influence quantities</b> .....                                  | <b>16</b> |
| 4.1 General.....                                                                        | 16        |
| 4.2 Overview.....                                                                       | 17        |
| 4.3 Influence quantities.....                                                           | 17        |
| <b>Annex A (informative) Components of a focus variation microscope</b> .....           | <b>19</b> |
| <b>Annex B (informative) Relation to the GPS matrix model</b> .....                     | <b>25</b> |
| <b>Bibliography</b> .....                                                               | <b>27</b> |

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT), see the following URL: [Foreword — Supplementary information](#).

The committee responsible for this document is ISO/TC 213, *Dimensional and geometrical product specifications and verification*.

ISO 25178 consists of the following parts, under the general title *Geometrical product specification (GPS) — Surface texture: Areal*:

- *Part 1: Indication des états de surface*
- *Part 2: Terms, definitions and surface texture parameters*
- *Part 3: Specification operators*
- *Part 6: Classification of methods for measuring surface texture*
- *Part 70: Material measures*
- *Part 71: Software measurement standards*
- *Part 72: Format de fichier XML x3p*
- *Part 601: Nominal characteristics of contact (stylus) instruments*
- *Part 602: Nominal characteristics of non-contact (confocal chromatic probe) instruments*
- *Part 603: Nominal characteristics of non-contact (phase-shifting interferometric microscopy) instruments*
- *Part 604: Nominal characteristics of non-contact (coherence scanning interferometry) instruments*
- *Part 605: Nominal characteristics of non-contact (point autofocus probe) instruments*
- *Part 606: Nominal characteristics of non-contact (focus variation) instruments*
- *Part 701: Calibration and measurement standards for contact (stylus) instruments*

The following parts are planned:

- *Part 73: Defects on material measures — Terms and definitions*
- *Part 600: Metrological characteristics for areal-topography measuring methods*
- *Part 607: Nominal characteristics of non-contact (imaging confocal microscopy) instruments*

STANDARDSISO.COM : Click to view the full PDF of ISO 25178-606:2015

## Introduction

This part of ISO 25178 is a geometrical product specification (GPS) standard and is to be regarded as a general GPS standard (see ISO/TR 14638). It influences the chain link 5 of the chain of standards on areal surface texture.

The ISO/GPS Masterplan given in ISO/TR 14638 gives an overview of the ISO/GPS system of which this part of ISO 25178 is a part of. The fundamental rules of ISO/GPS given in ISO 8015 apply to this part of ISO 25178 and the default decision rules given in ISO 14253-1 apply to specifications made in accordance with this part of ISO 25178, unless otherwise indicated.

For more detailed information of the relation of this part of ISO 25178 to other standards and the GPS matrix model, see [Annex B](#).

This part of ISO 25178 describes the metrological characteristics of focus variation microscopes designed for the measurement of surface topography maps.

For more detailed information on the focus variation technique, see [Annex A](#).

**NOTE** Portions of this part of ISO 25178, particularly the informative sections, describe patented systems and methods. This information is provided only to assist users in understanding the operating principles of focus variation. This part of ISO 25178 is not intended to establish priority for any intellectual property, nor does it imply a license to proprietary technologies described herein.

STANDARDSISO.COM : Click to view the full PDF of ISO 25178-606:2015

# Geometrical product specification (GPS) — Surface texture: Areal —

## Part 606: Nominal characteristics of non-contact (focus variation) instruments

### 1 Scope

This part of ISO 25178 defines the metrological characteristics of a particular non-contact method measuring surface texture using a focus variation (FV) sensor.

### 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3274:1996, *Geometrical Product Specifications (GPS) — Surface texture: Profile method — Nominal characteristics of contact (stylus) instruments*

ISO 4287:1997, *Geometrical Product Specifications (GPS) — Surface texture: Profile method — Terms, definitions and surface texture parameters*

ISO 10934-2:2007, *Optics and optical instruments — Vocabulary for microscopy — Part 2: Advanced techniques in light microscopy*

ISO 14978:2006, *Geometrical product specifications (GPS) — General concepts and requirements for GPS measuring equipment*

ISO 17450-1, *Geometrical product specifications (GPS) — General concepts — Part 1: Model for geometrical specification and verification*

ISO 25178-2:2012, *Geometrical product specifications (GPS) — Surface texture: Areal — Part 2: Terms, definitions and surface texture parameters*

ISO 25178-3:2012, *Geometrical product specifications (GPS) — Surface texture: Areal — Part 3: Specification operators*

ISO 25178-6:2010, *Geometrical product specifications (GPS) — Surface texture: Areal — Part 6: Classification of methods for measuring surface texture*

ISO 25178-601, *Geometrical product specifications (GPS) — Surface texture: Areal — Part 601: Nominal characteristics of contact (stylus) instruments*

ISO 25178-602, *Geometrical product specifications (GPS) — Surface texture: Areal — Part 602: Nominal characteristics of non-contact (confocal chromatic probe) instruments*

### 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 3274, ISO 4287, ISO 10934-2, ISO 17450-1, ISO 14978, ISO 25178-2, ISO 25178-3, ISO 25178-6, ISO 25178-601, ISO 25178-602, and the following apply.

### 3.1 Terms and definitions related to all areal surface texture measurement methods

#### 3.1.1

##### **areal reference**

component of the instrument that generates a reference surface with respect to which the surface topography is measured

#### 3.1.2

##### **coordinate system of the instrument**

right hand orthonormal system of axes  $(x, y, z)$  defined as:

- $(x, y)$  is the plane established by the *areal reference* (3.1.1) of the instrument (note that there are optical instruments that do not possess a physical areal guide);
- $z$ -axis is mounted parallel to the optical axis and is perpendicular to the  $(x, y)$  plane for an optical instrument

Note 1 to entry: See [Figure 1](#).

Note 2 to entry: Normally, the  $x$ -axis is the tracing axis and the  $y$ -axis is the stepping axis (this note is valid for instruments that scan in the horizontal plane).

Note 3 to entry: See also specification coordinate system [ISO 25178-2:2012, 3.1.2] and measurement coordinate system [ISO 25178-6:2010, 3.1.1].


#### 3.1.3

##### **measurement loop**

closed chain which comprises of all the components connecting the workpiece and the probe, e.g. the means of positioning, the work holding fixture, the measuring stand, the drive unit, and the *probing system* (3.5.3)

Note 1 to entry: See [Figure 1](#).

Note 2 to entry: The measurement loop will be subjected to external and internal disturbances that influence the measurement uncertainty.

**Key**

- 1 coordinate system of the instrument
- 2 measurement loop

**Figure 1 — Coordinate system and measurement loop of the instrument**

**3.1.4****real surface of a workpiece**

set of features which physically exist and separate the entire workpiece from the surrounding medium

Note 1 to entry: The real surface is a mathematical representation of the surface that is independent of the measurement process.

Note 2 to entry: See also mechanical surface [ISO 25178-2:2012, 3.1.1.1 or ISO 14406:2010, 3.1.1] and electromagnetic surface [ISO 25178-2:2012, 3.1.1.2 or ISO 14406:2010, 3.1.2].

Note 3 to entry: The electromagnetic surface considered for one type of optical instrument can be different from the electromagnetic surface for other types of optical instruments.

[SOURCE: ISO 17450-1:2011]

**3.1.5****surface probe**

device that converts the surface height into a signal during measurement

Note 1 to entry: In earlier International Standards, this was termed transducer.

**3.1.6****measuring volume**

range of the instrument stated in terms of the limits on all three coordinates measured by the instrument

Note 1 to entry: For areal surface texture measuring instruments, the measuring volume is defined by the measuring range of the x- and y- drive units and the measuring range of the z-probing system.

[SOURCE: ISO 25178-601:2010, 3.4.1]

### 3.1.7

#### response curve

$F_x, F_y, F_z$

graphical representation of the function that describes the relation between the actual quantity and the measured quantity

Note 1 to entry: See [Figure 2](#).

Note 2 to entry: An actual quantity in  $x$  (respectively  $y$  or  $z$ ) corresponds to a measured quantity  $x_M$  (respectively  $y_M$  or  $z_M$ ).

Note 3 to entry: The response curve can be used for adjustments and error corrections.

[SOURCE: ISO 25178-601:2010, 3.4.2]

### 3.1.8

#### amplification coefficient

$\alpha_x, \alpha_y, \alpha_z$

slope of the linear regression curve obtained from the *response curve* ([3.1.7](#))


Note 1 to entry: See [Figure 3](#).

Note 2 to entry: There will be amplification coefficients applicable to the  $x, y$ , and  $z$  quantities.

Note 3 to entry: The ideal response is a straight line with a slope equal to 1, which means that the values of the measurand are equal to the values of the input quantities.

Note 4 to entry: See also *sensitivity of a measuring system* (ISO/IEC Guide 99:2007, 4.12).


[SOURCE: ISO 25178-601:2010, 3.4.3, modified — Note 4 to entry has been added.]



#### Key

- 1 response curve
- 2 assessment of the linearity deviation by polynomial approximation
- 3 measured quantities
- 4 input quantities

Figure 2 — Example of a non-linear response curve

**Key**

- 1 measured quantities
- 2 input quantities
- 3 ideal response curve
- 4 linearization of the response curve of [Figure 2](#)
- 5 line from which the amplification coefficient  $\alpha$  (slope) is derived
- 6 local residual correction error

**Figure 3 — Example of the linearization of a response curve**

**3.1.9****instrument noise**
 $N_I$ 

internal noise added to the output signal caused by the instrument, if ideally placed in a noise-free environment

Note 1 to entry: Internal noise can be due to electronic noise, e.g. amplifiers, or to optical noise, e.g. stray light.

Note 2 to entry: This noise typically has high frequencies and it limits the ability of the instrument to detect small scale spatial wavelengths of the surface texture.

Note 3 to entry: The S-filter, according to ISO 25178-3:2012, can reduce this noise.

Note 4 to entry: For some instruments, instrument noise cannot be estimated because the instrument only takes data while moving.

**3.1.10****measurement noise**
 $N_M$ 

noise added to the output signal occurring during the normal use of the instrument

Note 1 to entry: Notes 2 and 3 of [3.1.9](#) apply as well to this definition.

Note 2 to entry: Measurement noise includes the *instrument noise* ([3.1.9](#)).

### 3.1.11

#### **surface topography measurement repeatability**

repeatability of topography map in successive measurements of the same surface under the same conditions of measurement

Note 1 to entry: Surface topography measurement repeatability provides a measure of the likely agreement between repeated measurements normally expressed as a standard deviation.

Note 2 to entry: See ISO/IEC Guide 99:2007, 2.15, and 2.21 for the general discussion of repeatability and related concepts.

Note 3 to entry: Evaluation of surface topography repeatability is a common method for determining the *measurement noise* ([3.1.10](#)).

### 3.1.12

#### **sampling interval in $x$ (respectively $y$ )**

$D_x$  ( $D_y$ )

distance between two adjacent measured points along the  $x$ -axis (respectively  $y$ -axis)

Note 1 to entry: In many microscopy systems, the sampling interval is determined through the optical magnification by the distance between sensor elements in a camera called pixels. For such systems, the terms pixel pitch and pixel spacing are often used interchangeably with the term sampling interval. Another term, pixel width, indicates a length associated with one side ( $x$  or  $y$ ) of the sensitive area of a single pixel and is always smaller than the pixel spacing. Yet another term, sampling zone, may be used to indicate the length or region over which a height sample is determined. This quantity could either be larger or smaller than the sampling interval.

### 3.1.13

#### **digitisation step in $z$**

$D_z$

smallest height variation along the  $z$ -axis between two ordinates of the extracted surface

### 3.1.14

#### **lateral resolution**

$R_l$

smallest distance between two features which can be detected

[SOURCE: ISO 25178-601:2010, 3.4.10]

### 3.1.15

#### **width limit for full height transmission**

$W_l$

width of the narrowest rectangular groove whose measured height remains unchanged by the measurement

[SOURCE: ISO 25178-601:2010, 3.4.11]

Note 1 to entry: Instrument properties such as the sampling interval in  $x$  and  $y$ , the digitization step in  $z$ , and the short wavelength cut-off filter can influence the *lateral resolution* ([3.1.14](#)) and the width limit for full height transmission.

Note 2 to entry: When determining this parameter by measurement, the depth of the rectangular groove should be close to that of the surface to be measured.

EXAMPLE 1 Measuring a grid, for which the grooves are wider than the width limit for full height transmission, leads to a correct measurement of the groove depth (see [Figure 4](#) and [Figure 5](#)).




Figure 4 — Grid with horizontal spacing where  $t$  is greater than or equal to  $W_l$




Figure 5 — Measurement of the grid in Figure 4; the spacing and depth of the grid are measured correctly

EXAMPLE 2 Measuring a grid, for which the grooves are narrower than the *width limit for full height transmission* (3.1.15), leads to an incorrect groove depth (see Figure 6 and Figure 7). In this situation, the signal is generally disturbed and may contain non-measured points.




Figure 6 — Grid with horizontal spacing  $t'$  smaller than  $W_l$



Figure 7 — Measurement of the grid in Figure 6; the spacing is measured correctly, but the depth is smaller ( $d' < d$ )

### 3.1.16

#### lateral period limit

$D_{\text{LIM}}$

the spatial period of a sinusoidal profile at which the height response of an instrument falls to 50%

Note 1 to entry: The lateral period limit is one metric for describing spatial or lateral resolution of a surface topography measuring instrument and its ability to distinguish and measure closely spaced surface features. Its value depends on the heights of surface features and on the method used to probe the surface. Typical values, mainly for noise suppression, are listed in ISO 25178-3:2012, Table 3, in comparison with the recommended values for short wavelength (s-filters), and sampling intervals.

Note 2 to entry: *Spatial period* is the same concept as *spatial wavelength* and is the *inverse of spatial frequency*.

Note 3 to entry: One factor related to the value of  $D_{\text{LIM}}$  for optical tools is, e.g. the *Rayleigh criterion* (3.3.7); another is the degree of focus of the objective on the surface.

Note 4 to entry: One factor related to the value of  $D_{\text{LIM}}$  for contact tools is the stylus tip radius,  $r_{\text{TIP}}$  (see ISO 25178-601).

Note 5 to entry: Other terms related to *lateral period limit* are *structural resolution* and *topographic spatial resolution*.

### 3.1.17

#### maximum local slope

greatest local slope of a surface feature that can be assessed by the probing system

Note 1 to entry: The term “local slope” is defined in ISO 4287:1997, 3.2.9.

### 3.1.18

#### instrument transfer function

ITF

$f_{\text{ITF}}$

function of spatial frequency describing how a surface topography measuring instrument responds to an object surface topography having a specific spatial frequency

Note 1 to entry: Ideally, the ITF tells us what the measured amplitude of a sinusoidal grating of a specified spatial frequency  $\nu$  would be relative to the true amplitude of the grating.

Note 2 to entry: For several types of optical instruments, the ITF may be a non-linear function of height, except for heights much smaller than the optical wavelength.

### 3.1.19

#### hysteresis

$X_{\text{HYS}}$ ,  $Y_{\text{HYS}}$ ,  $Z_{\text{HYS}}$

property of measuring equipment or characteristic, whereby the indication of the equipment or value of the characteristic depends on the orientation of the preceding stimuli

Note 1 to entry: Hysteresis can also depend, for example, on the distance travelled after the orientation of stimuli has changed.

Note 2 to entry: For *lateral scanning systems* (3.2.2), the hysteresis is mainly a repositioning error.

[SOURCE: ISO 14978:2006, 3.24]

### 3.1.20

#### metrological characteristic (of a measuring instrument)

<measuring equipment> characteristic of measuring equipment which may influence the results of the measurement

Note 1 to entry: Calibration of metrological characteristics may be necessary.

Note 2 to entry: The metrological characteristics have an immediate contribution to measurement uncertainty.

Note 3 to entry: Metrological characteristics for areal surface texture measuring instruments are given in [Table 1](#).

[SOURCE ISO 14978:2006, 3.12]

**Table 1 — List of metrological characteristics for surface texture measurement methods**

| Metrological characteristic | Symbol                         | Definition                                                                                                                                                                                         | Main potential error along |
|-----------------------------|--------------------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------|
| Amplification coefficient   | $\alpha_x, \alpha_y, \alpha_z$ | <a href="#">3.1.8</a> (see <a href="#">Figure 3</a> )                                                                                                                                              | $x, y, z$                  |
| Linearity deviation         | $l_x, l_y, l_z$                | Maximum local difference between the line from which the amplification coefficient is derived (see <a href="#">Figure 3</a> – key 5) and the response curve (see <a href="#">Figure 3</a> – key 4) | $x, y, z$                  |
| Residual flatness           | $z_{FLT}$                      | Flatness of the areal reference                                                                                                                                                                    | $z$                        |
| Measurement noise           | $N_M$                          | <a href="#">3.1.10</a>                                                                                                                                                                             | $z$                        |
| Lateral period limit        | $D_{LIM}$                      | <a href="#">3.1.16</a>                                                                                                                                                                             | $z$                        |
| Perpendicularity            | $\Delta_{PERxy}$               | Deviation from 90° of the angle between the $x$ - and $y$ -axes                                                                                                                                    | $x, y$                     |

## 3.2 Terms and definitions related to $x$ - and $y$ -scanning systems

### 3.2.1

#### areal reference guide

component(s) of the instrument that generate(s) the reference surface in which the probing system moves relative to the surface being measured according to a theoretically exact trajectory

Note 1 to entry: In the case of  $x$ - and  $y$ -scanning areal surface texture measuring instruments, the areal reference guide establishes a reference surface [ISO 25178-2:2012, 3.1.8]. It can be achieved through the use of two linear and perpendicular reference guides [ISO 3274:1996, 3.3.2] or one areal reference surface guide.

### 3.2.2

#### lateral scanning system

system that performs the scanning of the surface to be measured in the  $(x, y)$  plane

Note 1 to entry: There are essentially four aspects to a surface texture scanning instrument system; the  $x$ -axis drive, the  $y$ -axis drive, the  $z$ -measurement probe, and the surface to be measured. There are different ways in which these may be configured, and thus, there will be a difference between different configurations as explained in [Table 2](#).

Table 2 — Possible different configurations for reference guides (x and y)

|                |                                                          | Drive unit                     |           |           |                           |       |
|----------------|----------------------------------------------------------|--------------------------------|-----------|-----------|---------------------------|-------|
|                |                                                          | Two reference guides (x and y) |           |           | One areal reference guide |       |
|                |                                                          | Px o Cy <sup>a</sup>           | Px o Py   | Cx o Cy   | Pxy                       | Cxy   |
| Probing system | A: without arcuate error correction                      | Px o Cy-A                      | Px o Py-A | Cx o Cy-A | Pxy-A                     | Cxy-A |
|                | S: without arcuate error or with arcuate error corrected | Px o Cy-S                      | Px o Py-S | Cx o Cy-S | Pxy-S                     | Cxy-S |

NOTE For two given functions, f and g, f o g is the combination of these functions.

<sup>a</sup> Px = probing system moving along the x-axis  
Py = probing system moving along the y-axis  
Cx = component moving along the x-axis  
Cy = component moving along the y-axis

Note 2 to entry: When a measurement consists of a single field of view of a microscope, x- and y-scanning is not used. However, when several fields of view are linked together by stitching methods (see ISO 25178-601), the system is considered to be a scanning system.

### 3.2.3

#### drive unit x (respectively y)

component of the instrument that moves the probing system or the surface being measured along the reference guide on the x-axis (respectively y-axis) and returns the horizontal position of the measured point in terms of the lateral x-coordinate (respectively y-coordinate) of the profile

### 3.2.4

#### lateral position sensor

component of the drive unit that provides the lateral position of the measured point

Note 1 to entry: The lateral position can be measured or inferred by using, for example, a linear encoder, a laser interferometer, or a counting device coupled with a micrometer screw.

### 3.2.5

#### speed of measurement

$v_x$

speed of the probing system relative to the surface to be measured during the measurement along the x-axis

[SOURCE: ISO 25178-601:2010, 3.4.13]

### 3.2.6

#### static noise

$N_s$

combination of the *instrument* and *environmental noise* on the output signal when the instrument is not scanning laterally

Note 1 to entry: *Environmental noise* is caused by, e.g. seismic, sonic, and external electromagnetic disturbances.

Note 2 to entry: Notes 2 and 3 in 3.1.9 apply to this definition.

Note 3 to entry: Static noise is included in *measurement noise* (3.1.10).

**3.2.7****dynamic noise** $N_D$ 

noise occurring during the motion of the drive units on the output signal

Note 1 to entry: Notes 2 and 3 in [3.1.9](#) apply to this definition.

Note 2 to entry: Dynamic noise includes the *static noise* ([3.2.6](#)).

Note 3 to entry: Dynamic noise is included in *measurement noise* ([3.1.10](#)).

**3.3 Terms and definitions related to optical systems****3.3.1****light source**

optical device emitting an appropriate range of wavelengths in a specified spectral region

**3.3.2****measurement optical bandwidth** $B_{\lambda 0}$ 

range of wavelengths of light used to measure a surface

Note 1 to entry: Instruments may be constructed with light sources with a limited optical bandwidth and/or with additional filter elements to further limit the optical bandwidth.

**3.3.3****measurement optical wavelength** $\lambda_0$ 

effective value of the wavelength of the light used to measure a surface

Note 1 to entry: The measurement optical wavelength is affected by conditions such as the light source spectrum, spectral transmission of the optical components, and spectral response of the image sensor array.

**3.3.4****angular aperture**


angle of the cone of light entering an optical system from a point on the surface being measured

[SOURCE: ISO 25178-602:2010, 3.3.3]

**3.3.5****half aperture angle** $\alpha$ 

one half of the *angular aperture* ([3.3.4](#))

Note 1 to entry: This angle (see [Figure 8](#)) is sometimes also called as half cone angle.



**Key**

- L lens or optical system  
P focal point  
 $\alpha$  half aperture angle

**Figure 8 — Half aperture angle**

**3.3.6**

**numerical aperture**

$A_N$   
sine of the *half aperture angle* (3.3.5) multiplied by the refractive index  $n$  of the surrounding medium

Note 1 to entry:  $A_N = n \sin \alpha$

Note 2 to entry: In air for visible light,  $n \cong 1$ .

Note 3 to entry: The numerical aperture is dependent on the wavelength of light. Typically, the numerical aperture is specified for the wavelength that is in the middle of the *measurement optical bandwidth* (3.3.2).

**3.3.7**

**Rayleigh criterion**

quantity characterizing the spatial resolution of an optical system given by the separation of two-point sources at which the first diffraction minimum of the image of one point source coincides with the maximum of the other

Note 1 to entry: For a theoretically perfect, incoherent optical system with a filled objective pupil, the Rayleigh criterion of the optical system is equal to  $0,61 \lambda_0 / A_N$ .

Note 2 to entry: This parameter is useful for characterizing the instrument response to features with heights much less than  $\lambda_0$ .

**3.3.8**

**Sparrow criterion**

quantity characterizing the spatial resolution of an optical system given by the separation of two point sources at which the second derivative of the intensity distribution vanishes between the two-imaged points

Note 1 to entry: For a theoretically perfect incoherent optical system with a filled objective pupil, the Sparrow criterion of the optical system is equal to  $0,47 \lambda_0 / A_N$ , approximately 0,77 times the *Rayleigh criterion* (3.3.7).

Note 2 to entry: This parameter is useful for characterizing the instrument response to features with heights much less than  $\lambda_0$ .

Note 3 to entry: Under the same measurement conditions as the notes above, the Sparrow criterion is nearly equal to the spatial period of  $0,5 \lambda_0 / A_N$ , for which the theoretical instrument response falls to zero.

### 3.4 Terms and definitions related to optical properties of the workpiece

#### 3.4.1

##### **surface film**

material deposited onto another surface whose optical properties are different from that surface

Note 1 to entry: This concept may also be called *surface layer*.

#### 3.4.2

##### **thin film**

film whose thickness is such that the top and bottom surfaces cannot be readily separated by the optical measuring system

Note 1 to entry: For some measurement systems with special properties and algorithms, the thicknesses of thin films may be derived.

#### 3.4.3

##### **thick film**

film whose thickness is such that the top and bottom surfaces can be readily separated by the optical measuring system

#### 3.4.4

##### **optically smooth surface**

surface from which the reflected light is primarily specular and scattered light is not significant

Note 1 to entry: An optically smooth surface behaves locally like a mirror.

Note 2 to entry: A surface that acts as optically smooth under certain conditions such as wavelength range, *numerical aperture* (3.3.6), pixel resolution, etc. can act as optically rough when one or more of these conditions change.

#### 3.4.5

##### **optically rough surface**

surface that does not behave as an *optically smooth surface* (3.4.4), i.e. where scattered light is significant

Note 1 to entry: A surface that acts as optically rough under certain conditions such as wavelength range, numerical aperture, pixel resolution, etc. can act as optically smooth when one or more of these conditions change.

#### 3.4.6

##### **optically non-uniform material**

sample with different optical properties in different regions

Note 1 to entry: An optically non-uniform material may result in measured phase differences across the field of view that can be erroneously interpreted as differences in surface height.

### 3.5 Terms and definitions specific to focus variation instruments

#### 3.5.1

##### **focus variation microscopy**

##### **FV**

surface topography measurement method whereby the sharpness of the surface image (or another property of the reflected light at optimum focus) in an optical microscope is used to determine the surface height at each position along the surface

[SOURCE: ISO 25178-6:2010, 3.3.9]

#### 3.5.2

##### **focus variation sensor**

device that converts the height of points on the surface into signals during measurement using the focus variation method

### 3.5.3

#### **probing system**

<surface texture, *focus variation sensor* (3.5.2)> components of a focus variation instrument consisting of optical components, a vertical scanner, a digital optical sensor, an illumination system, and an optoelectronic controller

### 3.5.4

#### **focus variation measurement algorithm**

algorithm for analyzing the variation of focus in order to calculate the scan positions where each point is best in focus

### 3.5.5

#### **focus information**

measure to quantify the degree of focus at a specific lateral position in the surface image and at a specific vertical scan position

### 3.5.6

#### **focus information curve**

one-dimensional function where the x-axis contains the different vertical scan positions and the y-axis contains the corresponding *focus information* (3.5.5) at a specific lateral position in the surface image

### 3.5.7

#### **light source**

<focus variation sensor> source of light containing a continuum of wavelengths in a predefined spectral and spatial range

Note 1 to entry: Possible light sources are coaxial illumination, ring light, and external light sources.

### 3.5.8

#### **angular range of illumination**

$\alpha$

angular range from which the specimen is illuminated

#### **3.5.8.1**

##### **angular range of coaxial illumination**

$\alpha_1$

angular range from which the specimen is coaxially illuminated

Note 1 to entry: The value  $\alpha_1$  can be influenced by the choice of the objective.

Note 2 to entry: The value  $\alpha_1$  is often related to the angular range of detection (see [Figure 10](#)).

Note 3 to entry: In ordinary cases, the value  $\alpha_1$  can be derived from the numerical aperture of the objective.

Note 4 to entry: When special illumination sources are used (ring light, external light source, etc., see [Figure 9](#)), the *angular range of illumination* (3.5.8) can be much larger than  $\alpha_1$ .

#### **3.5.8.2**

##### **minimum incident angle of ring light illumination**

$\beta_{\text{Imin}}$


minimum incident angle of ring light illumination from which the specimen is illuminated

#### **3.5.8.3**


##### **maximum incident angle of ring light illumination**

$\beta_{\text{Imax}}$

maximum incident angle of ring light illumination from which the specimen is illuminated

**Key**

- L lens of optical system  
 RL ring light  
 $\alpha_l$  angular range of coaxial illumination  
 $\beta_{lmin}$  minimum incident angle of ring light illumination  
 $\beta_{lmax}$  maximum incident angle of ring light illumination  
 S specimen

**Figure 9 — Angular range of illumination****3.5.9****angular range of detection** $\alpha_D$ angular range of light rays that can be gathered by the objective (see [Figure 10](#))**Key**

- L lens of optical system  
 RL ring light  
 $\alpha_D$  angular range of detection  
 S specimen

**Figure 10 — Angular range of detection**

### 3.5.10

#### **scan height**

total range of physical path length traversed by the FV scan

Note 1 to entry: The scan height is usually synonymous with the total displacement of the FV sensor mechanically translated along its optical axis during data acquisition.

### 3.5.11

#### **polarization**

method which allows one to filter out light waves in certain polarization states by using special optical elements called *polarizers* (3.5.13) or analyzers

### 3.5.12

#### **analyzer**

optical element used to polarize the rays of the light after they have been reflected from the specimen and gathered by the objective

### 3.5.13

#### **polarizer**

optical element used to polarize the rays of the *light source* (3.5.7) before they are transmitted to the specimen

### 3.5.14

#### **polarization angle**

angle between the *polarization direction* (3.5.15) of the analyzer and the polarizer

### 3.5.15

#### **polarization direction**

direction of the electric vector the light waves that are transmitted by a polarizing optical element

### 3.5.16

#### **sensor settings**

settings that influence how the sensor converts the light information into a digital signal

Note 1 to entry: Typical settings are *exposure time* (3.5.18) and *gamma* (3.5.17).

### 3.5.17

#### **gamma**

sensor setting that performs a non-linear transformation of the sensor response in relation to exposure using an exponential function

### 3.5.18

#### **exposure time**

amount of time during which a photosensitive material is acted upon by light

### 3.5.19

#### **roughness threshold**

minimum  $S_q$  value of a workpiece at a certain short wavelength cut-off spatial frequency needed for proper measurements

## 4 Description of the influence quantities

### 4.1 General

Focus variation instruments provide a measurement of lateral ( $x$  and  $y$ ) and height ( $z$ ) values from which the surface texture parameters are calculated.

## 4.2 Overview

Focus variation instruments use the following measurement process:

- To perform a complete measurement of the surface, the optics is moved vertically along the optical axis while continuously capturing data from the surface. This means that each region of the object is sharply focused. Algorithms convert the acquired sensor data into 3D information and a true colour image with full-depth of field. The 3D information is then calculated by analysing the focus information (curve) along the vertical axis. (see [Annex A](#)).

## 4.3 Influence quantities

Influence quantities for focus variation instruments are given in [Table 3](#). The table indicates the metrological characteristics (see [3.1.20](#), [Table 1](#)) that are affected by deviations of influence quantities.

**Table 3 — Influence quantities for focus variation instruments**

| Component                 | Element                   | Influence quantities                   |                                                                                                                                                      | Metrological characteristic affected                                          |
|---------------------------|---------------------------|----------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------|
| Light source              |                           | $\lambda_0$                            | Measurement optical wavelength                                                                                                                       | $l_z, N_M, D_{LIM}, z_{FLT}$                                                  |
|                           |                           | $B_{\lambda 0}$                        | Measurement optical bandwidth                                                                                                                        | $l_z, N_M, D_{LIM}, z_{FLT}$                                                  |
|                           |                           | $\alpha_l, \beta_{lmin}, \beta_{lmax}$ | Angular range of illumination                                                                                                                        | $l_z, N_M, D_{LIM}, z_{FLT}$                                                  |
|                           |                           | S, P, C, U                             | The state of polarization of the light impinging on the measured surface. The polarization is typically described as S, P, circular, or unpolarized. | $l_z, N_M, D_{LIM}, z_{FLT}$                                                  |
| Microscope imaging system |                           | $A_N$                                  | Numerical aperture                                                                                                                                   | $l_z, N_M, D_{LIM}, z_{FLT}$                                                  |
|                           |                           | $M_{IMG}$                              | Magnification between object sizes on the surface and image sizes on the sensor                                                                      | $\alpha_x, \alpha_y$                                                          |
|                           |                           | $Q_{OPT}$                              | General quality of the optical components used including aberrations, transmission, alignment errors, etc.                                           | $\alpha_x, \alpha_y, \alpha_z, z_{FLT}, l_x, l_y, l_z, D_{LIM}, \Delta_{PER}$ |
|                           |                           | $F_{PSF}$                              | Point spread function of the microscope imaging system (see ISO 10934-2:2007, 2.35)                                                                  | $\alpha_x, \alpha_y, \alpha_z, z_{FLT}, l_x, l_y, l_z, D_{LIM}$               |
|                           |                           | $P_{DISXY}$                            | Lateral distortion of the magnified image on the camera                                                                                              | $\alpha_x, \alpha_y, l_x, l_y, \Delta_{PER}$                                  |
| Vertical scanning system  |                           | $F_z$                                  | Response curve of the vertical scanning system                                                                                                       | $\alpha_z, l_z, z_{FLT}$                                                      |
| Camera                    |                           | $\Delta_x$                             | x-pixel spacing                                                                                                                                      | $D_{LIM}$                                                                     |
|                           |                           | $\Delta_y$                             | y-pixel spacing                                                                                                                                      | $D_{LIM}$                                                                     |
| Controller                | Acquisition software      | $\Delta_z$                             | Z-scan increment                                                                                                                                     | $N_M$                                                                         |
|                           |                           | $A_{ACQ}$                              | Acquisition method – Manner in which the sensor data are acquired (e.g. continuously, discretely stepped)                                            | $l_z, N_M$                                                                    |
|                           |                           | $A_{NUM}$                              | Measurement algorithm – Number of images acquired                                                                                                    | $l_z, N_M, z_{FLT}$                                                           |
|                           | Focus variation algorithm | $A_{FOV}$                              | Procedure that constructs the surface heights from the acquired data using the focus variation method                                                | $\alpha_x, \alpha_y, \alpha_z, l_x, l_y, l_z, D_{LIM}, N_M$                   |
|                           |                           | $A_{CORR}$                             | Method to correct optical and stage aberrations                                                                                                      | $\alpha_x, \alpha_y, \alpha_z, z_{FLT}, l_x, l_y, l_z, D_{LIM}, \Delta_{PER}$ |
|                           |                           | $R_l$                                  | Lateral resolution                                                                                                                                   | $l_x, l_y, l_z, D_{LIM}, N_M$                                                 |

**Table 3** (continued)


| Component          | Element | Influence quantities |                                                                                                                          | Metrological characteristic affected |
|--------------------|---------|----------------------|--------------------------------------------------------------------------------------------------------------------------|--------------------------------------|
| Instrument overall |         | $D_x$ or $D_y$       | Lateral sampling interval equal to the lateral pixel spacing of the camera divided by the magnification $M_{\text{IMG}}$ | $D_{\text{LIM}}$                     |
|                    |         | $N_I$                | Instrument noise                                                                                                         | $N_M$                                |
|                    |         | $N_{\text{VIB}}$     | Environmental Vibration – Unwanted motion between the surface being measured and the optical system                      | $N_M$                                |
|                    |         | $T_I$                | Integration time required to complete a single scan in Z                                                                 | $N_M$                                |

## Annex A (informative)

### Components of a focus variation microscope

#### A.1 Typical configuration

Figure A.1 illustrates the typical configuration of a focus variation microscope.



#### Key

- 1 array detector
- 2 optical components
- 3 white light source
- 4 illumination beam splitter
- 5 objective
- 6 specimen
- 7 vertical scan
- 8 focus information curve with maximum position
- 9 light beam (...)
- 10 analyzer
- 11 polarizer
- 12 ring light
- 13 optical axis (...)

Figure A.1 — Schematic diagram of a typical measurement device based on focus variation

## A.2 Operation principle

Focus variation<sup>[7]</sup> combines the small-depth of focus of an optical system with vertical scanning to provide topographical information from the variation of focus. In the following, the operating principle is demonstrated for a typical focus variation microscope schematically shown in [Figure A.1](#). The main component of the system is the optical microscope containing various lenses that can be equipped with different objectives allowing measurements with different resolution. With a beam splitting mirror, light emerging from a white light source is inserted into the optical path of the system and focused onto the specimen through the objective. Depending on the topography of the specimen, the light is scattered into several directions as soon as it hits the specimen through the objective. If the topography shows diffuse reflective properties, the light is scattered strongly into all directions. In the case of specular reflection, the light is reflected mainly into one direction. All rays emerging from the specimen and hitting the objective lens are collected in the optics and gathered by a light sensitive sensor behind the beam splitting mirror. Due to the small depth of field of the optics, only small regions of the object are sharply imaged. To perform a complete detection of the surface with full depth of field, the optics is moved vertically along the optical axis while continuously capturing data from the surface. Each region of the object is sharply focused at one of the vertical positions of the scanner. Algorithms convert the acquired sensor data into 3D information and a true colour image with full-depth of the field. This is achieved by analysing the variation of focus along the vertical axis.

In addition to the scanned height data, the microscope also delivers colour information for each measured 3D point. This provides an optical colour image which eases measurements and identification of distinctive local surface features. The visual correlation between the optical colour image of the specimen surface and its depth information are often linked to each other and are, therefore, an essential aspect of meaningful 3D measurement.

For the description of the focus variation technique, see Reference [\[10\]](#). The focus variation technique is sometimes also referred to as shape from focus (see Reference [\[11\]](#)).

## A.3 Light source

In contrast to other optical techniques that are limited to coaxial illumination, the maximum local slope is not limited by the numerical aperture of the objective. Focus variation can be used with a large range of different illumination sources (such as a ring light) which allows the measurement of local slope angles up to 90°. Additionally, the light can be polarized using polarization filters (polarizer and analyser) which allow the removal of specular light components. This is especially helpful for the measurement of metallic surfaces containing steep and flat surface elements.

## A.4 Scanner

There are various methods for performing the vertical scan. These are moving the sample, moving the whole optics, and moving parts of the optics.

## A.5 Objectives

The objectives that are used for focus-variation should meet some basic requirements. They should have long working distances in order to allow the measurement of high and rough specimens. The numerical aperture should be as large as possible since the numerical aperture directly influences the depth of field which is the basic parameter for measurements based on focus-variation. Since the measurement is performed by white light, the objectives should be corrected for freedom of chromatic aberrations.






## A.6 Algorithm

Various methods exist to analyse the variation of focus usually based on the calculation of image sharpness at each scan position. Typically, such focus information is derived by evaluating the values of the surface image in a small local area. In general, an object point is focused more sharply when neighbouring points

have a larger variation of image intensity. As an example, the standard deviation of these values can be used as a simple measure for the focus information. This is demonstrated in [Table A.1](#), where focus information is calculated for a chess-pattern like object at five different scan positions.

The focus information curve consists of the focus information for each vertical scan position. By calculating the maximum of the focus information curve, the height information can be determined for one object point. Various methods exist to calculate the maximum of the focus curve. Three of them are summarized in [Table A.2](#) together with their speed and accuracy. The fastest, but least accurate method, is to simply use the scan position with the maximum focus information. More advanced methods fit polynomials or more complex functions with the focus information curve and calculate the maximum as the peak of the fitted polynomial or function.

**Table A.1 — Calculation of focus information using the standard deviation of the surface image within a 5×5 neighbourhood of points around the point of interest**

| Scan position                                                                                                                                                                | Surface image                                                                       | Standard deviation |
|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------|--------------------|
| Out of focus                                                                                                                                                                 |    | 10                 |
| Almost in focus                                                                                                                                                              |   | 20                 |
| In focus                                                                                                                                                                     |  | 50                 |
| Almost in focus                                                                                                                                                              |  | 20                 |
| Out of focus                                                                                                                                                                 |  | 10                 |
| <p>1 Point of interest for which the focus information is calculated.</p> <p>2 5×5 neighbourhood of points used to calculate the focus information (standard deviation).</p> |                                                                                     |                    |