
INTERNATIONAL STANDARD



3599

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Vernier callipers reading to 0,1 and 0,05 mm

Pieds à coulisse à vernier au 1/10 et au 1/20 mm

First edition — 1976-06-15

Corrected and reprinted — 1982-07-15

STANDARDSISO.COM : Click to view the full PDF of ISO 3599:1976

UDC 531.714.8

Ref. No. ISO 3599-1976 (E)

Descriptors : measuring instruments, mechanical measuring instruments, callipers, specifications, dimensions, precision, dimensional control.

Price based on 4 pages

FOREWORD

ISO (the International Organization for Standardization) is a worldwide federation of national standards institutes (ISO Member Bodies). The work of developing International Standards is carried out through ISO Technical Committees. Every Member Body interested in a subject for which a Technical Committee has been set up has the right to be represented on that Committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work.

Draft International Standards adopted by the Technical Committees are circulated to the Member Bodies for approval before their acceptance as International Standards by the ISO Council.

International Standard ISO 3599 was drawn up by Technical Committee ISO/TC 3, *Limits and fits*, and circulated to the Member Bodies in October 1974.

It has been approved by the Member Bodies of the following countries:

Australia	Finland	Poland
Austria	France	Romania
Belgium	Germany	South Africa, Rep. of
Bulgaria	Hungary	Sweden
Canada	India	Switzerland
Chile	Netherlands	Turkey
Czechoslovakia	New Zealand	U.S.S.R.

The Member Body of the following country expressed disapproval of the document on technical grounds :

United Kingdom

Vernier callipers reading to 0,1 and 0,05 mm

1 SCOPE AND FIELD OF APPLICATION

This International Standard specifies the most important dimensional, functional and quality characteristics of vernier callipers reading to 0,1 and 0,05 mm, with a maximum range of 1 000 mm. Methods for testing the accuracy of the instruments are given in an annex, for general information only.

NOTE — These vernier callipers are also commonly known as 1/10 and 1/20 vernier callipers respectively.

2 NOMENCLATURE AND DEFINITIONS

2.1 For the nomenclature for vernier callipers, see figures 1 and 2.

2.2 error of measurement: The algebraic difference between the measured size and the true size.

2.3 measuring uncertainty: The error margin corresponding to the inherent errors of measurement of a vernier calliper. It is defined as being equal to $\pm 2s$; i.e. for a normal distribution of the readings on the instrument, about 95 % of readings will not deviate from the mean size (true value) by more than twice the standard deviation s .

3 SPECIFICATION

3.1 Design features

3.1.1 Measuring ranges

For recommended measuring ranges, see table 1.

3.1.2 Material

The main parts of the calliper shall be of good quality steel (plain carbon steel or stainless steel).

3.1.3 Beam

The beam shall be long enough for the sliding jaw assembly not to overhang when measuring at the end of the measuring range.

3.1.4 Jaws

For the minimum projection of the jaws, $J_{\min}$, see table 1.

The maximum projection, $J_{\max}$, shall be equal to one-third of the measuring range but with a maximum of 200 mm.

The sliding jaw shall be a good sliding fit along the beam in order to permit fine adjustment to be made.

The slider shall be provided with a suitable clamp so that it may be effectively clamped to the beam without altering the setting.

For the minimum length of the faces for internal measurement ($l_{\min}$), see table 1.

The jaws may be provided with knife edges as shown in figure 2.

The nominal combined width of the jaws for internal measurement shall be 0*, 5, 10 or 20 mm. The faces for internal measurement (except the knife-edge faces) shall be of cylindrical form with a radius not exceeding one-half of their combined width (see figure 1).

3.1.5 Depth-measuring device

The vernier calliper may be provided with a depth-measuring blade which is connected to the slider and allows the measurement of depths with reference to the end face of the beam (see figure 2).

3.1.6 Scales

The beam shall be graduated in millimetres and the length of the scale shall be at least equal to the measuring range of the calliper plus the length of the vernier.

The length of the vernier scale may be 9, 19 or 39 mm (see figures 4, 5 and 6).

The scale lines of both the beam and the vernier shall be sharp, clear and perpendicular to the edge of the beam and their thickness shall be not less than 0,08 mm and not more than 0,2 mm.

The numbering on the beam and the vernier shall be such that the scale is easy to read.

* For jaws with knife edges.

The distance between the graduated face of the beam and the edge of the graduated, bevelled face of the vernier shall not exceed 0,3 mm (see figure 3).

4 ACCURACY

4.1 Measuring uncertainty

The permissible measuring uncertainty in micrometres at $\pm 2 s$, as given in table 2, is calculated from the following formula :

$$\pm (50 + 0,1 L)$$

where L is any measured length, in millimetres, within the measuring range.

4.2 Measuring faces

With the slider clamped to the beam at any position within the measuring range of the calliper, the faces for external measurement shall be flat to within $10 \mu\text{m}$ per 100 mm over their length. They shall be parallel to within $20 \mu\text{m}$ per 100 mm over their length.

The faces for internal measurement shall be parallel to within $10 \mu\text{m}$ over their length, and the permissible tolerance for their combined width (see 3.1.4) shall be : $+ 30 \mu\text{m}$
0

The measuring faces shall have a diamond pyramid hardness number of not less than

- 700 HV for plain carbon steel;
- 550 HV for stainless steel.

4.3 Scale lines

In any one instrument, the thickness of all scale lines on the main scale and vernier shall not differ by more than 0,03 mm.

5 TABLES

5.1 Dimensions of vernier callipers

TABLE 1

Dimensions in millimetres

External measuring range	Minimum projection of jaws $J_{\min}$	Minimum length of faces for internal measurement $l_{\min}$
0 to 135	35	6
0 to 160	40	6
0 to 200	50	8
0 to 250	50	10
0 to 300	60	10
0 to 500	80	15
0 to 750	80	15
0 to 1 000	100	20

5.2 Measuring uncertainty

TABLE 2

Measured length L	Measuring uncertainty at $\pm 2 s$ (95 %)
mm	μm
0	± 50
100	± 60
200	± 70
300	± 80
400	± 90
500	± 100
600	± 110
700	± 120
800	± 130
900	± 140
1 000	± 150