
Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy —

Part 4:
Calibration of area and personal dosimeters in low energy X reference radiation fields

Radioprotection — Rayonnements X et gamma de référence pour l'étalonnage des dosimètres et des débitmètres et pour la détermination de leur réponse en fonction de l'énergie des photons —

Partie 4: Étalonnage des dosimètres de zone et individuels dans des champs de référence X de faible énergie



STANDARDSISO.COM : Click to view the full PDF of ISO 4037-4:2019



COPYRIGHT PROTECTED DOCUMENT

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
Introduction.....	v
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	1
4 Symbols (and abbreviated terms).....	2
5 General procedures for calibrating and determining response.....	3
6 Characterisation and production of low energy X-ray reference radiations.....	3
6.1 General.....	3
6.2 Tube potential.....	4
6.3 Spectral fluence and conversion coefficients.....	4
7 Dosimetry of low energy reference radiations.....	4
7.1 General.....	4
7.2 Stability check facility.....	4
8 Calibration and determination of the response as a function of photon energy and angle of radiation incidence.....	4
8.1 General.....	4
8.2 Selection of calibration method.....	5
8.3 Calibration by using reference instruments for K_a	5
8.3.1 General.....	5
8.3.2 Conventional quantity value of the air kerma.....	5
8.3.3 Conventional quantity value of the dose equivalent quantities $H_p(0,07)$ and $H'(0,07)$	6
8.3.4 Conventional quantity value of the dose equivalent quantities $H_p(10)$ or $H^*(10)$ and $H_p(3)$ or $H'(3)$	6
8.3.5 Performing the calibration.....	8
8.4 Calibration by using reference instruments which measure the ICRU dose equivalent quantities.....	8
8.4.1 General.....	8
8.4.2 Conventional quantity value of the dose equivalent quantities $H_p(10)$ or $H^*(10)$ and $H_p(3)$ or $H'(3)$	9
8.4.3 Performing the calibration.....	10
8.5 Statement of uncertainty.....	11
Annex A (normative) Correction for air density.....	12
Bibliography.....	18

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 85, *Nuclear energy, nuclear technologies and radiological protection*, Subcommittee SC 2, *Radiological protection*.

This second edition cancels and replaces the first edition (ISO 4037-4:2004), which has been technically revised.

A list of all the parts in the ISO 4037 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

The maintenance release of this document adjusts this fourth part to the second edition of the first three parts. This includes the improvements on high voltage generators from 1996 to 2017 (e.g., the use of high frequency switching supplies providing nearly constant potential), and the spectral measurements at irradiation facilities equipped with such generators (e.g., the catalogue of X-ray spectra by Ankerhold^[1]). It also incorporates all published information with the aim to adjust the requirements for the technical parameters of the reference fields to the targeted overall uncertainty of about 6 % to 10 % for the phantom related operational quantities of the International Commission on Radiation Units and Measurements (ICRU)^[2]. It does not change the concept of ISO 4037.

ISO 4037, focusing on photon reference radiation fields, is divided into four parts. ISO 4037-1 gives the methods of production and characterization of reference radiation fields in terms of the quantities photon fluence and air kerma free-in-air. ISO 4037-2 describes the dosimetry of the reference radiation qualities in terms of air kerma and in terms of the phantom related operational quantities of the International Commission on Radiation Units and Measurements (ICRU)^[2]. ISO 4037-3 describes the methods for calibrating and determining the response of dosimeters and doserate meters in terms of the operational quantities of the ICRU^[2]. This document gives special considerations and additional requirements for calibration of area and personal dosimeters in low energy X reference radiation fields, which are reference fields with generating potential lower or equal to 30 kV.

The general procedures described in ISO 29661 including Amendment 1 are used as far as possible in this document. In addition, the symbols used are in line with ISO 29661.

NOTE For irradiation of the whole body, $H_p(10)$ and $H^*(10)$ are relevant for radiation protection, as long as they are closer to their limit than $H'(0,07)$ and $H_p(0,07)$. This is the case down to about 15 keV.

STANDARDSISO.COM : Click to view the full PDF of ISO 4037-4:2019

Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy —

Part 4:

Calibration of area and personal dosimeters in low energy X reference radiation fields

1 Scope

This document gives guidelines on additional aspects of the characterization of low energy photon radiations and on the procedures for calibration and determination of the response of area and personal dose(rate)meters as a function of photon energy and angle of incidence. This document concentrates on the accurate determination of conversion coefficients from air kerma to $H_p(10)$, $H^*(10)$, $H_p(3)$ and $H'(3)$ and for the spectra of low energy photon radiations. As an alternative to the use of conversion coefficients the direct calibration in terms of these quantities by means of appropriate reference instruments is described.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 4037-1, *Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy — Part 1: Radiation characteristics and production methods*

ISO 4037-2:2019, *Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy — Part 2: Dosimetry for radiation protection over the energy ranges from 8 keV to 1,3 MeV and 4 MeV to 9 MeV*

ISO 4037-3:2019, *Radiological protection — X and gamma reference radiation for calibrating dosimeters and doserate meters and for determining their response as a function of photon energy — Part 3: Calibration of area and personal dosimeters and the measurement of their response as a function of energy and angle of incidence*

ISO/IEC Guide 98-3:2008, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

ISO 29661, *Reference radiation fields for radiation protection — Definitions and fundamental concepts*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 4037-1, ISO 29661 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at <https://www.iso.org/obp>

— IEC Electropedia: available at <http://www.electropedia.org/>

3.1

low energy X-ray reference radiation

all radiation qualities with nominal tube potentials up to and including 30 kV

Note 1 to entry: These radiation qualities are as specified in ISO 4037-1 and all continuous reference filtered X radiations.

4 Symbols (and abbreviated terms)

The symbols (and abbreviated terms) used are given in Table 1.

Table 1 — Symbols (and abbreviated terms)

Symbol	Meaning	Unit
ρ	air density	kg/m ³
ρ_0	air density under reference conditions: $\rho_0 = 1,197\ 4\ \text{kg/m}^3$	kg/m ³
ρ_{irr}	air density prevailing during irradiation	kg/m ³
ρ_{con}	air density prevailing during determination of the conventional quantity value of the measurand	kg/m ³
ρ_{cal}	air density prevailing during calibration of the instrument	kg/m ³
ρ_{MC}	air density prevailing during calibration of the monitor chamber	kg/m ³
ρ_{spec}	air density prevailing during the spectral measurements	kg/m ³
$\Delta\rho$	change of air density	kg/m ³
α	angle of radiation incidence to the normal of the phantom surface	deg
$\Delta\alpha$	change of angle of radiation incidence	deg
U	tube potential	V
ΔU	change in tube potential	V
T	air temperature	K
T_0	air temperature under reference conditions: $T_0 = 293,15\ \text{K}$ (equivalent to 20 °C),	K
r	relative air humidity	—
r_0	relative air humidity under reference conditions: $r_0 = 0,65$ (equivalent to 65 %)	—
p	air pressure	kPa
p_0	air pressure under reference conditions: $p_0 = 101,3\ \text{kPa}$	kPa
m_{d}	gradient of the gradient $m(d_{\text{air}})$	m ² /kg
$m(d_{\text{air}})$	gradient for distance d_{air}	m ³ /kg
$m(1,0\ \text{m})$	gradient for distance 1,0 m	m ³ /kg
K_{a}	air kerma free-in-air	Gy
$k(\rho, M)$	air density correction factor for measurand M	—
$H_{\text{p}}(10)$	personal dose equivalent at 10 mm depth	Sv
$H_{\text{p}}(3)$	personal dose equivalent at 3 mm depth	Sv
$H_{\text{p}}(0,07)$	personal dose equivalent at 0,07 mm depth	Sv
$H^*(10)$	ambient dose equivalent at 10 mm depth	Sv
$H'(3)$	directional dose equivalent at 3 mm depth	Sv
$H'(0,07)$	directional dose equivalent at 0,07 mm depth	Sv
$h_{\text{pK}}(10; \alpha)$	conversion coefficient from K_{a} to $H_{\text{p}}(10)$ for angle of radiation incidence α	Sv/Gy
$h^*_{\text{p}}(10)$	conversion coefficient from K_{a} to $H^*(10)$	Sv/Gy
$h_{\text{pK}}(3; \alpha)$	conversion coefficient from K_{a} to $H_{\text{p}}(3)$ for angle of radiation incidence α	Sv/Gy

Table 1 (continued)

Symbol	Meaning	Unit
E	photon energy	eV
d_{MC}	distance from the beam exit window of the X-ray tube to the monitor chamber	m
d_{air}	distance from the beam exit window of the X-ray tube to the point of test	m
$\Phi_E(E)$	spectral fluence at the photon energy E	$m^{-2} eV^{-1}$
N	number of pulses generated in the detector	—
Q	charge Q generated in the detector by one photon	C
$R(E, Q)$	response function	$m^2 C^{-1}$

5 General procedures for calibrating and determining response

In ISO 4037-2, two methods are given to determine the phantom related dose equivalent quantities for low energy X reference fields. Both methods require a reference field according to ISO 4037-1. The first method, method I, requires the dosimetry with respect to air kerma free-in-air and after that, the selected operational quantity is derived by the application of a conversion coefficient that relates the air kerma free-in-air to the selected operational quantity. For matched reference fields this conversion coefficient is taken from ISO 4037-3, for characterized reference fields the conversion coefficient is determined using spectrometry. For the dose-equivalent quantities $H'(0,07)$ and $H_p(0,07)$, this procedure is associated with only a small additional uncertainty, because the conversion coefficients depend only slightly on the photon energy and angle of radiation incidence for the ranges given in ISO 4037-3. Therefore, for these dose equivalent quantities, no special attention is needed for the low energy X reference radiation fields. For the four other dose equivalent quantities $H_p(10)$, $H^*(10)$, $H_p(3)$ and $H'(3)$ this is different. For them, the use of conversion coefficients can be associated with large additional uncertainties if low energy X reference radiation fields are considered. This is because the conversion coefficients depend strongly on the photon energy and the angle of radiation incidence. A detailed description of all the measurements and methods necessary to avoid these additional uncertainties is given by Ankerhold et al.[3][4] and by Behrens[5].

The second method, method II, to determine the phantom related dose quantities is based on the use of (secondary) standards directly calibrated in terms of these dose equivalent quantities. This method can also be used for all non-validated radiation qualities, for which the recommended conversion coefficients cannot be used. This method is described in ISO 4037-2:2019, Clause 6.

If the reference field cannot be validated, then, method I can still be used if a spectrometer is used to measure the spectrum of the radiation quality under consideration. From this spectrum, the specific conversion coefficient can be calculated and applied to the measured value of the air kerma K_a free-in-air.

This document defines the conditions that shall be met to use one of the two methods and the experimental steps to be used for the selected method. If a monitor chamber (see ISO 4037-2:2019, 9.2) is used as a transfer device additional corrections shall be applied for differences in the air density prevailing during calibration of the monitor chamber and during calibration of the instrument under test. The standard does not give advice on the construction of the instruments necessary for both ways. Examples for the instruments and the experimental steps for both ways are given by Ankerhold et al.[3][4], Behrens[5] and Duftschmid et al.[6].

6 Characterisation and production of low energy X-ray reference radiations

6.1 General

This subclause specifies the characteristics by which a laboratory can produce the reference filtered X radiations specified in ISO 4037-1 for the given purposes. For various influence quantities, data are given on the required stability of these influence quantities. These data indicate how large the change in value of these influence quantities can be until a change of the measurand of 2 % is caused. These

data shall either be interpreted as limits for the deviation from its nominal value or, where possible, as a criterion for the necessity of corrections.

The requirements given in ISO 4037-1:2019, 4.2, consider partly the special requirements for low energy reference radiations for the quantities $H_p(10)$ or $H^*(10)$. These special requirements are, to less extent, also valid for the dose equivalent quantities $H'(3)$, $H_p(3)$. Therefore, this document focuses on the quantities $H_p(10)$ or $H^*(10)$ and assumes that for the dose equivalent quantities $H'(3)$, $H_p(3)$ nearly the same requirements are valid.

6.2 Tube potential

This subclause is relevant for method I and method II. The dose equivalent quantities $H_p(10)$, $H^*(10)$, $H_p(3)$ and $H'(3)$ are for low energy X radiation more sensitive to the tube potential than the air kerma, K_a , free-in-air. The requirements on the tube potential given in ISO 4037-1:2019, Table 7 are valid. This Table 7 gives values for the change of tube potential that cause a change in the value of the conversion coefficient of 2 % if all other parameters are unchanged.

6.3 Spectral fluence and conversion coefficients

This subclause is relevant for method I only. Knowledge of the spectral fluence is necessary to determine the conversion coefficient from air kerma to the measurand for every radiation quality of the X-ray facility. In ISO 4037-2:2019, Annex B, an example for the determination of the spectral fluence is given. The spectral fluence is converted to a spectral air kerma by multiplying the spectral fluence with the monoenergetic fluence to air kerma conversion coefficients. This spectral air kerma is then multiplied with the monoenergetic conversion coefficients for the respective measurand (see ISO 4037-3) to get the spectral $H_p(10)$, $H^*(10)$, $H_p(3)$ or $H'(3)$ distribution which is then integrated to get the actual conversion coefficient. The obtained conversion coefficients are valid only for the air density ρ_{spec} prevailing during the spectral measurements.

7 Dosimetry of low energy reference radiations

7.1 General

The instruments to be used shall be standard instruments as described in ISO 4037-2:2019, Clause 4. The general procedures in ISO 4037-2:2019, Clauses 5 and 6, and, where appropriate, the procedures applicable to ionization chambers in ISO 4037-2:2019, Clause 7, shall be followed.

7.2 Stability check facility

Where appropriate, a radioactive check source may be used to verify the satisfactory operation of the instrument prior to periods of use.

8 Calibration and determination of the response as a function of photon energy and angle of radiation incidence

8.1 General

The general methods given in ISO 4037-3 shall be followed. For an unsealed standard ionization chamber this includes correction for air temperature, pressure and humidity according to ISO 4037-2:2019, 7.4.2. In Cause 8, additional requirements and advice on the selection of calibration method are given. Moreover, for the dose equivalent quantity $H_p(10)$ limits are given for the adjustment of the angle of incidence.

8.2 Selection of calibration method

This subclause gives information, additional to ISO 4037-2, on the choice of dosimetric method, which can be used for determination of the conventional quantity value of the dose quantities of interest. As explained in [Clause 5](#), two methods are possible to determine the conventional quantity value of the dose quantities of interest.

For the highest level of dissemination of the phantom related quantities, e.g., by National Metrology Institutes, Method I, using spectrometry and reference instruments for K_a is required to achieve an uncertainty of the conventional quantity value of about 6 % ($k = 2$) or less. The air kerma, K_a , shall be determined by a primary or at least directly traceable standard and spectrometry of the reference field shall be performed according to ISO 4037-2:2019, Annex B, both at the point of test.

Method II, using secondary standard instruments, which measure directly dose equivalent quantities, may be used by all other laboratories. The achievable uncertainty is between 6 % and 10 % ($k = 2$) depending on the radiation quality.

The time period starting from the determination of the conventional quantity value of the measurand until the calibration of the instrument under test and determination of its response as function of photon energy and angle of radiation incidence has to be considered, because the stability of certain parameters over this period shall be maintained.

8.3 Calibration by using reference instruments for K_a

8.3.1 General

This subclause is relevant for method I only. During the potentially long time period between the determination of the conversion coefficient (see [6.3](#)) and the calibration of the instrument the requirements on tube potential of [6.2](#) shall be followed. In addition, the air density at all measuring events shall be constant within the limits given in [Table 2](#), otherwise the appropriate corrections, provided in [Annex A](#), shall be applied.

The additional corrections for the use of a monitor chamber as a transfer device are also provided in [Annex A](#).

As an example, [Table 2](#) gives values for the percentage change of air density that causes a change in the value of the air kerma, K_a and the conversion coefficients $h_{pK}(10, 0^\circ)$, $h^*_{pK}(10)$ and $h_{pK}(10, 60^\circ)$ of 2 % at 2,5 m distance of the point of test from the focus and for 0° and 60° radiation incidence. These conditions are representative for calibrations with respect to $H_p(10)$ performed on a ISO water slab phantom (see ISO 4037-3).

8.3.2 Conventional quantity value of the air kerma

Within the short time period (typically one or a few hours) from the measurement of the conventional quantity value of the air kerma and the determination of the required phantom related quantity value to the calibration of the instrument the air density shall not change by more than the limits given in [Table 2](#). These data are valid for a distance of 2,5 m, which is typical for calibrations with respect to $H_p(10)$ performed on an ISO water slab phantom. Normally, these air density requirements are fulfilled and no correction is necessary, in the other few cases the correction method given in the [Annex A](#) shall be applied as follows. If ρ_{con} is the air density prevailing during determination of the conventional quantity value of the air kerma K_a and ρ_{cal} those during calibration of the instrument, then the conventional quantity value of K_a during calibration is

$$K_a(\rho_{cal}) = \frac{k(\rho_{cal}, K_a)}{k(\rho_{con}, K_a)} K_a(\rho_{con}) \quad (1)$$

For the air density correction factor $k(\rho, K_a)$ for the quantity air kerma K_a see [Formula \(A.2\)](#).

If a monitor chamber is used as a transfer device for the measuring quantity air kerma K_a then the difference of the air density prevailing during the calibration of the monitor chamber and the air density prevailing during the calibration of the instrument shall be within the limits given in [Table 2](#). Otherwise, the correction method given in the [Annex A](#) shall be applied as follows. If the monitor chamber is mounted at a distance d_{MC} from the beam exit window, ρ_{MC} is the air density prevailing during calibration of the monitor chamber and ρ_{cal} those during calibration of the instrument at the distance d_{air} , then the conventional quantity value of K_a during calibration is (for the air density correction factor $k_{MC}(\rho, K_a)$ see [Formula \(A.5\)](#)):

$$K_a(\rho_{cal}) = \frac{k(\rho_{cal}, K_a)}{k_{MC}(\rho_{MC}, K_a)} K_a(\rho_{MC}) \quad (2)$$

Table 2 — Percentage change of air density that causes a change in the value of the air kerma, K_a , and the conversion coefficients $h_{pK}(10, 0^\circ)$ or $h^*_{pK}(10)$ and $h_{pK}(10, 60^\circ)$ of 2 % at 2,5 m distance of the point of test from the focus of the X-ray tube and 0° and 60° radiation incidence

Radiation quality	Tube potential U kV	$\Delta\rho/\rho$ for 2,5 m distance causing a change of 2 % of the value of		
		K_a %	$h_{pK}(10, 0^\circ), h^*_{pK}(10)$ %	$h_{pK}(10, 60^\circ)$ %
L-10	10	0,9	6,3	4,8
L-20	20	5,3	>20	>20
L-30	30	14	>20	>20
N-10	10	0,8	3,5	2,9
N-15	15	2,1	9,2	6,9
N-20	20	4,3	>20	18
N-25	25	8,0	>20	>20
N-30	30	12	>20	>20
H-10	10	0,7	2,4	2,0
H-20	20	1,9	3,7	3,2
H-30	30	4,4	11	9,1

8.3.3 Conventional quantity value of the dose equivalent quantities $H_p(0,07)$ and $H'(0,07)$

The determination of the conventional quantity value of the dose equivalent quantities $H_p(0,07)$ and $H'(0,07)$ is based on the determination of the conventional quantity value of the air kerma K_a plus the application of a conversion coefficient. For matched reference fields the conversion coefficients given in ISO 4037-3 for the dose equivalent quantities $H_p(0,07)$ and $H'(0,07)$ shall be applied, while for characterized reference fields the individually determined conversion coefficients shall be used. Using the conventional quantity value of the air kerma K_a as determined in [8.3.2](#), leads to:

$$H_p(0,07; \rho_{cal}) = h_{pK}(0,07) K_a(\rho_{cal}) \quad (3)$$

$$H'(0,07; \rho_{cal}) = h'_K(0,07) K_a(\rho_{cal}) \quad (4)$$

8.3.4 Conventional quantity value of the dose equivalent quantities $H_p(10)$ or $H^*(10)$ and $H_p(3)$ or $H'(3)$

8.3.4.1 Corrections of $h_{pK}(10, \alpha)$ or $h^*_{pK}(10)$ and $h_{pK}(3, \alpha)$ or $h'_K(3, \alpha)$ for air density

If the air density, ρ_{cal} , prevailing during calibration of the instrument differs from the air density, ρ_{spec} , prevailing during the determination of the conversion coefficient using spectrometry (see [6.3](#)) by more than the limits given in [Table 2](#), then, in addition to the correction of the air kerma, K_a , the correction

method given in [Annex A](#) shall also be applied to the conversion coefficients $h_{pK}(10, \alpha)$ or $h_{K}^*(10)$ and $h_{pK}(3, \alpha)$ or $h'_{K}(3, \alpha)$ as follows:

$$h_{pK}(10, \alpha, \rho_{\text{cal}}) = \frac{k[\rho_{\text{cal}}, h_{pK}(10, \alpha)]}{k[\rho_{\text{spec}}, h_{pK}(10, \alpha)]} h_{pK}(10, \alpha, \rho_{\text{spec}}) \text{ or} \quad (5)$$

$$h_{K}^*(10, \rho_{\text{cal}}) = \frac{k[\rho_{\text{cal}}, h_{K}^*(10)]}{k[\rho_{\text{spec}}, h_{K}^*(10)]} h_{K}^*(10, \rho_{\text{spec}}) \text{ and} \quad (6)$$

$$h_{pK}(3, \alpha, \rho_{\text{cal}}) = \frac{k[\rho_{\text{cal}}, h_{pK}(3, \alpha)]}{k[\rho_{\text{spec}}, h_{pK}(3, \alpha)]} h_{pK}(3, \alpha, \rho_{\text{spec}}) \text{ or} \quad (7)$$

$$h'_{K}(3, \alpha, \rho_{\text{cal}}) = \frac{k[\rho_{\text{cal}}, h'_{K}(3, \alpha)]}{k[\rho_{\text{spec}}, h'_{K}(3, \alpha)]} h'_{K}(3, \alpha, \rho_{\text{spec}}) \quad (8)$$

For the air density correction factors $k[\rho, h_{pK}(10, \alpha)]$ or $k[\rho, h_{K}^*(10)]$ and $k[\rho, h_{pK}(3, \alpha)]$ or $k[\rho, h'_{K}(3, \alpha)]$ for the conversion coefficients $h_{pK}(10, \alpha)$ or $h_{K}^*(10)$ and $h_{pK}(3, \alpha)$ or $h'_{K}(3, \alpha)$, respectively, see [A.2](#).

8.3.4.2 Evaluation of the effect of angle of radiation incidence α for $H_p(10)$, $H_p(3)$ and $H'(3)$

For a given value of K_a and parallel radiation incidence the conventional quantity value of the dose equivalent quantities $H_p(10)$, $H_p(3)$ and $H'(3)$ is changed if the angle of radiation incidence is changed; this is not the case for the dose equivalent quantity $H^*(10)$. [Table 3](#) gives for unidirectional radiation fields, as an example, values for the change of the angle of radiation incidence that cause a change in the value of the dose equivalent quantity $H_p(10)$ of 2 %. For simplicity, the same values are also assumed for $H_p(3)$ and $H'(3)$. The angle of radiation incidence shall be within the limits given in [Table 3](#), otherwise the uncertainty shall be determined individually, e.g., by performing special calculations.

NOTE 1 All calculations in [8.3.4.2](#) are based on the following assumption. For the purpose of calculating changes of the value of the dose equivalent quantity $H_p(10)$ for a given radiation quality, the respective conversion coefficient can be replaced by the monoenergetic one for the mean energy.

NOTE 2 The adjustment of the angle of radiation incidence α needs two steps, firstly the adjustment of 0° incidence and secondly a turn of the device of α . If the uncertainty of the second step is lower than that of the first step, then two measurements at two angles of radiation incidence of $+\alpha$ and $-\alpha$ are recommended. The mean value of the two measured values is taken as the value for the angle of radiation incidence α , which compensates (to the first order) the error of the adjustment of 0° incidence.

Table 3 — Change $\Delta\alpha$ of the angle of radiation incidence α that causes a change of $H_p(10)$ of 2 % at 2,5 m distance of the point of test from the focus of the X-ray tube

Radiation quality	Mean energy keV	$\Delta\alpha$ in degrees causing a change of $H_p(10)$ of 2 % for angle of incidence of					
		0°	15°	30°	45°	60°	75°
L-10	9,0	2,0	0,93	0,38	0,17	0,016	$(8,8 \cdot 10^{-6})^a$
L-20	17,3	10	4,8	1,9	0,90	0,41	0,083
L-30	26,6	16	10	4,2	1,9	0,83	0,33
N-10	8,5	1,8	0,85	0,34	0,15	0,011	$(2,7 \cdot 10^{-6})^a$
N-15	12,4	4,4	2,0	0,81	0,40	0,17	0,0078
N-20	16,3	10	4,2	1,7	0,79	0,36	0,066
N-25	20,3	17	7,1	2,6	1,2	0,54	0,15
N-30	24,6	15	9,3	3,7	1,7	0,75	0,28
H-10	8,0	1,6	0,80	0,31	0,13	0,0087	$(1,2 \cdot 10^{-6})^a$
H-20	13,1	6,4	2,6	1,0	0,52	0,24	0,021
H-30	19,7	17	6,9	2,5	1,2	0,53	0,14

^a Not achievable in practice.**8.3.4.3 Determination of the conventional quantity value of $H_p(10)$ or $H^*(10)$ and $H_p(3)$ or $H'(3)$**

The determination of the conventional quantity value of the dose equivalent quantities $H_p(10)$ or $H^*(10)$ and $H_p(3)$ or $H'(3)$ is based on the determination of the conventional quantity value of the air kerma, K_a , plus the application of a conversion coefficient. Using the conventional quantity value of the air kerma, K_a , as determined in 8.3.2 leads to:

$$H_p(10; \rho_{\text{cal}}) = h_{pK}(10, \alpha, \rho_{\text{cal}}) K_a(\rho_{\text{cal}}) \text{ or} \quad (9)$$

$$H^*(10; \rho_{\text{cal}}) = h_K^*(10, \rho_{\text{cal}}) K_a(\rho_{\text{cal}}) \text{ and} \quad (10)$$

$$H_p(3; \rho_{\text{cal}}) = h_{pK}(3, \alpha, \rho_{\text{cal}}) K_a(\rho_{\text{cal}}) \text{ or} \quad (11)$$

$$H'(3; \rho_{\text{cal}}) = h'_K(3, \alpha, \rho_{\text{cal}}) K_a(\rho_{\text{cal}}) \quad (12)$$

Formulae (9) to (12) are used to determine the conversion coefficients.

8.3.5 Performing the calibration

The calibration is done according to ISO 4037-3 using the conventional quantity values determined above.

8.4 Calibration by using reference instruments which measure the ICRU dose equivalent quantities**8.4.1 General**

This subclause is relevant for method II only. Within the time period of typically one hour from the measurement of the conventional quantity value of the ICRU dose equivalent quantities with the reference instrument to the calibration of the instrument under test the requirements on tube potential of 6.2 shall be followed. In addition, the air density shall be stable within the limits given in Table 4, otherwise the stated corrections, provided in Annex A, shall be applied.

The additional corrections for the use of a monitor chamber as a transfer device are also provided in Annex A.

As an example, [Table 4](#) gives values for the percentage change of air density that cause a change in the value of the dose equivalent quantities $H_p(10, 0^\circ)$ or $H^*(10)$ and $H_p(10, 60^\circ)$ of 2 % at 2,5 m distance of the point of test from the focus and for 0° and 60° radiation incidence. These conditions are representative for calibrations with respect to $H_p(10)$ performed on a ISO water slab phantom.

8.4.2 Conventional quantity value of the dose equivalent quantities $H_p(10)$ or $H^*(10)$ and $H_p(3)$ or $H'(3)$

8.4.2.1 Correction of $H_p(10)$ or $H^*(10)$ and $H_p(3)$ or $H'(3)$ for air density

Within the short time period (typically one or a few hours) from the measurement of the conventional quantity value of $H_p(10)$ or $H^*(10)$ and $H_p(3)$ or $H'(3)$ to the calibration of the instrument the air density shall not change by more than the limits given in [Table 4](#). For simplicity, the same values are also assumed for $H_p(3)$ or $H'(3)$. These data are valid for a distance of 2,5 m, which is typical for calibrations with respect to $H_p(10)$ performed on an ISO water slab phantom. Normally, these air density requirements are fulfilled and no correction is necessary, in the other few cases the correction method given in the [Annex A](#) shall be applied as follows. If ρ_{con} is the air density prevailing during determination of the conventional quantity value of $H_p(10)$ or $H^*(10)$ and $H_p(3)$ or $H'(3)$ and ρ_{cal} that during calibration of the instrument, then the conventional quantity value of $H_p(10)$ or $H^*(10)$ and $H_p(3)$ or $H'(3)$ during calibration is:

$$H_p(10, \rho_{\text{cal}}) = \frac{k[\rho_{\text{cal}}, H_p(10)]}{k[\rho_{\text{con}}, H_p(10)]} H_p(10, \rho_{\text{con}}) \quad (13)$$

$$H^*(10, \rho_{\text{cal}}) = \frac{k[\rho_{\text{cal}}, H^*(10)]}{k[\rho_{\text{con}}, H^*(10)]} H^*(10, \rho_{\text{con}}) \quad (14)$$

$$H_p(3, \rho_{\text{cal}}) = \frac{k[\rho_{\text{cal}}, H_p(3)]}{k[\rho_{\text{con}}, H_p(3)]} H_p(3, \rho_{\text{con}}) \quad (15)$$

$$H'(3, \rho_{\text{cal}}) = \frac{k[\rho_{\text{cal}}, H'(3)]}{k[\rho_{\text{con}}, H'(3)]} H'(3, \rho_{\text{con}}) \quad (16)$$

For the air density correction factors $k[\rho, H_p(10)]$ or $k[\rho, H^*(10)]$ and $k[\rho, H_p(3)]$ or $k[\rho, H'(10)]$ for the quantities $H_p(10)$ or $H^*(10)$ and $H_p(3)$ or $H'(3)$, respectively, see [A.2](#).

Table 4 — Percentage change of air density that causes a change in the value of $H_p(10, 0^\circ)$ or $H^*(10)$ and $H_p(10, 60^\circ)$ of 2 % at 2,5 m distance of the point of test from the focus and 0° and 60° radiation incidence

Radiation quality	Tube potential U kV	$\Delta\rho/\rho$ for 2,5 m distance causing a change of 2 % of the value of	
		$H_p(10, 0^\circ)$ or $H^*(10)$ %	$H_p(10, 60^\circ)$ %
L-10	10	1,1	1,2
L-20	20	5,9	6,2
L-30	30	15	15
N-10	10	1,1	1,2
N-15	15	2,8	3,1
N-20	20	5,2	5,7
N-25	25	8,3	8,8
N-30	30	12	13

Table 4 (continued)

Radiation quality	Tube potential <i>U</i> kV	$\Delta\rho/\rho$ for 2,5 m distance causing a change of 2 % of the value of	
		$H_p(10, 0^\circ)$ or $H^*(10)$ %	$H_p(10, 60^\circ)$ %
H-10	10	1,1	1,2
H-20	20	4,0	4,9
H-30	30	7,3	8,7

If a monitor chamber is used as a transfer device for the measuring quantity $H_p(10)$ or $H^*(10)$ and $H_p(3)$ or $H'(3)$ then the difference of the air density prevailing during the calibration of the monitor chamber and the air density prevailing during the calibration of the instrument shall be within the limits given in [Table 4](#). Otherwise, the correction method given in [Annex A](#) shall be applied as follows. If the monitor chamber is mounted at a distance d_{MC} from the beam exit window, ρ_{MC} is the air density prevailing during calibration of the monitor chamber and ρ_{cal} those during calibration of the instrument at the distance d_{air} , then the conventional quantity value of $H_p(10)$ or $H^*(10)$ and $H_p(3)$ or $H'(3)$ during calibration is:

$$H_p(10, \rho_{cal}) = \frac{k[\rho_{cal}, H_p(10)]}{k_{MC}[\rho_{MC}, H_p(10)]} H_p(10, \rho_{MC}) \text{ or} \quad (17)$$

$$H^*(10, \rho_{cal}) = \frac{k[\rho_{cal}, H^*(10)]}{k_{MC}[\rho_{MC}, H^*(10)]} H^*(10, \rho_{MC}) \text{ and} \quad (18)$$

$$H_p(3, \rho_{cal}) = \frac{k[\rho_{cal}, H_p(3)]}{k_{MC}[\rho_{MC}, H_p(3)]} H_p(3, \rho_{MC}) \text{ or} \quad (19)$$

$$H'(3, \rho_{cal}) = \frac{k[\rho_{cal}, H'(3)]}{k_{MC}[\rho_{MC}, H'(3)]} H'(3, \rho_{MC}) \quad (20)$$

For the air density correction factors $k[\rho, H_p(10)]$, $k_{MC}[\rho, H_p(10)]$ or $k[\rho, H^*(10)]$, $k_{MC}[\rho, H^*(10)]$ and $k[\rho, H_p(3)]$, $k_{MC}[\rho, H_p(3)]$ or $k[\rho, H'(3)]$, $k_{MC}[\rho, H'(3)]$ and for the quantities $H_p(10)$ or $H^*(10)$ and $H_p(3)$ or $H'(3)$, respectively, see [A.2](#).

8.4.2.2 Adjustment of angle of radiation incidence α for $H_p(10)$ and $H_p(3)$ or $H'(3)$

The value of the dose equivalent quantities $H_p(10)$ and $H_p(3)$ or $H'(3)$ depends on the angle of radiation incidence; this is not the case for the dose equivalent quantity $H^*(10)$. [Table 3](#) (see [8.3.4.2](#)) gives for unidirectional radiation values for the change of the angle of radiation incidence that causes a change in the value of the dose equivalent quantity $H_p(10)$ of 2 %. For simplicity, the same values are also assumed for $H_p(3)$ and $H'(3)$. The angle of radiation incidence shall be within the limits given in [Table 3](#), otherwise the uncertainty shall be determined individually.

8.4.2.3 Determination of the conventional quantity value of $H_p(10)$ or $H^*(10)$ and $H_p(3)$ or $H'(3)$

The conventional quantity values of $H_p(10)$ or $H^*(10)$ and $H_p(3)$ or $H'(3)$ are given by [Formulae \(13\)](#) to [\(16\)](#) or [\(17\)](#) to [\(20\)](#), respectively, without any additional corrections.

8.4.3 Performing the calibration

The calibration is done according to ISO 4037-3 using the conventional quantity values determined above.

8.5 Statement of uncertainty

The determination of uncertainty shall be done in accordance with ISO/IEC Guide 98-3:2008, 8.5, is relevant to methods I and II. Where only one method is affected by a particular point, this is mentioned. The guidelines given in ISO 4037-3:2019, 9.1, shall be followed and the component uncertainties given there shall be used. In addition to or replacing given component uncertainties the following relative standard uncertainties (1 s or $k = 1$) shall be taken into account. These values are only approximations and should not be taken directly. The uncertainties need to be evaluated by each laboratory for their facilities.

- uncertainty resulting from changes of the tube voltage: usually less than 2 % or corrected, should be assessed by the test laboratory;
- uncertainty due to changes in air density: usually less than 2 % or corrected, should be assessed by the test laboratory;
- uncertainty of the conversion coefficients, method I only: usually 1,5 %, should be assessed by the test laboratory;
- uncertainty of the conventional quantity value of $H_p(10)$ or $H^*(10)$, method II only: usually 2,5 %, should be given in the calibration certificate;
- uncertainty due to adjustment of the angle of incidence, $H_p(10)$ only: usually less than 2 %, should be assessed by the test laboratory.

Annex A (normative)

Correction for air density

A.1 General

Corrections for air density are given for all quantities defined in 10 mm and 3 mm depth in tissue for nominal tube potentials equal or greater 10 kV to less or equal 30 kV.

A.2 Method for air density correction

The climatic conditions, given by the air temperature, T , the air pressure, p , and the relative humidity, r , affect the value of the air kerma, K_a , of the conversion coefficients $h_{pK}(10, \alpha)$, $h_{pK}^*(10)$, $h_{pK}(3, \alpha)$ and $h_{pK}'(3)$ from air kerma K_a to the dose equivalent quantities $H_p(10)$, $H^*(10)$, $H_p(3)$ and $H'(3)$ and of their product, the dose equivalent quantities $H_p(10)$, $H^*(10)$, $H_p(3)$ and $H'(3)$ themselves, even if all other conditions of the irradiation facility are constant. The effect is due to the absorption of the photon radiation on the way from the beam exit window of the X-ray tube to the point of test and to the change of this absorption with photon energy. All influences increase with increasing air path. The absorption depends only on the air density ρ . For temperatures between 15 °C and 25 °C, ρ is determined by the following formula, see Drake and Böhm^[2] with modifications for new reference values given in Reference [3]:

$$\rho = \rho_0 \left[1,005\,699 \frac{p}{p_0} - \frac{1}{175,7} \frac{r}{r_0} \left(\frac{T}{T_0} \right)^{17,97} \right] \frac{T_0}{T} \quad (\text{A.1})$$

where

- p is the air pressure, $p_0 = 101,3$ kPa;
- T is the air temperature, $T_0 = 293,15$ K (equivalent to 20 °C);
- r is the relative air humidity, $r = 0,65$ (equivalent to 65 %);
- ρ is the air density for reference conditions, $\rho_0 = 1,197\,4$ kg/m³.

NOTE A change of the air density of 1 % is equivalent to a change of the air pressure from 100 kPa to 101 kPa (if temperature and humidity are unchanged) or to a change of the temperature from 293 K to 296 K (if air pressure and humidity are unchanged). The air pressure changes under normal conditions and altitudes below 1000 m by about -20 % to +10 %.

The correction of each of the above mentioned quantities for air density is performed as follows: For the air kerma K_a , taken as an example, the air density correction factor, $k(\rho, K_a)$, which is the quotient

of the value of the measurand at the air density ρ and the value of the measurand at the air density at reference conditions, ρ_0 , is calculated according to:

$$k(\rho, K_a) = \frac{K_a(\rho)}{K_a(\rho_0)} \quad (\text{A.2})$$

The value under irradiation conditions is then obtained from the value at reference conditions according to:

$$K_a(\rho_{\text{irr}}) = k(\rho_{\text{irr}}, K_a) \cdot K_a(\rho_0) \quad (\text{A.3})$$

where ρ_{irr} being the density of air during irradiation.

A linear approximation to the air density in the range $0,96 \text{ kg/m}^3 \leq \rho \leq 1,32 \text{ kg/m}^3$ of the correction factor leads to:

$$k(\rho) = 1 + m(d_{\text{air}}) \cdot (\rho - \rho_0) \quad (\text{A.4})$$

where d_{air} is the distance from the beam exit window to the reference point. If a monitor chamber at a distance d_{MC} from the beam exit window is used as a transfer device then the correction shall only be applied to the air path from the monitor chamber to the reference point. This leads to:

$$k_{\text{MC}}(\rho, K_a) = 1 + m(d_{\text{air}}) \cdot (\rho - \rho_0) \cdot \left(1 - \frac{d_{\text{MC}}}{d_{\text{air}}}\right) \quad (\text{A.5})$$

The gradients, $m(d_{\text{air}})$, are different for different air paths, d_{air} . For $1,0 \text{ m} \leq d_{\text{air}} \leq 3,0 \text{ m}$, the gradient $m(d_{\text{air}})$ can be approximated by a linear fit, to:

$$m(d_{\text{air}}) = m(1,0 \text{ m}) + (d_{\text{air}} - 1,0 \text{ m})m_d \quad (\text{A.6})$$

The uncertainty of the linear approximations using $m(1,0 \text{ m})$ and m_d compared with the values determined directly according to 8.4.2.1 is less than or equal to 1 % in the range of air density from $\rho = 1,10 \text{ kg/m}^3$ to $\rho = 1,27 \text{ kg/m}^3$. Inclusion of the uncertainties of the calculations themselves, leads to an overall uncertainty for the corrections, $|k(\rho)-1|$, of about 5 %, this in turn resulting in an overall uncertainty for the correction factors, $k(\rho)$, of about 2 %.

Values for these two parameters, $m(1,0 \text{ m})$ and m_d , are given in A.3 for the air kerma K_a and for the conversion coefficients $h_{\text{pK}}(10, \alpha)$ and $h^*_{\text{K}}(10)$, in A.4 for $H_{\text{p}}(10)$ and $H^*(10)$ and in A.5 for the conversion coefficients $h_{\text{pK}}(3, \alpha)$ and $h'_{\text{K}}(3, \alpha)$. No published data are available for $H_{\text{pK}}(3, \alpha)$ and $H'(3, \alpha)$. They may be calculated from the data given in A.3 for K_a and in A.5 for the conversion coefficients $h_{\text{pK}}(3, \alpha)$ and $h'_{\text{K}}(3, \alpha)$.

A.3 Air density correction parameters for K_a , $h_{\text{pK}}(10, \alpha)$ and $h^*_{\text{K}}(10)$

Tables A.1 and A.2 give, as an example, values for the two parameters $m(1,0 \text{ m})$ and m_d , for the quantity K_a and the conversion coefficients $h_{\text{pK}}(10, \alpha)$ and $h^*_{\text{K}}(10)$. Data are taken from Behrens[8].

NOTE The data can be slightly different from one X-ray facility to another, but the differences can be neglected for the range of air density from $\rho = 0,96 \text{ kg/m}^3$ to $\rho = 1,32 \text{ kg/m}^3$.

A.4 Air density correction parameters for $H_{\text{p}}(10)$ and $H^*(10)$

Tables A.3 and A.4 give, as an example, values for the two parameters $m(1,0 \text{ m})$ and m_d , for the quantities $H_{\text{p}}(10)$ and $H^*(10)$. Data are taken from Behrens[5].

NOTE The data can be slightly different from one X-ray facility to another, but the differences can be neglected for the range of air density from $\rho = 0,96 \text{ kg/m}^3$ to $\rho = 1,32 \text{ kg/m}^3$.

A.5 Air density correction parameters for $h_{pK}(3, \alpha)$ and $h'_{K}(3, \alpha)$

Tables A.5 to A.8 give, as an example, values for the two parameters $m(1,0 \text{ m})$ and m_d , for the conversion coefficients $h_{pK}(3, \alpha)$ and $h'_{K}(3, \alpha)$. Data are taken from Behrens[8].

NOTE The data can be slightly different from one X-ray facility to another, but the differences can be neglected for the range of air density from $\rho = 0,96 \text{ kg/m}^3$ to $\rho = 1,32 \text{ kg/m}^3$.

Table A.1 — Parameter $m(1,0 \text{ m})$ for the simple approximation of $m(d_{\text{air}})$ given in Formula (A.6)

Radiation quality	$m(1,0 \text{ m})$ in m^3/kg for the quantity							
	K_a	$h^*_{K}(10)$	$h_{pK}(10, 0^\circ)$	$h_{pK}(10, 15^\circ)$	$h_{pK}(10, 30^\circ)$	$h_{pK}(10, 45^\circ)$	$h_{pK}(10, 60^\circ)$	$h_{pK}(10, 75^\circ)$
L-10	-0,758	0,153	0,153	0,155	0,161	0,173	—	—
L-20	-0,123	0,009 98	0,009 95	0,010 1	0,010 9	0,012 4	0,015 4	0,020 9
L-30	-0,046 8	0,000 929	0,000 947	0,000 951	0,001 00	0,001 12	0,001 37	0,002 01
N-10	-0,970 ^a	0,342	0,342	0,346	0,355	0,371	—	—
N-15	-0,343	0,097 9	0,097 7	0,099 0	0,103	0,110	0,126	0,146
N-20	-0,163	0,032 0	0,032 2	0,032 7	0,034 6	0,038 1	0,044 3	0,054 3
N-25	-0,092 7	0,011 2	0,010 9	0,011 1	0,011 8	0,013 0	0,015 6	0,020 8
N-30	-0,058 7	0,003 41	0,003 38	0,003 39	0,003 60	0,004 04	0,004 96	0,007 19
W-30	-0,075 8	0,009 72	0,009 56	0,009 63	0,010 2	0,011 4	0,013 7	0,018 5
H-10	-1,26 ^a	0,625	0,625	0,628	0,640	0,659	—	—
H-20	-0,455 ^a	0,276	0,276	0,278	0,285	0,297	0,316	0,337
H-30	-0,187 ^a	0,091 1	0,090 8	0,091 5	0,094 4	0,099 3	0,108	0,120

NOTE All data given in this table are calculated from the spectra given by Ankerhold[1], for different spectra the data may be different.

^a These values are only valid in the range from $\rho = 1,10 \text{ kg/m}^3$ to $\rho = 1,27 \text{ kg/m}^3$.

Table A.2 — Parameter m_d for the simple approximation of $m(d_{\text{air}})$ given in Formula (A.6)

Radiation quality	m_d in m^2/kg for the quantity							
	K_a	$h^*_{K}(10)$	$h_{pK}(10, 0^\circ)$	$h_{pK}(10, 15^\circ)$	$h_{pK}(10, 30^\circ)$	$h_{pK}(10, 45^\circ)$	$h_{pK}(10, 60^\circ)$	$h_{pK}(10, 75^\circ)$
L-10	-0,661 ^a	0,0747	0,0747	0,076 1	0,080 8	0,089 3	—	—
L-20	-0,128	0,0169	0,016 9	0,017 0	0,017 8	0,019 2	0,021 8	0,026 8
L-30	-0,046 5	0,000 900	0,000 918	0,000 922	0,000 97	0,001 09	0,001 32	0,001 93
N-10	-0,689 ^a	0,088 0	0,088 0	0,089 9	0,095 9	0,107	—	—
N-15	-0,288	0,056 0	0,056 1	0,057 1	0,060 1	0,065 6	0,076 8	0,094 2
N-20	-0,149	0,023 4	0,023 5	0,023 8	0,025 3	0,028 1	0,033 1	0,041 7
N-25	-0,087 8	0,008 47	0,008 25	0,008 37	0,008 95	0,010 0	0,012 3	0,017 0
N-30	-0,058 1	0,003 39	0,003 36	0,003 38	0,003 58	0,004 01	0,004 9	0,007 03
W-30	-0,072 7	0,008 39	0,008 25	0,008 30	0,008 82	0,009 82	0,011 9	0,016 2
H-10	-0,663 ^b	0,051 8	0,051 7	0,054 2	0,061 3	0,074 1	—	—
H-20	-0,273 ^a	0,114	0,115	0,116	0,12	0,128	0,141	0,158
H-30	-0,125	0,038 1	0,037 9	0,038 3	0,040 3	0,043 8	0,050 1	0,060 5

NOTE All data given in this table are calculated from the spectra given by Ankerhold[1], for different spectra the data may be different.

^a These values are only valid in the range from $\rho = 1,10 \text{ kg/m}^3$ to $\rho = 1,27 \text{ kg/m}^3$.

^b These values are only valid in the range from $\rho = 1,15 \text{ kg/m}^3$ to $\rho = 1,23 \text{ kg/m}^3$.