
**Anaesthetic vaporizers — Agent-specific
filling systems**

*Évaporateurs d'anesthésie — Systèmes de remplissage spécifiques à
l'agent*

STANDARDSISO.COM : Click to view the full PDF of ISO 5360:2012



STANDARDSISO.COM : Click to view the full PDF of ISO 5360:2012



COPYRIGHT PROTECTED DOCUMENT

© ISO 2012

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Bottle.....	2
5 Bottle collar	4
6 Bottle adaptor	7
7 Filler receptacle	11
8 Filling rate	13
9 Leakage	15
10 Overfilling protection	15
11 Colour coding	15
12 Usability.....	15
13 Clinical evaluation	15
14 Information provided by the manufacturer.....	16
14.1 Marking.....	16
14.2 Labelling.....	16
14.3 Instructions for use	16
Annex A (informative) Recommendations on materials	18
Annex B (informative) Types of agent-specific filling systems.....	19
Annex C (normative) Determination of total leakage into atmosphere of anaesthetic agent during filling.....	20
Bibliography.....	21

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 5360 was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 1, *Breathing attachments and anaesthetic machines*.

This third edition cancels and replaces the second edition (ISO 5360:2006), of which it constitutes a minor revision. In particular, it

- indicates in the Scope that requirements of agent-specific filling systems for anaesthetic vaporizers (not merely the dimensions) are specified,
- transfers the recommendations on materials from the Scope to an informative annex,
- refers to substances which are carcinogenic, mutagenic or toxic to reproduction in Clause 9 (leakage),
- introduces new requirements on usability (Clause 12) and clinical evaluation (Clause 13), and
- amends the requirements on information provided by the manufacturer (renumbered Clause 14).

Anaesthetic vaporizers — Agent-specific filling systems

1 Scope

This International Standard specifies requirements, including dimensions, for agent-specific filling systems for agent-specific anaesthetic vaporizers.

This International Standard does not specify construction materials.

NOTE 1 For recommendations on materials, see Annex A.

Because of the unique properties of desflurane, dimensions for this agent have not been specified in this International Standard.

NOTE 2 Designs of connection systems, which only permit engagement of the agent-specific bottle adaptor to the bottle when the bottle collar is in place, are encouraged.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1101, *Geometrical Product Specifications (GPS) — Geometrical tolerancing — Tolerances of form, orientation, location and run-out*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

agent-specific

having both a prescribed configuration and prescribed dimensions, which are specific for a prescribed liquid anaesthetic agent

3.2

agent-specific filling system

functional system of agent-specific coded connections between an anaesthetic bottle and an agent-specific anaesthetic vaporizer, consisting of, for example, a threaded bottle neck with collar, bottle connector, male adaptor and filler receptacle

NOTE Different types of agent-specific filling systems are shown in Annex B.

3.3

anaesthetic vaporizer

device designed to facilitate the change of an anaesthetic agent from a liquid to a vapour


- 3.4
bottle adaptor**
assembly that is intended to connect a bottle for liquid anaesthetic agent to an agent-specific anaesthetic vaporizer
- 3.5
bottle collar**
agent-specific component on the neck of a bottle causing it to be agent-specific
- 3.6
bottle connector**
agent-specific component that fits the thread on the bottle neck and mates with the agent-specific bottle collar
- 3.7
bottle neck**
external threaded part of the bottle and the adjacent contour over which an agent-specific collar is fitted
- 3.8
filler receptacle**
receptacle for a bottle or a bottle adaptor on an agent-specific anaesthetic vaporizer
- 3.9
male adaptor**
part of a bottle adaptor that mates with a filler receptacle on an agent-specific vaporizer

4 Bottle

Each bottle shall have:

- a) the name of the anaesthetic agent with which it is intended to be used marked on it;
- b) either a bottle collar complying with Clause 5 and a threaded neck complying with Figure 1 and Table 1, or a permanently attached bottle adaptor complying with 6.2.

Dimensions in millimetres

**Key**

1 optional pouring lip (dimension not specified)

a Flatness and perpendicularity tolerances in accordance with ISO 1101.

NOTE The dimensions shown form part of this International Standard. Other features are for illustrative purposes only. See Table 1.

Figure 1 — Two examples of threaded necks of bottles for anaesthetic agents**Table 1 — Dimensions of threaded necks of bottles for anaesthetic agents**

Bottle type	Anaesthetic agent	h_1 $\pm 0,3$ mm	h_2^a min. mm	s $\pm 0,45$ mm	β	α min. at ϕD_1	p mm	Thread turns min.	D_1^a nom. mm	D_2^b $\pm 0,3$ mm	D_3^b $\pm 0,3$ mm	d max. mm
1	Isoflurane Enflurane	9,75	23	1,2	2° 35'	30°	3,2	1	28	23,6	21,5	28
2	Halothane	6,8	18,7	1,2	2° 15'	30°	2,54	1,25	24	21,45	19,7	28
3	Halothane (USA)	15	26,3	1	2° 50'	30°	3,2	1,75	24	21,7	19,5	28
4	Spare	9,05	20	1,15	3° 30'	30°	3,2	1,25	20	17,65	15,5	28
5	Spare	9,05	20	1,15	3° 7'	30°	3,2	1,25	22	19,65	17,5	28
6	Methoxy- flurane	9,8	20	1,15	2° 57'	30°	4,25	1,25	30	27,3	24,9	32
7	Spare	9,85	20	1,15	2° 31'	30°	4,25	1,25	34	31,8	29,4	32
8	Sevoflurane	8,9	23,9	1,3	2° 56'	30°	3,63	1,25	23,9	23,5	21,5	28

NOTE See Figure 1.

^a Recommended values.^b Summation of the tolerances of measures D_2 and D_3 shall be avoided. A maximum tolerance of $\pm 0,3$ mm for $(D_2 - D_3)$ should be required to avoid problems with the fitting of any bottle connector.

5 Bottle collar

5.1 Bottle collars shall comply with the configuration and dimensions shown in Figure 2 and angle, θ , specified in Table 2 for the anaesthetic agent with which it is intended to be used.

5.2 The position of the bottle collar relative to the screw thread of the bottle shall be as shown in Figure 3.

5.3 The bottle collar shall be attached to the bottle and shall be rotatable by hand.

Dimensions in millimetres





Figure 2 — Configuration of agent-specific bottle collars (continued)

Dimensions in millimetres



b) Bottle collar for large bottles, i.e. types 6 and 7

Key

- 1 face A
- a See Table 2.
- b May vary to suit bottle.

Figure 2 — Configuration of agent-specific bottle collars

Table 2 — Dimensions and colours of agent-specific bottle collars and connectors

Anaesthetic agent	θ^a $\pm 0^\circ 30'$	Specified colour ^b	Example of colour samples					
			Federal Standard 595 colour	BS 5252 colour	Pantone colour	SS 01 91 02 colour	Munsell colour ^c	DIN 6164-2 colour
Halothane	-20°	Red	11 105	04 E 56	200 C	NCS S 1080 R	5R4/14	8:7:2
Enflurane	$+20^\circ$	Orange	22 510	06 E 55	151 C	NCS S 0585-Y50R	2,5YR 6/16	5:5:1
Methoxy-flurane	0°	Green	14 187	14 E 53	334 C	NCS S 2060-B90G	10G 5/10	21:6:3
Desflurane	N.S. ^d	Blue	N/A ^e	18 E 53	3015 C	NCS S 3060 B	10B 4/10	18:4:3
Not for agent identification		White	37 875	18 B 15	5455 C	NCS S 0502-B	10B 9/1	N:0:0.5
Not for agent identification		Black	15 042	00 E 53	Process black C	NCS S 9000-N	N 0,5	N:0:9
Sevoflurane	$+50^\circ$	Yellow	N/A ^e	10 E 53	108 C	NCS S 0570-Y	6,25Y 8,5/12	2:6:1
Isoflurane	-40°	Purple	N/A ^e	24 E 53	254 C	NCS S 3055-R50B	7,5P4/12	11:4:4
Spare		Grey	16 251	00 A 09	Cool grey 9 C	NCS S 5502 B	5PB 5/1	N:0:4

^a The sign “+” means clockwise rotation and sign “-” means anticlockwise rotation, when viewed from the top.


^b If a colour is used on a vaporizer, bottle or package label to facilitate correct identification, it is important that only the colour for the appropriate anaesthetic agent be used.

^c Munsell colour is the original. Other colour systems show the nearest available colour sample.

^d N.S. means not specified.

^e N/A means not available.


Dimensions in millimetres



a) Position without clearance between collar and transfer ring

Figure 3 — Alternative positions of agent-specific bottle collar (continued)

Dimensions in millimetres



b) Position with clearance between collar and transfer ring

Key

- 1 face A (see Figure 2)
- a Clearance to suit bottle.

Figure 3 — Alternative positions of agent-specific bottle collar**6 Bottle adaptor**

6.1 If the bottle adaptor is not permanently attached to the bottle or the vaporizer (see Annex B), it shall include an agent-specific bottle connector complying with the configuration and dimensions specified in Figure 6 for the anaesthetic agent with which it is intended to be used. The bottle connector shall be designed so that the coding slots in the bottle connector engage with the bottle collar before a tight connection is obtained.

If an agent-specific male adaptor is used, it shall comply with the dimensions specified in Figure 4 or Figure 5 for the anaesthetic agent with which it is intended to be used.

6.2 If the bottle adaptor is permanently attached to the bottle and an agent-specific male adaptor is used, the agent-specific male adaptor shall comply with the dimensions specified in Figure 4 or Figure 5 for the anaesthetic agent with which it is intended to be used.

6.3 If the bottle adaptor is a permanent part of the vaporizer, it shall include an agent-specific bottle connector complying with the configuration and dimensions specified in Figure 6 for the anaesthetic agent with which it is intended to be used. The bottle connector shall be designed so that the coding slots in the bottle connector engage with the bottle collar before a tight connection is obtained.

6.4 Bottle adaptor threads shall be designed so that they

- a) ensure an engagement of at least 0,75 thread turns on a threaded neck [see 4 b)] of an anaesthetic bottle, and
- b) withstand, without visible damage, a tightening torque of $(3 \pm 0,3)$ N·m, when fitted to an appropriate bottle.

NOTE The intention of these requirements is to render the bottle adaptor unlikely to be accidentally displaced from the bottle during filling.

6.5 If the bottle adaptor is permanently attached to the bottle (see Annex B), and an agent-specific male adaptor complying with the configuration shown in Figure 4 or Figure 5 is used, means shall be provided for sealing the liquid and air/vapour passages on the adaptor when it is not inserted into the filler receptacle.

6.6 The bottle adaptor shall not break when dropped from a height of 1 m on to a hard surface.





Figure 4 — Configuration and dimensions of agent-specific male adaptors for use with enflurane, methoxyflurane and halothane

Table 3 — Details of male adaptors for use with enflurane, methoxyflurane and halothane

Anaesthetic agent	t $+0,1$ 0 mm	Slot in face
Enflurane	3,5	A
Methoxyflurane	7,5	B
Halothane	3,5	B
Spare	5,5	B
Spare	5,5	A
Spare	1,5	A
Spare	1,5	B

Dimensions in millimetres

**Key**

- | | | | | | |
|---|--------------|---|-----------------|---|--------|
| 1 | sealing face | 3 | air/vapour port | 5 | face C |
| 2 | liquid port | 4 | face A | 6 | face B |


NOTE See Table 4 and Figure 4.

Figure 5 — Configuration and dimensions of agent-specific male adaptors for use with isoflurane and sevoflurane

Table 4 — Details of male adaptors for use with isoflurane and sevoflurane

Anaesthetic agent	Slot position
Isoflurane	Faces A and C
Sevoflurane	Faces B and C


Dimensions in millimetres



a) Connector for small bottles i.e. types 1 to 5 and 8

- a Square corners optional.
- b See Table 2.
- c Space (dimension not specified) for sealing component.
- d May vary to suit bottle.

Figure 6 — Configuration and dimensions of agent-specific bottle connectors (continued)



b) Connector for large bottles i.e. types 6 and 7


- a Square corners optional.
- b See Table 2.
- c Space (dimension not specified) for sealing component.
- d May vary to suit bottle.

Figure 6 — Configuration and dimensions of agent-specific bottle connectors

7 Filler receptacle

7.1 The filler receptacle of the vaporizer shall

- a) comply with the configuration and dimensions shown in Figure 7 or Figure 8 for the anaesthetic agent with which it is intended to be used, and the design shall only permit the insertion of the agent-specific male adaptor complying with 6.1 or 6.2 into the front face of the filler receptacle as illustrated in Figure 7 or Figure 8, or
- b) comply with the configuration and dimensions of the bottle connector shown in Figure 6 and angle, θ , specified in Table 2 for the anaesthetic agent with which it is intended to be used.



Key

- 1 sealing face, space (dimension not specified) for sealing component
- 2 front face
- 3 air/vapour port
- 4 liquid port
- 5 pin
- 6 face A
- 7 face B

NOTE 1 See also Table 5.

NOTE 2 Port identification applies to filling procedure only.

Figure 7 — Configuration and dimensions of agent-specific filler receptacles for use with enflurane, methoxyflurane and halothane

Table 5 — Details of filler receptacles for use with enflurane, methoxyflurane and halothane


Anaesthetic agent	l +0,15 -0,10 mm	Pin in face
Enflurane	7,1	A
Methoxyflurane	11,1	B
Halothane	7,1	B
Spare	9,1	B
Spare	9,1	A
Spare	5,1	A
Spare	5,1	B

7.2 If the filler receptacle is of the type specified in 7.1 a), means shall be provided for tightening the male adaptor against the receptacle seal(s) when the adaptor is inserted into the filler receptacle.

7.3 The filler receptacle shall be provided with means of sealing the liquid and air/vapour passages in the receptacle while the bottle adaptor is not inserted.

8 Filling rate

When tested according to the manufacturer's instructions, the mean filling rate shall exceed 2 ml/s.



Key

- 1 sealing face, space (dimension not specified) for sealing component
- 2 front face
- 3 air/vapour port
- 4 liquid port
- 5 pin
- 6 face A
- 7 face B
- 8 face C

NOTE 1 See Table 6.

NOTE 2 See Figure 7 for all other details.

Figure 8 — Configuration and dimensions of agent-specific filler receptacles for use with isoflurane and sevoflurane

Table 6 — Details of filler receptacles for use with isoflurane and sevoflurane

Anaesthetic agent	Pin inserted in face C and adjacent to:
Isoflurane	Face A
Sevoflurane	Face B

9 Leakage

When measured in accordance with Annex C, the mean leakage of liquid or vaporized anaesthetic agent into the atmosphere shall not exceed 0,5 ml.

It is recognized that during disconnection of the male adaptor from the vaporizer and the bottle adaptor from the bottle, small amounts of anaesthetic agent escape to the environment. This should be noted in the user instruction manual.

Means should be provided to ensure that as little anaesthetic agent as possible escapes from the male adaptor to the environment when the adaptor is affixed to the bottle during storage.

NOTE Attention is drawn to substances which are carcinogenic, mutagenic or toxic to reproduction.

10 Overfilling protection

When filled in accordance with the manufacturer's instructions, it shall not be possible to overfill the vaporizer such that:

- a) its performance is affected;
- b) the fluid level is no longer visible.

11 Colour coding

The bottle collar and the bottle connector shall incorporate the colour coding using the colour specified by name in Table 2 for the anaesthetic agent intended.

If the filler receptacle is colour-coded, the colour shall comply with the colour specified by name in Table 2.

12 Usability

The manufacturer shall address, in a usability engineering process, the risk resulting from poor usability (see IEC 62366).

Check compliance by inspection of the usability engineering file.

13 Clinical evaluation

A clinical evaluation shall be performed and documented in the risk management file.

Check compliance by inspection of the risk management file.

14 Information provided by the manufacturer

14.1 Marking

Agent-specific filling systems or bottle collars or bottle adaptors supplied individually shall be marked with:

- a) the manufacturer's name and/or trademark and where the manufacturer does not have an address within the locale, the name and address of an authorized representative within the locale;
- b) the batch code or the serial number;
- c) the name of the anaesthetic agent with which it is intended to be used.

The use of the generic names of anaesthetic agents according to Table 2 is recommended.

14.2 Labelling

14.2.1 Agent-specific filling systems or components supplied individually shall provide the following information on the device itself, on the unit pack or on a leaflet accompanying the device:

- a) the name and address of the manufacturer/supplier and where the manufacturer does not have an address within the locale, the name and address of an authorized representative within the locale;
- b) the information necessary to identify the device or the contents of the package;
- c) the anaesthetic agent with which the device shall be used;
- d) if appropriate, an indication of the time limit for using the device safely, expressed as year/month;
- e) an indication if the device is for single use only;

NOTE Manufacturer's attention is drawn to consistent use of indication for single-use devices.

- f) any relevant particular storage and/or handling requirements.

If phthalates are incorporated in parts of the device coming directly or indirectly into contact with the patient, the device shall be labelled accordingly (see EN 15896).

14.2.2 The bottle adaptor shall have a leaflet enclosed with the device giving the following warning: "Caution: agent-specific filling cannot be assured when bottles without collars are used".

14.3 Instructions for use

Instructions for use of agent-specific filling systems or components thereof shall be provided by the vaporizer manufacturer or supplier and shall include:

- a) the details referred to in 14.2.1 with the exception of those in items c) and d);
- b) the warning given in 14.2.2;
- c) the information necessary to ensure that the agent-specific filling system is in safe and correct working order;
- d) the details on the nature and frequency of maintenance operations to ensure safe and correct operation at all times;
- e) a statement indicating compliance of the agent-specific filling system with this International Standard;