



INTERNATIONAL STANDARD ISO 5364:1986
TECHNICAL CORRIGENDUM 1

Published 1992-12-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Oropharyngeal airways

TECHNICAL CORRIGENDUM 1

Canules oro-pharyngées

RECTIFICATIF TECHNIQUE 1

Technical corrigendum 1 to International Standard ISO 5364:1986 was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Sub-Committee SC 2, *Tracheal tubes and other equipment*.

Page 2, clause 8

In the second line, after "nominal size", add "in accordance with table 1".

Add a note at the end of the paragraph:

"NOTE – The manufacturer's own size designation may additionally be given."

Table

Number the table "Table 1".

The Secretariat of ISO/TC 121/SC 2 adds the following rationale for these changes.

This Technical Corrigendum corrects an ambiguity in ISO 5364:1986. It has no effect on the technical content but is required to make it clear that while ISO 5364 requires size to be designated by length, in centimetres, manufacturers may in addition use their own size designation. Such additional means of size designation where familiar to particular users may assist in the selection of a device of appropriate size, and therefore safety.

UDC 615.473

Ref. No. ISO 5364:1986/Cor.1:1992 (E)

Descriptors: medical equipment, pharyngeal cannulae, specifications, dimensions, designation, marking, packing.

© ISO 1992

Printed in Switzerland