

# INTERNATIONAL STANDARD

**ISO  
5364**

Third edition  
2001-09-01

---

---

## **Anaesthetic and respiratory equipment — Oropharyngeal airways**

*Matériel d'anesthésie et de réanimation respiratoire — Canules  
oropharyngées*

STANDARDSISO.COM : Click to view the full PDF of ISO 5364:2001



Reference number  
ISO 5364:2001(E)

© ISO 2001

**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 5364:2001

© ISO 2001

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.ch](mailto:copyright@iso.ch)  
Web [www.iso.ch](http://www.iso.ch)

Printed in Switzerland

## Contents

|                                                             | Page |
|-------------------------------------------------------------|------|
| 1 Scope .....                                               | 1    |
| 2 Normative references .....                                | 1    |
| 3 Terms and definitions .....                               | 1    |
| 4 Size designation and dimensions .....                     | 2    |
| 5 Materials .....                                           | 3    |
| 6 Design .....                                              | 3    |
| 7 Performance requirements .....                            | 3    |
| 8 Sterility assurance .....                                 | 4    |
| 9 Packaging of oropharyngeal airways supplied sterile ..... | 4    |
| 10 Marking .....                                            | 4    |
| 11 Information to be supplied by the manufacturer .....     | 5    |

## Annexes

|                                                                     |    |
|---------------------------------------------------------------------|----|
| A Test method for resistance to collapse of the buccal portion..... | 6  |
| A.1 Principle .....                                                 | 6  |
| A.2 Apparatus .....                                                 | 6  |
| A.3 Test procedure .....                                            | 6  |
| A.4 Expression of result .....                                      | 6  |
| B Test method for patency of lumen .....                            | 8  |
| B.1 Principle .....                                                 | 8  |
| B.2 Apparatus .....                                                 | 8  |
| B.3 Test procedure .....                                            | 8  |
| B.4 Expression of results .....                                     | 8  |
| C Guidance on materials and design .....                            | 10 |
| Bibliography.....                                                   | 11 |

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

International Standard ISO 5364 was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 2, *Tracheal tubes and other equipment*.

This third edition cancels and replaces the second edition (ISO 5364:1986), which has been technically revised. This edition introduces a test for patency of lumen as well as a test for resistance to collapse of the buccal end. Metal airways have been excluded from this scope because of their association with dental trauma.

Annexes A and B form a normative part of this International Standard. Annex C is for information only.

## Introduction

This International Standard specifies dimensions and other requirements for oropharyngeal airways.

Airway size is designated by length, which is important when selecting an oropharyngeal airway to hold forward the base of the tongue to prevent obstruction of the human airway by the soft tissue.

Flammability of oropharyngeal airways, for example if flammable anaesthetics, electrosurgical units or lasers are used, is a well-recognized hazard<sup>1)</sup>. It is addressed by appropriate clinical management, which is outside the scope of this International Standard.

STANDARDSISO.COM : Click to view the full PDF of ISO 5364:2001

---

1) See ISO/TR 11991.

[STANDARDSISO.COM](https://standardsiso.com) : Click to view the full PDF of ISO 5364:2001

# Anaesthetic and respiratory equipment — Oropharyngeal airways

## 1 Scope

This International Standard specifies requirements for oropharyngeal airways of plastics materials and/or rubber, including those with a reinforcement insert made of plastics materials and/or metal.

This International Standard is not applicable to metal oropharyngeal airways, nor to requirements concerning flammability of oropharyngeal airways.

## 2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this International Standard. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing*

ISO 11607, *Packaging for terminally sterilized medical devices*

EN 556:1994, *Sterilization of medical devices — Requirements for medical devices to be labelled "STERILE"*

## 3 Terms and definitions

For the purposes of this International Standard, the following terms and definitions apply.

### 3.1

#### **oropharyngeal airway**

device intended to maintain a gas pathway through the oral cavity and pharynx

[ISO 4135]

### 3.2

#### **pharyngeal end**

that end of an oropharyngeal airway which is intended to be inserted into a patient's oropharynx

[ISO 4135]

### 3.3

#### **flanged end**

that end of an oropharyngeal airway which is flanged and is intended to be external to the teeth or gums


[ISO 4135]

## 4 Size designation and dimensions

### 4.1 Size designation

The size of oropharyngeal airways shall be designated by the nominal length (see  $l$ , Figure 1) expressed in centimetres, in accordance with Table 1.

NOTE The manufacturer's own size designation may additionally be given, but this is not recommended.



#### Key

- 1 Buccal portion
- 2 Reinforcement insert, if provided
- 3 Position for measuring minimum inside dimension (see Table 1)
- 4 Flanged end

<sup>a</sup> For  $l$  see 4.1 and 4.2.1.

**Figure 1 — Dimensions for size designation of oropharyngeal airways**

**Table 1 — Size designation of oropharyngeal airways — Dimensions and tolerances**

| Designated size<br>(nominal length)<br>cm | Length and tolerance<br>mm                              | Minimum inside<br>dimension<br>mm |
|-------------------------------------------|---------------------------------------------------------|-----------------------------------|
| 3                                         | $30 \pm 2,5$                                            | 2,5                               |
| 3,5                                       | $35 \pm 2,5$                                            | 3,0                               |
| 4                                         | $40 \pm 2,5$                                            | 3,0                               |
| 4,5                                       | $45 \pm 2,5$                                            | 3,0                               |
| 5                                         | $50 \pm 2,5$                                            | 3,5                               |
| 5,5                                       | $55 \pm 2,5$                                            | 3,5                               |
| 6                                         | $60 \pm 2,5$                                            | 4,0                               |
| 6,5                                       | $65 \pm 2,5$                                            | 4,0                               |
| 7                                         | $70 \begin{smallmatrix} +5,0 \\ -2,5 \end{smallmatrix}$ | 4,0                               |
| 8                                         | $80 \pm 5,0$                                            | 4,5                               |
| 9                                         | $90 \pm 5,0$                                            | 4,5                               |
| 10                                        | $100 \pm 5,0$                                           | 5,0                               |
| 11                                        | $110 \pm 5,0$                                           | 5,5                               |
| 12                                        | $120 \pm 5,0$                                           | 5,5                               |

## 4.2 Dimensions

**4.2.1** The length (see  $l$ , Figure 1) shall be in accordance with Table 1.

**4.2.2** The minimum inside dimension at any point along the length of the airway shall be not less than that specified in Table 1.

NOTE This dimension is relevant to the ability to pass other devices, e.g. a suction catheter, through the airway.

## 5 Materials

Oropharyngeal airways, in their ready-for-use state after any preparation for use recommended by the manufacturer, shall satisfy appropriate biological safety testing, as indicated in ISO 10993-1.

## 6 Design

Edges and corners intended to come into contact with the patient's tissues shall have a minimum radius of curvature of 0,5 mm.

## 7 Performance requirements

### 7.1 Resistance to collapse of the buccal portion

When tested in accordance with annex A, the minimum inside dimension of the buccal portion of the airway shall be not less than 75 % of that given in Table 1 for the size of airway being tested.

### 7.2 Patency of lumen

When tested in accordance with annex B, the patency of the oropharyngeal airway lumen shall be maintained.

## 8 Sterility assurance

Oropharyngeal airways supplied and marked "STERILE" shall satisfy the requirements of 4.1 of EN 556:1994.

## 9 Packaging of oropharyngeal airways supplied sterile

**9.1** Each oropharyngeal airway supplied and marked "STERILE" shall be contained in an individual pack.

**9.2** The pack shall serve as an effective barrier to the penetration of microorganisms and particulate material in accordance with ISO 11607.

**9.3** The pack shall permit the aseptic extraction of the contents and shall not be capable of re-closure without clearly revealing that it has been opened.

**9.4** The designated size of the airway shall be apparent on visual examination of the intact unit container.

**9.5** Individual packs shall be contained within a shelf or multi-unit pack.

## 10 Marking

### 10.1 General

Marking of oropharyngeal airways, of unit packs and of shelf or multi-unit packs and information to be supplied by the manufacturer should comply with EN 1041.

### 10.2 Use of symbols

The requirements of 10.4 and 10.5 may be met by use of appropriate symbols as given in ISO 7000 or EN 980.

### 10.3 Marking of oropharyngeal airways

The flanged end of the oropharyngeal airway shall be marked with the following:

- a) the designated size (nominal length, in centimetres) in accordance with 4.1 (see Figure 2);
- b) the name and/or trade mark of the manufacturer and/or supplier (see Figure 2).

Markings in accordance with a) and b) shall be visible from the flanged end.


### 10.4 Marking of unit packs

The marking of individual packs or a package insert shall include the following:

- a) the word "STERILE", if appropriate.

It is recommended that the method of sterilization be given;

- b) for oropharyngeal airways not intended for re-use, the words "single use" or equivalent;
- c) an indication of the presence of natural rubber (latex), if present in the device.

**Key**

- 1 Designated size  
2 Name or trademark of manufacturer or supplier

NOTE The designs shown in Figures 1 and 2 are intended to illustrate common types of oropharyngeal airway for the purpose of size designation and marking, but are for example only.

**Figure 2 — Typical marking locations on flanged end of oropharyngeal airways**

### 10.5 Marking of shelf or multi-unit packs

The marking of shelf or multi-unit packs shall include the following:

- a description of contents;
- the designated size in accordance with 4.1;
- the name and/or trademark and address of the manufacturer and/or supplier;
- the batch number;
- the word "STERILE", if appropriate.

It is recommended that the method of sterilization be given.

- for oropharyngeal airways not intended for re-use, the words "single use" or equivalent;

It is strongly recommended that the "use by" date be given.

## 11 Information to be supplied by the manufacturer

Unless the oropharyngeal airway is intended and marked as being for single use, the manufacturer shall recommend methods of cleaning and disinfection or sterilization.

For reusable devices, the manufacturer shall indicate the presence of natural rubber (latex), if present in the device.

## **Annex A**

(normative)

### **Test method for resistance to collapse of the buccal portion**

#### **A.1 Principle**

The resistance to collapse is tested by measuring the minimum inside dimension of the buccal portion during compression.

#### **A.2 Apparatus**

**A.2.1 Means of conditioning the oropharyngeal airway** at  $(34 \pm 2) ^\circ\text{C}$  and carrying out the test under the same conditions.

**A.2.2 Means of applying a force** of either  $(100 \pm 10)$  N or  $(200 \pm 20)$  N, as appropriate.

**A.2.3 Means of measuring the minimum inside dimension** of the buccal portion during compression, with an accuracy of  $\pm 0,10$  mm.

#### **A.3 Test procedure**

**A.3.1** Condition the oropharyngeal airway at  $(34 \pm 2) ^\circ\text{C}$  for 1 h and carry out the test under the same conditions.

**A.3.2** Compress the middle of the buccal portion (see Figure 1) of the oropharyngeal airway using blocks with a  $(60 \pm 2)^\circ$  included angle and radius of curvature of  $(1,0 \pm 0,5)$  mm to the mating surfaces and a width at least as great as that of the buccal portion of the airway being tested (see Figure A.1).


**A.3.3** For oropharyngeal airways of designated size 5,5 or smaller, apply a force of  $(100 \pm 10)$  N and maintain that force for 3 min whilst measuring the minimum dimension.

**A.3.4** For oropharyngeal airways of designated size 6,0 or larger, apply a force of  $(200 \pm 20)$  N and maintain that force for 3 min whilst measuring the minimum dimension.

#### **A.4 Expression of result**

Express the minimum inside dimension of the buccal portion during compression as a percentage of that value specified in Table 1 for the size of airway being tested.

Dimensions in millimetres

**Key**

1 Compressive force applied as in A.3.3 or A.3.4

<sup>a</sup> Width of blocks as great as buccal portion of airway being tested.**Figure A.1 — Apparatus for testing resistance to collapse of the buccal portion**

## **Annex B**

(normative)

### **Test method for patency of lumen**

#### **B.1 Principle**

The patency of the oropharyngeal airway lumen is tested by clamping the buccal portion, applying a force to the pharyngeal end, and passing a steel ball through the lumen of the oropharyngeal airway.

#### **B.2 Apparatus**

**B.2.1 Means of conditioning the oropharyngeal airway** at  $(34 \pm 2) ^\circ\text{C}$  and carrying out the test under the same conditions.

**B.2.2 Clamp**, for suspending the oropharyngeal airway.

**B.2.3 Means of applying a force** of  $(5,0 \pm 0,5) \text{ N}$ .

**B.2.4 Steel ball**, of diameter 75 % of the minimum inside dimension (see Table 1).

#### **B.3 Test procedure**

**B.3.1** Condition the oropharyngeal airway at  $(34 \pm 2) ^\circ\text{C}$  for 1 h and carry out the test under the same conditions.

**B.3.2** Clamp the buccal portion of the oropharyngeal airway so that it does not move, and apply a force of  $(5,0 \pm 0,5) \text{ N}$  to the pharyngeal end, maintaining this force for not less than 1 min (see Figure B.1).

**B.3.3** Pass a steel ball (B.2.4) through the lumen whilst the force is being applied.

#### **B.4 Expression of results**

Record whether or not the steel ball passes freely through the tube.