
**Implants for surgery — Metallic
skeletal pins and wires —**

**Part 1:
General requirements**

Implants chirurgicaux — Fils et broches pour os —

Partie 1: Matériaux et propriétés mécaniques



STANDARDSISO.COM : Click to view the full PDF of ISO 5838-1:2013



COPYRIGHT PROTECTED DOCUMENT

© ISO 2013

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

	Page
Foreword	iv
1 Scope	1
2 Normative references	1
3 Material	2
4 Mechanical properties	2
5 Method of manufacture	2
6 Finish	2
7 Sterilization	2
8 Packaging	3
9 Information supplied by the manufacturer	3
9.1 General	3
9.2 Instructions for use	3
9.3 Marking on implant	3

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 5838-1 was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, Subcommittee SC 5, *Osteosynthesis and spinal devices*.

This third edition cancels and replaces the second edition (ISO 5838-1:1995), which has been technically revised to add to the requirements for finish ([Clause 6](#)), to add a reference to ISO 14630 and to introduce new requirements for sterilization ([Clause 7](#)), packaging ([Clause 8](#)) and information supplied by the manufacturer ([Clause 9](#)).

ISO 5838 consists of the following parts, under the general title *Implants for surgery — Metallic skeletal pins and wires*:

- *Part 1: General requirements*
- *Part 2: Steinmann skeletal pins - Dimensions*
- *Part 3: Kirschner skeletal wires*

Implants for surgery — Metallic skeletal pins and wires —

Part 1: General requirements

1 Scope

This part of ISO 5838 specifies general requirements for metallic skeletal pins and wires for use in bone surgery, excluding wires for binding and twisting.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3651-2, *Determination of resistance to intergranular corrosion of stainless steels — Part 2: Ferritic, austenitic and ferritic-austenitic (duplex) stainless steels — Corrosion test in media containing sulfuric acid*

ISO 5832-1, *Implants for surgery — Metallic materials — Part 1: Wrought stainless steel*

ISO 5832-2, *Implants for surgery — Metallic materials — Part 2: Unalloyed titanium*

ISO 5832-3, *Implants for surgery — Metallic materials — Part 3: Wrought titanium 6-aluminium 4-vanadium alloy*

ISO 5832-5, *Implants for surgery — Metallic materials — Part 5: Wrought cobalt-chromium-tungsten-nickel alloy*

ISO 5832-6, *Implants for surgery — Metallic materials — Part 6: Wrought cobalt-nickel-chromium-molybdenum alloy*

ISO 5832-7, *Implants for surgery — Metallic materials — Part 7: Forgeable and cold-formed cobalt-chromium-nickel-molybdenum-iron alloy*

ISO 5832-8, *Implants for surgery — Metallic materials — Part 8: Wrought cobalt-nickel-chromium-molybdenum-tungsten-iron alloy*

ISO 5832-9, *Implants for surgery — Metallic materials — Part 9: Wrought high nitrogen stainless steel*

ISO 5832-11, *Implants for surgery — Metallic materials — Part 11: Wrought titanium 6-aluminium 7-niobium alloy*

ISO 5832-12, *Implants for surgery — Metallic materials — Part 12: Wrought cobalt-chromium-molybdenum alloy*

ISO 5832-14, *Implants for surgery — Metallic materials — Part 14: Wrought titanium 15-molybdenum 5-zirconium 3-aluminium alloy*

ISO 6892-1, *Metallic materials — Tensile testing — Part 1: Method of test at room temperature*

ISO 14630:2012, *Non-active surgical implants — General requirements*

3 Material

Skeletal pins and wires shall be made of wrought materials in accordance with ISO 5832-1, ISO 5832-2, ISO 5832-3, ISO 5832-5, ISO 5832-6, ISO 5832-7, ISO 5832-8, ISO 5832-9, ISO 5832-11, ISO 5832-12 and ISO 5832-14.

4 Mechanical properties

The mechanical properties shall be in accordance with the requirements given in [Table 1](#). The test methods in ISO 6892-1 shall be used in determining compliance with the tensile requirements.

Table 1 — Mechanical properties

Type of material	Diameter	Ultimate tensile strength	Elongation ^a
	d	min.	min.
	mm	MPa	%
Wrought stainless steel	$1 < d \leq 2,8$ $2,8 < d \leq 4$ $4 < d \leq 6$ ≤ 8	1 240 1 100 960 1 400 ^b	3 5 5 —
Wrought unalloyed titanium ^c	≤ 3 > 3	730 750	3 5
Wrought titanium alloys	≤ 6	1 030	3
Wrought cobalt-chromium base alloy	≤ 6	1 240	7
^a Gauge length = $5,65\sqrt{S_o}$ where S_o is the original cross-sectional area in mm ² or equal to 50 mm if possible. If this is not possible, for wires smaller in diameter than 2,5 mm, a gauge length of 100 mm or 200 mm, equal to the total length between the grips, may be used (ISO 6892-1). However, in this case, the minimum value of elongation shall be agreed between the interested parties. ^b ISO 5832-1 includes an extra-hard classification to accommodate the need for very high strength stainless steels for osteosynthesis devices. ^c In specific clinical applications, the cross-sectional dimensions of the pin or wire used should correspond to the strength of the material.			

5 Method of manufacture

Skeletal pins and wires shall be manufactured by a cold drawing or a cold working process. Welds are not permissible. If local heating is needed for re-shaping points made of stainless steel, this can only be permitted provided that the whole of the final product can be shown to satisfy the intergranular corrosion test according to ISO 3651-2.

6 Finish

The external surface finish shall be free from burs, scratches and other defects visible to the naked eye. The products shall be appropriately treated to provide surface passivation.

Special attention should be paid to screw threads, if applicable.

7 Sterilization

The requirements of ISO 14630:2012, Clause 9, shall apply.