



Auxiliary tables for vibration generators — Methods of describing equipment characteristics

TECHNICAL CORRIGENDUM 1

Tables auxiliaires pour générateurs de vibrations — Méthodes de description des caractéristiques

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to ISO 6070:1981/Cor.1:2006(E) was prepared by Technical Committee ISO/TC 108, *Mechanical vibration and shock*, Subcommittee SC 6, *Vibration and shock generating systems*.

Page 5:

Replace 6.2 with the following:

6.2 Test load m_1

A load permitting an acceleration of 10 m/s² (approximately 1 g_n) peak under sinusoidal conditions.

Page 5:

Replace 6.3 with the following:

6.3 Test load m_4

A load permitting an acceleration of 40 m/s² (approximately 4 g_n) peak under sinusoidal conditions.