
**Methods for the petrographic analysis of
coals —**

Part 5:
**Method of determining microscopically
the reflectance of vitrinite**

Méthodes d'analyse pétrographique des charbons —

*Partie 5: Détermination au microscope du pouvoir réflecteur de la
vitrinite*



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 7404-5:2009



COPYRIGHT PROTECTED DOCUMENT

© ISO 2009

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction.....	v
1 Scope	1
2 Normative references	1
3 Definitions	1
4 Principle	1
5 Reagents and materials	2
6 Apparatus	3
7 Preparation of coal sample	7
8 Procedure	7
9 Reporting of results	10
10 Precision	12
11 Test report	13
Bibliography	14

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 7404-5 was prepared by Technical Committee ISO/TC 27, *Solid mineral fuels*.

This third edition cancels and replaces the second edition (ISO 7404-5:1994), which has been technically revised.

ISO 7404 consists of the following parts, under the general title *Methods for the petrographic analysis of coals*:

- *Part 1: Vocabulary*¹⁾
- *Part 2: Methods of preparing coal samples*
- *Part 3: Method of determining maceral group composition*
- *Part 4: Method of determining microlithotype, carbominerite and minerite composition*¹⁾
- *Part 5: Method of determining microscopically the reflectance of vitrinite*

1) Parts 1 and 4 of this International Standard will be available under the original title, *Methods for the petrographic analysis of bituminous coal and anthracite*, until the revisions of these documents have reached the stage at which they are publicly available.

Introduction

Petrographic analyses have been recognized internationally as important in the context of the genesis, vertical and lateral variation, continuity, metamorphism and usage of coal. The International Committee for Coal and Organic Petrology (ICCP) has made recommendations concerning nomenclature and analytical methods and has published a comprehensive handbook that is continuously updated. The text of this part of ISO 7404 agrees substantially with the text of the handbook and incorporates many useful comments made by members of the ICCP and by member bodies of ISO/TC 27, *Solid mineral fuels*.

Petrographic analyses of single seam coals provide information about the rank, the maceral and microlithotype compositions and the distribution of minerals in the coal. The reflectance of vitrinite is a useful measure of coal rank and can provide information on the distribution of coals of different rank in a coal blend. Together with a maceral group analysis, it provides information about some important chemical and technological properties of the coal and the coal blend. The reflectance of vitrinite has various other applications, such as the characterization of bulk samples and cargoes. For coal blends, the measurement of the vitrinite reflectance profile can permit the identification of the component coals and permit the estimation of the relative abundance of the component coals within the blend.

ISO 7404 (all parts) is concerned with the methods of petrographic analysis currently employed in characterizing coal in the context of their technological use.

The method of determining the reflectance of vitrinite is applicable for low-, medium- and high-rank coals [7].

The properties of a given coal are determined by the proportions and associations of the macerals and minerals present (see ISO 7404-3 [3]) and by the rank of the coal and thus its type, grade and rank. The reflectance of the vitrinite in the coal can be used as an indicator of rank, independent of the petrographic composition. Vitrinite reflectance increases progressively with rank.

The reflectances of the macerals of the vitrinite group can vary significantly in a single coal seam and therefore the value of the reflectance obtained depends also on the choice of the macerals used for measurement. Reflectance measurements are made on one or more of the macerals of vitrinite and, in reporting the results, it is necessary to specify the macerals on which the measurement were made and the proportions of the overall value contributed by each of the macerals measured. Consequently, a vital step in the measurement of vitrinite reflectance is the identification of vitrinite and its various macerals or maceral varieties. For this purpose, reference can be made to ISO 7404-1 and the ICCP [1] handbook.

For rank determination of single-seam coals, normally the reflectance of collotelinite (eu-ulminite [6] in lignites, the equivalent of low-rank B and C [6]) is determined. In cases where collotelinite (or in low-rank coals, eu-ulminite) is not present in sufficient amounts, reflectance analysis on other vitrinite macerals is performed. Reflectance analysis on various vitrinite macerals can also be applied for technological purposes and to coal blends; see 8.3.1. The reflectance value obtained also depends on whether maximum or random reflectance measurements are made, so it is necessary to specify the type of measurement. All of these analysis procedures are applicable to single-coal seams or to blends providing that adequate (see 8.3.1) reflectance measurements are made in compliance with an unbiased sampling procedure on a representative sample.

An accreditation programme for vitrinite reflectance analysis of single-seam coals (SCAP) is run regularly by the ICCP for accrediting petrologists.

NOTE As this edition of ISO 7404 covers coals of all rank, the term “vitrinite” as used in this part of ISO 7404 includes vitrinite as well as huminite. Reference can be made to ISO 7404-1 for details. The equivalent to collotelinite in lignites is ulminite B. Reflectance measurement on lignites is performed on huminite.

STANDARDSISO.COM : Click to view the full PDF of ISO 7404-5:2009

Methods for the petrographic analysis of coals —

Part 5:

Method of determining microscopically the reflectance of vitrinite

1 Scope

This part of ISO 7404 specifies the methods for determining the reflectance of vitrinite of coals microscopically on the polished surfaces, immersed in oil. The methods are applicable to coals from single seams or coal blends covering the whole range of low-, medium- and high-rank coal.

Vitrinite reflectance measurements can be used to characterize the components within blends. Measures can be taken to correct for the vitrinite percentage within each of the components of the blend or to determine the proportion of components in a blend, particularly when the components have dissimilar vitrinite contents. This method necessitates the identification of vitrinite by the analyst.

Reflectance measurements on vitrinite obtained by interpreting the results from an automated system, are outside the scope of this part of ISO 7404.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 7404-1, *Methods for the petrographic analysis of bituminous coal and anthracite — Part 1: Vocabulary*

ISO 7404-2, *Methods for the petrographic analysis of coals — Part 2: Methods of preparing coal samples*

3 Definitions

For the purposes of this document, the definitions given in ISO 7404-1 apply.

4 Principle

The intensity of light reflected at near-normal incidence from a specified area of well polished vitrinite under oil immersion, measured at 546 nm using a photomultiplier (or similar device), is compared with the intensity of light reflected under identical conditions from a number of standards of known reflectance. Different vitrinite particles within a single-coal seam invariably differ slightly from one another in optical properties. Therefore, an adequate number (see 8.3.1) of readings on different particles is taken to ensure that the results are representative of the coal or coal blend.

5 Reagents and materials

5.1 Immersion oil, non-drying, non-corrosive type, with a refractive index of $1,518\ 0 \pm 0,000\ 4$ at $23\ ^\circ\text{C}$ and a wavelength of $546\ \text{nm}$.

Oil from a bottle opened more than one year ago should not be used.

In attempts to produce oils that are chemically and physically stable, toxic compounds such as polychlorinated bi-phenyls were used in some older products. Some more recent oils have been associated with allergies. The composition of the oil should, therefore, be checked to ensure that no toxic or other undesirable properties are associated with it.

5.2 Calibration standards

5.2.1 Reflectance standards, consisting of polished surfaces of materials that

- a) are isotropic (or basal sections of uniaxial minerals);
- b) are durable and resistant to corrosion;
- c) have a reflectance that is stable over a long period;
- d) are free from inclusions, grain boundaries, discontinuities, internal flaws and fractures;
- e) have negligibly low absorptance.

To avoid significant amounts of light other than that reflected from the top surface returning to the objective, the body of the standard shall be either thicker than $5\ \text{mm}$ or wedge-shaped. The lower surface shall be matt if it makes an angle of less than 10° with the upper polished surface. The sides shall be shielded from external light. The reflectance of the standards shall be of an order similar to that of the coal being measured. Use at least two such standards with well spaced reflectances. If a coal with a reflectance greater than $2,0\ \%$ is being measured, use one or more additional standards with reflectance greater than $2,0\ \%$.

Table 1 gives approximate mean values for reflectance standards or reflectance ranges as found between different standards, as calibrated against an ICCP Master Standard (see Note 2) in common use.

NOTE 1 For measuring a vitrinite reflectance of about $1,0\ \%$, a standard with reflectance below and a standard with reflectance above $1\ \%$ are used.

NOTE 2 An ICCP round robin exercise on reflectance standards demonstrated that variations between standards of nominally identical reflectances can be significant^[9]; since then, the ICCP has offered the calibration of standards against ICCP master standards.

NOTE 3 It is necessary that standards be carefully cleaned to avoid scratching the polished surface. If solvents are used to remove old oil, it is necessary to take care that the evaporation of the solvent does not leave a residue on the surface of the standard. Tarnishing can also occur with some standard materials, particularly glasses. When the surface becomes scratched, or when comparison with the other standards shows that the reflectance value has changed, polishing is necessary.

5.2.2 Calculation of reflectance standards

Some sources^[4] recommend calculating the reflectance, R , of a standard as given in Equation (1):

$$R = \left[\frac{(n - 1,518)^2 + n^2 \alpha^2}{(n + 1,518)^2 + n^2 \alpha^2} \right] \times 100 \quad (1)$$

where

n is the known refractive index of the standard material at a wavelength of 546 nm;

α is the known absorptance of the standard material at a wavelength of 546 nm.

NOTE The absorptance, α , is included only if it is significant.

However, the refractive index of the border phase is different from that of the interior of the standard. Consequently, solid reflectance reference materials should always be individually calibrated against a standard of known reflectance; see 5.2.1, Note 2.

Table 1 — Reflectance standards in common use

Designation	Reflectance %
Optical glasses	0,32 to 1,70
Spinel	≈ 0,42
Leucosaphire	≈ 0,59
Yttrium aluminium garnet (YAG)	0,895 to 0,916
Gadolinium gallium garnet (GGG)	1,60 to 1,80
Diamond	≈ 5,3
Silicon carbide	≈ 7,80
See also 5.2.1, Notes 1 to 3.	

5.2.3 Zero standard

A suitable non-reflecting standard consists of a coal or opaque resin block with a hole about 5 mm in diameter and 5 mm deep drilled in its upper surface and filled with immersion oil. Alternatively, optical glasses of refractive index lower than that of the immersion oil may be used.

6 Apparatus

6.1 Reflected light microscope, with photometer (or similar device), containing the following elements (key item numbers refer to Figure 1, which shows the optical parts of a typical reflectance-measuring microscope):

NOTE 1 The component parts might not always be in the same sequence as shown in Figure 1.

- light source (key item 1), with a stable output; a quartz halogen lamp with a rating of 100 W is recommended;
- polarizer (key item 5), either a sheet or prism polarizer (used if maximum reflectance is determined);
- light-controlling apertures, consisting of two variable diaphragms, one of which is focused on the back focal plane of the objective (illuminator aperture, key item 3) and the other on the surface of the specimen (field stop; key item 6); it shall be possible to centre both diaphragms on the optical axis of the microscope system;
- vertical illuminator (key item 8), Berek prism, simple coated glass plate or Smith illuminator (a combined mirror and glass plate);

NOTE 2 Typical light paths are shown in Figure 2.

- e) objective (key item 9), strain-free, designed for use with polarized light (for maximum reflectance analysis).

Magnifications higher than those achieved by the commonly used 32x to 50x objectives require numerical apertures that decrease the depth of focus to an extent that is undesirable and should, therefore, be avoided where possible;

- f) eyepieces (key item 12), one or two viewing eyepieces (oculars), one of which is fitted with crosshairs that can be scaled;

NOTE 3 An additional ocular (key item 13) can be necessary in the light path leading to the photomultiplier.

- g) microscope tube, with the following features:

- measuring aperture (key item 14), which restricts the light reaching the photomultiplier to that reflected from an area of the specimen (key item 10) less than $80 \mu\text{m}^2$ and that can be aligned with the crosshairs in the viewing eyepiece (key item 12),
- means of optically isolating the viewing eyepieces from the light path to the photomultiplier if the eyepieces permit the entry of extraneous light during measurement,
- adequate blackening of the inside of the tube to absorb stray light;

NOTE 4 Subject to the above precautions, part of the light beam can be diverted to the eyepieces or to a television camera for continuous observation during reflectance measurement.

- h) filter (key item 15), with a peak transmittance in the range of $546 \text{ nm} \pm 5 \text{ nm}$ and a half-peak transmittance band of less than 30 nm.

The filter should be inserted into the light path immediately before the photomultiplier;

- i) photomultiplier tube (key item 16), fitted in housing attached to the microscope, permitting the light passing through the measuring aperture and filter to fall onto the photomultiplier window.


The photomultiplier tube should be of a type recommended for low-light-level applications, and shall have adequate sensitivity at 546 nm with a low output of current when the front element of the photomultiplier is not illuminated (low dark-field current). It shall have a linear response over the range of the measurement and the output shall be stable over the duration of the analysis;

NOTE 5 Lately, semiconductor photodiodes and digital cameras of high sensitivity have been increasingly replacing photomultiplier tubes. The same requirements for precision, linearity and stability as for photomultipliers apply.

- j) microscope stage (key item 11), fitted with a mechanical stage capable of advancing the specimen in the X and Y directions. For maximum reflectance, the stage shall be capable of being rotated through 360° perpendicular to the optical axis.

6.2 Power supply unit, for the light source, stabilized d.c., for which the following characteristics have been found to be satisfactory:

- a) output variation of less than 0,02 % for a supply variation of 10 %;
- b) ripple content at full load of less than 0,07 % peak to peak.




Key

- | | |
|----------------------------------|-------------------------------------|
| 1 lamp | 9 objective |
| 2 collector lens | 10 specimen |
| 3 illuminator aperture diaphragm | 11 stage |
| 4 heat filter | 12 viewing eye-piece |
| 5 polarizer | 13 third eye-piece |
| 6 field stop | 14 measuring aperture |
| 7 field-stop focusing lens | 15 546 nm peak transmittance filter |
| 8 vertical illuminator | 16 photomultiplier tube |

NOTE This set up is commonly known as Köhler Illumination [2].

Figure 1 — Optical parts of a typical reflectance measuring microscope



Key

- | | |
|---|---|
| 1 lamp filament | 7 back focal plane of objective (position of images of the filament and the illuminator aperture) |
| 2 collector lens | 8 objective |
| 3 illuminator aperture (position of filament image) | 9 surface of specimen (position of image of field stop) |
| 4 field stop | 10 coated glass plate |
| 5 field-stop focusing lens | 11 thin glass reflector |
| 6 Berek prism | 12 mirror |

Figure 2 —Types of vertical illuminators

6.3 Power supply unit, for the photomultiplier tube, stabilized d.c. voltage, for which the following characteristics have been found to be satisfactory:

- output variation of less than 0,05 % for a 10 % variation in supply voltage;
- ripple content at full load of less than 0,07 % peak to peak;
- temperature coefficient of less than 0,05 % K⁻¹;
- change in load from zero to full rated load that causes less than 0,1 % variation in output voltage.

NOTE In situations when significant voltage fluctuations are suspected, a line conditioner can be placed before the stabilized power; see 6.2 and 6.3.

6.4 Display, comprised of one of the following devices:

- a) high-resistance galvanometer with a minimum sensitivity of 10^{-10} A/mm; a long-period unit with a 1 s or 0,5 s period should be used;
- b) chart recorder;
- c) digital voltmeter or digital indicator.

The device used shall be capable of resolution of 0,01 % reflectance. A facility for backing-off the small positive voltage due to glare and photomultiplier dark current shall be provided.

A low-noise, variable-gain amplifier may be used, if necessary, to amplify the signal from the photomultiplier before it is passed to the display.

6.5 Sample mounting apparatus, comprised of slides, modelling clay and levelling device.

7 Preparation of coal sample

Prepare coal in accordance with ISO 7404-2.

8 Procedure

8.1 Setting up the apparatus

8.1.1 Starting procedure

Ensure that the room temperature remains within the range 18 °C to 28 °C. Switch on the lamp, power supplies and other electrical parts of the apparatus. Set the power supply to the photomultiplier to within the voltage range recommended by the manufacturer of the particular photomultiplier tube. Allow about 30 min for the apparatus to stabilize prior to making any measurements.

8.1.2 Adjusting the microscope for random or maximum measurements

If random reflectance measurements are being made, remove the polarizer. If maximum reflectance measurements are being made, set the polarizer to zero if a glass plate or Smith illuminator is used, or to 45° if a Berek prism is used. If a sheet polarizer is used, check and replace it if it shows significant discolouration. If the microscope has an analyser, remove this from the light path.

8.1.3 Illumination

Apply immersion oil to the polished surface of a coal block or particulate coal block, mounted and levelled on a glass or metal slide, and place the specimen on the stage.

Check that the microscope is correctly adjusted for Köhler illumination. Ensure that the field stop is in focus. Adjust the illuminated field by means of the field stop (see key item 6 in Figure 1) so that its diameter is about one-third of the diameter of the full field. Adjust the illuminator intensity aperture (see key item 3 in Figure 1) to reduce glare but without reducing the light intensity excessively. Once adjusted, do not alter the size of the apertures.

8.1.4 Alignment

Centre and focus the image of the field stop, centre the objective (see key item 9 in Figure 1) with respect to the axis of rotation of the stage and adjust the centre of the measuring aperture (see key item 14 in Figure 1) to be coincident with either the crosshairs or a known datum point in the viewing system.

If it is not possible to view the image of the measuring aperture superimposed on the specimen, select a field of view that contains a small, bright inclusion, such as a small crystal of pyrite, and align this directly under the crosshairs. Adjust the centring of the measuring aperture (see key item 14 in Figure 1) until the photomultiplier reading is at its highest value.

8.2 Checking the reliability and calibration of the apparatus

8.2.1 Stability of the apparatus

Place the standard with the highest reflectance value under the microscope and focus under oil immersion.

Adjust the amplifier or the voltage to the photomultiplier until the reading on the display has the same numerical value as the reflectance of the standard.

EXAMPLE 173 mV can correspond to a reflectance of 1,73 %.

Ensure that the signal is stable by checking that the reading changes by desirably less than 1 % but under no circumstances more than 2 % relative to the first reading within a period of 15 min.

8.2.2 Variation in reading on rotating a reflectance standard on the stage

Place a standard with a reflectance in oil between 1,6 % and 2,0 % on the stage and focus under oil immersion. Slowly rotate the stage and verify that the maximum variation in the reading is less than 2 % relative to the reflectance of the standard being used. If this value is exceeded, check the levelling of the standard, and ensure that the stage is perpendicular to the optical axis and that it rotates in a fixed plane. If these checks do not reduce the variation to less than 2 %, it is necessary that the manufacturer check the mechanical stability of the stage and the geometry of the microscope.

8.2.3 Correction for parasitic reflections and photomultiplier dark current

Place the zero standard on the stage and note the reading that represents the sum of the photomultiplier dark-current and the parasitic reflections. If the photomultiplier dark-current and parasitic reflections exceed 0,04 % reflectance in total, determine their relative proportions by interrupting the light reaching the photomultiplier so that any residual signal is then due to the photomultiplier dark current. Check the setting of the illuminator aperture and change the photomultiplier tube and/or the objective, as appropriate, so that the total signal is below 0,04 % reflectance. When the total signal is below 0,04 % reflectance, adjust the display to zero using the backing-off control; see 6.4. The latter is not applicable for digital voltmeters. Continue making adjustments using the highest standard as in 8.2.1 and the zero standard in turn until no further adjustment is necessary.

8.2.4 Linearity of the signal from the photomultiplier

Measure the reflectance of the other standards whilst maintaining the constant settings of the voltage supplies and light-controlling apertures in order to check that the measuring system has a linear response in the range being measured and that the standards match their calculated values. Rotate each standard to ensure that the mean reading attained matches the calculated values. If the reading for any standard differs from its calculated reflectance value by more than 2 % relative to the calculated value, clean the standard and repeat the standardizing process. Re-polish any glass standard still displaying a reflectance differing from its calculated value by more than 2 % relative.

If the reflectances of the standards still do not give a linear plot, check the linearity of the photomultiplier signal using standards from other sources. If these fail to give a linear plot, check the linearity of the signal by means of several calibrated neutral-density filters to reduce the luminous flux by known amounts. If the signal from the photomultiplier is confirmed to be non-linear, reduce the photomultiplier voltage by 50 V and recheck. If rechecking still shows the signal to be non-linear, replace the photomultiplier tube and carry out further checks as necessary to achieve linearity of the signal.

8.3 Measurement of the reflectance of vitrinite

8.3.1 General

The procedure for measuring maximum reflectance is specified in 8.3.2 and that for random reflectance is specified in 8.3.3. In these subclauses the term vitrinite refers to the vitrinite group or one or more of the macerals in the vitrinite group. As explained in the introduction, the choice of the macerals on which the measurements are made affects the result. Consequently, it is important to decide on which macerals to measure the reflectance. Normally, reflectance measurements shall be made on collotelinite or eu-ulminite for single-seam coals. If other vitrinite macerals are used, this shall be specified in the analysis report. In blends, all vitrinites suitable for the analysis (homogeneous and of sufficient size) are measured.

Ensure that the measurements are evenly distributed over the surface of the block, but vitrinite grains near the edges of the polished block should be avoided.

The number of measurements required depends on the purpose (see introduction) of the analysis, the precision required and the range of vitrinite reflectance data in a sample. This number may vary between 50 and 1 000 (or more) depending on whether, for example, a rank determination of a single-seam coal is the purpose of the analysis or a precise determination of different rank vitrinites in complex coal blends.

8.3.2 Measurement of the maximum reflectance of vitrinite in oil

Ensure that the polarizer is fitted to the microscope as specified in 8.1.2 and that the appropriate checks on the apparatus have been made as specified in 8.1 and 8.2.

Immediately after calibrating the apparatus, place a levelled polished block with applied immersion oil on the mechanical stage to allow making measurements by starting close to one corner of the area being traversed and bring the specimen into focus.

Move the specimen using the mechanical stage until the crosshairs are focused on a suitable area of vitrinite. Ensure that the area being measured contains no cracks, polishing defects, mineral inclusions or relief effects, and is away from maceral boundaries.

Allow the light to pass to the photomultiplier and rotate the stage through 360°. Record the highest reflectance reading obtained during the stage rotation.

During rotation of the block through 360°, ideally two identical maximum readings should be obtained. If the two readings differ significantly, the reason should be sought and the fault corrected. Occasionally, air bubbles in the oil pass into the measuring area causing erratic readings. If air bubbles are seen or suspected, ignore the reading and remove the air bubbles.

After not more than 50 measurements, recheck the calibration of the apparatus using the standard nearest in reflectance to that of the highest reflecting vitrinite in the specimen; see 8.2.5. If the reflectance of the standard differs from its theoretical value by more than 2 % relative to the theoretical value, discard the last set of readings and repeat them after recalibrating the apparatus using the full range of standards.

Make reflectance measurements on the vitrinite until the required number (see 8.3.1) has been recorded.

8.3.3 Measurement of the random reflectance of vitrinite in oil

Adopt the same procedures as specified in 8.3.2, but make the measurements without the polarizer and without rotation of the sample block. Calibrate the apparatus as specified in 8.2.5.

Make reflectance measurements on the vitrinite until the required number (see 8.3.2) has been recorded.

9 Reporting of results

The results may be reported as individual values or as numbers of measurements in intervals of 0,05 % reflectance (1/2 V-step) or in intervals of 0,10 % reflectance (V-step). If individual readings are reported, calculate the mean maximum or mean random reflectance percentage, $\bar{R}$, from Equation (2) and the standard deviation, σ , from Equation (3):

$$\bar{R} = \frac{\sum R_i}{n} \quad (2)$$

$$\sigma = \sqrt{\frac{n \sum R_i^2 - (\sum R_i)^2}{n(n-1)}} \quad (3)$$

where

R is the mean maximum or random reflectance;

R_i is the i th maximum or random reflectance measurement;

σ is the standard deviation.

NOTE 1 Previous versions of this part of ISO 7404 included a method of calculating the mean from group data. This practice is undesirable from a mathematical point of view and, with the general availability of computers, it is no longer needed.

The identity of vitrinite measured should be recorded at the maceral or maceral sub group level. Results may be presented as a histogram. An example of a suitable method of expressing results is shown in Table 2 and the corresponding histogram is shown in Figure 3.

NOTE 2 A V-step has a range of 0,1 % reflectance and a half (1/2) V-step a range of 0,05 % reflectance. In order to avoid overlap of reflectance values expressed to two decimal places, the ranges of values belonging to selected V-steps and 1/2 V-steps are, for example, as follows:

- V-step: ranges of 0,60 to 0,69; 0,70 to 0,79, ...; with the mid-point of range 0,60 to 0,69 equal to 0,645;
- 1/2 V-step: ranges of 0,60 to 0,64; 0,65 to 0,69;, with the mid-point of range 0,60 to 0,64 equal to 0,62;

Table 2 — An example of presenting the results of a blend

Sample No.

Blend No. 1

Reflectance measured:

Random

Reflectance ^a	Number of observations	Percentage of observations	Reflectance ^a	Number of observations	Percentage of observations
0,40 to 0,44	—	—	1,20 to 1,24	47	9
0,45 to 0,49	—	—	1,25 to 1,29	39	8
0,50 to 0,54	—	—	1,30 to 1,34	18	4
0,55 to 0,59	—	—	1,35 to 1,39	20	4
0,60 to 0,64	—	—	1,40 to 1,44	23	5
0,65 to 0,69	—	—	1,45 to 1,49	29	6
0,70 to 0,74	—	—	1,50 to 1,54	66	13
0,75 to 0,79	—	—	1,55 to 1,59	65	13
0,80 to 0,84	—	—	1,60 to 1,64	13	3
0,85 to 0,89	2	—	1,65 to 1,69	8	2
0,90 to 0,94	12	2	1,70 to 1,74	—	—
0,95 to 0,99	12	2	1,75 to 1,79	—	—
1,00 to 1,04	15	3	1,80 to 1,84	—	—
1,05 to 1,09	14	3	1,85 to 1,89	—	—
1,10 to 1,14	39	8	1,90 to 1,94	—	—
1,15 to 1,19	78	16	1,95 to 1,99	—	—
Total number of measurements, n : 500					
Mean reflectance, R : 1,32 %					
Standard deviation of the distribution, σ : 0,200					
^a Upper and lower limits may be changed as appropriate.					