
International Standard



7888

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Water quality — Determination of electrical conductivity

Qualité de l'eau — Détermination de la conductivité électrique

First edition — 1985-05-15

STANDARDSISO.COM : Click to view the full PDF of ISO 7888:1985

UDC 543.3 : 541.133

Ref. No. ISO 7888-1985 (E)

Descriptors : water, quality, tests, determination, electrical properties, conductivity.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work.

Draft International Standards adopted by the technical committees are circulated to the member bodies for approval before their acceptance as International Standards by the ISO Council. They are approved in accordance with ISO procedures requiring at least 75 % approval by the member bodies voting.

International Standard ISO 7888 was prepared by Technical Committee ISO/TC 147, *Water quality*.

STANDARDSISO.COM : Click to view the full PDF of ISO 7888:1985

Water quality — Determination of electrical conductivity

1 Scope and field of application

This International Standard specifies a method for the measurement of the electrical conductivity of all types of water.

Electrical conductivity can be used to monitor the quality of

- a) surface waters;
- b) process waters in water supply and treatment plants;
- c) waste waters.

The completeness of analysis for ionic constituents^[1 to 3] can be checked using this method.

In some cases absolute values are important, in other cases only relative changes are of concern.

For interferences, see clause 9.

2 Definitions

2.1 specific conductance; electrical conductivity, γ :

The reciprocal of the resistance, measured under specified conditions, between the opposite faces of a unit cube of defined dimensions of an aqueous solution. For water quality examination, this is often expressed as "electrical conductivity" and may be used as a measure of the concentration of ionizable solutes present in the sample.

(Definition taken from ISO 6107/2.)

It is expressed in siemens per metre.¹⁾

NOTE — The symbols σ and κ are also used for electrical conductivity (see ISO 31/5).

2.2 cell constant, K : Quantity, in reciprocal metres, given by the equation

$$K = \frac{l}{A}$$

where

l is the length, in metres, of an electrical conductor;

A is the effective cross-sectional area, in square metres, of an electrical conductor.

The cell constant results from the geometry of the cell; it can be empirically determined.

2.3 temperature coefficient of electrical conductivity,²⁾

α : The temperature coefficient of conductivity $\alpha_{\theta, 25}$,^[4, 5] is given by the equation

$$\alpha_{\theta, 25} = \frac{1}{\gamma_{25}} \left(\frac{\gamma_{\theta} - \gamma_{25}}{\theta - 25} \right) \times 100$$

where 25 and θ °C are the temperatures at which the electrical conductivities γ_{25} and γ_{θ} respectively were measured.

2.4 temperature correction factors, f : Factors used to correct for the temperature dependence of electrical conductivity.

In order to make comparisons, it is essential that measurements are corrected to a chosen reference temperature, usually 25,0 °C, even if the temperature of the water sample differs only slightly from that temperature.

Conversions to the electrical conductivity at 25 °C, γ_{25} , can be made using the equation

$$\gamma_{25} = \frac{\gamma_{\theta}}{1 + (\alpha/100)(\theta - 25)}$$

where

α is the temperature coefficient of electrical conductivity;

γ_{θ} is the electrical conductivity at the measured temperature, θ ;

θ is the measuring temperature, in degrees Celsius, of the sample.

1) 1 S/m = 10⁴ μ S/cm = 10³ mS/m

2) The temperature coefficient of electrical conductivity can be expressed in reciprocal kelvin or % per °C.

3 Principle

Direct determination, using an appropriate instrument, of the electrical conductivity of aqueous solutions. The electrical conductivity is a measure of the current conducted by ions present in the water ("phenomenon of conductors of the second kind"), and depends on

- the concentration of the ions;
- the nature of the ions;
- the temperature of the solution;
- the viscosity of the solution.

Pure water as a result of its own dissociation has an electrical conductivity at 25 °C of 5,483 $\mu\text{S}/\text{m}$ ^[6] (0,005 483 mS/m).

4 Reagents

During the analysis, unless otherwise stated, use only reagents of recognized analytical grade.

4.1 Water for preparing solutions and dilutions. Double distilled or de-ionized water; the electrical conductivity shall be $\gamma_{25} \leq 0,1$ mS/m.

4.2 Potassium chloride standard solution A.^[7]
 $c(\text{KCl}) = 0,1$ mol/l.

Dry a few grams of potassium chloride at 105 °C for 2 h, and dissolve 7,456 g in water (4.1). Dilute to 1 000 ml.

The conductivity of this solution at 25 °C, γ_{25} , is 1 290 mS/m.

4.3 Potassium chloride standard solution B,
 $c(\text{KCl}) = 0,01$ mol/l.

Dilute 100 ml of solution A (4.2) with water (4.1) to 1 000 ml.

The conductivity of this solution at 25 °C, γ_{25} , is 141 mS/m.

4.4 Potassium chloride standard solution C,
 $c(\text{KCl}) = 0,001$ mol/l.

Dilute 100 ml of solution B (4.3) with water (4.1) to 1 000 ml. Immediately before preparing this solution the water shall be freed from carbon dioxide by purging with pure nitrogen or by boiling. During work with these solutions any contact with the atmosphere shall be minimized.

Prepare this solution shortly before use.

The conductivity of this solution at 25 °C, γ_{25} , is 14,7 mS/m.

NOTE — Table 1 gives alternative concentrations of potassium chloride that can be used as standards of conductivity.^[8, 9]

Table 1 — Electrical conductivity of potassium chloride solutions

Concentration of potassium chloride, $c(\text{KCl})$	Electrical conductivity at 25 °C, γ_{25}
mol/l	mS/m
0,000 5	7,4
0,001	14,7
0,005	72
0,01	141
0,02	277
0,05	670
0,1	1 290
0,2	2 480

4.5 Platinizing solution.

Dissolve 1,5 g of hydrogen hexachloroplatinate(IV) hexahydrate ($\text{H}_2\text{PtCl}_6 \cdot 6\text{H}_2\text{O}$) in 50 ml of water containing 0,012 5 g of lead(II) acetate [$\text{Pb}(\text{C}_2\text{H}_3\text{O}_2)_2$].

5 Apparatus

5.1 Instruments for measurement of electrical conductivity.

The instrument may be of either of the following types:

- instrument equipped with a flow- or dip-type conductivity cell fitted with two or more electrodes;
- instrument fitted with electrodes of the induction type.

Preferably instruments should be capable of discrete and continuous measurement both in the laboratory and in the field.

A flow-type conductivity cell from which air is excluded is essential for measurements of conductivities of less than 1 mS/m.

The recommended electrode cell constant can be chosen from table 2 for each measuring range.

Table 2 — Recommended cell constants for different ranges of electrical conductivity

Measuring range	Recommended cell constant
mS/m	m^{-1}
$\gamma < 2$	1
$0,1 < \gamma < 20$	10
$1 < \gamma < 200$	100
$10 < \gamma < 2 \times 10^3$	1 000
$100 < \gamma < 20 \times 10^3$	5 000

Some instruments are equipped with a cell constant control. If this is not the case, the reading must be multiplied by the cell constant.

5.2 Electrodes.

Whenever platinum electrodes are used for precision measurements the electrodes shall be platinized (see the note).

Unplatinized electrodes may be used only for field and routine laboratory testing.

NOTE — If platinization is necessary, the manufacturer's instructions should be followed, or proceed as follows.

Platinize the electrodes of the cell with platinizing solution (4.5). A suitable plating apparatus consists of a 6 V d.c. supply, a variable resistor, a milliammeter, and an electrode. The procedure for platinizing is not critical. Good platinized coatings are obtained using from 1,5 to 3 C/cm² of electrode area. For example for an electrode having a total area (both sides) of 10 cm², the plating time at a current of 20 mA would be from 12,5 to 25 min. The current density may be from 1 to 4 mA/cm² of electrode area. Plate the electrodes one at a time with the aid of an extra electrode. During the plating, agitate the solution gently. When not in use, fill the cells with water to prevent the drying out of the electrodes while in storage.

5.3 Thermometer, accurate to $\pm 0,1$ °C, within the temperature range of measurement, shall be used for precise determinations. For routine measurements, a thermometer accurate to $\pm 0,5$ °C is satisfactory.

5.4 Thermostatic bath, capable of being maintained at $25,0 \pm 0,1$ °C. For routine measurements, a tolerance of $\pm 0,5$ °C is satisfactory.

6 Sampling and samples

Collect the laboratory sample in a polyethylene bottle completely filled and tightly stoppered. Soda glass bottles shall not be used. Measurement of conductivity should be performed as soon as practicable, particularly when there is a possibility of an exchange of gases such as carbon dioxide or ammonia with the atmosphere, or a possibility of biological activity. Biological activity can be reduced by storing the samples in the dark at 4 °C; however, samples shall be brought to equilibrium at the reference temperature of 25 °C before the conductivity is measured. No suitable preservative is known for samples taken for conductivity measurements.

7 Procedure

7.1 General

Prepare the equipment for use as instructed by the manufacturer and ensure that an electrode cell of known cell constant

appropriate to the desired measuring range is fitted (see table 2). The test portion depends on the equipment used.

If the cell constant is not accurately known, determine the constant as given in clause 5 using the potassium chloride standard solutions (4.2 to 4.4) appropriate to each desired measuring range. Check the cell constant at least once every 6 months.

Many instruments incorporate cell constant correction as an integral function and thus a direct reading of electrical conductivity is obtained. Otherwise multiply the conductance value obtained by the cell constant to obtain electrical conductivity.

For high precision work, carry out the measurement of electrical conductivity when the sample and apparatus in direct contact with it has attained equilibrium at $25,0 \pm 0,1$ °C. Thus all sources of error that may arise from the use of temperature compensators, or from mathematical correction techniques, are eliminated.

7.2 Temperature correction

If measurement at $25,0 \pm 0,1$ °C is not possible, for example in field or plant work, measure the electrical conductivity of the sample at a known temperature, θ °C. Many instruments are fitted with temperature compensation devices, and with reference to the temperature coefficient of samples, may automatically correct measurements obtained over a range of temperatures to electrical conductivity at 25,0 °C. Such instruments shall be calibrated strictly in accordance with the manufacturer's instructions.

If the temperature coefficient of the sample is not known, it may be derived by substituting electrical conductivity values experimentally determined at $25,0 \pm 0,1$ °C and temperatures $\theta \pm 0,1$ °C (see 2.3).

Where instruments do not incorporate a temperature compensation device, the electrical conductivity measured at θ °C shall be corrected to 25,0 °C using the appropriate correction factor taken from table 3.

Whichever form of temperature compensation is applied to the measurement of electrical conductivity at θ °C, the result will be less accurate than that actually measured at the reference temperature of 25,0 °C.

In some aspects of routine field work, it may not be necessary to transpose values measured at θ °C to 25,0 °C. However, such measurements should be interpreted with great care and comparison with other values may be difficult or even meaningless.

Table 3 — Temperature correction factor, f_{25} , for the conversion of conductivity values of natural waters from θ °C to 25 °C^[5]

θ	f_{25}									
°C	,0	,1	,2	,3	,4	,5	,6	,7	,8	,9
0	1,918	1,912	1,906	1,899	1,893	1,887	1,881	1,875	1,869	1,863
1	1,857	1,851	1,845	1,840	1,834	1,829	1,822	1,817	1,811	1,805
2	1,800	1,794	1,788	1,783	1,777	1,772	1,766	1,761	1,756	1,750
3	1,745	1,740	1,734	1,729	1,724	1,719	1,713	1,708	1,703	1,698
4	1,693	1,688	1,683	1,678	1,673	1,668	1,663	1,658	1,653	1,648
5	1,643	1,638	1,634	1,629	1,624	1,619	1,615	1,610	1,605	1,601
6	1,596	1,591	1,587	1,582	1,578	1,573	1,569	1,564	1,560	1,555
7	1,551	1,547	1,542	1,538	1,534	1,529	1,525	1,521	1,516	1,512
8	1,508	1,504	1,500	1,496	1,491	1,487	1,483	1,479	1,475	1,471
9	1,467	1,463	1,459	1,455	1,451	1,447	1,443	1,439	1,436	1,432
10	1,428	1,424	1,420	1,416	1,413	1,409	1,405	1,401	1,398	1,394
11	1,390	1,387	1,383	1,379	1,376	1,372	1,369	1,365	1,362	1,358
12	1,354	1,351	1,347	1,344	1,341	1,337	1,334	1,330	1,327	1,323
13	1,320	1,317	1,313	1,310	1,307	1,303	1,300	1,297	1,294	1,290
14	1,287	1,284	1,281	1,278	1,274	1,271	1,268	1,265	1,262	1,259
15	1,256	1,253	1,249	1,246	1,243	1,240	1,237	1,234	1,231	1,228
16	1,225	1,222	1,219	1,216	1,214	1,211	1,208	1,205	1,202	1,199
17	1,196	1,193	1,191	1,188	1,185	1,182	1,179	1,177	1,174	1,171
18	1,168	1,166	1,163	1,160	1,157	1,155	1,152	1,149	1,147	1,144
19	1,141	1,139	1,136	1,134	1,131	1,128	1,126	1,123	1,121	1,118
20	1,116	1,113	1,111	1,108	1,105	1,103	1,101	1,098	1,096	1,093
21	1,091	1,088	1,086	1,083	1,081	1,079	1,076	1,074	1,071	1,069
22	1,067	1,064	1,062	1,060	1,057	1,055	1,053	1,051	1,048	1,046
23	1,044	1,041	1,039	1,037	1,035	1,032	1,030	1,028	1,026	1,024
24	1,021	1,019	1,017	1,015	1,013	1,011	1,008	1,006	1,004	1,002
25	1,000	0,998	0,996	0,994	0,992	0,990	0,987	0,985	0,983	0,981
26	0,979	0,977	0,975	0,973	0,971	0,969	0,967	0,965	0,963	0,961
27	0,959	0,957	0,955	0,953	0,952	0,950	0,948	0,946	0,944	0,942
28	0,940	0,938	0,936	0,934	0,933	0,931	0,929	0,927	0,925	0,923
29	0,921	0,920	0,918	0,916	0,914	0,912	0,911	0,909	0,907	0,905
30	0,903	0,902	0,900	0,898	0,896	0,895	0,893	0,891	0,889	0,888
31	0,886	0,884	0,883	0,881	0,879	0,877	0,876	0,874	0,872	0,871
32	0,869	0,867	0,866	0,864	0,863	0,861	0,859	0,858	0,856	0,854
33	0,853	0,851	0,850	0,848	0,846	0,845	0,843	0,842	0,840	0,839
34	0,837	0,835	0,834	0,832	0,831	0,829	0,828	0,826	0,825	0,823
35	0,822	0,820	0,819	0,817	0,816	0,814	0,813	0,811	0,810	0,808

NOTES TO TABLE 3

1 The values of the temperature correction factors given are the mean values of measurements of a number of natural waters. Note that they are only applicable for measurements of those waters having a γ_{25} of about 6 to 100 mS/m and a composition comparable to natural ground, well, or surface waters.

	Cations	Anions
Predominant	Ca ²⁺	HCO ₃ ⁻
Minor	Mg ²⁺	SO ₄ ²⁻ , Cl ⁻ , NO ₃ ⁻

2 Note especially that they are not applicable to potassium chloride solutions used for the calibration of conductivity cells. Therefore the calibration must always be carried out at the reference temperature 25,0 ± 0,1 °C.

3 The temperature correction factors, f_{25} , are calculated by the equations

$$f_{25} = \left[(1 - a) + a \left(\frac{\eta_{\theta}}{\eta_{25}} \right)^n \right] \times 1,116$$

and

$$\frac{\eta_{\theta}}{\eta_{25}} = A + \exp \left(B + \frac{C}{\theta + D} \right)$$

where

η is the viscosity of the solution;

θ is the temperature of the solution at which measurements were made.

Values of the constants

$$a = 0,962\ 144$$

$$n = 0,965\ 078$$

$$A = -0,198\ 058$$

$$B = -1,992\ 186$$

$$C = 231,176\ 28$$

$$D = 86,391\ 23$$

exp is the exponential function [$e = 2,718\ 28$ (base of natural logarithms)]

4 In the past electrical conductivity was often quoted as γ_{20} . If the water composition is comparable to that in note 1, it is possible to convert all such values to the present reference temperature of 25 °C using the temperature correction factor of 1,116 given in table 3.