
Neutron reference radiation fields —
Part 3:
Calibration of area and personal
dosemeters and determination
of their response as a function of
neutron energy and angle of incidence

Champs de rayonnement neutronique de référence —

*Partie 3: Étalonnage des dosimètres de zone et individuels et
détermination de leur réponse en fonction de l'énergie et de l'angle
d'incidence des neutrons*



STANDARDSISO.COM : Click to view the full PDF of ISO 8529-3:2023



COPYRIGHT PROTECTED DOCUMENT

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Procedures applicable to area and personal dosimeters	2
4.1 Neutron fields	2
4.2 Conversion coefficients	2
4.3 Determination of the response	3
4.4 Calibration procedures	3
5 Procedures for calibrating and determining the ambient dose equivalent response of area dosimeters	4
5.1 Quantity to be measured and conversion coefficients	4
5.2 Required response characteristics	4
5.3 Dosimeter conditions	4
5.4 Irradiation geometry	4
5.5 Evaluation of measurement	4
6 Procedures for calibrating and determining the personal dose equivalent response of personal dosimeters	5
6.1 Quantity to be measured and conversion coefficients	5
6.2 Required response characteristics	6
6.3 Calibration phantom	6
6.4 Dosimeter conditions	6
6.5 Irradiation geometry	6
6.6 Evaluation of measurement	7
7 Uncertainty	8
Annex A (informative) List of reference conditions and standard test conditions	9
Bibliography	10

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 85, *Nuclear energy, nuclear technologies, and radiological protection*, Subcommittee SC 2, *Radiation protection*.

This second edition cancels and replaces the first edition (ISO 8529-3:1998), which has been technically revised.

The main changes are as follows:

- The second and last edition of ISO 8529-1:2021 revised the neutron reference radiation fields produced with radionuclide sources as well as those produced with monoenergetic neutrons, thus requiring calculation of new conversion coefficients from neutron fluence to ambient dose equivalent or personal dose equivalent.

A list of all parts in the ISO 8529 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

This document is closely related to ISO 8529-1 and ISO 8529-2 concerning the calibration of dosimeters and dose rate meters for neutron radiation. ISO 8529-1 specifies the reference neutron radiation fields, in the energy range from thermal up to 20 MeV, and their production methods. ISO 8529-2 describes the calibration fundamentals of radiation protection devices related to basic quantities characterising the radiation field and specifies the procedures to be used for realising the calibration conditions of radiation protection devices produced by calibration sources with emphasis on correction for extraneous effects.

This document deals with dosimeters for area and individual monitoring. Unless differently specified, the word "dosimeter" always refers to both. Area dosimeters are often called area monitors or area survey meters, and dosimeters for individual monitoring are often called personal dosimeters or personal dosimeters. This document describes procedures for calibrating and determining the response in terms of the International Commission on Radiation Units and Measurements (ICRU) operational quantities. These are defined in ICRU Report 39^[1] and ICRU Report 51^[2]. For radiation protection purposes, these operational quantities are considered to be a sufficiently accurate approximation to the protection quantities. For the purposes of this document, the emphasis will be on the evaluation of the operational quantities at 10 mm depth defined in the body using conversion coefficients in the appropriate phantom. Cold neutrons may present special problems in dosimetry and are outside the scope of this document, as are the photon calibrations of instruments designed to measure both photons and neutrons.

The determination of the response of dosimeters is essentially a three-step process. First, a primary quantity such as the neutron fluence is determined at the point of test. Second, the reference point of the device being calibrated is then placed at the point of test to determine the fluence response. Third, the response of the device with respect to the appropriate operational quantity is then determined by the application of conversion coefficients that relate the physical quantity (the fluence) to the operational quantity (the dose equivalent). This document describes the methods and the conversion coefficients to be used for the determination of the response of area and personal dosimeters in terms of the respective ICRU operational quantities for neutrons.

STANDARDSISO.COM : Click to view the full PDF of ISO 8529-3:2023

Neutron reference radiation fields —

Part 3:

Calibration of area and personal dosimeters and determination of their response as a function of neutron energy and angle of incidence

1 Scope

This document provides guidance for those who calibrate protection-level dosimeters and doserate meters for area and individual monitoring with reference neutron radiation fields. This includes the determination of the response as a function of neutron energy and angle of incidence. The operational quantities recommended in ICRU Report 51^[2] are considered. In addition to the description of procedures, this document includes appropriate definitions and conversion coefficients and provides guidance on the statement of measurement uncertainties.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 8529-1, *Neutron reference radiations fields — Part 1: Characteristics and methods of production*

ISO 8529-2, *Reference neutron radiations — Part 2: Calibration fundamentals of radiation protection devices related to the basic quantities characterizing the radiation field*

ISO 12789-1, *Reference radiation fields — Simulated workplace neutron fields — Part 1: Characteristics and methods of production*

ISO 12789-2, *Reference radiation fields — Simulated workplace neutron fields — Part 2: Calibration fundamentals related to the basic quantities*

ISO 29661, *Reference radiation fields for radiation protection — Definitions and fundamental concepts*

ISO/IEC Guide 98-3, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

3 Terms and definitions

For the purposes of this document, the terms and definitions of ISO 29661 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

free-field neutron fluence

Φ

fluence that would be present at a point in absence of any dosimeter, calibration phantom if present, and scattering from the room and the air

3.2 uniform irradiation

irradiation of an area (extending over an area dosimeter or phantom with personal dosimeter) in which the uncertainty component introduced by differences in fluence or its energy distribution does not significantly impact the overall uncertainty of the calibration procedure

Note 1 to entry: The calibration laboratory determines the impact that can be regarded as significant.

4 Procedures applicable to area and personal dosimeters

4.1 Neutron fields

This document deals with neutron reference radiation fields that shall be chosen from and produced in accordance with ISO 8529-1 and characterized using the techniques of ISO 8529-2. In general, when selecting an appropriate neutron field, it will be useful to take into account the specified energy and dose or dose rate ranges of the dosimeter to be tested. The basic quantities characterizing the radiation fields (energy and direction distribution of the neutron fluence) shall be determined and all necessary corrections shall be considered in accordance with ISO 8529-2 to allow the use of the recommended conversion coefficients.

4.2 Conversion coefficients

The conversion coefficients given in this document refer to the monoenergetic energies or the continuous energy distributions given in ISO 8529-1 and were derived from those recommended by ICRU^[3].

All of the conversion coefficients given in [Tables 1](#) to [4](#) have been determined assuming broad parallel neutron beams or fields composed of such beams. If these coefficients are used for calibration and test purposes, the neutron fields used should be uniform over a broad area, i.e. extending over the whole item to be calibrated (area dosimeter or phantom with personal dosimeter) and also should be almost parallel or composed of almost-parallel beams. For calibrations of large devices in divergent beams as described in detail in ISO 8529-2, geometric corrections shall be used to correct for non-uniform irradiation of the item at close distances from point sources.

The fluence to which the conversion coefficients refer should be measured at the point of test.

If the fluence is uniform on the whole front face of the dosimeter or phantom the fluence-to-dose equivalent conversion coefficient can be applied without any further considerations.

If not, corrections should be applied for all influencing factors, especially for calibrations of multiple personal dosimeters:

- the calibration of multiple personal dosimeters in broad spectrum reference neutron fields, where the fluence and direction vary across the face of the phantom^[4];
- the calibration of multiple personal dosimeters in monoenergetic reference neutron fields, where the fluence, energy, and direction distributions vary across the face of the phantom^[5].

Numerical values of fluence-to-dose-equivalent conversion coefficients for various irradiation conditions are given in [Clauses 5](#) and [6](#).

If the calibration field differs in energy distribution from the reference fields given in ISO 8529-1, the average fluence to dose equivalent conversion coefficient, h_ϕ , should be derived from the energy distribution of the neutron fluence, Φ_E , and the energy-dependent fluence to dose equivalent conversion coefficients $h_\phi(E)$ ^[3] using [Formula \(1\)](#):

$$h_\phi = \frac{1}{\Phi} \int h_\phi(E) \Phi_E dE \quad (1)$$

where

Φ is the neutron fluence;

$h_{\Phi}(E)$ is the energy-dependent fluence to dose equivalent conversion coefficients.

4.3 Determination of the response

The response or calibration factor of a dosimeter is a unique property of the dosimeter, and will, in general, depend on the energy and direction distributions of the radiation field, but should not be a function of other characteristics of the calibration facility or of the experimental techniques employed. Hence, the procedures for calibration or determining the response should ensure that the results are independent of the technique, and other factors such as the source-to-device distance and room size. For determining their response or calibration factor, dosimeters are placed with their reference point at the point of test of a reference radiation field with known energy distribution and free-field fluence value. Personal dosimeters should be positioned on the appropriate phantom.

In accordance with the above, the dosimeter indication shall be corrected for all extraneous effects, including neutrons scattered by the air and by the calibration room (see ISO 8529-2) and neutrons having other than the desired energies, such as the target-scattered neutrons in accelerator-based reference monoenergetic neutron fields. These are type-S influence quantities for which corrections need to be applied when deriving the "corrected indication" defined in ISO 29661.

Once the free-field fluence and the corrected indication are obtained, the response or calibration factor can be calculated. If the calibration is performed in terms of the ICRU-recommended operational quantities, the operational quantities are obtained by multiplying the free-field fluence and the corresponding fluence-to-dose-equivalent conversion coefficients.

Since the dose equivalent response of neutron dosimeters is generally energy dependent, it is often useful to obtain a calibration factor that is not in one of the reference fields specified in ISO 8529-1, but is in a field that closely simulates the fields that exist in the workplace. These are called simulated workplace neutron fields. To derive such calibration factors, ISO 12789-1 and ISO 12789-2 shall be used. This is especially the case for the albedo neutron personal dosimeter, for which a calibration factor can only be valid for a field identical to the calibration field and can depend on parameters such as the source-detector distance or the room size. An investigation of this kind serves to test the suitability of the dosimeter in these workplace neutron fields rather than to determine a unique calibration factor or response.

4.4 Calibration procedures

Different calibration procedures are applied in radionuclide or reactor- and accelerator-based reference neutron fields.

For radionuclide reference neutron fields the quantities characterizing the field (fluence and its energy and direction distribution) should be known from previous investigations of the radiation source characteristics and be predictable with time. In other cases, additional measurements are needed such as monitoring variations in fluence or dose equivalent rate during the calibration procedure.

As explained in ISO 29661, the calibration procedure can rely on the usage of a standard instrument, either sequentially or simultaneously exposed with respect to the dosimeter to be calibrated.

Calibrations at reactor- or accelerator-based reference neutron fields require the use of a monitor to transfer the quantity value and to monitor variations of the fluence or dose equivalent rate during the calibration procedure. The monitor and the dosimeter are simultaneously exposed. The monitor and dosimeter should be placed such that each does not influence the response of the other.

5 Procedures for calibrating and determining the ambient dose equivalent response of area dosimeters

5.1 Quantity to be measured and conversion coefficients

The quantity to be measured in area monitoring is the ambient dose equivalent, $H^*(10)$. [Tables 1](#) and [2](#) contain conversion coefficients, h^*_ϕ , converting neutron fluence to ambient dose equivalent for the reference fields recommended in ISO 8529-1.

5.2 Required response characteristics

Ideally, the response of area dosimeters should be isotropic. Under this assumption the fluence response, R_ϕ , only depends on the energy distribution of the incident field, Φ_E , and can be derived using [Formula \(2\)](#):

$$R_\phi = \frac{1}{\Phi} \int R_\phi(E) \Phi_E dE \quad (2)$$

where

Φ is the incident neutron fluence;

$R_\phi(E)$ is the response to monoenergetic neutrons of energy E , usually derived via a combination of measurements and simulation.

To properly measure the ambient dose equivalent, the energy dependence of $R_\phi(E)$ should be similar to that of the fluence-to-ambient dose equivalent conversion coefficient^[3].

5.3 Dosimeter conditions

The measurement should be performed under a controlled set of conditions required by the manufacturer in the accompanying documents or by a product standard. [Annex A](#) lists standard test conditions and reference conditions for an electronic direct-reading area dosimeter.

5.4 Irradiation geometry

Calibrations or determinations of the response are ideally performed in broad, parallel beams providing a uniform irradiation of the total area of the dosimeter or, more precisely, of the parts to be irradiated, i.e. ideally, the dose rate is constant across the beam diameter.

With the use of point sources (accelerator targets, radionuclide neutron sources), this can, in general, only be achieved by having a sufficient distance between source and dosimeter, the minimum distance being dependent on the size of the dosimeter. For spherical devices, a geometry correction has been developed allowing smaller distances between source and dosimeter (see ISO 8529-2). If a narrow, collimated beam of neutrons is used, such as is usually found at reactors, a broad-beam irradiation shall be simulated by moving the dosimeter appropriately across the beam^[6].

5.5 Evaluation of measurement

The response (or calibration factor) of the area dosimeter under the conditions specified above is obtained by determining

- the dosimeter corrected indication. The influence of the scattering from the room and the air on the dosimeter indication shall be considered as indicated in ISO 8529-2;
- the free-field neutron fluence, Φ ;

and applying the appropriate fluence-to-ambient-dose-equivalent conversion coefficient referred to in 4.2 and given in Tables 1 and 2 for the ISO-recommended neutron reference fields.

Table 1 — Conversion coefficient $h^*_{\phi}(10; E)$ from neutron fluence ϕ to ambient dose equivalent $H^*(10)$ for monoenergetic neutron reference fields

Neutron energy MeV	$h^*_{\phi}(10; E)$ pSv cm ⁻²
$2,53 \times 10^{-8}$ ^a	10,6 ^a
0,002	7,7
0,008	9,4
0,024	19,3
0,144	127
0,250	203
0,565	343
1,2	425
2,5	416
2,8	413
5,0	405
14,8	536
17,0	563
19,0	584

^a The reported value refers to a monoenergetic field at $2,53 \times 10^{-8}$ MeV. Any practical thermal neutron facility typically exhibits a continuous Maxwellian spectrum superposed to an epithermal and a fast neutron continuum. Therefore specific calculations are needed to derive the field-specific fluence to ambient dose equivalent conversion coefficient [2][8].

Table 2 — Conversion coefficient $h^*_{\phi}(10; E)$ from neutron fluence ϕ to ambient dose equivalent $H^*(10)$ for ISO-recommended broad spectrum neutron reference fields

Neutron field	$h^*_{\phi}(10; E)$ pSv cm ⁻²
²⁵² Cf(D ₂ O-moderated) ^a	114
²⁵² Cf	385
²⁴¹ Am-Be small source	393
²⁴¹ Am-Be large source	387
²⁴¹ Am-Be ^b	391

^a At 100 cm distance from the source centre. As explained in ISO 8529-1, different distances or assembly constructions may lead to different coefficients.

^b Former versions of this document.

6 Procedures for calibrating and determining the personal dose equivalent response of personal dosimeters

6.1 Quantity to be measured and conversion coefficients

The quantity to be measured for individual monitoring is the personal dose equivalent, $H_p(10)$. Tables 3 and 4 contain conversion coefficients, $h_{p\phi}$, for converting neutron fluence to $H_p(10)$ for the neutron reference radiation fields recommended in ISO 8529-1.

The quantity $H_p(10)$, being defined in a body, only exists when the person or phantom is present. This means that the body becomes part of the irradiation situation and, in some cases (particularly

for neutrons), the radiation field can be modified at the position of the person or phantom by multiple scattering. This can be of special concern when considering calibration in simulated workplace fields^[9] ^[10]. For the calibration of a dosimeter in such radiation fields of broad direction distribution and/or scattered components, $H_p(10)$ will need to be calculated using the energy and direction distributions of the field incident at depth d in the phantom. The conversion coefficients tabulated in this document cannot be applied for such irradiation situations and a complete simulation of the irradiation set-up shall be done in accordance with ISO 12789-2, and as indicated in Reference ^[10].

6.2 Required response characteristics

Ideally, personal dosimeters should have a fluence response with an energy and angle dependence similar to that of the fluence-to-personal dose equivalent conversion coefficient, if it is fixed on the appropriate phantom (see 6.3). It is then assumed that it measures the personal dose equivalent when fixed on the body. The range of angles within which the tests are performed is given by national regulations or other standards.

6.3 Calibration phantom

For whole-body neutron personal dosimeters, the water-filled slab phantom specified in ISO 29661 should be used.

Routine calibrations can be performed, under simplified conditions, either to check the calibration carried out by the manufacturer or to check whether the calibration factor is sufficiently stable during continued long-term use of a dosimeter. These need not always be performed on the recommended phantom but may sometimes be done more simply (e.g. free-in-air) or even with a radiation field different to that which the dosimeter is intended to measure. Such simplifications, if they are to be applied, shall be justified prior to their adoption by demonstrating that they lead to results identical to those from procedures described in this document, or that any differences can be reliably corrected for. This may be done on the basis of type test. Calibration on a phantom is necessary if the dosimeter is sensitive to the radiation backscattered by the phantom.

6.4 Dosimeter conditions

The measurement should be performed under a controlled set of conditions required by the manufacturer in the accompanying documents or by a product standard. Annex A lists standard test conditions and reference conditions for an electronic direct-reading personal dosimeter.

6.5 Irradiation geometry

The point of test shall be chosen at a distance from the source such that the field size at this position is sufficiently large to allow the irradiation of the entire front surface of the phantom.

Ideally, the irradiation shall be done in broad, parallel fields, providing a uniform irradiation of the front surface of the phantom. With the use of point sources (accelerator targets, radionuclide neutron sources), this can generally be achieved approximately by having a sufficient distance between the source and the point of test.

If a narrow, collimated beam of neutrons is used, such as is usual at reactors, a broad-beam irradiation shall be simulated by moving the phantom appropriately across the beam^[6].

The variation of the angle of incidence is realised by rotating the phantom around a vertical axis which passes through the point of test.

For a simultaneous determination of the response of several dosimeters as a function of the angle, ideally the dosimeters should be positioned on their axis of rotation.

If several personal dosimeters are irradiated simultaneously on the front face of the phantom, their spatial distribution shall be chosen after considering the following factors of influence^[4]:

- fluence variations across the face of the phantom;
- effects of changes in backscatter from the phantom caused by partial shielding of the phantom by the array of dosimeters;
- mutual perturbation between dosimeters;
- further corrections are needed when calibrating multiple personal dosimeters in monoenergetic reference neutron fields, where the energy and direction distributions vary across the face of the phantom^[5] in addition to the fluence.

6.6 Evaluation of measurement

The response (or calibration factor) of the personal dosimeter under the conditions specified above is obtained by determining:

- the dosimeter corrected indication. The influence of the scattering from the room and the air on the dosimeter indication should be considered as indicated in ISO 8529-2;
- the free-field neutron fluence, Φ ;

and applying the appropriate fluence-to-personal-dose-equivalent conversion coefficients referred to in 4.2 and given in Tables 3 and 4 for the ISO-recommended neutron reference fields.

Table 3 — Conversion coefficient $h_{p\phi}(10; E, \alpha)$ from neutron fluence Φ to $H_p(10)$ for monoenergetic and parallel neutron radiation (expanded field)

Neutron energy MeV	$h_{p\phi}(10; E, \alpha)$ pSv cm ⁻² for angles of incidence α , of					
	0°	15°	30°	45°	60°	75°
$2,53 \times 10^{-8}$ ^a	11,4	10,6	9,11	6,61	4,04	1,73
0,002	8,72	8,22	7,27	5,43	3,46	1,67
0,008	10,4	9,89	8,36	6,43	3,99	1,73
0,024	20,2	19,9	17,2	13,6	7,85	2,38
0,144	134	131	121	102	69,9	22,9
0,250	215	214	201	173	125	47,0
0,565	355	349	347	313	245	115
1,2	433	427	440	412	355	210
2,5	437	434	454	441	410	294
2,8	433	431	451	441	412	302
5,0	420	418	437	435	409	331
14,8	561	563	581	572	576	517
17,0	587	587	609	600	606	549
19,0	600	596	621	614	620	568

^a The reported value refers to a monoenergetic field at $2,53 \times 10^{-8}$ MeV. Any practical thermal neutron facility typically exhibits a continuous Maxwellian spectrum superposed to an epithermal and a fast neutron continuum. Therefore specific calculations are needed to derive the field-specific fluence to personal dose equivalent conversion coefficient^{[7][8]}.

Table 4 — Conversion coefficient $h_{p\phi}(10; \alpha)$ from neutron fluence Φ to $H_p(10)$ for parallel neutron radiation (expanded field) and the ISO-recommended broad spectrum neutron reference fields

Neutron field	$h_{p\phi}(10; \alpha)$ $\mu\text{Sv cm}^{-2}$ for angles of incidence α , of					
	0°	15°	30°	45°	60°	75°
$^{252}\text{Cf}(\text{D}_2\text{O-moderated})^a$	119	118	119	111	95,0	60,6
^{252}Cf	400	397	409	389	346	230
$^{241}\text{Am-Be}$ small source	412	410	426	416	385	294
$^{241}\text{Am-Be}$ large source	406	404	419	409	377	287
$^{241}\text{Am-Be}^b$	411	409	424	415	383 ^c	293
^a At 100 cm distance from the source centre. As explained in ISO 8529-1, different distances or assembly constructions may lead to different coefficients. ^b Former versions of this document. ^c Value recalculated after detecting an error in the former versions of this document.						

7 Uncertainty

ISO/IEC Guide 98-3 shall be used for the determination of uncertainties and the statement of uncertainty shall be consistent with the approaches recommended in ISO 29661.

At least the following component uncertainties shall be taken into account:

- uncertainty of the conventional quantity value; as taken from the calibration certificate of the standard instrument or provided for the reference field;
- uncertainty in the exact positioning of standard and test instrument (see 4.4 and ISO 8529-2); to be assessed by the test laboratory;
- uncertainty of the conversion coefficient; by convention taken to be zero for monoenergetic neutrons, for broad spectra see also ISO 8529-2;
- uncertainty due to the non-uniformity of the field over the cross-sectional area of the beam in the plane of measurement owing to beam divergence;
- uncertainty due to simultaneous irradiation of several dosimeters; an estimate of the effect of absorption of the primary radiation by the dosimeters shall be made and added to the component uncertainties, where applicable. This component should not significantly impact the overall uncertainty;
- uncertainties due to simplified procedures; where applicable, to be assessed by the test laboratory (see 6.3). This component should not significantly impact the overall uncertainty.

For points e) and f), the calibration laboratory determines the impact that can be regarded as significant.