



INTERNATIONAL STANDARD ISO 8536-2:2001
TECHNICAL CORRIGENDUM 1

Published 2003-02-01

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Infusion equipment for medical use —

Part 2: Closures for infusion bottles

TECHNICAL CORRIGENDUM 1

Matériel de perfusion à usage médical —

Partie 2: Bouchons pour flacons de perfusion

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to ISO 8536-2:2001 was prepared by Technical Committee ISO/TC 76, *Transfusion, infusion and injection equipment for medical and pharmaceutical use*.

Page 5

A.3.3

Replace the last sentence with the following text:

“Cool to room temperature and remove from autoclave.”