

INTERNATIONAL STANDARD

**ISO
9713**

Second edition
2002-09-01

Neurosurgical implants — Self-closing intracranial aneurysm clips

*Implants neurochirurgicaux — Clips intracrâniens pour anévrisme à
autofermeture*



Reference number
ISO 9713:2002(E)

© ISO 2002

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 9713:2002

© ISO 2002

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.ch
Web www.iso.ch

Printed in Switzerland

Contents

	Page
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Description of aneurysm clips	3
5 Indication of dimensions	3
6 Materials	3
7 Determination of magnetic properties	6
8 Closing force	6
9 Marking of clips	6
10 Sterilization	7
11 Packaging	7
12 Labelling and accompanying documentation	7

STANDARDSISO.COM : Click to view the full PDF of ISO 9713:2002

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

International Standard ISO 9713 was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, Subcommittee SC 3, *Neurosurgical implants*.

This second edition cancels and replaces the first edition (ISO 9713:1990), which has been technically revised.

Introduction

Magnetic fields of considerable strength [e.g. 0,2 T to 2,0 T (tesla) or more] are used in medicine with increasing frequency as part of diagnostic techniques such as magnetic resonance imaging (MRI). Exposure to electromagnetic radiation may pose a hazard to patients who have intracranial aneurysm clips. Clips with magnetic properties (dia-, para-, antiferro-, ferro- and/or ferrimagnetic) become magnetized when subjected to a magnetic field and under this condition are liable to directing forces. These forces may result in the clip being removed from the aneurysm that it was intended to occlude and even being moved through the tissues. Because of the very high field strengths, even materials normally regarded as non-magnetic may exhibit some response to the magnetic field, such as minimal deflection or rotation. It is therefore essential that aneurysm clips have weakly or non-magnetic properties.

Compounds of certain non-magnetic elements may, when processed, have strong magnetic properties. The opposite also occurs. The work done at manufacture may have an additional effect. However, material normally regarded as non-magnetic may exhibit some response when subjected to MRI levels of field strength.

A secondary effect is that the presence of a metallic clip may interfere with the MRI process, resulting in deterioration of the quality of the scanning image.

One of the main intentions of this International Standard is to help to ensure that appropriate and comparable information is supplied for each clip in order to facilitate the choice of the correct clip by the surgeon. The closing force of the clip is an important factor in the selection process, and this International Standard requires that the manufacturers determine the closing force in a uniform manner and state this value on the labelling. The actuation of some types of clip can unduly result in a reduction of the closing force.

[STANDARDSISO.COM](https://standardsiso.com) : Click to view the full PDF of ISO 9713:2002

Neurosurgical implants — Self-closing intracranial aneurysm clips

1 Scope

This International Standard describes characteristics of self-closing aneurysm clips intended for permanent intracranial implantation and specifies requirements for their marking, packaging, sterilization and for labelling and accompanying documentation. In addition it gives a method for the measurement of closing force.

This International Standard is not applicable to malleable clips, or clips intended to be used during the course of surgery and removed before wound closure (temporary clips).

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this International Standard. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 5832-2, *Implants for surgery — Metallic materials — Part 2: Unalloyed titanium*

ISO 5832-3, *Implants for surgery — Metallic materials — Part 3: Wrought titanium 6-aluminium 4-vanadium alloy*

ISO 5832-5, *Implants for surgery — Metallic materials — Part 5: Wrought cobalt-chromium-tungsten-nickel alloy*

ISO 5832-6, *Implants for surgery — Metallic materials — Part 6: Wrought cobalt-nickel-chromium-molybdenum alloy*

ISO 5832-7, *Implants for surgery — Metallic materials — Part 7: Forgeable and cold-formed cobalt-chromium-nickel-molybdenum-iron alloy*

ISO 5832-8, *Implants for surgery — Metallic materials — Part 8: Wrought cobalt-nickel-chromium-molybdenum-tungsten-iron alloy*

ISO 14630:1997, *Non-active surgical implants — General requirements*

ISO 15223, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied*

ISO 16061, *Instrumentation for use in association with non-active surgical implants — General requirements*

3 Terms and definitions

For the purposes of this International Standard, the following terms and definitions apply.

3.1

accuracy

ability of a measuring instrument to give responses close to a true value

NOTE "Accuracy" is a qualitative concept.

3.2

aneurysm clip

device primarily intended for the permanent occlusion of the neck or sac of an intracranial aneurysm

3.3

closing force

force produced between the blades of the clip

3.3.1

nominal closing force

closing force defined by the manufacturer for each type of clip

3.3.2

actual closing force

closing force measured on each clip by the manufacturer before packaging

3.4

image artifact

inappropriate image signal in an MR image

NOTE Image artifact may be characterized as decreased signal intensity (voids) where signal should be produced, with or without geometric image distortion, but can also include abnormally increased signal intensity.

3.5

magnetic properties

property of a material to become magnetized when subjected to a magnetic field

NOTE 1 Materials which are ferro- or antiferromagnetic are strongly magnetic. Dia- and paramagnetic materials are weakly magnetic.

NOTE 2 Materials which can exhibit strongly magnetic properties are not suitable for the manufacturer of aneurysm clips.

3.6

magnetic induction

B

vector indicating both direction and magnitude of a magnetic field induced by an electric current flowing through conducting wire or wires

NOTE 1 It is expressed in teslas (T) or volt seconds per square metre.

NOTE 2 Values of magnetic inductance up to 2 T are used at the time of publication of this International Standard.

3.7

MRI safe

(of a device) demonstrated to present no additional risk to the patient when used in the MRI environment, but may affect the quality of the diagnostic information

NOTE MRI safe does not imply MRI compatibility in terms of magnetism.

3.8

repeatability

ability of a measuring instrument to provide closely similar indications for repeated applications of the same measurand under the same conditions

NOTE These conditions include

- reduction to a minimum of the variations due to the observer,
- the same measurement procedure,
- the same observer,
- the same measurement equipment, used under the same conditions,
- the same location,
- repetition over a short period of time.

4 Description of aneurysm clips

4.1 Mechanism of action

The description of certain clip mechanisms and their gripping action is shown in Figure 1.

4.2 Geometry

Diagrammatic representation (not to scale) of some examples of clip forms is indicated in Figure 2.

5 Indication of dimensions

The following dimensions of clips and components shall be indicated:

- a) the overall length;
- b) the length of the blades;
- c) the width of the blades giving, as appropriate, the width (disregarding any radius or taper at the tip) of blades of uniform width, the minimum and maximum widths of non-uniform blades, and the overall width of fenestrated blades;
- d) the internal diameter of any encircling or encompassing portions of the clip.

The variety of designs of clip does not make it feasible to specify the points between which the blade length should be measured. Manufacturers should indicate these points clearly on all diagrams. Examples of indication of dimensions are given in Figure 3. The diagrams are for illustration only and do not indicate a definitive requirement.

NOTE It is suggested that the blade length be indicated as that portion of the jaw which comes into contact with the other jaw when the clip is closed without a vessel in place or, for encircling clips, the longitudinal internal dimension of the clip when closed.

6 Materials

The materials shall comply with the requirements of ISO 5832-2, ISO 5832-3, ISO 5832-5, ISO 5832-6, ISO 5832-7 or ISO 5832-8.

Stainless steel is excluded as a material for aneurysm clips.




Figure 1 — Examples of clip mechanisms and their gripping action




Figure 2 — Examples of clip forms (not to scale)

Dimensions in millimetres




Figure 3 — Examples of dimensioned diagrams of clips