
**Guide for selection and calibration of
dosimetry systems for radiation
processing**

Guide de choix et d'étalonnage des appareils de mesure
dosimétrique pour le traitement par irradiation



Reference number
ISO/ASTM 51261:2002(E)

© ISO/ASTM International 2002

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. Neither the ISO Central Secretariat nor ASTM International accepts any liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies and ASTM members. In the unlikely event that a problem relating to it is found, please inform the ISO Central Secretariat or ASTM International at the addresses given below.

STANDARDSISO.COM : Click to view the full PDF of ISO/ASTM 51261:2002

© ISO/ASTM International 2002

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester. In the United States, such requests should be sent to ASTM International.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
E-mail copyright@iso.ch
Web www.iso.ch

ASTM International, 100 Barr Harbor Drive, PO Box C700,
West Conshohocken, PA 19428-2959, USA
Tel. +610 832 9634
Fax +610 832 9635
E-mail khooper@astm.org
Web www.astm.org

Printed in the United States

Contents

Page

1 Scope	1
2 Referenced documents	1
3 Terminology	2
4 Significance and use	2
5 Dosimeter classes and applications	3
6 Selection of dosimetry systems	4
7 Analytical instrument calibration and performance verification	4
8 Dosimetry system calibration	5
9 Interpretation of absorbed dose in a product	9
10 Minimum documentation requirements	9
11 Measurement uncertainty	10
12 Keywords	10
Annexes	10
Bibliography	18
Figure 1 Example of calibration package allowing alanine dosimeters to be placed on either side of thin film routine dosimeters	8
Figure 2 Example of calibration package allowing reference standard ampoules and routine dosimeters to be placed adjacent to each other	8
Figure 3 Example of calibration package allowing alanine dosimeters and thin-film dosimeters to be irradiated at the same position on the depth-dose curve	9
Figure A1.1 Ratios of mass energy absorption coefficients	12
Figure A1.2 Ratios of mass energy absorption coefficients—expanded view	12
Figure A1.3 Ratios of mass collision stopping powers	13
Figure A2.1 Transit dose effect	14
Table 1 Examples of reference standard dosimeters	3
Table 2 Examples of routine dosimeters	3
Table A1.1 Electron mass collision stopping powers	10
Table A1.2 Photon mass energy absorption coefficients	11
Table A4.1 Alanine/EPR dosimeter	15
Table A4.2 Calorimetric dosimetry systems	15
Table A4.3 Cellulose acetate dosimeter	15
Table A4.4 Ceric cerous sulfate dosimeter	15
Table A4.5 Potassium/silver dichromate dosimeter	16
Table A4.6 Polymethylmethacrylate dosimeter	16
Table A4.7 Ethanol chlorobenzene dosimeter	16
Table A4.8 Ferrous sulfate (Fricke) dosimeter	16
Table A4.9 Radiochromic liquid dosimeter	16
Table A4.10 Radiochromic film dosimeter	16
Table A4.11 Radiochromic optical waveguide dosimeter	17
Table A4.12 Thermoluminescence dosimeter (TLD)	17
Table A4.13 Ferrous cupric sulfate dosimeter	17
Table A4.14 MOSFET Dosimeter	17

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75% of the member bodies casting a vote.

ASTM International is one of the world's largest voluntary standards development organizations with global participation from affected stakeholders. ASTM technical committees follow rigorous due process balloting procedures.

A pilot project between ISO and ASTM International has been formed to develop and maintain a group of ISO/ASTM radiation processing dosimetry standards. Under this pilot project, ASTM Subcommittee E10.01, Dosimetry for Radiation Processing, is responsible for the development and maintenance of these dosimetry standards with unrestricted participation and input from appropriate ISO member bodies.

Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. Neither ISO nor ASTM International shall be held responsible for identifying any or all such patent rights.

International Standard ISO/ASTM 51261 was developed by ASTM Committee E10, Nuclear Technology and Applications, through Subcommittee E10.01, and by Technical Committee ISO/TC 85, Nuclear Energy.

Annexes A1, A2, A3, A4 and A5 of this International Standard are for information only.



Standard Guide for Selection and Calibration of Dosimetry Systems for Radiation Processing¹

This standard is issued under the fixed designation ISO/ASTM 51261; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision.

1. Scope

1.1 This guide covers the basis for selecting and calibrating dosimetry systems used to measure absorbed dose in gamma-ray or X-ray fields and in electron beams used for radiation processing. It discusses the types of dosimetry systems that may be employed during calibration or on a routine basis as part of quality assurance in commercial radiation processing of products. This guide also discusses interpretation of absorbed dose and briefly outlines measurements of the uncertainties associated with the dosimetry. The details of the calibration of the analytical instrumentation are addressed in individual dosimetry system standard practices.

1.2 The absorbed-dose range covered is up to 1 MGy (100 Mrad). Source energies covered are from 0.1 to 50 MeV photons and electrons.

1.3 This guide should be used along with standard practices and guides for specific dosimetry systems and applications covered in other standards.

1.4 Dosimetry for radiation processing with neutrons or heavy charged particles is not covered in this guide.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:

E 170 Terminology Relating to Radiation Measurements and Dosimetry²

E 178 Practice for Dealing with Outlying Observations³

E 456 Terminology Relating to Quality and Statistics³

E 666 Practice for Calculating Absorbed Dose from Gamma or X Radiation²

E 668 Practice for Application of Thermoluminescence-Dosimetry (TLD) Systems for Determining Absorbed Dose In Radiation-Hardness Testing of Electronic Devices²

E 1026 Practice for Using the Fricke Reference Standard Dosimetry System²

2.2 ISO/ASTM Standards:

51204 Practice for Dosimetry in Gamma Irradiation Facilities for Food Processing²

51205 Practice for Use of a Ceric-Cerous Sulfate Dosimetry System²

51275 Practice for Use of a Radiochromic Film Dosimetry System²

51276 Practice for Use of a Polymethylmethacrylate Dosimetry System²

51310 Practice for Use of a Radiochromic Optical Waveguide Dosimetry System²

51400 Practice for Characterization and Performance of a High-Dose Radiation Dosimetry Calibration Laboratory²

51401 Practice for Use of a Dichromate Dosimetry System²

51431 Practice for Dosimetry in Electron and Bremsstrahlung Irradiation Facilities for Food Processing²

51538 Practice for Use of the Ethanol-Chlorobenzene Dosimetry System²

51540 Practice for Use of a Radiochromic Liquid Dosimetry System²

51607 Practice for Use of the Alanine-EPR Dosimetry System²

51631 Practice for Use of Calorimetric Dosimetry Systems for Electron Beam Dose Measurements and Dosimeter Calibrations²

51649 Practice for Dosimetry in an Electron Beam Facility for Radiation Processing at Energies between 300 keV and 25 MeV²

51650 Practice for Use of Cellulose Acetate Dosimetry Systems²

51702 Practice for Dosimetry in a Gamma Irradiation Facility for Radiation Processing²

51707 Guide for Estimating Uncertainties in Dosimetry²

51956 Practice for the Use of Thermoluminescence-Dosimetry (TLD) Systems for Radiation Processing²

2.3 International Commission on Radiation Units and Measurements Reports:

ICRU Report 14 Radiation Dosimetry: X-rays and Gamma rays with Maximum Photon Energies Between 0.6 and 50 MeV⁴

ICRU Report 17 Radiation Dosimetry: X-rays Generated at Potentials of 5 to 150 kV⁴

ICRU Report 34 The Dosimetry of Pulsed Radiation⁴

¹ This guide is under the jurisdiction of ASTM Committee E10 on Nuclear Technology and Applications and is the direct responsibility of Subcommittee E10.01 on Dosimetry for Radiation Processing, and is also under the jurisdiction of ISO/TC 85/WG 3.

Current edition approved Jan. 22, 2002. Published March 15, 2002. Originally published as ASTM E 1261 – 88. Last previous ASTM edition E 1261 – 00. ASTM E 1261 – 94¹ was adopted by ISO in 1998 with the intermediate designation ISO 15556:1998(E). The present International Standard ISO/ASTM 51261:2002(E) is a revision of ISO 15556.

² Annual Book of ASTM Standards, Vol 12.02.

³ Annual Book of ASTM Standards, Vol 14.02.

⁴ Available from International Commission on Radiation Units and Measurements, 7910 Woodmont Avenue, Suite 800, Bethesda, MD 20814, USA.



ICRU Report 35 Radiation Dosimetry: Electron Beams with Energies between 1 and 50 MeV⁴

ICRU Report 37 Stopping Powers for Electrons and Positrons⁴

ICRU Report 60 Radiation Quantities and Units⁴

3. Terminology

3.1 Definitions:

3.1.1 *accredited dosimetry calibration laboratory*—a laboratory that meets specific performance criteria and has been tested and approved by a recognized accrediting organization.

3.1.2 *calibration curve*—graphical representation of the dosimetry system's response function.

3.1.3 *calibration facility*—combination of an ionizing radiation source and its associated instrumentation that provides a uniform and reproducible absorbed dose, or absorbed dose rate, traceable to national or international standards at a specified location and within a specific material, and that may be used to derive the dosimetry system's response function or calibration curve.

3.1.4 *charged particle equilibrium*—the condition that exists in an incremental volume within a material under irradiation if the kinetic energies and number of charged particles (of each type) entering the volume are equal to those leaving that volume.

3.1.4.1 *Discussion*—When electrons are the predominant charged particle, the term “electron equilibrium” is often used to describe charged-particle equilibrium. See also the discussions attached to the definitions of kerma and absorbed dose in E 170.

3.1.5 *dosimeter batch*—quantity of dosimeters made from a specific mass of material with uniform composition, fabricated in a single production run under controlled, consistent conditions, and having a unique identification code.

3.1.6 *dosimetry system*—system used to determine absorbed dose, consisting of dosimeters, measurement instruments and their associated reference standards, and procedures for the system's use.

3.1.7 *electron equilibrium*—charged particle equilibrium for electrons.

3.1.8 *measurement intercomparison*—a process by which an on-site measurement system is evaluated against a measurement of a standard reference device or material that is traceable to a nationally or internationally recognized standard.

3.1.8.1 *Discussion*—In radiation processing, reference standard or transfer-standard dosimeters are irradiated at one irradiation facility, and sent to another for analysis. Alternatively, an issuing laboratory may send dosimeters to an irradiation facility. The irradiated dosimeters are then sent back to the issuing laboratory for analysis.

3.1.9 *measurement quality assurance plan*—a documented program for the measurement process that assures on a continuing basis that the overall uncertainty meets the requirements of the specific application. This plan requires traceability to, and consistency with, nationally or internationally recognized standards.

3.1.10 *measurement traceability*—the ability to demonstrate by means of an unbroken chain of comparisons that a mea-

surement is in agreement within acceptable limits of uncertainty with comparable nationally or internationally recognized standards.

3.1.11 *primary-standard dosimeter*—dosimeter, of the highest metrological quality, established and maintained as an absorbed dose standard by a national or international standards organization.

3.1.12 *process load*—a volume of material with a specified loading configuration irradiated as a single entity.

3.1.13 *quality assurance*—all systematic actions necessary to provide adequate confidence that a calibration, measurement, or process is performed to a predefined level of quality.

3.1.14 *reference-standard dosimeter*—dosimeter of high metrological quality used as a standard to provide measurements traceable to, and consistent with, measurements made using primary-standard dosimeters.

3.1.15 *response function*—mathematical representation of the relationship between dosimeter response and absorbed dose for a given dosimetry system.

3.1.16 *routine dosimeter*—dosimeter calibrated against a primary, reference, or transfer-standard dosimeter and used for routine absorbed dose measurement.

3.1.17 *simulated product*—a mass of material with attenuation and scattering properties similar to those of the product, material or substance to be irradiated.

3.1.17.1 *Discussion*—Simulated product is used during irradiator characterization as a substitute for the actual product, material or substance to be irradiated. When used in routine production runs, it is sometimes referred to as a compensating dummy. When used for absorbed-dose mapping, the simulated product is sometimes referred to as phantom material.

3.1.18 *dosimeter stock*—part of a dosimeter batch held by the user.

3.1.19 *transfer-standard dosimeter*—dosimeter, often a reference-standard dosimeter, suitable for transport between different locations used to compare absorbed dose measurements.

3.1.20 *verification*—confirmation by examination of objective evidence that specified requirements have been met.

3.1.20.1 *Discussion*—In the case of measuring equipment, the result of verification leads to a decision to restore to service or to perform adjustments, repair, downgrade, or declare obsolete. In all cases it is required that a written trace of the verification performed be kept on the instrument's individual record.

3.2 Definitions of other terms used in this standard that pertain to radiation measurement and dosimetry may be found in ASTM Terminology E 170. Definitions in ASTM Terminology E 170 are compatible with ICRU 60; that document, therefore, may be used as an alternative reference.

4. Significance and Use

4.1 Ionizing radiation is used to produce various desired effects in products. Examples include the sterilization of medical products, processing of food, modification of polymers, irradiation of electronic devices, and curing of inks,



coatings, and adhesives (1, 2)⁵. The absorbed doses employed in these applications range from about 10 Gy to more than 100 kGy.

4.2 Regulations for sterilization of medical products and radiation processing of food exist in many countries. These regulations may require that the response of the dosimetry system be calibrated and traceable to national standards (3, 4, 5). Adequate dosimetry, with proper statistical controls and documentation, is necessary to ensure that the products are properly processed.

4.3 Proper dosimetric measurements must be employed to ensure that the product receives the desired absorbed dose. The dosimeters must be calibrated. Calibration of a routine dosimetry system can be carried out directly in a national or accredited standards laboratory by standardized irradiation of routine dosimeters. Alternatively, it may be carried out through the use of a local (in-house) calibration facility (6) or in a production irradiator. All possible factors that may affect the response of dosimeters, including environmental conditions and variations of such conditions within a processing facility, should be known and taken into account. The associated analytical instrumentation must also be calibrated.

5. Dosimeter Classes and Applications

5.1 Dosimeters may be divided into four basic classes in accordance with their relative quality and areas of applications as follows:

5.1.1 *Primary-Standard Dosimeter*—Primary-standard dosimeters are established and maintained by national standards laboratories for calibration of radiation environments (fields). Primary-standard dosimeters are typically used to calibrate or intercompare radiation environments in dosimetry calibration laboratories, and are not normally used as routine dosimeters. Discussions about the selection and calibration of primary-standard dosimeters are provided in 6.1 and 8.2.1, respectively. The two most commonly used primary-standard dosimeters are ionization chambers and calorimeters (for details, see ICRU Reports 14, 17, 34, and 35).

5.1.2 *Reference-Standard Dosimeter*—Reference-standard dosimeters are used to calibrate radiation environments and to calibrate routine dosimeters. Reference-standard dosimeters may also be used in routine dosimetry applications for radiation processing where higher quality dosimetry measurements are desired. Widely used reference dosimeters include the ferrous sulfate (Fricke) aqueous solution (see ASTM Practice E 1026) and the alanine-EPR dosimetry system (see ISO/ASTM Practice 51607). Discussions about the selection and calibration of reference-standard dosimeters are provided in 6.2 and 8.2.2, respectively. Devices used as primary-standard dosimeters may also be used as reference-standard dosimeters, in which case they shall be calibrated (see 8.2.2). Examples of reference dosimeters are listed in Table 1; more details of the characteristics of several systems may be found in Annex A4.

5.1.3 *Routine Dosimeters*—Routine dosimeters are used in radiation processing facilities for absorbed dose mapping and

TABLE 1 Examples of Reference-Standard Dosimeters

Dosimeter	Readout System	Useful Absorbed Dose, Gy	References ^a
Ionization chamber	Electrometer	10 ⁻⁴ to 10	(11,12)
Calorimeter	Thermometer	10 ² to 10 ⁵	(13)
Alanine	EPR spectrometer	1 to 10 ⁵	(14)
Ceric cerous sulfate solution	UV spectrophotometer or electrochemical potentiometer	10 ³ to 10 ⁵	(15,16)
Ethanol chlorobenzene solution	Spectrophotometer, color titration, high frequency conductivity	10 to 2 × 10 ⁶	(17, 18)
Ferrous sulfate solution	UV spectrophotometer	20 to 4 × 10 ²	(19)
Potassium/silver dichromate	UV/visible spectrophotometer	10 ³ to 10 ⁵	(20)

^aThese references are not exhaustive; others may be found in the literature.

process monitoring. Discussions about the selection and calibration of routine dosimeters are provided in 6.2 and 8.4, respectively. Examples of routine dosimeters are listed in Table 2; more details of the characteristics of several of these systems may be found in Annex A4.

5.1.4 *Transfer-Standard Dosimeters*—Transfer-standard dosimeters are specially selected dosimeters used for transferring dose information from an accredited or national standards laboratory to an irradiation facility in order to establish traceability for that calibration facility. Transfer-standard dosimeters should be used under conditions specified by the issuing laboratory. They may be either reference-standard dosimeters (Table 1) or routine dosimeters (Table 2) that have characteristics meeting the requirements of the particular application. In addition to the references given in Tables 1 and 2, relevant information on some other types of dosimeters may be found in ASTM Practice E 668 and in ISO/ASTM Practices 51275 and 51276.

NOTE 1—None of the reference-standard dosimeters or routine dosimeters listed have all of the desirable characteristics given in Section 6 for

TABLE 2 Examples of Routine Dosimeters

Dosimeter	Readout System	Useful Absorbed Dose, Gy	References ^a
Alanine	EPR spectrometer	1 to 10 ⁵	(14)
Dyed polymethylmethacrylate	Visible spectrophotometer	10 ² to 10 ⁵	(21,22,23)
Clear polymethylmethacrylate	UV spectrophotometer	10 ³ to 10 ⁵	(21,24)
Cellulose acetate	Spectrophotometer	10 ⁴ to 4 × 10 ⁵	(25)
Lithium borate, lithium fluoride	Thermoluminescence reader	10 ⁻⁴ to 10 ³	(26)
Lithium fluoride (optical grade)	UV/Visible spectrophotometer	10 ² to 10 ⁶	(27)
Radiochromic dye films, solutions, optical wave guide	Visible spectrophotometer	1 to 10 ⁵	(6,8,28)
Ceric cerous sulfate solution	Potentiometer or UV spectrophotometer	10 ³ to 10 ⁵	(15)
Ferrous cupric sulfate solution	UV spectrophotometer	10 ³ to 5 × 10 ³	(29)
Ethanol chlorobenzene solution	Spectrophotometer, color titration, high-frequency conductivity	10 to 2 × 10 ⁶	(18)
Amino acids	Lyoluminescence reader	10 ⁻⁵ to 10 ⁴	(30)
MOSFET	Voltmeter	1 to 2 × 10 ²	(31)

^aThese references are not exhaustive; others may be found in the literature.

⁵ The boldface numbers given in parentheses refer to the bibliography at the end of this guide.



an “ideal” transfer–standard dosimeter. However, such dosimeters may be used as transfer–standard dosimeters if the absence of one or more desirable characteristics has negligible effect on the response of the dosimeter, or if correction factors can be applied to bring the dosimeter’s response into conformity within the necessary limits of uncertainty for the application.

6. Selection of Dosimetry Systems

6.1 *Primary Standard Dosimetry System*—The criterion for the selection of a specific primary–standard dosimeter by a national laboratory depends on the specific measurement application requirement.

6.2 *Reference Standard, Transfer Standard and Routine Dosimetry Systems*

6.2.1 *General Criteria:*

6.2.1.1 Suitability of the dosimeter for the absorbed-dose range of interest and for use with a specific product,

6.2.1.2 Stability and reproducibility of the system,

6.2.1.3 Ease of system calibration,

6.2.1.4 Ability to control or correct system response for systematic error, such as those caused by temperature and humidity (for example, see Ref (6)),

6.2.1.5 Overall initial and operational cost of the system, including dosimeters, readout equipment, and labor,

6.2.1.6 Variance (that is, correlation coefficient) of the dosimetry system response data within established limits about a fitted calibration curve over the absorbed-dose range of interest,

6.2.1.7 Dependence of dosimeter response on environmental conditions (such as temperature, humidity, and light) before, during, and after calibration and production irradiation. Effects of environmental conditions on the dosimeter readout equipment shall also be considered.

6.2.1.8 Dependence of dosimeter response on absorbed-dose rate or incremental delivery of absorbed dose, or both,

6.2.1.9 Stability of dosimeter response both before and after irradiation,

6.2.1.10 Agreement of dosimeter response within a batch and between batches,

6.2.1.11 Effects of size, location, orientation, and composition of the dosimeter on the radiation field or the interpretation of the absorbed-dose measurement. In cases where it is desirable to measure absorbed dose at the interface of different materials (for example, at a bone-tissue interface or the surface of a product), dosimeters should be used that are thin compared to distances over which the absorbed-dose gradient is significant, and

6.2.1.12 Effects of differences in radiation energy spectra between calibration and product irradiation fields.

NOTE 2—Availability of adequate information on the performance characteristics of the dosimetry systems should be considered in selecting a dosimetry system.

6.2.2 *Additional Criteria Specific to Routine Dosimetry Systems:*

6.2.2.1 Ease and simplicity of use

6.2.2.2 Availability of dosimeters in reasonably large quantities

6.2.2.3 Time required for dosimeter response development,

and labor required for dosimeter readout and interpretation, and

6.2.2.4 Ruggedness of the system (resistance to damage during routine handling and use in a processing environment).

6.2.3 *Additional Criteria Specific to Transfer Standard Dosimetry Systems:*

6.2.3.1 Long pre-irradiation shelf life,

6.2.3.2 Post-irradiation response stability (ability to be archived), and

6.2.3.3 Portability, that is ability to withstand shipping to an irradiation facility and insensitivity to extremes of environmental conditions during transport.

7. Analytical Instrument Calibration and Performance Verification

7.1 Before the overall system is calibrated, and at periodic intervals between calibrations, the individual component instruments of the dosimetry system shall be calibrated or shall have their performance verified. These periodic checks should verify the stability of the components of the system, and should demonstrate that the components are performing as they were when the overall system was calibrated.

7.1.1 If appropriate standards exist, calibrate the individual instruments in accordance with documented procedures so that the instruments’ measurements are traceable to nationally or internationally recognized standards.

7.1.1.1 For optical absorbance measurements using a spectrophotometer, check and document that the wavelength and absorbance scales are within the documented specifications at or near the analysis wavelength using optical absorbance filters and wavelength standards traceable to national or international standards.

7.1.1.2 For thickness measurements using a thickness gauge, check and document that the instrument is within documented specifications using gauge blocks traceable to national or international standards.

7.1.2 For dosimetry system instruments where nationally or internationally recognized standards do not exist, verify correct instrument performance using documented industry or manufacturer’s procedures to demonstrate that the instrument is functioning in accordance with its own performance specifications.

NOTE 3—For example, the alanine dosimetry system employs electron paramagnetic resonance (EPR) spectroscopy for analysis. The proper operation of the EPR spectrometer instrumentation is verified with appropriate EPR spin standards such as irradiated alanine dosimeters, pitch sample, or Mn(II) in CaO (see ISO/ASTM Practice 51607 for details).

7.1.3 Repeat instrument calibration or instrument performance verification at periodic intervals between the overall system calibrations in accordance with documented procedures, and again if any maintenance or modification of the instrument occurs that may affect its performance.

NOTE 4—For some analytical instrumentation, correct performance can be demonstrated by showing that the readings of dosimeters given known absorbed doses are in agreement with the expected readings within the limits of the dosimetry system uncertainty. This method is only applicable for reference standard dosimetry systems where the long term stability of



the response has been demonstrated and documented.

7.1.4 Instrument calibrations and instrument performance verifications shall be conducted by qualified individuals in accordance with documented quality procedures.

7.1.5 Calibration or performance verification of each instrument shall demonstrate that the measurements are within specified limits over the full range of utilization.

8. Dosimetry System Calibration

8.1 General:

8.1.1 The calibration of a dosimetry system consists of the irradiation of dosimeters to a number of known absorbed doses over the range of use, analysis of the dosimeters using calibrated analytical equipment, and the generation of a calibration curve or response function. Calibration verification is performed periodically to confirm the continued validity of the calibration curve or response function. Calibration facilities shall meet the requirements specified in ISO/ASTM Practice 51400 and therefore shall have an absorbed-dose rate that has measurement traceability to nationally or internationally recognized standards.

NOTE 5—In several countries, the national standard is realized indirectly through a calibrated radiation field. For example, the absorbed-dose rate in the center of a ^{60}Co source array at the U.S. National Institute of Standards and Technology (NIST) has been well characterized by calorimetry and is one of the national standards used by NIST to irradiate reference standard, transfer standard, and routine dosimeters to known absorbed-dose levels.

8.1.2 Procedures, protocols, and training of personnel shall be provided to ensure that the correct absorbed dose is given to dosimeters.

8.2 Calibration of Dosimetry Classes:

8.2.1 *Primary—Standard Dosimeters*—These devices do not require calibration against other standards because their measurements are based on fundamental physical principles.

8.2.2 *Reference—Standard Dosimeters*—Calibration of reference—standard dosimeters is carried out by national or accredited laboratories using criteria specified in ISO/ASTM Practice 51400.

8.2.3 *Transfer—Standard Dosimeters*—Transfer—standard dosimeters may be selected from either reference—standard dosimeters or routine dosimeters (see 5.1.4). Transfer—standard dosimeters shall be calibrated by the class distinction requirements of the dosimeter type selected for dose measurement intercomparison transfer (see 8.2.2 and 8.2.4).

8.2.4 *Routine Dosimeters*—Calibration of routine dosimeters is performed by irradiation of the dosimeters in a calibration facility or in a production irradiator followed by analysis at the production irradiator.

8.3 Calibration Procedure:

8.3.1 The number of sets of dosimeters required to determine the calibration curve or response function of the dosimetry system depends on the absorbed-dose range of utilization. Use at least five sets for each factor of ten span of absorbed dose, or at least four sets if the range of utilization is less than a factor of ten.

NOTE 6—To determine mathematically the minimum number of sets to be used, divide the maximum dose in the range of utilization (D_{max}) by the

minimum dose (D_{min}), then, calculate log(base 10) of this ratio: $Q = \log(D_{\text{max}}/D_{\text{min}})$. If Q is equal to or greater than 1, calculate the product of $5 \times Q$, and round this up to the nearest integer value. This value represents the minimum number of sets to be used.

8.3.2 For each absorbed dose point, use the number of dosimeters required to achieve the desired confidence level (see ASTM Practice E 668).

8.3.3 Position the dosimeters in the calibration radiation field in a defined, reproducible location. The variation in absorbed-dose rate within the volume occupied by the dosimeters should be as low as practically possible.

8.3.4 When using a gamma-ray source or X-ray beam for calibration, surround the dosimeter with a sufficient amount of material to achieve approximate electron equilibrium conditions (7).

NOTE 7—The appropriate thickness of such material depends on the energy of the radiation (see ASTM Practices E 666 and E 668). For measurement of absorbed dose in water, use materials that have radiation-absorption properties essentially equivalent to water. For example, for a ^{60}Co source, 3 to 5 mm of solid polystyrene (or equivalent polymeric material) should surround the dosimeter in all directions.

8.3.5 Monitor and, if required, control the temperature of the dosimeters during irradiation.

NOTE 8—To minimize temperature extremes and to aid in the measurement of dosimeter temperature it is important that there be good thermal contact between the dosimeters and a heat sink especially for electron beam or x-ray irradiation.

8.3.6 If the response of the dosimeters is affected by humidity and they are not in a sealed container, monitor, and if required, control the relative humidity during irradiation.

8.3.7 Specify the calibration dose in terms of the material of interest. The calibration dose is usually specified in terms of absorbed dose in water. See Annex A1 for conversion factors for calculating the absorbed dose in different materials.

8.4 Absorbed Dose Rate Effects:

8.4.1 For some routine dosimetry systems, the dosimeter response at different absorbed-dose rates for the same given absorbed dose may differ over portions of the system's working range. This divergence may be dependent on several factors, such as the magnitude of the absorbed dose and type of radiation (gamma, electron beam, or X ray). If the absorbed dose rate effects are not known, divergence tests shall be performed. In these tests, other factors that could influence dosimeter response, for example irradiation temperature, should be either fixed or kept within a narrow range. The divergence may be checked by several methods. Two such methods are described in Annex A5.

8.5 Transit Dose Effects:

8.5.1 Transit dose effects occur when the timing of a calibration irradiation does not take into account the dose received during the movement of the dosimeters or the source into and out of the irradiation position. For example, the timer on a Gammacell-type irradiator does not start until the sample drawer reaches the fully-down (irradiate) position. Some dose is received as the drawer goes down and after the irradiation as the drawer goes up that is not accounted for in the timer setting. This transit dose can be significant for low-dose irradiations and should be determined experimentally and taken into



account when calculating timer settings. Two methods of determining transit dose correction are given in Annex A2.

8.6 *Frequency of Calibration and Verification:*

8.6.1 Calibrate the dosimetry system for each new batch of reference standard, transfer standard, or routine dosimeters prior to use.

8.6.2 At an interval not exceeding one year, re-calibrate the dosimetry system for each batch of reference standard, transfer standard, or routine dosimeters unless the dosimetry system is more frequently validated by measurement intercomparison. This re-calibration shall include the analytical instrumentation (see Section 7). Depending on seasonal variations in ambient conditions (for example, temperature and relative humidity) the interval between re-calibrations may be decreased (see Note 4).

8.6.3 Verify the calibration of the dosimetry system for each new stock of reference standard, transfer standard, and routine dosimeters using at least three absorbed doses over the range of application in order to confirm that their responses are the same as for the current stock.

8.7 *Calibration and Dose Measurement Uncertainties:*

8.7.1 The uncertainties in the calibration and absorbed dose measurement of a dosimetry system depend on the specific dosimetry system employed. Refer to ISO/ASTM Guide 51707 and the appropriate standard for a given dosimetry system for uncertainty statements. See 2.1 and Annex A4 for references to the appropriate dosimetry system standards.

8.8 *Irradiation of Dosimeters*—The irradiation of routine dosimeters may be performed by one of three different methods. One method (described in 8.8.1) calls for routine dosimeters to be irradiated in a gamma, electron beam, or X-ray (bremsstrahlung) calibration facility. The second method (described in 8.8.2) calls for routine dosimeters to be irradiated in an in-house calibration facility that has an absorbed-dose rate measured by reference or transfer-standard dosimeters. The third method (described in 8.8.3) calls for routine dosimeters to be irradiated together with reference or transfer-standard dosimeters in the production irradiator.

NOTE 9—The response of some routine dosimeters may be affected by the combined effect of several environmental factors such as absorbed dose rate, energy spectrum, temperature, or relative humidity (including seasonal variations in temperature or relative humidity). For these cases, it may not be possible to take these combined effects into account by applying a correction factor. Therefore, the calibration of routine dosimeters should be performed using irradiation conditions similar to those in the actual production irradiator.

NOTE 10—The response of reference and transfer-standard dosimeters to environmental effects such as temperature and relative humidity, absorbed-dose rate, and energy spectrum, should be documented. Differences in the dosimeter response between calibration and use conditions should be taken into account using known correction factors.

8.8.1 *Irradiation Using a Calibration Facility*—The calibration of routine dosimeters using a gamma, electron beam, or X-ray calibration facility meeting the requirements of ISO/ASTM Practice 51400 has the advantage that the dosimeters are irradiated to accurately known absorbed doses under well-controlled and documented conditions. However, use of these routine dosimeters under different environmental condi-

tions in a production irradiator may introduce uncertainties that are difficult to quantify. Transporting of the dosimeters to and from the calibration facility may also introduce uncertainties from pre- and post-irradiation storage effects.

8.8.1.1 Irradiate routine dosimeters to known absorbed doses in a calibration facility.

8.8.1.2 Specify to the calibration facility that the irradiation conditions should be as similar as possible to those in the actual production irradiator. These conditions, including the energy spectrum, absorbed-dose rate, and irradiation temperature, should be as close as practical to those encountered during routine use. The variation of these conditions should be documented and incorporated into the uncertainty analysis.

NOTE 11—The temperature and absorbed dose rate experienced by dosimeters may vary during a production run. However, it is usually not practical to perform the calibration of the dosimeters under the same varying conditions. To approximate production irradiator conditions, the calibration should be performed at a fixed temperature somewhere between the average and the maximum temperatures encountered during routine production; and, if a calibration facility with the desired dose rate is available, at a fixed dose rate somewhere between the average and the maximum dose rates experienced during routine production. The possible effects of differences in temperature and dose rate between the calibration facility and production irradiator can be minimized by performing the calibration using the method described in 8.8.3.

8.8.1.3 Adverse environmental conditions (such as high or low temperature and humidity) during transport of the dosimeters to and from the calibration facility may affect the dosimeter response.

8.8.1.4 Package dosimeters to minimize the effects of environmental conditions during transport.

8.8.1.5 Include maximum temperature indicators in the dosimeter package during transport to determine if the calibration has been compromised and is therefore invalid.

8.8.1.6 Confirm that the environmental conditions during transport have not changed the response of the dosimeter. This may be achieved by sending a set of dosimeters irradiated to known absorbed doses along with the set of dosimeters sent for calibration. The readings of the two sets of dosimeters should be compared when they are returned with readings from additional dosimeters given the same absorbed dose and stored under controlled conditions.

8.8.1.7 Calibration curves for routine dosimetry systems obtained by irradiating dosimeters in a calibration facility irradiator shall be verified for the actual conditions of use in the production irradiator (see 8.8.1.8).

8.8.1.8 Calibration verification may be performed by irradiating the routine dosimeters together with reference standard or transfer-standard dosimeters to a minimum of three different absorbed doses in the production irradiator. The reference-standard or transfer-standard dosimeters should be of a different type than the routine dosimeters to reduce the probability that both types of dosimeters are influenced by the same combined environmental effects. Ensure that the routine and reference standard or transfer dosimeters receive the same absorbed dose (see 8.7 and 8.8.3.5 for guidance). Corrective action will be required if the differences in dose readings between the routine dosimeters and the reference standard or



transfer dosimeters are significant in comparison to the accuracy required for the application. Appropriate corrective action may include: repeating the calibration using more appropriate environmental conditions, carrying out a full calibration in a production irradiator (see 8.8.3), or applying a correction factor in cases where a single factor is applicable over the entire dose range of interest.

8.8.2 Irradiation Using an In-House Calibration Facility—The calibration of routine dosimeters using an in-house calibration facility has the advantage that the pre- and post-irradiation storage conditions of the dosimeters can be controlled so that they are similar to those encountered during routine production. However, it may not be possible for the in-house calibration facility to meet all the irradiation requirements of ISO/ASTM Practice 51400.

8.8.2.1 Irradiate routine dosimetry systems to known absorbed doses in an in-house calibration facility.

8.8.2.2 The absorbed-dose rate of the in-house calibration facility shall be demonstrated to be traceable to nationally or internationally recognized standards by direct measurement intercomparisons or calibrations using transfer-standard dosimeters supplied by a nationally recognized radiation dosimetry calibration laboratory.

8.8.2.3 Measurement intercomparisons or calibrations of absorbed-dose rates of the in-house calibration facility shall be performed at least once every three years and after any change in source activity or geometry.

8.8.2.4 The requirements of 8.8.1.2 are equally applicable to an in-house calibration facility.

8.8.2.5 Calibration of routine dosimeters using an in-house calibration facility reduces the possibility of changes in the response due to adverse storage conditions during transport of dosimeters. After irradiation, the irradiated dosimeters should be stored under similar conditions to those encountered in the production irradiator and read at approximately the same time after irradiation as the dosimeters used in routine production dosimetry.

8.8.2.6 Calibration curves for routine dosimetry systems obtained by irradiating dosimeters using an in-house calibration facility shall be verified (see 8.8.2.7) for the actual conditions of use in the production irradiator.

8.8.2.7 Calibration verification may be performed by irradiating the routine dosimeters together with reference standard or transfer-standard dosimeters to three different absorbed doses in the production irradiator. The reference standard or transfer-standard dosimeters should be of a different type than the routine dosimeters to reduce the probability that both types of dosimeters are influenced by the same combined environmental effects. Ensure that the routine and reference standard or transfer dosimeters receive the same absorbed dose (see 8.7 and 8.8.3.5 for guidance). Corrective action will be required if the differences in dose readings between the routine dosimeters and the reference standard or transfer dosimeters are significant in comparison to the accuracy required for the application. Appropriate corrective action may include: repeating the calibration using more appropriate environmental conditions, carrying out a full calibration in a production irradiator (see 8.8.3),

or applying a correction factor in cases where a single factor is applicable over the entire dose range of interest.

8.8.3 Irradiation Using a Production Irradiator—The calibration of routine dosimeters by irradiating the dosimeters together with reference or transfer-standard dosimeters in the production irradiator has the advantage that the environmental conditions are similar to those encountered during routine production, reducing the requirement to make corrections to the routine dosimeter response for environmental effects. However, great care must be taken to ensure that the routine and reference or transfer-standard dosimeters irradiated together receive the same absorbed dose (8). For some production irradiators, calibration by this technique may not be practical because of limitations in covering the full range of utilization during irradiator operations.

8.8.3.1 Calibration of routine dosimeters in the production irradiator using reference or transfer-standard dosimeters provides a calibration curve or response function valid for the actual production irradiation conditions existing during the calibration. This method takes combined environmental factors into account to the extent that the reference or transfer dosimeter response can be corrected for differences in environmental factors between the calibration facility and production irradiator.

8.8.3.2 Use reference-standard or transfer-standard dosimeters supplied and analyzed by a nationally recognized radiation dosimetry calibration laboratory to demonstrate traceability to national standards.

8.8.3.3 Reference-standard or transfer-standard dosimeters obtained commercially or prepared in accordance with published standards and analyzed on site may be used provided that the reference standard or transfer standard dosimetry systems have been calibrated in accordance with 8.2 and 8.3.

8.8.3.4 Calibrate routine dosimeters by irradiating them together with reference-standard or transfer-standard dosimeters under actual production irradiation conditions over the entire range of normal use.

8.8.3.5 Design a calibration package that minimizes the difference between the absorbed doses received by the routine and reference standard or transfer dosimeters. The package should hold the two types of dosimeters so that they do not significantly shield each other, that the geometry is appropriate for the radiation source employed, and that they are as close together as possible. See Figs. 1 and 2 for examples of such packages employed for gamma-ray or X-ray sources. See Fig. 3 for an example of a package suitable for electron-beam sources.

NOTE 12—The absorbed dose variation discussed in 8.8.3.5 can be determined by irradiating calibration packages containing transfer or reference-standard dosimeters in all dosimeter positions within the calibration package.

8.8.3.6 When thick and thin dosimeters are irradiated together, surround the thin dosimeters by sufficient polymeric material to ensure that the attenuation characteristics are similar and to ensure that the dosimeters receive the same dose.

8.8.3.7 To calibrate dosimeters under conditions similar to those used for processing, place the calibration packages with




FIG. 1 Example of Calibration Package Allowing Alanine Dosimeters to be Placed on Either Side of Thin Film Routine Dosimeters




FIG. 2 Example of Calibration Package Allowing Reference Standard Ampoules and Routine Dosimeters to be Placed Adjacent to Each Other

product or simulated product in process loads where the absorbed-dose variation over the area containing dosimeters is within specified limits.

NOTE 13—The absorbed-dose rate, temperature, and energy spectrum may vary depending on the location of the dosimeter on or in the process load. These factors may have to be considered when evaluating overall dosimeter uncertainties during routine use.

8.8.3.8 All guidance given in 8.8.1.3-8.8.1.6 shall be met for the transport of the reference- or transfer-standard dosimeters.

8.9 Analysis of Dosimeters:

8.9.1 Check the performance of the analytical instrumentation (see 7.1).

8.9.2 Analyze dosimeters using analytical instrumentation with calibration traceable to national standards.

8.9.3 If the dosimeter response changes with time after


irradiation, analyze dosimeters at approximately the time after irradiation when dosimeters will be analyzed during routine production. Alternately, the time dependence of the dosimeter response after irradiation can be characterized and a correction factor applied for the time of analysis.

8.9.4 Document and retain all analysis data.

8.10 Analysis of Calibration Data:

8.10.1 Calculate and document the mean response, $\bar{k}$, and the sample standard deviation (S_{n-1}) for each set of dosimeters at each absorbed-dose value. The sample standard deviation, S_{n-1} , is calculated from the sample data set of n values as follows:

$$S_{n-1} = \sqrt{\frac{\sum(k_i - \bar{k})^2}{n-1}} \quad (1)$$



Electron Beam Phantom

FIG. 3 Example of Calibration Package Allowing Alanine Dosimeters and Thin-Film Dosimeters to be Irradiated at the Same Position on the Depth-Dose Curve

where:

k_i = i^{th} value of k .

8.10.2 Calculate the coefficient of variation, CV, for each absorbed dose value as follows:

$$CV = \frac{S_{n-1}}{k} \times 100(\%) \quad (2)$$

NOTE 14—In general, if any CV values are greater than a value prescribed for a specific dosimetry system and application, then a redetermination of the data should be considered, or the stock of dosimeters should be rejected.

8.10.3 Graphically plot dosimeter response versus absorbed dose to obtain the calibration curve, or use a suitable computer code, or both, to derive the equivalent response function in mathematical form. Choose an analytical form (for example, linear, polynomial, or exponential) that provides an appropriate fit to the measured data (see ISO/ASTM Guide 51707).

8.10.4 Examine the resulting calibration curve or response function for goodness of fit within specified limits.

8.10.5 Repeat this calibration procedure to the extent necessary if any response value exceeds accepted statistical limits of the determined curve, and if discarding this value would result in there being insufficient data to adequately define the curve (see ASTM Practice E 178 for guidance on dealing with outliers).

8.10.6 Calibration curves or response functions generated using one analytical instrument shall not be used with another instrument unless it has been demonstrated that the measurements of the dosimeters' response agree within specified limits over the full absorbed dose range.

9. Interpretation of Absorbed Dose in a Product

9.1 Generally, the absorbed dose in an irradiated food or

polymer product is specified in terms of absorbed dose in water because most food or polymer products are nearly water-equivalent in terms of radiation absorption properties. If the radiation absorption properties of the product differ from those of water, then interpretation of the absorbed dose in the product may be accomplished by means of cavity theory (see Annex A1).

9.2 Electron equilibrium conditions may not exist within dosimeters placed throughout the product under actual processing conditions. This particularly is the case near interfaces of different materials, for example, at bone-tissue interfaces or on the surface of a product package. Absorbed dose measured under non-equilibrium conditions is sometimes used to monitor the absorbed dose within the product using the procedures described in ISO/ASTM Practices 51204, 51431, and 51702.

10. Minimum Documentation Requirements

10.1 Document the dosimetry system selected including the dosimeter manufacturer, type and batch number, and instruments used for analysis.

10.2 Document the dosimeter calibration data, including date, reference standard or transfer standard, and description of the facility used.

10.3 Document or reference a description of the radiation source used in processing, including the type, nominal activity or beam parameters, and any available information on the energy spectrum.

10.4 Document the irradiation temperature and relative humidity for dosimetry systems whose performance is affected by these environmental conditions (see 8.8, and Notes 9 and 10).

10.5 Document or reference the method used to convert



dosimetry measurements to absorbed-dose values in water.

10.6 Document the value and the assigned uncertainty of the measurements of absorbed dose.

10.7 Document or reference the measurement quality assurance plan used for the routine dosimetry.

11. Measurement Uncertainty

11.1 To be meaningful, a measurement of absorbed dose shall be accompanied by an estimate of uncertainty. Components of uncertainty shall be identified as belonging to one of two groups:

11.1.1 Those which are evaluated by statistical methods, or

11.1.2 Those which are evaluated by other means. Additional information is given in ISO/ASTM Guide 51707 and Refs (9 and 10), where these components are referred to as Type A and Type B, respectively. In reporting uncertainty, other classifications such as precision and bias may be useful.

NOTE 15—The identification of Type A and Type B uncertainties are based on the methodology adopted in 1993 by the International Organization for Standardization (ISO) for estimating uncertainty. This is different from the way uncertainty has been traditionally expressed in terms of precision and bias, where precision is a measure of the extent to which replicate measurements made under specified conditions are in

agreement, and bias is a systematic error (see ASTM Terminologies E 170 and E 456 and ASTM Practice E 177). The purpose of using the method of expressing uncertainties as Type A and Type B recommended in the ISO Guide to the Expression of Uncertainty in Measurement (10) is to promote an understanding of how uncertainty statements are arrived at and to provide a basis for the international comparison of measurement results.

NOTE 16—ISO/ASTM Guide 51707 defines possible sources of error in dosimetry performed in radiation processing facilities and offers procedures for estimating the resulting magnitude of the uncertainties in the measurement results. Basic concepts of measurement, estimate of the measured value of a quantity, “true” value, error and uncertainty are defined and discussed. Components of uncertainty are discussed and methods are given for evaluating and estimating their values. Their contributions to the standard uncertainty in the reported values of absorbed dose are considered and methods are given for calculating the combined standard uncertainty and an estimate of overall (expanded) uncertainty.

12. Keywords

12.1 absorbed dose; calibration facility; dose traceability; dosimeter; dosimeter selection; dosimeter calibration; dosimetry system; electron beam; gamma ray; ionizing radiation; primary-standard dosimeter; quality assurance; radiation processing; reference-standard dosimeter; routine dosimeter; transfer-standard dosimeter; ICS 17.240

ANNEXES

(informative)

A1. INTERPRETATION OF ABSORBED DOSE

A1.1 For irradiations using a photon source, the dosimeter may be considered as a cavity (7) in the material interest, and the interpretation of absorbed dose in materials is as follows:

A1.1.1 If the sensitive region of the dosimeter is very thin compared to the range of the highest energy secondary electrons, then most of the energy deposited in the dosimeter and in the material surrounding it results from secondary electrons produced outside the dosimeter (that is, in the equilibrium layer of material). Thus, the absorbed dose in the material, D_m is given by:

$$D_m = \frac{(S/\rho)_m}{(S/\rho)_d} D_d \quad (\text{A1.1})$$

where:

$(S/\rho)_m$ and $(S/\rho)_d$ = mass collision stopping power for the surrounding material and dosimeter, respectively, and

D_d = absorbed dose in the dosimeter.

A1.1.1.1 Values of mass collision stopping powers are given in Table A1.1 (see ICRU Report 37).

A1.1.2 If the sensitive region of the dosimeter has a thickness much greater than the range of the highest energy secondary electrons, then most of the energy deposited in it results from the secondary electrons produced within the dosimeter itself. Thus, the absorbed dose in the material is given by:

TABLE A1.1 Electron Mass Collision Stopping Powers, S/ρ (MeV cm²/g)

Electron Energy (MeV)	Air	Water	Polystyrene	Bone	LiF
0.01	19.75	22.56	22.23	19.71	17.96
0.02	11.57	13.17	12.96	11.61	10.55
0.04	6.848	7.777	7.637	6.903	6.252
0.06	5.111	5.797	5.688	5.163	4.670
0.08	4.198	4.757	4.666	4.246	3.838
0.1	3.633	4.115	4.034	3.678	3.323
0.2	2.470	2.793	2.735	2.507	2.261
0.4	1.902	2.145	2.101	1.931	1.737
0.6	1.743	1.956	1.911	1.760	1.583
0.8	1.683	1.879	1.832	1.690	1.521
1	1.661	1.844	1.794	1.658	1.491
2	1.684	1.825	1.768	1.643	1.474
4	1.790	1.877	1.816	1.697	1.513
6	1.870	1.919	1.859	1.740	1.547
8	1.931	1.951	1.891	1.773	1.572
10	1.979	1.976	1.916	1.799	1.592
20	2.134	2.051	1.989	1.874	1.654
40	2.282	2.120	2.053	1.942	1.711
60	2.347	2.157	2.089	1.978	1.742

$$D_m = \frac{(\mu_{en}/\rho)_m}{(\mu_{en}/\rho)_d} D_d \quad (\text{A1.2})$$

where:



$(\mu_{en}/\rho)_m$ and $(\mu_{en}/\rho)_d$ = mass energy-absorption coefficients for the material and dosimeter, respectively.

A1.1.2.1 Values of mass energy absorption coefficient are given in Table A1.2 and Refs (1) and (32).

A1.1.3 If the sensitive region of the dosimeter has a thickness between the two limits discussed in A1.1.1 and A1.1.2, then Eq A1.1 and Eq A1.2 may be combined with appropriate weighing factors to reflect the relative contribution of each term (11).

NOTE A1.1—The collision stopping powers and energy absorption coefficients are energy dependent; however, for low atomic number materials and for the energy range considered in this guide (0.1 to 50 MeV), the ratios of the stopping powers and energy absorption coefficients do not vary significantly as a function of energy. For these conditions, the difference in the absorbed dose given in Eq A1.1 and Eq A1.2 is usually less than 10 %. Ratios of mass energy absorption coefficients are plotted in Fig. A1.1 and Fig. A1.2. Ratios of mass collision stopping powers are plotted in Fig. A1.3.

A1.1.4 If the absorbed dose has been determined under equilibrium conditions in a material, m_1 by A1.1.1, A1.1.2, or

A1.1.3, then the absorbed dose in another material of interest, m_2 , irradiated under essentially identical conditions, is given by:

$$D_{m2} = \frac{(\mu_{en}/\rho)_{m2}}{(\mu_{en}/\rho)_{m1}} D_{m1} \quad (\text{A1.3})$$

where:

D_{m1} = absorbed dose in material m_1 ,
 D_{m2} = absorbed dose in material of interest, m_2 , and
 $(\mu_{en}/\rho)_m$ = mass energy absorption coefficient for the material m .

NOTE A1.2—For a photon energy spectrum including energies down to 50 keV, the ratios of mass energy absorption coefficient in Eq A1.2 and A1.3 are essentially equal to unity for most food or polymer materials compared to water.

A1.1.5 A correction for radiation energy dependence may be necessary in the case of materials having photon absorption properties greatly different from water (such as bone or metal). If the photon energy spectrum at the point of interest has a significant component below 0.2 MeV and, if the spectrum is known, then a more accurate absorbed-dose value can be obtained by integration of Eq A1.2 or Eq A1.3, or both, over the entire photon energy spectrum (see A1.1.3).

A1.2 For irradiations using an electron source, the absorbed dose in the material can be interpreted as follows:

A1.2.1 At any given depth in a material of interest the dose may be determined with a dosimeter that is thin compared to the range of the incident electrons. The absorbed dose may be calculated using Eq A1.1. However, all of the following conditions are necessary: (a) the depth in the material shall be less than the incident electron range; (b) the ratio of mass collision stopping powers (material/dosimeter) shall be essentially constant; and (c) the degraded electrons at a given depth shall have sufficient energy to traverse the dosimeter thickness. Therefore, for materials considered in this guide, Eq A1.1 is valid for electron energies down to about 0.05 to 0.1 MeV (33).

NOTE A1.3—For most cases, this method of absorbed-dose interpretation requires an incident beam energy of at least 0.5 MeV. However, energies down to 0.1 MeV may be used for surface treatment of products. Measurements of absorbed dose in a material at such low incident energies may be difficult to interpret.

TABLE A1.2 Photon Mass Energy Absorption Coefficients, μ_{en}/ρ (cm^2/g) (32)

Photon Energy (MeV)	Air	Water	Polystyrene	Bone	LiF
0.01	4.742	4.944	1.918	26.80	5.733
0.02	0.5389	0.5503	0.2075	3.601	0.6494
0.04	0.06833	0.06947	0.03264	0.4507	0.07890
0.06	0.03041	0.03190	0.02172	0.1400	0.03223
0.08	0.02407	0.02597	0.02160	0.06896	0.02385
0.1	0.02325	0.02546	0.02296	0.04585	0.02229
0.2	0.02672	0.03192	0.02857	0.03003	0.02484
0.4	0.02949	0.03279	0.03175	0.03069	0.02734
0.6	0.02953	0.03284	0.03182	0.03052	0.02736
0.8	0.02882	0.03206	0.03106	0.02973	0.02671
1	0.02789	0.03103	0.03006	0.02875	0.02585
2	0.02345	0.02608	0.02524	0.02421	0.02173
4	0.01870	0.02066	0.01979	0.01975	0.01733
6	0.01647	0.01806	0.01708	0.01788	0.01528
8	0.01525	0.01658	0.01550	0.01695	0.01414
10	0.01450	0.01566	0.01448	0.01644	0.01345
20	0.01311	0.01382	0.01232	0.01568	0.01211
40	0.01262	0.01298	0.01128	0.01546	0.01154
60	0.01242	0.01261	0.01086	0.01511	0.01123




FIG. A1.1 Ratios of Mass Energy Absorption Coefficients




FIG. A1.2 Ratios of Mass Energy Absorption Coefficients — Expanded View




FIG. A1.3 Ratios of Mass Collision Stopping Powers

A2. DETERMINATION OF TRANSIT DOSE

A2.1 Two methods for the determination of transit dose are as follows:

A2.1.1 In the first method, select a series of incremental irradiation times that start as close to zero as is possible with the specific dosimetry system employed. It may be necessary to utilize a dosimetry system with a higher sensitivity (lower dose capability such as alanine) than the system normally employed for routine use. Use at least five different irradiation times. Analyze the results using linear regression analysis. A graph of the results should look similar to Fig. A2.1. Note that the intercept with the Y axis gives the value of the transit dose.

A2.1.2 In the second method, select a high sensitivity dosimeter such as the Fricke. Set the irradiator timer at zero and perform ten irradiation cycles. The dosimeter receives ten times the transit dose; hence divide the total dose by ten to obtain the transit dose. Repeat this process three times, and

calculate the average of the three measurements to get the transit dose.

A2.2 To calculate a corrected timer setting for future irradiations, first convert the transit dose into an equivalent irradiation time (transit time, t_t); that is, divide the transit dose by the current dose rate ($\dot{D}$). Subsequent timer settings can be calculated using the formula:

$$t = \frac{D_T}{\dot{D}} - t_t \quad (\text{A2.1})$$

where:

D_T = the target dose and t is the timer setting required to achieve that dose.




FIG. A2.1 Transit Dose Effect

A3. SOURCES OF UNCERTAINTY

A3.1 This annex lists some of the factors that could contribute to the uncertainty in the measurement of absorbed dose for a specific dosimetry system. Because of the diverse types of dosimeters covered in this guide and the large number of possible applications, no attempt is made to quantify the potential uncertainties for a specific dosimeter type. Not all of the factors listed are applicable to any specific dosimeter system, but they illustrate the range of possibilities that may need to be considered.

A3.2 Possible factors that may contribute to the uncertainty in absorbed dose measurements are listed below. These lists are not intended to be exhaustive, but are illustrative of some of the more common sources of uncertainties in radiation dosimetry.

A3.2.1 Variation in response of a group of dosimeters irradiated to the same dose level,

A3.2.2 Variation in thickness of individual dosimeters,

A3.2.3 Dosimeter system analytical instrumentation (for example, variation in absorbance reading, wavelength setting, etc.),

A3.2.4 Reproducibility of absorbed dose delivered to dosimeters (for example, timing of irradiations, transit dose),

A3.2.5 Irradiation source decay correction for given date,

A3.2.6 Absorbed dose delivered to dosimeters during calibration,

A3.2.7 Calibration of radiation source used to irradiate dosimeters during calibration,

A3.2.8 Determination of dosimetry system response function (calibration curve),

A3.2.9 Dosimetry response dependence on environmental conditions (for example, temperature, humidity, atmosphere), before, during, and after irradiation,

A3.2.10 Stability of dosimeter response before and after irradiation,

A3.2.11 Dosimeter response dependence on dose rate or dose fractionation (dose delivered in discrete steps), or both,

A3.2.12 Dosimeter response dependence on geometrical effects (for example, the size and orientation of the dosimeter),

A3.2.13 Dosimeter response dependence on differences in energy spectrum between calibration and production radiation sources,

A3.2.14 Calibration of dosimetry system analytical equipment,

A3.2.15 Correction for attenuation in equilibrium material surrounding dosimeters, and

A3.2.16 Value of half-life of the source.

A4. EXAMPLES OF DOSIMETER CHARACTERISTICS

TABLE A4.1 Alanine/EPR Dosimeter

NOTE—For more information on this dosimetry system, see ISO/ASTM Practice 51607.

Applicable Dose Range: 1 to 10^5 Gy.
Applicable Dose Rate: $<10^8$ Gy s ⁻¹ .
Use: Electron/gamma ray/X ray
Physical Characteristics:
This dosimeter is used in the form of tablets, small rods, or rope of 3 to 5-mm diameter and various lengths, consisting primarily of α-alanine and a small amount paraffin or other binder.
Instrumentation Characteristics: EPR spectrometer.
Environmental Factors:
Temperature: Irradiation temperature coefficient varies with alanine formulation
Humidity: Somewhat sensitive to humidity. May require preconditioning.
Ambient Light: Not generally sensitive to UV light.

TABLE A4.2 Calorimetric Dosimetry Systems

NOTE—For more information on this dosimetry system, see ISO/ASTM Practice 51631.

Dose Range: Approximately 10^2 to 10^5 Gy.
Applicable Dose Rate: >10 Gy s ⁻¹ .
Use: Electron/gamma ray/X ray.
Physical Characteristics:
Generally designed in cylindrical shapes. The dimensions depend on the energy of the electron beam, gamma, or X-ray spectrum.
Instrumentation Characteristics: High sensitivity voltage or current meters.
Environmental Factors:
Temperature: Not applicable.
Humidity: Not applicable.
Ambient Light: Not applicable.

TABLE A4.3 Cellulose Acetate Dosimeter

NOTE—For more information on this dosimetry system, see ISO/ASTM Practice 51650.

Applicable Dose Range: 5×10^3 to 10^6 Gy.
Applicable Dose Rate: 3×10^{-2} to 3×10^7 Gy s ⁻¹ .
Use:
Electron/gamma ray/X-ray. Appropriate corrections factors must be applied between electron and gamma ray use.
Physical Characteristics:
Most cellulose acetate films are produced in 8-mm wide rolls especially for dosimetry.
Instrumentation Requirements: UV spectrophotometer (various wavelengths).
Environmental Factors:
Temperature: Irradiation temperature coefficient approximately +0.5 % (°C) ⁻¹ .
Humidity: This dosimeter is sensitive to changes in humidity.
Ambient Light: These dosimeters are not sensitive to ambient light conditions.

TABLE A4.4 Ceric Cerous Sulfate Dosimeter

NOTE 1—The solutions may be evaluated by platinum electrodes in terms of changes in cerous concentration according to millivolt readings of a potentiometer.

NOTE 2—For more information on this dosimetry system, see ISO/ASTM Practice 51205.

Applicable Dose Range: 5×10^2 to 10^5 Gy.
Applicable Dose Rate: $<10^6$ Gy s ⁻¹
Use: Electron/gamma ray/X ray.
Physical Characteristics:
The typical dosimeter consists of an aqueous solution of 1.5×10^{-2} mol dm ⁻³ Ce(SO ₄) ₂ , Ce ₂ (SO ₄) ₃ and 0.4 mol dm ⁻³ H ₂ SO ₄ . The dosimeter is usually irradiated in sealed 2-mL glass ampoules of 10-mm inner diameter. Before spectrophotometry, solutions are diluted 50 or 100 times with 0.4 mol dm ⁻³ sulfuric acid.
Instrumentation Requirements:
UV spectrophotometer (usual wavelength: 320 nm); electrochemical potentiometer (see Note 1).
Environmental Factors:
Temperature: Irradiation temperature coefficient varies with Ce ³⁺ ion concentration.
Humidity: Not applicable.
Ambient Light: These dosimeters are not sensitive to normal ambient light conditions.