
**Tissue-engineered medical
products — Evaluation of anisotropic
structure of articular cartilage using
DT (Diffusion Tensor)-MR Imaging**

*Produits médicaux à base de tissus — Évaluation de la structure
anisotrope du cartilage articulaire en utilisant l'imagerie en tenseur
de diffusion (IRM-TD)*



STANDARDSISO.COM : Click to view the full PDF of ISO/TR 16379:2014



COPYRIGHT PROTECTED DOCUMENT

© ISO 2014

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Terms and definitions	1
3 Principle	3
4 Diffusion tensor magnetic resonance imaging (DT-MRI) data observation in articular cartilage	3
4.1 DT-MRI measurement process	3
4.2 Notes on setting of DT-MRI imaging parameters for articular cartilage	4
4.3 Measurement indices for structural evaluation of articular cartilage by DT-MRI	5
Annex A (informative) Measurement Results	10
Bibliography	23

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: Foreword - Supplementary information

The committee responsible for this document is ISO/TC 150, *Implants for Surgery*, Subcommittee SC 7, *Tissue-engineered Medical Products*.

Introduction

Structural evaluation of articular cartilage with conventional diagnostic technologies is challenging, and Nihon University has developed technologies (see Reference [1]) and collected relevant data for *in vivo* evaluation of articular cartilage structure by means of diffusion tensor magnetic resonance imaging (DT-MRI) using 1,5 Tesla or 3 Tesla MRI equipment employed for treatment in hospital settings. These data are released in this Technical Report prepared for reference in treatment settings.

This work is part of “Development of Cartilage Observation and Evaluation Technologies for Regenerative Medicine Processes”, an activity managed by the University under “Development of Evaluation Technology for Early Introduction of Regenerative Medicine”, a project contracted by the New Energy and Industrial Technology Development Organization (NEDO) to the National Institute of Advanced Industrial Science and Technology (AIST) and its Technology Research Association of Medical Welfare Apparatus.

STANDARDSISO.COM : Click to view the full PDF of ISO/TR 16379:2014

[STANDARDSISO.COM](https://standardsiso.com) : Click to view the full PDF of ISO/TR 16379:2014

Tissue-engineered medical products — Evaluation of anisotropic structure of articular cartilage using DT (Diffusion Tensor)-MR Imaging

1 Scope

This Technical Report has been prepared for evaluation of therapeutic courses for articular cartilage disease and summarizes results from structural evaluation of knee joint cartilage by diffusion tensor imaging, an MRI applied technology allowing non-invasive observation of soft tissue morphology *in vivo*.

This Technical Report is intended for use in areas such as regenerative medicine for knee joint cartilage disease.

After *in vivo* transplant of cartilage cells or tissue as a regenerative treatment, longitudinal diagnosis is needed to assess regeneration as articular cartilage, but arthroscopes used primarily for this purpose are invasive and also do not allow evaluation of structure by simple observation of surfacial characteristics. Radiography and CT do not visualize articular cartilage and also entail the problem of exposure. Collagen fibres, the primary component of articular cartilage, have a surfacial layer parallel to the articular surface to serve a lubricating function for the articular surface, a middle layer with a randomized structure to distribute loads, and deep layers oriented vertically to support loads. The anisotropy of this three-layer structure is a characteristic feature of hyaline cartilage structures and a mechanism demonstrating a lubricating function for articular cartilage. We can then ask whether articular cartilage can be assessed by evaluating the anisotropy of collagen.

MRI techniques allow non-invasive visualization of soft tissue form and function *in vivo*, and DT-MRI conveys the direction of water molecule motion. In fibrous tissues, the direction of water molecule motion is restricted to the direction of fibre orientation; consequently, the direction of water molecule motion matches that of fibre orientation. The use of DT-MRI therefore does allow evaluation of collagen fibre orientation and anisotropy in articular cartilage.

DT-MRI is thus used to observe articular cartilage anisotropy data for use as standardized data in longitudinal diagnosis following transplant of articular cartilage as a regenerative treatment.

2 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

2.1 diffusion tensor

DT
tensor expressing the orientation and magnitude of diffused proton signals

2.2 sequence

protocol for performance of MRI

2.3 spin-echo echo-planar imaging SE-EPI

method of high-speed imaging in which gradient fields are flipped continuously at high speed to produce echoes continuously by means of a spin-echo pulse sequence

2.4

field of view

FOV

width and height of an imaged region (expressed in cm or mm)

2.5

matrix

pixel resolution for acquisition of MR signals in a field of view

2.6

echo time

TE

time after RF pulse application until an echo is produced

[SOURCE: JIS K 3611]

2.7

radio frequency pulse

RF pulse

short duration, high-frequency electromagnetic wave in pulse form

[SOURCE: JIS K 3611]

2.8

repetition time

TR

time interval for repetition of the basic unit of magnetic resonance pulse sequences

[SOURCE: JIS K 3611]

2.9

slice thickness

thickness of the imaging plane

2.10

number of averages

NA

number of times an identical MR signal is repeated

2.11

b value

maximum value of the parameter indicating level of diffusion weighting

2.12

motion probing gradient

MPG

gradient field applied to detect diffusion

2.13

parallel imaging

high-speed imaging method making use of the difference in sensitivities provided by multiple coils

2.14

fractional anisotropy

FA

number indicating level of structural anisotropy

2.15

mean diffusivity

MD

mean of diffusion coefficients along the three primary axes of a diffusion tensor

2.16**signal-to-noise ratio****SNR**

value expressing the proportion of signal to noise; greater values indicate higher image quality

2.17**voxel**

three-dimensional cuboid representing the minimum unit comprising a three-dimensional image

3 Principle

In the regeneration medicine for artificial cartilage, it is important to evaluate whether the implanted tissues regenerate as an artificial cartilage with time. However, an arthroscope is invasive and only monitors the surface texture of articular cartilage. X-ray and CT cannot project the cartilage tissue and have an exposure problem.

An articular cartilage has anisotropy by differential orientation of collagen fibres to exert a lubrication property as a joint. In the superficial layer, collagen fibres are oriented parallel to the joint surface. Next, in the middle layer, collagen fibres are randomly distributed for loading and oriented vertically in the deep layer. Such three-layer structure is a feature of articular cartilage, which is based on the biomechanical property of articular cartilage. Thus, it is possible to evaluate whether or not the articular cartilage by observation of the anisotropy structure.

In DT-MRI, in MRI techniques, it is possible to know the direction of proton movement. In fibrotic tissues, as the direction of water molecule movement is limited along the orientation of collagen fibres, the fibre direction is consistent with the direction of proton movement. Therefore, the direction of collagen structure can be evaluated with DT-MRI.

In this draft, DT-MRI data obtained from healthy males by using several MR devices shows for use as reference data to evaluate the process after regenerative treatment of articular cartilage.

4 Diffusion tensor magnetic resonance imaging (DT-MRI) data observation in articular cartilage**4.1 DT-MRI measurement process**

The process shown in the flowchart in [Figure 1](#) is used for acquisition, measurement, and observation of data to evaluate articular cartilage structure. This Technical Report envisions that different models of MRI equipment are used in different hospital facilities, and thus observational data are shown for three types of MRI apparatus produced by different manufacturers. [Table 1](#) presents the MRI apparatus, signal-receiving coils, and imaging parameters used in observation.




Figure 1 — DT-MRI data measurement process

Table 1 — MRI apparatus, signal-receiving coils, and imaging parameters used in DT-MRI

Imaging facility	Kyoto University Institute for Frontier Medical Sciences	Nihon University School of Dentistry at Matsudo Hospital	Hiroshima University Hospital
Apparatus (Manufacturer)	SONATA 1,5T (Siemens)	Achieva 1,5T (Philips)	Signa Excite 3T (General Electric)
Signal-receiving coil	4 ch flex array	8 ch sense knee	Lower extremity
Sequence	SE-EPI	SE-EPI	SE-EPI
FOV [mm]	192 × 192	150 × 150	128 × 128
Matrix	192 × 192	144 × 142	128 × 128
TR [ms]	2 200	2 200	2 200
TE [ms]	70	68	68
Slice thickness [mm]	3	5	5
Number of averages	24	20	12
b-value	600	600	400
No. MPG axes	6	6	6,15
Parallel imaging	GRAPPA ^a	SENSE ^b	n/a
Image slice	Sagittal plane	Sagittal plane	Sagittal plane
Pixels	384 × 384	400 × 400	256 × 256

^a Generalized rapid acquisition with partially parallel acquisition.
^b Sensitivity encoding.

4.2 Notes on setting of DT-MRI imaging parameters for articular cartilage

4.2.1 Imaging resolution

The resolution at which MR signals are acquired corresponds to a voxel size determined by matrix and slice thickness. Larger voxel sizes correspond to higher SNR, which also increases data reliability. Conversely, smaller voxel sizes increase resolution but decreased SNR of MR signals, which decreases data reliability.

Because articular cartilage has a three-layer structure with differing collagen fibre orientations (see Reference [2]), the matrix and slice thickness parameters used for this Technical Report were selected to allow acquisition of three-layer data.

4.2.2 Repetition time

Longer TR increases the number of slices in an image but lengthens imaging time. The parameter in this Technical Report was selected based on Reference [1].

4.2.3 Number of averages

A greater number of signal averages raises SNR and also increases data reliability but lengthens imaging time. The parameter in this Technical Report was selected based on Reference [1].

4.2.4 b-value

Higher b-values lengthen TE, which might lead to under-representation of diffusivity. Conversely, lower b-values can lead to over representation of diffusivity, and this parameter shall be set with consideration for the imaging target. The parameter in this Technical Report was selected based on Reference [1].

4.2.5 Motion probing gradient (MPG)

MPG shall be applied in a minimum of six directions to determine diffusion tensors. As MPG is applied to more axes, more complex structures can be analysed, but imaging time lengthens. But for application of MPG to equal numbers of axes, a higher static field intensity of the MRI apparatus will shorten imaging time. In consideration of the imaging times envisioned in hospital use, this Technical Report compares measurement data for imaging performed in six directions with an apparatus having a 1,5 T static field intensity, and in six directions and 15 directions with an apparatus having a 3 T static field intensity.

4.2.6 Parallel imaging

Imaging time can be shortened by application of parallel imaging. Signal-receiving coil is required for application of parallel imaging.

4.3 Measurement indices for structural evaluation of articular cartilage by DT-MRI

4.3.1 DT-MRI imaging slices and data measurement sites

A measurement slice is selected for acquisition of DT-MRI imaging data on articular cartilage. In a coronal slice of the knee joint as shown in Figure 2, DT-MRI imaging is performed in the sagittal plane at the location most inferior to the lateral condyle of the femur, and the image data acquired are used to create anisotropy distribution image and a mean diffusivity distribution graph, and fractional anisotropy and mean diffusivity are measured. The measured data are graphed, with the horizontal axis representing depth from the cartilage surface.

But as shown in Figure 3, depth from the cartilage surface is a relative measure relating the pixel centre where data are measured to the thickness of the cartilage concerned. [See the following Formula (1)]

$$ND = \frac{D}{T} \times 100\% \quad (1)$$

where

ND is the depth from surface (%);

D is the distance to pixel centre of data measurement (mm);

T is the cartilage thickness (mm).




Figure 2 — DT-MRI imaging slice and data measurement site




Figure 3 — Pixel intensity measurement site and depth from cartilage surface

NOTE Enlargement of measurement site at right in Figure 2.

4.3.2 Measurement of fractional anisotropy

4.3.2.1 Purpose of measurement

Fractional anisotropy is measured to evaluate the strength of structural anisotropy corresponding to collagen fibre orientation in articular cartilage.

4.3.2.2 Measurement method

The knee joint is imaged by DT-MRI, an anisotropy distribution image of the imaging slice is created from the image data acquired, and the fractional anisotropy of the articular cartilage depicted is measured. The value of fractional anisotropy is a measurement of the value at three pixels parallel to the femoral aspect of the cartilage surface in the fractional anisotropy distribution diagram, and five pixels

in the direction of depth. The fractional anisotropy distribution image is a pixel luminance distribution diagram displaying pixel luminance in 256 gradations of shading from 0 to 255, and measured pixel luminance is thus converted to fractional anisotropy by Formula (2). [Figure 4](#) shows an example of a fractional anisotropy distribution image.

$$FA = \frac{I}{256} \quad (2)$$

where

FA is the fractional anisotropy;

I is the pixel luminance

A graph is produced with fractional anisotropy on the vertical axis and depth from cartilage surface on the horizontal axis.




Figure 4 — Example of anisotropy distribution image

Fractional anisotropy (FA) is an index showing the strength of anisotropy and is represented as a value from 0 to 1, with stronger anisotropy approaching one, and weaker anisotropy tending to isotropy approaching 0. In the fractional anisotropy distribution image shown in [Figure 4](#), stronger anisotropy appears as white colouring with high pixel luminance, and weaker anisotropy appears dark, with low pixel luminance. In fibrous tissue, greater fibre alignment in one orientation strengthens anisotropy, which approaches 1, and more random fibre orientation leads to isotropy and weak anisotropy approaching zero.

The relationship between collagen fibre orientation and fractional anisotropy in articular cartilage has been ascertained and proven in basic investigations using knee joint cartilage from miniature pigs.

4.3.3 Measurement of mean diffusivity

4.3.3.1 Purpose of measurement

Mean diffusivity is measured to ascertain the condition of tissue based on the extent of moisture diffusion in articular cartilage tissue.

4.3.3.2 Measurement method

The knee joint is imaged by DT-MRI, a mean diffusivity distribution image is created from the image data acquired, and the mean diffusivity of the articular cartilage depicted is measured. As in the case of fractional anisotropy, three pixels parallel to the surfacial portion of femoral articular cartilage, and five pixels in the direction of depth are measured. The mean diffusivity distribution image is a pixel luminance distribution diagram displaying pixel luminance in 256 gradations of shading from 0 to 255, and measured pixel luminance is thus converted to mean diffusivity by Formula (3). [Figure 5](#) shows an example of a mean diffusivity distribution image.

$$MD = \frac{\frac{I}{255} \times C}{3} \times 1000 \quad (3)$$

where

MD is the mean diffusivity ($\times 10^{-3} \text{ mm}^2/\text{s}$);

I is the pixel luminance;

C is the conversion factor.

A graph is produced with mean diffusivity on the vertical axis, and depth from cartilage surface on the horizontal axis.




Figure 5 — Example of mean diffusivity distribution image

Mean diffusivity (MD) is the mean value of diffusivity in the three directions of the primary axes of diffusion tensors and expresses the quantitative magnitude of diffusion. Because MD expresses the phenomenon of diffusion quantitatively, it is used to investigate tissue properties. In the mean diffusivity distribution image shown in [Figure 5](#), greater mean diffusivity appears as white colouring with high pixel luminance, and lower mean diffusivity appears dark, with low pixel luminance.

STANDARDSISO.COM : Click to view the full PDF of ISO/TR 16379:2014

Annex A (informative)

Measurement Results

A.1 Measurement subjects

In this data acquisition, six healthy adult male subjects with no prior history of knee-joint disease underwent DT-MRI of the lateral condyles of the knee joint, and data pertaining to the femoral articular cartilage were measured. [Table A.1](#) presents the height, body weight, and age in years of subjects at the time of imaging.

Table A.1 — Height, body weight, and age in years of measured subjects

Subject	Age (years)	Height (cm)	Body weight (kg)
A	35	178	64
B	23	160	62
C	24	170	63
D	30	166	58
E	25	173	70
F	45	172	78
Mean	31	170	65

NOTE 1 Data on Subject A were acquired at all facilities.

NOTE 2 Data on Subjects B and C were acquired only at the Kyoto University Institute for Frontier Medical Sciences.

NOTE 3 Data on Subjects D and E were acquired only at the Nihon University School of Dentistry at Matsudo Hospital.

NOTE 4 Data on Subject F were acquired only at Hiroshima University Hospital.

A.2 Fractional anisotropy


A.2.1 SONATA 1,5 T (Kyoto University Institute for Frontier Medical Sciences)

Data for three subjects were acquired with a SONATA 1,5 T apparatus (Siemens) at the Kyoto University Institute for Frontier Medical Sciences. The acquired data in [Table A.2](#) are shown as a comparison graph in [Figure A.1](#).

The measurement results showed that fractional anisotropy was lowest in the middle layer, where collagen fibres are oriented isotropically, and higher in the surface and deep layers, where collagen fibres are oriented in regular directions. These results reflect patterns similar to those reported in the References [1], [4], and [5].

Table A.2 — Data acquired by SONATA 1,5 T apparatus

Item		Depth from surface (%)				
		10	30	50	70	90
Subject A	Measurement	0,25	0,13	0,07	0,20	0,37
		0,24	0,09	0,10	0,21	0,34
		0,24	0,12	0,10	0,14	0,24
	Mean	0,24	0,11	0,09	0,18	0,32
	Standard deviation	0,01	0,02	0,02	0,04	0,07
Subject B	Measurement	0,15	0,11	0,08	0,10	0,13
		0,14	0,07	0,07	0,16	0,20
		0,12	0,05	0,10	0,19	0,26
	Mean	0,13	0,08	0,08	0,15	0,19
	Standard deviation	0,02	0,03	0,01	0,05	0,06
Subject C	Measurement	0,36	0,15	0,15	0,12	0,34
		0,28	0,20	0,11	0,12	0,19
		0,17	0,09	0,09	0,07	0,12
	Mean	0,27	0,15	0,12	0,10	0,21
	Standard deviation	0,09	0,06	0,03	0,03	0,11

**Figure A.1 — Comparison graph of SONATA 1,5 T data, by subject****A.2.2 Achieva 1,5 T (Nihon University School of Dentistry at Matsudo Hospital)**

Data for three subjects were acquired with an Achieva 1,5 T (Philips) apparatus at the Nihon University School of Dentistry at Matsudo Hospital. The acquired data in [Table A.3](#) are shown as a comparison graph in [Figure A.2](#).

The measurement results showed changes in fractional anisotropy when moving from the surface layer to the deep layer of articular cartilage, just as in the case of measurement results from the Kyoto University Institute for Frontier Medical Sciences.

Table A.3 — Data acquired by Achieva 1,5 T apparatus

Item		Depth from surface (%)				
		10	30	50	70	90
Subject A	Measurement	0,32	0,18	0,08	0,18	0,32
		0,26	0,14	0,07	0,18	0,37
		0,22	0,11	0,10	0,21	0,41
	Mean	0,27	0,14	0,08	0,19	0,37
	Standard deviation	0,05	0,04	0,02	0,02	0,05
Subject D	Measurement	0,27	0,19	0,15	0,16	0,20
		0,18	0,15	0,13	0,18	0,25
		0,15	0,12	0,15	0,18	0,26
	Mean	0,20	0,15	0,14	0,17	0,24
	Standard deviation	0,06	0,04	0,01	0,01	0,03
Subject E	Measurement	0,24	0,13	0,06	0,11	0,25
		0,22	0,15	0,09	0,10	0,26
		0,21	0,15	0,11	0,10	0,29
	Mean	0,22	0,14	0,09	0,10	0,27
	Standard deviation	0,01	0,01	0,03	0,00	0,02




Figure A.2 — Comparison graph of Achieva 1,5 T data, by subject

A.2.3 Signa Excite 3 T (Hiroshima University Hospital)

Data for two subjects were acquired, in each case, for six directions and 15 directions of MPG using a Signa Excite 3 T (General Electric) apparatus at the Hiroshima University Hospital. [Table A.4](#) and [Figure A.3](#) present acquired data and a comparison graph for six axes, and [Table A.5](#) and [Figure A.4](#) present the same for 15 axes. [Figure A.5](#) also presents graphs comparing the number of MPG in six directions and 15 directions respectively.

The measurement results showed a pattern of changes in fractional anisotropy in articular cartilage. Results from *t*-testing of measurement data for these subjects also demonstrated no significant differences in fractional anisotropy according to number of MPG axes.

Table A.4 — Data for 6 MPG axes acquired by Signa Excite 3 T apparatus

Item		Depth from surface (%)				
		10	30	50	70	90
Subject A	Measurement	0,09	0,05	0,04	0,08	0,18
		0,11	0,07	0,08	0,14	0,28
		0,12	0,08	0,13	0,24	0,46
	Mean	0,11	0,07	0,08	0,15	0,31
	Standard deviation	0,01	0,02	0,04	0,08	0,14
Subject F	Measurement	0,13	0,06	0,05	0,11	0,20
		0,09	0,07	0,08	0,12	0,24
		0,12	0,10	0,08	0,13	0,27
	Mean	0,11	0,08	0,07	0,12	0,24
	Standard deviation	0,02	0,02	0,02	0,01	0,03





Figure A.3 — Comparison graph for 6 MPG axes by subject, acquired by Signa Excite 3 T apparatus

Table A.5 — Data for 15 MPG axes acquired by Signa Excite 3 T apparatus

Item		Depth from surface (%)				
		10	30	50	70	90
Subject A	Measurement	0,20	0,07	0,09	0,13	0,23
		0,16	0,10	0,11	0,16	0,33
		0,17	0,12	0,13	0,19	0,50
	Mean	0,18	0,10	0,11	0,16	0,35
	Standard deviation	0,03	0,02	0,02	0,03	0,14
Subject F	Measurement	0,12	0,10	0,09	0,11	0,17
		0,12	0,09	0,08	0,10	0,29
		0,10	0,09	0,09	0,15	0,34
	Mean	0,11	0,09	0,09	0,12	0,27
	Standard deviation	0,01	0,01	0,01	0,02	0,09

**Figure A.4 — Comparison graph for 15 MPG axes by subject, acquired by Signa Excite 3 T apparatus**




Figure A.5 — Comparison by difference in number of MPG axes

A.2.4 Comparison of same-subject data, by apparatus

[Table A.6](#) presents a graph comparing different facility data for Subject A in each case

Apart from results at 10 % of depth from the surface, results from *t*-testing of same-subject measurement data demonstrated no significant differences according to MRI apparatus.




Figure A.6 — Comparison graph of fractional anisotropy in Subject A, by apparatus

A.3 Mean diffusivity

A.3.1 SONATA 1,5 T (Kyoto University Institute for Frontier Medical Sciences)

Table A.6 presents data for three subjects acquired with a SONATA 1,5 T apparatus (Siemens) at the Kyoto University Institute for Frontier Medical Sciences, and Figure A.7 presents a comparison graph of the same.

For data moving from the surface layer to the deep layer, the measurement results showed patterns similar to those reported in the References [1], [4], and [5], and no significant differences were observed in results for *t*-testing of measurement data among subjects.

Table A.6 — Data acquired by SONATA 1,5 T apparatus

Item		Depth from surface (%)				
		10	30	50	70	90
Subject A	Measurement ($\times 10^{-3}\text{mm}^2/\text{s}$)	1,08	1,32	1,46	1,35	0,87
		1,42	1,67	1,78	1,54	0,94
		1,61	1,91	1,99	1,75	1,15
	Mean	1,37	1,63	1,74	1,54	0,99
	Standard deviation	0,27	0,30	0,26	0,20	0,15
Subject B	Measurement ($\times 10^{-3}\text{mm}^2/\text{s}$)	1,52	1,51	1,49	1,42	1,24
		1,59	1,58	1,50	1,40	1,22
		1,67	1,63	1,51	1,40	1,24
	Mean	1,59	1,57	1,50	1,41	1,23
	Standard deviation	0,07	0,06	0,01	0,01	0,01
Subject C	Measurement ($\times 10^{-3}\text{mm}^2/\text{s}$)	1,31	1,58	1,55	1,30	0,88
		1,19	1,55	1,54	1,33	0,98
		1,02	1,46	1,49	1,33	1,04
	Mean	1,17	1,53	1,53	1,32	0,97
	Standard deviation	0,14	0,07	0,03	0,02	0,08


**Figure A.7 — Comparison graph of SONATA 1,5 T data, by subject****A.3.2 Achieva 1,5 T (Nihon University School of Dentistry at Matsudo Hospital)**

Table A.7 presents data for three subjects acquired with an Achieva 1,5 T (Philips) apparatus at the Nihon University School of Dentistry at Matsudo Hospital, and Figure A.8 presents a comparison graph of the same.

For data moving from the surface layer to the deep layer, the measurement results showed patterns similar to those reported in the References [1], [4], and [5], and no significant differences were observed in results for *t*-testing of measurement data among subjects.

Table A.7 — Data acquired by Achieva 1,5 T apparatus

Item		Depth from surface (%)				
		10	30	50	70	90
Subject A	Measurement ($\times 10^{-3}\text{mm}^2/\text{s}$)	1,72	1,78	1,82	1,76	1,52
		1,62	1,67	1,72	1,64	1,31
		1,53	1,56	1,61	1,49	0,97
	Mean	1,62	1,67	1,72	1,63	1,27
	Standard deviation	0,09	0,11	0,11	0,14	0,28
Subject D	Measurement ($\times 10^{-3}\text{mm}^2/\text{s}$)	1,92	2,00	1,94	1,73	1,55
		1,98	1,96	1,87	1,71	1,53
		2,05	1,92	1,80	1,64	1,51
	Mean	1,98	1,96	1,87	1,69	1,53
	Standard deviation	0,06	0,04	0,07	0,05	0,02
Subject E	Measurement ($\times 10^{-3}\text{mm}^2/\text{s}$)	1,81	1,94	1,89	1,73	1,49
		1,67	1,81	1,77	1,59	1,27
		1,53	1,70	1,67	1,49	1,06
	Mean	1,67	1,82	1,78	1,61	1,27
	Standard deviation	0,14	0,12	0,11	0,12	0,22




Figure A.8 — Comparison graph of Achieva 1,5 T data, by subject

A.3.3 Signa Excite 3 T (Hiroshima University Hospital)

Data for two subjects were acquired, in each case, for six directions and 15 directions of MPG using a Signa Excite 3 T (General Electric) apparatus at the Hiroshima University Hospital. [Table A.8](#) and [Figure A.9](#) present acquired data and a comparison graph for six axes, and [Table A.9](#) and [Figure A.10](#) present the same for 15 axes. [Figure A.11](#) also presents graphs comparing number of MPG for six axes and 15 axes respectively.