
**Hydraulic fluid power — Impact and
use of ISO 11171:2016 $\mu\text{m}(\text{b})$ and
 $\mu\text{m}(\text{c})$ particle size designations on
particle count and filter test data**

STANDARDSISO.COM : Click to view the full PDF of ISO/TR 22681:2019



STANDARDSISO.COM : Click to view the full PDF of ISO/TR 22681:2019



COPYRIGHT PROTECTED DOCUMENT

© ISO 2019

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword.....	iv
Introduction.....	v
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	1
4 Origins of the particle size shift.....	1
5 Implications of the particle size shift.....	3
6 Use of revised ISO 11171.....	6
Bibliography.....	8

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 131, *Fluid power systems*, Subcommittee SC 6, *Contamination control*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

A minor revision to ISO 11171:2010 was approved during 2016. The revision was necessitated by a particle size shift resulting from the replacement of the particle counter calibration material, National Institute of Standards and Technology (NIST) SRM 2806a, with SRM 2806b. Prior to this revision, particle count data based upon ISO 11171 was reported in units of $\mu\text{m}(\text{c})$. Following the introduction of SRM 2806b, users of particle count data in the absence of this revision could not discern whether particle sizes being reported were based upon SRM 2806a or SRM 2806b. Hypothetically, a particle size reported as 20 $\mu\text{m}(\text{c})$ could actually be as large as 20 μm or as small as 18 μm depending upon whether SRM 2806a or SRM 2806b were used. This approximately 10 % shift in particle size can become significant in terms of the actual numbers of particles counted.

To minimize confusion and provide for clear communication, ISO 11171:2010, 6.8 and 7.1 were revised to provide a means for reporting particle size that clearly identifies the basis for reported particle size and provides the industry with tools to relate past SRM 2806a and new SRM 2806b data without extensive revisions to existing standards, specifications, and other literature. It provides a historically consistent, traceable definition of $\mu\text{m}(\text{c})$, while allowing an option to report sizes in terms of a defined $\mu\text{m}(\text{b})$ as needed. This document summarizes the underlying reasons for the minor revision and its practical impact on the industry.

STANDARDSISO.COM : Click to view the full PDF of ISO/TR 22681:2019

Hydraulic fluid power — Impact and use of ISO 11171:2016 $\mu\text{m}(\text{b})$ and $\mu\text{m}(\text{c})$ particle size designations on particle count and filter test data

1 Scope

This document explains the use of the two acceptable methods of reporting particle size, $\mu\text{m}(\text{c})$ and $\mu\text{m}(\text{b})$, that are defined in ISO 11171:2016. It also explains the reasons for the existence of two alternative size reporting methods and its implications with respect to particle count and filter Beta Ratio data.

2 Normative references

There are no normative references in this document.

3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

4 Origins of the particle size shift

ISO 11171 provides direct traceability to the standard international (SI) definition of a metre through NIST SRM 2806. ISO 11171 has been widely used to calibrate automatic particle counters (APCs) for hydraulic, lube oil, diesel fuel, and other non-aqueous liquid applications since 1999. NIST SRM 2806 is the primary calibration material used in this document.

During 2010, the original supply of SRM 2806, also sold as SRM 2806a, was exhausted. For ease of communication, the term “SRM 2806a” is used henceforth in this document to refer to both designations, SRM 2806 and SRM 2806a, of the original calibration material.

During 2014, its replacement, SRM 2806b, was released. The SRM 2806b production and certification process was overseen by an international group of experts from ISO/TC 131/SC 6. The specifications for SRM 2806b were better defined than those for SRM 2806a and it was produced by a different supplier. Advances in sample preparation and metrology were used to produce and certify SRM 2806b. A critical difference between SRM 2806a and SRM 2806b is that particle sizing was done manually from SEM micrographs with SRM 2806a, while automated image analysis was used with SRM 2806b.

This allowed an order of magnitude of more particles to be analysed and in a manner not subject to human bias. The end result was a certified calibration material, SRM 2806b, with better precision in the size distribution and a reduction in uncertainty compared to its predecessor, SRM 2806a.

An international round robin of ISO 11171 found that particle size, as defined using SRM 2806b compared to SRM 2806a, had shifted. The relationship between SRM 2806a and SRM 2806b sizes was determined from data submitted by 15 participating labs and is given by the following formula:

$$d_c = 0,898 d_b$$

where

d_c is the particle size in $\mu\text{m(c)}$ obtained using an APC calibrated with SRM 2806 and SRM 2806a;

d_b is the particle size in $\mu\text{m(b)}$ obtained using an APC calibrated with SRM 2806b.

As a result of the minor revision to ISO 11171, particle size reported in $\mu\text{m(c)}$ is traceable to SRM 2806a using the conversion equation, or by direct measurement using an APC calibrated with SRM 2806a or SRM 2806a-traceable secondary samples in the case of data generated prior to 2016. Particle sizes traceable to SRM 2806b are reported in units of $\mu\text{m(b)}$. Users of particle count data have the option to report particle size in units of either $\mu\text{m(c)}$ or $\mu\text{m(b)}$, or both, as appropriate. This particle size shift is within the published uncertainty of SRM 2806a, but it is great enough to affect particle count and filter test data. The magnitude of the particle size shift is illustrated graphically in [Figure 1](#).

In [Figure 1](#), particle size in $\mu\text{m(c)}$ as determined using ISO 11171 and SRM 2806a is plotted on the X-axis. The corresponding particle size as determined using other means, i.e. using SRM 2806b or AC Fine Test Dust, is plotted on the Y-axis. The **bold** line shows the relationship between $\mu\text{m(c)}$ and $\mu\text{m(b)}$. The relationship passes through the origin with a slope of 1,114.

For reference, the thin line shows the relationship between $\mu\text{m(c)}$ and μm sizes determined using the obsolete ISO 4402¹⁾ calibration method and AC Fine Test Dust. This relationship does not pass through the origin and the slope is steeper. [Figure 1](#) demonstrates that the particle size shift with SRM 2806b is significant, yet less than what was observed when the industry transitioned from ISO 4402 to ISO 11171.

1) Withdrawn standard.




Figure 1 — Relationship between particle sizes defined using SRM 2806a and those obtained using SRM 2806b, and using AC Fine Test Dust

5 Implications of the particle size shift

Prior to the revision in ISO 11171, there was no convenient method for users of particle count data to determine whether or not their data was affected by the particle size shift. Users could not tell whether particle count results were based upon SRM 2806a or SRM 2806b calibrations. As a result of the revision, in most cases particle count and size data reported in units of μm (c) can be used in the same manner that was done prior to the release of SRM 2806b. In those cases where μm (b) is used or compared to μm (c) data, the direction and magnitude of the correction are clearly defined and identified. The implications of the particle size shift can be discussed in terms of its impact upon particle size, particle count, ISO Code, and filter Beta Ratio results.

The impact upon particle size is illustrated in [Figure 1](#), as previously discussed. For the particle size range from 0 to 38 μm (b), the corresponding μm (c) sizes are 10,2 % smaller. The μm (c) sizes are obtained by multiplying μm (b) sizes by 0,898. At sizes larger than 38 μm (b), latex is used for calibration and the particle size does not require correction.

The magnitude of the impact of the size shift on observed particle count data depends upon the particle size distribution of the sample in question as shown in [Figure 2](#).

[Figure 2](#) shows particle count data for two different dusts, ISO MTD (thin lines) and ISO UFTD (**bold lines**). ISO MTD is used in ISO 16889 filter testing. ISO UFTD is a finer dust with a steeper particle size distribution such as might be found downstream of a high performance filter. For illustrative purposes, the gravimetric dust concentration was chosen arbitrarily. The results for each dust are plotted in two different ways. The dashed lines show the results reported in μm (c) while the solid lines show results reported in μm (b). For any given particle size, μm (b) particle counts tend to be greater than the corresponding μm (c) results. The magnitude of the increase becomes greater as the slope of the particle size distribution becomes steeper (more negative) as per ISO UFTD example.

The revision to ISO 11171 clearly informs users of the underlying basis of particle count data, SRM 2806a or SRM 2806b, and enables accurate interpretations of results to be made. Furthermore, the revision provides for mathematical conversion of particle diameters derived from SRM 2806b calibration, $\mu\text{m}(\text{b})$, to those obtained from SRM 2806a calibration, $\mu\text{m}(\text{c})$. As a result, existing specifications and standards based upon $\mu\text{m}(\text{c})$ can be used without modification, other than to reference the most recent version of ISO 11171.





Figure 2 — Impact on revision on particle count results for two different dust size distributions

The magnitude of the impact of the particle size shift on cleanliness levels reported as ISO 4406 Code numbers depends upon the actual particle size distribution of a sample and how near the particle concentration is to cross-over points between ISO Codes. [Figures 3 a\)](#) and [3 b\)](#), and [Table 1](#) are used to illustrate this. [Figures 3 a\)](#) and [3 b\)](#) graphically illustrate the particle size distribution of arbitrary concentrations of ISO MTD and ISO UFTD, and of a sample taken downstream of a filter ([Figure 3 a\)](#) shows the data reported in $\mu\text{m}(\text{c})$ and [Figure 3 b\)](#) shows data from the same sample reported in $\mu\text{m}(\text{b})$).

[Table 1](#) shows this data reported as ISO 4406 codes rather than particle counts. Using the revised ISO 11171, observed ISO 4406 codes do not change provided that $\mu\text{m}(\text{c})$ particle sizes are used as shown in the second column of [Table 1](#). On the other, ISO 4406 cannot be directly used with $\mu\text{m}(\text{b})$ sizes as the code requires $\mu\text{m}(\text{c})$ sizes. Sizes of 4,0 $\mu\text{m}(\text{b})$, 6,0 $\mu\text{m}(\text{b})$ and 14,0 $\mu\text{m}(\text{b})$ correspond to sizes of 3,6 $\mu\text{m}(\text{c})$, 5,4 $\mu\text{m}(\text{c})$ and 14,0 $\mu\text{m}(\text{c})$. If particle sizes of 4,0 $\mu\text{m}(\text{b})$, 6,0 $\mu\text{m}(\text{b})$ and 14,0 $\mu\text{m}(\text{b})$ are erroneously used with the ISO 4406 code, the examples in [Table 1](#) (column 3) show that ISO codes typically increase 1 or 2 code units, depending upon the particle concentration and size distribution.

The following magnitude of the change in ISO code data is dependent upon the particle size distribution. The steeper the particle size distribution, the greater for the change. For this reason, a greater shift could occur with certain field samples such as those collected downstream of fine filters.

In practical terms, such an error could result in systems failing to meet cleanliness recommendations simply due to a failure to use the proper particle size reference, i.e. incorrect use of $\mu\text{m}(\text{b})$ instead of the correct $\mu\text{m}(\text{c})$. This could result in unnecessary service or corrective action being taken to solve problems that don't exist. Prior to the revision, users had no convenient way of determining the basis for calibration, putting them at risk of making inappropriate contamination control decisions.



Key

Y particles/mL > indicated size
 X particle size, $\mu\text{m(c)}$ S
 Z ISO 4406 code
 1 ISO UFTD
 2 ISO MTD
 3 Down stream

Y particles/mL > indicated size
 X particle size, $\mu\text{m(b)}$
 Z ISO 4406 code
 1 ISO UFTD
 2 ISO MTD
 3 Downstream

Figure 3 — Effect of calibration material on ISO code for various types of samples

NOTE [Figure 3 a\)](#) shows the data reported in $\mu\text{m(c)}$ and [Figure 3 b\)](#) shows data from the same sample reported in $\mu\text{m(b)}$

Table 1 — Effect of calibration material upon reported ISO code^a

Sample description	ISO Code	
	$\mu\text{m(c)}$	$\mu\text{m(b)}$
MTD	18/17/13	19/18/14
UFTD	20/19/11	21/20/13
Downstream of filter	12/9/3	14/10/4


^a Refer to [Figure 3](#) for particle size distribution of each sample.

The particle size shift also impacts observed filter Beta Ratio data. The magnitude of the impact depends upon the characteristics of the filter in question. When particle size is reported in terms of $\mu\text{m(c)}$ according to the revised ISO 11171, Beta Ratio results are unaffected. New and historical data based upon ISO 11171 can be compared without conversion. On the other hand, direct comparison of $\mu\text{m(c)}$ and $\mu\text{m(b)}$ Beta Ratios is problematic. This is illustrated in [Figure 4](#) using results from three different types of filters, referred to as A, B and C.

Results for each filter are plotted using particle sizes as $\mu\text{m(c)}$ (thin lines) and as $\mu\text{m(b)}$ (**bold lines**) according to the revised ISO 11171. In general Beta Ratios appear smaller and filter performance appears worse using $\mu\text{m(b)}$ sizes because the corresponding $\mu\text{m(c)}$ sizes are numerically 10,2 % smaller. Since the magnitude of the decrease in Beta Ratio depends upon the characteristics of the filter, there is no generic Beta Ratio correction available. Prior to the revision users of Beta Ratio data had no clear indication of the cause of this decrease and would have assumed that filter performance had deteriorated even though no actual change in filter performance had occurred.

The financial implications of this are significant and include potential rejection of filter shipments based upon a perceived failure to meet the specification and extensive troubleshooting efforts to eliminate failure modes that do not exist. The revision provides a means to avoid such problem by

providing a clear means to report filter performance using historically consistent $\mu\text{m(c)}$ sizes. This allows filter customers and manufacturers to evaluate products in the context of historic performance. Thus, existing standards and specifications referring to particle size in $\mu\text{m(c)}$ do not require revision and historical data based upon ISO 11171 can be directly compared to $\mu\text{m(c)}$ data obtained using the revised ISO 11171.



Key	
Y	Beta Ratio
X	particle size
—	$\mu\text{m(b)}$
- - -	$\mu\text{m(c)}$

Figure 4 — Filter Beta Ratios expressed as function of $\mu\text{m(c)}$ and $\mu\text{m(b)}$

6 Use of revised ISO 11171

The particle size shift resulting from the introduction of new APC calibration material posed a major challenge for the industry requiring immediate action. In the absence of a standard means for expressing particle size, there was inconsistency in terms of how data was reported and erroneous contamination control decisions made. In response, a minor revision to ISO 11171 affecting the reporting of data was passed by ISO TC 131/SC 6.

Since the supply of SRM 2806a and the shelf life of SRM 2806a secondary calibration samples had expired, it was not possible to revert back to the previous SRM 2806a calibration material. Future calibration according to ISO 11171 must be made using SRM 2806b traceable material. The revision provides the industry with a mathematical method for converting $\mu\text{m(b)}$ sizes obtained from APCs calibrated with SRM 2806b to $\mu\text{m(c)}$ sizes that are traceable to SRM 2806a. There has been no technical change to ISO 11171. Only the reporting method has been revised. The revised ISO 11171 does not require particle size or count data to be expressed in any particular manner. It provides two mathematically related options for reporting sizes. The use $\mu\text{m(c)}$ or $\mu\text{m(b)}$ is at the discretion of user, as is how to implement their choice.

In most instances, organizations will find it least disruptive to continue reporting particle size in units of $\mu\text{m(c)}$. There is no inherent advantage in reporting in units of $\mu\text{m(b)}$, while standards, product and